

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3422

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170232BL		WATER		Batch Number:	PB170232		Prep Date:	10/23/2025	
	Mercury	0.076	<0.2	U	0.20	CV	10/24/2025	10:50	LB137632

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: AECOM

SDG No.: Q3422

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170225BL	WATER			Batch Number:	PB170225		Prep Date:	10/23/2025	
	Aluminum	5.67	<25	U	50.0	P	10/24/2025	15:20	LB137662
	Antimony	3.38	<12.5	U	25.0	P	10/24/2025	15:20	LB137662
	Arsenic	2.56	<5	U	10.0	P	10/24/2025	15:20	LB137662
	Barium	7.28	<25	U	50.0	P	10/24/2025	15:20	LB137662
	Beryllium	0.28	<1.5	U	3.00	P	10/24/2025	15:20	LB137662
	Cadmium	0.25	<1.5	U	3.00	P	10/24/2025	15:20	LB137662
	Calcium	117	<500	U	1000	P	10/24/2025	15:20	LB137662
	Chromium	1.06	<2.5	U	5.00	P	10/24/2025	15:20	LB137662
	Cobalt	1.13	<7.5	U	15.0	P	10/24/2025	15:20	LB137662
	Copper	2.30	<5	U	10.0	P	10/24/2025	15:20	LB137662
	Iron	11.7	<25	U	50.0	P	10/24/2025	15:20	LB137662
	Lead	1.15	<3	U	6.00	P	10/24/2025	15:20	LB137662
	Magnesium	122	<500	U	1000	P	10/24/2025	15:20	LB137662
	Manganese	2.97	<5	U	10.0	P	10/24/2025	15:20	LB137662
	Nickel	1.53	<10	U	20.0	P	10/24/2025	15:20	LB137662
	Potassium	459	<500	U	1000	P	10/24/2025	15:20	LB137662
	Selenium	4.82	<5	U	10.0	P	10/24/2025	15:20	LB137662
	Silver	0.81	<2.5	U	5.00	P	10/24/2025	15:20	LB137662
	Sodium	434	<500	U	1000	P	10/24/2025	15:20	LB137662
	Thallium	2.19	<10	U	20.0	P	10/24/2025	15:20	LB137662
	Vanadium	3.13	<10	U	20.0	P	10/24/2025	15:20	LB137662
	Zinc	8.33	<10	U	20.0	P	10/24/2025	15:20	LB137662