

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>AECOM</u>	level: <u>low</u>	sdg no.: <u>Q3422</u>
contract: <u>AECO02</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3420-05</u>	client id: <u>COMPOSITEMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3420-05MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	2560		1360		1000	120		P
Antimony	ug/L	75 - 125	507		25.0	U	400	127	N	P
Arsenic	ug/L	75 - 125	563		51.5		400	128	N	P
Barium	ug/L	75 - 125	111		50.0	U	100	111		P
Beryllium	ug/L	75 - 125	96.0		3.00	U	100	96		P
Cadmium	ug/L	75 - 125	108		3.00	U	100	108		P
Calcium	ug/L	75 - 125	1360		809	J	500	110		P
Chromium	ug/L	75 - 125	221		15.9		200	103		P
Cobalt	ug/L	75 - 125	108		15.0	U	100	108		P
Copper	ug/L	75 - 125	149		7.37	J	150	94		P
Iron	ug/L	75 - 125	2550		473		1500	138	N	P
Lead	ug/L	75 - 125	492		1.84	J	500	98		P
Magnesium	ug/L	75 - 125	1640		642	J	1000	100		P
Manganese	ug/L	75 - 125	112		10.1		100	102		P
Nickel	ug/L	75 - 125	282		15.0	J	250	107		P
Potassium	ug/L	75 - 125	30200		11400		5000	378	N	P
Selenium	ug/L	75 - 125	1880		492		1000	139	N	P
Silver	ug/L	75 - 125	55.1		5.00	U	37.5	147	N	P
Sodium	ug/L	75 - 125	5870000	D	6230000	D	1500	-24427		P
Thallium	ug/L	75 - 125	1020		20.0	U	1000	102		P
Vanadium	ug/L	75 - 125	161		20.0	U	150	107		P
Zinc	ug/L	75 - 125	413		266		100	146	N	P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>AECOM</u>	level: <u>low</u>	sdg no.: <u>Q3422</u>
contract: <u>AECO02</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3420-05</u>	client id: <u>COMPOSITEMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3420-05MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	2410		1360		1000	105		P
Antimony	ug/L	75 - 125	460		25.0	U	400	115		P
Arsenic	ug/L	75 - 125	513		51.5		400	115		P
Barium	ug/L	75 - 125	99.6		50.0	U	100	100		P
Beryllium	ug/L	75 - 125	81.9		3.00	U	100	82		P
Cadmium	ug/L	75 - 125	97.2		3.00	U	100	97		P
Calcium	ug/L	75 - 125	1300		809	J	500	97		P
Chromium	ug/L	75 - 125	200		15.9		200	92		P
Cobalt	ug/L	75 - 125	97.6		15.0	U	100	98		P
Copper	ug/L	75 - 125	135		7.37	J	150	85		P
Iron	ug/L	75 - 125	2420		473		1500	130	N	P
Lead	ug/L	75 - 125	441		1.84	J	500	88		P
Magnesium	ug/L	75 - 125	1500		642	J	1000	86		P
Manganese	ug/L	75 - 125	101		10.1		100	91		P
Nickel	ug/L	75 - 125	255		15.0	J	250	96		P
Potassium	ug/L	75 - 125	32500		11400		5000	422	N	P
Selenium	ug/L	75 - 125	1760		492		1000	127	N	P
Silver	ug/L	75 - 125	49.9		5.00	U	37.5	133	N	P
Sodium	ug/L	75 - 125	5840000	D	6230000	D	1500	-26205		P
Thallium	ug/L	75 - 125	931		20.0	U	1000	93		P
Vanadium	ug/L	75 - 125	144		20.0	U	150	96		P
Zinc	ug/L	75 - 125	401		266		100	134	N	P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: AECOM level: low sdg no.: Q3422
contract: AECO02 lab code: ACE
matrix: Water sample id: Q3434-02 client id: PR132-WC2-20251022MS
Percent Solids for Sample: NA Spiked ID: Q3434-02MS Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	4.65		1.03		4.0	91		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: AECOM **level:** low **sdg no.:** Q3422
contract: AECO02 **lab code:** ACE
matrix: Water **sample id:** Q3434-02 **client id:** PR132-WC2-20251022MSD
Percent Solids for Sample: NA **Spiked ID:** Q3434-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	4.62		1.03		4.0	90		CV