

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3422
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114632.D **Time Analyzed:** 15:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.493	5.392	5.592	588.730	500.000	17.7
Aroclor-1016-2	5.514	5.414	5.614	557.630	500.000	11.5
Aroclor-1016-3	5.576	5.476	5.676	557.680	500.000	11.5
Aroclor-1016-4	5.672	5.571	5.771	588.150	500.000	17.6
Aroclor-1016-5	5.963	5.863	6.063	584.150	500.000	16.8
Aroclor-1260-1	7.077	6.975	7.175	562.700	500.000	12.5
Aroclor-1260-2	7.331	7.229	7.429	545.370	500.000	9.1
Aroclor-1260-3	7.688	7.585	7.785	461.520	500.000	-7.7
Aroclor-1260-4	7.910	7.807	8.007	502.890	500.000	0.6
Aroclor-1260-5	8.217	8.113	8.313	477.970	500.000	-4.4
Decachlorobiphenyl	9.940	9.834	10.034	57.850	50.000	15.7
Tetrachloro-m-xylene	4.354	4.255	4.455	58.210	50.000	16.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3422
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114632.D **Time Analyzed:** 15:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.736	4.643	4.843	520.760	500.000	4.2
Aroclor-1016-2	4.756	4.662	4.862	519.830	500.000	4.0
Aroclor-1016-3	4.930	4.836	5.036	554.200	500.000	10.8
Aroclor-1016-4	4.972	4.879	5.079	558.310	500.000	11.7
Aroclor-1016-5	5.184	5.092	5.292	549.740	500.000	9.9
Aroclor-1260-1	6.216	6.125	6.325	587.290	500.000	17.5
Aroclor-1260-2	6.403	6.313	6.513	621.300	500.000	24.3
Aroclor-1260-3	6.556	6.466	6.666	559.760	500.000	12.0
Aroclor-1260-4	7.028	6.939	7.139	615.160	500.000	23.0
Aroclor-1260-5	7.268	7.179	7.379	593.860	500.000	18.8
Decachlorobiphenyl	8.655	8.571	8.771	49.330	50.000	-1.3
Tetrachloro-m-xylene	3.656	3.560	3.760	51.300	50.000	2.6