

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3422
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114646.D **Time Analyzed:** 20:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.492	5.392	5.592	585.330	500.000	17.1
Aroclor-1016-2	5.514	5.414	5.614	552.310	500.000	10.5
Aroclor-1016-3	5.575	5.476	5.676	552.620	500.000	10.5
Aroclor-1016-4	5.671	5.571	5.771	583.120	500.000	16.6
Aroclor-1016-5	5.962	5.863	6.063	588.600	500.000	17.7
Aroclor-1260-1	7.076	6.975	7.175	564.780	500.000	13.0
Aroclor-1260-2	7.330	7.229	7.429	541.640	500.000	8.3
Aroclor-1260-3	7.687	7.585	7.785	481.280	500.000	-3.7
Aroclor-1260-4	7.909	7.807	8.007	521.950	500.000	4.4
Aroclor-1260-5	8.215	8.113	8.313	486.820	500.000	-2.6
Decachlorobiphenyl	9.937	9.834	10.034	58.680	50.000	17.4
Tetrachloro-m-xylene	4.354	4.255	4.455	57.910	50.000	15.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3422
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114646.D **Time Analyzed:** 20:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.735	4.643	4.843	508.970	500.000	1.8
Aroclor-1016-2	4.754	4.662	4.862	506.890	500.000	1.4
Aroclor-1016-3	4.929	4.836	5.036	544.440	500.000	8.9
Aroclor-1016-4	4.971	4.879	5.079	552.230	500.000	10.4
Aroclor-1016-5	5.183	5.092	5.292	515.180	500.000	3.0
Aroclor-1260-1	6.215	6.125	6.325	567.310	500.000	13.5
Aroclor-1260-2	6.402	6.313	6.513	599.240	500.000	19.8
Aroclor-1260-3	6.555	6.466	6.666	541.940	500.000	8.4
Aroclor-1260-4	7.026	6.939	7.139	581.650	500.000	16.3
Aroclor-1260-5	7.267	7.179	7.379	577.660	500.000	15.5
Decachlorobiphenyl	8.654	8.571	8.771	52.290	50.000	4.6
Tetrachloro-m-xylene	3.656	3.560	3.760	50.920	50.000	1.8