

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3422
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071098.D **Time Analyzed:** 13:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	547.800	500.000	9.6
Aroclor-1016-2	4.498	4.399	4.599	544.520	500.000	8.9
Aroclor-1016-3	4.556	4.456	4.656	550.240	500.000	10.0
Aroclor-1016-4	4.646	4.546	4.746	556.420	500.000	11.3
Aroclor-1016-5	4.931	4.831	5.031	553.620	500.000	10.7
Aroclor-1260-1	6.020	5.919	6.119	549.350	500.000	9.9
Aroclor-1260-2	6.276	6.175	6.375	554.480	500.000	10.9
Aroclor-1260-3	6.627	6.526	6.726	559.460	500.000	11.9
Aroclor-1260-4	6.841	6.740	6.940	571.200	500.000	14.2
Aroclor-1260-5	7.146	7.045	7.245	565.790	500.000	13.2
Decachlorobiphenyl	8.498	8.396	8.596	56.680	50.000	13.4
Tetrachloro-m-xylene	3.399	3.300	3.500	54.440	50.000	8.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3422
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071098.D **Time Analyzed:** 13:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	453.960	500.000	-9.2
Aroclor-1016-2	3.792	3.693	3.893	452.590	500.000	-9.5
Aroclor-1016-3	3.951	3.853	4.053	464.930	500.000	-7.0
Aroclor-1016-4	4.000	3.902	4.102	469.520	500.000	-6.1
Aroclor-1016-5	4.195	4.097	4.297	470.610	500.000	-5.9
Aroclor-1260-1	5.188	5.089	5.289	476.540	500.000	-4.7
Aroclor-1260-2	5.377	5.278	5.478	468.050	500.000	-6.4
Aroclor-1260-3	5.518	5.419	5.619	457.860	500.000	-8.4
Aroclor-1260-4	5.978	5.879	6.079	473.740	500.000	-5.3
Aroclor-1260-5	6.222	6.124	6.324	469.640	500.000	-6.1
Decachlorobiphenyl	7.532	7.433	7.633	55.130	50.000	10.3
Tetrachloro-m-xylene	2.776	2.678	2.878	49.120	50.000	-1.8