

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3422
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071119.D **Time Analyzed:** 08:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	529.290	500.000	5.9
Aroclor-1016-2	4.499	4.399	4.599	527.760	500.000	5.6
Aroclor-1016-3	4.557	4.456	4.656	540.370	500.000	8.1
Aroclor-1016-4	4.647	4.546	4.746	539.790	500.000	8.0
Aroclor-1016-5	4.931	4.831	5.031	531.950	500.000	6.4
Aroclor-1260-1	6.021	5.919	6.119	517.090	500.000	3.4
Aroclor-1260-2	6.277	6.175	6.375	516.680	500.000	3.3
Aroclor-1260-3	6.628	6.526	6.726	542.990	500.000	8.6
Aroclor-1260-4	6.842	6.740	6.940	552.950	500.000	10.6
Aroclor-1260-5	7.147	7.045	7.245	543.020	500.000	8.6
Decachlorobiphenyl	8.499	8.396	8.596	54.590	50.000	9.2
Tetrachloro-m-xylene	3.400	3.300	3.500	52.510	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3422
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071119.D **Time Analyzed:** 08:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	423.330	500.000	-15.3
Aroclor-1016-2	3.791	3.693	3.893	423.970	500.000	-15.2
Aroclor-1016-3	3.952	3.853	4.053	438.320	500.000	-12.3
Aroclor-1016-4	4.000	3.902	4.102	444.030	500.000	-11.2
Aroclor-1016-5	4.195	4.097	4.297	436.450	500.000	-12.7
Aroclor-1260-1	5.187	5.089	5.289	433.170	500.000	-13.4
Aroclor-1260-2	5.376	5.278	5.478	431.050	500.000	-13.8
Aroclor-1260-3	5.517	5.419	5.619	433.100	500.000	-13.4
Aroclor-1260-4	5.977	5.879	6.079	438.310	500.000	-12.3
Aroclor-1260-5	6.222	6.124	6.324	424.780	500.000	-15.0
Decachlorobiphenyl	7.532	7.433	7.633	51.170	50.000	2.3
Tetrachloro-m-xylene	2.776	2.678	2.878	45.960	50.000	-8.1