

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3422

Instrument ID: ECD_D Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	32564400
		2	6.44	6.34	6.54	48621800
		3	7.14	7.04	7.24	85710600
		4	7.56	7.46	7.66	95737400
		5	7.92	7.82	8.02	60480500



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: AECO02
Lab Code: ACE SDG NO.: Q3422
Instrument ID: ECD_D Date(s) Analyzed: 10/17/2025 10/17/2025
GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.47	5.37	5.57	263702000
		2	5.64	5.54	5.74	176560000
		3	6.75	6.65	6.85	658883000
		4	7.19	7.09	7.29	495104000
		5	7.32	7.22	7.42	246033000