

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3422

Client ID : SSTDCCC040

Date Analyzed: 10/27/2025

Lab File ID: BG064583.D

Time Analyzed: 11:57

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	39737	8.22	173874	11.05	136456	14.84
UPPER LIMIT	79474	8.72	347748	11.546	272912	15.342
LOWER LIMIT	19868.5	7.72	86937	10.546	68228	14.342
EPA SAMPLE NO.						
01 PR132-S15-000008-20251021	39403	8.22	157676	11.04	121417	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.