

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064582.D
Instrument ID: BNA_G

Contract: AECO02
SDG NO.: Q3422
DFTPP Injection Date: 10/27/2025
DFTPP Injection Time: 11:16

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.2) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
365	Greater than 1% of mass 198	5.2
441	Present, but less than mass 443	81.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.7 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG064583.D	10/27/2025	11:57
PR132-S15-000008-20251021	Q3422-01	BG064597.D	10/27/2025	21:31