

Report of Analysis

Client: AECOM
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
Client Sample ID: VX1103WBSD01
Lab Sample ID: VX1103WBSD01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3422
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	16.7		1	0.22	1.00	ug/L	11/03/25 12:44	VX110325
74-87-3	Chloromethane	19.6		1	0.32	1.00	ug/L	11/03/25 12:44	VX110325
75-01-4	Vinyl Chloride	18.4		1	0.26	1.00	ug/L	11/03/25 12:44	VX110325
74-83-9	Bromomethane	20.8		1	1.40	5.00	ug/L	11/03/25 12:44	VX110325
75-00-3	Chloroethane	17.2		1	0.47	1.00	ug/L	11/03/25 12:44	VX110325
75-69-4	Trichlorofluoromethane	18.5		1	0.33	1.00	ug/L	11/03/25 12:44	VX110325
76-13-1	1,1,2-Trichlorotrifluoroethane	17.7		1	0.25	1.00	ug/L	11/03/25 12:44	VX110325
75-35-4	1,1-Dichloroethene	16.5		1	0.23	1.00	ug/L	11/03/25 12:44	VX110325
67-64-1	Acetone	99.0		1	1.50	5.00	ug/L	11/03/25 12:44	VX110325
75-15-0	Carbon Disulfide	16.5		1	0.21	1.00	ug/L	11/03/25 12:44	VX110325
1634-04-4	Methyl tert-butyl Ether	17.4		1	0.16	1.00	ug/L	11/03/25 12:44	VX110325
79-20-9	Methyl Acetate	16.0		1	0.27	1.00	ug/L	11/03/25 12:44	VX110325
75-09-2	Methylene Chloride	16.8		1	0.28	1.00	ug/L	11/03/25 12:44	VX110325
156-60-5	trans-1,2-Dichloroethene	16.7		1	0.23	1.00	ug/L	11/03/25 12:44	VX110325
75-34-3	1,1-Dichloroethane	16.8		1	0.23	1.00	ug/L	11/03/25 12:44	VX110325
110-82-7	Cyclohexane	19.5		1	1.50	5.00	ug/L	11/03/25 12:44	VX110325
78-93-3	2-Butanone	98.3		1	0.98	5.00	ug/L	11/03/25 12:44	VX110325
56-23-5	Carbon Tetrachloride	21.6		1	0.25	1.00	ug/L	11/03/25 12:44	VX110325
156-59-2	cis-1,2-Dichloroethene	19.3		1	0.19	1.00	ug/L	11/03/25 12:44	VX110325
74-97-5	Bromochloromethane	17.8		1	0.22	1.00	ug/L	11/03/25 12:44	VX110325
67-66-3	Chloroform	19.1		1	0.25	1.00	ug/L	11/03/25 12:44	VX110325
71-55-6	1,1,1-Trichloroethane	19.1		1	0.20	1.00	ug/L	11/03/25 12:44	VX110325
108-87-2	Methylcyclohexane	19.3		1	0.16	1.00	ug/L	11/03/25 12:44	VX110325
71-43-2	Benzene	22.8		1	0.15	1.00	ug/L	11/03/25 12:44	VX110325
107-06-2	1,2-Dichloroethane	21.1		1	0.22	1.00	ug/L	11/03/25 12:44	VX110325
79-01-6	Trichloroethene	19.1		1	0.090	1.00	ug/L	11/03/25 12:44	VX110325
78-87-5	1,2-Dichloropropane	19.5		1	0.20	1.00	ug/L	11/03/25 12:44	VX110325
75-27-4	Bromodichloromethane	21.7		1	0.22	1.00	ug/L	11/03/25 12:44	VX110325
108-10-1	4-Methyl-2-Pentanone	100		1	0.68	5.00	ug/L	11/03/25 12:44	VX110325
108-88-3	Toluene	19.7		1	0.14	1.00	ug/L	11/03/25 12:44	VX110325
10061-02-6	t-1,3-Dichloropropene	21.7		1	0.17	1.00	ug/L	11/03/25 12:44	VX110325
10061-01-5	cis-1,3-Dichloropropene	21.0		1	0.16	1.00	ug/L	11/03/25 12:44	VX110325
79-00-5	1,1,2-Trichloroethane	20.4		1	0.21	1.00	ug/L	11/03/25 12:44	VX110325
591-78-6	2-Hexanone	110		1	0.89	5.00	ug/L	11/03/25 12:44	VX110325
124-48-1	Dibromochloromethane	21.0		1	0.18	1.00	ug/L	11/03/25 12:44	VX110325
106-93-4	1,2-Dibromoethane	20.8		1	0.15	1.00	ug/L	11/03/25 12:44	VX110325
127-18-4	Tetrachloroethene	19.9		1	0.23	1.00	ug/L	11/03/25 12:44	VX110325
108-90-7	Chlorobenzene	20.0		1	0.12	1.00	ug/L	11/03/25 12:44	VX110325
100-41-4	Ethyl Benzene	20.3		1	0.13	1.00	ug/L	11/03/25 12:44	VX110325
179601-23-1	m/p-Xylenes	40.7		1	0.24	2.00	ug/L	11/03/25 12:44	VX110325
95-47-6	o-Xylene	19.8		1	0.12	1.00	ug/L	11/03/25 12:44	VX110325
100-42-5	Styrene	20.6		1	0.15	1.00	ug/L	11/03/25 12:44	VX110325
75-25-2	Bromoform	21.5		1	0.19	1.00	ug/L	11/03/25 12:44	VX110325

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	VX1103WBSD01	SDG No.:	Q3422
Lab Sample ID:	VX1103WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 mL	Test:	VOC-TCLVOA-10
	Level: LOW		
	Final Vol: 5000 uL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	17.5		1	0.12	1.00	ug/L	11/03/25 12:44	VX110325
79-34-5	1,1,2,2-Tetrachloroethane	17.6		1	0.26	1.00	ug/L	11/03/25 12:44	VX110325
541-73-1	1,3-Dichlorobenzene	19.7		1	0.16	1.00	ug/L	11/03/25 12:44	VX110325
106-46-7	1,4-Dichlorobenzene	19.9		1	0.19	1.00	ug/L	11/03/25 12:44	VX110325
95-50-1	1,2-Dichlorobenzene	20.2		1	0.16	1.00	ug/L	11/03/25 12:44	VX110325
96-12-8	1,2-Dibromo-3-Chloropropane	20.0		1	0.53	1.00	ug/L	11/03/25 12:44	VX110325
120-82-1	1,2,4-Trichlorobenzene	19.5		1	0.20	1.00	ug/L	11/03/25 12:44	VX110325
87-61-6	1,2,3-Trichlorobenzene	20.6		1	0.20	1.00	ug/L	11/03/25 12:44	VX110325
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	49.2			74 - 125	98%	SPK: 50		
1868-53-7	Dibromofluoromethane	59.5			75 - 124	119%	SPK: 50		
2037-26-5	Toluene-d8	50.7			86 - 113	101%	SPK: 50		
460-00-4	4-Bromofluorobenzene	50.5			77 - 121	101%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	167000							
540-36-3	1,4-Difluorobenzene	243000							
3114-55-4	Chlorobenzene-d5	214000							
3855-82-1	1,4-Dichlorobenzene-d4	125000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products