

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3422
 Lab File ID: VX048445.D Date Analyzed: 11/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:14
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	116056	12				
UPPER LIMIT	232112	12.5				
LOWER LIMIT	58028	11.5				
EPA SAMPLE NO.						
PR132-FB2-20251021	106541	12.01				
VX1103WBL01	117306	12.01				
VX1103WBS01	110522	12.01				
VX1103WBSD01	125216	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.