

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3422
 Lab File ID: VY023543.D Date Analyzed: 10/22/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:30
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	373234	13.346				
UPPER LIMIT	746468	13.846				
LOWER LIMIT	186617	12.846				
EPA SAMPLE NO.						
PR132-S15-000008-20251021	514846	13.35				
VY1022SBL01	484557	13.35				
VY1022SBS01	365280	13.35				
VY1022SBSD01	364513	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.