

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3422

Lab File ID: VX048444.D

BFB Injection Date: 11/03/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:20

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.4
75	30.0 - 60.0% of mass 95	57.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1 (1.3) 1
174	50.0 - 100.0% of mass 95	74.3
175	5.0 - 9.0% of mass 174	5.4 (7.2) 1
176	95.0 - 101.0% of mass 174	72 (96.9) 1
177	5.0 - 9.0% of mass 176	4.8 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX048445.D	11/03/2025	09:14
VX1103WBL01	VX1103WBL01	VX048447.D	11/03/2025	11:25
VX1103WBS01	VX1103WBS01	VX048449.D	11/03/2025	12:21
VX1103WBSD01	VX1103WBSD01	VX048450.D	11/03/2025	12:44
PR132-FB2-20251021	Q3422-02	VX048458.D	11/03/2025	15:41