

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3422

Lab File ID: VY023542.D

BFB Injection Date: 10/22/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 09:00

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.8
75	30.0 - 60.0% of mass 95	47.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1 (1) 1
174	50.0 - 100.0% of mass 95	99.3
175	5.0 - 9.0% of mass 174	7.2 (7.2) 1
176	95.0 - 101.0% of mass 174	97.5 (98.2) 1
177	5.0 - 9.0% of mass 176	6.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023543.D	10/22/2025	09:30
VY1022SBL01	VY1022SBL01	VY023544.D	10/22/2025	10:00
VY1022SBS01	VY1022SBS01	VY023545.D	10/22/2025	10:28
VY1022SBSD01	VY1022SBSD01	VY023546.D	10/22/2025	10:51
PR132-S15-000008-20251021	Q3422-01	VY023554.D	10/22/2025	16:22