



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: AECOM

SDG No.: Q3424

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	100	70 - 130	CV	10/23/2025	14:07	LB137626
CRI01	Aluminum	21.2	20.0	106	70 - 130	P	10/23/2025	16:18	LB137649
	Antimony	2.07	2.0	104	70 - 130	P	10/23/2025	16:18	LB137649
	Arsenic	1.29	1.0	129	70 - 130	P	10/23/2025	16:18	LB137649
	Barium	9.50	10.0	95	70 - 130	P	10/23/2025	16:18	LB137649
	Beryllium	1.08	1.0	108	70 - 130	P	10/23/2025	16:18	LB137649
	Cadmium	0.94	1.0	94	70 - 130	P	10/23/2025	16:18	LB137649
	Calcium	507	500	101	70 - 130	P	10/23/2025	16:18	LB137649
	Chromium	2.02	2.0	101	70 - 130	P	10/23/2025	16:18	LB137649
	Cobalt	1.05	1.0	105	50 - 150	P	10/23/2025	16:18	LB137649
	Copper	2.02	2.0	101	70 - 130	P	10/23/2025	16:18	LB137649
	Iron	57.7	50.0	115	70 - 130	P	10/23/2025	16:18	LB137649
	Lead	0.87	1.0	87	70 - 130	P	10/23/2025	16:18	LB137649
	Magnesium	476	500	95	70 - 130	P	10/23/2025	16:18	LB137649
	Manganese	1.01	1.0	101	50 - 150	P	10/23/2025	16:18	LB137649
	Nickel	0.86	1.0	86	70 - 130	P	10/23/2025	16:18	LB137649
	Potassium	519	500	104	70 - 130	P	10/23/2025	16:18	LB137649
	Selenium	4.86	5.0	97	70 - 130	P	10/23/2025	16:18	LB137649
	Silver	1.10	1.0	110	70 - 130	P	10/23/2025	16:18	LB137649
	Sodium	510	500	102	70 - 130	P	10/23/2025	16:18	LB137649
	Thallium	1.04	1.0	104	70 - 130	P	10/23/2025	16:18	LB137649
	Vanadium	4.89	5.0	98	70 - 130	P	10/23/2025	16:18	LB137649
	Zinc	5.22	5.0	104	50 - 150	P	10/23/2025	16:18	LB137649