



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-DP3-20251021

Lab Name: Alliance
Lab Code: ACE
Lab Sample ID: Q3424-01
Instrument ID (1): ECD_O

Contract: AECO02
SDG NO.: Q3424
Date(s) Analyzed: 10/23/2025 10/23/2025
Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)
Data file PO114642.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD	
			FROM	TO				
Aroclor-1242	1	5.491	5.441	5.541	34.2	37.2	9.72	
	2	5.513	5.463	5.563	39.8			
	3	5.574	5.524	5.624	23.8			
	4	5.669	5.619	5.719	39.5			
	5	6.398	6.348	6.448	48.7			
COLUMN 1								
	1	4.735	4.685	4.785	31.2	41.0		
	2	4.755	4.705	4.805	34.5			
	3	4.93	4.88	4.98	34.8			
	4	5.013	4.963	5.063	45.9			
5	5.536	5.486	5.586	58.8				
COLUMN 2								
	1	6.334	6.284	6.384	22.3	13.8	16	
	2	6.552	6.502	6.602	15.8			
	3	6.908	6.858	6.958	12.5			
	4	7.192	7.142	7.242	8.40			
5	7.607	7.557	7.657	9.70				
Aroclor-1254								
	1	5.536	5.486	5.586	27.1	16.2		
	2	5.681	5.631	5.731	18.1			
	3	6.084	6.034	6.134	16.2			
	4	6.312	6.262	6.362	12.4			
5	6.729	6.679	6.779	7.44				
COLUMN 1								
COLUMN 2								

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-S11-010080-2025

Lab Name: Alliance

Contract: AE002

Lab Code: ACE

SDG NO.: Q3424

Lab Sample ID: Q3424-02

Date(s) Analyzed: 10/23/2025 10/23/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO114643.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1242	1	5.492	5.442	5.542	26.7	32.4	1.24
COLUMN 1	2	5.513	5.463	5.563	31.1		
	3	5.576	5.526	5.626	24.9		
	4	5.669	5.619	5.719	44.3		
	5	6.398	6.348	6.448	35.2		
	1	4.734	4.684	4.784	22.5		
COLUMN 2	2	4.754	4.704	4.804	26.2	32.0	
	3	4.93	4.88	4.98	25.0		
	4	5.014	4.964	5.064	34.6		
	5	5.534	5.484	5.584	51.8		
Aroclor-1254	1	6.334	6.284	6.384	16.6	10.8	14.59
COLUMN 1	2	6.551	6.501	6.601	12.3		
	3	6.908	6.858	6.958	12.0		
	4	7.193	7.143	7.243	7.51		
	5	7.609	7.559	7.659	5.70		
	1	5.534	5.484	5.584	23.9		
COLUMN 2	2	5.681	5.631	5.731	14.9	12.5	
	3	6.083	6.033	6.133	10.4		
	4	6.314	6.264	6.364	8.57		
	5	6.728	6.678	6.778	4.91		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-S14-00094-20251

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3424

Lab Sample ID: Q3424-03

Date(s) Analyzed: 10/23/2025 10/23/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO114644.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1242	1	5.492	5.442	5.542	38.4	40.6	0.25
COLUMN 1	2	5.513	5.463	5.563	39.7		
	3	5.575	5.525	5.625	28.4		
	4	5.669	5.619	5.719	46.4		
	5	6.397	6.347	6.447	50.2		
	1	4.734	4.684	4.784	28.4		
COLUMN 2	2	4.754	4.704	4.804	32.3	40.7	
	3	4.928	4.878	4.978	38.5		
	4	5.012	4.962	5.062	43.5		
	5	5.534	5.484	5.584	60.5		
Aroclor-1254	1	6.333	6.283	6.383	16.3	14.8	1.36
COLUMN 1	2	6.55	6.5	6.6	18.6		
	3	6.91	6.86	6.96	18.8		
	4	7.193	7.143	7.243	11.1		
	5	7.607	7.557	7.657	9.15		
	1	5.534	5.484	5.584	26.6		
COLUMN 2	2	5.68	5.63	5.73	15.6	14.6	
	3	6.084	6.034	6.134	12.0		
	4	6.31	6.26	6.36	9.79		
	5	6.728	6.678	6.778	8.87		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-S07-000102-2025

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3424

Lab Sample ID: Q3395-03MS

Date(s) Analyzed: 10/23/2025 10/23/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071107.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.479	4.429	4.529	337	351	19.38
COLUMN 1	2	4.498	4.448	4.548	356		
	3	4.555	4.505	4.605	339		
	4	4.644	4.594	4.694	362		
	5	4.93	4.88	4.98	360		
	1	3.775	3.725	3.825	270		
COLUMN 2	2	3.79	3.74	3.84	278	289	
	3	3.951	3.901	4.001	301		
	4	3.999	3.949	4.049	285		
	5	4.195	4.145	4.245	309		
Aroclor-1254	1	5.292	5.242	5.342	175	198	10.64
COLUMN 1	2	5.507	5.457	5.557	142		
	3	5.855	5.805	5.905	92.9		
	4	6.134	6.084	6.184	71.3		
	5	6.542	6.492	6.592	510		
	1	4.531	4.481	4.581	175		
COLUMN 2	2	4.677	4.627	4.727	226	178	
	3	5.059	5.009	5.109	62.9		
	4	5.284	5.234	5.334	48.9		
	5	5.687	5.637	5.737	375		



Aroclor-1260	1	6.02	5.97	6.07	344		
	2	6.276	6.226	6.326	372		
	3	6.627	6.577	6.677	301		
	4	6.841	6.791	6.891	319		
	5	7.146	7.096	7.196	322		
COLUMN 1						332	18.78
	1	5.187	5.137	5.237	304		
	2	5.377	5.327	5.427	289		
	3	5.517	5.467	5.567	294		
	4	5.977	5.927	6.027	246		
5	6.222	6.172	6.272	241	275		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-S07-000102-2025

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3424

Lab Sample ID: Q3395-04MSD

Date(s) Analyzed: 10/23/2025 10/23/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071108.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.479	4.429	4.529	365	352	18.29
COLUMN 1	2	4.498	4.448	4.548	342		
	3	4.556	4.506	4.606	339		
	4	4.645	4.595	4.695	358		
	5	4.931	4.881	4.981	356		
	1	3.775	3.725	3.825	312		
COLUMN 2	2	3.791	3.741	3.841	271	293	
	3	3.951	3.901	4.001	301		
	4	4	3.95	4.05	277		
	5	4.195	4.145	4.245	305		
Aroclor-1254	1	5.293	5.243	5.343	173	197	10.7
COLUMN 1	2	5.507	5.457	5.557	140		
	3	5.855	5.805	5.905	93.1		
	4	6.135	6.085	6.185	71.2		
	5	6.544	6.494	6.594	509		
	1	4.532	4.482	4.582	174		
COLUMN 2	2	4.677	4.627	4.727	225	177	
	3	5.059	5.009	5.109	62.4		
	4	5.284	5.234	5.334	48.7		
	5	5.688	5.638	5.738	372		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170227BS

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3424

Lab Sample ID: PB170227BS

Date(s) Analyzed: 10/24/2025 10/24/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071123.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.485	4.435	4.535	175	178	21.12
COLUMN 1	2	4.503	4.453	4.553	180		
	3	4.561	4.511	4.611	181		
	4	4.65	4.6	4.7	181		
	5	4.935	4.885	4.985	175		
	1	3.775	3.725	3.825	142		
COLUMN 2	2	3.791	3.741	3.841	144	144	
	3	3.951	3.901	4.001	146		
	4	4	3.95	4.05	144		
	5	4.195	4.145	4.245	142		
Aroclor-1260	1	6.025	5.975	6.075	194	181	22.09
COLUMN 1	2	6.281	6.231	6.331	198		
	3	6.631	6.581	6.681	163		
	4	6.845	6.795	6.895	176		
	5	7.151	7.101	7.201	173		
	1	5.187	5.137	5.237	158		
COLUMN 2	2	5.377	5.327	5.427	150	145	
	3	5.517	5.467	5.567	150		
	4	5.978	5.928	6.028	136		
	5	6.222	6.172	6.272	131		