

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3424
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071113.D **Time Analyzed:** 17:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	544.580	500.000	8.9
Aroclor-1016-2	4.499	4.399	4.599	548.660	500.000	9.7
Aroclor-1016-3	4.556	4.456	4.656	555.030	500.000	11.0
Aroclor-1016-4	4.646	4.546	4.746	561.360	500.000	12.3
Aroclor-1016-5	4.930	4.831	5.031	552.730	500.000	10.5
Aroclor-1260-1	6.020	5.919	6.119	554.140	500.000	10.8
Aroclor-1260-2	6.276	6.175	6.375	560.510	500.000	12.1
Aroclor-1260-3	6.628	6.526	6.726	569.440	500.000	13.9
Aroclor-1260-4	6.841	6.740	6.940	581.630	500.000	16.3
Aroclor-1260-5	7.147	7.045	7.245	575.760	500.000	15.2
Decachlorobiphenyl	8.498	8.396	8.596	57.760	50.000	15.5
Tetrachloro-m-xylene	3.400	3.300	3.500	54.630	50.000	9.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3424
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071113.D **Time Analyzed:** 17:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	447.270	500.000	-10.5
Aroclor-1016-2	3.791	3.693	3.893	454.990	500.000	-9.0
Aroclor-1016-3	3.952	3.853	4.053	468.020	500.000	-6.4
Aroclor-1016-4	4.000	3.902	4.102	465.860	500.000	-6.8
Aroclor-1016-5	4.195	4.097	4.297	471.500	500.000	-5.7
Aroclor-1260-1	5.187	5.089	5.289	475.400	500.000	-4.9
Aroclor-1260-2	5.377	5.278	5.478	464.790	500.000	-7.0
Aroclor-1260-3	5.518	5.419	5.619	460.820	500.000	-7.8
Aroclor-1260-4	5.978	5.879	6.079	472.380	500.000	-5.5
Aroclor-1260-5	6.223	6.124	6.324	455.280	500.000	-8.9
Decachlorobiphenyl	7.532	7.433	7.633	55.270	50.000	10.5
Tetrachloro-m-xylene	2.776	2.678	2.878	48.960	50.000	-2.1