

Surrogate Summary

SDG No.: Q3424

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114312.D	PIBLK-PO114312.D	Tetrachloro-m-xyl	1	20	18.6	93		60	140
		Decachlorobiphen	1	20	19.2	96		60	140
		Tetrachloro-m-xyl	2	20	18.9	94		60	140
		Decachlorobiphen	2	20	19.0	95		60	140
I.BLK-PO114636.D	PIBLK-PO114636.D	Tetrachloro-m-xyl	1	20	21.1	105		60	140
		Decachlorobiphen	1	20	19.7	98		60	140
		Tetrachloro-m-xyl	2	20	19.5	97		60	140
		Decachlorobiphen	2	20	17.7	88		60	140
Q3424-01	PR132-DP3-20251021	Tetrachloro-m-xyl	1	20	22.8	114		32	144
		Decachlorobiphen	1	20	17.1	85		32	175
		Tetrachloro-m-xyl	2	20	22.6	113		32	144
		Decachlorobiphen	2	20	15.4	77		32	175
Q3424-02	PR132-S11-010080-20251021	Tetrachloro-m-xyl	1	20	27.7	139		32	144
		Decachlorobiphen	1	20	22.9	115		32	175
		Tetrachloro-m-xyl	2	20	26.7	134		32	144
		Decachlorobiphen	2	20	21.3	107		32	175
Q3424-03	PR132-S14-000094-20251021	Tetrachloro-m-xyl	1	20	26.9	135		32	144
		Decachlorobiphen	1	20	22.1	110		32	175
		Tetrachloro-m-xyl	2	20	27.2	136		32	144
		Decachlorobiphen	2	20	20.8	104		32	175
I.BLK-PO114650.D	PIBLK-PO114650.D	Tetrachloro-m-xyl	1	20	21.1	106		60	140
		Decachlorobiphen	1	20	19.9	99		60	140
		Tetrachloro-m-xyl	2	20	19.2	96		60	140
		Decachlorobiphen	2	20	19.1	96		60	140
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	19.4	97		60	140
I.BLK-PQ071102.D	PIBLK-PQ071102.D	Tetrachloro-m-xyl	1	20	20.6	103		60	140
		Decachlorobiphen	1	20	23.2	116		60	140
		Tetrachloro-m-xyl	2	20	19.1	95		60	140
		Decachlorobiphen	2	20	22.9	114		60	140
PB170227BL	PB170227BL	Tetrachloro-m-xyl	1	20	24.4	122		32	144
		Decachlorobiphen	1	20	27.7	138		32	175
		Tetrachloro-m-xyl	2	20	22.8	114		32	144
		Decachlorobiphen	2	20	26.7	134		32	175
Q3395-03MS	PR132-S07-000102-20251017M	Tetrachloro-m-xyl	1	20	26.2	131		32	144
		Decachlorobiphen	1	20	29.6	148		32	175
		Tetrachloro-m-xyl	2	20	22.4	112		32	144
		Decachlorobiphen	2	20	27.5	138		32	175
Q3395-04MSD	PR132-S07-000102-20251017M	Tetrachloro-m-xyl	1	20	26.3	131		32	144

Surrogate Summary

SDG No.: Q3424

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3395-04MSD	PR132-S07-000102-20251017MS	Decachlorobiphen	1	20	29.8	149		32	175
		Tetrachloro-m-xyl	2	20	22.1	110		32	144
		Decachlorobiphen	2	20	27.6	138		32	175
I.BLK-PQ071117.D	PIBLK-PQ071117.D	Tetrachloro-m-xyl	1	20	20.7	104		60	140
		Decachlorobiphen	1	20	23.6	118		60	140
		Tetrachloro-m-xyl	2	20	19.3	97		60	140
I.BLK-PQ071120.D	PIBLK-PQ071120.D	Decachlorobiphen	2	20	23.0	115		60	140
		Tetrachloro-m-xyl	1	20	20.3	101		60	140
		Decachlorobiphen	1	20	23.2	116		60	140
PB170227BS	PB170227BS	Tetrachloro-m-xyl	2	20	18.0	90		60	140
		Decachlorobiphen	2	20	21.9	110		60	140
		Tetrachloro-m-xyl	1	20	22.2	111		32	144
I.BLK-PQ071128.D	PIBLK-PQ071128.D	Decachlorobiphen	1	20	25.4	127		32	175
		Tetrachloro-m-xyl	2	20	18.1	91		32	144
		Decachlorobiphen	2	20	23.8	119		32	175
		Tetrachloro-m-xyl	1	20	20.8	104		60	140
		Decachlorobiphen	1	20	24.2	121		60	140
		Tetrachloro-m-xyl	2	20	18.4	92		60	140
		Decachlorobiphen	2	20	22.9	115		60	140