

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3424</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>10/28/2025 13:17</u>
Lab File ID:	<u>BF144097.D</u>	Init. Calib. Date(s):	<u>10/24/2025 10/24/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:13 16:24</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.264	1.303		3.1	
Benzaldehyde	1.084	1.105		1.9	
Phenol-d6	1.396	1.466		5.0	
Phenol	1.722	1.775		3.1	20.0
bis(2-Chloroethyl)ether	1.245	1.329		6.7	
2-Chlorophenol	1.245	1.300		4.4	
2-Methylphenol	0.957	1.028		7.4	
2,2-oxybis(1-Chloropropane)	2.339	2.513		7.4	
Acetophenone	0.432	0.450		4.2	
3+4-Methylphenols	1.132	1.216		7.4	
n-Nitroso-di-n-propylamine	0.903	1.004	0.050	11.2	
Nitrobenzene-d5	0.365	0.371		1.6	
Hexachloroethane	0.541	0.558		3.1	
Nitrobenzene	0.386	0.408		5.7	
Isophorone	0.638	0.714		11.9	
2-Nitrophenol	0.184	0.203		10.3	20.0
2,4-Dimethylphenol	0.273	0.303		11.0	
bis(2-Chloroethoxy)methane	0.398	0.444		11.6	
2,4-Dichlorophenol	0.317	0.338		6.6	20.0
Naphthalene	0.944	0.987		4.6	
4-Chloroaniline	0.329	0.364		10.6	
Hexachlorobutadiene	0.267	0.269		0.7	20.0
Caprolactam	0.080	0.097		21.3	
4-Chloro-3-methylphenol	0.280	0.322		15.0	20.0
2-Methylnaphthalene	0.622	0.666		7.1	
Hexachlorocyclopentadiene	0.212	0.238	0.050	12.3	
2,4,6-Trichlorophenol	0.447	0.474		6.0	20.0
2-Fluorobiphenyl	1.421	1.379		-3.0	
2,4,5-Trichlorophenol	0.469	0.522		11.3	
1,1-Biphenyl	1.432	1.391		-2.9	
2-Chloronaphthalene	1.199	1.218		1.6	
2-Nitroaniline	0.356	0.388		9.0	
Dimethylphthalate	1.311	1.432		9.2	
Acenaphthylene	1.617	1.684		4.1	
2,6-Dinitrotoluene	0.259	0.297		14.7	
3-Nitroaniline	0.288	0.326		13.2	
Acenaphthene	1.080	1.111		2.9	20.0
2,4-Dinitrophenol	0.140	0.166	0.050	18.6	
4-Nitrophenol	0.195	0.221	0.050	13.3	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3424</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>10/28/2025 13:17</u>
Lab File ID:	<u>BF144097.D</u>	Init. Calib. Date(s):	<u>10/24/2025 10/24/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:13 16:24</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.499	1.565		4.4	
2,4-Dinitrotoluene	0.350	0.407		16.3	
Diethylphthalate	1.233	1.355		9.9	
4-Chlorophenyl-phenylether	0.643	0.693		7.8	
Fluorene	1.143	1.207		5.6	
4-Nitroaniline	0.273	0.304		11.4	
4,6-Dinitro-2-methylphenol	0.128	0.136		6.3	
n-Nitrosodiphenylamine	0.639	0.653		2.2	20.0
2,4,6-Tribromophenol	0.255	0.287		12.5	
4-Bromophenyl-phenylether	0.250	0.266		6.4	
Hexachlorobenzene	0.307	0.322		4.9	
Atrazine	0.210	0.208		-1.0	
Pentachlorophenol	0.146	0.170		16.4	20.0
Phenanthrene	0.994	1.013		1.9	
Anthracene	1.026	1.035		0.9	
Carbazole	0.874	0.869		-0.6	
Di-n-butylphthalate	1.071	1.082		1.0	
Fluoranthene	1.098	1.000		-8.9	20.0
Pyrene	1.438	2.068		43.8	
Terphenyl-d14	1.144	1.573		37.5	
Butylbenzylphthalate	0.544	0.651		19.7	
3,3-Dichlorobenzidine	0.470	0.517		10.0	
Benzo (a) anthracene	1.352	1.480		9.5	
Chrysene	1.268	1.334		5.2	
Bis(2-ethylhexyl)phthalate	0.743	0.775		4.3	
Di-n-octyl phthalate	1.277	1.323		3.6	20.0
Benzo (b) fluoranthene	1.131	1.126		-0.4	
Benzo (k) fluoranthene	1.043	1.047		0.4	
Benzo (a) pyrene	1.014	1.094		7.9	20.0
Indeno (1,2,3-cd) pyrene	1.373	1.546		12.6	
Dibenzo (a,h) anthracene	1.087	1.218		12.1	
Benzo (g,h,i) perylene	1.093	1.265		15.7	
1,2,4,5-Tetrachlorobenzene	0.687	0.668		-2.8	
1,4-Dioxane	0.546	0.520		-4.8	20.0
2,3,4,6-Tetrachlorophenol	0.330	0.372		12.7	

All other compounds must meet a minimum RRF of 0.010.