

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
 Data File : BG064618.D
 Acq On : 29 Oct 2025 11:22
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 29 12:00:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 24 16:40:26 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.528	0.560	-6.1	97	0.00
3	Pyridine	1.584	1.622	-2.4	89	0.00
4	n-Nitrosodimethylamine	0.850	0.861	-1.3	87	0.00
5 S	2-Fluorophenol	1.192	1.266	-6.2	89	0.00
6	Aniline	2.225	2.273	-2.2	87	0.00
7 S	Phenol-d6	1.635	1.656	-1.3	85	0.00
8	2-Chlorophenol	1.247	1.377	-10.4	91	0.00
9	Benzaldehyde	0.893	0.838	6.2	81	0.00
10 C	Phenol	1.808	1.838	-1.7	86	0.00
11	bis(2-Chloroethyl)ether	2.225	2.273	-2.2	87	0.00
12	1,3-Dichlorobenzene	1.435	1.546	-7.7	93	0.00
13 C	1,4-Dichlorobenzene	1.460	1.553	-6.4	92	0.00
14	1,2-Dichlorobenzene	1.408	1.517	-7.7	93	0.00
15	Benzyl Alcohol	1.348	1.422	-5.5	91	0.00
16	2,2'-oxybis(1-Chloropropane	3.487	3.674	-5.4	90	0.00
17	2-Methylphenol	0.674	0.720	-6.8	88	0.00
18	Hexachloroethane	0.512	0.567	-10.7	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.130	1.223	-8.2	91	0.00
20	3+4-Methylphenols	1.688	1.742	-3.2	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
22	Acetophenone	0.547	0.581	-6.2	90	0.00
23 S	Nitrobenzene-d5	0.333	0.401	-20.4	96	0.00
24	Nitrobenzene	0.362	0.429	-18.5	95	0.00
25	Isophorone	0.799	0.833	-4.3	87	0.00
26 C	2-Nitrophenol	0.121	0.168	-38.8#	111	0.00
27	2,4-Dimethylphenol	0.303	0.311	-2.6	84	0.00
28	bis(2-Chloroethoxy)methane	0.455	0.480	-5.5	89	0.00
29 C	2,4-Dichlorophenol	0.320	0.352	-10.0	88	0.00
30	1,2,4-Trichlorobenzene	0.363	0.399	-9.9	92	0.00
31	Naphthalene	1.109	1.159	-4.5	89	0.00
32	Benzoic acid	0.189	0.203	-7.4	90	-0.03
33	4-Chloroaniline	0.431	0.435	-0.9	86	0.00
34 C	Hexachlorobutadiene	0.284	0.322	-13.4	97	0.00
35	Caprolactam	0.123	0.108	12.2	73	-0.03
36 C	4-Chloro-3-methylphenol	0.384	0.389	-1.3	83	0.00
37	2-Methylnaphthalene	0.814	0.836	-2.7	87	0.00
38	1-Methylnaphthalene	0.801	0.825	-3.0	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
40	1,2,4,5-Tetrachlorobenzene	0.671	0.754	-12.4	90	0.00
41 P	Hexachlorocyclopentadiene	0.302	0.400	-32.5#	103	0.00
42 S	2,4,6-Tribromophenol	0.289	0.312	-8.0	82	0.00
43 C	2,4,6-Trichlorophenol	0.384	0.455	-18.5	90	0.00
44	2,4,5-Trichlorophenol	0.443	0.492	-11.1	85	0.00
45 S	2-Fluorobiphenyl	1.319	1.473	-11.7	89	0.00
46	1,1'-Biphenyl	1.422	1.535	-7.9	86	0.00
47	2-Chloronaphthalene	1.073	1.175	-9.5	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
 Data File : BG064618.D
 Acq On : 29 Oct 2025 11:22
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 29 12:00:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 24 16:40:26 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.311	0.378	-21.5	90	0.00
49	Acenaphthylene	1.693	1.770	-4.5	83	0.00
50	Dimethylphthalate	1.530	1.552	-1.4	79	0.00
51	2,6-Dinitrotoluene	0.218	0.272	-24.8	92	0.00
52 C	Acenaphthene	1.157	1.225	-5.9	84	0.00
53	3-Nitroaniline	0.263	0.288	-9.5	82	0.00
54 P	2,4-Dinitrophenol	0.132	0.164	-24.2	100	0.00
55	Dibenzofuran	1.870	1.930	-3.2	82	0.00
56 P	4-Nitrophenol	0.244	0.241	1.2	77	0.00
57	2,4-Dinitrotoluene	0.340	0.396	-16.5	86	0.00
58	Fluorene	1.460	1.476	-1.1	80	0.00
59	2,3,4,6-Tetrachlorophenol	0.424	0.469	-10.6	84	0.00
60	Diethylphthalate	1.563	1.507	3.6	76	0.00
61	4-Chlorophenyl-phenylether	0.820	0.840	-2.4	81	0.00
62	4-Nitroaniline	0.293	0.306	-4.4	78	0.00
63	Azobenzene	1.434	1.404	2.1	76	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.00
65	4,6-Dinitro-2-methylphenol	0.079	0.104	-31.6#	97	0.00
66 c	n-Nitrosodiphenylamine	0.540	0.582	-7.8	77	0.00
67	4-Bromophenyl-phenylether	0.237	0.262	-10.5	78	0.00
68	Hexachlorobenzene	0.270	0.300	-11.1	79	0.00
69	Atrazine	0.232	0.233	-0.4	73	0.00
70 C	Pentachlorophenol	0.169	0.189	-11.8	81	0.00
71	Phenanthrene	1.094	1.132	-3.5	75	0.00
72	Anthracene	1.113	1.125	-1.1	73	0.00
73	Carbazole	0.981	0.970	1.1	73	0.00
74	Di-n-butylphthalate	1.134	1.140	-0.5	71	0.00
75 C	Fluoranthene	1.370	1.343	2.0	72	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
77	Benzydine	0.448	0.509	-13.6	81	0.00
78	Pyrene	1.307	1.322	-1.1	73	0.00
79 S	Terphenyl-d14	1.060	1.074	-1.3	74	0.00
80	Butylbenzylphthalate	0.395	0.462	-17.0	80	0.00
81	Benzo(a)anthracene	1.326	1.395	-5.2	77	0.00
82	3,3'-Dichlorobenzidine	0.443	0.484	-9.3	80	0.00
83	Chrysene	1.262	1.298	-2.9	76	0.00
84	Bis(2-ethylhexyl)phthalate	0.590	0.671	-13.7	78	0.00
85 c	Di-n-octyl phthalate	0.995	1.203	-20.9#	87	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	85	0.00
87	Indeno(1,2,3-cd)pyrene	1.471	1.519	-3.3	82	-0.02
88	Benzo(b)fluoranthene	1.226	1.236	-0.8	79	0.00
89	Benzo(k)fluoranthene	1.227	1.235	-0.7	82	0.00
90 C	Benzo(a)pyrene	1.127	1.157	-2.7	82	-0.02
91	Dibenzo(a,h)anthracene	1.195	1.241	-3.8	83	-0.03
92	Benzo(g,h,i)perylene	1.471	1.519	-3.3	82	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 12:00:42 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 2