

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064568.D
Instrument ID: BNA_G

Contract: AECO02
SDG NO.: Q3424
DFTPP Injection Date: 10/24/2025
DFTPP Injection Time: 08:20

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.2) 1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.2
365	Greater than 1% of mass 198	5.1
441	Present, but less than mass 443	86.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.5 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BG064569.D	10/24/2025	09:00
SSTDICC005	SSTDICC005	BG064570.D	10/24/2025	09:41
SSTDICC010	SSTDICC010	BG064571.D	10/24/2025	11:02
SSTDICC020	SSTDICC020	BG064572.D	10/24/2025	12:23
SSTDICCC040	SSTDICCC040	BG064573.D	10/24/2025	13:03
SSTDICC060	SSTDICC060	BG064574.D	10/24/2025	13:44
SSTDICC080	SSTDICC080	BG064575.D	10/24/2025	14:24
SSTDICC050	SSTDICC050	BG064576.D	10/24/2025	15:32