

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>AECO02</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3424</u>
Instrument ID: <u>MSVOA_Y</u>	Calibration Date(s): <u>10/07/2025</u> <u>10/07/2025</u>
Heated Purge: (Y/N) <u>Y</u>	Calibration Time(s): <u>09:17</u> <u>12:32</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:		RRF005 = VY023384.D		RRF010 = VY023385.D		RRF020 = VY023386.D			
		RRF050 = VY023387.D		RRF100 = VY023388.D		RRF150 = VY023389.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dichlorodifluoromethane	0.397	0.399	0.400	0.321	0.310	0.333	0.360	12	
Chloromethane	0.518	0.557	0.512	0.449	0.438	0.467	0.490	9.4	
Vinyl Chloride	0.660	0.684	0.659	0.598	0.597	0.633	0.639	5.6	
Bromomethane	0.601	0.596	0.548	0.469	0.483	0.518	0.536	10.4	
Chloroethane	0.441	0.462	0.453	0.414	0.413	0.429	0.435	4.7	
Trichlorofluoromethane	0.884	0.920	0.946	0.842	0.850	0.877	0.887	4.5	
1,1,2-Trichlorotrifluoroethane	0.484	0.486	0.479	0.437	0.429	0.445	0.460	5.6	
1,1-Dichloroethene	0.449	0.475	0.456	0.427	0.426	0.443	0.446	4.1	
Acetone	0.128	0.117	0.094	0.084	0.085	0.089	0.100	18.7	
Carbon Disulfide	1.291	1.367	1.320	1.225	1.210	1.246	1.277	4.7	
Methyl tert-butyl Ether	1.006	1.075	1.075	1.020	1.091	1.165	1.072	5.3	
Methyl Acetate	0.244	0.265	0.260	0.245	0.266	0.286	0.261	6	
Methylene Chloride	0.625	0.634	0.576	0.458	0.463	0.473	0.538	15.4	
trans-1,2-Dichloroethene	0.494	0.504	0.507	0.470	0.478	0.494	0.491	2.9	
1,1-Dichloroethane	0.800	0.835	0.827	0.759	0.765	0.792	0.796	3.9	
Cyclohexane	0.797	0.717	0.697	0.643	0.644	0.676	0.696	8.3	
2-Butanone	0.133	0.134	0.115	0.105	0.115	0.126	0.122	9.6	
Carbon Tetrachloride	0.483	0.501	0.504	0.481	0.487	0.503	0.493	2.1	
cis-1,2-Dichloroethene	0.542	0.584	0.582	0.542	0.563	0.591	0.567	3.8	
Bromochloromethane	0.332	0.323	0.320	0.267	0.270	0.273	0.297	10.2	
Chloroform	0.886	0.912	0.901	0.835	0.842	0.874	0.875	3.5	
1,1,1-Trichloroethane	0.808	0.814	0.822	0.767	0.764	0.796	0.795	3.1	
Methylcyclohexane	0.491	0.486	0.537	0.527	0.554	0.583	0.530	7	
Benzene	1.253	1.302	1.322	1.259	1.281	1.312	1.288	2.2	
1,2-Dichloroethane	0.328	0.346	0.343	0.315	0.324	0.338	0.332	3.6	
Trichloroethene	0.380	0.381	0.386	0.367	0.372	0.376	0.377	1.8	
1,2-Dichloropropane	0.276	0.289	0.292	0.274	0.283	0.293	0.285	2.9	
Bromodichloromethane	0.440	0.463	0.459	0.439	0.449	0.468	0.453	2.7	
4-Methyl-2-Pentanone	0.141	0.159	0.161	0.159	0.180	0.196	0.166	11.6	
Toluene	0.777	0.818	0.855	0.827	0.855	0.885	0.836	4.5	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q3424

Instrument ID: MSVOA_Y

Calibration Date(s): 10/07/2025 10/07/2025

Heated Purge: (Y/N) Y

Calibration Time(s): 09:17 12:32

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY023384.D		RRF010 = VY023385.D		RRF020 = VY023386.D			
		RRF050 = VY023387.D		RRF100 = VY023388.D		RRF150 = VY023389.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.357	0.380	0.390	0.385	0.415	0.438	0.394	7.2	
cis-1,3-Dichloropropene	0.424	0.460	0.465	0.455	0.481	0.505	0.465	5.9	
1,1,2-Trichloroethane	0.234	0.247	0.241	0.231	0.244	0.254	0.242	3.5	
2-Hexanone	0.108	0.117	0.110	0.112	0.126	0.138	0.119	9.9	
Dibromochloromethane	0.319	0.337	0.340	0.327	0.347	0.365	0.339	4.7	
1,2-Dibromoethane	0.215	0.229	0.231	0.223	0.237	0.249	0.231	5.2	
Tetrachloroethene	0.538	0.517	0.561	0.523	0.502	0.456	0.516	6.9	
Chlorobenzene	1.086	1.095	1.092	1.053	1.083	1.116	1.087	1.9	
Ethyl Benzene	1.606	1.711	1.753	1.777	1.824	1.891	1.760	5.5	
m/p-Xylenes	0.632	0.675	0.715	0.719	0.734	0.763	0.706	6.5	
o-Xylene	0.588	0.623	0.653	0.670	0.701	0.738	0.662	8.1	
Styrene	0.902	1.026	1.098	1.109	1.166	1.233	1.089	10.5	
Bromoform	0.225	0.242	0.238	0.230	0.244	0.262	0.240	5.4	
Isopropylbenzene	2.967	3.153	3.353	3.317	3.425	3.555	3.295	6.3	
1,1,2,2-Tetrachloroethane	0.530	0.548	0.502	0.499	0.543	0.611	0.539	7.6	
1,3-Dichlorobenzene	1.667	1.698	1.718	1.657	1.701	1.794	1.706	2.9	
1,4-Dichlorobenzene	1.703	1.703	1.694	1.624	1.653	1.732	1.685	2.3	
1,2-Dichlorobenzene	1.485	1.498	1.511	1.446	1.490	1.543	1.496	2.1	
1,2-Dibromo-3-Chloropropane	0.087	0.093	0.083	0.083	0.088	0.096	0.088	6.2	
1,2,4-Trichlorobenzene	0.846	0.867	0.901	0.918	1.007	1.054	0.932	8.7	
1,2,3-Trichlorobenzene	0.719	0.753	0.795	0.793	0.866	0.920	0.808	9.1	
1,2-Dichloroethane-d4	0.455	0.436	0.429	0.392	0.392	0.405	0.418	6.2	
Dibromofluoromethane	0.311	0.314	0.317	0.298	0.299	0.303	0.307	2.6	
Toluene-d8	1.132	1.170	1.177	1.129	1.122	1.135	1.144	2	
4-Bromofluorobenzene	0.401	0.370	0.373	0.365	0.372	0.386	0.378	3.5	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.