

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q3424

Instrument ID: MSVOA_Y

Calibration Date/Time: 10/22/2025 09:30

Lab File ID: VY023543.D

Init. Calib. Date(s): 10/07/2025 10/07/2025

Heated Purge: (Y/N) Y

Init. Calib. Time(s): 09:17 12:32

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.360	0.308		-14.44	20
Chloromethane	0.490	0.409	0.1	-16.53	20
Vinyl Chloride	0.639	0.552		-13.61	20
Bromomethane	0.536	0.456		-14.93	20
Chloroethane	0.435	0.403		-7.36	20
Trichlorofluoromethane	0.887	0.857		-3.38	20
1,1,2-Trichlorotrifluoroethane	0.460	0.469		1.96	20
1,1-Dichloroethene	0.446	0.456		2.24	20
Acetone	0.100	0.096		-4	20
Carbon Disulfide	1.277	1.206		-5.56	20
Methyl tert-butyl Ether	1.072	1.142		6.53	20
Methyl Acetate	0.261	0.294		12.64	20
Methylene Chloride	0.538	0.496		-7.81	20
trans-1,2-Dichloroethene	0.491	0.510		3.87	20
1,1-Dichloroethane	0.796	0.815	0.1	2.39	20
Cyclohexane	0.696	0.650		-6.61	20
2-Butanone	0.122	0.118		-3.28	20
Carbon Tetrachloride	0.493	0.522		5.88	20
cis-1,2-Dichloroethene	0.567	0.611		7.76	20
Bromochloromethane	0.297	0.257		-13.47	20
Chloroform	0.875	0.928		6.06	20
1,1,1-Trichloroethane	0.795	0.839		5.53	20
Methylcyclohexane	0.530	0.561		5.85	20
Benzene	1.288	1.371		6.44	20
1,2-Dichloroethane	0.332	0.340		2.41	20
Trichloroethene	0.377	0.411		9.02	20
1,2-Dichloropropane	0.285	0.303		6.32	20
Bromodichloromethane	0.453	0.487		7.51	20
4-Methyl-2-Pentanone	0.166	0.175		5.42	20
Toluene	0.836	0.926		10.77	20
t-1,3-Dichloropropene	0.394	0.433		9.9	20
cis-1,3-Dichloropropene	0.465	0.509		9.46	20
1,1,2-Trichloroethane	0.242	0.264		9.09	20
2-Hexanone	0.119	0.129		8.4	20
Dibromochloromethane	0.339	0.377		11.21	20
1,2-Dibromoethane	0.231	0.255		10.39	20
Tetrachloroethene	0.516	0.580		12.4	20
Chlorobenzene	1.087	1.168	0.3	7.45	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3424</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>10/22/2025</u> <u>09:30</u>
Lab File ID:	<u>VY023543.D</u>	Init. Calib. Date(s):	<u>10/07/2025</u> <u>10/07/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:17</u> <u>12:32</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.760	1.922		9.2	20
m/p-Xylenes	0.706	0.784		11.05	20
o-Xylene	0.662	0.739		11.63	20
Styrene	1.089	1.228		12.76	20
Bromoform	0.240	0.264	0.1	10	20
Isopropylbenzene	3.295	3.699		12.26	20
1,1,2,2-Tetrachloroethane	0.539	0.562	0.3	4.27	20
1,3-Dichlorobenzene	1.706	1.845		8.15	20
1,4-Dichlorobenzene	1.685	1.798		6.71	20
1,2-Dichlorobenzene	1.496	1.634		9.23	20
1,2-Dibromo-3-Chloropropane	0.088	0.092		4.55	20
1,2,4-Trichlorobenzene	0.932	1.020		9.44	20
1,2,3-Trichlorobenzene	0.808	0.876		8.42	20
1,2-Dichloroethane-d4	0.418	0.381		-8.85	20
Dibromofluoromethane	0.307	0.307		0	20
Toluene-d8	1.144	1.136		-0.7	20
4-Bromofluorobenzene	0.378	0.382		1.06	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.