284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QCBatch ID # VY100725

Review By	Semsettin Yesilyurt	Review On	10/10/2025 3:00:22 AM		
Supervise By	Mahesh Dadoda	Supervise On	10/10/2025 4:38:02 AM		
SubDirectory	VY100725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135766
Initial Calibration Stds	VP135767,VP135768,VP135769,VP135770,VP135771,VP135772
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP135773
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023383.D	07 Oct 2025 08:43		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023384.D	07 Oct 2025 09:17		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY023385.D	07 Oct 2025 10:02		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY023386.D	07 Oct 2025 11:24		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY023387.D	07 Oct 2025 11:47		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY023388.D	07 Oct 2025 12:09		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY023389.D	07 Oct 2025 12:32		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023390.D	07 Oct 2025 13:02		SY/MD	Ok
9	VSTDICV050	ICVVY100725	VY023391.D	07 Oct 2025 14:12		SY/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QCBatch ID # VY102225

Review By	John Carlone	Review On	10/24/2025 2:07:41 PM		
Supervise By	Mahesh Dadoda	Supervise On	10/24/2025 2:54:27 PM		
SubDirectory	VY102225	HP Acquire Method	MSVOA_Y	HP Processing Method	82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135879
Initial Calibration Stds	
CCC	VP135880,VP135881
Internal Standard/PEM	VP133934
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023542.D	22 Oct 2025 09:00		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023543.D	22 Oct 2025 09:30		SY/MD	Ok,M
3	VY1022SBL01	VY1022SBL01	VY023544.D	22 Oct 2025 10:00		SY/MD	Ok
4	VY1022SBS01	VY1022SBS01	VY023545.D	22 Oct 2025 10:28		SY/MD	Ok,M
5	VY1022SBSD01	VY1022SBSD01	VY023546.D	22 Oct 2025 10:51		SY/MD	Ok,M
6	Q3395-04MSD	PR132-S07-000102-20	VY023547.D	22 Oct 2025 11:28	vial-B Internal Standard Fail	SY/MD	Not Ok
7	Q3416-01	NB-08-10212025	VY023548.D	22 Oct 2025 12:03	vial-A	SY/MD	Ok
8	Q3404-01RE	MOO-25-0285-0288RE	VY023549.D	22 Oct 2025 12:52	vial-B Internal standard fail	SY/MD	Confirms
9	Q3427-03	TP-8-VOC	VY023550.D	22 Oct 2025 13:43	vial-A	SY/MD	Ok
10	Q3419-01	PR-132-S16-015094-20	VY023551.D	22 Oct 2025 14:07	vial-A	SY/MD	Ok
11	Q3419-04	PR132-S16-119133-20	VY023552.D	22 Oct 2025 15:34	vial-A	SY/MD	Ok
12	Q3419-05	PR132-S16-094119-20	VY023553.D	22 Oct 2025 15:58	vial-A	SY/MD	Ok
13	Q3422-01	PR132-S15-000008-20	VY023554.D	22 Oct 2025 16:22	vial-A	SY/MD	Ok
14	Q3424-01	PR132-DP3-20251021	VY023555.D	22 Oct 2025 16:46	vial-A	SY/MD	Ok
15	Q3424-02	PR132-S11-010080-20	VY023556.D	22 Oct 2025 17:10	vial-A	SY/MD	Ok
16	Q3424-03	PR132-S14-000094-20	VY023557.D	22 Oct 2025 17:35	vial-A	SY/MD	Ok

M : Manual Integration