

Surrogate Summary

SDG No.: Q3424

Client: AECOM

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3424-01	PR132-DP3-20251021	1,2-Dichloroethane-d4	50	53.2	106		63	155
		Dibromofluoromethane	50	50.5	101		70	134
		Toluene-d8	50	49.8	100		74	123
		4-Bromofluorobenzene	50	45.5	91		17	146
Q3424-02	PR132-S11-010080-20251021	1,2-Dichloroethane-d4	50	58.0	116		63	155
		Dibromofluoromethane	50	51.2	102		70	134
		Toluene-d8	50	49.7	99		74	123
		4-Bromofluorobenzene	50	47.3	95		17	146
Q3424-03	PR132-S14-000094-20251021	1,2-Dichloroethane-d4	50	51.1	102		63	155
		Dibromofluoromethane	50	50.8	102		70	134
		Toluene-d8	50	49.9	100		74	123
		4-Bromofluorobenzene	50	44.3	89		17	146
VY1022SBL01	VY1022SBL01	1,2-Dichloroethane-d4	50	48.3	97		63	155
		Dibromofluoromethane	50	49.8	100		70	134
		Toluene-d8	50	48.6	97		74	123
		4-Bromofluorobenzene	50	41.1	82		17	146
VY1022SBS01	VY1022SBS01	1,2-Dichloroethane-d4	50	45.7	91		63	155
		Dibromofluoromethane	50	49.3	99		70	134
		Toluene-d8	50	49.8	100		74	123
		4-Bromofluorobenzene	50	49.3	99		17	146
VY1022SBSD01	VY1022SBSD01	1,2-Dichloroethane-d4	50	45.1	90		63	155
		Dibromofluoromethane	50	49.3	99		70	134
		Toluene-d8	50	48.6	97		74	123
		4-Bromofluorobenzene	50	48.7	97		17	146