



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102825\
Data File : VN088126.D
Acq On : 28 Oct 2025 09:26
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Title : SW846 8260
Last Update : Sat Oct 25 00:15:22 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1839

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.1	20512	PASS
75	95	30	60	57.4	48912	PASS
95	95	100	100	100.0	85157	PASS
96	95	5	9	6.2	5307	PASS
173	174	0.00	2	1.1	601	PASS
174	95	50	100	66.3	56432	PASS
175	174	5	9	7.6	4276	PASS
176	174	95	101	98.0	55320	PASS
177	176	5	9	6.5	3593	PASS

Average of 12.823 to 12.835 min.: VN088126.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.95	1079	48.00	688	62.00	5129	74.00	1640
36.95	5754	49.00	4342	62.95	4207	75.00	4891
38.00	5296	50.00	20512	63.95	389	76.00	435
39.00	2142	51.00	5773	64.95	171	76.90	48
39.90	31	51.95	277	66.80	327	77.10	146
40.90	55	54.95	417	68.00	9986	77.95	378
42.70	62	55.95	1949	68.95	10569	78.90	3928
44.00	855	57.00	3328	70.00	921	79.75	686
44.95	814	57.80	72	71.80	375	80.05	743
47.05	1032	59.95	1014	72.00	183	80.90	4262
47.80	187	61.00	5136	72.90	4795	81.85	913

Instrument :

MSVOA_N

ClientSampleId :

BFB

Average of 12.823 to 12.835 min.: VN088126.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
82.90	61	103.00	51	116.95	433	139.80	86
87.00	3231	103.75	540	117.95	560	140.80	432
87.95	3099	104.80	104	118.60	235	140.95	728
90.70	89	105.10	90	119.05	596	141.75	126
90.95	370	105.80	421	127.75	352	142.70	324
92.00	2901	107.10	55	128.90	109	142.90	949
93.00	4193	110.90	58	129.80	121	144.80	56
94.00	11466	114.95	167	130.70	54	147.00	54
95.00	85157	115.70	142	134.70	97	147.75	161
96.05	5307	115.85	298	136.80	122	150.00	54
96.95	302	116.70	217	137.00	78	153.00	55

Average of 12.823 to 12.835 min.: VN088126.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
153.90	50	174.95	4276				
154.70	92	176.00	55320				
154.90	110	177.00	3593				
156.90	176	178.10	118				
158.80	84	206.80	88				
171.60	84	207.05	2				
171.90	219	281.00	50				
172.10	90						
172.80	150						
173.00	601						
174.00	56432						