

Surrogate Summary

SDG No.: **Q3426**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3426-01	TWP2	1,2-Dichloroethane-d4	50	59.6	119		74	125
		Dibromofluoromethane	50	49.5	99		75	124
		Toluene-d8	50	49.2	98		86	113
		4-Bromofluorobenzene	50	61.9	124	*	77	121
Q3426-01DL	TWP2DL	1,2-Dichloroethane-d4	50	52.6	105		74	125
		Dibromofluoromethane	50	45.1	90		75	124
		Toluene-d8	50	43.8	88		86	113
		4-Bromofluorobenzene	50	50.8	102		77	121
Q3426-02	TWP3	1,2-Dichloroethane-d4	50	54.2	108		74	125
		Dibromofluoromethane	50	46.6	93		75	124
		Toluene-d8	50	44.4	89		86	113
		4-Bromofluorobenzene	50	48.9	98		77	121
Q3426-03	TWP5	1,2-Dichloroethane-d4	50	56.5	113		74	125
		Dibromofluoromethane	50	46.6	93		75	124
		Toluene-d8	50	44.8	90		86	113
		4-Bromofluorobenzene	50	49.7	99		77	121
Q3426-04	MW2	1,2-Dichloroethane-d4	50	60.6	121		74	125
		Dibromofluoromethane	50	48.9	98		75	124
		Toluene-d8	50	45.8	92		86	113
		4-Bromofluorobenzene	50	52.5	105		77	121
VN1027WBL01	VN1027WBL01	1,2-Dichloroethane-d4	50	61.1	122		74	125
		Dibromofluoromethane	50	52.8	106		75	124
		Toluene-d8	50	49.9	100		86	113
		4-Bromofluorobenzene	50	52.0	104		77	121
VN1027WBS01	VN1027WBS01	1,2-Dichloroethane-d4	50	56.2	112		74	125
		Dibromofluoromethane	50	53.2	106		75	124
		Toluene-d8	50	51.1	102		86	113
		4-Bromofluorobenzene	50	52.6	105		77	121
VN1027WBSD0	VN1027WBSD01	1,2-Dichloroethane-d4	50	56.4	113		74	125
		Dibromofluoromethane	50	53.0	106		75	124
		Toluene-d8	50	51.0	102		86	113
		4-Bromofluorobenzene	50	52.9	106		77	121
VN1028WBL01	VN1028WBL01	1,2-Dichloroethane-d4	50	54.8	110		74	125
		Dibromofluoromethane	50	46.8	94		75	124
		Toluene-d8	50	44.8	90		86	113
		4-Bromofluorobenzene	50	48.3	97		77	121
VN1028WBS02	VN1028WBS02	1,2-Dichloroethane-d4	50	48.6	97		74	125
		Dibromofluoromethane	50	47.6	95		75	124
		Toluene-d8	50	46.6	93		86	113
		4-Bromofluorobenzene	50	47.9	96		77	121