

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q3426
Lab File ID: VN088102.D Date Analyzed: 10/27/2025
Instrument ID: MSVOA_N Time Analyzed: 10:40
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	321965	8.21	587119	9.08	522069	11.84
UPPER LIMIT	643930	8.706	1174240	9.582	1044140	12.341
LOWER LIMIT	160983	7.706	293560	8.582	261035	11.341
EPA SAMPLE NO.						
TWP2	270716	8.21	619054	9.08	604894	11.84
TWP3	224334	8.21	500872	9.08	474497	11.84
TWP5	246197	8.21	561002	9.08	537996	11.84
MW2	257811	8.21	580168	9.08	561441	11.84
VN1027WBL01	198801	8.21	446179	9.08	422679	11.84
VN1027WBS01	300941	8.21	563874	9.08	505093	11.85
VN1027WBSD01	293635	8.21	537984	9.08	482887	11.85

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.