

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q3426
Lab File ID: VN088127.D Date Analyzed: 10/28/2025
Instrument ID: MSVOA_N Time Analyzed: 10:23
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	299407	8.21	575460	9.08	507990	11.85
UPPER LIMIT	598814	8.706	1150920	9.582	1015980	12.347
LOWER LIMIT	149704	7.706	287730	8.582	253995	11.347
EPA SAMPLE NO.						
TWP2DL	238689	8.21	533116	9.08	499340	11.84
VN1028WBL01	231261	8.21	516618	9.08	484454	11.85
VN1028WBS02	325166	8.21	590689	9.08	518318	11.85

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.