

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3428

Project: Monthly 2025

Run No.: LB137623

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137623BS							
BOD5	mg/L	198	186		94	1	84.6-115.4	10/23/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3428

Project: Monthly 2025

Run No.: LB137648

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137648BS							
TSS	mg/L	550	531		96	1	90-110	10/24/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3428

Project: Monthly 2025

Run No.: LB137651

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137651BS							
TPH	mg/L	20.0	16.7		84	1	78-114	10/27/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3428

Project: Monthly 2025

Run No.: LB137651

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137651BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	10/27/2025