

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3432

Project: AEC-2025-0013- CSC 4026 Easthampton

RunNo.: LB137643

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: COD	ICV	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: COD	CCV1	mg/L	51.184	50	102	95-105	10/24/2025
Sample ID: COD	CCV2	mg/L	51.184	50	102	95-105	10/24/2025