

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3432

Project: AEC-2025-0013- CSC 4026 Easthampton

Run No.: LB137643

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137643BS							
COD	mg/L	50	49.2		98	1	90-110	10/24/2025

Laboratory Control Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3432
Project:	AEC-2025-0013- CSC 4026 Easthampton	Run No.:	LB137648

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137648BS							
TSS	mg/L	550	531		96	1	90-110	10/24/2025