

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137618

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/22/2025	15:10
2	CAL2	1	Water	NA	NA	20.2	7.01	10/22/2025	15:11
3	CAL3	1	Water	NA	NA	20.3	10.02	10/22/2025	15:15
4	ICV	1	Water	NA	NA	20.1	7.00	10/22/2025	15:20
5	CCV1	1	Water	NA	NA	20.2	2.01	10/22/2025	15:22
6	Q3427-04	1	Solid	20.02	20	20.8	6.20	10/22/2025	15:35
7	Q3427-04DUP	1	Solid	20.03	20	20.9	6.22	10/22/2025	15:37
8	Q3434-01	1	Solid	20.02	20	20.1	8.62	10/22/2025	15:44
9	CCV2	1	Water	NA	NA	20.2	12.02	10/22/2025	15:45

WORKLIST(Hardcopy Internal Chain)

9137618

WorkList Name : corrosivity q3427 WorkList ID : 192613 Department : Wet-Chemistry Date : 10-22-2025 12:38:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3427-04	TP-8	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	10/21/2025	9045D
Q3434-01	PR132-WC1-20251022	Solid	Corrosivity	Cool 4 deg C	AECO02	D41	10/22/2025	9045D

Date/Time 10/22/25 15:00

Raw Sample Received by: SPC1

Raw Sample Relinquished by: SPC1

Date/Time 10/22/25

Raw Sample Received by: SPC1

Raw Sample Relinquished by: SPC1

17:40