

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137621

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.02	10/23/2025	08:35
2	CAL2	1	Water	NA	NA	20.2	7.01	10/23/2025	08:36
3	CAL3	1	Water	NA	NA	20.2	10.02	10/23/2025	08:37
4	ICV	1	Water	NA	NA	20.3	7.00	10/23/2025	08:39
5	CCV1	1	Water	NA	NA	20.3	2.01	10/23/2025	08:40
6	Q3434-02	1	Water	NA	NA	20.7	7.77	10/23/2025	08:50
7	Q3434-02DUP	1	Water	NA	NA	20.3	7.79	10/23/2025	08:52
8	CCV2	1	Water	NA	NA	20.3	12.02	10/23/2025	08:55

VB 137621

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W Q3434 WorkList ID : 192628 Department : Wet-Chemistry Date : 10-23-2025 08:18:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3434-02	PR132-WC2-20251022	Water	pH	Cool 4 deg C	AECO02	D41	10/22/2025	9040C

Date/Time 10/23/25 08:25
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 10/23/25 13:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]