

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137627

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	10/23/2025	09:00
2	Q3434-02		1	13	100.00	>212.0	>212.0	10/23/2025	09:30
3	Q3434-02DUP		1	13	100.00	>212.0	>212.0	10/23/2025	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB137627

WorkList Name : fp-10-23.

WorkList ID : 192645

Department : Wet-Chemistry

Date : 10-23-2025 08:15:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3434-02	PR132-WC2-20251022	Water	Flash Point	Cool 4 deg C	AECO02	D41	10/22/2025	1010B

Date/Time 10/23/2025 08:35
Raw Sample Received by: RM CWS
Raw Sample Relinquished by: RM CWS

Date/Time 10/23/2025 11:15
Raw Sample Received by: RM CWS
Raw Sample Relinquished by: RM CWS