



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3434

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137618

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.00	7	100	90-110	10/22/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	10/22/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	10/22/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3434

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137621

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	10/23/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	10/23/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	10/23/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3434

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137627

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.7	81	102	78-84	10/23/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3434

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137639

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.096	0.099	97	85-115	10/24/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/24/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/24/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	10/24/2025