

Laboratory Control Sample Summary

Client:	AECOM	SDG No.:	Q3434
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Run No.:	LB137651

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137651BS							
TPH	mg/L	20.0	16.7		84	1	78-114	10/27/2025

Laboratory Control Sample Summary

Client:	AECOM	SDG No.:	Q3434
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Run No.:	LB137651

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137651BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	10/27/2025