

Surrogate Summary

SDG No.: Q3434

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		57	171
		Tetrachloro-m-xyl	1	20	19.4	97		61	148
		Decachlorobiphen	2	20	20.3	102		57	171
		Tetrachloro-m-xyl	2	20	20.0	100		61	148
I.BLK-PD090775.D	PIBLK-PD090775.D	Decachlorobiphen	1	20	21.7	109		57	171
		Tetrachloro-m-xyl	1	20	20.3	101		61	148
		Decachlorobiphen	2	20	24.9	124		57	171
		Tetrachloro-m-xyl	2	20	21.1	106		61	148
PB170213BL	PB170213BL	Decachlorobiphen	1	20	20.7	103		20	144
		Tetrachloro-m-xyl	1	20	18.7	94		19	148
		Decachlorobiphen	2	20	23.8	119		20	144
		Tetrachloro-m-xyl	2	20	20.1	101		19	148
PB170213BS	PB170213BS	Decachlorobiphen	1	20	22.1	110		20	144
		Tetrachloro-m-xyl	1	20	17.6	88		19	148
		Decachlorobiphen	2	20	26.0	130		20	144
		Tetrachloro-m-xyl	2	20	18.3	91		19	148
Q3395-03MS	PR132-S07-000102-20251017MS	Decachlorobiphen	1	20	15.2	76		20	144
		Tetrachloro-m-xyl	1	20	14.4	72		19	148
		Decachlorobiphen	2	20	15.5	77		20	144
		Tetrachloro-m-xyl	2	20	15.5	78		19	148
Q3395-04MSD	PR132-S07-000102-20251017MSD	Decachlorobiphen	1	20	16.4	82		20	144
		Tetrachloro-m-xyl	1	20	15.3	76		19	148
		Decachlorobiphen	2	20	16.4	82		20	144
		Tetrachloro-m-xyl	2	20	16.4	82		19	148
I.BLK-PD090788.D	PIBLK-PD090788.D	Decachlorobiphen	1	20	20.9	105		57	171
		Tetrachloro-m-xyl	1	20	20.0	100		61	148
		Decachlorobiphen	2	20	23.5	117		57	171
		Tetrachloro-m-xyl	2	20	20.8	104		61	148
I.BLK-PD090799.D	PIBLK-PD090799.D	Decachlorobiphen	1	20	21.1	105		57	171
		Tetrachloro-m-xyl	1	20	19.7	99		61	148
		Decachlorobiphen	2	20	23.4	117		57	171
		Tetrachloro-m-xyl	2	20	20.3	101		61	148
Q3434-01	PR132-WC1-20251022	Decachlorobiphen	1	20	20.5	102		20	144
		Tetrachloro-m-xyl	1	20	19.0	95		19	148
		Decachlorobiphen	2	20	19.6	98		20	144
		Tetrachloro-m-xyl	2	20	19.7	98		19	148
I.BLK-PD090805.D	PIBLK-PD090805.D	Decachlorobiphen	1	20	21.4	107		57	171
		Tetrachloro-m-xyl	1	20	20.0	100		61	148
		Decachlorobiphen	2	20	22.2	111		57	171
		Tetrachloro-m-xyl	2	20	20.4	102		61	148
I.BLK-PD090879.D	PIBLK-PD090879.D	Decachlorobiphen	1	20	23.1	115		57	171

Surrogate Summary

SDG No.: Q3434

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090879.D	PIBLK-PD090879.D	Tetrachloro-m-xyl	1	20	21.6	108		61	148
		Decachlorobiphen	2	20	24.9	125		57	171
PB170287BL	PB170287BL	Tetrachloro-m-xyl	2	20	21.8	109		61	148
		Decachlorobiphen	1	20	20.5	102		57	171
		Tetrachloro-m-xyl	1	20	20.1	100		61	148
		Decachlorobiphen	2	20	18.5	93		57	171
		Tetrachloro-m-xyl	2	20	20.4	102		61	148
		Decachlorobiphen	1	20	19.3	96		57	171
PB170287BS	PB170287BS	Tetrachloro-m-xyl	1	20	15.6	78		61	148
		Decachlorobiphen	2	20	18.9	95		57	171
		Tetrachloro-m-xyl	2	20	15.3	76		61	148
		Decachlorobiphen	1	20	19.7	98		57	171
PB170287BSD	PB170287BSD	Tetrachloro-m-xyl	1	20	14.6	73		61	148
		Decachlorobiphen	2	20	19.6	98		57	171
		Tetrachloro-m-xyl	2	20	15.1	76		61	148
		Decachlorobiphen	1	20	13.7	69		57	171
		Tetrachloro-m-xyl	1	20	20.9	105		61	148
		Decachlorobiphen	2	20	13.8	69		57	171
Q3434-02	PR132-WC2-20251022	Tetrachloro-m-xyl	2	20	19.6	98		61	148
		Decachlorobiphen	1	20	22.8	114		57	171
		Tetrachloro-m-xyl	1	20	21.5	108		61	148
		Decachlorobiphen	2	20	23.0	115		57	171
I.BLK-PD090891.D	PIBLK-PD090891.D	Tetrachloro-m-xyl	2	20	21.6	108		61	148
		Decachlorobiphen	2	20	21.6	108		61	148