

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3434

Analytical Method: 8081B

Client: AECOM

Datafile : PD090889.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170287BSD (Column 1)	alpha-BHC	0.5	0.47	ug/L	94	2			85	130	20
	beta-BHC	0.5	0.46	ug/L	91	1			83	126	20
	delta-BHC	0.5	0.48	ug/L	96	1			69	141	20
	gamma-BHC (Lindane)	0.5	0.46	ug/L	93	1			82	129	20
	Heptachlor	0.5	0.55	ug/L	109	2			79	127	20
	Aldrin	0.5	0.47	ug/L	93	2			79	126	20
	Heptachlor epoxide	0.5	0.49	ug/L	97	2			81	124	20
	Endosulfan I	0.5	0.47	ug/L	94	0			85	122	20
	Dieldrin	0.5	0.47	ug/L	94	0			83	125	20
	4,4'-DDE	0.5	0.46	ug/L	93	2			80	127	20
	Endrin	0.5	0.45	ug/L	90	1			81	128	20
	Endosulfan II	0.5	0.46	ug/L	91	3			82	123	20
	4,4'-DDD	0.5	0.50	ug/L	100	1			77	131	20
	Endosulfan sulfate	0.5	0.46	ug/L	93	1			76	129	20
	4,4'-DDT	0.5	0.44	ug/L	89	1			80	133	20
	Methoxychlor	0.5	0.45	ug/L	90	1			78	108	20
	Endrin ketone	0.5	0.49	ug/L	97	1			80	131	20
	Endrin aldehyde	0.5	0.48	ug/L	96	1			82	127	20
	alpha-Chlordane	0.5	0.50	ug/L	101	3			82	125	20
	gamma-Chlordane	0.5	0.46	ug/L	91	2			82	125	20
	alpha-BHC	0.5	0.45	ug/L	90	1			85	130	20
	beta-BHC	0.5	0.45	ug/L	89	1			83	126	20
	delta-BHC	0.5	0.45	ug/L	91	1			69	141	20
	gamma-BHC (Lindane)	0.5	0.45	ug/L	90	1			82	129	20
PB170287BSD (Column 2)	Heptachlor	0.5	0.46	ug/L	91	2			79	127	20
	Aldrin	0.5	0.45	ug/L	90	1			79	126	20
	Heptachlor epoxide	0.5	0.45	ug/L	90	0			81	124	20
	Endosulfan I	0.5	0.44	ug/L	88	1			85	122	20
	Dieldrin	0.5	0.45	ug/L	91	0			83	125	20
	4,4'-DDE	0.5	0.45	ug/L	91	0			80	127	20
	Endrin	0.5	0.44	ug/L	88	0			81	128	20
	Endosulfan II	0.5	0.45	ug/L	90	0			82	123	20
	4,4'-DDD	0.5	0.48	ug/L	96	0			77	131	20
	Endosulfan sulfate	0.5	0.45	ug/L	91	1			76	129	20
	4,4'-DDT	0.5	0.43	ug/L	86	0			80	133	20
	Methoxychlor	0.5	0.44	ug/L	87	1			78	108	20
	Endrin ketone	0.5	0.47	ug/L	93	2			80	131	20
	Endrin aldehyde	0.5	0.46	ug/L	92	2			82	127	20
	alpha-Chlordane	0.5	0.45	ug/L	89	1			82	125	20
	gamma-Chlordane	0.5	0.45	ug/L	90	1			82	125	20