

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: AECO02
ProjectID: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 6
Lab Code: ACE SDG No.: Q3434
DataFile: FF016632.D Analyst Name: YP\AJ Analyst Date: 10-30-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	84691500	99637	106517	6.459