

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3434</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>10/29/2025 11:30</u>
Lab File ID:	<u>BF144121.D</u>	Init. Calib. Date(s):	<u>10/24/2025 10/24/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:13 16:24</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18 (mm)</u>

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.264	1.346		6.5	
Benzaldehyde	1.084	1.051		-3.0	
Phenol-d6	1.396	1.460		4.6	
Phenol	1.722	1.843		7.0	20.0
bis(2-Chloroethyl)ether	1.245	1.332		7.0	
2-Chlorophenol	1.245	1.284		3.1	
2-Methylphenol	0.957	1.014		6.0	
2,2-oxybis(1-Chloropropane)	2.339	2.460		5.2	
Acetophenone	0.432	0.429		-0.7	
3+4-Methylphenols	1.132	1.224		8.1	
n-Nitroso-di-n-propylamine	0.903	0.927	0.050	2.7	
Nitrobenzene-d5	0.365	0.361		-1.1	
Hexachloroethane	0.541	0.545		0.7	
Nitrobenzene	0.386	0.394		2.1	
Isophorone	0.638	0.649		1.7	
2-Nitrophenol	0.184	0.193		4.9	20.0
2,4-Dimethylphenol	0.273	0.290		6.2	
bis(2-Chloroethoxy)methane	0.398	0.409		2.8	
2,4-Dichlorophenol	0.317	0.328		3.5	20.0
Naphthalene	0.944	0.957		1.4	
4-Chloroaniline	0.329	0.349		6.1	
Hexachlorobutadiene	0.267	0.260		-2.6	20.0
Caprolactam	0.080	0.084		5.0	
4-Chloro-3-methylphenol	0.280	0.296		5.7	20.0
2-Methylnaphthalene	0.622	0.630		1.3	
Hexachlorocyclopentadiene	0.212	0.181	0.050	-14.6	
2,4,6-Trichlorophenol	0.447	0.466		4.3	20.0
2-Fluorobiphenyl	1.421	1.368		-3.7	
2,4,5-Trichlorophenol	0.469	0.490		4.5	
1,1-Biphenyl	1.432	1.430		-0.1	
2-Chloronaphthalene	1.199	1.242		3.6	
2-Nitroaniline	0.356	0.357		0.3	
Dimethylphthalate	1.311	1.355		3.4	
Acenaphthylene	1.617	1.679		3.8	
2,6-Dinitrotoluene	0.259	0.272		5.0	
3-Nitroaniline	0.288	0.311		8.0	
Acenaphthene	1.080	1.076		-0.4	20.0
2,4-Dinitrophenol	0.140	0.135	0.050	-3.6	
4-Nitrophenol	0.195	0.193	0.050	-1.0	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3434</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>10/29/2025 11:30</u>
Lab File ID:	<u>BF144121.D</u>	Init. Calib. Date(s):	<u>10/24/2025 10/24/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:13 16:24</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.499	1.529		2.0	
2,4-Dinitrotoluene	0.350	0.369		5.4	
Diethylphthalate	1.233	1.231		-0.2	
4-Chlorophenyl-phenylether	0.643	0.660		2.6	
Fluorene	1.143	1.160		1.5	
4-Nitroaniline	0.273	0.277		1.5	
4,6-Dinitro-2-methylphenol	0.128	0.128		0.0	
n-Nitrosodiphenylamine	0.639	0.669		4.7	20.0
2,4,6-Tribromophenol	0.255	0.252		-1.2	
4-Bromophenyl-phenylether	0.250	0.265		6.0	
Hexachlorobenzene	0.307	0.310		1.0	
Atrazine	0.210	0.208		-1.0	
Pentachlorophenol	0.146	0.162		11.0	20.0
Phenanthrene	0.994	1.031		3.7	
Anthracene	1.026	1.051		2.4	
Carbazole	0.874	0.888		1.6	
Di-n-butylphthalate	1.071	1.096		2.3	
Fluoranthene	1.098	1.097		-0.1	20.0
Pyrene	1.438	1.509		4.9	
Terphenyl-d14	1.144	1.166		1.9	
Butylbenzylphthalate	0.544	0.562		3.3	
3,3-Dichlorobenzidine	0.470	0.520		10.6	
Benzo (a) anthracene	1.352	1.439		6.4	
Chrysene	1.268	1.339		5.6	
Bis(2-ethylhexyl)phthalate	0.743	0.791		6.5	
Di-n-octyl phthalate	1.277	1.456		14.0	20.0
Benzo (b) fluoranthene	1.131	1.146		1.3	
Benzo (k) fluoranthene	1.043	1.131		8.4	
Benzo (a) pyrene	1.014	1.085		7.0	20.0
Indeno (1,2,3-cd) pyrene	1.373	1.487		8.3	
Dibenzo (a,h) anthracene	1.087	1.186		9.1	
Benzo (g,h,i) perylene	1.093	1.183		8.2	
1,2,4,5-Tetrachlorobenzene	0.687	0.695		1.2	
1,4-Dioxane	0.546	0.560		2.6	20.0
2,3,4,6-Tetrachlorophenol	0.330	0.352		6.7	

All other compounds must meet a minimum RRF of 0.010.