284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

DATA PACKAGE

SEMI-VOLATILE ORGANICS

PROJECT NAME : AE1-CTY 3.2.Z-PR-DOCK-DES-ODC - 60693795-1719206

AECOM

605 3rd Avenue

29th Floor

New York, NY - 10158

Phone No: 212-973-2900

ORDER ID : Q3434

ATTENTION : Rob Forstner

Laboratory Certification ID # 20012

Table Of Contents for Q3434

1) SEMI-VOLATILE DATA 2
2) Signature Page 4
3) Case Narrative 5
4) Qualifier Page 7
5) Conformance/Non Conformance 8
6) QA Checklist 10
7) Chronicle 11
8) Hit Summary 12
9) QC Data Summary For SVOC-TCL BNA -20 13
9.1) Deuterated Monitoring Compound Summary 14
9.2) MS/MSD Summary 16
9.3) LCS/LCSD Summary 20
9.4) Method Blank Summary 26
9.5) GS/MS Tune Summary 28
9.6) Internal Standard Area and RT Summary 34
10) Sample Data 42
10.1) PR132-WC1-20251022 43
10.2) PR132-WC2-20251022 59
11) Calibration Data Summary 77
11.1) Initial Calibration Data 3
11.1.1) BF102425 78
11.1.2) BG102425 481
11.2) Continued Calibration Data 890
11.2.1) BF144097.D 890
11.2.2) BF144121.D 947
11.2.3) BG064618.D 1004
11.2.4) BG064636.D 1061
12) QC Sample Data 1118
12.1) Tune Raw Data 1119
12.2) Method Blank Data 1143
12.3) LCS Data 1173
12.4) LCSD Data 1277
12.5) MS Data 1329
12.6) MSD Data 1381
13) Manual Integration 1433

14) Analytical Runlogs 1439
15) Percent Solid 1458
16) Extraction Logs 1460
16.1) PB170215.pdf 1460
16.2) PB170215IC.pdf 1462
16.3) PB170286.pdf 1463
16.4) PB170286IC.pdf 1465
17) Standard Prep Logs 1466
18) Shipping Document 1541
18.1) Chain Of Custody 1542
18.2) Lab Certificate 1543
18.3) Internal COC 1544

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q3434

Project ID : AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

Client : AECOM

Lab Sample Number	Client Sample Number
Q3434-01	PR132-WC1-20251022
Q3434-02	PR132-WC2-20251022
Q3434-03	TB

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 11/5/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

AECOM
Project Name: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
Project # N/A
Order ID # Q3434
Test Name: SVOC-TCL BNA -20

A. Number of Samples and Date of Receipt:
1 Solid sample was received on 10/22/2025.
1 Water sample was received on 10/22/2025.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: SVOC-TCL BNA -20. This data package contains results for SVOC-TCL BNA -20.

C. Analytical Techniques:
The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The samples were analyzed on instrument BNA_G using GC Column ZB-Semi Volatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:
The Holding Times were met for all analysis.
The Surrogate recoveries were met for all analysis except for PB170215BL [Nitrobenzene-d5 - 110%] marginally high therefore no corrective action was taken.

The Internal Standards Areas were met for all analysis.
The Retention Times were met for all analysis.
The MS recoveries met the requirements for all compounds.
The MSD recoveries met the requirements for all compounds.
The RPD were met for all analysis.
The Blank Spike met requirements for all compounds.
The Blank Spike Duplicate met requirements for all compounds.
The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Method 8270-BG102421.M for 2-Nitrophenol, 2-Nitroaniline, 2,6-Dinitrotoluene, 2,4-Dinitrophenol, 2,4-Dinitrotoluene, 4,6-Dinitro-2-methylphenolthese Compounds are passing on Linear regression,

The Continuous Calibration File ID BF144097.D met the requirements except for Caprolactam,Pyrene and Terphenyl-d14 , The associate samples have no positive hit for these compounds therefore no corrective action was taken.

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Continuous Calibration File ID BG064618.D met the requirements except for Nitrobenzene-d5, Di-n-octyl phthalate, marginally high, therefore no corrective action was taken, and Hexachlorocyclopentadiene failing high but the associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Continuous Calibration File ID BG064636.D met the requirements except for 2,4-Dinitrophenol,2-Nitrophenol,4,6-Dinitro-2-methylphenol, Hexachlorocyclopentadiene, Nitrobenzene and Nitrobenzene-d5, The associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:
The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.
The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:
Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value**.** This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3434 MATRIX: /Water

METHOD: 8270E/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.		✓	
	The %RSD is greater than 20% in the Method 8270-BG102421.M for 2-Nitrophenol, 2-Nitroaniline, 2,6-Dinitrotoluene, 2,4-Dinitrophenol, 2,4-Dinitrotoluene, 4,6-Dinitro-2-methylphenolthese Compounds are passing on Linear regression,			
	The Continuous Calibration File ID BF144097.D met the requirements except for Caprolactam ,Pyrene and Terphenyl-d14 , The associate samples have no positive hit for these compounds therefore no corrective action was taken.			
	The Continuous Calibration File ID BG064618.D met the requirements except for Nitrobenzene-d5, Di-n-octyl phthalate, marginally high, therefore no corrective action was taken, and Hexachlorocyclopentadiene failing high but the associate samples have no positive hit for these compounds therefore no corrective action was taken.			
	The Continuous Calibration File ID BG064636.D met the requirements except for 2,4-Dinitrophenol,2-Nitrophenol,4,6-Dinitro-2-methylphenol,Hexachlorocyclopentadiene,Nitrobenzene and Nitrobenzene-d5, The associate samples have no positive hit for these compounds therefore no corrective action was taken.			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

		NA	NO	YES
7.	Surrogate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries were met for all analysis except for PB170215BL [Nitrobenzene-d5 - 110%] marginally high therefore no corrective action was taken.			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The MS recoveries met the requirements for all compounds. The MSD recoveries met the requirements for all compounds. The Blank Spike met requirements for all compounds. The Blank Spike Duplicate met requirements for all compounds.			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The soil samples results are based on a dry weight basis.

_____ _____

QA REVIEW Date

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3434

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: SOHIL JODHANI **Date:** 11/05/2025

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q3434	**OrderDate:**	10/22/2025 1:40:00 PM
Client:	AECOM	**Project:**	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
Contact:	Rob Forstner	**Location:**	D41,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3434-01	**PR132-WC1-20251022**	**SOIL**			**10/22/25**			**10/22/25**
			SVOC-TCL BNA -20	8270E		10/23/25	10/28/25	
Q3434-02	**PR132-WC2-20251022**	**Water**			**10/22/25**			**10/22/25**
			SVOC-TCL BNA -20	8270E		10/28/25	10/29/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3434

Client: AECOM

Sample ID	Client ID	Matrix	Parameter		Concentration	C	MDL	RDL	Units
Client ID :	**PR132-WC1-20251022**								
Q3434-01	PR132-WC1-20251022	SOIL	Pyrene		120.000	J	62.2	290	ug/Kg
			Total Svoc :		**120.00**				
Q3434-01	PR132-WC1-20251022	SOIL	Benzophenone	*	330.000	J	0	0	ug/Kg
			Total Tics :		**330.00**				
			Total Concentration:		**450.00**				
Client ID :	**PR132-WC2-20251022**								
Q3434-02	PR132-WC2-20251022	WATER	1-Hexacosene	*	3.800	J	0	0	ug/L
Q3434-02	PR132-WC2-20251022	WATER	Benzophenone	*	4.100	J	0	0	ug/L
Q3434-02	PR132-WC2-20251022	WATER	Ethanol, 2-(2-butoxyethoxy)-	*	13.100	J	0	0	ug/L
Q3434-02	PR132-WC2-20251022	WATER	n-Hexadecanoic acid	*	6.600	J	0	0	ug/L
			Total Tics :		**27.60**				
			Total Concentration:		**27.60**				

QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: Q3434

Client: AECOM

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%) Low	Limits (%) High
PB170215BL	PB170215BL	2-Fluorophenol	150	139	92		18	112
		Phenol-d6	150	139	92		15	107
		Nitrobenzene-d5	100	110	110	*	18	107
		2-Fluorobiphenyl	100	96.8	97		20	109
		2,4,6-Tribromophenol	150	161	108		10	116
		Terphenyl-d14	100	90.8	91		10	105
PB170215BS	PB170215BS	2-Fluorophenol	150	125	83		18	112
		Phenol-d6	150	125	83		15	107
		Nitrobenzene-d5	100	95.4	95		18	107
		2-Fluorobiphenyl	100	79.6	80		20	109
		2,4,6-Tribromophenol	150	144	96		10	116
		Terphenyl-d14	100	83.3	83		10	105
Q3419-02MS	PR-132-S16-015094-20251021MS	2-Fluorophenol	150	95.5	64		18	112
		Phenol-d6	150	99.7	66		15	107
		Nitrobenzene-d5	100	76.9	77		18	107
		2-Fluorobiphenyl	100	60.3	60		20	109
		2,4,6-Tribromophenol	150	123	82		10	116
		Terphenyl-d14	100	60.4	60		10	105
Q3419-03MSD	PR-132-S16-015094-20251021MSD	2-Fluorophenol	150	92.9	62		18	112
		Phenol-d6	150	92.6	62		15	107
		Nitrobenzene-d5	100	74.9	75		18	107
		2-Fluorobiphenyl	100	58.4	58		20	109
		2,4,6-Tribromophenol	150	118	79		10	116
		Terphenyl-d14	100	57.1	57		10	105
Q3434-01	PR132-WC1-20251022	2-Fluorophenol	150	94.7	63		18	112
		Phenol-d6	150	99.7	66		15	107
		Nitrobenzene-d5	100	66.2	66		18	107
		2-Fluorobiphenyl	100	61.3	61		20	109
		2,4,6-Tribromophenol	150	102	68		10	116
		Terphenyl-d14	100	76.5	76		10	105

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: Q3434

Client: AECOM

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%) Low	Limits (%) High
PB170286BL	PB170286BL	2-Fluorophenol	150	133	88		23	138
		Phenol-d6	150	126	84		10	134
		Nitrobenzene-d5	100	102	102		67	132
		2-Fluorobiphenyl	100	87.4	87		52	132
		2,4,6-Tribromophenol	150	148	99		44	137
		Terphenyl-d14	100	79.4	79		42	152
PB170286BS	PB170286BS	2-Fluorophenol	150	121	81		23	138
		Phenol-d6	150	116	78		10	134
		Nitrobenzene-d5	100	93.9	94		67	132
		2-Fluorobiphenyl	100	79.9	80		52	132
		2,4,6-Tribromophenol	150	144	96		44	137
		Terphenyl-d14	100	79.5	79		42	152
PB170286BSD	PB170286BSD	2-Fluorophenol	150	122	82		23	138
		Phenol-d6	150	122	81		10	134
		Nitrobenzene-d5	100	97.0	97		67	132
		2-Fluorobiphenyl	100	79.8	80		52	132
		2,4,6-Tribromophenol	150	143	95		44	137
		Terphenyl-d14	100	80.6	81		42	152
Q3434-02	PR132-WC2-20251022	2-Fluorophenol	150	83.2	55		23	138
		Phenol-d6	150	62.0	41		10	134
		Nitrobenzene-d5	100	92.1	92		67	132
		2-Fluorobiphenyl	100	86.9	87		52	132
		2,4,6-Tribromophenol	150	138	92		44	137
		Terphenyl-d14	100	82.9	83		42	152

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** SW8270E

Client: AECOM **DataFile:** BG064627.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: Q3419-02MS		**Client Sample ID: PR-132-S16-015094-20251021MS**									
Benzaldehyde	3100	0	2100	ug/Kg	68				10	171	
Phenol	3100	0	3000	ug/Kg	97				51	122	
bis(2-Chloroethyl)ether	3100	0	2800	ug/Kg	90				54	125	
2-Chlorophenol	3100	0	3200	ug/Kg	103				51	121	
2-Methylphenol	3100	0	3100	ug/Kg	100				47	125	
2,2-oxybis(1-Chloropropane)	3100	0	2800	ug/Kg	90				46	119	
Acetophenone	3100	0	3300	ug/Kg	106				55	128	
3+4-Methylphenols	3100	0	3000	ug/Kg	97				49	125	
N-Nitroso-di-n-propylamine	3100	0	3000	ug/Kg	97				59	119	
Hexachloroethane	3100	0	3200	ug/Kg	103				51	116	
Nitrobenzene	3100	0	3400	ug/Kg	110				47	124	
Isophorone	3100	0	3000	ug/Kg	97				49	127	
2-Nitrophenol	3100	0	3800	ug/Kg	123				43	131	
2,4-Dimethylphenol	3100	0	3500	ug/Kg	113				63	151	
bis(2-Chloroethoxy)methane	3100	0	2900	ug/Kg	94				51	119	
2,4-Dichlorophenol	3100	0	3500	ug/Kg	113				50	122	
Naphthalene	3100	0	3000	ug/Kg	97				51	121	
4-Chloroaniline	3100	0	1300	ug/Kg	42				10	100	
Hexachlorobutadiene	3100	0	3200	ug/Kg	103				44	126	
Caprolactam	3100	0	3100	ug/Kg	100				51	134	
4-Chloro-3-methylphenol	3100	0	3200	ug/Kg	103				57	132	
2-Methylnaphthalene	3100	0	2800	ug/Kg	90				59	123	
Hexachlorocyclopentadiene	6100	0	9500	ug/Kg	156				10	175	
2,4,6-Trichlorophenol	3100	0	3600	ug/Kg	116				33	141	
2,4,5-Trichlorophenol	3100	0	3400	ug/Kg	110				38	135	
1,1-Biphenyl	3100	0	3400	ug/Kg	110				55	131	
2-Chloronaphthalene	3100	0	3100	ug/Kg	100				48	124	
2-Nitroaniline	3100	0	3400	ug/Kg	110				47	134	
Dimethylphthalate	3100	0	3100	ug/Kg	100				54	120	
Acenaphthylene	3100	0	3500	ug/Kg	113				57	125	
2,6-Dinitrotoluene	3100	0	3700	ug/Kg	119				48	127	
3-Nitroaniline	3100	0	2500	ug/Kg	81				10	112	
Acenaphthene	3100	0	3500	ug/Kg	113				70	121	
2,4-Dinitrophenol	6100	0	8100	ug/Kg	133				10	155	
4-Nitrophenol	6100	0	7500	ug/Kg	123				10	175	
Dibenzofuran	3100	0	3100	ug/Kg	100				52	114	
2,4-Dinitrotoluene	3100	0	3600	ug/Kg	116				41	140	
Diethylphthalate	3100	0	3100	ug/Kg	100				51	119	
4-Chlorophenyl-phenylether	3100	0	3100	ug/Kg	100				48	122	
Fluorene	3100	0	3100	ug/Kg	100				53	118	
4-Nitroaniline	3100	0	3600	ug/Kg	116				29	140	
4,6-Dinitro-2-methylphenol	3100	0	4000	ug/Kg	129				10	160	
N-Nitrosodiphenylamine	3100	0	3100	ug/Kg	100				73	118	
4-Bromophenyl-phenylether	3100	0	3200	ug/Kg	103				65	121	
Hexachlorobenzene	3100	0	3200	ug/Kg	103				67	118	
Atrazine	3100	0	3400	ug/Kg	110				45	175	

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** SW8270E

Client: AECOM **DataFile:** BG064627.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Pentachlorophenol	6100	0	7000	ug/Kg	115				13	153	
Phenanthrene	3100	0	3100	ug/Kg	100				52	128	
Anthracene	3100	0	3100	ug/Kg	100				62	124	
Carbazole	3100	0	3100	ug/Kg	100				59	119	
Di-n-butylphthalate	3100	0	3200	ug/Kg	103				55	125	
Fluoranthene	3100	0	3100	ug/Kg	100				44	125	
Pyrene	3100	0	2900	ug/Kg	94				37	122	
Butylbenzylphthalate	3100	0	3500	ug/Kg	113				44	135	
3,3-Dichlorobenzidine	3100	0	1700	ug/Kg	55				15	112	
Benzo(a)anthracene	3100	0	3100	ug/Kg	100				53	119	
Chrysene	3100	0	3100	ug/Kg	100				57	121	
bis(2-Ethylhexyl)phthalate	3100	0	3400	ug/Kg	110				42	169	
Di-n-octyl phthalate	3100	0	3500	ug/Kg	113				51	156	
Benzo(b)fluoranthene	3100	0	3100	ug/Kg	100				52	117	
Benzo(k)fluoranthene	3100	0	3200	ug/Kg	103				57	134	
Benzo(a)pyrene	3100	0	3200	ug/Kg	103				70	142	
Indeno(1,2,3-cd)pyrene	3100	0	3100	ug/Kg	100				40	129	
Dibenz(a,h)anthracene	3100	0	3200	ug/Kg	103				43	123	
Benzo(g,h,i)perylene	3100	0	3200	ug/Kg	103				24	125	
1,2,4,5-Tetrachlorobenzene	3100	0	3500	ug/Kg	113				52	134	
1,4-Dioxane	3100	0	2500	ug/Kg	81				46	112	
2,3,4,6-Tetrachlorophenol	3100	0	3800	ug/Kg	123				24	146	

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** SW8270E

Client: AECOM **DataFile:** BG064628.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: Q3419-03MSD		**Client Sample ID: PR-132-S16-015094-20251021MSD**									
Benzaldehyde	3100	0	2100	ug/Kg	68	0			10	171	20
Phenol	3100	0	2900	ug/Kg	94	3			51	122	20
bis(2-Chloroethyl)ether	3100	0	2600	ug/Kg	84	7			54	125	20
2-Chlorophenol	3100	0	3000	ug/Kg	97	6			51	121	20
2-Methylphenol	3100	0	2900	ug/Kg	94	6			47	125	20
2,2-oxybis(1-Chloropropane)	3100	0	2700	ug/Kg	87	3			46	119	20
Acetophenone	3100	0	3100	ug/Kg	100	6			55	128	20
3+4-Methylphenols	3100	0	2900	ug/Kg	94	3			49	125	20
N-Nitroso-di-n-propylamine	3100	0	2700	ug/Kg	87	11			59	119	20
Hexachloroethane	3100	0	3000	ug/Kg	97	6			51	116	20
Nitrobenzene	3100	0	3300	ug/Kg	106	4			47	124	20
Isophorone	3100	0	2800	ug/Kg	90	7			49	127	20
2-Nitrophenol	3100	0	3800	ug/Kg	123	0			43	131	20
2,4-Dimethylphenol	3100	0	3400	ug/Kg	110	3			63	151	20
bis(2-Chloroethoxy)methane	3100	0	2800	ug/Kg	90	4			51	119	20
2,4-Dichlorophenol	3100	0	3300	ug/Kg	106	6			50	122	20
Naphthalene	3100	0	2900	ug/Kg	94	3			51	121	20
4-Chloroaniline	3100	0	1200	ug/Kg	39	7			10	100	20
Hexachlorobutadiene	3100	0	3100	ug/Kg	100	3			44	126	20
Caprolactam	3100	0	2800	ug/Kg	90	11			51	134	20
4-Chloro-3-methylphenol	3100	0	3100	ug/Kg	100	3			57	132	20
2-Methylnaphthalene	3100	0	2700	ug/Kg	87	3			59	123	20
Hexachlorocyclopentadiene	6100	0	8900	ug/Kg	146	7			10	175	20
2,4,6-Trichlorophenol	3100	0	3500	ug/Kg	113	3			33	141	20
2,4,5-Trichlorophenol	3100	0	3400	ug/Kg	110	0			38	135	20
1,1-Biphenyl	3100	0	3300	ug/Kg	106	4			55	131	20
2-Chloronaphthalene	3100	0	3000	ug/Kg	97	3			48	124	20
2-Nitroaniline	3100	0	3300	ug/Kg	106	4			47	134	20
Dimethylphthalate	3100	0	2900	ug/Kg	94	6			54	120	20
Acenaphthylene	3100	0	3300	ug/Kg	106	6			57	125	20
2,6-Dinitrotoluene	3100	0	3600	ug/Kg	116	3			48	127	20
3-Nitroaniline	3100	0	2400	ug/Kg	77	5			10	112	20
Acenaphthene	3100	0	3400	ug/Kg	110	3			70	121	20
2,4-Dinitrophenol	6100	0	7800	ug/Kg	128	4			10	155	20
4-Nitrophenol	6100	0	6900	ug/Kg	113	8			10	175	20
Dibenzofuran	3100	0	2900	ug/Kg	94	6			52	114	20
2,4-Dinitrotoluene	3100	0	3300	ug/Kg	106	9			41	140	20
Diethylphthalate	3100	0	2900	ug/Kg	94	6			51	119	20
4-Chlorophenyl-phenylether	3100	0	3000	ug/Kg	97	3			48	122	20
Fluorene	3100	0	3000	ug/Kg	97	3			53	118	20
4-Nitroaniline	3100	0	3500	ug/Kg	113	3			29	140	20
4,6-Dinitro-2-methylphenol	3100	0	3900	ug/Kg	126	2			10	160	20
N-Nitrosodiphenylamine	3100	0	3100	ug/Kg	100	0			73	118	20
4-Bromophenyl-phenylether	3100	0	3000	ug/Kg	97	6			65	121	20
Hexachlorobenzene	3100	0	3100	ug/Kg	100	3			67	118	20
Atrazine	3100	0	3300	ug/Kg	106	4			45	175	20

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** SW8270E

Client: AECOM **DataFile:** BG064628.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Pentachlorophenol	6100	0	6800	ug/Kg	111	4			13	153	20
Phenanthrene	3100	0	3000	ug/Kg	97	3			52	128	20
Anthracene	3100	0	2900	ug/Kg	94	6			62	124	20
Carbazole	3100	0	3000	ug/Kg	97	3			59	119	20
Di-n-butylphthalate	3100	0	3100	ug/Kg	100	3			55	125	20
Fluoranthene	3100	0	3100	ug/Kg	100	0			44	125	20
Pyrene	3100	0	2700	ug/Kg	87	8			37	122	20
Butylbenzylphthalate	3100	0	3400	ug/Kg	110	3			44	135	20
3,3-Dichlorobenzidine	3100	0	1700	ug/Kg	55	0			15	112	20
Benzo(a)anthracene	3100	0	3000	ug/Kg	97	3			53	119	20
Chrysene	3100	0	2900	ug/Kg	94	6			57	121	20
bis(2-Ethylhexyl)phthalate	3100	0	3400	ug/Kg	110	0			42	169	20
Di-n-octyl phthalate	3100	0	3300	ug/Kg	106	6			51	156	20
Benzo(b)fluoranthene	3100	0	3000	ug/Kg	97	3			52	117	20
Benzo(k)fluoranthene	3100	0	3000	ug/Kg	97	6			57	134	20
Benzo(a)pyrene	3100	0	3100	ug/Kg	100	3			70	142	20
Indeno(1,2,3-cd)pyrene	3100	0	3000	ug/Kg	97	3			40	129	20
Dibenz(a,h)anthracene	3100	0	3000	ug/Kg	97	6			43	123	20
Benzo(g,h,i)perylene	3100	0	3100	ug/Kg	100	3			24	125	20
1,2,4,5-Tetrachlorobenzene	3100	0	3400	ug/Kg	110	3			52	134	20
1,4-Dioxane	3100	0	2300	ug/Kg	74	9			46	112	20
2,3,4,6-Tetrachlorophenol	3100	0	3600	ug/Kg	116	6			24	146	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** 8270E

Client: AECOM **DataFile:** BG064622.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB170215BS	Benzaldehyde	1700	1100	ug/Kg	65				10	133	
	Phenol	1700	1500	ug/Kg	88				62	112	
	bis(2-Chloroethyl)ether	1700	1300	ug/Kg	76				60	101	
	2-Chlorophenol	1700	1600	ug/Kg	94				65	112	
	2-Methylphenol	1700	1500	ug/Kg	88				61	108	
	2,2-oxybis(1-Chloropropane)	1700	1300	ug/Kg	76				51	100	
	Acetophenone	1700	1600	ug/Kg	94				66	98	
	3+4-Methylphenols	1700	1500	ug/Kg	88				58	111	
	N-Nitroso-di-n-propylamine	1700	1400	ug/Kg	82				63	95	
	Hexachloroethane	1700	1500	ug/Kg	88				72	108	
	Nitrobenzene	1700	1600	ug/Kg	94				57	101	
	Isophorone	1700	1400	ug/Kg	82				59	99	
	2-Nitrophenol	1700	1700	ug/Kg	100				61	111	
	2,4-Dimethylphenol	1700	1700	ug/Kg	100				46	141	
	bis(2-Chloroethoxy)methane	1700	1400	ug/Kg	82				66	97	
	2,4-Dichlorophenol	1700	1600	ug/Kg	94				62	107	
	Naphthalene	1700	1400	ug/Kg	82				62	100	
	4-Chloroaniline	1700	790	ug/Kg	46				16	100	
	Hexachlorobutadiene	1700	1500	ug/Kg	88				53	98	
	Caprolactam	1700	1500	ug/Kg	88				67	110	
	4-Chloro-3-methylphenol	1700	1600	ug/Kg	94				58	112	
	2-Methylnaphthalene	1700	1300	ug/Kg	76				60	104	
	Hexachlorocyclopentadiene	3300	4100	ug/Kg	124				45	165	
	2,4,6-Trichlorophenol	1700	1700	ug/Kg	100				59	102	
	2,4,5-Trichlorophenol	1700	1600	ug/Kg	94				61	98	
	1,1-Biphenyl	1700	1600	ug/Kg	94				57	103	
	2-Chloronaphthalene	1700	1400	ug/Kg	82				58	99	
	2-Nitroaniline	1700	1600	ug/Kg	94				66	101	
	Dimethylphthalate	1700	1500	ug/Kg	88				61	99	
	Acenaphthylene	1700	1700	ug/Kg	100				63	101	
	2,6-Dinitrotoluene	1700	1700	ug/Kg	100				61	104	
	3-Nitroaniline	1700	1200	ug/Kg	71				28	100	
	Acenaphthene	1700	1600	ug/Kg	94				57	104	
	2,4-Dinitrophenol	3300	3600	ug/Kg	109				37	128	
	4-Nitrophenol	3300	3400	ug/Kg	103				48	119	
	Dibenzofuran	1700	1500	ug/Kg	88				63	99	
	2,4-Dinitrotoluene	1700	1700	ug/Kg	100				60	106	
	Diethylphthalate	1700	1500	ug/Kg	88				60	101	
	4-Chlorophenyl-phenylether	1700	1500	ug/Kg	88				58	98	
	Fluorene	1700	1500	ug/Kg	88				61	101	
	4-Nitroaniline	1700	1700	ug/Kg	100				64	103	
	4,6-Dinitro-2-methylphenol	1700	1800	ug/Kg	106				76	113	
	N-Nitrosodiphenylamine	1700	1500	ug/Kg	88				71	99	
	4-Bromophenyl-phenylether	1700	1500	ug/Kg	88				66	102	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** 8270E

Client: AECOM **DataFile:** BG064622.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB170215BS	Hexachlorobenzene	1700	1500	ug/Kg	88				64	98	
	Atrazine	1700	1600	ug/Kg	94				47	152	
	Pentachlorophenol	3300	3300	ug/Kg	100				67	105	
	Phenanthrene	1700	1500	ug/Kg	88				59	103	
	Anthracene	1700	1500	ug/Kg	88				61	105	
	Carbazole	1700	1500	ug/Kg	88				61	99	
	Di-n-butylphthalate	1700	1500	ug/Kg	88				58	104	
	Fluoranthene	1700	1500	ug/Kg	88				57	107	
	Pyrene	1700	1400	ug/Kg	82				59	103	
	Butylbenzylphthalate	1700	1600	ug/Kg	94				55	103	
	3,3-Dichlorobenzidine	1700	1100	ug/Kg	65				42	91	
	Benzo(a)anthracene	1700	1500	ug/Kg	88				60	102	
	Chrysene	1700	1500	ug/Kg	88				59	101	
	bis(2-Ethylhexyl)phthalate	1700	1600	ug/Kg	94				54	135	
	Di-n-octyl phthalate	1700	1700	ug/Kg	100				52	137	
	Benzo(b)fluoranthene	1700	1500	ug/Kg	88				62	109	
	Benzo(k)fluoranthene	1700	1500	ug/Kg	88				62	109	
	Benzo(a)pyrene	1700	1600	ug/Kg	94				63	103	
	Indeno(1,2,3-cd)pyrene	1700	1500	ug/Kg	88				63	101	
	Dibenz(a,h)anthracene	1700	1500	ug/Kg	88				61	112	
	Benzo(g,h,i)perylene	1700	1500	ug/Kg	88				70	108	
	1,2,4,5-Tetrachlorobenzene	1700	1600	ug/Kg	94				53	101	
	1,4-Dioxane	1700	1100	ug/Kg	65				50	96	
	2,3,4,6-Tetrachlorophenol	1700	1800	ug/Kg	106				59	108	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** 8270E

Client: AECOM **DataFile:** BG064638.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	Limits RPD
PB170286BS	Benzaldehyde	50	11.4	ug/L	23				10	162	
	Phenol	50	38.9	ug/L	78				66	118	
	bis(2-Chloroethyl)ether	50	35.9	ug/L	72				62	103	
	2-Chlorophenol	50	43.7	ug/L	87				70	117	
	2-Methylphenol	50	40.8	ug/L	82				69	109	
	2,2-oxybis(1-Chloropropane)	50	35.8	ug/L	72				65	100	
	Acetophenone	50	40.5	ug/L	81				60	104	
	3+4-Methylphenols	50	38.9	ug/L	78				67	106	
	N-Nitroso-di-n-propylamine	50	38.1	ug/L	76				57	107	
	Hexachloroethane	50	40.6	ug/L	81				76	118	
	Nitrobenzene	50	43.9	ug/L	88				58	106	
	Isophorone	50	37.1	ug/L	74				61	102	
	2-Nitrophenol	50	47.8	ug/L	96				70	115	
	2,4-Dimethylphenol	50	45.1	ug/L	90				42	142	
	bis(2-Chloroethoxy)methane	50	38.1	ug/L	76				58	109	
	2,4-Dichlorophenol	50	44.2	ug/L	88				66	115	
	Naphthalene	50	38.0	ug/L	76				64	107	
	4-Chloroaniline	50	19.7	ug/L	39				10	85	
	Hexachlorobutadiene	50	40.7	ug/L	81				69	101	
	Caprolactam	50	39.1	ug/L	78				58	128	
	4-Chloro-3-methylphenol	50	41.8	ug/L	84				65	114	
	2-Methylnaphthalene	50	34.6	ug/L	69				64	107	
	Hexachlorocyclopentadiene	100	120	ug/L	120				36	160	
	2,4,6-Trichlorophenol	50	47.1	ug/L	94				61	110	
	2,4,5-Trichlorophenol	50	44.9	ug/L	90				70	106	
	1,1-Biphenyl	50	42.2	ug/L	84				72	98	
	2-Chloronaphthalene	50	39.4	ug/L	79				59	106	
	2-Nitroaniline	50	45.5	ug/L	91				73	114	
	Dimethylphthalate	50	40.5	ug/L	81				64	103	
	Acenaphthylene	50	44.6	ug/L	89				79	103	
	2,6-Dinitrotoluene	50	50.0	ug/L	100				64	110	
	3-Nitroaniline	50	33.5	ug/L	67				28	100	
	Acenaphthene	50	45.1	ug/L	90				59	113	
	2,4-Dinitrophenol	100	110	ug/L	110				36	166	
	4-Nitrophenol	100	95.8	ug/L	96				45	147	
	Dibenzofuran	50	40.5	ug/L	81				65	106	
	2,4-Dinitrotoluene	50	48.4	ug/L	97				60	115	
	Diethylphthalate	50	39.9	ug/L	80				63	105	
	4-Chlorophenyl-phenylether	50	40.7	ug/L	81				61	104	
	Fluorene	50	41.0	ug/L	82				64	107	
	4-Nitroaniline	50	47.7	ug/L	95				55	125	
	4,6-Dinitro-2-methylphenol	50	53.3	ug/L	107				62	132	
	N-Nitrosodiphenylamine	50	41.6	ug/L	83				61	109	
	4-Bromophenyl-phenylether	50	41.5	ug/L	83				73	103	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** 8270E

Client: AECOM **DataFile:** BG064638.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB170286BS	Hexachlorobenzene	50	41.8	ug/L	84				73	106	
	Atrazine	50	42.1	ug/L	84				76	120	
	Pentachlorophenol	100	93.8	ug/L	94				47	114	
	Phenanthrene	50	40.4	ug/L	81				62	109	
	Anthracene	50	40.2	ug/L	80				65	110	
	Carbazole	50	41.3	ug/L	83				62	106	
	Di-n-butylphthalate	50	39.4	ug/L	79				64	106	
	Fluoranthene	50	41.6	ug/L	83				64	110	
	Pyrene	50	38.0	ug/L	76				71	103	
	Butylbenzylphthalate	50	43.4	ug/L	87				61	105	
	3,3-Dichlorobenzidine	50	30.5	ug/L	61				43	108	
	Benzo(a)anthracene	50	40.7	ug/L	81				62	107	
	Chrysene	50	40.2	ug/L	80				61	108	
	bis(2-Ethylhexyl)phthalate	50	42.9	ug/L	86				59	110	
	Di-n-octyl phthalate	50	44.5	ug/L	89				52	139	
	Benzo(b)fluoranthene	50	43.3	ug/L	87				77	113	
	Benzo(k)fluoranthene	50	43.1	ug/L	86				77	105	
	Benzo(a)pyrene	50	45.1	ug/L	90				72	131	
	Indeno(1,2,3-cd)pyrene	50	43.3	ug/L	87				72	105	
	Dibenz(a,h)anthracene	50	43.7	ug/L	87				78	115	
	Benzo(g,h,i)perylene	50	45.2	ug/L	90				75	118	
	1,2,4,5-Tetrachlorobenzene	50	43.1	ug/L	86				72	101	
	1,4-Dioxane	50	30.4	ug/L	61				38	125	
	2,3,4,6-Tetrachlorophenol	50	50.0	ug/L	100				63	116	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** 8270E

Client: AECOM **DataFile:** BG064639.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB170286BSD	Benzaldehyde	50	12.8	ug/L	26				10	162	20
	Phenol	50	40.5	ug/L	81				66	118	20
	bis(2-Chloroethyl)ether	50	36.7	ug/L	73				62	103	20
	2-Chlorophenol	50	42.6	ug/L	85				70	117	20
	2-Methylphenol	50	41.6	ug/L	83				69	109	20
	2,2-oxybis(1-Chloropropane)	50	36.9	ug/L	74				65	100	20
	Acetophenone	50	42.7	ug/L	85				60	104	20
	3+4-Methylphenols	50	40.6	ug/L	81				67	106	20
	N-Nitroso-di-n-propylamine	50	40.3	ug/L	81				57	107	20
	Hexachloroethane	50	42.4	ug/L	85				76	118	20
	Nitrobenzene	50	45.1	ug/L	90				58	106	20
	Isophorone	50	39.1	ug/L	78				61	102	20
	2-Nitrophenol	50	51.5	ug/L	103				70	115	20
	2,4-Dimethylphenol	50	47.6	ug/L	95				42	142	20
	bis(2-Chloroethoxy)methane	50	39.6	ug/L	79				58	109	20
	2,4-Dichlorophenol	50	45.9	ug/L	92				66	115	20
	Naphthalene	50	39.4	ug/L	79				64	107	20
	4-Chloroaniline	50	21.2	ug/L	42				10	85	20
	Hexachlorobutadiene	50	43.1	ug/L	86				69	101	20
	Caprolactam	50	40.3	ug/L	81				58	128	20
	4-Chloro-3-methylphenol	50	43.6	ug/L	87				65	114	20
	2-Methylnaphthalene	50	36.4	ug/L	73				64	107	20
	Hexachlorocyclopentadiene	100	120	ug/L	120				36	160	20
	2,4,6-Trichlorophenol	50	46.3	ug/L	93				61	110	20
	2,4,5-Trichlorophenol	50	44.9	ug/L	90				70	106	20
	1,1-Biphenyl	50	41.0	ug/L	82				72	98	20
	2-Chloronaphthalene	50	38.9	ug/L	78				59	106	20
	2-Nitroaniline	50	45.9	ug/L	92				73	114	20
	Dimethylphthalate	50	40.2	ug/L	80				64	103	20
	Acenaphthylene	50	44.5	ug/L	89				79	103	20
	2,6-Dinitrotoluene	50	49.0	ug/L	98				64	110	20
	3-Nitroaniline	50	33.1	ug/L	66				28	100	20
	Acenaphthene	50	45.3	ug/L	91				59	113	20
	2,4-Dinitrophenol	100	110	ug/L	110				36	166	20
	4-Nitrophenol	100	95.8	ug/L	96				45	147	20
	Dibenzofuran	50	39.8	ug/L	80				65	106	20
	2,4-Dinitrotoluene	50	47.9	ug/L	96				60	115	20
	Diethylphthalate	50	38.7	ug/L	77				63	105	20
	4-Chlorophenyl-phenylether	50	40.1	ug/L	80				61	104	20
	Fluorene	50	40.0	ug/L	80				64	107	20
	4-Nitroaniline	50	47.3	ug/L	95				55	125	20
	4,6-Dinitro-2-methylphenol	50	55.1	ug/L	110				62	132	20
	N-Nitrosodiphenylamine	50	41.8	ug/L	84				61	109	20
	4-Bromophenyl-phenylether	50	41.3	ug/L	83				73	103	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3434 **Analytical Method:** 8270E

Client: AECOM **DataFile:** BG064639.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	Limits RPD
PB170286BSD	Hexachlorobenzene	50	41.3	ug/L	83				73	106	20
	Atrazine	50	42.2	ug/L	84				76	120	20
	Pentachlorophenol	100	93.3	ug/L	93				47	114	20
	Phenanthrene	50	40.3	ug/L	81				62	109	20
	Anthracene	50	40.1	ug/L	80				65	110	20
	Carbazole	50	41.3	ug/L	83				62	106	20
	Di-n-butylphthalate	50	39.2	ug/L	78				64	106	20
	Fluoranthene	50	41.3	ug/L	83				64	110	20
	Pyrene	50	37.9	ug/L	76				71	103	20
	Butylbenzylphthalate	50	43.3	ug/L	87				61	105	20
	3,3-Dichlorobenzidine	50	31.1	ug/L	62				43	108	20
	Benzo(a)anthracene	50	40.7	ug/L	81				62	107	20
	Chrysene	50	40.5	ug/L	81				61	108	20
	bis(2-Ethylhexyl)phthalate	50	42.3	ug/L	85				59	110	20
	Di-n-octyl phthalate	50	44.3	ug/L	89				52	139	20
	Benzo(b)fluoranthene	50	43.9	ug/L	88				77	113	20
	Benzo(k)fluoranthene	50	43.0	ug/L	86				77	105	20
	Benzo(a)pyrene	50	44.8	ug/L	90				72	131	20
	Indeno(1,2,3-cd)pyrene	50	43.3	ug/L	87				72	105	20
	Dibenz(a,h)anthracene	50	44.4	ug/L	89				78	115	20
	Benzo(g,h,i)perylene	50	44.6	ug/L	89				75	118	20
	1,2,4,5-Tetrachlorobenzene	50	42.2	ug/L	84				72	101	20
	1,4-Dioxane	50	31.6	ug/L	63				38	125	20
	2,3,4,6-Tetrachlorophenol	50	49.8	ug/L	100				63	116	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID: PB170215BL

Lab Name: Alliance | Contract: AECO02

Lab Code: ACE | SDG NO.: Q3434

Lab File ID: BG064621.D | Lab Sample ID: PB170215BL

Instrument ID: BNA_G | Date Extracted: 10/23/2025

Matrix: (soil/water) SOIL | Date Analyzed: 10/29/2025

Level: (low/med) LOW | Time Analyzed: 13:25

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB170215BS	PB170215BS	BG064622.D	10/29/2025
PR-132-S16-015094-20251021MS	Q3419-02MS	BG064627.D	10/29/2025
PR-132-S16-015094-20251021MSD	Q3419-03MSD	BG064628.D	10/29/2025
PR132-WC1-20251022	Q3434-01	BF144106.D	10/28/2025

COMMENTS:

Form IV SV

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID: PB170286BL

Lab Name: Alliance | Contract: AECO02

Lab Code: ACE | SDG NO.: Q3434

Lab File ID: BG064637.D | Lab Sample ID: PB170286BL

Instrument ID: BNA_G | Date Extracted: 10/28/2025

Matrix: (soil/water) Water | Date Analyzed: 10/30/2025

Level: (low/med) LOW | Time Analyzed: 12:17

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB170286BS	PB170286BS	BG064638.D	10/30/2025
PB170286BSD	PB170286BSD	BG064639.D	10/30/2025
PR132-WC2-20251022	Q3434-02	BF144132.D	10/29/2025

COMMENTS:

Form IV SV

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3434

Lab File ID: BF144055.D DFTPP Injection Date: 10/24/2025

Instrument ID: BNA_F DFTPP Injection Time: 09:45

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.0 (0.0) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.1
365	Greater than 1% of mass 198	4.5
441	Present, but less than mass 443	89.0
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.8 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF144056.D	10/24/2025	10:13
SSTDICC005	SSTDICC005	BF144057.D	10/24/2025	10:42
SSTDICC010	SSTDICC010	BF144058.D	10/24/2025	11:39
SSTDICC020	SSTDICC020	BF144059.D	10/24/2025	12:36
SSTDICCC040	SSTDICCC040	BF144060.D	10/24/2025	13:05
SSTDICC050	SSTDICC050	BF144061.D	10/24/2025	13:34
SSTDICC080	SSTDICC080	BF144063.D	10/24/2025	14:32
SSTDICC060	SSTDICC060	BF144064.D	10/24/2025	16:24

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3434

Lab File ID: BF144096.D DFTPP Injection Date: 10/28/2025

Instrument ID: BNA_F DFTPP Injection Time: 12:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (1.7) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.0 (0.0) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	88.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.5 (18.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF144097.D	10/28/2025	13:17
PR132-WC1-20251022	Q3434-01	BF144106.D	10/28/2025	17:44

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3434

Lab File ID: BF144120.D DFTPP Injection Date: 10/29/2025

Instrument ID: BNA_F DFTPP Injection Time: 11:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.3 (1.3) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.0 (0.0) 1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	84.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18 (18) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF144121.D	10/29/2025	11:30
PR132-WC2-20251022	Q3434-02	BF144132.D	10/29/2025	17:02

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance | Contract: AECO02

Lab Code: ACE | SDG NO.: Q3434

Lab File ID: BG064568.D | DFTPP Injection Date: 10/24/2025

Instrument ID: BNA_G | DFTPP Injection Time: 08:20

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	35.3
70	Less than 2.0% of mass 69	0.1 (0.2) 1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.2
365	Greater than 1% of mass 198	5.1
441	Present, but less than mass 443	15.1
442	Greater than 50% of mass 198	92.7
443	15.0 - 24.0% of mass 442	17.5 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BG064569.D	10/24/2025	09:00
SSTDICC005	SSTDICC005	BG064570.D	10/24/2025	09:41
SSTDICC010	SSTDICC010	BG064571.D	10/24/2025	11:02
SSTDICC020	SSTDICC020	BG064572.D	10/24/2025	12:23
SSTDICCC040	SSTDICCC040	BG064573.D	10/24/2025	13:03
SSTDICC060	SSTDICC060	BG064574.D	10/24/2025	13:44
SSTDICC080	SSTDICC080	BG064575.D	10/24/2025	14:24
SSTDICC050	SSTDICC050	BG064576.D	10/24/2025	15:32

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3434

Lab File ID: BG064617.D DFTPP Injection Date: 10/29/2025

Instrument ID: BNA_G DFTPP Injection Time: 10:41

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.3 (1) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.3
365	Greater than 1% of mass 198	5.3
441	Present, but less than mass 443	91.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.5 (18) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG064618.D	10/29/2025	11:22
PB170215BL	PB170215BL	BG064621.D	10/29/2025	13:25
PB170215BS	PB170215BS	BG064622.D	10/29/2025	14:06
PR-132-S16-015094-20251021MS	Q3419-02MS	BG064627.D	10/29/2025	17:34
PR-132-S16-015094-20251021MSD	Q3419-03MSD	BG064628.D	10/29/2025	18:15

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG NO.: Q3434

Lab File ID: BG064635.D DFTPP Injection Date: 10/30/2025

Instrument ID: BNA_G DFTPP Injection Time: 10:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.2 (0.7) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.0 (0.0) 1
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.3
365	Greater than 1% of mass 198	5.1
441	Present, but less than mass 443	80.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.8 (20) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG064636.D	10/30/2025	11:36
PB170286BL	PB170286BL	BG064637.D	10/30/2025	12:17
PB170286BS	PB170286BS	BG064638.D	10/30/2025	12:58
PB170286BSD	PB170286BSD	BG064639.D	10/30/2025	13:39

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q3434

Client ID : SSTDCCC040 Date Analyzed: 10/28/2025

Lab File ID: BF144097.D Time Analyzed: 13:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	66521	6.881	250146	8.16	147935	9.92
	UPPER LIMIT	133042	7.381	500292	8.663	295870	10.416
	LOWER LIMIT	33260.5	6.381	125073	7.663	73967.5	9.416
	EPA SAMPLE NO.						
01	PR132-WC1-20251022	65453	6.88	248016	8.16	145676	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q3434

Client ID: SSTDCCC040 Date Analyzed: 10/28/2025

Lab File ID: BF144097.D Time Analyzed: 13:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	264042	11.41	128248	14.057	169914	15.539
	UPPER LIMIT	528084	11.91	256496	14.557	339828	16.039
	LOWER LIMIT	132021	10.91	64124	13.557	84957	15.039
	EPA SAMPLE NO.						
01	PR132-WC1-20251022	242686	11.40	126847	14.05	179897	15.53

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q3434

Client ID : SSTDCCC040 Date Analyzed: 10/29/2025

Lab File ID: BF144121.D Time Analyzed: 11:30

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	61447	6.881	232371	8.16	124035	9.92
	UPPER LIMIT	122894	7.381	464742	8.663	248070	10.416
	LOWER LIMIT	30723.5	6.381	116186	7.663	62017.5	9.416
	EPA SAMPLE NO.						
01	PR132-WC2-20251022	59068	6.88	222374	8.16	127295	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q3434

Client ID: SSTDCCC040 Date Analyzed: 10/29/2025

Lab File ID: BF144121.D Time Analyzed: 11:30

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	207074	11.41	153131	14.057	204043	15.539
	UPPER LIMIT	414148	11.91	306262	14.557	408086	16.039
	LOWER LIMIT	103537	10.91	76565.5	13.557	102022	15.039
	EPA SAMPLE NO.						
01	PR132-WC2-20251022	211645	11.40	165479	14.05	206243	15.53

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q3434

Client ID : SSTDCCC040 Date Analyzed: 10/29/2025

Lab File ID: BG064618.D Time Analyzed: 11:22

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	36613	8.223	148961	11.04	108882	14.84
	UPPER LIMIT	73226	8.723	297922	11.543	217764	15.339
	LOWER LIMIT	18306.5	7.723	74480.5	10.543	54441	14.339
	EPA SAMPLE NO.						
01	PB170215BL	33974	8.22	131027	11.04	98338	14.84
02	PB170215BS	30857	8.22	126314	11.04	102616	14.84
03	PR-132-S16-015094-20251021MS	30488	8.22	121320	11.04	93614	14.84
04	PR-132-S16-015094-20251021MSD	31803	8.21	124534	11.04	94989	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q3434

Client ID: SSTDCCC040 Date Analyzed: 10/29/2025

Lab File ID: BG064618.D Time Analyzed: 11:22

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	241683	17.577	261016	21.854	298207	25.197
	UPPER LIMIT	483366	18.077	522032	22.354	596414	25.697
	LOWER LIMIT	120842	17.077	130508	21.354	149104	24.697
	EPA SAMPLE NO.						
01	PB170215BL	248117	17.58	318243	21.85	369003	25.19
02	PB170215BS	252758	17.58	286271	21.85	311364	25.19
03	PR-132-S16-015094-20251021I	229619	17.58	271892	21.85	296767	25.19
04	PR-132-S16-015094-20251021I	230638	17.58	282250	21.85	308600	25.19

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q3434

Client ID : SSTDCCC040 Date Analyzed: 10/30/2025

Lab File ID: BG064636.D Time Analyzed: 11:36

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	32501	8.219	127691	11.05	94503	14.84
	UPPER LIMIT	65002	8.719	255382	11.545	189006	15.341
	LOWER LIMIT	16250.5	7.719	63845.5	10.545	47251.5	14.341
	EPA SAMPLE NO.						
01	PB170286BL	25136	8.22	99198	11.04	73729	14.84
02	PB170286BS	30784	8.22	124677	11.04	97021	14.84
03	PB170286BSD	30191	8.22	121649	11.04	100532	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q3434

Client ID: SSTDCCC040 Date Analyzed: 10/30/2025

Lab File ID: BG064636.D Time Analyzed: 11:36

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	210658	17.579	257813	21.856	295961	25.199
	UPPER LIMIT	421316	18.079	515626	22.356	591922	25.699
	LOWER LIMIT	105329	17.079	128907	21.356	147981	24.699
	EPA SAMPLE NO.						
01	PB170286BL	191571	17.58	261407	21.85	302809	25.19
02	PB170286BS	242858	17.58	288177	21.85	302049	25.20
03	PB170286BSD	246375	17.58	288758	21.86	302411	25.19

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

SAMPLE DATA

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/22/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/22/25
Client Sample ID:	PR132-WC1-20251022	SDG No.:	Q3434
Lab Sample ID:	Q3434-01	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	57.9
Sample Wt/Vol:	30.01 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	570	U	1	270	570	ug/Kg	10/28/25 17:44	PB170215
108-95-2	Phenol	290	U	1	38.2	290	ug/Kg	10/28/25 17:44	PB170215
111-44-4	bis(2-Chloroethyl)ether	290	U	1	42.0	290	ug/Kg	10/28/25 17:44	PB170215
95-57-8	2-Chlorophenol	290	U	1	42.1	290	ug/Kg	10/28/25 17:44	PB170215
95-48-7	2-Methylphenol	290	U	1	51.6	290	ug/Kg	10/28/25 17:44	PB170215
108-60-1	2,2-oxybis(1-Chloropropane)	290	U	1	64.7	290	ug/Kg	10/28/25 17:44	PB170215
98-86-2	Acetophenone	290	U	1	50.9	290	ug/Kg	10/28/25 17:44	PB170215
65794-96-9	3+4-Methylphenols	570	U	1	71.0	570	ug/Kg	10/28/25 17:44	PB170215
621-64-7	n-Nitroso-di-n-propylamine	140	U	1	81.8	140	ug/Kg	10/28/25 17:44	PB170215
67-72-1	Hexachloroethane	290	U	1	30.4	290	ug/Kg	10/28/25 17:44	PB170215
98-95-3	Nitrobenzene	290	U	1	31.6	290	ug/Kg	10/28/25 17:44	PB170215
78-59-1	Isophorone	290	U	1	56.6	290	ug/Kg	10/28/25 17:44	PB170215
88-75-5	2-Nitrophenol	290	U	1	100	290	ug/Kg	10/28/25 17:44	PB170215
105-67-9	2,4-Dimethylphenol	290	U	1	110	290	ug/Kg	10/28/25 17:44	PB170215
111-91-1	bis(2-Chloroethoxy)methane	290	U	1	53.2	290	ug/Kg	10/28/25 17:44	PB170215
120-83-2	2,4-Dichlorophenol	290	U	1	48.9	290	ug/Kg	10/28/25 17:44	PB170215
91-20-3	Naphthalene	290	U	1	39.2	290	ug/Kg	10/28/25 17:44	PB170215
106-47-8	4-Chloroaniline	290	U	1	61.1	290	ug/Kg	10/28/25 17:44	PB170215
87-68-3	Hexachlorobutadiene	290	U	1	43.7	290	ug/Kg	10/28/25 17:44	PB170215
105-60-2	Caprolactam	570	U	1	90.0	570	ug/Kg	10/28/25 17:44	PB170215
59-50-7	4-Chloro-3-methylphenol	290	U	1	49.6	290	ug/Kg	10/28/25 17:44	PB170215
91-57-6	2-Methylnaphthalene	290	U	1	44.2	290	ug/Kg	10/28/25 17:44	PB170215
77-47-4	Hexachlorocyclopentadiene	570	U	1	200	570	ug/Kg	10/28/25 17:44	PB170215
88-06-2	2,4,6-Trichlorophenol	290	U	1	34.2	290	ug/Kg	10/28/25 17:44	PB170215
95-95-4	2,4,5-Trichlorophenol	290	U	1	50.2	290	ug/Kg	10/28/25 17:44	PB170215
92-52-4	1,1-Biphenyl	290	U	1	37.6	290	ug/Kg	10/28/25 17:44	PB170215
91-58-7	2-Chloronaphthalene	290	U	1	38.8	290	ug/Kg	10/28/25 17:44	PB170215
88-74-4	2-Nitroaniline	290	U	1	83.0	290	ug/Kg	10/28/25 17:44	PB170215
131-11-3	Dimethylphthalate	290	U	1	46.8	290	ug/Kg	10/28/25 17:44	PB170215
208-96-8	Acenaphthylene	290	U	1	49.9	290	ug/Kg	10/28/25 17:44	PB170215
606-20-2	2,6-Dinitrotoluene	290	U	1	58.0	290	ug/Kg	10/28/25 17:44	PB170215
99-09-2	3-Nitroaniline	290	U	1	79.4	290	ug/Kg	10/28/25 17:44	PB170215
83-32-9	Acenaphthene	290	U	1	36.8	290	ug/Kg	10/28/25 17:44	PB170215
51-28-5	2,4-Dinitrophenol	570	U	1	400	570	ug/Kg	10/28/25 17:44	PB170215
100-02-7	4-Nitrophenol	570	U	1	180	570	ug/Kg	10/28/25 17:44	PB170215
132-64-9	Dibenzofuran	290	U	1	39.2	290	ug/Kg	10/28/25 17:44	PB170215
121-14-2	2,4-Dinitrotoluene	290	U	1	86.5	290	ug/Kg	10/28/25 17:44	PB170215
84-66-2	Diethylphthalate	290	U	1	48.9	290	ug/Kg	10/28/25 17:44	PB170215
7005-72-3	4-Chlorophenyl-phenylether	290	U	1	46.1	290	ug/Kg	10/28/25 17:44	PB170215
86-73-7	Fluorene	290	U	1	43.7	290	ug/Kg	10/28/25 17:44	PB170215
100-01-6	4-Nitroaniline	290	U	1	110	290	ug/Kg	10/28/25 17:44	PB170215
534-52-1	4,6-Dinitro-2-methylphenol	570	U	1	180	570	ug/Kg	10/28/25 17:44	PB170215

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/22/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/22/25
Client Sample ID:	PR132-WC1-20251022	SDG No.:	Q3434
Lab Sample ID:	Q3434-01	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	57.9
Sample Wt/Vol:	30.01 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	290	U	1	56.8	290	ug/Kg	10/28/25 17:44	PB170215
101-55-3	4-Bromophenyl-phenylether	290	U	1	48.0	290	ug/Kg	10/28/25 17:44	PB170215
118-74-1	Hexachlorobenzene	290	U	1	43.7	290	ug/Kg	10/28/25 17:44	PB170215
1912-24-9	Atrazine	290	U	1	58.7	290	ug/Kg	10/28/25 17:44	PB170215
87-86-5	Pentachlorophenol	570	U	1	88.6	570	ug/Kg	10/28/25 17:44	PB170215
85-01-8	Phenanthrene	290	U	1	36.1	290	ug/Kg	10/28/25 17:44	PB170215
120-12-7	Anthracene	290	U	1	57.5	290	ug/Kg	10/28/25 17:44	PB170215
86-74-8	Carbazole	290	U	1	53.9	290	ug/Kg	10/28/25 17:44	PB170215
84-74-2	Di-n-butylphthalate	290	U	1	82.7	290	ug/Kg	10/28/25 17:44	PB170215
206-44-0	Fluoranthene	290	U	1	51.8	290	ug/Kg	10/28/25 17:44	PB170215
129-00-0	Pyrene	120	J	1	62.2	290	ug/Kg	10/28/25 17:44	PB170215
85-68-7	Butylbenzylphthalate	290	U	1	120	290	ug/Kg	10/28/25 17:44	PB170215
91-94-1	3,3-Dichlorobenzidine	570	U	1	63.4	570	ug/Kg	10/28/25 17:44	PB170215
56-55-3	Benzo(a)anthracene	290	U	1	39.7	290	ug/Kg	10/28/25 17:44	PB170215
218-01-9	Chrysene	290	U	1	34.4	290	ug/Kg	10/28/25 17:44	PB170215
117-81-7	Bis(2-ethylhexyl)phthalate	290	U	1	100	290	ug/Kg	10/28/25 17:44	PB170215
117-84-0	Di-n-octyl phthalate	570	U	1	150	570	ug/Kg	10/28/25 17:44	PB170215
205-99-2	Benzo(b)fluoranthene	290	U	1	32.8	290	ug/Kg	10/28/25 17:44	PB170215
207-08-9	Benzo(k)fluoranthene	290	U	1	38.7	290	ug/Kg	10/28/25 17:44	PB170215
50-32-8	Benzo(a)pyrene	290	U	1	50.9	290	ug/Kg	10/28/25 17:44	PB170215
193-39-5	Indeno(1,2,3-cd)pyrene	290	U	1	50.2	290	ug/Kg	10/28/25 17:44	PB170215
53-70-3	Dibenzo(a,h)anthracene	290	U	1	47.3	290	ug/Kg	10/28/25 17:44	PB170215
191-24-2	Benzo(g,h,i)perylene	290	U	1	44.4	290	ug/Kg	10/28/25 17:44	PB170215
95-94-3	1,2,4,5-Tetrachlorobenzene	290	U	1	44.2	290	ug/Kg	10/28/25 17:44	PB170215
123-91-1	1,4-Dioxane	290	U	1	78.0	290	ug/Kg	10/28/25 17:44	PB170215
58-90-2	2,3,4,6-Tetrachlorophenol	290	U	1	47.3	290	ug/Kg	10/28/25 17:44	PB170215
SURROGATES									
367-12-4	2-Fluorophenol	94.7			18 - 112	63%	SPK: 150		
13127-88-3	Phenol-d6	99.7			15 - 107	66%	SPK: 150		
4165-60-0	Nitrobenzene-d5	66.2			18 - 107	66%	SPK: 100		
321-60-8	2-Fluorobiphenyl	61.3			20 - 109	61%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	102			10 - 116	68%	SPK: 150		
1718-51-0	Terphenyl-d14	76.5			10 - 105	76%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	65500							
1146-65-2	Naphthalene-d8	248000							
15067-26-2	Acenaphthene-d10	146000							
1517-22-2	Phenanthrene-d10	243000							
1719-03-5	Chrysene-d12	127000							
1520-96-3	Perylene-d12	180000							
TENTATIVE IDENTIFIED COMPOUNDS									
000119-61-9	Benzophenone	330	J			10.6	ug/Kg		

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/22/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/22/25
Client Sample ID:	PR132-WC1-20251022	SDG No.:	Q3434
Lab Sample ID:	Q3434-01	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	57.9
Sample Wt/Vol:	30.01 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144106.D
Acq On : 28 Oct 2025 17:44
Operator : RC/JU
Sample : Q3434-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Quant Time: Oct 28 18:04:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	65453	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	248016	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	145676	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	242686	20.000	ng	0.00
76) Chrysene-d12	14.051	240	126847	20.000	ng	0.00
86) Perylene-d12	15.533	264	179897	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	391672	94.674	ng	0.00
7) Phenol-d6	6.504	99	455555	99.706	ng	0.00
23) Nitrobenzene-d5	7.439	82	299877	66.243	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	188656	101.667	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	634513	61.293	ng	0.00
79) Terphenyl-d14	12.992	244	554862	76.472	ng	0.00
Target Compounds						Qvalue
78) Pyrene	12.851	202	18545	2.034	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144106.D
Acq On : 28 Oct 2025 17:44
Operator : RC/JU
Sample : Q3434-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Quant Time: Oct 28 18:04:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Tgt Ion:152 Resp: 65453

Ion	Ratio	Lower	Upper
152	100		
150	159.3	124.4	186.6
115	63.8	49.4	74.0







#5
2-Fluorophenol
Concen: 94.674 ng
RT: 5.498 min Scan# 562
Delta R.T. 0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Tgt Ion:112 Resp: 391672

Ion	Ratio	Lower	Upper
112	100		
64	63.2	49.8	74.8
63	33.8	27.4	41.2







#7
Phenol-d6
Concen: 99.706 ng
RT: 6.504 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Tgt Ion: 99 Resp: 455555

Ion	Ratio	Lower	Upper
99	100		
42	29.0	23.9	35.9
71	43.2	34.4	51.6







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Tgt Ion:136 Resp: 248016

Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.6	13.0
54	8.2	6.6	9.8
68	5.4	4.6	6.8







#23
Nitrobenzene-d5
Concen: 66.243 ng
RT: 7.439 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Tgt Ion: 82 Resp: 299877

Ion	Ratio	Lower	Upper
82	100		
128	41.3	32.4	48.6
54	64.4	54.0	81.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Tgt Ion:164 Resp: 145676

Ion	Ratio	Lower	Upper
164	100		
162	99.7	79.4	119.2
160	44.3	36.7	55.1







#42
2,4,6-Tribromophenol
Concen: 101.667 ng
RT: 10.710 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Tgt Ion:330 Resp: 188656

Ion	Ratio	Lower	Upper
330	100		
332	95.2	77.0	115.6
141	23.9	19.4	29.0







#45
2-Fluorobiphenyl
Concen: 61.293 ng
RT: 9.233 min Scan# 1197
Delta R.T. -0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Tgt Ion:172 Resp: 634513

Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.5	41.3
170	24.2	19.0	28.6







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Tgt Ion:188 Resp: 242686

Ion	Ratio	Lower	Upper
188	100		
94	7.4	5.4	8.0
80	7.9	6.0	9.0







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Tgt Ion:240 Resp: 126847

Ion	Ratio	Lower	Upper
240	100		
120	7.9	6.5	9.7
236	24.6	20.1	30.1







#78
Pyrene
Concen: 2.034 ng
RT: 12.851 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Tgt Ion:202 Resp: 18545

Ion	Ratio	Lower	Upper
202	100		
200	20.0	17.2	25.8
203	15.3	13.5	20.3







#79
Terphenyl-d14
Concen: 76.472 ng
RT: 12.992 min Scan# 1836
Delta R.T. -0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Tgt Ion:244 Resp: 554862

Ion	Ratio	Lower	Upper
244	100		
212	6.5	5.8	8.6
122	6.6	4.9	7.3






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.533 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF144106.D
Acq: 28 Oct 2025 17:44

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Tgt Ion:264 Resp: 179897

Ion	Ratio	Lower	Upper
264	100		
260	21.4	17.0	25.4
265	21.0	16.6	24.8



Sample Results: **BF144106.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144106.D
Acq On : 28 Oct 2025 17:44
Operator : RC/JU
Sample : Q3434-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Integration Parameters: rteint.p
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC: BF144106.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.110	492	496	502	rVB	15056	23551	1.33%	0.197%
2	5.498	556	562	584	rBV	1084461	1459535	82.63%	12.215%
3	6.504	727	733	751	rBV	1104840	1528656	86.54%	12.793%
4	6.875	791	796	803	rBV	313821	388089	21.97%	3.248%
5	7.439	886	892	902	rBV	717805	948236	53.68%	7.936%
6	8.157	1009	1014	1021	rBV	386572	501254	28.38%	4.195%
7	9.233	1192	1197	1202	rBV	1405179	1766447	100.00%	14.783%
8	9.627	1259	1264	1272	rVB3	16946	27436	1.55%	0.230%
9	9.916	1307	1313	1318	rBV	474393	592076	33.52%	4.955%
10	10.563	1418	1423	1429	rBV	13199	23162	1.31%	0.194%
11	10.627	1429	1434	1439	rVV	129199	167147	9.46%	1.399%
12	10.704	1442	1447	1460	rVV	921272	1258390	71.24%	10.531%
13	10.827	1464	1468	1475	rVB	30712	45167	2.56%	0.378%
14	10.939	1483	1487	1491	rBV	28513	35261	2.00%	0.295%
15	11.292	1543	1547	1554	rVB6	19689	32520	1.84%	0.272%
16	11.404	1561	1566	1576	rBV	447440	633885	35.88%	5.305%
17	11.927	1649	1655	1662	rBV3	63450	106416	6.02%	0.891%
18	12.021	1668	1671	1680	rVB6	14634	28195	1.60%	0.236%
19	12.427	1736	1740	1743	rBV	14996	20306	1.15%	0.170%
20	12.621	1769	1773	1777	rBV	27690	32483	1.84%	0.272%
21	12.851	1808	1812	1819	rBV	29032	38293	2.17%	0.320%
22	12.992	1831	1836	1844	rBV	1064392	1357826	76.87%	11.363%
23	13.915	1988	1993	2004	rBV10	19629	61399	3.48%	0.514%
24	14.051	2008	2016	2027	rVB	258773	392372	22.21%	3.284%
25	14.898	2157	2160	2165	rVB2	14960	19192	1.09%	0.161%
26	15.533	2263	2268	2280	rVB	278123	461792	26.14%	3.865%

Sum of corrected areas: 11949086


Sample Results: **BF144106.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144106.D
Acq On : 28 Oct 2025 17:44
Operator : RC/JU
Sample : Q3434-01
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P







Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144106.D
Acq On : 28 Oct 2025 17:44
Operator : RC/JU
Sample : Q3434-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION


TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P


****

Peak Number 1 Benzophenone Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
10.627	5.65 ng	167147	Acenaphthene-d10	9.916

Hit# of 5	Tentative ID	MW	MolForm	CAS#	Qual
1	Benzophenone	182	C13H10O	000119-61-9	97
2	Phenyl 4-pyridyl ketone	183	C12H9NO	014548-46-0	47
3	Benzilic acid	228	C14H12O3	000076-93-7	47
4	1,2-Propanedione, 1-phenyl-	148	C9H8O2	000579-07-7	43
5	N-(1-Cyanovinyl)benzamide	172	C10H8N2O	1000186-19-1	43





Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144106.D
Acq On : 28 Oct 2025 17:44
Operator : RC/JU
Sample : Q3434-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC1-20251022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

					--Internal Standard---			
TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc
Benzophenone	10.627	5.7	ng	167147	3	9.916	592076	20.0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/22/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/22/25
Client Sample ID:	PR132-WC2-20251022	SDG No.:	Q3434
Lab Sample ID:	Q3434-02	Matrix:	Water
Analytical Method:	8270E Level: LOW	% Solid:	0
Sample Wt/Vol:	880 mL Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C Prep Date: 10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	11.4	U	1	4.40	11.4	ug/L	10/29/25 17:02	PB170286
108-95-2	Phenol	5.70	U	1	1.00	5.70	ug/L	10/29/25 17:02	PB170286
111-44-4	bis(2-Chloroethyl)ether	5.70	U	1	0.92	5.70	ug/L	10/29/25 17:02	PB170286
95-57-8	2-Chlorophenol	5.70	U	1	0.66	5.70	ug/L	10/29/25 17:02	PB170286
95-48-7	2-Methylphenol	5.70	U	1	1.30	5.70	ug/L	10/29/25 17:02	PB170286
108-60-1	2,2-oxybis(1-Chloropropane)	5.70	U	1	1.50	5.70	ug/L	10/29/25 17:02	PB170286
98-86-2	Acetophenone	5.70	U	1	0.84	5.70	ug/L	10/29/25 17:02	PB170286
65794-96-9	3+4-Methylphenols	11.4	U	1	1.30	11.4	ug/L	10/29/25 17:02	PB170286
621-64-7	n-Nitroso-di-n-propylamine	2.80	U	1	1.60	2.80	ug/L	10/29/25 17:02	PB170286
67-72-1	Hexachloroethane	5.70	U	1	0.74	5.70	ug/L	10/29/25 17:02	PB170286
98-95-3	Nitrobenzene	5.70	U	1	0.86	5.70	ug/L	10/29/25 17:02	PB170286
78-59-1	Isophorone	5.70	U	1	0.85	5.70	ug/L	10/29/25 17:02	PB170286
88-75-5	2-Nitrophenol	5.70	U	1	2.00	5.70	ug/L	10/29/25 17:02	PB170286
105-67-9	2,4-Dimethylphenol	5.70	U	1	2.10	5.70	ug/L	10/29/25 17:02	PB170286
111-91-1	bis(2-Chloroethoxy)methane	5.70	U	1	0.77	5.70	ug/L	10/29/25 17:02	PB170286
120-83-2	2,4-Dichlorophenol	5.70	U	1	0.59	5.70	ug/L	10/29/25 17:02	PB170286
91-20-3	Naphthalene	5.70	U	1	0.57	5.70	ug/L	10/29/25 17:02	PB170286
106-47-8	4-Chloroaniline	5.70	U	1	0.95	5.70	ug/L	10/29/25 17:02	PB170286
87-68-3	Hexachlorobutadiene	5.70	U	1	0.61	5.70	ug/L	10/29/25 17:02	PB170286
105-60-2	Caprolactam	11.4	U	1	1.30	11.4	ug/L	10/29/25 17:02	PB170286
59-50-7	4-Chloro-3-methylphenol	5.70	U	1	0.67	5.70	ug/L	10/29/25 17:02	PB170286
91-57-6	2-Methylnaphthalene	5.70	U	1	0.64	5.70	ug/L	10/29/25 17:02	PB170286
77-47-4	Hexachlorocyclopentadiene	11.4	U	1	4.10	11.4	ug/L	10/29/25 17:02	PB170286
88-06-2	2,4,6-Trichlorophenol	5.70	U	1	0.58	5.70	ug/L	10/29/25 17:02	PB170286
95-95-4	2,4,5-Trichlorophenol	5.70	U	1	0.70	5.70	ug/L	10/29/25 17:02	PB170286
92-52-4	1,1-Biphenyl	5.70	U	1	0.60	5.70	ug/L	10/29/25 17:02	PB170286
91-58-7	2-Chloronaphthalene	5.70	U	1	0.69	5.70	ug/L	10/29/25 17:02	PB170286
88-74-4	2-Nitroaniline	5.70	U	1	1.40	5.70	ug/L	10/29/25 17:02	PB170286
131-11-3	Dimethylphthalate	5.70	U	1	0.69	5.70	ug/L	10/29/25 17:02	PB170286
208-96-8	Acenaphthylene	5.70	U	1	0.85	5.70	ug/L	10/29/25 17:02	PB170286
606-20-2	2,6-Dinitrotoluene	5.70	U	1	1.00	5.70	ug/L	10/29/25 17:02	PB170286
99-09-2	3-Nitroaniline	5.70	U	1	1.20	5.70	ug/L	10/29/25 17:02	PB170286
83-32-9	Acenaphthene	5.70	U	1	0.63	5.70	ug/L	10/29/25 17:02	PB170286
51-28-5	2,4-Dinitrophenol	11.4	U	1	6.80	11.4	ug/L	10/29/25 17:02	PB170286
100-02-7	4-Nitrophenol	11.4	U	1	2.70	11.4	ug/L	10/29/25 17:02	PB170286
132-64-9	Dibenzofuran	5.70	U	1	0.69	5.70	ug/L	10/29/25 17:02	PB170286
121-14-2	2,4-Dinitrotoluene	5.70	U	1	1.40	5.70	ug/L	10/29/25 17:02	PB170286
84-66-2	Diethylphthalate	5.70	U	1	0.78	5.70	ug/L	10/29/25 17:02	PB170286
7005-72-3	4-Chlorophenyl-phenylether	5.70	U	1	0.77	5.70	ug/L	10/29/25 17:02	PB170286
86-73-7	Fluorene	5.70	U	1	0.72	5.70	ug/L	10/29/25 17:02	PB170286
100-01-6	4-Nitroaniline	5.70	U	1	1.70	5.70	ug/L	10/29/25 17:02	PB170286
534-52-1	4,6-Dinitro-2-methylphenol	11.4	U	1	3.30	11.4	ug/L	10/29/25 17:02	PB170286

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/22/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/22/25
Client Sample ID:	PR132-WC2-20251022	SDG No.:	Q3434
Lab Sample ID:	Q3434-02	Matrix:	Water
Analytical Method:	8270E Level: LOW	% Solid:	0
Sample Wt/Vol:	880 mL Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C Prep Date: 10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	5.70	U	1	0.66	5.70	ug/L	10/29/25 17:02	PB170286
101-55-3	4-Bromophenyl-phenylether	5.70	U	1	0.45	5.70	ug/L	10/29/25 17:02	PB170286
118-74-1	Hexachlorobenzene	5.70	U	1	0.59	5.70	ug/L	10/29/25 17:02	PB170286
1912-24-9	Atrazine	5.70	U	1	1.10	5.70	ug/L	10/29/25 17:02	PB170286
87-86-5	Pentachlorophenol	11.4	U	1	1.80	11.4	ug/L	10/29/25 17:02	PB170286
85-01-8	Phenanthrene	5.70	U	1	0.57	5.70	ug/L	10/29/25 17:02	PB170286
120-12-7	Anthracene	5.70	U	1	0.69	5.70	ug/L	10/29/25 17:02	PB170286
86-74-8	Carbazole	5.70	U	1	0.82	5.70	ug/L	10/29/25 17:02	PB170286
84-74-2	Di-n-butylphthalate	5.70	U	1	1.40	5.70	ug/L	10/29/25 17:02	PB170286
206-44-0	Fluoranthene	5.70	U	1	0.93	5.70	ug/L	10/29/25 17:02	PB170286
129-00-0	Pyrene	5.70	U	1	0.57	5.70	ug/L	10/29/25 17:02	PB170286
85-68-7	Butylbenzylphthalate	5.70	U	1	2.20	5.70	ug/L	10/29/25 17:02	PB170286
91-94-1	3,3-Dichlorobenzidine	11.4	U	1	1.10	11.4	ug/L	10/29/25 17:02	PB170286
56-55-3	Benzo(a)anthracene	5.70	U	1	0.51	5.70	ug/L	10/29/25 17:02	PB170286
218-01-9	Chrysene	5.70	U	1	0.50	5.70	ug/L	10/29/25 17:02	PB170286
117-81-7	Bis(2-ethylhexyl)phthalate	5.70	U	1	1.80	5.70	ug/L	10/29/25 17:02	PB170286
117-84-0	Di-n-octyl phthalate	11.4	U	1	2.70	11.4	ug/L	10/29/25 17:02	PB170286
205-99-2	Benzo(b)fluoranthene	5.70	U	1	0.56	5.70	ug/L	10/29/25 17:02	PB170286
207-08-9	Benzo(k)fluoranthene	5.70	U	1	0.55	5.70	ug/L	10/29/25 17:02	PB170286
50-32-8	Benzo(a)pyrene	5.70	U	1	0.63	5.70	ug/L	10/29/25 17:02	PB170286
193-39-5	Indeno(1,2,3-cd)pyrene	5.70	U	1	0.67	5.70	ug/L	10/29/25 17:02	PB170286
53-70-3	Dibenzo(a,h)anthracene	5.70	U	1	0.76	5.70	ug/L	10/29/25 17:02	PB170286
191-24-2	Benzo(g,h,i)perylene	5.70	U	1	0.78	5.70	ug/L	10/29/25 17:02	PB170286
95-94-3	1,2,4,5-Tetrachlorobenzene	5.70	U	1	0.59	5.70	ug/L	10/29/25 17:02	PB170286
123-91-1	1,4-Dioxane	5.70	U	1	1.10	5.70	ug/L	10/29/25 17:02	PB170286
58-90-2	2,3,4,6-Tetrachlorophenol	5.70	U	1	0.82	5.70	ug/L	10/29/25 17:02	PB170286
SURROGATES									
367-12-4	2-Fluorophenol	83.2			23 - 138	55%	SPK: 150		
13127-88-3	Phenol-d6	62.0			10 - 134	41%	SPK: 150		
4165-60-0	Nitrobenzene-d5	92.1			67 - 132	92%	SPK: 100		
321-60-8	2-Fluorobiphenyl	86.9			52 - 132	87%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	138			44 - 137	92%	SPK: 150		
1718-51-0	Terphenyl-d14	82.9			42 - 152	83%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	59100							
1146-65-2	Naphthalene-d8	222000							
15067-26-2	Acenaphthene-d10	127000							
1517-22-2	Phenanthrene-d10	212000							
1719-03-5	Chrysene-d12	165000							
1520-96-3	Perylene-d12	206000							
TENTATIVE IDENTIFIED COMPOUNDS									
000112-34-5	Ethanol, 2-(2-butoxyethoxy)-	13.1	J			8.06	ug/L		
000119-61-9	Benzophenone	4.10	J			10.6	ug/L		

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/22/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/22/25
Client Sample ID:	PR132-WC2-20251022	SDG No.:	Q3434
Lab Sample ID:	Q3434-02	Matrix:	Water
Analytical Method:	8270E Level: LOW	% Solid:	0
Sample Wt/Vol:	880 mL Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C Prep Date: 10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
000057-10-3	n-Hexadecanoic acid	6.60	J			11.9	ug/L		
018835-33-1	1-Hexacosene	3.80	J			13.9	ug/L		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Quant Time: Oct 29 17:30:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	59068	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	222374	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	127295	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	211645	20.000	ng	0.00
76) Chrysene-d12	14.051	240	165479	20.000	ng	0.00
86) Perylene-d12	15.533	264	206243	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	310598	83.192	ng	0.00
7) Phenol-d6	6.498	99	255503	61.966	ng	0.00
23) Nitrobenzene-d5	7.439	82	373695	92.069	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	223716	137.969	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	785918	86.881	ng	0.00
79) Terphenyl-d14	12.992	244	784373	82.866	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Quant Time: Oct 29 17:30:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Tgt Ion:152 Resp: 59068

Ion	Ratio	Lower	Upper
152	100		
150	154.5	124.4	186.6
115	65.8	49.4	74.0







#5
2-Fluorophenol
Concen: 83.192 ng
RT: 5.493 min Scan# 561
Delta R.T. 0.000 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Tgt Ion:112 Resp: 310598

Ion	Ratio	Lower	Upper
112	100		
64	61.7	49.8	74.8
63	32.3	27.4	41.2







#7
Phenol-d6
Concen: 61.966 ng
RT: 6.498 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Tgt Ion: 99 Resp: 255503

Ion	Ratio	Lower	Upper
99	100		
42	26.5	23.9	35.9
71	40.1	34.4	51.6







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Tgt Ion:136 Resp: 222374

Ion	Ratio	Lower	Upper
136	100		
137	10.5	8.6	13.0
54	8.6	6.6	9.8
68	5.4	4.6	6.8







#23
Nitrobenzene-d5
Concen: 92.069 ng
RT: 7.439 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Tgt Ion: 82 Resp: 373695

Ion	Ratio	Lower	Upper
82	100		
128	41.4	32.4	48.6
54	64.6	54.0	81.0







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Tgt Ion:164 Resp: 127295

Ion	Ratio	Lower	Upper
164	100		
162	99.7	79.4	119.2
160	44.2	36.7	55.1







#42
2,4,6-Tribromophenol
Concen: 137.969 ng
RT: 10.710 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Tgt Ion:330 Resp: 223716

Ion	Ratio	Lower	Upper
330	100		
332	96.0	77.0	115.6
141	22.7	19.4	29.0







#45
2-Fluorobiphenyl
Concen: 86.881 ng
RT: 9.233 min Scan# 1197
Delta R.T. -0.006 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Tgt Ion:172 Resp: 785918

Ion	Ratio	Lower	Upper
172	100		
171	35.1	27.5	41.3
170	23.7	19.0	28.6







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Tgt Ion:188 Resp: 211645

Ion	Ratio	Lower	Upper
188	100		
94	7.9	5.4	8.0
80	7.7	6.0	9.0







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Tgt Ion:240 Resp: 165479

Ion	Ratio	Lower	Upper
240	100		
120	8.0	6.5	9.7
236	27.0	20.1	30.1







#79
Terphenyl-d14
Concen: 82.866 ng
RT: 12.992 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Tgt Ion:244 Resp: 784373

Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.8	8.6
122	6.8	4.9	7.3






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.533 min Scan# 2268
Delta R.T. -0.006 min
Lab File: BF144132.D
Acq: 29 Oct 2025 17:02

Tgt Ion:264 Resp: 206243

Ion	Ratio	Lower	Upper
264	100		
260	22.4	17.0	25.4
265	22.5	16.6	24.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Integration Parameters: rteint.p
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC: BF144132.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.352	21	27	34	rVB	19624	32137	1.47%	0.239%
2	2.969	125	132	139	rBV	25710	49438	2.27%	0.368%
3	5.098	489	494	502	rVB	57862	89126	4.09%	0.664%
4	5.416	541	548	555	rBV	42731	68382	3.14%	0.510%
5	5.493	555	561	573	rVV	845685	1151494	52.83%	8.581%
6	5.828	611	618	629	rBV	433184	584085	26.80%	4.353%
7	6.498	727	732	751	rBV	603350	812651	37.28%	6.056%
8	6.710	764	768	775	rBV	22397	34576	1.59%	0.258%
9	6.875	791	796	804	rBV	287807	358998	16.47%	2.675%
10	7.398	880	885	887	rBV	36421	55492	2.55%	0.414%
11	7.440	887	892	904	rVV	900111	1236637	56.73%	9.216%
12	7.534	904	908	915	rVB	14782	26929	1.24%	0.201%
13	8.057	992	997	1009	rBV	163455	259733	11.92%	1.936%
14	8.157	1009	1014	1022	rVB	359202	449185	20.61%	3.348%
15	9.234	1191	1197	1202	rBV	1698215	2179825	100.00%	16.245%
16	9.416	1224	1228	1237	rVB2	23339	36202	1.66%	0.270%
17	9.916	1307	1313	1317	rBV	416717	529979	24.31%	3.950%
18	9.945	1317	1318	1322	rVB	29147	24352	1.12%	0.181%
19	10.322	1378	1382	1388	rVB3	12111	23587	1.08%	0.176%
20	10.628	1429	1434	1440	rBV	70538	94770	4.35%	0.706%
21	10.704	1442	1447	1457	rBV	1081130	1459612	66.96%	10.878%
22	10.828	1464	1468	1474	rVB3	18370	31486	1.44%	0.235%
23	11.292	1543	1547	1554	rVB4	16233	29802	1.37%	0.222%
24	11.404	1561	1566	1576	rBV	396223	547689	25.13%	4.082%
25	11.516	1581	1585	1589	rBV5	19932	29778	1.37%	0.222%
26	11.563	1589	1593	1599	rVB2	42754	71822	3.29%	0.535%
27	11.927	1650	1655	1668	rBV2	93189	159288	7.31%	1.187%
28	12.716	1785	1789	1795	rBV5	16968	30290	1.39%	0.226%
29	12.992	1831	1836	1842	rBV	1466351	1922189	88.18%	14.325%
30	13.910	1986	1992	2004	rBV4	27916	75570	3.47%	0.563%
31	14.051	2008	2016	2024	rVB	333138	456127	20.92%	3.399%
32	15.533	2262	2268	2282	rVB	310919	507149	23.27%	3.780%


Sum of corrected areas: 13418380


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P







Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022


Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION


TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

****
Peak Number 5 Ethanol, 2-(2-butoxyethoxy)- Concentration Rank 2

R.T.	EstConc	Area	Relative to ISTD	R.T.
8.057	11.56 ng	259733	Naphthalene-d8	8.157

Hit# of 5	Tentative ID	MW	MolForm	CAS#	Qual
1	Ethanol, 2-(2-butoxyethoxy)-	162	C8H18O3	000112-34-5	90
2	Ethanol, 1-(2-butoxyethoxy)-	162	C8H18O3	054446-78-5	83
3	Ethylene glycol monoisobutyl ether	118	C6H14O2	004439-24-1	64
4	Di-sec-Butyl ether	130	C8H18O	006863-58-7	59
5	DL-Lactamide, butyl ether	145	C7H15NO2	1000452-56-5	53





Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION


TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P


****

Peak Number 6 Benzophenone Concentration Rank 6

R.T.	EstConc	Area	Relative to ISTD	R.T.
10.628	3.58 ng	94770	Acenaphthene-d10	9.916

Hit# of 5	Tentative ID	MW	MolForm	CAS#	Qual
1	Benzophenone	182	C13H10O	000119-61-9	95
2	Phenyl 4-pyridyl ketone	183	C12H9NO	014548-46-0	53
3	Benzilic acid	228	C14H12O3	000076-93-7	43
4	2-(2-Oxo-2-phenyl-ethyl)-malonon...	184	C11H8N2O	1000296-76-9	43
5	Benzoylformic acid	150	C8H6O3	000611-73-4	43





Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION


TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P


****

Peak Number 8 n-Hexadecanoic acid Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
11.927	5.82 ng	159288	Phenanthrene-d10	11.404

Hit# of 5	Tentative ID	MW	MolForm	CAS#	Qual
1	n-Hexadecanoic acid	256	C16H32O2	000057-10-3	98
2	Tridecanoic acid	214	C13H26O2	000638-53-9	91
3	Pentadecanoic acid	242	C15H30O2	001002-84-2	81
4	Tetradecanoic acid	228	C14H28O2	000544-63-8	64
5	n-Decanoic acid	172	C10H20O2	000334-48-5	58





Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION


TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P


****

Peak Number 9 1-Hexacosene Concentration Rank 7

R.T.	EstConc	Area	Relative to ISTD	R.T.
13.910	3.31 ng	75570	Chrysene-d12	14.051

Hit# of 5	Tentative ID	MW	MolForm	CAS#	Qual
1	1-Hexacosene	364	C26H52	018835-33-1	93
2	1-Nonadecene	266	C19H38	018435-45-5	87
3	Trifluoroacetoxy hexadecane	338	C18H33F3O2	006222-03-3	87
4	Trifluoroacetic acid, pentadecyl...	324	C17H31F3O2	959010-23-2	81
5	Pentafluoropropionic acid, penta...	374	C18H31F5O2	959092-08-1	81





Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144132.D
Acq On : 29 Oct 2025 17:02
Operator : RC/JU
Sample : Q3434-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PR132-WC2-20251022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

					--Internal Standard---			
TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc
Ethanol, 2-(2-b...	8.057	11.6	ng	259733	2	8.157	449185	20.0
Benzophenone	10.628	3.6	ng	94770	3	9.916	529979	20.0
n-Hexadecanoic ...	11.927	5.8	ng	159288	4	11.404	547689	20.0
1-Hexacosene	13.910	3.3	ng	75570	5	14.051	456127	20.0

CALIBRATION SUMMARY

Response Factor Report BNA_F

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 24 17:30:25 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BF144056.D 5 =BF144057.D 10 =BF144058.D 20 =BF144059.D 40 =BF144060.D 50 =BF144061.D 60 =BF144064.D 80 =BF1440
63.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------	ISTD	----------							
2)		1,4-Dioxane		0.723	0.494	0.485	0.550	0.519	0.572	0.481	0.546	15.58
3)		Pyridine		1.395	1.384	1.530	1.527	1.427	1.604	1.370	1.462	6.19
4)		n-Nitrosodimet...			0.805	0.963	0.912	0.835	1.017	0.860	0.899	8.98
5)	S	2-Fluorophenol		1.324	1.322	1.328	1.303	1.163	1.322	1.087	1.264	7.76
6)		Aniline		2.173	2.152	2.106	2.016	1.809	2.050	1.675	1.997	9.36
7)	S	Phenol-d6		1.548	1.443	1.481	1.408	1.289	1.412	1.193	1.396	8.56
8)		2-Chlorophenol		1.398	1.310	1.306	1.251	1.124	1.258	1.067	1.245	9.17
9)		Benzaldehyde			1.093	0.965	1.043	0.959	1.406	1.039	1.084	15.28
10)	C	Phenol		1.896	1.860	1.820	1.736	1.575	1.725	1.445	1.722	9.41
11)		bis(2-Chloroet...		1.406	1.278	1.285	1.248	1.128	1.285	1.082	1.245	8.72
12)		1,3-Dichlorobe...		1.681	1.582	1.640	1.565	1.415	1.566	1.300	1.536	8.66
13)	C	1,4-Dichlorobe...		1.645	1.567	1.592	1.551	1.371	1.575	1.295	1.513	8.52
14)		1,2-Dichlorobe...		1.588	1.548	1.534	1.480	1.355	1.520	1.225	1.464	8.81
15)		Benzyl Alcohol			1.129	1.234	1.145	1.056	1.222	1.025	1.135	7.47
16)		2,2'-oxybis(1-...		2.561	2.432	2.505	2.376	2.149	2.313	2.038	2.339	8.08
17)		2-Methylphenol		1.052	1.010	1.032	0.931	0.874	0.981	0.823	0.957	8.89
18)		Hexachloroethane		0.553	0.544	0.569	0.544	0.506	0.595	0.477	0.541	7.21
19)	P	n-Nitroso-di-n...	0.900	0.971	0.972	0.949	0.876	0.801	0.943	0.809	0.903	7.60
20)		3+4-Methylphenols			1.270	1.190	1.167	1.036	1.159	0.969	1.132	9.69
21)	I	Naphthalene-d8	----------	ISTD	----------							
22)		Acetophenone		0.473	0.459	0.433	0.437	0.397	0.456	0.371	0.432	8.41
23)	S	Nitrobenzene-d5		0.377	0.378	0.374	0.368	0.335	0.393	0.330	0.365	6.48
24)		Nitrobenzene		0.395	0.384	0.404	0.392	0.360	0.417	0.351	0.386	6.05
25)		Isophorone		0.665	0.662	0.645	0.643	0.584	0.686	0.582	0.638	6.29
26)	C	2-Nitrophenol		0.190	0.184	0.196	0.191	0.167	0.194	0.165	0.184	6.96
27)		2,4-Dimethylph...		0.289	0.279	0.262	0.287	0.254	0.293	0.247	0.273	6.73
28)		bis(2-Chloroet...		0.430	0.408	0.413	0.404	0.365	0.416	0.351	0.398	7.25
29)	C	2,4-Dichloroph...		0.334	0.345	0.325	0.323	0.293	0.328	0.275	0.317	7.75
30)		1,2,4-Trichlor...		0.344	0.339	0.334	0.335	0.301	0.348	0.286	0.327	7.19
31)		Naphthalene		1.054	1.002	0.962	0.965	0.852	0.983	0.788	0.944	9.72
32)		Benzoic acid			0.217	0.200	0.146	0.147	0.168	0.162	0.173	16.81
33)		4-Chloroaniline		0.332	0.356	0.349	0.336	0.298	0.344	0.286	0.329	8.10
34)	C	Hexachlorobuta...		0.279	0.276	0.274	0.270	0.241	0.296	0.234	0.267	8.19
35)		Caprolactam			0.078	0.082	0.085	0.072	0.086	0.075	0.080	7.03
36)	C	4-Chloro-3-met...		0.302	0.289	0.285	0.282	0.257	0.296	0.250	0.280	6.98
37)		2-Methylnaphth...		0.718	0.654	0.631	0.644	0.549	0.637	0.524	0.622	10.55
38)		1-Methylnaphth...		0.685	0.627	0.591	0.591	0.524	0.611	0.499	0.590	10.60

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF102425.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.777	0.707	0.693	0.707	0.629	0.716	0.583	0.687	9.17
41)	P	Hexachlorocycl...		0.148	0.208	0.218	0.205	0.273	0.222	0.212	18.88
42)	S	2,4,6-Tribromo...	0.251	0.249	0.256	0.262	0.241	0.289	0.236	0.255	6.77
43)	C	2,4,6-Trichlor...	0.430	0.447	0.480	0.473	0.427	0.485	0.386	0.447	7.98
44)		2,4,5-Trichlor...	0.488	0.485	0.484	0.467	0.427	0.511	0.419	0.469	7.19
45)	S	2-Fluorobiphenyl	1.654	1.503	1.411	1.451	1.291	1.473	1.164	1.421	11.04
46)		1,1'-Biphenyl	1.613	1.532	1.441	1.472	1.317	1.470	1.176	1.432	10.09
47)		2-Chloronaphth...	1.336	1.264	1.233	1.222	1.101	1.250	0.989	1.199	9.68
48)		2-Nitroaniline	0.325	0.359	0.368	0.380	0.333	0.399	0.328	0.356	8.00
49)		Acenaphthylene	1.780	1.693	1.664	1.703	1.476	1.668	1.337	1.617	9.54
50)		Dimethylphthalate	1.398	1.328	1.332	1.366	1.219	1.387	1.146	1.311	7.15
51)		2,6-Dinitrotol...	0.228	0.257	0.275	0.276	0.252	0.288	0.240	0.259	8.31
52)	C	Acenaphthene	1.125	1.153	1.134	1.114	0.999	1.122	0.914	1.080	8.21
53)		3-Nitroaniline	0.276	0.294	0.293	0.322	0.281	0.301	0.250	0.288	7.79
54)	P	2,4-Dinitrophenol		0.161	0.162	0.124	0.116	0.139	0.140	0.140	13.44
55)		Dibenzofuran	1.685	1.572	1.512	1.559	1.391	1.535	1.242	1.499	9.55
56)	P	4-Nitrophenol		0.224	0.199	0.187	0.175	0.202	0.184	0.195	8.90
57)		2,4-Dinitrotol...	0.302	0.343	0.370	0.385	0.336	0.390	0.323	0.350	9.44
58)		Fluorene	1.359	1.169	1.133	1.188	1.041	1.162	0.945	1.143	11.27
59)		2,3,4,6-Tetrac...	0.317	0.318	0.333	0.360	0.322	0.366	0.297	0.330	7.53
60)		Diethylphthalate	1.286	1.264	1.244	1.290	1.145	1.318	1.083	1.233	6.98
61)		4-Chlorophenyl...	0.706	0.667	0.656	0.675	0.591	0.674	0.529	0.643	9.49
62)		4-Nitroaniline	0.262	0.266	0.275	0.296	0.260	0.299	0.253	0.273	6.70
63)		Azobenzene	1.192	1.178	1.187	1.236	1.103	1.253	1.020	1.167	6.91
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.133	0.151	0.123	0.115	0.132	0.116	0.128	10.63
66)	c	n-Nitrosodiphe...	0.716	0.666	0.657	0.643	0.590	0.659	0.542	0.639	8.82
67)		4-Bromophenyl-...	0.274	0.256	0.247	0.260	0.230	0.270	0.215	0.250	8.52
68)		Hexachlorobenzene	0.311	0.338	0.309	0.321	0.275	0.328	0.268	0.307	8.59
69)		Atrazine	0.229	0.223	0.218	0.208	0.185	0.235	0.169	0.210	11.52
70)	C	Pentachlorophenol		0.123	0.149	0.150	0.143	0.168	0.144	0.146	9.94
71)		Phenanthrene	1.073	1.040	1.009	1.017	0.918	1.043	0.854	0.994	7.89
72)		Anthracene	1.171	1.077	1.029	1.053	0.932	1.065	0.857	1.026	10.01
73)		Carbazole	0.943	0.920	0.883	0.905	0.818	0.907	0.737	0.874	8.20
74)		Di-n-butylphth...	1.090	1.091	1.116	1.118	1.002	1.135	0.945	1.071	6.58
75)	C	Fluoranthene	1.171	1.151	1.121	1.123	0.997	1.167	0.954	1.098	7.88
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine		0.689	0.614	0.665	0.513	0.787	0.466	0.622	18.99
78)		Pyrene	1.603	1.466	1.426	1.521	1.295	1.519	1.236	1.438	9.08
79)	S	Terphenyl-d14	1.280	1.134	1.121	1.212	1.002	1.241	1.018	1.144	9.39
80)		Butylbenzylpht...	0.508	0.542	0.566	0.601	0.494	0.595	0.501	0.544	8.25
81)		Benzo(a)anthra...	1.341	1.322	1.359	1.516	1.206	1.492	1.226	1.352	8.79
82)		3,3'-Dichlorob...		0.469	0.485	0.510	0.404	0.529	0.427	0.470	10.21
83)		Chrysene	1.330	1.263	1.253	1.319	1.179	1.366	1.170	1.268	5.93
84)		Bis(2-ethylhex...	0.745	0.725	0.761	0.795	0.690	0.813	0.670	0.743	7.01
85)	c	Di-n-octyl pht...		1.228	1.295	1.399	1.179	1.382	1.180	1.277	7.65

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF102425.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.351	1.331	1.394	1.477	1.295	1.501	1.263	1.373	6.53
88)		Benzo(b)fluora...	1.134	1.108	1.153	1.264	1.055	1.151	1.050	1.131	6.39
89)		Benzo(k)fluora...	1.206	1.077	1.010	1.059	0.906	1.157	0.886	1.043	11.44
90)	C	Benzo(a)pyrene	1.049	0.990	1.036	1.089	0.942	1.086	0.909	1.014	6.88
91)		Dibenzo(a,h)an...	1.058	1.061	1.107	1.166	1.024	1.195	0.997	1.087	6.72
92)		Benzo(g,h,i)pe...	1.051	1.051	1.124	1.195	1.023	1.192	1.015	1.093	7.08

(#) = Out of Range

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144056.D
Acq On : 24 Oct 2025 10:13
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Oct 24 17:00:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	63395	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	236986	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	128411	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	216866	20.000	ng	0.00
76) Chrysene-d12	14.057	240	165196	20.000	ng	0.00
86) Perylene-d12	15.539	264	182261	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.281	70	7133	2.493	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144056.D
Acq On : 24 Oct 2025 10:13
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Oct 24 17:00:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 79
Delta R.T. -0.000 min
Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 63395

Ion	Ratio	Lower	Upper
152	100		
150	159.5	124.4	186.6
115	63.2	49.4	74.0







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.49 min

Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	62.3
63	34.3






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.50 min


Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99

Sig	Exp Ratio
99	100
42	29.9
71	43.0








#19
n-Nitroso-di-n-propylamine
Concen: 2.493 ng
RT: 7.281 min Scan# 865
Delta R.T. -0.006 min
Lab File: BF144056.D
Acq: 24 Oct 2025 10:13


Tgt Ion: 70 Resp: 7133

Ion	Ratio	Lower	Upper
70	100		
42	80.5	67.4	101.0
101	9.6	7.1	10.7
130	16.2	13.5	20.3











#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 236986

Ion	Ratio	Lower	Upper
136	100		
137	11.4	8.6	13.0
54	8.8	6.6	9.8
68	5.4	4.6	6.8







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 7.44 min

Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	40.5
54	67.5







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 128411

Ion	Ratio	Lower	Upper
164	100		
162	102.0	79.4	119.2
160	44.9	36.7	55.1







#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 10.71 min

Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	96.3
141	24.2






#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 9.24 min

Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172


Sig	Exp Ratio
172	100
171	34.4
170	23.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Tgt Ion:188 Resp: 216866

Ion	Ratio	Lower	Upper
188	100		
94	6.9	5.4	8.0
80	7.6	6.0	9.0







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 165196

Ion	Ratio	Lower	Upper
240	100		
120	7.4	6.5	9.7
236	26.6	20.1	30.1







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 13.00 min

Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	7.2
122	6.1






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144056.D
Acq: 24 Oct 2025 10:13

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 182261

Ion	Ratio	Lower	Upper
264	100		
260	22.9	17.0	25.4
265	22.3	16.6	24.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144057.D
Acq On : 24 Oct 2025 10:42
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Oct 24 17:01:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	64729	20.000	ng		0.00
21) Naphthalene-d8	8.163	136	239688	20.000	ng		0.00
39) Acenaphthene-d10	9.916	164	127182	20.000	ng		0.00
64) Phenanthrene-d10	11.410	188	215493	20.000	ng		0.00
76) Chrysene-d12	14.057	240	170453	20.000	ng		0.00
86) Perylene-d12	15.539	264	190551	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.492	112	42841	10.471	ng		0.00
7) Phenol-d6	6.498	99	50099	11.088	ng		0.00
23) Nitrobenzene-d5	7.434	82	45136	10.317	ng		0.00
42) 2,4,6-Tribromophenol	10.710	330	15954	9.848	ng		0.00
45) 2-Fluorobiphenyl	9.233	172	105210	11.641	ng		0.00
79) Terphenyl-d14	12.992	244	109089	11.189	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.640	88	11703	6.619	ng		95
3) Pyridine	3.428	79	22571	4.769	ng	#	95
6) Aniline	6.539	93	35169	5.441	ng		99
8) 2-Chlorophenol	6.663	128	22623	5.615	ng		93
10) Phenol	6.516	94	30686	5.505	ng		92
11) bis(2-Chloroethyl)ether	6.610	93	22757	5.650	ng		99
12) 1,3-Dichlorobenzene	6.822	146	27208	5.474	ng		98
13) 1,4-Dichlorobenzene	6.898	146	26614	5.433	ng		98
14) 1,2-Dichlorobenzene	7.045	146	25700	5.423	ng		95
16) 2,2'-oxybis(1-Chloropr...	7.145	45	41443	5.474	ng		94
17) 2-Methylphenol	7.128	107	17019	5.492	ng		91
18) Hexachloroethane	7.392	117	8945	5.108	ng		99
19) n-Nitroso-di-n-propyla...	7.281	70	15719	5.380	ng		92
22) Acetophenone	7.281	105	28341	5.471	ng		94
24) Nitrobenzene	7.457	77	23655	5.111	ng		97
25) Isophorone	7.692	82	39860	5.211	ng		98
26) 2-Nitrophenol	7.775	139	11373	5.166	ng	#	89
27) 2,4-Dimethylphenol	7.810	122	17297	5.287	ng		99
28) bis(2-Chloroethoxy)met...	7.904	93	25751	5.399	ng		99
29) 2,4-Dichlorophenol	8.022	162	20009	5.259	ng		96
30) 1,2,4-Trichlorobenzene	8.098	180	20600	5.263	ng		99
31) Naphthalene	8.181	128	63185	5.585	ng		99
33) 4-Chloroaniline	8.233	127	19875	5.047	ng		94
34) Hexachlorobutadiene	8.298	225	16702	5.218	ng		97
36) 4-Chloro-3-methylphenol	8.710	107	18077	5.384	ng		95
37) 2-Methylnaphthalene	8.875	142	43006	5.766	ng		98
38) 1-Methylnaphthalene	8.975	142	41029	5.806	ng		95
40) 1,2,4,5-Tetrachloroben...	9.039	216	24701	5.651	ng		96
43) 2,4,6-Trichlorophenol	9.151	196	13673	4.813	ng		93
44) 2,4,5-Trichlorophenol	9.198	196	15530	5.208	ng		94
46) 1,1'-Biphenyl	9.333	154	51299	5.635	ng		98
47) 2-Chloronaphthalene	9.363	162	42492	5.571	ng		97
48) 2-Nitroaniline	9.463	65	10347	4.568	ng		94
49) Acenaphthylene	9.775	152	56586	5.502	ng		99
50) Dimethylphthalate	9.627	163	44452	5.332	ng		99
51) 2,6-Dinitrotoluene	9.698	165	7239	4.392	ng		93

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144057.D
Acq On : 24 Oct 2025 10:42
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Oct 24 17:01:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
52) Acenaphthene	9.951	154	35769	5.208	ng		98
53) 3-Nitroaniline	9.875	138	8780	4.792	ng		97
55) Dibenzofuran	10.122	168	53591	5.620	ng		95
57) 2,4-Dinitrotoluene	10.104	165	9591	4.312	ng	#	94
58) Fluorene	10.469	166	43217	5.948	ng		98
59) 2,3,4,6-Tetrachlorophenol	10.245	232	10071	4.793	ng	#	97
60) Diethylphthalate	10.333	149	40882	5.215	ng		97
61) 4-Chlorophenyl-phenyle...	10.457	204	22434	5.491	ng		97
62) 4-Nitroaniline	10.480	138	8315	4.789	ng		87
63) Azobenzene	10.616	77	37885	5.106	ng		96
66) n-Nitrosodiphenylamine	10.574	169	38552	5.599	ng		95
67) 4-Bromophenyl-phenylether	10.951	248	14748	5.466	ng		94
68) Hexachlorobenzene	11.016	284	16776	5.070	ng		96
69) Atrazine	11.098	200	12331	5.458	ng		95
71) Phenanthrene	11.433	178	57785	5.398	ng		99
72) Anthracene	11.486	178	63077	5.704	ng		97
73) Carbazole	11.645	167	50818	5.399	ng		98
74) Di-n-butylphthalate	11.969	149	58713	5.088	ng		99
75) Fluoranthene	12.627	202	63073	5.333	ng		97
78) Pyrene	12.857	202	68289	5.573	ng		98
80) Butylbenzylphthalate	13.474	149	21644	4.669	ng		94
81) Benzo(a)anthracene	14.045	228	57138	4.960	ng		99
83) Chrysene	14.080	228	56694	5.245	ng		98
84) Bis(2-ethylhexyl)phtha...	14.033	149	31727	5.013	ng		99
87) Indeno(1,2,3-cd)pyrene	17.039	276	64345	4.918	ng		98
88) Benzo(b)fluoranthene	15.104	252	54015	5.014	ng		97
89) Benzo(k)fluoranthene	15.133	252	57466	5.782	ng		99
90) Benzo(a)pyrene	15.474	252	49993	5.173	ng	#	99
91) Dibenzo(a,h)anthracene	17.056	278	50377	4.865	ng		98
92) Benzo(g,h,i)perylene	17.486	276	50046	4.807	ng		98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144057.D
Acq On : 24 Oct 2025 10:42
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Oct 24 17:01:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 79
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:152 Resp: 64729

Ion	Ratio	Lower	Upper
152	100		
150	159.2	124.4	186.6
115	64.1	49.4	74.0







#2
1,4-Dioxane
Concen: 6.619 ng
RT: 2.640 min Scan# 76
Delta R.T. 0.012 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion: 88 Resp: 11703

Ion	Ratio	Lower	Upper
88	100		
58	91.4	76.5	114.7
43	39.6	35.3	52.9







#3
Pyridine
Concen: 4.769 ng
RT: 3.428 min Scan# 21
Delta R.T. 0.029 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 22571

Ion	Ratio	Lower	Upper
79	100		
52	84.2	67.5	101.3
51	51.2	33.6	50.4#







#5
2-Fluorophenol
Concen: 10.471 ng
RT: 5.492 min Scan# 561
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:112 Resp: 42841

Ion	Ratio	Lower	Upper
112	100		
64	61.9	49.8	74.8
63	29.9	27.4	41.2







#6
Aniline
Concen: 5.441 ng
RT: 6.539 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 35169

Ion	Ratio	Lower	Upper
93	100		
66	45.3	35.7	53.5
65	21.6	17.0	25.6







#7
Phenol-d6
Concen: 11.088 ng
RT: 6.498 min Scan# 732
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion: 99 Resp: 50099

Ion	Ratio	Lower	Upper
99	100		
42	26.7	23.9	35.9
71	37.5	34.4	51.6







#8
2-Chlorophenol
Concen: 5.615 ng
RT: 6.663 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:128 Resp: 22623

Ion	Ratio	Lower	Upper
128	100		
130	32.9	10.7	50.7
64	48.4	34.7	74.7







#10
Phenol
Concen: 5.505 ng
RT: 6.516 min Scan# 735
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion: 94 Resp: 30686

Ion	Ratio	Lower	Upper
94	100		
65	29.4	13.1	53.1
66	43.9	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 5.650 ng
RT: 6.610 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 22757

Ion	Ratio	Lower	Upper
93	100		
63	80.0	60.2	100.2
95	34.5	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 5.474 ng
RT: 6.822 min Scan# 787
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:146 Resp: 27208

Ion	Ratio	Lower	Upper
146	100		
148	63.9	50.6	75.8
75	30.3	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 5.433 ng
RT: 6.898 min Scan# 80
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:146 Resp: 26614

Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.0	76.6
111	40.2	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 5.423 ng
RT: 7.045 min Scan# 825
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:146 Resp: 25700

Ion	Ratio	Lower	Upper
146	100		
148	68.2	50.8	76.2
111	45.3	34.7	52.1







#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.474 ng
RT: 7.145 min Scan# 84
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 45 Resp: 41443

Ion	Ratio	Lower	Upper
45	100		
77	12.7	0.0	30.9
79	10.8	0.0	28.2







#17
2-Methylphenol
Concen: 5.492 ng
RT: 7.128 min Scan# 839
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:107 Resp: 17019

Ion	Ratio	Lower	Upper
107	100		
108	112.5	96.4	144.6
77	49.2	46.8	70.2
79	50.6	45.8	68.6







#18
Hexachloroethane
Concen: 5.108 ng
RT: 7.392 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:117 Resp: 8945

Ion	Ratio	Lower	Upper
117	100		
119	96.9	76.6	115.0
201	124.4	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 5.380 ng
RT: 7.281 min Scan# 865
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion: 70 Resp: 15719

Ion	Ratio	Lower	Upper
70	100		
42	74.8	67.4	101.0
101	7.9	7.1	10.7
130	17.9	13.5	20.3







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 10
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:136 Resp: 239688

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	13.0
54	8.4	6.6	9.8
68	5.6	4.6	6.8







#22
Acetophenone
Concen: 5.471 ng
RT: 7.281 min Scan# 865
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:105 Resp: 28341

Ion	Ratio	Lower	Upper
105	100		
71	9.0	6.9	10.3
51	45.7	40.6	60.8
120	23.6	16.8	25.2







#23
Nitrobenzene-d5
Concen: 10.317 ng
RT: 7.434 min Scan# 89
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 82 Resp: 45136

Ion	Ratio	Lower	Upper
82	100		
128	40.9	32.4	48.6
54	62.7	54.0	81.0







#24
Nitrobenzene
Concen: 5.111 ng
RT: 7.457 min Scan# 895
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion: 77 Resp: 23655

Ion	Ratio	Lower	Upper
77	100		
123	46.1	35.0	52.6
65	15.4	12.2	18.2







#25
Isophorone
Concen: 5.211 ng
RT: 7.692 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 82 Resp: 39860

Ion	Ratio	Lower	Upper
82	100		
95	7.7	6.6	9.8
138	18.2	13.8	20.8







#26
2-Nitrophenol
Concen: 5.166 ng
RT: 7.775 min Scan# 949
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:139 Resp: 11373

Ion	Ratio	Lower	Upper
139	100		
109	24.1	24.6	37.0#
65	50.3	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 5.287 ng
RT: 7.810 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:122 Resp: 17297

Ion	Ratio	Lower	Upper
122	100		
107	115.2	91.8	137.8
121	55.0	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 5.399 ng
RT: 7.904 min Scan# 971
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion: 93 Resp: 25751

Ion	Ratio	Lower	Upper
93	100		
95	30.5	24.6	36.8
123	10.6	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 5.259 ng
RT: 8.022 min Scan# 99
Delta R.T. 0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:162 Resp: 20009

Ion	Ratio	Lower	Upper
162	100		
164	65.5	42.1	82.1
98	33.3	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 5.263 ng
RT: 8.098 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:180 Resp: 20600

Ion	Ratio	Lower	Upper
180	100		
182	108.9	87.1	130.7
145	33.4	25.3	37.9







#31
Naphthalene
Concen: 5.585 ng
RT: 8.181 min Scan# 10
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:128 Resp: 63185

Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.4	12.6
127	12.7	9.9	14.9







#33
4-Chloroaniline
Concen: 5.047 ng
RT: 8.233 min Scan# 1027
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:127 Resp: 19875

Ion	Ratio	Lower	Upper
127	100		
129	38.2	25.6	38.4
65	35.0	30.9	46.3
92	24.0	18.9	28.3







#34
Hexachlorobutadiene
Concen: 5.218 ng
RT: 8.298 min Scan# 10
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:225 Resp: 16702

Ion	Ratio	Lower	Upper
225	100		
223	64.8	50.6	76.0
227	68.4	51.8	77.8







#36
4-Chloro-3-methylphenol
Concen: 5.384 ng
RT: 8.710 min Scan# 1108
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:107 Resp: 18077

Ion	Ratio	Lower	Upper
107	100		
144	31.5	21.6	32.4
142	85.4	65.4	98.0







#37
2-Methylnaphthalene
Concen: 5.766 ng
RT: 8.875 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:142 Resp: 43006

Ion	Ratio	Lower	Upper
142	100		
141	86.0	67.6	101.4







#38
1-Methylnaphthalene
Concen: 5.806 ng
RT: 8.975 min Scan# 1153
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:142 Resp: 41029

Ion	Ratio	Lower	Upper
142	100		
141	87.5	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:164 Resp: 127182

Ion	Ratio	Lower	Upper
164	100		
162	104.2	79.4	119.2
160	45.0	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.651 ng
RT: 9.039 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:216 Resp: 24701

Ion	Ratio	Lower	Upper
216	100		
214	75.6	63.2	94.8
179	17.3	15.2	22.8
108	13.5	11.9	17.9







#42
2,4,6-Tribromophenol
Concen: 9.848 ng
RT: 10.710 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:330 Resp: 15954

Ion	Ratio	Lower	Upper
330	100		
332	89.7	77.0	115.6
141	21.6	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 4.813 ng
RT: 9.151 min Scan# 1183
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:196 Resp: 13673

Ion	Ratio	Lower	Upper
196	100		
198	99.3	74.4	111.6
200	34.0	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 5.208 ng
RT: 9.198 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:196 Resp: 15530

Ion	Ratio	Lower	Upper
196	100		
198	92.9	78.0	117.0
97	40.7	37.7	56.5
132	28.5	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 11.641 ng
RT: 9.233 min Scan# 1197
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:172 Resp: 105210

Ion	Ratio	Lower	Upper
172	100		
171	33.5	27.5	41.3
170	24.2	19.0	28.6







#46
1,1'-Biphenyl
Concen: 5.635 ng
RT: 9.333 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 51299

Ion	Ratio	Lower	Upper
154	100		
153	38.6	19.7	59.7
76	11.6	0.0	31.1







#47
2-Chloronaphthalene
Concen: 5.571 ng
RT: 9.363 min Scan# 1219
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:162 Resp: 42492

Ion	Ratio	Lower	Upper
162	100		
127	36.4	31.6	47.4
164	33.0	26.5	39.7







#48
2-Nitroaniline
Concen: 4.568 ng
RT: 9.463 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 65 Resp: 10347

Ion	Ratio	Lower	Upper
65	100		
92	65.5	52.3	78.5
138	107.7	78.3	117.5







#49
Acenaphthylene
Concen: 5.502 ng
RT: 9.775 min Scan# 1289
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:152 Resp: 56586

Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.9	23.9
153	12.6	10.4	15.6







#50
Dimethylphthalate
Concen: 5.332 ng
RT: 9.627 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:163 Resp: 44452

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.0
164	9.5	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 4.392 ng
RT: 9.698 min Scan# 1276
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:165 Resp: 7239

Ion	Ratio	Lower	Upper
165	100		
63	61.0	50.4	75.6
89	43.6	41.4	62.2







#52
Acenaphthene
Concen: 5.208 ng
RT: 9.951 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 35769

Ion	Ratio	Lower	Upper
154	100		
153	114.6	89.9	134.9
152	52.1	42.0	63.0







#53
3-Nitroaniline
Concen: 4.792 ng
RT: 9.875 min Scan# 1306
Delta R.T. 0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:138 Resp: 8780

Ion	Ratio	Lower	Upper
138	100		
108	11.1	7.9	11.9
92	111.8	92.0	138.0







#55
Dibenzofuran
Concen: 5.620 ng
RT: 10.122 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:168 Resp: 53591

Ion	Ratio	Lower	Upper
168	100		
139	37.1	33.2	49.8
169	15.2	12.6	19.0







#57
2,4-Dinitrotoluene
Concen: 4.312 ng
RT: 10.104 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:165 Resp: 9591

Ion	Ratio	Lower	Upper
165	100		
63	49.3	35.4	53.2
89	62.0	52.7	79.1
182	0.0	2.3	3.5#







#58
Fluorene
Concen: 5.948 ng
RT: 10.469 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:166 Resp: 43217

Ion	Ratio	Lower	Upper
166	100		
165	98.3	77.0	115.4
167	13.0	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 4.793 ng
RT: 10.245 min Scan# 1369
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:232 Resp: 10071

Ion	Ratio	Lower	Upper
232	100		
131	36.6	31.9	47.9
130	0.0	1.1	1.7#
166	26.4	21.0	31.4







#60
Diethylphthalate
Concen: 5.215 ng
RT: 10.333 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 40882

Ion	Ratio	Lower	Upper
149	100		
177	24.4	17.8	26.6
150	13.1	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 5.491 ng
RT: 10.457 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:204 Resp: 22434

Ion	Ratio	Lower	Upper
204	100		
206	35.9	27.8	41.6
141	54.1	45.4	68.2







#62
4-Nitroaniline
Concen: 4.789 ng
RT: 10.480 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:138 Resp: 8315

Ion	Ratio	Lower	Upper
138	100		
92	42.8	30.5	70.5
108	64.5	56.2	96.2







#63
Azobenzene
Concen: 5.106 ng
RT: 10.616 min Scan# 1432
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion: 77 Resp: 37885

Ion	Ratio	Lower	Upper
77	100		
182	29.9	9.0	49.0
105	24.4	1.4	41.4
51	34.6	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:188 Resp: 215493

Ion	Ratio	Lower	Upper
188	100		
94	7.7	5.4	8.0
80	7.5	6.0	9.0







#66
n-Nitrosodiphenylamine
Concen: 5.599 ng
RT: 10.574 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:169 Resp: 38552

Ion	Ratio	Lower	Upper
169	100		
168	55.0	47.4	71.2
167	34.1	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 5.466 ng
RT: 10.951 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:248 Resp: 14748

Ion	Ratio	Lower	Upper
248	100		
250	106.1	77.8	116.6
141	62.8	50.5	75.7







#68
Hexachlorobenzene
Concen: 5.070 ng
RT: 11.016 min Scan# 1500
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:284 Resp: 16776

Ion	Ratio	Lower	Upper
284	100		
142	25.5	20.0	30.0
249	31.5	22.2	33.2







#69
Atrazine
Concen: 5.458 ng
RT: 11.098 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:200 Resp: 12331

Ion	Ratio	Lower	Upper
200	100		
173	26.2	4.9	44.9
215	44.1	28.2	68.2







#71
Phenanthrene
Concen: 5.398 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:178 Resp: 57785

Ion	Ratio	Lower	Upper
178	100		
176	21.2	16.3	24.5
179	16.0	12.6	18.8







#72
Anthracene
Concen: 5.704 ng
RT: 11.486 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:178 Resp: 63077

Ion	Ratio	Lower	Upper
178	100		
176	17.6	15.0	22.4
179	13.2	11.8	17.6







#73
Carbazole
Concen: 5.399 ng
RT: 11.645 min Scan# 1607
Delta R.T. 0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:167 Resp: 50818

Ion	Ratio	Lower	Upper
167	100		
166	21.6	16.1	24.1
139	12.6	9.8	14.6







#74
Di-n-butylphthalate
Concen: 5.088 ng
RT: 11.969 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 58713

Ion	Ratio	Lower	Upper
149	100		
150	8.7	7.3	10.9
104	4.4	3.6	5.4







#75
Fluoranthene
Concen: 5.333 ng
RT: 12.627 min Scan# 1774
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:202 Resp: 63073

Ion	Ratio	Lower	Upper
202	100		
101	7.5	0.0	27.7
203	15.7	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2(
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:240 Resp: 170453

Ion	Ratio	Lower	Upper
240	100		
120	7.5	6.5	9.7
236	24.8	20.1	30.1







#78
Pyrene
Concen: 5.573 ng
RT: 12.857 min Scan# 1813
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:202 Resp: 68289

Ion	Ratio	Lower	Upper
202	100		
200	20.4	17.2	25.8
203	16.5	13.5	20.3







#79
Terphenyl-d14
Concen: 11.189 ng
RT: 12.992 min Scan# 18
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:244 Resp: 109089

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.8	8.6
122	7.7	4.9	7.3#







#80
Butylbenzylphthalate
Concen: 4.669 ng
RT: 13.474 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:149 Resp: 21644

Ion	Ratio	Lower	Upper
149	100		
91	66.1	56.6	84.8
206	25.6	18.3	27.5







#81
Benzo(a)anthracene
Concen: 4.960 ng
RT: 14.045 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:228 Resp: 57138

Ion	Ratio	Lower	Upper
228	100		
226	26.3	21.4	32.2
229	19.7	15.3	22.9







#83
Chrysene
Concen: 5.245 ng
RT: 14.080 min Scan# 2021
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:228 Resp: 56694

Ion	Ratio	Lower	Upper
228	100		
226	28.2	23.8	35.6
229	19.2	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 5.013 ng
RT: 14.033 min Scan# 20
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 31727

Ion	Ratio	Lower	Upper
149	100		
167	27.3	22.5	33.7
279	4.5	3.6	5.4







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:264 Resp: 190551

Ion	Ratio	Lower	Upper
264	100		
260	21.4	17.0	25.4
265	21.5	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 4.918 ng
RT: 17.039 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:276 Resp: 64345

Ion	Ratio	Lower	Upper
276	100		
138	18.4	13.0	19.6
277	24.2	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 5.014 ng
RT: 15.104 min Scan# 2195
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:252 Resp: 54015

Ion	Ratio	Lower	Upper
252	100		
253	19.7	17.2	25.8
125	6.8	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 5.782 ng
RT: 15.133 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 57466

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.2
125	6.7	4.9	7.3







#90
Benzo(a)pyrene
Concen: 5.173 ng
RT: 15.474 min Scan# 2258
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:252 Resp: 49993

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.1	25.7
125	8.5	5.4	8.2#







#91
Dibenzo(a,h)anthracene
Concen: 4.865 ng
RT: 17.056 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:278 Resp: 50377

Ion	Ratio	Lower	Upper
278	100		
139	11.9	8.5	12.7
279	24.2	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 4.807 ng
RT: 17.486 min Scan# 2600
Delta R.T. -0.012 min
Lab File: BF144057.D
Acq: 24 Oct 2025 10:42

Tgt Ion:276 Resp: 50046

Ion	Ratio	Lower	Upper
276	100		
277	20.9	17.9	26.9
138	14.5	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144058.D
Acq On : 24 Oct 2025 11:39
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Oct 24 17:01:53 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	93506	20.000	ng		0.00
21) Naphthalene-d8	8.163	136	345633	20.000	ng		0.00
39) Acenaphthene-d10	9.916	164	181955	20.000	ng		0.00
64) Phenanthrene-d10	11.410	188	301308	20.000	ng		0.00
76) Chrysene-d12	14.057	240	243705	20.000	ng		0.00
86) Perylene-d12	15.539	264	290045	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.493	112	123660	20.923	ng		0.00
7) Phenol-d6	6.498	99	134938	20.673	ng		0.00
23) Nitrobenzene-d5	7.440	82	130813	20.735	ng		0.00
42) 2,4,6-Tribromophenol	10.704	330	45367	19.574	ng		0.00
45) 2-Fluorobiphenyl	9.234	172	273504	21.152	ng		0.00
79) Terphenyl-d14	12.992	244	276477	19.833	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.634	88	23088	9.040	ng		96
3) Pyridine	3.410	79	64703	9.463	ng		95
4) n-Nitrosodimethylamine	3.340	42	37657	8.963	ng		98
6) Aniline	6.540	93	100630	10.776	ng		97
8) 2-Chlorophenol	6.663	128	61265	10.526	ng		96
9) Benzaldehyde	6.428	77	51089	10.081	ng		99
10) Phenol	6.516	94	86962	10.799	ng		95
11) bis(2-Chloroethyl)ether	6.610	93	59764	10.271	ng		99
12) 1,3-Dichlorobenzene	6.816	146	73976	10.303	ng		99
13) 1,4-Dichlorobenzene	6.898	146	73246	10.352	ng		96
14) 1,2-Dichlorobenzene	7.045	146	72385	10.573	ng		97
15) Benzyl Alcohol	7.016	79	52766	9.942	ng		95
16) 2,2'-oxybis(1-Chloropr...	7.145	45	113682	10.395	ng		98
17) 2-Methylphenol	7.128	107	47217	10.548	ng		93
18) Hexachloroethane	7.393	117	25444	10.058	ng		95
19) n-Nitroso-di-n-propyla...	7.281	70	45435	10.764	ng		92
20) 3+4-Methylphenols	7.281	107	59370	11.219	ng	#	82
22) Acetophenone	7.287	105	79330	10.620	ng		91
24) Nitrobenzene	7.457	77	66345	9.941	ng		95
25) Isophorone	7.692	82	114333	10.366	ng		99
26) 2-Nitrophenol	7.775	139	31721	9.992	ng		98
27) 2,4-Dimethylphenol	7.810	122	48192	10.215	ng		100
28) bis(2-Chloroethoxy)met...	7.904	93	70437	10.241	ng		96
29) 2,4-Dichlorophenol	8.016	162	59570	10.858	ng		96
30) 1,2,4-Trichlorobenzene	8.098	180	58514	10.367	ng		99
31) Naphthalene	8.181	128	173217	10.618	ng		99
32) Benzoic acid	7.904	122	37467	12.520	ng	#	81
33) 4-Chloroaniline	8.234	127	61541	10.837	ng		96
34) Hexachlorobutadiene	8.298	225	47643	10.323	ng		96
35) Caprolactam	8.575	113	13454	9.779	ng		92
36) 4-Chloro-3-methylphenol	8.704	107	50000	10.327	ng		95
37) 2-Methylnaphthalene	8.875	142	113005	10.507	ng		98
38) 1-Methylnaphthalene	8.975	142	108437	10.642	ng		99
40) 1,2,4,5-Tetrachloroben...	9.039	216	64302	10.283	ng		99
41) Hexachlorocyclopentadiene	9.028	237	13467	6.972	ng		97
43) 2,4,6-Trichlorophenol	9.151	196	40647	10.000	ng		96

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144058.D
Acq On : 24 Oct 2025 11:39
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Oct 24 17:01:53 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	44114	10.341	ng	96
46) 1,1'-Biphenyl	9.334	154	139423	10.704	ng	99
47) 2-Chloronaphthalene	9.363	162	115014	10.540	ng	94
48) 2-Nitroaniline	9.457	65	32649	10.076	ng	99
49) Acenaphthylene	9.775	152	154058	10.470	ng	99
50) Dimethylphthalate	9.628	163	120856	10.133	ng	99
51) 2,6-Dinitrotoluene	9.698	165	23341	9.899	ng	93
52) Acenaphthene	9.951	154	104855	10.671	ng	99
53) 3-Nitroaniline	9.869	138	26729	10.196	ng	97
54) 2,4-Dinitrophenol	9.981	184	14660	11.496	ng	91
55) Dibenzofuran	10.122	168	142976	10.481	ng	98
56) 4-Nitrophenol	10.028	139	20383	11.482	ng	94
57) 2,4-Dinitrotoluene	10.098	165	31163	9.793	ng	96
58) Fluorene	10.469	166	106355	10.232	ng	96
59) 2,3,4,6-Tetrachlorophenol	10.239	232	28893	9.612	ng	99
60) Diethylphthalate	10.334	149	115029	10.256	ng	99
61) 4-Chlorophenyl-phenyle...	10.457	204	60704	10.385	ng	97
62) 4-Nitroaniline	10.481	138	24222	9.752	ng	98
63) Azobenzene	10.616	77	107163	10.095	ng	98
65) 4,6-Dinitro-2-methylph...	10.510	198	20093	10.394	ng	95
66) n-Nitrosodiphenylamine	10.575	169	100334	10.422	ng	97
67) 4-Bromophenyl-phenylether	10.951	248	38639	10.242	ng	99
68) Hexachlorobenzene	11.016	284	50850	10.990	ng	96
69) Atrazine	11.098	200	33605	10.638	ng	99
70) Pentachlorophenol	11.216	266	18468	8.386	ng	97
71) Phenanthrene	11.433	178	156663	10.467	ng	96
72) Anthracene	11.486	178	162203	10.491	ng	100
73) Carbazole	11.639	167	138634	10.534	ng	99
74) Di-n-butylphthalate	11.969	149	164302	10.183	ng	100
75) Fluoranthene	12.622	202	173397	10.486	ng	98
77) Benzidine	12.751	184	83945	11.068	ng	99
78) Pyrene	12.857	202	178638	10.196	ng	99
80) Butylbenzylphthalate	13.469	149	66048	9.964	ng	97
81) Benzo(a)anthracene	14.045	228	161042	9.778	ng	99
82) 3,3'-Dichlorobenzidine	14.010	252	57120	9.963	ng	97
83) Chrysene	14.080	228	153845	9.954	ng	99
84) Bis(2-ethylhexyl)phtha...	14.033	149	88365	9.765	ng	99
85) Di-n-octyl phthalate	14.645	149	149635	9.616	ng	97
87) Indeno(1,2,3-cd)pyrene	17.039	276	193064	9.695	ng	99
88) Benzo(b)fluoranthene	15.104	252	160748	9.803	ng	100
89) Benzo(k)fluoranthene	15.133	252	156128	10.320	ng	100
90) Benzo(a)pyrene	15.474	252	143530	9.757	ng	99
91) Dibenzo(a,h)anthracene	17.051	278	153941	9.766	ng	98
92) Benzo(g,h,i)perylene	17.486	276	152368	9.614	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144058.D
Acq On : 24 Oct 2025 11:39
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Oct 24 17:01:53 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 93506

Ion	Ratio	Lower	Upper
152	100		
150	154.9	124.4	186.6
115	62.6	49.4	74.0







#2
1,4-Dioxane
Concen: 9.040 ng
RT: 2.634 min Scan# 75
Delta R.T. 0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion: 88 Resp: 23088

Ion	Ratio	Lower	Upper
88	100		
58	89.8	76.5	114.7
43	43.6	35.3	52.9







#3
Pyridine
Concen: 9.463 ng
RT: 3.410 min Scan# 20
Delta R.T. 0.012 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 64703

Ion	Ratio	Lower	Upper
79	100		
52	80.8	67.5	101.3
51	37.1	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 8.963 ng
RT: 3.340 min Scan# 195
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion: 42 Resp: 37657

Ion	Ratio	Lower	Upper
42	100		
74	93.6	76.4	114.6
44	7.0	5.2	7.8







#5
2-Fluorophenol
Concen: 20.923 ng
RT: 5.493 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:112 Resp: 123660

Ion	Ratio	Lower	Upper
112	100		
64	55.8	49.8	74.8
63	30.2	27.4	41.2







#6
Aniline
Concen: 10.776 ng
RT: 6.540 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion: 93 Resp: 100630

Ion	Ratio	Lower	Upper
93	100		
66	41.9	35.7	53.5
65	20.6	17.0	25.6







#7
Phenol-d6
Concen: 20.673 ng
RT: 6.498 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 99 Resp: 134938

Ion	Ratio	Lower	Upper
99	100		
42	29.9	23.9	35.9
71	41.6	34.4	51.6







#8
2-Chlorophenol
Concen: 10.526 ng
RT: 6.663 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:128 Resp: 61265

Ion	Ratio	Lower	Upper
128	100		
130	28.8	10.7	50.7
64	51.5	34.7	74.7







#9
Benzaldehyde
Concen: 10.081 ng
RT: 6.428 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 51089

Ion	Ratio	Lower	Upper
77	100		
105	91.6	72.2	112.2
106	88.5	67.9	107.9







#10
Phenol
Concen: 10.799 ng
RT: 6.516 min Scan# 735
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion: 94 Resp: 86962

Ion	Ratio	Lower	Upper
94	100		
65	30.2	13.1	53.1
66	46.9	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 10.271 ng
RT: 6.610 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 59764

Ion	Ratio	Lower	Upper
93	100		
63	79.2	60.2	100.2
95	32.3	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 10.303 ng
RT: 6.816 min Scan# 786
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:146 Resp: 73976

Ion	Ratio	Lower	Upper
146	100		
148	63.9	50.6	75.8
75	32.3	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 10.352 ng
RT: 6.898 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:146 Resp: 73246

Ion	Ratio	Lower	Upper
146	100		
148	66.0	51.0	76.6
111	39.5	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 10.573 ng
RT: 7.045 min Scan# 825
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:146 Resp: 72385

Ion	Ratio	Lower	Upper
146	100		
148	61.1	50.8	76.2
111	42.3	34.7	52.1







#15
Benzyl Alcohol
Concen: 9.942 ng
RT: 7.016 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 52766

Ion	Ratio	Lower	Upper
79	100		
108	73.5	54.8	82.2
77	58.8	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.395 ng
RT: 7.145 min Scan# 842
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion: 45 Resp: 113682

Ion	Ratio	Lower	Upper
45	100		
77	12.4	0.0	30.9
79	8.2	0.0	28.2







#17
2-Methylphenol
Concen: 10.548 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:107 Resp: 47217

Ion	Ratio	Lower	Upper
107	100		
108	112.7	96.4	144.6
77	55.0	46.8	70.2
79	51.6	45.8	68.6







#18
Hexachloroethane
Concen: 10.058 ng
RT: 7.393 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:117 Resp: 25444

Ion	Ratio	Lower	Upper
117	100		
119	94.8	76.6	115.0
201	117.6	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 10.764 ng
RT: 7.281 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 70 Resp: 45435

Ion	Ratio	Lower	Upper
70	100		
42	75.2	67.4	101.0
101	7.8	7.1	10.7
130	15.8	13.5	20.3







#20
3+4-Methylphenols
Concen: 11.219 ng
RT: 7.281 min Scan# 865
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:107 Resp: 59370

Ion	Ratio	Lower	Upper
107	100		
108	90.1	68.0	108.0
77	166.0	108.7	148.7#
79	34.8	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:136 Resp: 345633

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	13.0
54	8.5	6.6	9.8
68	5.4	4.6	6.8







#22
Acetophenone
Concen: 10.620 ng
RT: 7.287 min Scan# 866
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:105 Resp: 79330

Ion	Ratio	Lower	Upper
105	100		
71	7.8	6.9	10.3
51	42.0	40.6	60.8
120	21.7	16.8	25.2







#23
Nitrobenzene-d5
Concen: 20.735 ng
RT: 7.440 min Scan# 89
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 130813

Ion	Ratio	Lower	Upper
82	100		
128	41.0	32.4	48.6
54	62.2	54.0	81.0







#24
Nitrobenzene
Concen: 9.941 ng
RT: 7.457 min Scan# 895
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion: 77 Resp: 66345

Ion	Ratio	Lower	Upper
77	100		
123	47.4	35.0	52.6
65	16.4	12.2	18.2







#25
Isophorone
Concen: 10.366 ng
RT: 7.692 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 114333

Ion	Ratio	Lower	Upper
82	100		
95	7.9	6.6	9.8
138	16.8	13.8	20.8







#26
2-Nitrophenol
Concen: 9.992 ng
RT: 7.775 min Scan# 949
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:139 Resp: 31721

Ion	Ratio	Lower	Upper
139	100		
109	31.1	24.6	37.0
65	56.3	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 10.215 ng
RT: 7.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:122 Resp: 48192

Ion	Ratio	Lower	Upper
122	100		
107	114.9	91.8	137.8
121	56.9	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 10.241 ng
RT: 7.904 min Scan# 971
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion: 93 Resp: 70437

Ion	Ratio	Lower	Upper
93	100		
95	32.9	24.6	36.8
123	12.9	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 10.858 ng
RT: 8.016 min Scan# 99
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 59570

Ion	Ratio	Lower	Upper
162	100		
164	59.2	42.1	82.1
98	32.7	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 10.367 ng
RT: 8.098 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:180 Resp: 58514

Ion	Ratio	Lower	Upper
180	100		
182	107.4	87.1	130.7
145	30.9	25.3	37.9







#31
Naphthalene
Concen: 10.618 ng
RT: 8.181 min Scan# 10
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:128 Resp: 173217

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.4	12.6
127	12.6	9.9	14.9







#32
Benzoic acid
Concen: 12.520 ng
RT: 7.904 min Scan# 971
Delta R.T. -0.012 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:122 Resp: 37467

Ion	Ratio	Lower	Upper
122	100		
105	111.5	111.3	151.3
77	92.4	96.7	136.7#







#33
4-Chloroaniline
Concen: 10.837 ng
RT: 8.234 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:127 Resp: 61541

Ion	Ratio	Lower	Upper
127	100		
129	31.5	25.6	38.4
65	34.2	30.9	46.3
92	21.8	18.9	28.3







#34
Hexachlorobutadiene
Concen: 10.323 ng
RT: 8.298 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:225 Resp: 47643

Ion	Ratio	Lower	Upper
225	100		
223	60.7	50.6	76.0
227	61.5	51.8	77.8







#35
Caprolactam
Concen: 9.779 ng
RT: 8.575 min Scan# 10
Delta R.T. -0.017 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:113 Resp: 13454

Ion	Ratio	Lower	Upper
113	100		
55	194.9	160.1	200.1
56	151.9	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 10.327 ng
RT: 8.704 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:107 Resp: 50000

Ion	Ratio	Lower	Upper
107	100		
144	26.5	21.6	32.4
142	76.6	65.4	98.0







#37
2-Methylnaphthalene
Concen: 10.507 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:142 Resp: 113005

Ion	Ratio	Lower	Upper
142	100		
141	86.2	67.6	101.4







#38
1-Methylnaphthalene
Concen: 10.642 ng
RT: 8.975 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:142 Resp: 108437

Ion	Ratio	Lower	Upper
142	100		
141	92.4	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:164 Resp: 181955

Ion	Ratio	Lower	Upper
164	100		
162	100.4	79.4	119.2
160	44.5	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.283 ng
RT: 9.039 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:216 Resp: 64302

Ion	Ratio	Lower	Upper
216	100		
214	78.8	63.2	94.8
179	21.3	15.2	22.8
108	14.4	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 6.972 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:237 Resp: 13467

Ion	Ratio	Lower	Upper
237	100		
235	61.7	44.3	84.3
272	14.4	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 19.574 ng
RT: 10.704 min Scan# 1447
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:330 Resp: 45367

Ion	Ratio	Lower	Upper
330	100		
332	94.9	77.0	115.6
141	24.4	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 10.000 ng
RT: 9.151 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:196 Resp: 40647

Ion	Ratio	Lower	Upper
196	100		
198	96.6	74.4	111.6
200	31.4	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 10.341 ng
RT: 9.192 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:196 Resp: 44114

Ion	Ratio	Lower	Upper
196	100		
198	94.2	78.0	117.0
97	41.9	37.7	56.5
132	25.0	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 21.152 ng
RT: 9.234 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:172 Resp: 273504

Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.5	41.3
170	23.3	19.0	28.6







#46
1,1'-Biphenyl
Concen: 10.704 ng
RT: 9.334 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:154 Resp: 139423

Ion	Ratio	Lower	Upper
154	100		
153	39.2	19.7	59.7
76	11.1	0.0	31.1







#47
2-Chloronaphthalene
Concen: 10.540 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 115014

Ion	Ratio	Lower	Upper
162	100		
127	34.7	31.6	47.4
164	31.5	26.5	39.7







#48
2-Nitroaniline
Concen: 10.076 ng
RT: 9.457 min Scan# 1235
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion: 65 Resp: 32649

Ion	Ratio	Lower	Upper
65	100		
92	66.4	52.3	78.5
138	96.6	78.3	117.5







#49
Acenaphthylene
Concen: 10.470 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 154058

Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.9	23.9
153	12.7	10.4	15.6







#50
Dimethylphthalate
Concen: 10.133 ng
RT: 9.628 min Scan# 1264
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:163 Resp: 120856

Ion	Ratio	Lower	Upper
163	100		
194	4.5	3.4	5.0
164	10.1	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 9.899 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 23341

Ion	Ratio	Lower	Upper
165	100		
63	56.7	50.4	75.6
89	48.3	41.4	62.2







#52
Acenaphthene
Concen: 10.671 ng
RT: 9.951 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:154 Resp: 104855

Ion	Ratio	Lower	Upper
154	100		
153	113.5	89.9	134.9
152	53.0	42.0	63.0







#53
3-Nitroaniline
Concen: 10.196 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:138 Resp: 26729

Ion	Ratio	Lower	Upper
138	100		
108	10.1	7.9	11.9
92	111.9	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 11.496 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:184 Resp: 14660

Ion	Ratio	Lower	Upper
184	100		
63	58.3	52.1	78.1
154	67.4	47.9	71.9







#55
Dibenzofuran
Concen: 10.481 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:168 Resp: 142976

Ion	Ratio	Lower	Upper
168	100		
139	40.3	33.2	49.8
169	15.5	12.6	19.0







#56
4-Nitrophenol
Concen: 11.482 ng
RT: 10.028 min Scan# 1332
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:139 Resp: 20383

Ion	Ratio	Lower	Upper
139	100		
109	69.1	53.8	93.8
65	109.7	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 9.793 ng
RT: 10.098 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 31163

Ion	Ratio	Lower	Upper
165	100		
63	49.5	35.4	53.2
89	64.7	52.7	79.1
182	2.5	2.3	3.5







#58
Fluorene
Concen: 10.232 ng
RT: 10.469 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:166 Resp: 106355

Ion	Ratio	Lower	Upper
166	100		
165	100.5	77.0	115.4
167	13.0	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 9.612 ng
RT: 10.239 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:232 Resp: 28893

Ion	Ratio	Lower	Upper
232	100		
131	39.6	31.9	47.9
130	1.4	1.1	1.7
166	27.6	21.0	31.4







#60
Diethylphthalate
Concen: 10.256 ng
RT: 10.334 min Scan# 1384
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:149 Resp: 115029

Ion	Ratio	Lower	Upper
149	100		
177	22.9	17.8	26.6
150	12.8	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 10.385 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:204 Resp: 60704

Ion	Ratio	Lower	Upper
204	100		
206	33.2	27.8	41.6
141	54.3	45.4	68.2







#62
4-Nitroaniline
Concen: 9.752 ng
RT: 10.481 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:138 Resp: 24222

Ion	Ratio	Lower	Upper
138	100		
92	48.7	30.5	70.5
108	74.9	56.2	96.2







#63
Azobenzene
Concen: 10.095 ng
RT: 10.616 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 107163

Ion	Ratio	Lower	Upper
77	100		
182	27.6	9.0	49.0
105	22.0	1.4	41.4
51	35.7	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:188 Resp: 301308

Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.4	8.0
80	8.2	6.0	9.0







#65
4,6-Dinitro-2-methylphenol
Concen: 10.394 ng
RT: 10.510 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:198 Resp: 20093

Ion	Ratio	Lower	Upper
198	100		
51	51.6	25.8	65.8
105	37.3	17.2	57.2







#66
n-Nitrosodiphenylamine
Concen: 10.422 ng
RT: 10.575 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:169 Resp: 100334

Ion	Ratio	Lower	Upper
169	100		
168	57.1	47.4	71.2
167	33.9	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 10.242 ng
RT: 10.951 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:248 Resp: 38639

Ion	Ratio	Lower	Upper
248	100		
250	98.5	77.8	116.6
141	62.5	50.5	75.7







#68
Hexachlorobenzene
Concen: 10.990 ng
RT: 11.016 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:284 Resp: 50850

Ion	Ratio	Lower	Upper
284	100		
142	22.0	20.0	30.0
249	26.5	22.2	33.2







#69
Atrazine
Concen: 10.638 ng
RT: 11.098 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:200 Resp: 33605

Ion	Ratio	Lower	Upper
200	100		
173	24.7	4.9	44.9
215	47.7	28.2	68.2







#70
Pentachlorophenol
Concen: 8.386 ng
RT: 11.216 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:266 Resp: 18468

Ion	Ratio	Lower	Upper
266	100		
268	60.6	48.2	72.4
264	57.7	49.4	74.0







#71
Phenanthrene
Concen: 10.467 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:178 Resp: 156663

Ion	Ratio	Lower	Upper
178	100		
176	18.9	16.3	24.5
179	14.1	12.6	18.8







#72
Anthracene
Concen: 10.491 ng
RT: 11.486 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:178 Resp: 162203

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.4
179	14.5	11.8	17.6







#73
Carbazole
Concen: 10.534 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:167 Resp: 138634

Ion	Ratio	Lower	Upper
167	100		
166	21.0	16.1	24.1
139	12.4	9.8	14.6







#74
Di-n-butylphthalate
Concen: 10.183 ng
RT: 11.969 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:149 Resp: 164302

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	4.5	3.6	5.4







#75
Fluoranthene
Concen: 10.486 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:202 Resp: 173397

Ion	Ratio	Lower	Upper
202	100		
101	9.2	0.0	27.7
203	17.0	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:240 Resp: 243705

Ion	Ratio	Lower	Upper
240	100		
120	7.6	6.5	9.7
236	26.4	20.1	30.1







#77
Benzidine
Concen: 11.068 ng
RT: 12.751 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:184 Resp: 83945

Ion	Ratio	Lower	Upper
184	100		
185	14.1	11.6	17.4
183	10.7	8.5	12.7







#78
Pyrene
Concen: 10.196 ng
RT: 12.857 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:202 Resp: 178638

Ion	Ratio	Lower	Upper
202	100		
200	20.6	17.2	25.8
203	17.1	13.5	20.3







#79
Terphenyl-d14
Concen: 19.833 ng
RT: 12.992 min Scan# 18
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:244 Resp: 276477

Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.8	8.6
122	7.6	4.9	7.3#







#80
Butylbenzylphthalate
Concen: 9.964 ng
RT: 13.469 min Scan# 1917
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:149 Resp: 66048

Ion	Ratio	Lower	Upper
149	100		
91	69.3	56.6	84.8
206	19.9	18.3	27.5







#81
Benzo(a)anthracene
Concen: 9.778 ng
RT: 14.045 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 161042

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.4	32.2
229	20.0	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 9.963 ng
RT: 14.010 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:252 Resp: 57120

Ion	Ratio	Lower	Upper
252	100		
254	64.8	49.9	74.9
126	9.1	6.5	9.7







#83
Chrysene
Concen: 9.954 ng
RT: 14.080 min Scan# 20
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 153845

Ion	Ratio	Lower	Upper
228	100		
226	30.4	23.8	35.6
229	20.3	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 9.765 ng
RT: 14.033 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:149 Resp: 88365

Ion	Ratio	Lower	Upper
149	100		
167	28.5	22.5	33.7
279	4.9	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 9.616 ng
RT: 14.645 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:149 Resp: 149635

Ion	Ratio	Lower	Upper
149	100		
167	1.4	1.2	1.8
43	12.8	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:264 Resp: 290045

Ion	Ratio	Lower	Upper
264	100		
260	22.1	17.0	25.4
265	21.5	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 9.695 ng
RT: 17.039 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:276 Resp: 193064

Ion	Ratio	Lower	Upper
276	100		
138	17.8	13.0	19.6
277	24.2	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 9.803 ng
RT: 15.104 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:252 Resp: 160748

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.2	25.8
125	6.5	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 10.320 ng
RT: 15.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:252 Resp: 156128

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.2
125	6.0	4.9	7.3







#90
Benzo(a)pyrene
Concen: 9.757 ng
RT: 15.474 min Scan# 2258
Delta R.T. -0.006 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:252 Resp: 143530

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.1	25.7
125	7.6	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 9.766 ng
RT: 17.051 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 153941

Ion	Ratio	Lower	Upper
278	100		
139	12.0	8.5	12.7
279	22.8	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 9.614 ng
RT: 17.486 min Scan# 2600
Delta R.T. -0.012 min
Lab File: BF144058.D
Acq: 24 Oct 2025 11:39

Tgt Ion:276 Resp: 152368

Ion	Ratio	Lower	Upper
276	100		
277	21.5	17.9	26.9
138	14.3	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144059.D
Acq On : 24 Oct 2025 12:36
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Oct 24 17:02:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	112429	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	417749	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	217706	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	369584	20.000	ng	0.00
76) Chrysene-d12	14.063	240	292368	20.000	ng	0.00
86) Perylene-d12	15.545	264	364339	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	298616	42.021	ng	0.00
7) Phenol-d6	6.504	99	332945	42.423	ng	0.00
23) Nitrobenzene-d5	7.445	82	312746	41.016	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	111292	40.132	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	614559	39.724	ng	0.00
79) Terphenyl-d14	12.998	244	655393	39.189	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.634	88	54561	17.768	ng	96
3) Pyridine	3.398	79	172002	20.922	ng	97
4) n-Nitrosodimethylamine	3.351	42	108274	21.434	ng	98
6) Aniline	6.545	93	236820	21.092	ng	99
8) 2-Chlorophenol	6.663	128	146877	20.988	ng	100
9) Benzaldehyde	6.434	77	108472	17.802	ng	99
10) Phenol	6.522	94	204649	21.136	ng	99
11) bis(2-Chloroethyl)ether	6.616	93	144490	20.652	ng	99
12) 1,3-Dichlorobenzene	6.822	146	184376	21.358	ng	99
13) 1,4-Dichlorobenzene	6.898	146	178962	21.035	ng	99
14) 1,2-Dichlorobenzene	7.051	146	172463	20.951	ng	100
15) Benzyl Alcohol	7.022	79	138784	21.748	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.151	45	281645	21.419	ng	99
17) 2-Methylphenol	7.128	107	115994	21.550	ng	94
18) Hexachloroethane	7.392	117	63937	21.020	ng	95
19) n-Nitroso-di-n-propyla...	7.292	70	106671	21.019	ng	97
20) 3+4-Methylphenols	7.281	107	133833	21.033	ng	95
22) Acetophenone	7.286	105	181010	20.050	ng	97
24) Nitrobenzene	7.463	77	168641	20.907	ng	99
25) Isophorone	7.698	82	269356	20.206	ng	98
26) 2-Nitrophenol	7.781	139	81754	21.306	ng	97
27) 2,4-Dimethylphenol	7.810	122	109494	19.202	ng	95
28) bis(2-Chloroethoxy)met...	7.910	93	172327	20.729	ng	99
29) 2,4-Dichlorophenol	8.016	162	135856	20.487	ng	99
30) 1,2,4-Trichlorobenzene	8.104	180	139458	20.442	ng	99
31) Naphthalene	8.186	128	401788	20.378	ng	99
32) Benzoic acid	7.928	122	83503	23.086	ng	87
33) 4-Chloroaniline	8.233	127	145706	21.228	ng	99
34) Hexachlorobutadiene	8.298	225	114614	20.546	ng	98
35) Caprolactam	8.598	113	34241	20.591	ng	93
36) 4-Chloro-3-methylphenol	8.710	107	119127	20.357	ng	95
37) 2-Methylnaphthalene	8.875	142	263628	20.280	ng	98
38) 1-Methylnaphthalene	8.975	142	246748	20.035	ng	99
40) 1,2,4,5-Tetrachloroben...	9.039	216	150836	20.160	ng	99
41) Hexachlorocyclopentadiene	9.028	237	45275	19.591	ng	97
43) 2,4,6-Trichlorophenol	9.151	196	104484	21.484	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144059.D
Acq On : 24 Oct 2025 12:36
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Oct 24 17:02:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	105424	20.655	ng		98
46) 1,1'-Biphenyl	9.339	154	313680	20.128	ng		99
47) 2-Chloronaphthalene	9.363	162	268452	20.561	ng		96
48) 2-Nitroaniline	9.463	65	80171	20.678	ng		97
49) Acenaphthylene	9.780	152	362218	20.575	ng		99
50) Dimethylphthalate	9.633	163	290025	20.324	ng		99
51) 2,6-Dinitrotoluene	9.698	165	59905	21.235	ng		100
52) Acenaphthene	9.951	154	246935	21.003	ng		99
53) 3-Nitroaniline	9.875	138	63852	20.358	ng		97
54) 2,4-Dinitrophenol	9.980	184	35229	23.089	ng		94
55) Dibenzofuran	10.127	168	329079	20.162	ng		99
56) 4-Nitrophenol	10.027	139	43333	20.402	ng		90
57) 2,4-Dinitrotoluene	10.104	165	80566	21.160	ng		96
58) Fluorene	10.469	166	246753	19.841	ng		96
59) 2,3,4,6-Tetrachlorophenol	10.245	232	72554	20.172	ng	#	96
60) Diethylphthalate	10.339	149	270864	20.184	ng		99
61) 4-Chlorophenyl-phenyle...	10.463	204	142719	20.406	ng		99
62) 4-Nitroaniline	10.486	138	59903	20.157	ng		95
63) Azobenzene	10.622	77	258435	20.346	ng		98
65) 4,6-Dinitro-2-methylph...	10.516	198	55850	23.553	ng		98
66) n-Nitrosodiphenylamine	10.580	169	242811	20.563	ng		98
67) 4-Bromophenyl-phenylether	10.951	248	91405	19.753	ng		98
68) Hexachlorobenzene	11.016	284	114304	20.141	ng		96
69) Atrazine	11.104	200	80666	20.818	ng		98
70) Pentachlorophenol	11.216	266	55240	20.449	ng		95
71) Phenanthrene	11.433	178	372967	20.315	ng		97
72) Anthracene	11.486	178	380455	20.060	ng		99
73) Carbazole	11.639	167	326491	20.225	ng		98
74) Di-n-butylphthalate	11.969	149	412471	20.841	ng		100
75) Fluoranthene	12.627	202	414382	20.430	ng		99
77) Benzidine	12.751	184	179581	19.736	ng		99
78) Pyrene	12.857	202	416839	19.831	ng		99
80) Butylbenzylphthalate	13.474	149	165613	20.827	ng		98
81) Benzo(a)anthracene	14.045	228	397414	20.114	ng		98
82) 3,3'-Dichlorobenzidine	14.010	252	141681	20.599	ng		98
83) Chrysene	14.086	228	366241	19.753	ng		99
84) Bis(2-ethylhexyl)phtha...	14.033	149	222446	20.491	ng		98
85) Di-n-octyl phthalate	14.645	149	378673	20.284	ng		99
87) Indeno(1,2,3-cd)pyrene	17.045	276	507947	20.307	ng		100
88) Benzo(b)fluoranthene	15.104	252	420103	20.396	ng		99
89) Benzo(k)fluoranthene	15.133	252	368139	19.372	ng		100
90) Benzo(a)pyrene	15.480	252	377387	20.423	ng		99
91) Dibenzo(a,h)anthracene	17.062	278	403214	20.364	ng		98
92) Benzo(g,h,i)perylene	17.503	276	409358	20.563	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144059.D
Acq On : 24 Oct 2025 12:36
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Oct 24 17:02:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 79
Delta R.T. -0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 112429

Ion	Ratio	Lower	Upper
152	100		
150	163.4	124.4	186.6
115	66.6	49.4	74.0







#2
1,4-Dioxane
Concen: 17.768 ng
RT: 2.634 min Scan# 75
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion: 88 Resp: 54561

Ion	Ratio	Lower	Upper
88	100		
58	98.8	76.5	114.7
43	47.5	35.3	52.9







#3
Pyridine
Concen: 20.922 ng
RT: 3.398 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 172002

Ion	Ratio	Lower	Upper
79	100		
52	80.9	67.5	101.3
51	43.3	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 21.434 ng
RT: 3.351 min Scan# 197
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion: 42 Resp: 108274

Ion	Ratio	Lower	Upper
42	100		
74	93.2	76.4	114.6
44	5.8	5.2	7.8







#5
2-Fluorophenol
Concen: 42.021 ng
RT: 5.492 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:112 Resp: 298616

Ion	Ratio	Lower	Upper
112	100		
64	61.6	49.8	74.8
63	34.0	27.4	41.2







#6
Aniline
Concen: 21.092 ng
RT: 6.545 min Scan# 740
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion: 93 Resp: 236820

Ion	Ratio	Lower	Upper
93	100		
66	44.1	35.7	53.5
65	20.7	17.0	25.6







#7
Phenol-d6
Concen: 42.423 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 99 Resp: 332945

Ion	Ratio	Lower	Upper
99	100		
42	30.7	23.9	35.9
71	44.2	34.4	51.6







#8
2-Chlorophenol
Concen: 20.988 ng
RT: 6.663 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:128 Resp: 146877

Ion	Ratio	Lower	Upper
128	100		
130	31.3	10.7	50.7
64	54.6	34.7	74.7







#9
Benzaldehyde
Concen: 17.802 ng
RT: 6.434 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 108472

Ion	Ratio	Lower	Upper
77	100		
105	91.1	72.2	112.2
106	86.4	67.9	107.9







#10
Phenol
Concen: 21.136 ng
RT: 6.522 min Scan# 736
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion: 94 Resp: 204649

Ion	Ratio	Lower	Upper
94	100		
65	32.4	13.1	53.1
66	49.7	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 20.652 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 93 Resp: 144490

Ion	Ratio	Lower	Upper
93	100		
63	81.4	60.2	100.2
95	33.1	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 21.358 ng
RT: 6.822 min Scan# 787
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:146 Resp: 184376

Ion	Ratio	Lower	Upper
146	100		
148	62.0	50.6	75.8
75	33.4	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 21.035 ng
RT: 6.898 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:146 Resp: 178962

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.6
111	42.9	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 20.951 ng
RT: 7.051 min Scan# 826
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:146 Resp: 172463

Ion	Ratio	Lower	Upper
146	100		
148	63.3	50.8	76.2
111	43.5	34.7	52.1







#15
Benzyl Alcohol
Concen: 21.748 ng
RT: 7.022 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 138784

Ion	Ratio	Lower	Upper
79	100		
108	66.5	54.8	82.2
77	62.4	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 21.419 ng
RT: 7.151 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion: 45 Resp: 281645

Ion	Ratio	Lower	Upper
45	100		
77	11.4	0.0	30.9
79	8.5	0.0	28.2







#17
2-Methylphenol
Concen: 21.550 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:107 Resp: 115994

Ion	Ratio	Lower	Upper
107	100		
108	113.6	96.4	144.6
77	53.7	46.8	70.2
79	54.7	45.8	68.6







#18
Hexachloroethane
Concen: 21.020 ng
RT: 7.392 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:117 Resp: 63937

Ion	Ratio	Lower	Upper
117	100		
119	94.0	76.6	115.0
201	117.0	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 21.019 ng
RT: 7.292 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 70 Resp: 106671

Ion	Ratio	Lower	Upper
70	100		
42	81.2	67.4	101.0
101	8.3	7.1	10.7
130	15.5	13.5	20.3







#20
3+4-Methylphenols
Concen: 21.033 ng
RT: 7.281 min Scan# 865
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:107 Resp: 133833

Ion	Ratio	Lower	Upper
107	100		
108	91.4	68.0	108.0
77	121.6	108.7	148.7
79	35.9	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:136 Resp: 417749

Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.6	13.0
54	9.2	6.6	9.8
68	5.1	4.6	6.8







#22
Acetophenone
Concen: 20.050 ng
RT: 7.286 min Scan# 866
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:105 Resp: 181010

Ion	Ratio	Lower	Upper
105	100		
71	8.4	6.9	10.3
51	53.6	40.6	60.8
120	20.9	16.8	25.2







#23
Nitrobenzene-d5
Concen: 41.016 ng
RT: 7.445 min Scan# 89
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 312746

Ion	Ratio	Lower	Upper
82	100		
128	40.2	32.4	48.6
54	67.2	54.0	81.0







#24
Nitrobenzene
Concen: 20.907 ng
RT: 7.463 min Scan# 896
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion: 77 Resp: 168641

Ion	Ratio	Lower	Upper
77	100		
123	44.1	35.0	52.6
65	14.4	12.2	18.2







#25
Isophorone
Concen: 20.206 ng
RT: 7.698 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 269356

Ion	Ratio	Lower	Upper
82	100		
95	7.7	6.6	9.8
138	18.0	13.8	20.8







#26
2-Nitrophenol
Concen: 21.306 ng
RT: 7.781 min Scan# 950
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:139 Resp: 81754

Ion	Ratio	Lower	Upper
139	100		
109	30.3	24.6	37.0
65	55.3	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 19.202 ng
RT: 7.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:122 Resp: 109494

Ion	Ratio	Lower	Upper
122	100		
107	120.9	91.8	137.8
121	58.9	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 20.729 ng
RT: 7.910 min Scan# 972
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion: 93 Resp: 172327

Ion	Ratio	Lower	Upper
93	100		
95	31.3	24.6	36.8
123	10.8	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 20.487 ng
RT: 8.016 min Scan# 99
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 135856

Ion	Ratio	Lower	Upper
162	100		
164	62.7	42.1	82.1
98	36.7	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 20.442 ng
RT: 8.104 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:180 Resp: 139458

Ion	Ratio	Lower	Upper
180	100		
182	108.4	87.1	130.7
145	32.7	25.3	37.9







#31
Naphthalene
Concen: 20.378 ng
RT: 8.186 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:128 Resp: 401788

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.4	12.6
127	12.7	9.9	14.9







#32
Benzoic acid
Concen: 23.086 ng
RT: 7.928 min Scan# 975
Delta R.T. 0.012 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:122 Resp: 83503

Ion	Ratio	Lower	Upper
122	100		
105	120.9	111.3	151.3
77	97.1	96.7	136.7







#33
4-Chloroaniline
Concen: 21.228 ng
RT: 8.233 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:127 Resp: 145706

Ion	Ratio	Lower	Upper
127	100		
129	31.1	25.6	38.4
65	38.1	30.9	46.3
92	22.8	18.9	28.3







#34
Hexachlorobutadiene
Concen: 20.546 ng
RT: 8.298 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:225 Resp: 114614

Ion	Ratio	Lower	Upper
225	100		
223	63.1	50.6	76.0
227	62.7	51.8	77.8







#35
Caprolactam
Concen: 20.591 ng
RT: 8.598 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:113 Resp: 34241

Ion	Ratio	Lower	Upper
113	100		
55	187.8	160.1	200.1
56	157.4	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 20.357 ng
RT: 8.710 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:107 Resp: 119127

Ion	Ratio	Lower	Upper
107	100		
144	28.0	21.6	32.4
142	76.0	65.4	98.0







#37
2-Methylnaphthalene
Concen: 20.280 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:142 Resp: 263628

Ion	Ratio	Lower	Upper
142	100		
141	86.1	67.6	101.4







#38
1-Methylnaphthalene
Concen: 20.035 ng
RT: 8.975 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:142 Resp: 246748

Ion	Ratio	Lower	Upper
142	100		
141	92.8	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:164 Resp: 217706

Ion	Ratio	Lower	Upper
164	100		
162	100.9	79.4	119.2
160	44.6	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.160 ng
RT: 9.039 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:216 Resp: 150836

Ion	Ratio	Lower	Upper
216	100		
214	79.3	63.2	94.8
179	20.2	15.2	22.8
108	15.4	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 19.591 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:237 Resp: 45275

Ion	Ratio	Lower	Upper
237	100		
235	62.0	44.3	84.3
272	14.5	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 40.132 ng
RT: 10.710 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:330 Resp: 111292

Ion	Ratio	Lower	Upper
330	100		
332	96.7	77.0	115.6
141	24.9	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 21.484 ng
RT: 9.151 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:196 Resp: 104484

Ion	Ratio	Lower	Upper
196	100		
198	91.0	74.4	111.6
200	28.8	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 20.655 ng
RT: 9.192 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:196 Resp: 105424

Ion	Ratio	Lower	Upper
196	100		
198	95.9	78.0	117.0
97	44.7	37.7	56.5
132	25.2	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 39.724 ng
RT: 9.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:172 Resp: 614559

Ion	Ratio	Lower	Upper
172	100		
171	34.5	27.5	41.3
170	24.0	19.0	28.6







#46
1,1'-Biphenyl
Concen: 20.128 ng
RT: 9.339 min Scan# 1215
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:154 Resp: 313680

Ion	Ratio	Lower	Upper
154	100		
153	39.2	19.7	59.7
76	11.1	0.0	31.1







#47
2-Chloronaphthalene
Concen: 20.561 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 268452

Ion	Ratio	Lower	Upper
162	100		
127	37.1	31.6	47.4
164	30.7	26.5	39.7







#48
2-Nitroaniline
Concen: 20.678 ng
RT: 9.463 min Scan# 1236
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion: 65 Resp: 80171

Ion	Ratio	Lower	Upper
65	100		
92	66.6	52.3	78.5
138	102.3	78.3	117.5







#49
Acenaphthylene
Concen: 20.575 ng
RT: 9.780 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 362218

Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.9	23.9
153	13.2	10.4	15.6







#50
Dimethylphthalate
Concen: 20.324 ng
RT: 9.633 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:163 Resp: 290025

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.0
164	10.1	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 21.235 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 59905

Ion	Ratio	Lower	Upper
165	100		
63	63.0	50.4	75.6
89	51.7	41.4	62.2







#52
Acenaphthene
Concen: 21.003 ng
RT: 9.951 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:154 Resp: 246935

Ion	Ratio	Lower	Upper
154	100		
153	112.0	89.9	134.9
152	53.7	42.0	63.0







#53
3-Nitroaniline
Concen: 20.358 ng
RT: 9.875 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:138 Resp: 63852

Ion	Ratio	Lower	Upper
138	100		
108	11.0	7.9	11.9
92	118.5	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 23.089 ng
RT: 9.980 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:184 Resp: 35229

Ion	Ratio	Lower	Upper
184	100		
63	64.6	52.1	78.1
154	68.6	47.9	71.9







#55
Dibenzofuran
Concen: 20.162 ng
RT: 10.127 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:168 Resp: 329079

Ion	Ratio	Lower	Upper
168	100		
139	40.4	33.2	49.8
169	15.7	12.6	19.0







#56
4-Nitrophenol
Concen: 20.402 ng
RT: 10.027 min Scan# 1332
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:139 Resp: 43333

Ion	Ratio	Lower	Upper
139	100		
109	68.6	53.8	93.8
65	103.2	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 21.160 ng
RT: 10.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 80566

Ion	Ratio	Lower	Upper
165	100		
63	48.1	35.4	53.2
89	63.1	52.7	79.1
182	2.6	2.3	3.5







#58
Fluorene
Concen: 19.841 ng
RT: 10.469 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:166 Resp: 246753

Ion	Ratio	Lower	Upper
166	100		
165	99.8	77.0	115.4
167	13.5	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 20.172 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:232 Resp: 72554

Ion	Ratio	Lower	Upper
232	100		
131	43.5	31.9	47.9
130	1.7	1.1	1.7#
166	26.9	21.0	31.4







#60
Diethylphthalate
Concen: 20.184 ng
RT: 10.339 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:149 Resp: 270864

Ion	Ratio	Lower	Upper
149	100		
177	22.3	17.8	26.6
150	11.4	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 20.406 ng
RT: 10.463 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:204 Resp: 142719

Ion	Ratio	Lower	Upper
204	100		
206	33.9	27.8	41.6
141	57.2	45.4	68.2







#62
4-Nitroaniline
Concen: 20.157 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:138 Resp: 59903

Ion	Ratio	Lower	Upper
138	100		
92	55.0	30.5	70.5
108	79.8	56.2	96.2







#63
Azobenzene
Concen: 20.346 ng
RT: 10.622 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 258435

Ion	Ratio	Lower	Upper
77	100		
182	28.8	9.0	49.0
105	20.3	1.4	41.4
51	38.3	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:188 Resp: 369584

Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.4	8.0
80	7.8	6.0	9.0







#65
4,6-Dinitro-2-methylphenol
Concen: 23.553 ng
RT: 10.516 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:198 Resp: 55850

Ion	Ratio	Lower	Upper
198	100		
51	46.8	25.8	65.8
105	35.7	17.2	57.2







#66
n-Nitrosodiphenylamine
Concen: 20.563 ng
RT: 10.580 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:169 Resp: 242811

Ion	Ratio	Lower	Upper
169	100		
168	56.9	47.4	71.2
167	32.0	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 19.753 ng
RT: 10.951 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:248 Resp: 91405

Ion	Ratio	Lower	Upper
248	100		
250	98.0	77.8	116.6
141	66.6	50.5	75.7







#68
Hexachlorobenzene
Concen: 20.141 ng
RT: 11.016 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:284 Resp: 114304

Ion	Ratio	Lower	Upper
284	100		
142	26.3	20.0	30.0
249	30.2	22.2	33.2







#69
Atrazine
Concen: 20.818 ng
RT: 11.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:200 Resp: 80666

Ion	Ratio	Lower	Upper
200	100		
173	26.1	4.9	44.9
215	49.1	28.2	68.2







#70
Pentachlorophenol
Concen: 20.449 ng
RT: 11.216 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:266 Resp: 55240

Ion	Ratio	Lower	Upper
266	100		
268	64.3	48.2	72.4
264	64.7	49.4	74.0







#71
Phenanthrene
Concen: 20.315 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:178 Resp: 372967

Ion	Ratio	Lower	Upper
178	100		
176	19.2	16.3	24.5
179	14.6	12.6	18.8







#72
Anthracene
Concen: 20.060 ng
RT: 11.486 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:178 Resp: 380455

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.4
179	15.4	11.8	17.6







#73
Carbazole
Concen: 20.225 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:167 Resp: 326491

Ion	Ratio	Lower	Upper
167	100		
166	21.0	16.1	24.1
139	12.9	9.8	14.6







#74
Di-n-butylphthalate
Concen: 20.841 ng
RT: 11.969 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:149 Resp: 412471

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	4.6	3.6	5.4







#75
Fluoranthene
Concen: 20.430 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:202 Resp: 414382

Ion	Ratio	Lower	Upper
202	100		
101	7.5	0.0	27.7
203	17.0	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.063 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:240 Resp: 292368

Ion	Ratio	Lower	Upper
240	100		
120	7.3	6.5	9.7
236	25.6	20.1	30.1







#77
Benzidine
Concen: 19.736 ng
RT: 12.751 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:184 Resp: 179581

Ion	Ratio	Lower	Upper
184	100		
185	13.8	11.6	17.4
183	10.9	8.5	12.7







#78
Pyrene
Concen: 19.831 ng
RT: 12.857 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:202 Resp: 416839

Ion	Ratio	Lower	Upper
202	100		
200	20.9	17.2	25.8
203	17.1	13.5	20.3







#79
Terphenyl-d14
Concen: 39.189 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:244 Resp: 655393

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.8	8.6
122	6.8	4.9	7.3







#80
Butylbenzylphthalate
Concen: 20.827 ng
RT: 13.474 min Scan# 1918
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:149 Resp: 165613

Ion	Ratio	Lower	Upper
149	100		
91	69.0	56.6	84.8
206	23.7	18.3	27.5







#81
Benzo(a)anthracene
Concen: 20.114 ng
RT: 14.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 397414

Ion	Ratio	Lower	Upper
228	100		
226	28.0	21.4	32.2
229	19.9	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 20.599 ng
RT: 14.010 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:252 Resp: 141681

Ion	Ratio	Lower	Upper
252	100		
254	64.0	49.9	74.9
126	7.3	6.5	9.7







#83
Chrysene
Concen: 19.753 ng
RT: 14.086 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 366241

Ion	Ratio	Lower	Upper
228	100		
226	29.2	23.8	35.6
229	19.3	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 20.491 ng
RT: 14.033 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:149 Resp: 222446

Ion	Ratio	Lower	Upper
149	100		
167	27.0	22.5	33.7
279	4.6	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 20.284 ng
RT: 14.645 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:149 Resp: 378673

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	12.2	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.545 min Scan# 2270
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:264 Resp: 364339

Ion	Ratio	Lower	Upper
264	100		
260	22.0	17.0	25.4
265	20.4	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 20.307 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:276 Resp: 507947

Ion	Ratio	Lower	Upper
276	100		
138	16.3	13.0	19.6
277	24.5	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 20.396 ng
RT: 15.104 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:252 Resp: 420103

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.2	25.8
125	6.4	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 19.372 ng
RT: 15.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:252 Resp: 368139

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.2
125	6.4	4.9	7.3







#90
Benzo(a)pyrene
Concen: 20.423 ng
RT: 15.480 min Scan# 2259
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:252 Resp: 377387

Ion	Ratio	Lower	Upper
252	100		
253	20.7	17.1	25.7
125	6.6	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 20.364 ng
RT: 17.062 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:278 Resp: 403214

Ion	Ratio	Lower	Upper
278	100		
139	10.5	8.5	12.7
279	24.4	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 20.563 ng
RT: 17.503 min Scan# 2603
Delta R.T. 0.006 min
Lab File: BF144059.D
Acq: 24 Oct 2025 12:36

Tgt Ion:276 Resp: 409358

Ion	Ratio	Lower	Upper
276	100		
277	22.8	17.9	26.9
138	14.6	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144060.D
Acq On : 24 Oct 2025 13:05
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Quant Time: Oct 24 17:03:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	49457	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	178191	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	91554	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	159613	20.000	ng	0.00
76) Chrysene-d12	14.057	240	123616	20.000	ng	0.00
86) Perylene-d12	15.539	264	152855	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	257786	82.465	ng	0.00
7) Phenol-d6	6.504	99	278572	80.690	ng	0.00
23) Nitrobenzene-d5	7.439	82	262114	80.590	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	95908	82.238	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	531491	81.692	ng	0.00
79) Terphenyl-d14	12.998	244	599175	84.737	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.628	88	54397	40.269	ng	100
3) Pyridine	3.398	79	151037	41.765	ng	100
4) n-Nitrosodimethylamine	3.346	42	90217	40.599	ng	100
6) Aniline	6.539	93	199395	40.371	ng	100
8) 2-Chlorophenol	6.663	128	123726	40.190	ng	100
9) Benzaldehyde	6.434	77	103170	38.490	ng	100
10) Phenol	6.522	94	171675	40.307	ng	100
11) bis(2-Chloroethyl)ether	6.616	93	123456	40.113	ng	100
12) 1,3-Dichlorobenzene	6.822	146	154816	40.768	ng	100
13) 1,4-Dichlorobenzene	6.898	146	153441	41.000	ng	100
14) 1,2-Dichlorobenzene	7.051	146	146404	40.430	ng	100
15) Benzyl Alcohol	7.016	79	113241	40.340	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.151	45	235025	40.632	ng	100
17) 2-Methylphenol	7.128	107	92050	38.877	ng	100
18) Hexachloroethane	7.392	117	53836	40.236	ng	100
19) n-Nitroso-di-n-propyla...	7.286	70	86672	38.823	ng	100
20) 3+4-Methylphenols	7.281	107	115474	41.255	ng	100
22) Acetophenone	7.286	105	155572	40.399	ng	100
24) Nitrobenzene	7.463	77	139675	40.596	ng	100
25) Isophorone	7.698	82	229089	40.288	ng	100
26) 2-Nitrophenol	7.775	139	68074	41.591	ng	100
27) 2,4-Dimethylphenol	7.810	122	102275	42.049	ng	100
28) bis(2-Chloroethoxy)met...	7.910	93	144080	40.631	ng	100
29) 2,4-Dichlorophenol	8.016	162	115057	40.677	ng	100
30) 1,2,4-Trichlorobenzene	8.104	180	119224	40.971	ng	100
31) Naphthalene	8.186	128	344062	40.910	ng	100
32) Benzoic acid	7.916	122	51883	33.628	ng	100
33) 4-Chloroaniline	8.233	127	119889	40.949	ng	100
34) Hexachlorobutadiene	8.298	225	96082	40.380	ng	100
35) Caprolactam	8.592	113	30206	42.585	ng	100
36) 4-Chloro-3-methylphenol	8.710	107	100429	40.234	ng	100
37) 2-Methylnaphthalene	8.875	142	229345	41.362	ng	100
38) 1-Methylnaphthalene	8.975	142	210460	40.062	ng	100
40) 1,2,4,5-Tetrachloroben...	9.039	216	129417	41.130	ng	100
41) Hexachlorocyclopentadiene	9.028	237	39857	41.012	ng	100
43) 2,4,6-Trichlorophenol	9.151	196	86607	42.346	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144060.D
Acq On : 24 Oct 2025 13:05
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Quant Time: Oct 24 17:03:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	85520	39.842	ng	100
46) 1,1'-Biphenyl	9.339	154	269550	41.129	ng	100
47) 2-Chloronaphthalene	9.363	162	223762	40.753	ng	100
48) 2-Nitroaniline	9.457	65	69638	42.711	ng	100
49) Acenaphthylene	9.780	152	311770	42.112	ng	100
50) Dimethylphthalate	9.633	163	250045	41.666	ng	100
51) 2,6-Dinitrotoluene	9.698	165	50463	42.535	ng	100
52) Acenaphthene	9.951	154	203905	41.240	ng	100
53) 3-Nitroaniline	9.869	138	58978	44.713	ng	100
54) 2,4-Dinitrophenol	9.980	184	22640	35.283	ng	100
55) Dibenzofuran	10.122	168	285490	41.592	ng	100
56) 4-Nitrophenol	10.027	139	34191	38.279	ng	100
57) 2,4-Dinitrotoluene	10.104	165	70572	44.075	ng	100
58) Fluorene	10.469	166	217594	41.604	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.245	232	65926	43.585	ng	100
60) Diethylphthalate	10.339	149	236119	41.838	ng	100
61) 4-Chlorophenyl-phenyle...	10.457	204	123596	42.022	ng	100
62) 4-Nitroaniline	10.480	138	54231	43.392	ng	100
63) Azobenzene	10.622	77	226383	42.381	ng	100
65) 4,6-Dinitro-2-methylph...	10.516	198	39142	38.223	ng	100
66) n-Nitrosodiphenylamine	10.574	169	205138	40.226	ng	100
67) 4-Bromophenyl-phenylether	10.951	248	83105	41.585	ng	100
68) Hexachlorobenzene	11.016	284	102619	41.869	ng	100
69) Atrazine	11.104	200	66524	39.753	ng	100
70) Pentachlorophenol	11.216	266	47778	40.953	ng	100
71) Phenanthrene	11.433	178	324774	40.961	ng	100
72) Anthracene	11.486	178	336252	41.053	ng	100
73) Carbazole	11.639	167	289029	41.458	ng	100
74) Di-n-butylphthalate	11.969	149	357025	41.770	ng	100
75) Fluoranthene	12.627	202	358431	40.917	ng	100
77) Benzidine	12.751	184	164527	42.766	ng	100
78) Pyrene	12.857	202	375923	42.300	ng	100
80) Butylbenzylphthalate	13.474	149	148621	44.204	ng	100
81) Benzo(a)anthracene	14.045	228	374791	44.863	ng	100
82) 3,3'-Dichlorobenzidine	14.010	252	126028	43.337	ng	100
83) Chrysene	14.086	228	326041	41.590	ng	100
84) Bis(2-ethylhexyl)phtha...	14.033	149	196471	42.805	ng	100
85) Di-n-octyl phthalate	14.645	149	345760	43.804	ng	100
87) Indeno(1,2,3-cd)pyrene	17.045	276	451467	43.020	ng	100
88) Benzo(b)fluoranthene	15.104	252	386268	44.700	ng	100
89) Benzo(k)fluoranthene	15.133	252	323867	40.621	ng	100
90) Benzo(a)pyrene	15.480	252	332791	42.927	ng	100
91) Dibenzo(a,h)anthracene	17.062	278	356487	42.914	ng	100
92) Benzo(g,h,i)perylene	17.498	276	365458	43.758	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144060.D
Acq On : 24 Oct 2025 13:05
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Quant Time: Oct 24 17:03:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 79
Delta R.T. -0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 49457

Ion	Ratio	Lower	Upper
152	100		
150	155.5	124.4	186.6
115	61.7	49.4	74.0







#2
1,4-Dioxane
Concen: 40.269 ng
RT: 2.628 min Scan# 74
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion: 88 Resp: 54397

Ion	Ratio	Lower	Upper
88	100		
58	95.6	76.5	114.7
43	44.1	35.3	52.9







#3
Pyridine
Concen: 41.765 ng
RT: 3.398 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 151037

Ion	Ratio	Lower	Upper
79	100		
52	84.4	67.5	101.3
51	42.0	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 40.599 ng
RT: 3.346 min Scan# 196
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion: 42 Resp: 90217

Ion	Ratio	Lower	Upper
42	100		
74	95.5	76.4	114.6
44	6.5	5.2	7.8







#5
2-Fluorophenol
Concen: 82.465 ng
RT: 5.492 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:112 Resp: 257786

Ion	Ratio	Lower	Upper
112	100		
64	62.3	49.8	74.8
63	34.3	27.4	41.2







#6
Aniline
Concen: 40.371 ng
RT: 6.539 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion: 93 Resp: 199395

Ion	Ratio	Lower	Upper
93	100		
66	44.6	35.7	53.5
65	21.3	17.0	25.6







#7
Phenol-d6
Concen: 80.690 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 99 Resp: 278572

Ion	Ratio	Lower	Upper
99	100		
42	29.9	23.9	35.9
71	43.0	34.4	51.6







#8
2-Chlorophenol
Concen: 40.190 ng
RT: 6.663 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:128 Resp: 123726

Ion	Ratio	Lower	Upper
128	100		
130	30.7	10.7	50.7
64	54.7	34.7	74.7







#9
Benzaldehyde
Concen: 38.490 ng
RT: 6.434 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 103170

Ion	Ratio	Lower	Upper
77	100		
105	92.2	72.2	112.2
106	87.9	67.9	107.9







#10
Phenol
Concen: 40.307 ng
RT: 6.522 min Scan# 736
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion: 94 Resp: 171675

Ion	Ratio	Lower	Upper
94	100		
65	33.1	13.1	53.1
66	50.6	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 40.113 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 93 Resp: 123456

Ion	Ratio	Lower	Upper
93	100		
63	80.2	60.2	100.2
95	32.4	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 40.768 ng
RT: 6.822 min Scan# 787
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:146 Resp: 154816

Ion	Ratio	Lower	Upper
146	100		
148	63.2	50.6	75.8
75	33.3	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 41.000 ng
RT: 6.898 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:146 Resp: 153441

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.0	76.6
111	43.8	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 40.430 ng
RT: 7.051 min Scan# 826
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:146 Resp: 146404

Ion	Ratio	Lower	Upper
146	100		
148	63.5	50.8	76.2
111	43.4	34.7	52.1







#15
Benzyl Alcohol
Concen: 40.340 ng
RT: 7.016 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 113241

Ion	Ratio	Lower	Upper
79	100		
108	68.5	54.8	82.2
77	61.7	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.632 ng
RT: 7.151 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion: 45 Resp: 235025

Ion	Ratio	Lower	Upper
45	100		
77	10.9	0.0	30.9
79	8.2	0.0	28.2







#17
2-Methylphenol
Concen: 38.877 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:107 Resp: 92050

Ion	Ratio	Lower	Upper
107	100		
108	120.5	96.4	144.6
77	58.5	46.8	70.2
79	57.2	45.8	68.6







#18
Hexachloroethane
Concen: 40.236 ng
RT: 7.392 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:117 Resp: 53836

Ion	Ratio	Lower	Upper
117	100		
119	95.8	76.6	115.0
201	126.1	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 38.823 ng
RT: 7.286 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 70 Resp: 86672

Ion	Ratio	Lower	Upper
70	100		
42	84.2	67.4	101.0
101	8.9	7.1	10.7
130	16.9	13.5	20.3







#20
3+4-Methylphenols
Concen: 41.255 ng
RT: 7.281 min Scan# 865
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:107 Resp: 115474

Ion	Ratio	Lower	Upper
107	100		
108	88.0	68.0	108.0
77	128.7	108.7	148.7
79	36.7	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:136 Resp: 178191

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	13.0
54	8.2	6.6	9.8
68	5.7	4.6	6.8







#22
Acetophenone
Concen: 40.399 ng
RT: 7.286 min Scan# 866
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:105 Resp: 155572

Ion	Ratio	Lower	Upper
105	100		
71	8.6	6.9	10.3
51	50.7	40.6	60.8
120	21.0	16.8	25.2







#23
Nitrobenzene-d5
Concen: 80.590 ng
RT: 7.439 min Scan# 89
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 262114

Ion	Ratio	Lower	Upper
82	100		
128	40.5	32.4	48.6
54	67.5	54.0	81.0







#24
Nitrobenzene
Concen: 40.596 ng
RT: 7.463 min Scan# 896
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion: 77 Resp: 139675

Ion	Ratio	Lower	Upper
77	100		
123	43.8	35.0	52.6
65	15.2	12.2	18.2







#25
Isophorone
Concen: 40.288 ng
RT: 7.698 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 229089

Ion	Ratio	Lower	Upper
82	100		
95	8.2	6.6	9.8
138	17.3	13.8	20.8







#26
2-Nitrophenol
Concen: 41.591 ng
RT: 7.775 min Scan# 949
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:139 Resp: 68074

Ion	Ratio	Lower	Upper
139	100		
109	30.8	24.6	37.0
65	58.2	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 42.049 ng
RT: 7.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:122 Resp: 102275

Ion	Ratio	Lower	Upper
122	100		
107	114.8	91.8	137.8
121	56.5	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 40.631 ng
RT: 7.910 min Scan# 972
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion: 93 Resp: 144080

Ion	Ratio	Lower	Upper
93	100		
95	30.7	24.6	36.8
123	11.4	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 40.677 ng
RT: 8.016 min Scan# 99
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 115057

Ion	Ratio	Lower	Upper
162	100		
164	62.1	42.1	82.1
98	36.0	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 40.971 ng
RT: 8.104 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:180 Resp: 119224

Ion	Ratio	Lower	Upper
180	100		
182	108.9	87.1	130.7
145	31.6	25.3	37.9







#31
Naphthalene
Concen: 40.910 ng
RT: 8.186 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:128 Resp: 344062

Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.4	12.6
127	12.4	9.9	14.9







#32
Benzoic acid
Concen: 33.628 ng
RT: 7.916 min Scan# 973
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:122 Resp: 51883

Ion	Ratio	Lower	Upper
122	100		
105	131.3	111.3	151.3
77	116.7	96.7	136.7







#33
4-Chloroaniline
Concen: 40.949 ng
RT: 8.233 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:127 Resp: 119889

Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.6	38.4
65	38.6	30.9	46.3
92	23.6	18.9	28.3







#34
Hexachlorobutadiene
Concen: 40.380 ng
RT: 8.298 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:225 Resp: 96082

Ion	Ratio	Lower	Upper
225	100		
223	63.3	50.6	76.0
227	64.8	51.8	77.8







#35
Caprolactam
Concen: 42.585 ng
RT: 8.592 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:113 Resp: 30206

Ion	Ratio	Lower	Upper
113	100		
55	180.1	160.1	200.1
56	144.9	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 40.234 ng
RT: 8.710 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:107 Resp: 100429

Ion	Ratio	Lower	Upper
107	100		
144	27.0	21.6	32.4
142	81.7	65.4	98.0







#37
2-Methylnaphthalene
Concen: 41.362 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:142 Resp: 229345

Ion	Ratio	Lower	Upper
142	100		
141	84.5	67.6	101.4







#38
1-Methylnaphthalene
Concen: 40.062 ng
RT: 8.975 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:142 Resp: 210460

Ion	Ratio	Lower	Upper
142	100		
141	91.8	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:164 Resp: 91554

Ion	Ratio	Lower	Upper
164	100		
162	99.3	79.4	119.2
160	45.9	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.130 ng
RT: 9.039 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:216 Resp: 129417

Ion	Ratio	Lower	Upper
216	100		
214	79.0	63.2	94.8
179	19.0	15.2	22.8
108	14.9	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 41.012 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:237 Resp: 39857

Ion	Ratio	Lower	Upper
237	100		
235	64.3	44.3	84.3
272	15.2	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 82.238 ng
RT: 10.710 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:330 Resp: 95908

Ion	Ratio	Lower	Upper
330	100		
332	96.3	77.0	115.6
141	24.2	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 42.346 ng
RT: 9.151 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:196 Resp: 86607

Ion	Ratio	Lower	Upper
196	100		
198	93.0	74.4	111.6
200	29.8	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 39.842 ng
RT: 9.192 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:196 Resp: 85520

Ion	Ratio	Lower	Upper
196	100		
198	97.5	78.0	117.0
97	47.1	37.7	56.5
132	25.3	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 81.692 ng
RT: 9.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:172 Resp: 531491

Ion	Ratio	Lower	Upper
172	100		
171	34.4	27.5	41.3
170	23.8	19.0	28.6







#46
1,1'-Biphenyl
Concen: 41.129 ng
RT: 9.339 min Scan# 1215
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:154 Resp: 269550

Ion	Ratio	Lower	Upper
154	100		
153	39.7	19.7	59.7
76	11.1	0.0	31.1







#47
2-Chloronaphthalene
Concen: 40.753 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 223762

Ion	Ratio	Lower	Upper
162	100		
127	39.5	31.6	47.4
164	33.1	26.5	39.7







#48
2-Nitroaniline
Concen: 42.711 ng
RT: 9.457 min Scan# 1235
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion: 65 Resp: 69638

Ion	Ratio	Lower	Upper
65	100		
92	65.4	52.3	78.5
138	97.9	78.3	117.5







#49
Acenaphthylene
Concen: 42.112 ng
RT: 9.780 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 311770

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.0	10.4	15.6







#50
Dimethylphthalate
Concen: 41.666 ng
RT: 9.633 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:163 Resp: 250045

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	9.8	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 42.535 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 50463

Ion	Ratio	Lower	Upper
165	100		
63	63.0	50.4	75.6
89	51.8	41.4	62.2







#52
Acenaphthene
Concen: 41.240 ng
RT: 9.951 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:154 Resp: 203905

Ion	Ratio	Lower	Upper
154	100		
153	112.4	89.9	134.9
152	52.5	42.0	63.0







#53
3-Nitroaniline
Concen: 44.713 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:138 Resp: 58978

Ion	Ratio	Lower	Upper
138	100		
108	9.9	7.9	11.9
92	115.0	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 35.283 ng
RT: 9.980 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:184 Resp: 22640

Ion	Ratio	Lower	Upper
184	100		
63	65.1	52.1	78.1
154	59.9	47.9	71.9







#55
Dibenzofuran
Concen: 41.592 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:168 Resp: 285490

Ion	Ratio	Lower	Upper
168	100		
139	41.5	33.2	49.8
169	15.8	12.6	19.0







#56
4-Nitrophenol
Concen: 38.279 ng
RT: 10.027 min Scan# 1332
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:139 Resp: 34191

Ion	Ratio	Lower	Upper
139	100		
109	73.8	53.8	93.8
65	115.9	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 44.075 ng
RT: 10.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 70572

Ion	Ratio	Lower	Upper
165	100		
63	44.3	35.4	53.2
89	65.9	52.7	79.1
182	2.9	2.3	3.5







#58
Fluorene
Concen: 41.604 ng
RT: 10.469 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:166 Resp: 217594

Ion	Ratio	Lower	Upper
166	100		
165	96.2	77.0	115.4
167	12.8	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 43.585 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:232 Resp: 65926

Ion	Ratio	Lower	Upper
232	100		
131	39.9	31.9	47.9
130	1.4	1.1	1.7
166	26.2	21.0	31.4







#60
Diethylphthalate
Concen: 41.838 ng
RT: 10.339 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:149 Resp: 236119

Ion	Ratio	Lower	Upper
149	100		
177	22.2	17.8	26.6
150	12.6	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 42.022 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:204 Resp: 123596

Ion	Ratio	Lower	Upper
204	100		
206	34.7	27.8	41.6
141	56.8	45.4	68.2







#62
4-Nitroaniline
Concen: 43.392 ng
RT: 10.480 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:138 Resp: 54231

Ion	Ratio	Lower	Upper
138	100		
92	50.5	30.5	70.5
108	76.2	56.2	96.2







#63
Azobenzene
Concen: 42.381 ng
RT: 10.622 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 226383

Ion	Ratio	Lower	Upper
77	100		
182	29.0	9.0	49.0
105	21.4	1.4	41.4
51	36.9	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:188 Resp: 159613

Ion	Ratio	Lower	Upper
188	100		
94	6.7	5.4	8.0
80	7.5	6.0	9.0






#65
4,6-Dinitro-2-methylphenol
Concen: 38.223 ng
RT: 10.516 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:198 Resp: 39142


Ion	Ratio	Lower	Upper
198	100		
51	45.8	25.8	65.8
105	37.2	17.2	57.2



#66
n-Nitrosodiphenylamine
Concen: 40.226 ng
RT: 10.574 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:169 Resp: 205138

Ion	Ratio	Lower	Upper
169	100		
168	59.3	47.4	71.2
167	31.9	25.5	38.3





#67
4-Bromophenyl-phenylether
Concen: 41.585 ng
RT: 10.951 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:248 Resp: 83105

Ion	Ratio	Lower	Upper
248	100		
250	97.2	77.8	116.6
141	63.1	50.5	75.7







#68
Hexachlorobenzene
Concen: 41.869 ng
RT: 11.016 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:284 Resp: 102619

Ion	Ratio	Lower	Upper
284	100		
142	25.0	20.0	30.0
249	27.7	22.2	33.2







#69
Atrazine
Concen: 39.753 ng
RT: 11.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:200 Resp: 66524

Ion	Ratio	Lower	Upper
200	100		
173	24.9	4.9	44.9
215	48.2	28.2	68.2







#70
Pentachlorophenol
Concen: 40.953 ng
RT: 11.216 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:266 Resp: 47778

Ion	Ratio	Lower	Upper
266	100		
268	60.3	48.2	72.4
264	61.7	49.4	74.0







#71
Phenanthrene
Concen: 40.961 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:178 Resp: 324774

Ion	Ratio	Lower	Upper
178	100		
176	20.4	16.3	24.5
179	15.7	12.6	18.8







#72
Anthracene
Concen: 41.053 ng
RT: 11.486 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:178 Resp: 336252

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	14.7	11.8	17.6







#73
Carbazole
Concen: 41.458 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:167 Resp: 289029

Ion	Ratio	Lower	Upper
167	100		
166	20.1	16.1	24.1
139	12.2	9.8	14.6







#74
Di-n-butylphthalate
Concen: 41.770 ng
RT: 11.969 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:149 Resp: 357025

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	4.5	3.6	5.4







#75
Fluoranthene
Concen: 40.917 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:202 Resp: 358431

Ion	Ratio	Lower	Upper
202	100		
101	7.7	0.0	27.7
203	17.3	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:240 Resp: 123616

Ion	Ratio	Lower	Upper
240	100		
120	8.1	6.5	9.7
236	25.1	20.1	30.1







#77
Benzidine
Concen: 42.766 ng
RT: 12.751 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:184 Resp: 164527

Ion	Ratio	Lower	Upper
184	100		
185	14.5	11.6	17.4
183	10.6	8.5	12.7







#78
Pyrene
Concen: 42.300 ng
RT: 12.857 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:202 Resp: 375923

Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.2	25.8
203	16.9	13.5	20.3







#79
Terphenyl-d14
Concen: 84.737 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:244 Resp: 599175

Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.8	8.6
122	6.1	4.9	7.3







#80
Butylbenzylphthalate
Concen: 44.204 ng
RT: 13.474 min Scan# 1918
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:149 Resp: 148621

Ion	Ratio	Lower	Upper
149	100		
91	70.7	56.6	84.8
206	22.9	18.3	27.5







#81
Benzo(a)anthracene
Concen: 44.863 ng
RT: 14.045 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 374791

Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.4	32.2
229	19.1	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 43.337 ng
RT: 14.010 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:252 Resp: 126028

Ion	Ratio	Lower	Upper
252	100		
254	62.4	49.9	74.9
126	8.1	6.5	9.7







#83
Chrysene
Concen: 41.590 ng
RT: 14.086 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 326041

Ion	Ratio	Lower	Upper
228	100		
226	29.7	23.8	35.6
229	19.5	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 42.805 ng
RT: 14.033 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:149 Resp: 196471

Ion	Ratio	Lower	Upper
149	100		
167	28.1	22.5	33.7
279	4.5	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 43.804 ng
RT: 14.645 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:149 Resp: 345760

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	11.6	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:264 Resp: 152855

Ion	Ratio	Lower	Upper
264	100		
260	21.2	17.0	25.4
265	20.7	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 43.020 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:276 Resp: 451467

Ion	Ratio	Lower	Upper
276	100		
138	16.3	13.0	19.6
277	24.2	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 44.700 ng
RT: 15.104 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:252 Resp: 386268

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.2	25.8
125	6.3	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 40.621 ng
RT: 15.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:252 Resp: 323867

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.2
125	6.1	4.9	7.3







#90
Benzo(a)pyrene
Concen: 42.927 ng
RT: 15.480 min Scan# 2259
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:252 Resp: 332791

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.1	25.7
125	6.8	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 42.914 ng
RT: 17.062 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:278 Resp: 356487

Ion	Ratio	Lower	Upper
278	100		
139	10.6	8.5	12.7
279	23.3	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 43.758 ng
RT: 17.498 min Scan# 2602
Delta R.T. 0.000 min
Lab File: BF144060.D
Acq: 24 Oct 2025 13:05

Tgt Ion:276 Resp: 365458

Ion	Ratio	Lower	Upper
276	100		
277	22.4	17.9	26.9
138	15.0	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144061.D
Acq On : 24 Oct 2025 13:34
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Quant Time: Oct 24 17:04:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	52198	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	187852	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	95790	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	163068	20.000	ng	0.00
76) Chrysene-d12	14.057	240	133555	20.000	ng	0.00
86) Perylene-d12	15.539	264	170761	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	303472	91.982	ng	0.00
7) Phenol-d6	6.504	99	336307	92.298	ng	0.00
23) Nitrobenzene-d5	7.445	82	314744	91.795	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	115501	94.659	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	618497	90.861	ng	0.00
79) Terphenyl-d14	12.998	244	669013	87.573	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.628	88	67694	47.481	ng	99
3) Pyridine	3.399	79	186251	48.798	ng	99
4) n-Nitrosodimethylamine	3.346	42	108941	46.451	ng	96
6) Aniline	6.545	93	236051	45.283	ng	97
8) 2-Chlorophenol	6.663	128	146647	45.134	ng	100
9) Benzaldehyde	6.434	77	125111	44.224	ng	99
10) Phenol	6.522	94	205466	45.707	ng	99
11) bis(2-Chloroethyl)ether	6.616	93	147198	45.316	ng	100
12) 1,3-Dichlorobenzene	6.822	146	184614	46.062	ng	99
13) 1,4-Dichlorobenzene	6.898	146	178849	45.279	ng	99
14) 1,2-Dichlorobenzene	7.051	146	176825	46.267	ng	98
15) Benzyl Alcohol	7.022	79	137745	46.492	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.151	45	280440	45.937	ng	100
17) 2-Methylphenol	7.128	107	114018	45.627	ng	96
18) Hexachloroethane	7.392	117	66036	46.762	ng	94
19) n-Nitroso-di-n-propyla...	7.292	70	104565	44.378	ng	99
20) 3+4-Methylphenols	7.281	107	135141	45.746	ng	96
22) Acetophenone	7.287	105	186463	45.930	ng	99
24) Nitrobenzene	7.463	77	169181	46.643	ng	97
25) Isophorone	7.698	82	274478	45.788	ng	99
26) 2-Nitrophenol	7.781	139	78217	45.330	ng	96
27) 2,4-Dimethylphenol	7.810	122	119169	46.475	ng	97
28) bis(2-Chloroethoxy)met...	7.910	93	171275	45.816	ng	99
29) 2,4-Dichlorophenol	8.016	162	137586	46.140	ng	98
30) 1,2,4-Trichlorobenzene	8.104	180	141259	46.047	ng	100
31) Naphthalene	8.187	128	400308	45.150	ng	99
32) Benzoic acid	7.922	122	68843	42.326	ng	90
33) 4-Chloroaniline	8.234	127	139897	45.326	ng	99
34) Hexachlorobutadiene	8.298	225	113173	45.117	ng	99
35) Caprolactam	8.598	113	33776	45.169	ng	91
36) 4-Chloro-3-methylphenol	8.710	107	120661	45.853	ng	96
37) 2-Methylnaphthalene	8.875	142	257754	44.094	ng	95
38) 1-Methylnaphthalene	8.975	142	245890	44.400	ng	99
40) 1,2,4,5-Tetrachloroben...	9.039	216	150659	45.764	ng	99
41) Hexachlorocyclopentadiene	9.028	237	49167	48.354	ng	96
43) 2,4,6-Trichlorophenol	9.151	196	102347	47.829	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144061.D
Acq On : 24 Oct 2025 13:34
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Quant Time: Oct 24 17:04:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	102322	45.562	ng		99
46) 1,1'-Biphenyl	9.339	154	315397	45.996	ng		99
47) 2-Chloronaphthalene	9.363	162	263764	45.914	ng		98
48) 2-Nitroaniline	9.463	65	79650	46.691	ng		93
49) Acenaphthylene	9.781	152	353529	45.641	ng		100
50) Dimethylphthalate	9.633	163	291936	46.495	ng		99
51) 2,6-Dinitrotoluene	9.698	165	60313	48.590	ng		98
52) Acenaphthene	9.951	154	239254	46.250	ng		98
53) 3-Nitroaniline	9.869	138	67246	48.727	ng		98
54) 2,4-Dinitrophenol	9.981	184	27746	41.328	ng		95
55) Dibenzofuran	10.122	168	333154	46.390	ng		100
56) 4-Nitrophenol	10.028	139	41829	44.760	ng		98
57) 2,4-Dinitrotoluene	10.104	165	80539	48.075	ng		96
58) Fluorene	10.469	166	249298	45.558	ng		97
59) 2,3,4,6-Tetrachlorophenol	10.245	232	77175	48.766	ng	#	97
60) Diethylphthalate	10.339	149	274281	46.451	ng		99
61) 4-Chlorophenyl-phenyle...	10.457	204	141601	46.015	ng		96
62) 4-Nitroaniline	10.486	138	62266	47.618	ng		99
63) Azobenzene	10.622	77	264030	47.243	ng		99
65) 4,6-Dinitro-2-methylph...	10.516	198	46938	44.864	ng		95
66) n-Nitrosodiphenylamine	10.575	169	240557	46.171	ng		97
67) 4-Bromophenyl-phenylether	10.951	248	93911	45.996	ng		98
68) Hexachlorobenzene	11.016	284	111943	44.706	ng		96
69) Atrazine	11.104	200	75502	44.162	ng		98
70) Pentachlorophenol	11.216	266	58380	48.981	ng		98
71) Phenanthrene	11.433	178	374285	46.205	ng		99
72) Anthracene	11.486	178	380096	45.423	ng		99
73) Carbazole	11.639	167	333654	46.845	ng		98
74) Di-n-butylphthalate	11.969	149	408630	46.795	ng		99
75) Fluoranthene	12.627	202	406402	45.411	ng		99
77) Benzidine	12.751	184	171303	41.214	ng		99
78) Pyrene	12.857	202	432441	45.038	ng		99
80) Butylbenzylphthalate	13.474	149	164976	45.417	ng		99
81) Benzo(a)anthracene	14.045	228	402629	44.609	ng		99
82) 3,3'-Dichlorobenzidine	14.010	252	134912	42.940	ng		95
83) Chrysene	14.086	228	393508	46.461	ng		99
84) Bis(2-ethylhexyl)phtha...	14.027	149	230343	46.450	ng		96
85) Di-n-octyl phthalate	14.645	149	393623	46.156	ng		98
87) Indeno(1,2,3-cd)pyrene	17.045	276	552781	47.151	ng		100
88) Benzo(b)fluoranthene	15.104	252	450302	46.646	ng		98
89) Benzo(k)fluoranthene	15.133	252	386955	43.445	ng		99
90) Benzo(a)pyrene	15.480	252	402167	46.436	ng		99
91) Dibenzo(a,h)anthracene	17.062	278	437275	47.119	ng		100
92) Benzo(g,h,i)perylene	17.504	276	436689	46.804	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144061.D
Acq On : 24 Oct 2025 13:34
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Quant Time: Oct 24 17:04:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 79
Delta R.T. -0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 52198

Ion	Ratio	Lower	Upper
152	100		
150	167.1	124.4	186.6
115	60.5	49.4	74.0







#2
1,4-Dioxane
Concen: 47.481 ng
RT: 2.628 min Scan# 74
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion: 88 Resp: 67694

Ion	Ratio	Lower	Upper
88	100		
58	95.6	76.5	114.7
43	46.0	35.3	52.9







#3
Pyridine
Concen: 48.798 ng
RT: 3.399 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 186251

Ion	Ratio	Lower	Upper
79	100		
52	83.5	67.5	101.3
51	43.3	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 46.451 ng
RT: 3.346 min Scan# 196
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion: 42 Resp: 108941

Ion	Ratio	Lower	Upper
42	100		
74	91.9	76.4	114.6
44	7.1	5.2	7.8







#5
2-Fluorophenol
Concen: 91.982 ng
RT: 5.493 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:112 Resp: 303472

Ion	Ratio	Lower	Upper
112	100		
64	64.1	49.8	74.8
63	35.1	27.4	41.2







#6
Aniline
Concen: 45.283 ng
RT: 6.545 min Scan# 740
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion: 93 Resp: 236051

Ion	Ratio	Lower	Upper
93	100		
66	42.4	35.7	53.5
65	20.7	17.0	25.6







#7
Phenol-d6
Concen: 92.298 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 99 Resp: 336307

Ion	Ratio	Lower	Upper
99	100		
42	30.5	23.9	35.9
71	43.3	34.4	51.6







#8
2-Chlorophenol
Concen: 45.134 ng
RT: 6.663 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:128 Resp: 146647

Ion	Ratio	Lower	Upper
128	100		
130	30.9	10.7	50.7
64	54.4	34.7	74.7







#9
Benzaldehyde
Concen: 44.224 ng
RT: 6.434 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 125111

Ion	Ratio	Lower	Upper
77	100		
105	92.5	72.2	112.2
106	88.6	67.9	107.9







#10
Phenol
Concen: 45.707 ng
RT: 6.522 min Scan# 736
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion: 94 Resp: 205466

Ion	Ratio	Lower	Upper
94	100		
65	33.6	13.1	53.1
66	50.0	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 45.316 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 93 Resp: 147198

Ion	Ratio	Lower	Upper
93	100		
63	80.2	60.2	100.2
95	31.8	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 46.062 ng
RT: 6.822 min Scan# 787
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:146 Resp: 184614

Ion	Ratio	Lower	Upper
146	100		
148	62.8	50.6	75.8
75	33.8	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 45.279 ng
RT: 6.898 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:146 Resp: 178849

Ion	Ratio	Lower	Upper
146	100		
148	64.7	51.0	76.6
111	43.4	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 46.267 ng
RT: 7.051 min Scan# 826
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:146 Resp: 176825

Ion	Ratio	Lower	Upper
146	100		
148	61.8	50.8	76.2
111	42.6	34.7	52.1







#15
Benzyl Alcohol
Concen: 46.492 ng
RT: 7.022 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 137745

Ion	Ratio	Lower	Upper
79	100		
108	67.2	54.8	82.2
77	62.4	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 45.937 ng
RT: 7.151 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion: 45 Resp: 280440

Ion	Ratio	Lower	Upper
45	100		
77	11.1	0.0	30.9
79	8.3	0.0	28.2







#17
2-Methylphenol
Concen: 45.627 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:107 Resp: 114018

Ion	Ratio	Lower	Upper
107	100		
108	114.6	96.4	144.6
77	56.6	46.8	70.2
79	56.2	45.8	68.6







#18
Hexachloroethane
Concen: 46.762 ng
RT: 7.392 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:117 Resp: 66036

Ion	Ratio	Lower	Upper
117	100		
119	91.6	76.6	115.0
201	117.9	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 44.378 ng
RT: 7.292 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 70 Resp: 104565

Ion	Ratio	Lower	Upper
70	100		
42	83.3	67.4	101.0
101	9.3	7.1	10.7
130	16.7	13.5	20.3







#20
3+4-Methylphenols
Concen: 45.746 ng
RT: 7.281 min Scan# 865
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:107 Resp: 135141

Ion	Ratio	Lower	Upper
107	100		
108	88.6	68.0	108.0
77	120.4	108.7	148.7
79	36.7	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:136 Resp: 187852

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	13.0
54	8.7	6.6	9.8
68	5.3	4.6	6.8







#22
Acetophenone
Concen: 45.930 ng
RT: 7.287 min Scan# 866
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:105 Resp: 186463

Ion	Ratio	Lower	Upper
105	100		
71	8.9	6.9	10.3
51	51.7	40.6	60.8
120	20.2	16.8	25.2







#23
Nitrobenzene-d5
Concen: 91.795 ng
RT: 7.445 min Scan# 89
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 314744

Ion	Ratio	Lower	Upper
82	100		
128	41.7	32.4	48.6
54	66.8	54.0	81.0







#24
Nitrobenzene
Concen: 46.643 ng
RT: 7.463 min Scan# 896
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion: 77 Resp: 169181

Ion	Ratio	Lower	Upper
77	100		
123	41.9	35.0	52.6
65	14.6	12.2	18.2







#25
Isophorone
Concen: 45.788 ng
RT: 7.698 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 274478

Ion	Ratio	Lower	Upper
82	100		
95	8.0	6.6	9.8
138	16.5	13.8	20.8







#26
2-Nitrophenol
Concen: 45.330 ng
RT: 7.781 min Scan# 950
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:139 Resp: 78217

Ion	Ratio	Lower	Upper
139	100		
109	32.5	24.6	37.0
65	55.0	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 46.475 ng
RT: 7.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:122 Resp: 119169

Ion	Ratio	Lower	Upper
122	100		
107	119.3	91.8	137.8
121	55.2	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 45.816 ng
RT: 7.910 min Scan# 972
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion: 93 Resp: 171275

Ion	Ratio	Lower	Upper
93	100		
95	31.5	24.6	36.8
123	11.0	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 46.140 ng
RT: 8.016 min Scan# 99
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 137586

Ion	Ratio	Lower	Upper
162	100		
164	62.8	42.1	82.1
98	34.3	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 46.047 ng
RT: 8.104 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:180 Resp: 141259

Ion	Ratio	Lower	Upper
180	100		
182	109.0	87.1	130.7
145	32.5	25.3	37.9







#31
Naphthalene
Concen: 45.150 ng
RT: 8.187 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:128 Resp: 400308

Ion	Ratio	Lower	Upper
128	100		
129	10.4	8.4	12.6
127	12.9	9.9	14.9







#32
Benzoic acid
Concen: 42.326 ng
RT: 7.922 min Scan# 974
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:122 Resp: 68843

Ion	Ratio	Lower	Upper
122	100		
105	122.6	111.3	151.3
77	102.9	96.7	136.7







#33
4-Chloroaniline
Concen: 45.326 ng
RT: 8.234 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:127 Resp: 139897

Ion	Ratio	Lower	Upper
127	100		
129	31.6	25.6	38.4
65	38.0	30.9	46.3
92	23.2	18.9	28.3







#34
Hexachlorobutadiene
Concen: 45.117 ng
RT: 8.298 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:225 Resp: 113173

Ion	Ratio	Lower	Upper
225	100		
223	63.7	50.6	76.0
227	63.0	51.8	77.8







#35
Caprolactam
Concen: 45.169 ng
RT: 8.598 min Scan# 10
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:113 Resp: 33776

Ion	Ratio	Lower	Upper
113	100		
55	194.0	160.1	200.1
56	155.4	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 45.853 ng
RT: 8.710 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:107 Resp: 120661

Ion	Ratio	Lower	Upper
107	100		
144	26.9	21.6	32.4
142	77.0	65.4	98.0







#37
2-Methylnaphthalene
Concen: 44.094 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:142 Resp: 257754

Ion	Ratio	Lower	Upper
142	100		
141	88.9	67.6	101.4







#38
1-Methylnaphthalene
Concen: 44.400 ng
RT: 8.975 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:142 Resp: 245890

Ion	Ratio	Lower	Upper
142	100		
141	91.2	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:164 Resp: 95790

Ion	Ratio	Lower	Upper
164	100		
162	102.3	79.4	119.2
160	45.5	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 45.764 ng
RT: 9.039 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:216 Resp: 150659

Ion	Ratio	Lower	Upper
216	100		
214	78.9	63.2	94.8
179	20.9	15.2	22.8
108	14.6	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 48.354 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:237 Resp: 49167

Ion	Ratio	Lower	Upper
237	100		
235	61.8	44.3	84.3
272	13.0	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 94.659 ng
RT: 10.710 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:330 Resp: 115501

Ion	Ratio	Lower	Upper
330	100		
332	93.6	77.0	115.6
141	24.0	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 47.829 ng
RT: 9.151 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:196 Resp: 102347

Ion	Ratio	Lower	Upper
196	100		
198	94.8	74.4	111.6
200	29.3	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 45.562 ng
RT: 9.192 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:196 Resp: 102322

Ion	Ratio	Lower	Upper
196	100		
198	96.6	78.0	117.0
97	45.1	37.7	56.5
132	25.3	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 90.861 ng
RT: 9.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:172 Resp: 618497

Ion	Ratio	Lower	Upper
172	100		
171	33.9	27.5	41.3
170	23.7	19.0	28.6







#46
1,1'-Biphenyl
Concen: 45.996 ng
RT: 9.339 min Scan# 1215
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:154 Resp: 315397

Ion	Ratio	Lower	Upper
154	100		
153	39.3	19.7	59.7
76	10.8	0.0	31.1







#47
2-Chloronaphthalene
Concen: 45.914 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 263764

Ion	Ratio	Lower	Upper
162	100		
127	38.2	31.6	47.4
164	32.2	26.5	39.7







#48
2-Nitroaniline
Concen: 46.691 ng
RT: 9.463 min Scan# 1236
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion: 65 Resp: 79650

Ion	Ratio	Lower	Upper
65	100		
92	67.9	52.3	78.5
138	106.7	78.3	117.5







#49
Acenaphthylene
Concen: 45.641 ng
RT: 9.781 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 353529

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.1	10.4	15.6







#50
Dimethylphthalate
Concen: 46.495 ng
RT: 9.633 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:163 Resp: 291936

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.4	5.0
164	9.6	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 48.590 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 60313

Ion	Ratio	Lower	Upper
165	100		
63	65.2	50.4	75.6
89	52.8	41.4	62.2







#52
Acenaphthene
Concen: 46.250 ng
RT: 9.951 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:154 Resp: 239254

Ion	Ratio	Lower	Upper
154	100		
153	110.7	89.9	134.9
152	53.7	42.0	63.0







#53
3-Nitroaniline
Concen: 48.727 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:138 Resp: 67246

Ion	Ratio	Lower	Upper
138	100		
108	9.7	7.9	11.9
92	112.5	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 41.328 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:184 Resp: 27746

Ion	Ratio	Lower	Upper
184	100		
63	66.6	52.1	78.1
154	66.0	47.9	71.9







#55
Dibenzofuran
Concen: 46.390 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:168 Resp: 333154

Ion	Ratio	Lower	Upper
168	100		
139	41.3	33.2	49.8
169	15.6	12.6	19.0







#56
4-Nitrophenol
Concen: 44.760 ng
RT: 10.028 min Scan# 1332
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:139 Resp: 41829

Ion	Ratio	Lower	Upper
139	100		
109	71.1	53.8	93.8
65	114.7	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 48.075 ng
RT: 10.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 80539

Ion	Ratio	Lower	Upper
165	100		
63	48.0	35.4	53.2
89	63.1	52.7	79.1
182	2.5	2.3	3.5







#58
Fluorene
Concen: 45.558 ng
RT: 10.469 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:166 Resp: 249298

Ion	Ratio	Lower	Upper
166	100		
165	99.0	77.0	115.4
167	13.8	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 48.766 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:232 Resp: 77175

Ion	Ratio	Lower	Upper
232	100		
131	42.1	31.9	47.9
130	1.9	1.1	1.7#
166	25.4	21.0	31.4







#60
Diethylphthalate
Concen: 46.451 ng
RT: 10.339 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:149 Resp: 274281

Ion	Ratio	Lower	Upper
149	100		
177	22.6	17.8	26.6
150	11.9	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 46.015 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:204 Resp: 141601

Ion	Ratio	Lower	Upper
204	100		
206	32.5	27.8	41.6
141	60.2	45.4	68.2







#62
4-Nitroaniline
Concen: 47.618 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:138 Resp: 62266

Ion	Ratio	Lower	Upper
138	100		
92	49.8	30.5	70.5
108	76.6	56.2	96.2







#63
Azobenzene
Concen: 47.243 ng
RT: 10.622 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 264030

Ion	Ratio	Lower	Upper
77	100		
182	27.9	9.0	49.0
105	21.0	1.4	41.4
51	36.1	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:188 Resp: 163068

Ion	Ratio	Lower	Upper
188	100		
94	7.7	5.4	8.0
80	8.0	6.0	9.0







#65
4,6-Dinitro-2-methylphenol
Concen: 44.864 ng
RT: 10.516 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:198 Resp: 46938

Ion	Ratio	Lower	Upper
198	100		
51	49.8	25.8	65.8
105	34.5	17.2	57.2







#66
n-Nitrosodiphenylamine
Concen: 46.171 ng
RT: 10.575 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:169 Resp: 240557

Ion	Ratio	Lower	Upper
169	100		
168	56.2	47.4	71.2
167	32.1	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 45.996 ng
RT: 10.951 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:248 Resp: 93911

Ion	Ratio	Lower	Upper
248	100		
250	99.5	77.8	116.6
141	64.5	50.5	75.7







#68
Hexachlorobenzene
Concen: 44.706 ng
RT: 11.016 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:284 Resp: 111943

Ion	Ratio	Lower	Upper
284	100		
142	25.9	20.0	30.0
249	30.3	22.2	33.2







#69
Atrazine
Concen: 44.162 ng
RT: 11.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:200 Resp: 75502

Ion	Ratio	Lower	Upper
200	100		
173	25.4	4.9	44.9
215	46.0	28.2	68.2







#70
Pentachlorophenol
Concen: 48.981 ng
RT: 11.216 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:266 Resp: 58380

Ion	Ratio	Lower	Upper
266	100		
268	61.6	48.2	72.4
264	62.8	49.4	74.0







#71
Phenanthrene
Concen: 46.205 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:178 Resp: 374285

Ion	Ratio	Lower	Upper
178	100		
176	19.6	16.3	24.5
179	15.5	12.6	18.8







#72
Anthracene
Concen: 45.423 ng
RT: 11.486 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:178 Resp: 380096

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.0	22.4
179	15.0	11.8	17.6







#73
Carbazole
Concen: 46.845 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:167 Resp: 333654

Ion	Ratio	Lower	Upper
167	100		
166	20.9	16.1	24.1
139	12.8	9.8	14.6







#74
Di-n-butylphthalate
Concen: 46.795 ng
RT: 11.969 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:149 Resp: 408630

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.3	10.9
104	4.6	3.6	5.4







#75
Fluoranthene
Concen: 45.411 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:202 Resp: 406402

Ion	Ratio	Lower	Upper
202	100		
101	7.4	0.0	27.7
203	17.6	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:240 Resp: 133555

Ion	Ratio	Lower	Upper
240	100		
120	8.2	6.5	9.7
236	25.3	20.1	30.1







#77
Benzidine
Concen: 41.214 ng
RT: 12.751 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:184 Resp: 171303

Ion	Ratio	Lower	Upper
184	100		
185	14.9	11.6	17.4
183	10.5	8.5	12.7







#78
Pyrene
Concen: 45.038 ng
RT: 12.857 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:202 Resp: 432441

Ion	Ratio	Lower	Upper
202	100		
200	21.0	17.2	25.8
203	17.0	13.5	20.3







#79
Terphenyl-d14
Concen: 87.573 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:244 Resp: 669013

Ion	Ratio	Lower	Upper
244	100		
212	6.9	5.8	8.6
122	6.6	4.9	7.3







#80
Butylbenzylphthalate
Concen: 45.417 ng
RT: 13.474 min Scan# 1918
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:149 Resp: 164976

Ion	Ratio	Lower	Upper
149	100		
91	70.0	56.6	84.8
206	23.0	18.3	27.5







#81
Benzo(a)anthracene
Concen: 44.609 ng
RT: 14.045 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 402629

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.4	32.2
229	19.7	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 42.940 ng
RT: 14.010 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:252 Resp: 134912

Ion	Ratio	Lower	Upper
252	100		
254	66.9	49.9	74.9
126	7.5	6.5	9.7







#83
Chrysene
Concen: 46.461 ng
RT: 14.086 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 393508

Ion	Ratio	Lower	Upper
228	100		
226	30.0	23.8	35.6
229	19.9	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 46.450 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:149 Resp: 230343

Ion	Ratio	Lower	Upper
149	100		
167	25.8	22.5	33.7
279	4.4	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 46.156 ng
RT: 14.645 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:149 Resp: 393623

Ion	Ratio	Lower	Upper
149	100		
167	1.3	1.2	1.8
43	12.3	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:264 Resp: 170761

Ion	Ratio	Lower	Upper
264	100		
260	21.4	17.0	25.4
265	22.4	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 47.151 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:276 Resp: 552781

Ion	Ratio	Lower	Upper
276	100		
138	16.4	13.0	19.6
277	24.2	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 46.646 ng
RT: 15.104 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:252 Resp: 450302

Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.2	25.8
125	6.6	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 43.445 ng
RT: 15.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:252 Resp: 386955

Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.4	26.2
125	6.0	4.9	7.3







#90
Benzo(a)pyrene
Concen: 46.436 ng
RT: 15.480 min Scan# 2259
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:252 Resp: 402167

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.1	25.7
125	6.1	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 47.119 ng
RT: 17.062 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 437275

Ion	Ratio	Lower	Upper
278	100		
139	10.4	8.5	12.7
279	23.5	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 46.804 ng
RT: 17.504 min Scan# 2603
Delta R.T. 0.006 min
Lab File: BF144061.D
Acq: 24 Oct 2025 13:34

Tgt Ion:276 Resp: 436689

Ion	Ratio	Lower	Upper
276	100		
277	23.5	17.9	26.9
138	15.1	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144063.D
Acq On : 24 Oct 2025 14:32
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Quant Time: Oct 24 17:05:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	58739	20.000	ng		0.00
21) Naphthalene-d8	8.163	136	209353	20.000	ng		0.00
39) Acenaphthene-d10	9.922	164	112793	20.000	ng		0.00
64) Phenanthrene-d10	11.410	188	191608	20.000	ng		0.00
76) Chrysene-d12	14.062	240	151514	20.000	ng		0.00
86) Perylene-d12	15.539	264	200004	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.498	112	510582	137.523	ng		0.00
7) Phenol-d6	6.516	99	560479	136.693	ng		0.01
23) Nitrobenzene-d5	7.451	82	552167	144.501	ng		0.01
42) 2,4,6-Tribromophenol	10.716	330	212808	148.116	ng		0.00
45) 2-Fluorobiphenyl	9.239	172	1050410	131.050	ng		0.00
79) Terphenyl-d14	13.004	244	1234056	142.390	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.628	88	113019	70.445	ng		97
3) Pyridine	3.398	79	321948	74.957	ng		97
4) n-Nitrosodimethylamine	3.369	42	201953	76.521	ng		97
6) Aniline	6.545	93	393504	67.082	ng		92
8) 2-Chlorophenol	6.669	128	250635	68.549	ng		97
9) Benzaldehyde	6.434	77	244035	76.656	ng		96
10) Phenol	6.528	94	339521	67.118	ng		87
11) bis(2-Chloroethyl)ether	6.622	93	254160	69.532	ng		99
12) 1,3-Dichlorobenzene	6.822	146	305526	67.741	ng		98
13) 1,4-Dichlorobenzene	6.898	146	304178	68.434	ng		99
14) 1,2-Dichlorobenzene	7.051	146	287786	66.915	ng		99
15) Benzyl Alcohol	7.028	79	240927	72.263	ng		98
16) 2,2'-oxybis(1-Chloropr...	7.151	45	478809	69.697	ng		93
17) 2-Methylphenol	7.134	107	193458	68.795	ng		97
18) Hexachloroethane	7.392	117	112050	70.510	ng		93
19) n-Nitroso-di-n-propyla...	7.298	70	190167	71.721	ng		99
20) 3+4-Methylphenols	7.292	107	227718	68.500	ng	#	86
22) Acetophenone	7.298	105	310700	68.672	ng		99
24) Nitrobenzene	7.469	77	294221	72.786	ng		98
25) Isophorone	7.710	82	487685	73.000	ng		100
26) 2-Nitrophenol	7.781	139	138311	71.925	ng		97
27) 2,4-Dimethylphenol	7.816	122	207224	72.515	ng		98
28) bis(2-Chloroethoxy)met...	7.910	93	293996	70.566	ng		98
29) 2,4-Dichlorophenol	8.022	162	230075	69.232	ng		99
30) 1,2,4-Trichlorobenzene	8.104	180	239831	70.150	ng		97
31) Naphthalene	8.186	128	660164	66.812	ng		98
32) Benzoic acid	7.945	122	135524	74.766	ng		88
33) 4-Chloroaniline	8.233	127	239182	69.535	ng		97
34) Hexachlorobutadiene	8.304	225	196083	70.141	ng		98
35) Caprolactam	8.622	113	62887	75.462	ng		92
36) 4-Chloro-3-methylphenol	8.716	107	209253	71.352	ng		96
37) 2-Methylnaphthalene	8.880	142	438984	67.385	ng		97
38) 1-Methylnaphthalene	8.975	142	418223	67.761	ng		100
40) 1,2,4,5-Tetrachloroben...	9.045	216	263109	67.873	ng		99
41) Hexachlorocyclopentadiene	9.028	237	99937	83.468	ng		99
43) 2,4,6-Trichlorophenol	9.157	196	173959	69.040	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144063.D
Acq On : 24 Oct 2025 14:32
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Quant Time: Oct 24 17:05:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	189248	71.565	ng		96
46) 1,1'-Biphenyl	9.345	154	530392	65.690	ng		99
47) 2-Chloronaphthalene	9.369	162	446307	65.979	ng		99
48) 2-Nitroaniline	9.463	65	148164	73.762	ng		99
49) Acenaphthylene	9.780	152	603127	66.126	ng		100
50) Dimethylphthalate	9.639	163	517194	69.954	ng		99
51) 2,6-Dinitrotoluene	9.704	165	108081	73.947	ng		97
52) Acenaphthene	9.957	154	412422	67.707	ng		98
53) 3-Nitroaniline	9.875	138	112811	69.421	ng		88
54) 2,4-Dinitrophenol	9.986	184	62962	79.646	ng		94
55) Dibenzofuran	10.127	168	560264	66.253	ng		99
56) 4-Nitrophenol	10.033	139	83091	75.509	ng		99
57) 2,4-Dinitrotoluene	10.110	165	145667	73.844	ng		97
58) Fluorene	10.474	166	426390	66.175	ng		98
59) 2,3,4,6-Tetrachlorophenol	10.245	232	133917	71.865	ng	#	97
60) Diethylphthalate	10.345	149	488605	70.274	ng		99
61) 4-Chlorophenyl-phenyle...	10.463	204	238684	65.871	ng		98
62) 4-Nitroaniline	10.492	138	113935	73.997	ng		98
63) Azobenzene	10.622	77	460222	69.934	ng		96
65) 4,6-Dinitro-2-methylph...	10.522	198	88545	72.027	ng		92
66) n-Nitrosodiphenylamine	10.580	169	415732	67.908	ng		98
67) 4-Bromophenyl-phenylether	10.957	248	164677	68.643	ng		100
68) Hexachlorobenzene	11.022	284	205267	69.766	ng		97
69) Atrazine	11.110	200	129646	64.536	ng		98
70) Pentachlorophenol	11.216	266	110717	79.055	ng		97
71) Phenanthrene	11.439	178	654557	68.769	ng		99
72) Anthracene	11.492	178	656473	66.765	ng		99
73) Carbazole	11.645	167	565232	67.538	ng		97
74) Di-n-butylphthalate	11.969	149	723962	70.556	ng		99
75) Fluoranthene	12.627	202	731393	69.552	ng		99
77) Benzidine	12.751	184	282392	59.887	ng		99
78) Pyrene	12.857	202	749362	68.794	ng		99
80) Butylbenzylphthalate	13.474	149	303486	73.645	ng		97
81) Benzo(a)anthracene	14.051	228	743007	72.564	ng		99
82) 3,3'-Dichlorobenzidine	14.015	252	258546	72.536	ng		100
83) Chrysene	14.092	228	708889	73.777	ng		99
84) Bis(2-ethylhexyl)phtha...	14.033	149	406325	72.226	ng		99
85) Di-n-octyl phthalate	14.651	149	714909	73.894	ng		96
87) Indeno(1,2,3-cd)pyrene	17.062	276	1010103	73.561	ng		98
88) Benzo(b)fluoranthene	15.109	252	840121	74.302	ng		99
89) Benzo(k)fluoranthene	15.145	252	708890	67.953	ng		99
90) Benzo(a)pyrene	15.486	252	727345	71.704	ng		99
91) Dibenzo(a,h)anthracene	17.080	278	797474	73.369	ng		99
92) Benzo(g,h,i)perylene	17.521	276	811620	74.269	ng		98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144063.D
Acq On : 24 Oct 2025 14:32
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Quant Time: Oct 24 17:05:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 79
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 58739

Ion	Ratio	Lower	Upper
152	100		
150	159.3	124.4	186.6
115	63.2	49.4	74.0







#2
1,4-Dioxane
Concen: 70.445 ng
RT: 2.628 min Scan# 74
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion: 88 Resp: 113019

Ion	Ratio	Lower	Upper
88	100		
58	94.6	76.5	114.7
43	48.3	35.3	52.9







#3
Pyridine
Concen: 74.957 ng
RT: 3.398 min Scan# 20
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 321948

Ion	Ratio	Lower	Upper
79	100		
52	87.3	67.5	101.3
51	44.5	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 76.521 ng
RT: 3.369 min Scan# 200
Delta R.T. 0.024 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion: 42 Resp: 201953

Ion	Ratio	Lower	Upper
42	100		
74	92.7	76.4	114.6
44	6.7	5.2	7.8







#5
2-Fluorophenol
Concen: 137.523 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:112 Resp: 510582

Ion	Ratio	Lower	Upper
112	100		
64	64.4	49.8	74.8
63	35.5	27.4	41.2







#6
Aniline
Concen: 67.082 ng
RT: 6.545 min Scan# 740
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion: 93 Resp: 393504

Ion	Ratio	Lower	Upper
93	100		
66	50.0	35.7	53.5
65	25.4	17.0	25.6







#7
Phenol-d6
Concen: 136.693 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 99 Resp: 560479

Ion	Ratio	Lower	Upper
99	100		
42	32.7	23.9	35.9
71	46.2	34.4	51.6







#8
2-Chlorophenol
Concen: 68.549 ng
RT: 6.669 min Scan# 761
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:128 Resp: 250635

Ion	Ratio	Lower	Upper
128	100		
130	31.5	10.7	50.7
64	57.2	34.7	74.7







#9
Benzaldehyde
Concen: 76.656 ng
RT: 6.434 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 244035

Ion	Ratio	Lower	Upper
77	100		
105	87.9	72.2	112.2
106	84.6	67.9	107.9







#10
Phenol
Concen: 67.118 ng
RT: 6.528 min Scan# 737
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion: 94 Resp: 339521

Ion	Ratio	Lower	Upper
94	100		
65	37.9	13.1	53.1
66	61.1	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 69.532 ng
RT: 6.622 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 93 Resp: 254160

Ion	Ratio	Lower	Upper
93	100		
63	80.9	60.2	100.2
95	32.6	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 67.741 ng
RT: 6.822 min Scan# 787
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:146 Resp: 305526

Ion	Ratio	Lower	Upper
146	100		
148	63.3	50.6	75.8
75	37.2	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 68.434 ng
RT: 6.898 min Scan# 80
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:146 Resp: 304178

Ion	Ratio	Lower	Upper
146	100		
148	62.4	51.0	76.6
111	43.3	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 66.915 ng
RT: 7.051 min Scan# 826
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:146 Resp: 287786

Ion	Ratio	Lower	Upper
146	100		
148	63.4	50.8	76.2
111	44.8	34.7	52.1







#15
Benzyl Alcohol
Concen: 72.263 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 240927

Ion	Ratio	Lower	Upper
79	100		
108	67.0	54.8	82.2
77	63.6	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 69.697 ng
RT: 7.151 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion: 45 Resp: 478809

Ion	Ratio	Lower	Upper
45	100		
77	13.9	0.0	30.9
79	10.6	0.0	28.2







#17
2-Methylphenol
Concen: 68.795 ng
RT: 7.134 min Scan# 84
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:107 Resp: 193458

Ion	Ratio	Lower	Upper
107	100		
108	116.9	96.4	144.6
77	61.4	46.8	70.2
79	58.2	45.8	68.6







#18
Hexachloroethane
Concen: 70.510 ng
RT: 7.392 min Scan# 884
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:117 Resp: 112050

Ion	Ratio	Lower	Upper
117	100		
119	92.7	76.6	115.0
201	115.5	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 71.721 ng
RT: 7.298 min Scan# 8
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 70 Resp: 190167

Ion	Ratio	Lower	Upper
70	100		
42	84.9	67.4	101.0
101	8.2	7.1	10.7
130	15.2	13.5	20.3







#20
3+4-Methylphenols
Concen: 68.500 ng
RT: 7.292 min Scan# 867
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:107 Resp: 227718

Ion	Ratio	Lower	Upper
107	100		
108	91.1	68.0	108.0
77	156.8	108.7	148.7#
79	40.1	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 10
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:136 Resp: 209353

Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.6	13.0
54	8.7	6.6	9.8
68	5.1	4.6	6.8







#22
Acetophenone
Concen: 68.672 ng
RT: 7.298 min Scan# 868
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:105 Resp: 310700

Ion	Ratio	Lower	Upper
105	100		
71	9.2	6.9	10.3
51	50.9	40.6	60.8
120	21.3	16.8	25.2







#23
Nitrobenzene-d5
Concen: 144.501 ng
RT: 7.451 min Scan# 89
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 552167

Ion	Ratio	Lower	Upper
82	100		
128	38.2	32.4	48.6
54	68.4	54.0	81.0







#24
Nitrobenzene
Concen: 72.786 ng
RT: 7.469 min Scan# 897
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion: 77 Resp: 294221

Ion	Ratio	Lower	Upper
77	100		
123	42.5	35.0	52.6
65	16.0	12.2	18.2







#25
Isophorone
Concen: 73.000 ng
RT: 7.710 min Scan# 9
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 487685

Ion	Ratio	Lower	Upper
82	100		
95	8.2	6.6	9.8
138	17.6	13.8	20.8







#26
2-Nitrophenol
Concen: 71.925 ng
RT: 7.781 min Scan# 950
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:139 Resp: 138311

Ion	Ratio	Lower	Upper
139	100		
109	33.8	24.6	37.0
65	59.6	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 72.515 ng
RT: 7.816 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:122 Resp: 207224

Ion	Ratio	Lower	Upper
122	100		
107	117.7	91.8	137.8
121	56.8	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 70.566 ng
RT: 7.910 min Scan# 972
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion: 93 Resp: 293996

Ion	Ratio	Lower	Upper
93	100		
95	32.1	24.6	36.8
123	10.8	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 69.232 ng
RT: 8.022 min Scan# 99
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 230075

Ion	Ratio	Lower	Upper
162	100		
164	62.6	42.1	82.1
98	36.6	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 70.150 ng
RT: 8.104 min Scan# 1005
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:180 Resp: 239831

Ion	Ratio	Lower	Upper
180	100		
182	105.4	87.1	130.7
145	32.6	25.3	37.9







#31
Naphthalene
Concen: 66.812 ng
RT: 8.186 min Scan# 10
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:128 Resp: 660164

Ion	Ratio	Lower	Upper
128	100		
129	11.4	8.4	12.6
127	13.4	9.9	14.9







#32
Benzoic acid
Concen: 74.766 ng
RT: 7.945 min Scan# 978
Delta R.T. 0.029 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:122 Resp: 135524

Ion	Ratio	Lower	Upper
122	100		
105	122.4	111.3	151.3
77	97.7	96.7	136.7







#33
4-Chloroaniline
Concen: 69.535 ng
RT: 8.233 min Scan# 10
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:127 Resp: 239182

Ion	Ratio	Lower	Upper
127	100		
129	31.6	25.6	38.4
65	42.1	30.9	46.3
92	24.4	18.9	28.3







#34
Hexachlorobutadiene
Concen: 70.141 ng
RT: 8.304 min Scan# 1039
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:225 Resp: 196083

Ion	Ratio	Lower	Upper
225	100		
223	63.2	50.6	76.0
227	62.5	51.8	77.8







#35
Caprolactam
Concen: 75.462 ng
RT: 8.622 min Scan# 10
Delta R.T. 0.029 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:113 Resp: 62887

Ion	Ratio	Lower	Upper
113	100		
55	190.2	160.1	200.1
56	157.0	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 71.352 ng
RT: 8.716 min Scan# 1109
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:107 Resp: 209253

Ion	Ratio	Lower	Upper
107	100		
144	26.7	21.6	32.4
142	77.1	65.4	98.0







#37
2-Methylnaphthalene
Concen: 67.385 ng
RT: 8.880 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:142 Resp: 438984

Ion	Ratio	Lower	Upper
142	100		
141	87.0	67.6	101.4







#38
1-Methylnaphthalene
Concen: 67.761 ng
RT: 8.975 min Scan# 1153
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:142 Resp: 418223

Ion	Ratio	Lower	Upper
142	100		
141	91.6	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:164 Resp: 112793

Ion	Ratio	Lower	Upper
164	100		
162	100.9	79.4	119.2
160	42.2	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 67.873 ng
RT: 9.045 min Scan# 1165
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:216 Resp: 263109

Ion	Ratio	Lower	Upper
216	100		
214	78.5	63.2	94.8
179	19.4	15.2	22.8
108	14.5	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 83.468 ng
RT: 9.028 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:237 Resp: 99937

Ion	Ratio	Lower	Upper
237	100		
235	64.1	44.3	84.3
272	14.0	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 148.116 ng
RT: 10.716 min Scan# 1449
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:330 Resp: 212808

Ion	Ratio	Lower	Upper
330	100		
332	93.0	77.0	115.6
141	23.8	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 69.040 ng
RT: 9.157 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:196 Resp: 173959

Ion	Ratio	Lower	Upper
196	100		
198	94.1	74.4	111.6
200	30.3	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 71.565 ng
RT: 9.198 min Scan# 1191
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:196 Resp: 189248

Ion	Ratio	Lower	Upper
196	100		
198	93.4	78.0	117.0
97	44.9	37.7	56.5
132	24.3	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 131.050 ng
RT: 9.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:172 Resp: 1050410

Ion	Ratio	Lower	Upper
172	100		
171	34.7	27.5	41.3
170	24.1	19.0	28.6







#46
1,1'-Biphenyl
Concen: 65.690 ng
RT: 9.345 min Scan# 1216
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:154 Resp: 530392

Ion	Ratio	Lower	Upper
154	100		
153	38.9	19.7	59.7
76	10.8	0.0	31.1







#47
2-Chloronaphthalene
Concen: 65.979 ng
RT: 9.369 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 446307

Ion	Ratio	Lower	Upper
162	100		
127	38.8	31.6	47.4
164	33.4	26.5	39.7







#48
2-Nitroaniline
Concen: 73.762 ng
RT: 9.463 min Scan# 1236
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion: 65 Resp: 148164

Ion	Ratio	Lower	Upper
65	100		
92	64.3	52.3	78.5
138	96.5	78.3	117.5







#49
Acenaphthylene
Concen: 66.126 ng
RT: 9.780 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 603127

Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.9	23.9
153	13.2	10.4	15.6







#50
Dimethylphthalate
Concen: 69.954 ng
RT: 9.639 min Scan# 1266
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:163 Resp: 517194

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.0
164	10.1	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 73.947 ng
RT: 9.704 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 108081

Ion	Ratio	Lower	Upper
165	100		
63	65.9	50.4	75.6
89	53.8	41.4	62.2







#52
Acenaphthene
Concen: 67.707 ng
RT: 9.957 min Scan# 1320
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:154 Resp: 412422

Ion	Ratio	Lower	Upper
154	100		
153	114.0	89.9	134.9
152	55.3	42.0	63.0







#53
3-Nitroaniline
Concen: 69.421 ng
RT: 9.875 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:138 Resp: 112811

Ion	Ratio	Lower	Upper
138	100		
108	11.2	7.9	11.9
92	128.6	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 79.646 ng
RT: 9.986 min Scan# 1325
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:184 Resp: 62962

Ion	Ratio	Lower	Upper
184	100		
63	59.8	52.1	78.1
154	63.3	47.9	71.9







#55
Dibenzofuran
Concen: 66.253 ng
RT: 10.127 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:168 Resp: 560264

Ion	Ratio	Lower	Upper
168	100		
139	42.0	33.2	49.8
169	15.8	12.6	19.0







#56
4-Nitrophenol
Concen: 75.509 ng
RT: 10.033 min Scan# 1333
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:139 Resp: 83091

Ion	Ratio	Lower	Upper
139	100		
109	76.3	53.8	93.8
65	115.9	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 73.844 ng
RT: 10.110 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 145667

Ion	Ratio	Lower	Upper
165	100		
63	49.6	35.4	53.2
89	66.3	52.7	79.1
182	2.6	2.3	3.5







#58
Fluorene
Concen: 66.175 ng
RT: 10.474 min Scan# 1408
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:166 Resp: 426390

Ion	Ratio	Lower	Upper
166	100		
165	98.2	77.0	115.4
167	13.0	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 71.865 ng
RT: 10.245 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:232 Resp: 133917

Ion	Ratio	Lower	Upper
232	100		
131	42.4	31.9	47.9
130	1.9	1.1	1.7#
166	27.6	21.0	31.4







#60
Diethylphthalate
Concen: 70.274 ng
RT: 10.345 min Scan# 1386
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:149 Resp: 488605

Ion	Ratio	Lower	Upper
149	100		
177	22.2	17.8	26.6
150	11.9	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 65.871 ng
RT: 10.463 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:204 Resp: 238684

Ion	Ratio	Lower	Upper
204	100		
206	33.7	27.8	41.6
141	58.4	45.4	68.2







#62
4-Nitroaniline
Concen: 73.997 ng
RT: 10.492 min Scan# 1411
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:138 Resp: 113935

Ion	Ratio	Lower	Upper
138	100		
92	52.3	30.5	70.5
108	77.5	56.2	96.2







#63
Azobenzene
Concen: 69.934 ng
RT: 10.622 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 460222

Ion	Ratio	Lower	Upper
77	100		
182	27.0	9.0	49.0
105	19.6	1.4	41.4
51	39.4	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:188 Resp: 191608

Ion	Ratio	Lower	Upper
188	100		
94	8.0	5.4	8.0
80	8.7	6.0	9.0







#65
4,6-Dinitro-2-methylphenol
Concen: 72.027 ng
RT: 10.522 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:198 Resp: 88545

Ion	Ratio	Lower	Upper
198	100		
51	54.7	25.8	65.8
105	38.1	17.2	57.2







#66
n-Nitrosodiphenylamine
Concen: 67.908 ng
RT: 10.580 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:169 Resp: 415732

Ion	Ratio	Lower	Upper
169	100		
168	57.3	47.4	71.2
167	33.0	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 68.643 ng
RT: 10.957 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:248 Resp: 164677

Ion	Ratio	Lower	Upper
248	100		
250	97.3	77.8	116.6
141	63.3	50.5	75.7







#68
Hexachlorobenzene
Concen: 69.766 ng
RT: 11.022 min Scan# 1501
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:284 Resp: 205267

Ion	Ratio	Lower	Upper
284	100		
142	24.2	20.0	30.0
249	30.3	22.2	33.2







#69
Atrazine
Concen: 64.536 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:200 Resp: 129646

Ion	Ratio	Lower	Upper
200	100		
173	27.3	4.9	44.9
215	48.3	28.2	68.2







#70
Pentachlorophenol
Concen: 79.055 ng
RT: 11.216 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:266 Resp: 110717

Ion	Ratio	Lower	Upper
266	100		
268	63.6	48.2	72.4
264	62.9	49.4	74.0







#71
Phenanthrene
Concen: 68.769 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:178 Resp: 654557

Ion	Ratio	Lower	Upper
178	100		
176	19.7	16.3	24.5
179	15.2	12.6	18.8







#72
Anthracene
Concen: 66.765 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:178 Resp: 656473

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.0	22.4
179	15.1	11.8	17.6







#73
Carbazole
Concen: 67.538 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:167 Resp: 565232

Ion	Ratio	Lower	Upper
167	100		
166	21.4	16.1	24.1
139	13.4	9.8	14.6







#74
Di-n-butylphthalate
Concen: 70.556 ng
RT: 11.969 min Scan# 1662
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:149 Resp: 723962

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.3	10.9
104	4.9	3.6	5.4







#75
Fluoranthene
Concen: 69.552 ng
RT: 12.627 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:202 Resp: 731393

Ion	Ratio	Lower	Upper
202	100		
101	7.5	0.0	27.7
203	17.6	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.062 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:240 Resp: 151514

Ion	Ratio	Lower	Upper
240	100		
120	7.2	6.5	9.7
236	26.0	20.1	30.1







#77
Benzidine
Concen: 59.887 ng
RT: 12.751 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:184 Resp: 282392

Ion	Ratio	Lower	Upper
184	100		
185	15.1	11.6	17.4
183	11.2	8.5	12.7







#78
Pyrene
Concen: 68.794 ng
RT: 12.857 min Scan# 1813
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:202 Resp: 749362

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	17.7	13.5	20.3







#79
Terphenyl-d14
Concen: 142.390 ng
RT: 13.004 min Scan# 18
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:244 Resp: 1234056

Ion	Ratio	Lower	Upper
244	100		
212	6.9	5.8	8.6
122	5.8	4.9	7.3







#80
Butylbenzylphthalate
Concen: 73.645 ng
RT: 13.474 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:149 Resp: 303486

Ion	Ratio	Lower	Upper
149	100		
91	72.9	56.6	84.8
206	21.7	18.3	27.5







#81
Benzo(a)anthracene
Concen: 72.564 ng
RT: 14.051 min Scan# 20
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:228 Resp: 743007

Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.4	32.2
229	19.3	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 72.536 ng
RT: 14.015 min Scan# 2010
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:252 Resp: 258546

Ion	Ratio	Lower	Upper
252	100		
254	62.4	49.9	74.9
126	7.1	6.5	9.7







#83
Chrysene
Concen: 73.777 ng
RT: 14.092 min Scan# 20
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:228 Resp: 708889

Ion	Ratio	Lower	Upper
228	100		
226	29.2	23.8	35.6
229	19.4	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 72.226 ng
RT: 14.033 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:149 Resp: 406325

Ion	Ratio	Lower	Upper
149	100		
167	27.4	22.5	33.7
279	4.7	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 73.894 ng
RT: 14.651 min Scan# 21
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 714909

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	13.2	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:264 Resp: 200004

Ion	Ratio	Lower	Upper
264	100		
260	20.8	17.0	25.4
265	22.1	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 73.561 ng
RT: 17.062 min Scan# 2
Delta R.T. 0.018 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:276 Resp: 1010103

Ion	Ratio	Lower	Upper
276	100		
138	15.6	13.0	19.6
277	25.4	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 74.302 ng
RT: 15.109 min Scan# 2196
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:252 Resp: 840121

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.2	25.8
125	5.5	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 67.953 ng
RT: 15.145 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 708890

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.4	26.2
125	5.2	4.9	7.3







#90
Benzo(a)pyrene
Concen: 71.704 ng
RT: 15.486 min Scan# 2260
Delta R.T. 0.006 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:252 Resp: 727345

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.1	25.7
125	6.4	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 73.369 ng
RT: 17.080 min Scan# 2
Delta R.T. 0.018 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Tgt Ion:278 Resp: 797474

Ion	Ratio	Lower	Upper
278	100		
139	9.0	8.5	12.7
279	23.3	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 74.269 ng
RT: 17.521 min Scan# 2606
Delta R.T. 0.024 min
Lab File: BF144063.D
Acq: 24 Oct 2025 14:32

Tgt Ion:276 Resp: 811620

Ion	Ratio	Lower	Upper
276	100		
277	22.8	17.9	26.9
138	13.6	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144064.D
Acq On : 24 Oct 2025 16:24
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Quant Time: Oct 24 17:06:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	40936	20.000	ng		0.00
21) Naphthalene-d8	8.163	136	144864	20.000	ng		0.00
39) Acenaphthene-d10	9.922	164	76683	20.000	ng		0.00
64) Phenanthrene-d10	11.410	188	132038	20.000	ng		0.00
76) Chrysene-d12	14.062	240	106050	20.000	ng		0.00
86) Perylene-d12	15.539	264	140075	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.492	112	324791	125.526	ng		0.00
7) Phenol-d6	6.510	99	346692	121.325	ng		0.00
23) Nitrobenzene-d5	7.445	82	341936	129.319	ng		0.00
42) 2,4,6-Tribromophenol	10.710	330	132798	135.953	ng		0.00
45) 2-Fluorobiphenyl	9.239	172	677674	124.361	ng		0.00
79) Terphenyl-d14	12.998	244	789728	130.186	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.628	88	70224	62.807	ng		98
3) Pyridine	3.398	79	196981	65.807	ng		97
4) n-Nitrosodimethylamine	3.351	42	124870	67.890	ng		95
6) Aniline	6.545	93	251710	61.571	ng		99
8) 2-Chlorophenol	6.663	128	154530	60.645	ng		97
9) Benzaldehyde	6.434	77	172641	77.814	ng		98
10) Phenol	6.522	94	211848	60.092	ng		92
11) bis(2-Chloroethyl)ether	6.616	93	157748	61.925	ng		100
12) 1,3-Dichlorobenzene	6.822	146	192311	61.183	ng		96
13) 1,4-Dichlorobenzene	6.898	146	193362	62.421	ng		99
14) 1,2-Dichlorobenzene	7.051	146	186696	62.289	ng		99
15) Benzyl Alcohol	7.022	79	150113	64.606	ng		96
16) 2,2'-oxybis(1-Chloropr...	7.151	45	284081	59.335	ng		98
17) 2-Methylphenol	7.128	107	120515	61.494	ng		97
18) Hexachloroethane	7.392	117	73031	65.943	ng		92
19) n-Nitroso-di-n-propyla...	7.292	70	115863	62.701	ng		99
20) 3+4-Methylphenols	7.286	107	142326	61.433	ng	#	85
22) Acetophenone	7.292	105	198031	63.255	ng		99
24) Nitrobenzene	7.463	77	181356	64.837	ng		96
25) Isophorone	7.704	82	298273	64.523	ng		99
26) 2-Nitrophenol	7.781	139	84384	63.417	ng		92
27) 2,4-Dimethylphenol	7.810	122	127412	64.435	ng		97
28) bis(2-Chloroethoxy)met...	7.910	93	180863	62.737	ng		98
29) 2,4-Dichlorophenol	8.022	162	142524	61.979	ng		97
30) 1,2,4-Trichlorobenzene	8.104	180	151390	63.994	ng		97
31) Naphthalene	8.186	128	427166	62.477	ng		99
32) Benzoic acid	7.928	122	73142	58.314	ng		85
33) 4-Chloroaniline	8.233	127	149397	62.767	ng		98
34) Hexachlorobutadiene	8.298	225	128615	66.488	ng		98
35) Caprolactam	8.604	113	37413	64.879	ng		89
36) 4-Chloro-3-methylphenol	8.710	107	128805	63.473	ng		94
37) 2-Methylnaphthalene	8.875	142	276923	61.432	ng		96
38) 1-Methylnaphthalene	8.975	142	265532	62.174	ng		99
40) 1,2,4,5-Tetrachloroben...	9.039	216	164703	62.495	ng		99
41) Hexachlorocyclopentadiene	9.027	237	62879	77.248	ng		98
43) 2,4,6-Trichlorophenol	9.151	196	111523	65.103	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144064.D
Acq On : 24 Oct 2025 16:24
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Quant Time: Oct 24 17:06:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	117536	65.377	ng		96
46) 1,1'-Biphenyl	9.339	154	338254	61.621	ng		100
47) 2-Chloronaphthalene	9.369	162	287490	62.514	ng		99
48) 2-Nitroaniline	9.463	65	91867	67.272	ng		97
49) Acenaphthylene	9.780	152	383800	61.895	ng		99
50) Dimethylphthalate	9.633	163	319101	63.485	ng		100
51) 2,6-Dinitrotoluene	9.704	165	66198	66.619	ng		97
52) Acenaphthene	9.957	154	258125	62.331	ng		98
53) 3-Nitroaniline	9.874	138	69206	62.642	ng		93
54) 2,4-Dinitrophenol	9.980	184	31982	59.508	ng		88
55) Dibenzofuran	10.127	168	353228	61.440	ng		99
56) 4-Nitrophenol	10.033	139	46485	62.136	ng		95
57) 2,4-Dinitrotoluene	10.104	165	89630	66.833	ng		94
58) Fluorene	10.469	166	267202	60.997	ng		96
59) 2,3,4,6-Tetrachlorophenol	10.245	232	84250	66.502	ng	#	96
60) Diethylphthalate	10.339	149	303176	64.138	ng		100
61) 4-Chlorophenyl-phenyle...	10.463	204	155019	62.927	ng		99
62) 4-Nitroaniline	10.486	138	68899	65.819	ng		90
63) Azobenzene	10.621	77	288180	64.412	ng		96
65) 4,6-Dinitro-2-methylph...	10.516	198	52340	61.785	ng		91
66) n-Nitrosodiphenylamine	10.580	169	261183	61.911	ng		98
67) 4-Bromophenyl-phenylether	10.951	248	106864	64.641	ng		97
68) Hexachlorobenzene	11.021	284	129804	64.021	ng		98
69) Atrazine	11.104	200	93028	67.200	ng		96
70) Pentachlorophenol	11.216	266	66439	68.842	ng		98
71) Phenanthrene	11.433	178	413305	63.013	ng		98
72) Anthracene	11.486	178	421890	62.266	ng		99
73) Carbazole	11.645	167	359155	62.276	ng		97
74) Di-n-butylphthalate	11.968	149	449708	63.601	ng		99
75) Fluoranthene	12.627	202	462087	63.767	ng		99
77) Benzidine	12.751	184	250383	75.863	ng		99
78) Pyrene	12.857	202	483140	63.369	ng		99
80) Butylbenzylphthalate	13.474	149	189412	65.668	ng		99
81) Benzo(a)anthracene	14.051	228	474589	66.219	ng		99
82) 3,3'-Dichlorobenzidine	14.010	252	168368	67.487	ng		99
83) Chrysene	14.086	228	434514	64.608	ng		99
84) Bis(2-ethylhexyl)phtha...	14.033	149	258560	65.663	ng		99
85) Di-n-octyl phthalate	14.645	149	439771	64.942	ng		98
87) Indeno(1,2,3-cd)pyrene	17.056	276	630927	65.606	ng		98
88) Benzo(b)fluoranthene	15.109	252	483612	61.071	ng		98
89) Benzo(k)fluoranthene	15.139	252	486238	66.551	ng		99
90) Benzo(a)pyrene	15.480	252	456250	64.222	ng		99
91) Dibenzo(a,h)anthracene	17.068	278	502376	65.994	ng		99
92) Benzo(g,h,i)perylene	17.509	276	500846	65.439	ng		98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144064.D
Acq On : 24 Oct 2025 16:24
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Quant Time: Oct 24 17:06:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:52:41 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 79
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 40936

Ion	Ratio	Lower	Upper
152	100		
150	161.1	124.4	186.6
115	64.2	49.4	74.0







#2
1,4-Dioxane
Concen: 62.807 ng
RT: 2.628 min Scan# 74
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion: 88 Resp: 70224

Ion	Ratio	Lower	Upper
88	100		
58	95.3	76.5	114.7
43	47.2	35.3	52.9







#3
Pyridine
Concen: 65.807 ng
RT: 3.398 min Scan# 20
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 196981

Ion	Ratio	Lower	Upper
79	100		
52	82.8	67.5	101.3
51	44.9	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 67.890 ng
RT: 3.351 min Scan# 197
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion: 42 Resp: 124870

Ion	Ratio	Lower	Upper
42	100		
74	90.0	76.4	114.6
44	6.8	5.2	7.8







#5
2-Fluorophenol
Concen: 125.526 ng
RT: 5.492 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:112 Resp: 324791

Ion	Ratio	Lower	Upper
112	100		
64	64.0	49.8	74.8
63	35.7	27.4	41.2







#6
Aniline
Concen: 61.571 ng
RT: 6.545 min Scan# 740
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion: 93 Resp: 251710

Ion	Ratio	Lower	Upper
93	100		
66	44.1	35.7	53.5
65	21.5	17.0	25.6







#7
Phenol-d6
Concen: 121.325 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 99 Resp: 346692

Ion	Ratio	Lower	Upper
99	100		
42	33.3	23.9	35.9
71	48.4	34.4	51.6







#8
2-Chlorophenol
Concen: 60.645 ng
RT: 6.663 min Scan# 760
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:128 Resp: 154530

Ion	Ratio	Lower	Upper
128	100		
130	31.4	10.7	50.7
64	57.4	34.7	74.7







#9
Benzaldehyde
Concen: 77.814 ng
RT: 6.434 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 172641

Ion	Ratio	Lower	Upper
77	100		
105	91.4	72.2	112.2
106	84.9	67.9	107.9







#10
Phenol
Concen: 60.092 ng
RT: 6.522 min Scan# 736
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion: 94 Resp: 211848

Ion	Ratio	Lower	Upper
94	100		
65	36.0	13.1	53.1
66	57.4	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 61.925 ng
RT: 6.616 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 93 Resp: 157748

Ion	Ratio	Lower	Upper
93	100		
63	80.0	60.2	100.2
95	32.0	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 61.183 ng
RT: 6.822 min Scan# 787
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:146 Resp: 192311

Ion	Ratio	Lower	Upper
146	100		
148	64.9	50.6	75.8
75	37.1	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 62.421 ng
RT: 6.898 min Scan# 80
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:146 Resp: 193362

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.6
111	43.0	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 62.289 ng
RT: 7.051 min Scan# 826
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:146 Resp: 186696

Ion	Ratio	Lower	Upper
146	100		
148	61.9	50.8	76.2
111	43.2	34.7	52.1







#15
Benzyl Alcohol
Concen: 64.606 ng
RT: 7.022 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 150113

Ion	Ratio	Lower	Upper
79	100		
108	65.8	54.8	82.2
77	64.7	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 59.335 ng
RT: 7.151 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion: 45 Resp: 284081

Ion	Ratio	Lower	Upper
45	100		
77	12.3	0.0	30.9
79	8.5	0.0	28.2







#17
2-Methylphenol
Concen: 61.494 ng
RT: 7.128 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:107 Resp: 120515

Ion	Ratio	Lower	Upper
107	100		
108	116.1	96.4	144.6
77	59.3	46.8	70.2
79	59.6	45.8	68.6







#18
Hexachloroethane
Concen: 65.943 ng
RT: 7.392 min Scan# 884
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:117 Resp: 73031

Ion	Ratio	Lower	Upper
117	100		
119	91.0	76.6	115.0
201	114.5	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 62.701 ng
RT: 7.292 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 70 Resp: 115863

Ion	Ratio	Lower	Upper
70	100		
42	83.4	67.4	101.0
101	8.3	7.1	10.7
130	16.1	13.5	20.3







#20
3+4-Methylphenols
Concen: 61.433 ng
RT: 7.286 min Scan# 866
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:107 Resp: 142326

Ion	Ratio	Lower	Upper
107	100		
108	90.5	68.0	108.0
77	158.2	108.7	148.7#
79	40.5	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:136 Resp: 144864

Ion	Ratio	Lower	Upper
136	100		
137	10.2	8.6	13.0
54	8.5	6.6	9.8
68	5.2	4.6	6.8







#22
Acetophenone
Concen: 63.255 ng
RT: 7.292 min Scan# 867
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:105 Resp: 198031

Ion	Ratio	Lower	Upper
105	100		
71	8.6	6.9	10.3
51	49.4	40.6	60.8
120	21.3	16.8	25.2







#23
Nitrobenzene-d5
Concen: 129.319 ng
RT: 7.445 min Scan# 89
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 341936

Ion	Ratio	Lower	Upper
82	100		
128	38.5	32.4	48.6
54	70.2	54.0	81.0







#24
Nitrobenzene
Concen: 64.837 ng
RT: 7.463 min Scan# 896
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion: 77 Resp: 181356

Ion	Ratio	Lower	Upper
77	100		
123	40.7	35.0	52.6
65	16.0	12.2	18.2







#25
Isophorone
Concen: 64.523 ng
RT: 7.704 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 298273

Ion	Ratio	Lower	Upper
82	100		
95	8.7	6.6	9.8
138	17.1	13.8	20.8







#26
2-Nitrophenol
Concen: 63.417 ng
RT: 7.781 min Scan# 950
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:139 Resp: 84384

Ion	Ratio	Lower	Upper
139	100		
109	36.1	24.6	37.0
65	63.2	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 64.435 ng
RT: 7.810 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:122 Resp: 127412

Ion	Ratio	Lower	Upper
122	100		
107	120.0	91.8	137.8
121	56.9	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 62.737 ng
RT: 7.910 min Scan# 972
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion: 93 Resp: 180863

Ion	Ratio	Lower	Upper
93	100		
95	31.8	24.6	36.8
123	10.6	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 61.979 ng
RT: 8.022 min Scan# 99
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 142524

Ion	Ratio	Lower	Upper
162	100		
164	64.7	42.1	82.1
98	35.1	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 63.994 ng
RT: 8.104 min Scan# 1005
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:180 Resp: 151390

Ion	Ratio	Lower	Upper
180	100		
182	105.9	87.1	130.7
145	32.3	25.3	37.9







#31
Naphthalene
Concen: 62.477 ng
RT: 8.186 min Scan# 10
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:128 Resp: 427166

Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.4	12.6
127	13.0	9.9	14.9







#32
Benzoic acid
Concen: 58.314 ng
RT: 7.928 min Scan# 975
Delta R.T. 0.012 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:122 Resp: 73142

Ion	Ratio	Lower	Upper
122	100		
105	117.2	111.3	151.3
77	97.6	96.7	136.7







#33
4-Chloroaniline
Concen: 62.767 ng
RT: 8.233 min Scan# 10
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 149397

Ion	Ratio	Lower	Upper
127	100		
129	31.5	25.6	38.4
65	40.5	30.9	46.3
92	24.4	18.9	28.3







#34
Hexachlorobutadiene
Concen: 66.488 ng
RT: 8.298 min Scan# 1038
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:225 Resp: 128615

Ion	Ratio	Lower	Upper
225	100		
223	63.1	50.6	76.0
227	61.3	51.8	77.8







#35
Caprolactam
Concen: 64.879 ng
RT: 8.604 min Scan# 10
Delta R.T. 0.012 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:113 Resp: 37413

Ion	Ratio	Lower	Upper
113	100		
55	193.3	160.1	200.1
56	160.7	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 63.473 ng
RT: 8.710 min Scan# 1108
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:107 Resp: 128805

Ion	Ratio	Lower	Upper
107	100		
144	26.5	21.6	32.4
142	75.0	65.4	98.0







#37
2-Methylnaphthalene
Concen: 61.432 ng
RT: 8.875 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:142 Resp: 276923

Ion	Ratio	Lower	Upper
142	100		
141	88.0	67.6	101.4







#38
1-Methylnaphthalene
Concen: 62.174 ng
RT: 8.975 min Scan# 1153
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:142 Resp: 265532

Ion	Ratio	Lower	Upper
142	100		
141	90.5	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:164 Resp: 76683

Ion	Ratio	Lower	Upper
164	100		
162	103.7	79.4	119.2
160	44.6	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 62.495 ng
RT: 9.039 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:216 Resp: 164703

Ion	Ratio	Lower	Upper
216	100		
214	79.2	63.2	94.8
179	19.9	15.2	22.8
108	15.7	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 77.248 ng
RT: 9.027 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:237 Resp: 62879

Ion	Ratio	Lower	Upper
237	100		
235	65.4	44.3	84.3
272	13.5	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 135.953 ng
RT: 10.710 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:330 Resp: 132798

Ion	Ratio	Lower	Upper
330	100		
332	93.2	77.0	115.6
141	24.0	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 65.103 ng
RT: 9.151 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:196 Resp: 111523

Ion	Ratio	Lower	Upper
196	100		
198	92.0	74.4	111.6
200	28.2	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 65.377 ng
RT: 9.198 min Scan# 1191
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:196 Resp: 117536

Ion	Ratio	Lower	Upper
196	100		
198	94.5	78.0	117.0
97	43.6	37.7	56.5
132	23.3	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 124.361 ng
RT: 9.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:172 Resp: 677674

Ion	Ratio	Lower	Upper
172	100		
171	34.9	27.5	41.3
170	23.7	19.0	28.6







#46
1,1'-Biphenyl
Concen: 61.621 ng
RT: 9.339 min Scan# 1215
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:154 Resp: 338254

Ion	Ratio	Lower	Upper
154	100		
153	39.8	19.7	59.7
76	10.6	0.0	31.1







#47
2-Chloronaphthalene
Concen: 62.514 ng
RT: 9.369 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 287490

Ion	Ratio	Lower	Upper
162	100		
127	38.6	31.6	47.4
164	33.1	26.5	39.7







#48
2-Nitroaniline
Concen: 67.272 ng
RT: 9.463 min Scan# 1236
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion: 65 Resp: 91867

Ion	Ratio	Lower	Upper
65	100		
92	65.7	52.3	78.5
138	93.1	78.3	117.5







#49
Acenaphthylene
Concen: 61.895 ng
RT: 9.780 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 383800

Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.9	23.9
153	12.6	10.4	15.6







#50
Dimethylphthalate
Concen: 63.485 ng
RT: 9.633 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:163 Resp: 319101

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.4	5.0
164	9.9	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 66.619 ng
RT: 9.704 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 66198

Ion	Ratio	Lower	Upper
165	100		
63	63.8	50.4	75.6
89	55.7	41.4	62.2







#52
Acenaphthene
Concen: 62.331 ng
RT: 9.957 min Scan# 1320
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:154 Resp: 258125

Ion	Ratio	Lower	Upper
154	100		
153	114.5	89.9	134.9
152	54.6	42.0	63.0







#53
3-Nitroaniline
Concen: 62.642 ng
RT: 9.874 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:138 Resp: 69206

Ion	Ratio	Lower	Upper
138	100		
108	10.0	7.9	11.9
92	123.3	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 59.508 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:184 Resp: 31982

Ion	Ratio	Lower	Upper
184	100		
63	76.4	52.1	78.1
154	67.6	47.9	71.9







#55
Dibenzofuran
Concen: 61.440 ng
RT: 10.127 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:168 Resp: 353228

Ion	Ratio	Lower	Upper
168	100		
139	42.2	33.2	49.8
169	16.2	12.6	19.0







#56
4-Nitrophenol
Concen: 62.136 ng
RT: 10.033 min Scan# 1333
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:139 Resp: 46485

Ion	Ratio	Lower	Upper
139	100		
109	82.7	53.8	93.8
65	113.8	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 66.833 ng
RT: 10.104 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 89630

Ion	Ratio	Lower	Upper
165	100		
63	50.4	35.4	53.2
89	69.4	52.7	79.1
182	2.4	2.3	3.5







#58
Fluorene
Concen: 60.997 ng
RT: 10.469 min Scan# 1407
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:166 Resp: 267202

Ion	Ratio	Lower	Upper
166	100		
165	101.0	77.0	115.4
167	13.5	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 66.502 ng
RT: 10.245 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:232 Resp: 84250

Ion	Ratio	Lower	Upper
232	100		
131	42.7	31.9	47.9
130	1.8	1.1	1.7#
166	27.4	21.0	31.4







#60
Diethylphthalate
Concen: 64.138 ng
RT: 10.339 min Scan# 1385
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:149 Resp: 303176

Ion	Ratio	Lower	Upper
149	100		
177	22.3	17.8	26.6
150	12.5	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 62.927 ng
RT: 10.463 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:204 Resp: 155019

Ion	Ratio	Lower	Upper
204	100		
206	34.4	27.8	41.6
141	57.6	45.4	68.2







#62
4-Nitroaniline
Concen: 65.819 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:138 Resp: 68899

Ion	Ratio	Lower	Upper
138	100		
92	53.3	30.5	70.5
108	88.1	56.2	96.2







#63
Azobenzene
Concen: 64.412 ng
RT: 10.621 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 288180

Ion	Ratio	Lower	Upper
77	100		
182	27.3	9.0	49.0
105	20.1	1.4	41.4
51	39.4	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:188 Resp: 132038

Ion	Ratio	Lower	Upper
188	100		
94	7.0	5.4	8.0
80	7.2	6.0	9.0







#65
4,6-Dinitro-2-methylphenol
Concen: 61.785 ng
RT: 10.516 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:198 Resp: 52340

Ion	Ratio	Lower	Upper
198	100		
51	55.8	25.8	65.8
105	37.5	17.2	57.2







#66
n-Nitrosodiphenylamine
Concen: 61.911 ng
RT: 10.580 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:169 Resp: 261183

Ion	Ratio	Lower	Upper
169	100		
168	58.0	47.4	71.2
167	32.8	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 64.641 ng
RT: 10.951 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:248 Resp: 106864

Ion	Ratio	Lower	Upper
248	100		
250	95.9	77.8	116.6
141	67.7	50.5	75.7







#68
Hexachlorobenzene
Concen: 64.021 ng
RT: 11.021 min Scan# 1501
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:284 Resp: 129804

Ion	Ratio	Lower	Upper
284	100		
142	23.3	20.0	30.0
249	28.1	22.2	33.2







#69
Atrazine
Concen: 67.200 ng
RT: 11.104 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:200 Resp: 93028

Ion	Ratio	Lower	Upper
200	100		
173	26.3	4.9	44.9
215	45.6	28.2	68.2







#70
Pentachlorophenol
Concen: 68.842 ng
RT: 11.216 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:266 Resp: 66439

Ion	Ratio	Lower	Upper
266	100		
268	62.4	48.2	72.4
264	61.4	49.4	74.0







#71
Phenanthrene
Concen: 63.013 ng
RT: 11.433 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:178 Resp: 413305

Ion	Ratio	Lower	Upper
178	100		
176	19.6	16.3	24.5
179	15.1	12.6	18.8







#72
Anthracene
Concen: 62.266 ng
RT: 11.486 min Scan# 1580
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:178 Resp: 421890

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.0	22.4
179	15.2	11.8	17.6







#73
Carbazole
Concen: 62.276 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:167 Resp: 359155

Ion	Ratio	Lower	Upper
167	100		
166	21.5	16.1	24.1
139	13.7	9.8	14.6







#74
Di-n-butylphthalate
Concen: 63.601 ng
RT: 11.968 min Scan# 1662
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:149 Resp: 449708

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	4.8	3.6	5.4







#75
Fluoranthene
Concen: 63.767 ng
RT: 12.627 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:202 Resp: 462087

Ion	Ratio	Lower	Upper
202	100		
101	6.9	0.0	27.7
203	17.3	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.062 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:240 Resp: 106050

Ion	Ratio	Lower	Upper
240	100		
120	6.6	6.5	9.7
236	26.7	20.1	30.1







#77
Benzidine
Concen: 75.863 ng
RT: 12.751 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:184 Resp: 250383

Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.6	17.4
183	11.0	8.5	12.7







#78
Pyrene
Concen: 63.369 ng
RT: 12.857 min Scan# 1813
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:202 Resp: 483140

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	17.3	13.5	20.3







#79
Terphenyl-d14
Concen: 130.186 ng
RT: 12.998 min Scan# 18
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:244 Resp: 789728

Ion	Ratio	Lower	Upper
244	100		
212	6.9	5.8	8.6
122	6.2	4.9	7.3







#80
Butylbenzylphthalate
Concen: 65.668 ng
RT: 13.474 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:149 Resp: 189412

Ion	Ratio	Lower	Upper
149	100		
91	71.6	56.6	84.8
206	22.1	18.3	27.5







#81
Benzo(a)anthracene
Concen: 66.219 ng
RT: 14.051 min Scan# 20
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:228 Resp: 474589

Ion	Ratio	Lower	Upper
228	100		
226	27.6	21.4	32.2
229	19.7	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 67.487 ng
RT: 14.010 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:252 Resp: 168368

Ion	Ratio	Lower	Upper
252	100		
254	63.4	49.9	74.9
126	7.7	6.5	9.7







#83
Chrysene
Concen: 64.608 ng
RT: 14.086 min Scan# 20
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:228 Resp: 434514

Ion	Ratio	Lower	Upper
228	100		
226	29.6	23.8	35.6
229	19.0	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 65.663 ng
RT: 14.033 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:149 Resp: 258560

Ion	Ratio	Lower	Upper
149	100		
167	27.4	22.5	33.7
279	5.0	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 64.942 ng
RT: 14.645 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 439771

Ion	Ratio	Lower	Upper
149	100		
167	1.3	1.2	1.8
43	12.7	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:264 Resp: 140075

Ion	Ratio	Lower	Upper
264	100		
260	22.4	17.0	25.4
265	22.5	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 65.606 ng
RT: 17.056 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:276 Resp: 630927

Ion	Ratio	Lower	Upper
276	100		
138	14.2	13.0	19.6
277	24.4	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 61.071 ng
RT: 15.109 min Scan# 2196
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:252 Resp: 483612

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.2	25.8
125	5.2	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 66.551 ng
RT: 15.139 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 486238

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.4	26.2
125	5.4	4.9	7.3







#90
Benzo(a)pyrene
Concen: 64.222 ng
RT: 15.480 min Scan# 2259
Delta R.T. -0.000 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:252 Resp: 456250

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.1	25.7
125	6.4	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 65.994 ng
RT: 17.068 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:278 Resp: 502376

Ion	Ratio	Lower	Upper
278	100		
139	9.1	8.5	12.7
279	23.4	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 65.439 ng
RT: 17.509 min Scan# 2604
Delta R.T. 0.012 min
Lab File: BF144064.D
Acq: 24 Oct 2025 16:24

Tgt Ion:276 Resp: 500846

Ion	Ratio	Lower	Upper
276	100		
277	23.3	17.9	26.9
138	13.7	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	39660	20.000	ng		0.00
21) Naphthalene-d8	8.163	136	138565	20.000	ng		0.00
39) Acenaphthene-d10	9.916	164	76145	20.000	ng		0.00
64) Phenanthrene-d10	11.410	188	127305	20.000	ng		0.00
76) Chrysene-d12	14.057	240	97298	20.000	ng		0.00
86) Perylene-d12	15.539	264	122912	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.493	112	176184	70.283	ng		0.00
7) Phenol-d6	6.504	99	190899	68.955	ng		0.00
23) Nitrobenzene-d5	7.440	82	178800	70.696	ng		0.00
42) 2,4,6-Tribromophenol	10.710	330	66490	68.551	ng		0.00
45) 2-Fluorobiphenyl	9.234	172	366420	67.717	ng		0.00
79) Terphenyl-d14	12.998	244	409300	73.542	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.634	88	38586	35.621	ng		94
3) Pyridine	3.405	79	94470	32.576	ng		96
4) n-Nitrosodimethylamine	3.346	42	65742	36.893	ng		93
6) Aniline	6.540	93	143853	36.320	ng		98
8) 2-Chlorophenol	6.663	128	82897	33.579	ng		98
9) Benzaldehyde	6.434	77	74914	34.852	ng		98
10) Phenol	6.516	94	119040	34.853	ng		98
11) bis(2-Chloroethyl)ether	6.610	93	83544	33.851	ng		98
12) 1,3-Dichlorobenzene	6.822	146	112585	36.971	ng		99
13) 1,4-Dichlorobenzene	6.898	146	111592	37.183	ng		99
14) 1,2-Dichlorobenzene	7.051	146	104166	35.872	ng		99
15) Benzyl Alcohol	7.016	79	81782	36.330	ng		97
16) 2,2'-oxybis(1-Chloropr...	7.151	45	160104	34.517	ng		100
17) 2-Methylphenol	7.128	107	66743	35.152	ng		96
18) Hexachloroethane	7.392	117	37228	34.696	ng		98
19) n-Nitroso-di-n-propyla...	7.287	70	62662	35.002	ng		99
20) 3+4-Methylphenols	7.281	107	81018	36.095	ng		92
22) Acetophenone	7.287	105	115151	38.453	ng		97
24) Nitrobenzene	7.457	77	98332	36.753	ng		97
25) Isophorone	7.698	82	160608	36.322	ng		100
26) 2-Nitrophenol	7.775	139	46278	36.360	ng		95
27) 2,4-Dimethylphenol	7.810	122	74530	39.405	ng		97
28) bis(2-Chloroethoxy)met...	7.904	93	97014	35.182	ng		97
29) 2,4-Dichlorophenol	8.016	162	78967	35.901	ng		99
30) 1,2,4-Trichlorobenzene	8.104	180	87660	38.739	ng		99
31) Naphthalene	8.181	128	234353	35.834	ng		99
32) Benzoic acid	7.922	122	52854	44.054	ng		88
33) 4-Chloroaniline	8.234	127	94319	41.428	ng		99
34) Hexachlorobutadiene	8.298	225	70793	38.260	ng		97
35) Caprolactam	8.587	113	19488	35.331	ng	#	84
36) 4-Chloro-3-methylphenol	8.710	107	70096	36.112	ng		94
37) 2-Methylnaphthalene	8.875	142	143296	33.233	ng		94
38) 1-Methylnaphthalene	8.975	142	146838	35.945	ng		97
40) 1,2,4,5-Tetrachloroben...	9.039	216	91910	35.121	ng		99
41) Hexachlorocyclopentadiene	9.028	237	39270	48.585	ng		94
43) 2,4,6-Trichlorophenol	9.151	196	57810	33.986	ng		98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	64300	36.018	ng		94
46) 1,1'-Biphenyl	9.339	154	192660	35.345	ng		99
47) 2-Chloronaphthalene	9.363	162	159410	34.908	ng		99
48) 2-Nitroaniline	9.457	65	49659	36.621	ng		95
49) Acenaphthylene	9.781	152	243055	39.474	ng		99
50) Dimethylphthalate	9.634	163	170189	34.098	ng		99
51) 2,6-Dinitrotoluene	9.698	165	34990	35.461	ng		96
52) Acenaphthene	9.951	154	144520	35.145	ng		99
53) 3-Nitroaniline	9.869	138	39278	35.804	ng		98
54) 2,4-Dinitrophenol	9.981	184	21935	41.102	ng		94
55) Dibenzofuran	10.122	168	194900	34.140	ng		98
56) 4-Nitrophenol	10.028	139	28418	38.254	ng		95
57) 2,4-Dinitrotoluene	10.104	165	49275	37.002	ng		96
58) Fluorene	10.469	166	151653	34.864	ng		96
59) 2,3,4,6-Tetrachlorophenol	10.239	232	49354	39.232	ng	#	96
60) Diethylphthalate	10.333	149	166357	35.442	ng		99
61) 4-Chlorophenyl-phenyle...	10.457	204	86789	35.479	ng		99
62) 4-Nitroaniline	10.481	138	35087	33.756	ng		90
63) Azobenzene	10.616	77	161807	36.422	ng		98
65) 4,6-Dinitro-2-methylph...	10.516	198	30929	37.867	ng		97
66) n-Nitrosodiphenylamine	10.575	169	146733	36.075	ng		98
67) 4-Bromophenyl-phenylether	10.951	248	56286	35.313	ng		96
68) Hexachlorobenzene	11.016	284	70184	35.903	ng		98
69) Atrazine	11.104	200	48908	36.643	ng		99
70) Pentachlorophenol	11.216	266	30927	33.237	ng		95
71) Phenanthrene	11.433	178	228423	36.120	ng		98
72) Anthracene	11.486	178	233757	35.782	ng		99
73) Carbazole	11.639	167	196006	35.250	ng		98
74) Di-n-butylphthalate	11.969	149	245199	35.967	ng		100
75) Fluoranthene	12.622	202	244043	34.929	ng		99
77) Benzidine	12.745	184	146518	48.386	ng		99
78) Pyrene	12.857	202	246902	35.297	ng		100
80) Butylbenzylphthalate	13.469	149	94583	35.741	ng		98
81) Benzo(a)anthracene	14.045	228	234218	35.620	ng		98
82) 3,3'-Dichlorobenzidine	14.010	252	97608	42.643	ng		99
83) Chrysene	14.080	228	224422	36.371	ng		99
84) Bis(2-ethylhexyl)phtha...	14.033	149	130456	36.110	ng		98
85) Di-n-octyl phthalate	14.645	149	218335	35.142	ng	#	99
87) Indeno(1,2,3-cd)pyrene	17.039	276	302575	35.856	ng		100
88) Benzo(b)fluoranthene	15.104	252	234274	33.715	ng		99
89) Benzo(k)fluoranthene	15.133	252	240517	37.516	ng		99
90) Benzo(a)pyrene	15.474	252	229237	36.773	ng		99
91) Dibenzo(a,h)anthracene	17.057	278	240551	36.012	ng		98
92) Benzo(g,h,i)perylene	17.492	276	248094	36.942	ng		98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 79
Delta R.T. -0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:152 Resp: 39660

Ion	Ratio	Lower	Upper
152	100		
150	157.4	124.4	186.6
115	61.2	49.4	74.0







#2
1,4-Dioxane
Concen: 35.621 ng
RT: 2.634 min Scan# 75
Delta R.T. 0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion: 88 Resp: 38586

Ion	Ratio	Lower	Upper
88	100		
58	89.3	76.5	114.7
43	46.7	35.3	52.9







#3
Pyridine
Concen: 32.576 ng
RT: 3.405 min Scan# 20
Delta R.T. 0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 79 Resp: 94470

Ion	Ratio	Lower	Upper
79	100		
52	87.0	67.5	101.3
51	46.3	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 36.893 ng
RT: 3.346 min Scan# 196
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion: 42 Resp: 65742

Ion	Ratio	Lower	Upper
42	100		
74	88.1	76.4	114.6
44	6.0	5.2	7.8







#5
2-Fluorophenol
Concen: 70.283 ng
RT: 5.493 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:112 Resp: 176184

Ion	Ratio	Lower	Upper
112	100		
64	61.4	49.8	74.8
63	34.3	27.4	41.2







#6
Aniline
Concen: 36.320 ng
RT: 6.540 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion: 93 Resp: 143853

Ion	Ratio	Lower	Upper
93	100		
66	43.3	35.7	53.5
65	20.7	17.0	25.6







#7
Phenol-d6
Concen: 68.955 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 99 Resp: 190899

Ion	Ratio	Lower	Upper
99	100		
42	29.8	23.9	35.9
71	46.7	34.4	51.6







#8
2-Chlorophenol
Concen: 33.579 ng
RT: 6.663 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:128 Resp: 82897

Ion	Ratio	Lower	Upper
128	100		
130	32.9	10.7	50.7
64	54.4	34.7	74.7







#9
Benzaldehyde
Concen: 34.852 ng
RT: 6.434 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 77 Resp: 74914

Ion	Ratio	Lower	Upper
77	100		
105	90.9	72.2	112.2
106	85.3	67.9	107.9







#10
Phenol
Concen: 34.853 ng
RT: 6.516 min Scan# 735
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion: 94 Resp: 119040

Ion	Ratio	Lower	Upper
94	100		
65	34.2	13.1	53.1
66	52.2	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 33.851 ng
RT: 6.610 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 93 Resp: 83544

Ion	Ratio	Lower	Upper
93	100		
63	82.7	60.2	100.2
95	31.7	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 36.971 ng
RT: 6.822 min Scan# 787
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:146 Resp: 112585

Ion	Ratio	Lower	Upper
146	100		
148	63.0	50.6	75.8
75	31.9	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 37.183 ng
RT: 6.898 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:146 Resp: 111592

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.0	76.6
111	41.5	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 35.872 ng
RT: 7.051 min Scan# 826
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:146 Resp: 104166

Ion	Ratio	Lower	Upper
146	100		
148	62.4	50.8	76.2
111	43.8	34.7	52.1







#15
Benzyl Alcohol
Concen: 36.330 ng
RT: 7.016 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 79 Resp: 81782

Ion	Ratio	Lower	Upper
79	100		
108	67.4	54.8	82.2
77	64.8	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 34.517 ng
RT: 7.151 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion: 45 Resp: 160104

Ion	Ratio	Lower	Upper
45	100		
77	10.8	0.0	30.9
79	8.3	0.0	28.2







#17
2-Methylphenol
Concen: 35.152 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:107 Resp: 66743

Ion	Ratio	Lower	Upper
107	100		
108	114.1	96.4	144.6
77	57.5	46.8	70.2
79	56.1	45.8	68.6







#18
Hexachloroethane
Concen: 34.696 ng
RT: 7.392 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:117 Resp: 37228

Ion	Ratio	Lower	Upper
117	100		
119	92.8	76.6	115.0
201	128.1	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 35.002 ng
RT: 7.287 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 70 Resp: 62662

Ion	Ratio	Lower	Upper
70	100		
42	85.3	67.4	101.0
101	8.5	7.1	10.7
130	16.2	13.5	20.3







#20
3+4-Methylphenols
Concen: 36.095 ng
RT: 7.281 min Scan# 865
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:107 Resp: 81018

Ion	Ratio	Lower	Upper
107	100		
108	88.9	68.0	108.0
77	145.2	108.7	148.7
79	36.3	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:136 Resp: 138565

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.6	13.0
54	9.0	6.6	9.8
68	4.6	4.6	6.8







#22
Acetophenone
Concen: 38.453 ng
RT: 7.287 min Scan# 866
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:105 Resp: 115151

Ion	Ratio	Lower	Upper
105	100		
71	9.1	6.9	10.3
51	47.7	40.6	60.8
120	21.3	16.8	25.2







#23
Nitrobenzene-d5
Concen: 70.696 ng
RT: 7.440 min Scan# 89
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 82 Resp: 178800

Ion	Ratio	Lower	Upper
82	100		
128	40.4	32.4	48.6
54	66.9	54.0	81.0







#24
Nitrobenzene
Concen: 36.753 ng
RT: 7.457 min Scan# 895
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion: 77 Resp: 98332

Ion	Ratio	Lower	Upper
77	100		
123	41.4	35.0	52.6
65	15.1	12.2	18.2







#25
Isophorone
Concen: 36.322 ng
RT: 7.698 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 82 Resp: 160608

Ion	Ratio	Lower	Upper
82	100		
95	8.5	6.6	9.8
138	17.2	13.8	20.8







#26
2-Nitrophenol
Concen: 36.360 ng
RT: 7.775 min Scan# 949
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:139 Resp: 46278

Ion	Ratio	Lower	Upper
139	100		
109	32.2	24.6	37.0
65	62.6	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 39.405 ng
RT: 7.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:122 Resp: 74530

Ion	Ratio	Lower	Upper
122	100		
107	118.1	91.8	137.8
121	54.8	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 35.182 ng
RT: 7.904 min Scan# 971
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion: 93 Resp: 97014

Ion	Ratio	Lower	Upper
93	100		
95	32.8	24.6	36.8
123	11.2	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 35.901 ng
RT: 8.016 min Scan# 99
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:162 Resp: 78967

Ion	Ratio	Lower	Upper
162	100		
164	61.6	42.1	82.1
98	34.8	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 38.739 ng
RT: 8.104 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:180 Resp: 87660

Ion	Ratio	Lower	Upper
180	100		
182	107.3	87.1	130.7
145	31.7	25.3	37.9







#31
Naphthalene
Concen: 35.834 ng
RT: 8.181 min Scan# 10
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:128 Resp: 234353

Ion	Ratio	Lower	Upper
128	100		
129	10.2	8.4	12.6
127	13.0	9.9	14.9







#32
Benzoic acid
Concen: 44.054 ng
RT: 7.922 min Scan# 974
Delta R.T. 0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:122 Resp: 52854

Ion	Ratio	Lower	Upper
122	100		
105	121.1	111.3	151.3
77	100.0	96.7	136.7







#33
4-Chloroaniline
Concen: 41.428 ng
RT: 8.234 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:127 Resp: 94319

Ion	Ratio	Lower	Upper
127	100		
129	31.0	25.6	38.4
65	38.2	30.9	46.3
92	24.4	18.9	28.3







#34
Hexachlorobutadiene
Concen: 38.260 ng
RT: 8.298 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:225 Resp: 70793

Ion	Ratio	Lower	Upper
225	100		
223	64.2	50.6	76.0
227	61.1	51.8	77.8







#35
Caprolactam
Concen: 35.331 ng
RT: 8.587 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:113 Resp: 19488

Ion	Ratio	Lower	Upper
113	100		
55	205.9	160.1	200.1#
56	160.2	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 36.112 ng
RT: 8.710 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:107 Resp: 70096

Ion	Ratio	Lower	Upper
107	100		
144	27.7	21.6	32.4
142	74.8	65.4	98.0







#37
2-Methylnaphthalene
Concen: 33.233 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:142 Resp: 143296

Ion	Ratio	Lower	Upper
142	100		
141	89.8	67.6	101.4







#38
1-Methylnaphthalene
Concen: 35.945 ng
RT: 8.975 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:142 Resp: 146838

Ion	Ratio	Lower	Upper
142	100		
141	89.1	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:164 Resp: 76145

Ion	Ratio	Lower	Upper
164	100		
162	99.9	79.4	119.2
160	44.0	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 35.121 ng
RT: 9.039 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:216 Resp: 91910

Ion	Ratio	Lower	Upper
216	100		
214	79.3	63.2	94.8
179	20.8	15.2	22.8
108	15.4	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 48.585 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:237 Resp: 39270

Ion	Ratio	Lower	Upper
237	100		
235	58.8	44.3	84.3
272	13.6	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 68.551 ng
RT: 10.710 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:330 Resp: 66490

Ion	Ratio	Lower	Upper
330	100		
332	94.2	77.0	115.6
141	25.7	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 33.986 ng
RT: 9.151 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:196 Resp: 57810

Ion	Ratio	Lower	Upper
196	100		
198	95.5	74.4	111.6
200	28.9	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 36.018 ng
RT: 9.192 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:196 Resp: 64300

Ion	Ratio	Lower	Upper
196	100		
198	90.7	78.0	117.0
97	43.7	37.7	56.5
132	24.2	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 67.717 ng
RT: 9.234 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:172 Resp: 366420

Ion	Ratio	Lower	Upper
172	100		
171	35.7	27.5	41.3
170	23.9	19.0	28.6







#46
1,1'-Biphenyl
Concen: 35.345 ng
RT: 9.339 min Scan# 1215
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:154 Resp: 192660

Ion	Ratio	Lower	Upper
154	100		
153	39.3	19.7	59.7
76	10.2	0.0	31.1







#47
2-Chloronaphthalene
Concen: 34.908 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:162 Resp: 159410

Ion	Ratio	Lower	Upper
162	100		
127	39.2	31.6	47.4
164	31.6	26.5	39.7







#48
2-Nitroaniline
Concen: 36.621 ng
RT: 9.457 min Scan# 1235
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion: 65 Resp: 49659

Ion	Ratio	Lower	Upper
65	100		
92	62.6	52.3	78.5
138	91.6	78.3	117.5







#49
Acenaphthylene
Concen: 39.474 ng
RT: 9.781 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:152 Resp: 243055

Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.9	23.9
153	12.4	10.4	15.6







#50
Dimethylphthalate
Concen: 34.098 ng
RT: 9.634 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:163 Resp: 170189

Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.4	5.0
164	10.0	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 35.461 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:165 Resp: 34990

Ion	Ratio	Lower	Upper
165	100		
63	65.4	50.4	75.6
89	55.7	41.4	62.2







#52
Acenaphthene
Concen: 35.145 ng
RT: 9.951 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:154 Resp: 144520

Ion	Ratio	Lower	Upper
154	100		
153	111.9	89.9	134.9
152	54.8	42.0	63.0







#53
3-Nitroaniline
Concen: 35.804 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:138 Resp: 39278

Ion	Ratio	Lower	Upper
138	100		
108	10.2	7.9	11.9
92	117.0	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 41.102 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:184 Resp: 21935

Ion	Ratio	Lower	Upper
184	100		
63	67.5	52.1	78.1
154	66.5	47.9	71.9







#55
Dibenzofuran
Concen: 34.140 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:168 Resp: 194900

Ion	Ratio	Lower	Upper
168	100		
139	43.0	33.2	49.8
169	16.0	12.6	19.0







#56
4-Nitrophenol
Concen: 38.254 ng
RT: 10.028 min Scan# 1332
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:139 Resp: 28418

Ion	Ratio	Lower	Upper
139	100		
109	79.1	53.8	93.8
65	110.6	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 37.002 ng
RT: 10.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:165 Resp: 49275

Ion	Ratio	Lower	Upper
165	100		
63	47.3	35.4	53.2
89	63.3	52.7	79.1
182	2.5	2.3	3.5







#58
Fluorene
Concen: 34.864 ng
RT: 10.469 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:166 Resp: 151653

Ion	Ratio	Lower	Upper
166	100		
165	100.7	77.0	115.4
167	13.2	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 39.232 ng
RT: 10.239 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:232 Resp: 49354

Ion	Ratio	Lower	Upper
232	100		
131	36.9	31.9	47.9
130	1.7	1.1	1.7#
166	24.1	21.0	31.4







#60
Diethylphthalate
Concen: 35.442 ng
RT: 10.333 min Scan# 1384
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:149 Resp: 166357

Ion	Ratio	Lower	Upper
149	100		
177	21.8	17.8	26.6
150	12.1	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 35.479 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:204 Resp: 86789

Ion	Ratio	Lower	Upper
204	100		
206	34.4	27.8	41.6
141	56.0	45.4	68.2







#62
4-Nitroaniline
Concen: 33.756 ng
RT: 10.481 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:138 Resp: 35087

Ion	Ratio	Lower	Upper
138	100		
92	56.0	30.5	70.5
108	86.6	56.2	96.2







#63
Azobenzene
Concen: 36.422 ng
RT: 10.616 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion: 77 Resp: 161807

Ion	Ratio	Lower	Upper
77	100		
182	27.4	9.0	49.0
105	20.0	1.4	41.4
51	37.1	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:188 Resp: 127305

Ion	Ratio	Lower	Upper
188	100		
94	6.8	5.4	8.0
80	6.9	6.0	9.0







#65
4,6-Dinitro-2-methylphenol
Concen: 37.867 ng
RT: 10.516 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:198 Resp: 30929

Ion	Ratio	Lower	Upper
198	100		
51	45.8	25.8	65.8
105	33.6	17.2	57.2







#66
n-Nitrosodiphenylamine
Concen: 36.075 ng
RT: 10.575 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:169 Resp: 146733

Ion	Ratio	Lower	Upper
169	100		
168	57.9	47.4	71.2
167	31.4	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 35.313 ng
RT: 10.951 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:248 Resp: 56286

Ion	Ratio	Lower	Upper
248	100		
250	100.6	77.8	116.6
141	66.4	50.5	75.7







#68
Hexachlorobenzene
Concen: 35.903 ng
RT: 11.016 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:284 Resp: 70184

Ion	Ratio	Lower	Upper
284	100		
142	23.9	20.0	30.0
249	28.2	22.2	33.2







#69
Atrazine
Concen: 36.643 ng
RT: 11.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:200 Resp: 48908

Ion	Ratio	Lower	Upper
200	100		
173	25.0	4.9	44.9
215	47.7	28.2	68.2







#70
Pentachlorophenol
Concen: 33.237 ng
RT: 11.216 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:266 Resp: 30927

Ion	Ratio	Lower	Upper
266	100		
268	56.2	48.2	72.4
264	58.3	49.4	74.0







#71
Phenanthrene
Concen: 36.120 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:178 Resp: 228423

Ion	Ratio	Lower	Upper
178	100		
176	19.3	16.3	24.5
179	15.0	12.6	18.8







#72
Anthracene
Concen: 35.782 ng
RT: 11.486 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:178 Resp: 233757

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.0	22.4
179	14.7	11.8	17.6







#73
Carbazole
Concen: 35.250 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:167 Resp: 196006

Ion	Ratio	Lower	Upper
167	100		
166	20.9	16.1	24.1
139	13.3	9.8	14.6







#74
Di-n-butylphthalate
Concen: 35.967 ng
RT: 11.969 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:149 Resp: 245199

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	4.5	3.6	5.4







#75
Fluoranthene
Concen: 34.929 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:202 Resp: 244043

Ion	Ratio	Lower	Upper
202	100		
101	7.8	0.0	27.7
203	18.0	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:240 Resp: 97298

Ion	Ratio	Lower	Upper
240	100		
120	7.0	6.5	9.7
236	27.5	20.1	30.1







#77
Benzidine
Concen: 48.386 ng
RT: 12.745 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:184 Resp: 146518

Ion	Ratio	Lower	Upper
184	100		
185	14.9	11.6	17.4
183	10.1	8.5	12.7







#78
Pyrene
Concen: 35.297 ng
RT: 12.857 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:202 Resp: 246902

Ion	Ratio	Lower	Upper
202	100		
200	21.7	17.2	25.8
203	17.1	13.5	20.3







#79
Terphenyl-d14
Concen: 73.542 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:244 Resp: 409300

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.8	8.6
122	6.1	4.9	7.3







#80
Butylbenzylphthalate
Concen: 35.741 ng
RT: 13.469 min Scan# 1917
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:149 Resp: 94583

Ion	Ratio	Lower	Upper
149	100		
91	72.6	56.6	84.8
206	22.0	18.3	27.5







#81
Benzo(a)anthracene
Concen: 35.620 ng
RT: 14.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:228 Resp: 234218

Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.4	32.2
229	20.3	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 42.643 ng
RT: 14.010 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:252 Resp: 97608

Ion	Ratio	Lower	Upper
252	100		
254	62.7	49.9	74.9
126	7.1	6.5	9.7







#83
Chrysene
Concen: 36.371 ng
RT: 14.080 min Scan# 20
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:228 Resp: 224422

Ion	Ratio	Lower	Upper
228	100		
226	29.8	23.8	35.6
229	18.4	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 36.110 ng
RT: 14.033 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:149 Resp: 130456

Ion	Ratio	Lower	Upper
149	100		
167	26.7	22.5	33.7
279	4.9	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 35.142 ng
RT: 14.645 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:149 Resp: 218335

Ion	Ratio	Lower	Upper
149	100		
167	1.2	1.2	1.8#
43	12.0	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:264 Resp: 122912

Ion	Ratio	Lower	Upper
264	100		
260	21.5	17.0	25.4
265	22.0	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 35.856 ng
RT: 17.039 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:276 Resp: 302575

Ion	Ratio	Lower	Upper
276	100		
138	16.0	13.0	19.6
277	24.4	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 33.715 ng
RT: 15.104 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:252 Resp: 234274

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.2	25.8
125	5.7	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 37.516 ng
RT: 15.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:252 Resp: 240517

Ion	Ratio	Lower	Upper
252	100		
253	21.0	17.4	26.2
125	5.8	4.9	7.3







#90
Benzo(a)pyrene
Concen: 36.773 ng
RT: 15.474 min Scan# 2258
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:252 Resp: 229237

Ion	Ratio	Lower	Upper
252	100		
253	21.1	17.1	25.7
125	7.1	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 36.012 ng
RT: 17.057 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Tgt Ion:278 Resp: 240551

Ion	Ratio	Lower	Upper
278	100		
139	9.8	8.5	12.7
279	22.3	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 36.942 ng
RT: 17.492 min Scan# 2601
Delta R.T. -0.006 min
Lab File: BF144065.D
Acq: 24 Oct 2025 17:22

Tgt Ion:276 Resp: 248094

Ion	Ratio	Lower	Upper
276	100		
277	22.3	17.9	26.9
138	13.5	12.0	18.0



Evaluate Continuing Calibration Report

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	80	0.00
2		1,4-Dioxane	40.000	35.621	10.9	71	0.00
3		Pyridine	40.000	32.576	18.6	63	0.00
4		n-Nitrosodimethylamine	40.000	36.893	7.8	73	0.00
5	S	2-Fluorophenol	80.000	70.283	12.1	68	0.00
6		Aniline	40.000	36.320	9.2	72	0.00
7	S	Phenol-d6	80.000	68.955	13.8	69	0.00
8		2-Chlorophenol	40.000	33.579	16.1	67	0.00
9		Benzaldehyde	40.000	34.852	12.9	73	0.00
10	C	Phenol	40.000	34.853	12.9	69	0.00
11		bis(2-Chloroethyl)ether	40.000	33.851	15.4	68	0.00
12		1,3-Dichlorobenzene	40.000	36.971	7.6	73	0.00
13	C	1,4-Dichlorobenzene	40.000	37.183	7.0	73	0.00
14		1,2-Dichlorobenzene	40.000	35.872	10.3	71	0.00
15		Benzyl Alcohol	40.000	36.330	9.2	72	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	34.517	13.7	68	0.00
17		2-Methylphenol	40.000	35.152	12.1	73	0.00
18		Hexachloroethane	40.000	34.696	13.3	69	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	35.002	12.5	72	0.00
20		3+4-Methylphenols	40.000	36.095	9.8	70	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	78	0.00
22		Acetophenone	40.000	38.453	3.9	74	0.00
23	S	Nitrobenzene-d5	80.000	70.696	11.6	68	0.00
24		Nitrobenzene	40.000	36.753	8.1	70	0.00
25		Isophorone	40.000	36.322	9.2	70	0.00
26	C	2-Nitrophenol	40.000	36.360	9.1	68	0.00
27		2,4-Dimethylphenol	40.000	39.405	1.5	73	0.00
28		bis(2-Chloroethoxy)methane	40.000	35.182	12.0	67	0.00
29	C	2,4-Dichlorophenol	40.000	35.901	10.2	69	0.00
30		1,2,4-Trichlorobenzene	40.000	38.739	3.2	74	0.00
31		Naphthalene	40.000	35.834	10.4	68	0.00
32		Benzoic acid	40.000	44.054	-10.1	102	0.00
33		4-Chloroaniline	40.000	41.428	-3.6	79	0.00
34	C	Hexachlorobutadiene	40.000	38.260	4.4	74	0.00
35		Caprolactam	40.000	35.331	11.7	65	0.00
36	C	4-Chloro-3-methylphenol	40.000	36.112	9.7	70	0.00
37		2-Methylnaphthalene	40.000	33.233	16.9	62	0.00
38		1-Methylnaphthalene	40.000	35.945	10.1	70	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	83	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	35.121	12.2	71	0.00
41	P	Hexachlorocyclopentadiene	40.000	48.585	-21.5	99	0.00
42	S	2,4,6-Tribromophenol	80.000	68.551	14.3	69	0.00
43	C	2,4,6-Trichlorophenol	40.000	33.986	15.0	67	0.00
44		2,4,5-Trichlorophenol	40.000	36.018	10.0	75	0.00
45	S	2-Fluorobiphenyl	80.000	67.717	15.4	69	0.00
46		1,1'-Biphenyl	40.000	35.345	11.6	71	0.00
47		2-Chloronaphthalene	40.000	34.908	12.7	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	36.621	8.4	71	0.00
49		Acenaphthylene	40.000	39.474	1.3	78	0.00
50		Dimethylphthalate	40.000	34.098	14.8	68	0.00
51		2,6-Dinitrotoluene	40.000	35.461	11.3	69	0.00
52	C	Acenaphthene	40.000	35.145	12.1	71	0.00
53		3-Nitroaniline	40.000	35.804	10.5	67	0.00
54	P	2,4-Dinitrophenol	40.000	41.102	-2.8	97	0.00
55		Dibenzofuran	40.000	34.140	14.6	68	0.00
56	P	4-Nitrophenol	40.000	38.254	4.4	83	0.00
57		2,4-Dinitrotoluene	40.000	37.002	7.5	70	0.00
58		Fluorene	40.000	34.864	12.8	70	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	39.232	1.9	75	0.00
60		Diethylphthalate	40.000	35.442	11.4	70	0.00
61		4-Chlorophenyl-phenylether	40.000	35.479	11.3	70	0.00
62		4-Nitroaniline	40.000	33.756	15.6	65	0.00
63		Azobenzene	40.000	36.422	8.9	71	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	80	0.00
65		4,6-Dinitro-2-methylphenol	40.000	37.867	5.3	79	0.00
66	c	n-Nitrosodiphenylamine	40.000	36.075	9.8	72	0.00
67		4-Bromophenyl-phenylether	40.000	35.313	11.7	68	0.00
68		Hexachlorobenzene	40.000	35.903	10.2	68	0.00
69		Atrazine	40.000	36.643	8.4	74	0.00
70	C	Pentachlorophenol	40.000	33.237	16.9	65	0.00
71		Phenanthrene	40.000	36.120	9.7	70	0.00
72		Anthracene	40.000	35.782	10.5	70	0.00
73		Carbazole	40.000	35.250	11.9	68	0.00
74		Di-n-butylphthalate	40.000	35.967	10.1	69	0.00
75	C	Fluoranthene	40.000	34.929	12.7	68	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	79	0.00
77		Benzidine	40.000	48.386	-21.0	89	0.00
78		Pyrene	40.000	35.297	11.8	66	0.00
79	S	Terphenyl-d14	80.000	73.542	8.1	68	0.00
80		Butylbenzylphthalate	40.000	35.741	10.6	64	0.00
81		Benzo(a)anthracene	40.000	35.620	11.0	62	0.00
82		3,3'-Dichlorobenzidine	40.000	42.643	-6.6	77	0.00
83		Chrysene	40.000	36.371	9.1	69	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	36.110	9.7	66	0.00
85	c	Di-n-octyl phthalate	40.000	35.142	12.1	63	0.00
86	I	Perylene-d12	20.000	20.000	0.0	80	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	35.856	10.4	67	0.00
88		Benzo(b)fluoranthene	40.000	33.715	15.7	61	0.00
89		Benzo(k)fluoranthene	40.000	37.516	6.2	74	0.00
90	C	Benzo(a)pyrene	40.000	36.773	8.1	69	0.00
91		Dibenzo(a,h)anthracene	40.000	36.012	10.0	67	0.00
92		Benzo(g,h,i)perylene	40.000	36.942	7.6	68	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2		1,4-Dioxane	0.546	0.486	11.0	71	0.00
3		Pyridine	1.462	1.191	18.5	63	0.00
4		n-Nitrosodimethylamine	0.899	0.829	7.8	73	0.00
5	S	2-Fluorophenol	1.264	1.111	12.1	68	0.00
6		Aniline	1.997	1.814	9.2	72	0.00
7	S	Phenol-d6	1.396	1.203	13.8	69	0.00
8		2-Chlorophenol	1.245	1.045	16.1	67	0.00
9		Benzaldehyde	1.084	0.944	12.9	73	0.00
10	C	Phenol	1.722	1.501	12.8	69	0.00
11		bis(2-Chloroethyl)ether	1.245	1.053	15.4	68	0.00
12		1,3-Dichlorobenzene	1.536	1.419	7.6	73	0.00
13	C	1,4-Dichlorobenzene	1.513	1.407	7.0	73	0.00
14		1,2-Dichlorobenzene	1.464	1.313	10.3	71	0.00
15		Benzyl Alcohol	1.135	1.031	9.2	72	0.00
16		2,2'-oxybis(1-Chloropropane	2.339	2.018	13.7	68	0.00
17		2-Methylphenol	0.957	0.841	12.1	73	0.00
18		Hexachloroethane	0.541	0.469	13.3	69	0.00
19	P	n-Nitroso-di-n-propylamine	0.903	0.790	12.5	72	0.00
20		3+4-Methylphenols	1.132	1.021	9.8	70	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
22		Acetophenone	0.432	0.416	3.7	74	0.00
23	S	Nitrobenzene-d5	0.365	0.323	11.5	68	0.00
24		Nitrobenzene	0.386	0.355	8.0	70	0.00
25		Isophorone	0.638	0.580	9.1	70	0.00
26	C	2-Nitrophenol	0.184	0.167	9.2	68	0.00
27		2,4-Dimethylphenol	0.273	0.269	1.5	73	0.00
28		bis(2-Chloroethoxy)methane	0.398	0.350	12.1	67	0.00
29	C	2,4-Dichlorophenol	0.317	0.285	10.1	69	0.00
30		1,2,4-Trichlorobenzene	0.327	0.316	3.4	74	0.00
31		Naphthalene	0.944	0.846	10.4	68	0.00
32		Benzoic acid	0.173	0.191	-10.4	102	0.00
33		4-Chloroaniline	0.329	0.340	-3.3	79	0.00
34	C	Hexachlorobutadiene	0.267	0.255	4.5	74	0.00
35		Caprolactam	0.080	0.070	12.5	65	0.00
36	C	4-Chloro-3-methylphenol	0.280	0.253	9.6	70	0.00
37		2-Methylnaphthalene	0.622	0.517	16.9	62	0.00
38		1-Methylnaphthalene	0.590	0.530	10.2	70	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
40		1,2,4,5-Tetrachlorobenzene	0.687	0.604	12.1	71	0.00
41	P	Hexachlorocyclopentadiene	0.212	0.258	-21.7	99	0.00
42	S	2,4,6-Tribromophenol	0.255	0.218	14.5	69	0.00
43	C	2,4,6-Trichlorophenol	0.447	0.380	15.0	67	0.00
44		2,4,5-Trichlorophenol	0.469	0.422	10.0	75	0.00
45	S	2-Fluorobiphenyl	1.421	1.203	15.3	69	0.00
46		1,1'-Biphenyl	1.432	1.265	11.7	71	0.00
47		2-Chloronaphthalene	1.199	1.047	12.7	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.356	0.326	8.4	71	0.00
49		Acenaphthylene	1.617	1.596	1.3	78	0.00
50		Dimethylphthalate	1.311	1.118	14.7	68	0.00
51		2,6-Dinitrotoluene	0.259	0.230	11.2	69	0.00
52	C	Acenaphthene	1.080	0.949	12.1	71	0.00
53		3-Nitroaniline	0.288	0.258	10.4	67	0.00
54	P	2,4-Dinitrophenol	0.140	0.144	-2.9	97	0.00
55		Dibenzofuran	1.499	1.280	14.6	68	0.00
56	P	4-Nitrophenol	0.195	0.187	4.1	83	0.00
57		2,4-Dinitrotoluene	0.350	0.324	7.4	70	0.00
58		Fluorene	1.143	0.996	12.9	70	0.00
59		2,3,4,6-Tetrachlorophenol	0.330	0.324	1.8	75	0.00
60		Diethylphthalate	1.233	1.092	11.4	70	0.00
61		4-Chlorophenyl-phenylether	0.643	0.570	11.4	70	0.00
62		4-Nitroaniline	0.273	0.230	15.8	65	0.00
63		Azobenzene	1.167	1.062	9.0	71	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
65		4,6-Dinitro-2-methylphenol	0.128	0.121	5.5	79	0.00
66	c	n-Nitrosodiphenylamine	0.639	0.576	9.9	72	0.00
67		4-Bromophenyl-phenylether	0.250	0.221	11.6	68	0.00
68		Hexachlorobenzene	0.307	0.276	10.1	68	0.00
69		Atrazine	0.210	0.192	8.6	74	0.00
70	C	Pentachlorophenol	0.146	0.121	17.1	65	0.00
71		Phenanthrene	0.994	0.897	9.8	70	0.00
72		Anthracene	1.026	0.918	10.5	70	0.00
73		Carbazole	0.874	0.770	11.9	68	0.00
74		Di-n-butylphthalate	1.071	0.963	10.1	69	0.00
75	C	Fluoranthene	1.098	0.958	12.8	68	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	79	0.00
77		Benzidine	0.622	0.753	-21.1	89	0.00
78		Pyrene	1.438	1.269	11.8	66	0.00
79	S	Terphenyl-d14	1.144	1.052	8.0	68	0.00
80		Butylbenzylphthalate	0.544	0.486	10.7	64	0.00
81		Benzo(a)anthracene	1.352	1.204	10.9	62	0.00
82		3,3'-Dichlorobenzidine	0.470	0.502	-6.8	77	0.00
83		Chrysene	1.268	1.153	9.1	69	0.00
84		Bis(2-ethylhexyl)phthalate	0.743	0.670	9.8	66	0.00
85	c	Di-n-octyl phthalate	1.277	1.122	12.1	63	0.00
86	I	Perylene-d12	1.000	1.000	0.0	80	0.00
87		Indeno(1,2,3-cd)pyrene	1.373	1.231	10.3	67	0.00
88		Benzo(b)fluoranthene	1.131	0.953	15.7	61	0.00
89		Benzo(k)fluoranthene	1.043	0.978	6.2	74	0.00
90	C	Benzo(a)pyrene	1.014	0.933	8.0	69	0.00
91		Dibenzo(a,h)anthracene	1.087	0.979	9.9	67	0.00
92		Benzo(g,h,i)perylene	1.093	1.009	7.7	68	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144065.D
Acq On : 24 Oct 2025 17:22
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102425

Quant Time: Oct 24 17:42:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Response Factor Report GC/MS Ins

Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
Method File : 8270-BG102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Nov 03 18:16:26 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BG064569.D 5 =BG064570.D 10 =BG064571.D 20 =BG064572.D 40 =BG064573.D 50 =BG064576.D 60 =BG064574.D 80 =BG0645
75.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------ISTD----------									
2)		1,4-Dioxane		0.570	0.581	0.524	0.525	0.474	0.559	0.460	0.528	8.85
3)		Pyridine		1.630	1.608	1.552	1.669	1.471	1.733	1.427	1.584	6.84
4)		n-Nitrosodimet...			0.912	0.807	0.900	0.780	0.921	0.776	0.850	8.08
5)	S	2-Fluorophenol		1.098	1.207	1.155	1.292	1.139	1.318	1.132	1.192	7.09
6)		Aniline		2.203	2.320	2.131	2.386	2.090	2.440	2.004	2.225	7.28
7)	S	Phenol-d6		1.469	1.654	1.566	1.771	1.597	1.864	1.524	1.635	8.55
8)		2-Chlorophenol		1.051	1.207	1.197	1.373	1.225	1.461	1.216	1.247	10.65
9)		Benzaldehyde			0.942	0.994	0.940	0.836	0.981	0.662	0.893	14.09
10)	C	Phenol		1.665	1.820	1.748	1.944	1.761	2.027	1.693	1.808	7.36
11)		bis(2-Chloroet...		1.307	1.409	1.292	1.383	1.273	1.443	1.198	1.329	6.47
12)		1,3-Dichlorobe...		1.435	1.516	1.394	1.520	1.342	1.542	1.299	1.435	6.62
13)	C	1,4-Dichlorobe...		1.431	1.546	1.442	1.534	1.381	1.561	1.323	1.460	6.21
14)		1,2-Dichlorobe...		1.371	1.518	1.364	1.483	1.316	1.520	1.283	1.408	6.97
15)		Benzyl Alcohol			1.281	1.263	1.425	1.306	1.533	1.283	1.348	7.97
16)		2,2'-oxybis(1-...		3.421	3.604	3.396	3.733	3.301	3.802	3.154	3.487	6.74
17)		2-Methylphenol		1.103	1.225	1.178	1.273	1.159	1.345	1.120	1.201	7.22
18)		Hexachloroethane		0.468	0.535	0.505	0.541	0.496	0.557	0.483	0.512	6.41
19)	P	n-Nitroso-di-n...	1.044	1.038	1.141	1.129	1.227	1.115	1.289	1.057	1.130	7.93
20)		3+4-Methylphenols			1.652	1.581	1.818	1.612	1.895	1.571	1.688	8.04
21)	I	Naphthalene-d8	----------ISTD----------									
22)		Acetophenone		0.535	0.565	0.528	0.584	0.516	0.590	0.509	0.547	6.04
23)	S	Nitrobenzene-d5		0.239	0.286	0.314	0.377	0.356	0.403	0.356	0.333	17.02
24)		Nitrobenzene		0.274	0.312	0.344	0.410	0.377	0.438	0.380	0.362	15.64
25)		Isophorone		0.751	0.796	0.771	0.869	0.764	0.891	0.751	0.799	7.25
26)	C	2-Nitrophenol		0.076	0.095	0.110	0.137	0.135	0.152	0.144	0.121	23.20
27)		2,4-Dimethylph...		0.236	0.294	0.285	0.335	0.297	0.343	0.290	0.297	11.93
28)		bis(2-Chloroet...		0.452	0.464	0.443	0.490	0.426	0.493	0.417	0.455	6.48
29)	C	2,4-Dichloroph...		0.243	0.299	0.315	0.360	0.327	0.373	0.322	0.320	13.34
30)		1,2,4-Trichlor...		0.338	0.369	0.348	0.391	0.349	0.399	0.349	0.363	6.44
31)		Naphthalene		1.083	1.144	1.084	1.177	1.036	1.201	1.037	1.109	5.96
32)		Benzoic acid			0.142	0.164	0.223	0.215	0.260	0.237	0.207	21.74
33)		4-Chloroaniline		0.390	0.456	0.411	0.459	0.410	0.486	0.405	0.431	8.29
34)	C	Hexachlorobuta...		0.266	0.291	0.282	0.299	0.271	0.306	0.273	0.284	5.33
35)		Caprolactam			0.112	0.114	0.134	0.118	0.142	0.116	0.123	9.85
36)	C	4-Chloro-3-met...		0.323	0.381	0.361	0.426	0.376	0.454	0.370	0.384	11.18
37)		2-Methylnaphth...		0.764	0.858	0.797	0.865	0.772	0.888	0.755	0.814	6.74
38)		1-Methylnaphth...		0.798	0.849	0.776	0.853	0.752	0.878	0.735	0.806	6.85


Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
Method File : 8270-BG102425.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.657	0.684	0.648	0.702	0.633	0.717	0.657	0.671	4.55
41)	P	Hexachlorocycl...		0.268	0.279	0.324	0.297	0.340	0.305	0.302	8.97
42)	S	2,4,6-Tribromo...	0.221	0.266	0.272	0.318	0.292	0.346	0.306	0.289	14.04
43)	C	2,4,6-Trichlor...	0.294	0.333	0.375	0.423	0.399	0.456	0.408	0.384	14.41
44)		2,4,5-Trichlor...	0.339	0.425	0.439	0.484	0.443	0.517	0.454	0.443	12.48
45)	S	2-Fluorobiphenyl	1.310	1.389	1.306	1.384	1.235	1.384	1.228	1.319	5.24
46)		1,1'-Biphenyl	1.456	1.450	1.401	1.489	1.326	1.498	1.333	1.422	4.98
47)		2-Chloronaphth...	1.065	1.082	1.053	1.132	1.013	1.145	1.023	1.073	4.71
48)		2-Nitroaniline	0.201	0.234	0.277	0.348	0.343	0.404	0.367	0.311	24.06
49)		Acenaphthylene	1.667	1.726	1.647	1.786	1.593	1.827	1.604	1.693	5.30
50)		Dimethylphthalate	1.389	1.570	1.488	1.641	1.462	1.699	1.462	1.530	7.21
51)		2,6-Dinitrotol...	0.127	0.180	0.193	0.247	0.244	0.283	0.254	0.218	24.69
52)	C	Acenaphthene	1.126	1.178	1.115	1.223	1.093	1.261	1.101	1.157	5.63
53)		3-Nitroaniline	0.180	0.211	0.252	0.294	0.281	0.325	0.297	0.263	19.70
54)	P	2,4-Dinitrophenol		0.086	0.110	0.137	0.139	0.165	0.154	0.132	22.11
55)		Dibenzofuran	1.881	1.972	1.826	1.963	1.734	1.988	1.725	1.870	5.96
56)	P	4-Nitrophenol		0.191	0.213	0.261	0.250	0.283	0.267	0.244	14.40
57)		2,4-Dinitrotol...	0.223	0.269	0.300	0.385	0.373	0.434	0.396	0.340	22.67
58)		Fluorene	1.471	1.559	1.426	1.531	1.356	1.549	1.329	1.460	6.38
59)		2,3,4,6-Tetrac...	0.312	0.385	0.402	0.466	0.444	0.508	0.449	0.424	15.04
60)		Diethylphthalate	1.417	1.645	1.514	1.657	1.486	1.718	1.502	1.563	7.06
61)		4-Chlorophenyl...	0.800	0.864	0.792	0.866	0.763	0.888	0.763	0.820	6.38
62)		4-Nitroaniline	0.194	0.252	0.286	0.327	0.310	0.355	0.324	0.293	18.55
63)		Azobenzene	1.362	1.504	1.377	1.538	1.338	1.564	1.353	1.434	6.80
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.055	0.064	0.083	0.083	0.097	0.093	0.079	21.00
66)	c	n-Nitrosodiphe...	0.510	0.551	0.525	0.583	0.518	0.594	0.500	0.540	6.79
67)		4-Bromophenyl-...	0.218	0.236	0.230	0.258	0.229	0.264	0.224	0.237	7.28
68)		Hexachlorobenzene	0.254	0.266	0.262	0.292	0.260	0.302	0.256	0.270	6.99
69)		Atrazine	0.215	0.227	0.249	0.245	0.219	0.261	0.211	0.232	8.27
70)	C	Pentachlorophenol		0.132	0.152	0.180	0.172	0.199	0.177	0.169	13.91
71)		Phenanthrene	1.084	1.133	1.073	1.162	1.020	1.179	1.005	1.094	6.17
72)		Anthracene	1.087	1.144	1.101	1.187	1.041	1.204	1.025	1.113	6.17
73)		Carbazole	0.982	1.037	0.982	1.031	0.910	1.028	0.897	0.981	5.90
74)		Di-n-butylphth...	1.011	1.156	1.179	1.240	1.084	1.198	1.075	1.134	7.11
75)	C	Fluoranthene	1.385	1.463	1.380	1.440	1.273	1.403	1.247	1.370	5.91
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine		0.392	0.475	0.490	0.416	0.533	0.383	0.448	13.40
78)		Pyrene	1.287	1.331	1.284	1.411	1.233	1.406	1.196	1.307	6.24
79)	S	Terphenyl-d14	1.074	1.120	1.066	1.142	0.987	1.102	0.928	1.060	7.21
80)		Butylbenzylpht...	0.284	0.359	0.388	0.450	0.408	0.467	0.410	0.395	15.43
81)		Benzo(a)anthra...	1.294	1.355	1.292	1.416	1.253	1.430	1.243	1.326	5.71
82)		3,3'-Dichlorob...		0.418	0.428	0.477	0.424	0.486	0.424	0.443	6.88
83)		Chrysene	1.249	1.291	1.229	1.344	1.188	1.359	1.173	1.262	5.76
84)		Bis(2-ethylhex...	0.484	0.555	0.581	0.671	0.585	0.667	0.588	0.590	10.95
85)	c	Di-n-octyl pht...		0.871	0.932	1.087	0.976	1.117	0.986	0.995	9.32


Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
Method File : 8270-BG102425.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.383	1.491	1.410	1.568	1.396	1.642	1.410	1.471	6.80
88)		Benzo(b)fluora...	1.138	1.246	1.169	1.322	1.173	1.353	1.183	1.226	6.77
89)		Benzo(k)fluora...	1.175	1.260	1.206	1.285	1.151	1.364	1.149	1.227	6.51
90)	C	Benzo(a)pyrene	1.061	1.144	1.093	1.202	1.067	1.246	1.079	1.127	6.42
91)		Dibenzo(a,h)an...	1.120	1.223	1.158	1.268	1.130	1.324	1.143	1.195	6.55
92)		Benzo(g,h,i)pe...	1.080	1.162	1.089	1.213	1.078	1.280	1.092	1.142	6.95


(#) = Out of Range




Response = 1.495e-001 * Amt - 2.340e-002
Coef of Det (r^2) = 0.995585 Curve Fit: wlr(1/a)
Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Calibration Table Last Updated: Fri Oct 24 16:40:26 2025




Response = 3.863e-001 * Amt - 6.228e-002
Coef of Det (r^2) = 0.991893 Curve Fit: wlr(1/a)
Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Calibration Table Last Updated: Fri Oct 24 16:40:26 2025




Response = 2.705e-001 * Amt - 4.319e-002
Coef of Det (r^2) = 0.992850 Curve Fit: wlr(1/a)
Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Calibration Table Last Updated: Fri Oct 24 16:40:26 2025



Response = 1.666e-001 * Amt - 4.640e-002
Coef of Det (r^2) = 0.992196 Curve Fit: wlr(1/a)
Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Calibration Table Last Updated: Fri Oct 24 16:40:26 2025



Response = 4.169e-001 * Amt - 6.360e-002
Coef of Det (r^2) = 0.992833 Curve Fit: wlr(1/a)
Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Calibration Table Last Updated: Fri Oct 24 16:40:26 2025



Response = 9.922e-002 * Amt - 2.692e-002
Coef of Det (r^2) = 0.993360 Curve Fit: wlr(1/a)
Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Calibration Table Last Updated: Fri Oct 24 16:40:26 2025

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064569.D
Acq On : 24 Oct 2025 9:00
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Quant Time: Nov 03 17:57:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.223	152	43718	20.000	ng	0.00
21) Naphthalene-d8	11.049	136	174680	20.000	ng	0.00
39) Acenaphthene-d10	14.839	164	136814	20.000	ng	0.00
64) Phenanthrene-d10	17.577	188	341404	20.000	ng	0.00
76) Chrysene-d12	21.854	240	378586	20.000	ng	0.00
86) Perylene-d12	25.203	264	398350	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	9.016	70	5707	2.310	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064569.D
Acq On : 24 Oct 2025 9:00
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Quant Time: Nov 03 17:57:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.223 min Scan# 8
Delta R.T. -0.000 min
Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 43718

Ion	Ratio	Lower	Upper
152	100		
150	151.2	122.2	183.4
115	59.5	50.8	76.2







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.75 min

Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	75.9
63	41.7






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 7.35 min


Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99

Sig	Exp Ratio
99	100
42	22.9
71	35.8








#19
n-Nitroso-di-n-propylamine
Concen: 2.310 ng
RT: 9.016 min Scan# 1009
Delta R.T. -0.012 min
Lab File: BG064569.D
Acq: 24 Oct 2025 9:00


Tgt Ion: 70 Resp: 5707

Ion	Ratio	Lower	Upper
70	100		
42	73.3	65.2	97.8
101	9.1	7.9	11.9
130	10.8	14.3	21.5#











#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.049 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 174680

Ion	Ratio	Lower	Upper
136	100		
137	11.9	9.2	13.8
54	12.7	12.6	19.0
68	6.7	5.7	8.5







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 9.39 min

Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	39.1
54	70.7







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.839 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 136814

Ion	Ratio	Lower	Upper
164	100		
162	97.0	77.0	115.6
160	45.0	34.4	51.6







#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 16.32 min


Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	97.5
141	27.2








#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 13.48 min


Lab File: BG064569.D
Acq: 24 Oct 2025 9:00


Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172

Sig	Exp Ratio
172	100
171	36.7
170	23.9









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.577 min Scan# 2466
Delta R.T. -0.006 min
Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Tgt Ion:188 Resp: 341404

Ion	Ratio	Lower	Upper
188	100		
94	8.1	5.7	8.5
80	8.9	6.8	10.2







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.854 min Scan# 31
Delta R.T. -0.006 min
Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 378586

Ion	Ratio	Lower	Upper
240	100		
120	7.6	6.5	9.7
236	25.4	20.2	30.2







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 20.17 min

Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	7.6
122	7.0






#86
Perylene-d12
Concen: 20.000 ng
RT: 25.203 min Scan# 3
Delta R.T. 0.006 min
Lab File: BG064569.D
Acq: 24 Oct 2025 9:00

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 398350

Ion	Ratio	Lower	Upper
264	100		
260	23.3	19.0	28.4
265	21.3	17.6	26.4



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064570.D
Acq On : 24 Oct 2025 9:41
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Nov 03 17:58:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.219	152	40837	20.000	ng		0.00
21) Naphthalene-d8	11.045	136	168287	20.000	ng	#	0.00
39) Acenaphthene-d10	14.840	164	131387	20.000	ng		0.00
64) Phenanthrene-d10	17.578	188	331372	20.000	ng		0.00
76) Chrysene-d12	21.856	240	381389	20.000	ng		0.00
86) Perylene-d12	25.199	264	401350	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.745	112	22421	9.216	ng		0.00
7) Phenol-d6	7.349	99	30003	8.987	ng		0.00
23) Nitrobenzene-d5	9.382	82	20097	7.173	ng		0.00
42) 2,4,6-Tribromophenol	16.315	330	14514	7.654	ng		0.00
45) 2-Fluorobiphenyl	13.471	172	86089	9.932	ng		0.00
79) Terphenyl-d14	20.169	244	204719	10.130	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.589	88	5823	5.406	ng		95
3) Pyridine	4.000	79	16637	5.143	ng		89
6) Aniline	7.525	93	22494	4.952	ng		98
8) 2-Chlorophenol	7.778	128	10732	4.214	ng		96
10) Phenol	7.373	94	16995	4.603	ng		96
11) bis(2-Chloroethyl)ether	7.619	93	13341	4.915	ng		95
12) 1,3-Dichlorobenzene	8.113	146	14650	4.998	ng		93
13) 1,4-Dichlorobenzene	8.260	146	14610m	4.902	ng		
14) 1,2-Dichlorobenzene	8.583	146	13994	4.868	ng		96
16) 2,2'-oxybis(1-Chloropr...	8.736	45	34925	4.905	ng		97
17) 2-Methylphenol	8.648	107	11259	4.593	ng		97
18) Hexachloroethane	9.329	117	4773	4.564	ng		97
19) n-Nitroso-di-n-propyla...	9.024	70	10596	4.592	ng		92
22) Acetophenone	9.035	105	22505	4.892	ng	#	92
24) Nitrobenzene	9.423	77	11526	3.782	ng	#	95
25) Isophorone	9.946	82	31609	4.701	ng	#	93
26) 2-Nitrophenol	10.140	139	3186	5.664	ng		99
27) 2,4-Dimethylphenol	10.193	122	9925	3.969	ng		92
28) bis(2-Chloroethoxy)met...	10.434	93	18999	4.963	ng		95
29) 2,4-Dichlorophenol	10.680	162	10223	3.798	ng		89
30) 1,2,4-Trichlorobenzene	10.904	180	14238	4.656	ng		97
31) Naphthalene	11.098	128	45559	4.883	ng		98
33) 4-Chloroaniline	11.186	127	16429	4.530	ng		93
34) Hexachlorobutadiene	11.391	225	11212	4.688	ng		95
36) 4-Chloro-3-methylphenol	12.290	107	13604	4.205	ng		94
37) 2-Methylnaphthalene	12.690	142	32129	4.690	ng		97
38) 1-Methylnaphthalene	12.901	142	33565m	4.951	ng		
40) 1,2,4,5-Tetrachloroben...	13.048	216	21576	4.893	ng		99
43) 2,4,6-Trichlorophenol	13.277	196	9650	3.826	ng		93
44) 2,4,5-Trichlorophenol	13.342	196	11151	3.831	ng	#	91
46) 1,1'-Biphenyl	13.683	154	47840	5.122	ng		95
47) 2-Chloronaphthalene	13.724	162	34977	4.961	ng		96
48) 2-Nitroaniline	13.906	65	6601	5.826	ng		87
49) Acenaphthylene	14.564	152	54748	4.923	ng		98
50) Dimethylphthalate	14.282	163	45639	4.540	ng		98
51) 2,6-Dinitrotoluene	14.394	165	4161	5.535	ng	#	91

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064570.D
Acq On : 24 Oct 2025 9:41
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 17:58:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
52) Acenaphthene	14.905	154	36982	4.867	ng		96
53) 3-Nitroaniline	14.717	138	5903	3.417	ng	#	96
55) Dibenzofuran	15.234	168	61779	5.029	ng		97
57) 2,4-Dinitrotoluene	15.175	165	7312	5.721	ng	#	80
58) Fluorene	15.880	166	48305	5.035	ng		98
59) 2,3,4,6-Tetrachlorophenol	15.451	232	10241	3.679	ng	#	95
60) Diethylphthalate	15.639	149	46534	4.533	ng		98
61) 4-Chlorophenyl-phenyle...	15.874	204	26270	4.880	ng		97
62) 4-Nitroaniline	15.869	138	6383	3.321	ng		90
63) Azobenzene	16.162	77	44729	4.749	ng		98
66) n-Nitrosodiphenylamine	16.080	169	42291	4.725	ng		98
67) 4-Bromophenyl-phenylether	16.762	248	18064	4.600	ng		95
68) Hexachlorobenzene	16.885	284	21076	4.705	ng		95
69) Atrazine	17.014	200	17783	4.616	ng		94
71) Phenanthrene	17.619	178	89771	4.954	ng		99
72) Anthracene	17.713	178	90063	4.886	ng		99
73) Carbazole	17.966	167	81350	5.005	ng		98
74) Di-n-butylphthalate	18.530	149	83713	4.454	ng		99
75) Fluoranthene	19.617	202	114724	5.054	ng		98
78) Pyrene	19.975	202	122676	4.923	ng		98
80) Butylbenzylphthalate	20.857	149	27072	3.595	ng		94
81) Benzo(a)anthracene	21.832	228	123345	4.878	ng		98
83) Chrysene	21.903	228	119080	4.949	ng		97
84) Bis(2-ethylhexyl)phtha...	21.750	149	46117	4.098	ng		98
87) Indeno(1,2,3-cd)pyrene	29.030	276	138758	4.700	ng	#	88
88) Benzo(b)fluoranthene	24.129	252	114203	4.640	ng		98
89) Benzo(k)fluoranthene	24.200	252	117893	4.787	ng		99
90) Benzo(a)pyrene	25.040	252	106414	4.704	ng		99
91) Dibenzo(a,h)anthracene	29.100	278	112398	4.686	ng		96
92) Benzo(g,h,i)perylene	30.216	276	108407	4.731	ng		97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064570.D
Acq On : 24 Oct 2025 9:41
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1


Quant Time: Nov 03 17:58:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.219 min Scan# 8
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:152 Resp: 4083


Ion	Ratio	Lower	Upper
152	100		
150	153.0	122.2	183.4
115	62.5	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 5.406 ng
RT: 3.589 min Scan# 85
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 88 Resp: 5823

Ion	Ratio	Lower	Upper
88	100		
58	108.0	92.4	138.6
43	41.6	33.0	49.6






#3
Pyridine
Concen: 5.143 ng
RT: 4.000 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 79 Resp: 1663


Ion	Ratio	Lower	Upper
79	100		
52	106.6	94.2	141.2
51	48.4	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#5
2-Fluorophenol
Concen: 9.216 ng
RT: 5.745 min Scan# 452
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:112 Resp: 22421

Ion	Ratio	Lower	Upper
112	100		
64	68.9	60.7	91.1
63	35.8	33.4	50.0






#6
Aniline
Concen: 4.952 ng
RT: 7.525 min Scan# 7
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 93 Resp: 2249


Ion	Ratio	Lower	Upper
93	100		
66	44.4	37.0	55.6
65	22.0	18.3	27.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#7
Phenol-d6
Concen: 8.987 ng
RT: 7.349 min Scan# 725
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 99 Resp: 30003

Ion	Ratio	Lower	Upper
99	100		
42	24.5	18.3	27.5
71	35.1	28.6	43.0






#8
2-Chlorophenol
Concen: 4.214 ng
RT: 7.778 min Scan# 79
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:128 Resp: 1073


Ion	Ratio	Lower	Upper
128	100		
130	39.4	13.5	53.5
64	65.2	45.2	85.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 4.603 ng
RT: 7.373 min Scan# 729
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 94 Resp: 16995

Ion	Ratio	Lower	Upper
94	100		
65	38.8	15.8	55.8
66	45.9	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 4.915 ng
RT: 7.619 min Scan# 7
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 93 Resp: 1334


Ion	Ratio	Lower	Upper
93	100		
63	117.0	91.3	131.3
95	32.3	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 4.998 ng
RT: 8.113 min Scan# 855
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:146 Resp: 14650

Ion	Ratio	Lower	Upper
146	100		
148	64.6	55.3	82.9
75	28.7	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 4.902 ng m
RT: 8.260 min Scan# 8
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:146 Resp: 1461


Ion	Ratio	Lower	Upper
146	100		
148	54.9	56.2	84.4
111	38.6	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 4.868 ng
RT: 8.583 min Scan# 935
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:146 Resp: 13994

Ion	Ratio	Lower	Upper
146	100		
148	61.9	51.5	77.3
111	40.0	35.0	52.6





#16
2,2'-oxybis(1-Chloropropane)
Concen: 4.905 ng
RT: 8.736 min Scan# 9
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41


Tgt Ion: 45 Resp: 3492


Ion	Ratio	Lower	Upper
45	100		
77	7.1	0.0	28.5
79	5.0	0.0	25.6

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#17
2-Methylphenol
Concen: 4.593 ng
RT: 8.648 min Scan# 946
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:107 Resp: 11259

Ion	Ratio	Lower	Upper
107	100		
108	111.6	90.0	135.0
77	46.6	37.8	56.8
79	43.4	40.5	60.7






#18
Hexachloroethane
Concen: 4.564 ng
RT: 9.329 min Scan# 10
Delta R.T. 0.008 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:117 Resp: 477


Ion	Ratio	Lower	Upper
117	100		
119	88.0	75.4	113.2
201	124.2	99.2	148.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#19
n-Nitroso-di-n-propylamine
Concen: 4.592 ng
RT: 9.024 min Scan# 1010
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 70 Resp: 10596

Ion	Ratio	Lower	Upper
70	100		
42	74.1	65.2	97.8
101	10.4	7.9	11.9
130	14.5	14.3	21.5






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.045 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:136 Resp: 16828


Ion	Ratio	Lower	Upper
136	100		
137	10.4	9.2	13.8
54	15.4	12.6	19.0
68	5.6	5.7	8.5#



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 4.892 ng
RT: 9.035 min Scan# 1012
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:105 Resp: 22505

Ion	Ratio	Lower	Upper
105	100		
71	2.8	1.5	2.3#
51	38.4	33.7	50.5
120	14.6	16.1	24.1#






#23
Nitrobenzene-d5
Concen: 7.173 ng
RT: 9.382 min Scan# 10
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 82 Resp: 2009


Ion	Ratio	Lower	Upper
82	100		
128	39.6	31.3	46.9
54	71.0	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 3.782 ng
RT: 9.423 min Scan# 1078
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 77 Resp: 11526

Ion	Ratio	Lower	Upper
77	100		
123	36.6	31.0	46.6
65	12.6	12.7	19.1#






#25
Isophorone
Concen: 4.701 ng
RT: 9.946 min Scan# 11
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 82 Resp: 3160


Ion	Ratio	Lower	Upper
82	100		
95	7.2	6.1	9.1
138	12.3	13.0	19.4#



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 5.664 ng
RT: 10.140 min Scan# 1200
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:139 Resp: 3186

Ion	Ratio	Lower	Upper
139	100		
109	32.1	24.6	37.0
65	58.9	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 3.969 ng
RT: 10.193 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:122 Resp: 992


Ion	Ratio	Lower	Upper
122	100		
107	121.0	103.0	154.4
121	51.7	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 4.963 ng
RT: 10.434 min Scan# 1250
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 93 Resp: 18999

Ion	Ratio	Lower	Upper
93	100		
95	35.3	25.4	38.2
123	11.5	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 3.798 ng
RT: 10.680 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:162 Resp: 1022


Ion	Ratio	Lower	Upper
162	100		
164	79.1	48.7	88.7
98	32.4	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 4.656 ng
RT: 10.904 min Scan# 1330
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:180 Resp: 14238

Ion	Ratio	Lower	Upper
180	100		
182	92.9	76.9	115.3
145	25.7	22.1	33.1






#31
Naphthalene
Concen: 4.883 ng
RT: 11.098 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:128 Resp: 4555


Ion	Ratio	Lower	Upper
128	100		
129	10.4	8.8	13.2
127	14.3	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#33
4-Chloroaniline
Concen: 4.530 ng
RT: 11.186 min Scan# 1378
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:127 Resp: 16429

Ion	Ratio	Lower	Upper
127	100		
129	29.8	25.5	38.3
65	33.5	32.2	48.4
92	25.1	18.2	27.2






#34
Hexachlorobutadiene
Concen: 4.688 ng
RT: 11.391 min Scan# 14
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:225 Resp: 1121


Ion	Ratio	Lower	Upper
225	100		
223	66.4	48.8	73.2
227	63.9	49.5	74.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 4.205 ng
RT: 12.290 min Scan# 1566
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:107 Resp: 13604

Ion	Ratio	Lower	Upper
107	100		
144	27.6	19.1	28.7
142	68.9	58.6	87.8






#37
2-Methylnaphthalene
Concen: 4.690 ng
RT: 12.690 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:142 Resp: 3212


Ion	Ratio	Lower	Upper
142	100		
141	89.3	68.9	103.3
115	39.3	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 4.951 ng m
RT: 12.901 min Scan# 1670
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:142 Resp: 33565

Ion	Ratio	Lower	Upper
142	100		
141	89.1	73.3	109.9
116	5.4	3.1	4.7#






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.840 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:164 Resp: 13138


Ion	Ratio	Lower	Upper
164	100		
162	103.0	77.0	115.6
160	48.2	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.893 ng
RT: 13.048 min Scan# 1695
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:216 Resp: 21576

Ion	Ratio	Lower	Upper
216	100		
214	78.0	61.4	92.2
179	19.5	15.4	23.0
108	15.8	11.8	17.6






#42
2,4,6-Tribromophenol
Concen: 7.654 ng
RT: 16.315 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:330 Resp: 1451


Ion	Ratio	Lower	Upper
330	100		
332	95.3	78.0	117.0
141	25.5	21.8	32.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#43
2,4,6-Trichlorophenol
Concen: 3.826 ng
RT: 13.277 min Scan# 1734
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:196 Resp: 9650

Ion	Ratio	Lower	Upper
196	100		
198	103.9	76.6	115.0
200	33.5	26.0	39.0






#44
2,4,5-Trichlorophenol
Concen: 3.831 ng
RT: 13.342 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:196 Resp: 1115


Ion	Ratio	Lower	Upper
196	100		
198	88.6	74.2	111.4
97	35.1	37.0	55.6#
132	18.5	18.5	27.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#45
2-Fluorobiphenyl
Concen: 9.932 ng
RT: 13.471 min Scan# 1767
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:172 Resp: 86089

Ion	Ratio	Lower	Upper
172	100		
171	35.1	29.4	44.0
170	24.3	19.1	28.7






#46
1,1'-Biphenyl
Concen: 5.122 ng
RT: 13.683 min Scan# 18
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:154 Resp: 4784


Ion	Ratio	Lower	Upper
154	100		
153	40.1	22.4	62.4
76	10.3	0.0	33.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#47
2-Chloronaphthalene
Concen: 4.961 ng
RT: 13.724 min Scan# 1810
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:162 Resp: 34977

Ion	Ratio	Lower	Upper
162	100		
127	44.2	32.2	48.2
164	34.7	27.5	41.3






#48
2-Nitroaniline
Concen: 5.826 ng
RT: 13.906 min Scan# 18
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 65 Resp: 6601


Ion	Ratio	Lower	Upper
65	100		
92	52.8	48.3	72.5
138	79.6	74.9	112.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#49
Acenaphthylene
Concen: 4.923 ng
RT: 14.564 min Scan# 1953
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:152 Resp: 54748

Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.4	23.2
153	12.0	11.0	16.4





#50
Dimethylphthalate
Concen: 4.540 ng
RT: 14.282 min Scan# 19
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41


Tgt Ion:163 Resp: 4563


Ion	Ratio	Lower	Upper
163	100		
194	3.1	2.8	4.2
164	10.5	7.7	11.5

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#51
2,6-Dinitrotoluene
Concen: 5.535 ng
RT: 14.394 min Scan# 1924
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:165 Resp: 4161

Ion	Ratio	Lower	Upper
165	100		
63	70.7	57.4	86.0
89	73.2	47.8	71.8#






#52
Acenaphthene
Concen: 4.867 ng
RT: 14.905 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:154 Resp: 3698


Ion	Ratio	Lower	Upper
154	100		
153	111.0	92.2	138.4
152	52.2	43.2	64.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#53
3-Nitroaniline
Concen: 3.417 ng
RT: 14.717 min Scan# 1979
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:138 Resp: 5903

Ion	Ratio	Lower	Upper
138	100		
108	27.2	11.4	17.0#
92	120.0	95.3	142.9






#55
Dibenzofuran
Concen: 5.029 ng
RT: 15.234 min Scan# 20
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:168 Resp: 61779


Ion	Ratio	Lower	Upper
168	100		
139	32.7	27.8	41.8
169	13.7	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#57
2,4-Dinitrotoluene
Concen: 5.721 ng
RT: 15.175 min Scan# 2057
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:165 Resp: 7312

Ion	Ratio	Lower	Upper
165	100		
63	35.1	41.8	62.8#
89	57.9	58.1	87.1#
182	3.3	1.4	2.0#






#58
Fluorene
Concen: 5.035 ng
RT: 15.880 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:166 Resp: 4830


Ion	Ratio	Lower	Upper
166	100		
165	103.5	81.5	122.3
167	11.6	11.4	17.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#59
2,3,4,6-Tetrachlorophenol
Concen: 3.679 ng
RT: 15.451 min Scan# 2104
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:232 Resp: 10241

Ion	Ratio	Lower	Upper
232	100		
131	29.7	26.7	40.1
130	1.1	1.6	2.4#
166	24.3	20.5	30.7






#60
Diethylphthalate
Concen: 4.533 ng
RT: 15.639 min Scan# 21
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:149 Resp: 4653


Ion	Ratio	Lower	Upper
149	100		
177	23.1	18.0	27.0
150	13.9	10.2	15.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#61
4-Chlorophenyl-phenylether
Concen: 4.880 ng
RT: 15.874 min Scan# 2176
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:204 Resp: 26270

Ion	Ratio	Lower	Upper
204	100		
206	32.3	25.6	38.4
141	48.0	40.6	60.8






#62
4-Nitroaniline
Concen: 3.321 ng
RT: 15.869 min Scan# 21
Delta R.T. -0.016 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:138 Resp: 638


Ion	Ratio	Lower	Upper
138	100		
92	55.4	36.1	76.1
108	80.1	75.6	115.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#63
Azobenzene
Concen: 4.749 ng
RT: 16.162 min Scan# 2225
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion: 77 Resp: 44729

Ion	Ratio	Lower	Upper
77	100		
182	25.7	4.9	44.9
105	21.4	1.3	41.3
51	37.7	19.0	59.0






#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.578 min Scan# 24
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:188 Resp: 331371


Ion	Ratio	Lower	Upper
188	100		
94	8.3	5.7	8.5
80	9.4	6.8	10.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#66
n-Nitrosodiphenylamine
Concen: 4.725 ng
RT: 16.080 min Scan# 2211
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:169 Resp: 42291

Ion	Ratio	Lower	Upper
169	100		
168	71.9	56.3	84.5
167	38.5	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 4.600 ng
RT: 16.762 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:248 Resp: 1806


Ion	Ratio	Lower	Upper
248	100		
250	97.5	77.7	116.5
141	69.6	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 4.705 ng
RT: 16.885 min Scan# 2348
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:284 Resp: 21076

Ion	Ratio	Lower	Upper
284	100		
142	28.2	21.0	31.6
249	34.7	24.9	37.3






#69
Atrazine
Concen: 4.616 ng
RT: 17.014 min Scan# 2
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:200 Resp: 1778


Ion	Ratio	Lower	Upper
200	100		
173	26.6	0.9	40.9
215	44.8	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#71
Phenanthrene
Concen: 4.954 ng
RT: 17.619 min Scan# 2473
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:178 Resp: 89771

Ion	Ratio	Lower	Upper
178	100		
176	19.0	14.7	22.1
179	15.8	12.2	18.4






#72
Anthracene
Concen: 4.886 ng
RT: 17.713 min Scan# 24
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:178 Resp: 9006


Ion	Ratio	Lower	Upper
178	100		
176	17.7	14.4	21.6
179	14.8	12.4	18.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#73
Carbazole
Concen: 5.005 ng
RT: 17.966 min Scan# 2532
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:167 Resp: 81350

Ion	Ratio	Lower	Upper
167	100		
166	23.7	17.9	26.9
139	12.7	9.7	14.5






#74
Di-n-butylphthalate
Concen: 4.454 ng
RT: 18.530 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:149 Resp: 8371


Ion	Ratio	Lower	Upper
149	100		
150	8.4	7.0	10.6
104	5.2	4.1	6.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#75
Fluoranthene
Concen: 5.054 ng
RT: 19.617 min Scan# 2813
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:202 Resp: 114724

Ion	Ratio	Lower	Upper
202	100		
101	10.2	0.0	28.7
203	17.7	0.0	37.0






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.856 min Scan# 31
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:240 Resp: 381389


Ion	Ratio	Lower	Upper
240	100		
120	7.7	6.5	9.7
236	25.4	20.2	30.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 4.923 ng
RT: 19.975 min Scan# 2874
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:202 Resp: 122676

Ion	Ratio	Lower	Upper
202	100		
200	22.0	17.0	25.4
203	18.2	13.8	20.6





#79
Terphenyl-d14
Concen: 10.130 ng
RT: 20.169 min Scan# 29
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41


Tgt Ion:244 Resp: 204719


Ion	Ratio	Lower	Upper
244	100		
212	6.9	6.1	9.1
122	6.8	5.6	8.4

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#80
Butylbenzylphthalate
Concen: 3.595 ng
RT: 20.857 min Scan# 3024
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:149 Resp: 27072

Ion	Ratio	Lower	Upper
149	100		
91	67.8	59.2	88.8
206	23.1	18.1	27.1






#81
Benzo(a)anthracene
Concen: 4.878 ng
RT: 21.832 min Scan# 3
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:228 Resp: 12334


Ion	Ratio	Lower	Upper
228	100		
226	27.9	21.9	32.9
229	20.7	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#83
Chrysene
Concen: 4.949 ng
RT: 21.903 min Scan# 3202
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:228 Resp: 119080

Ion	Ratio	Lower	Upper
228	100		
226	31.2	23.3	34.9
229	20.1	15.2	22.8





#84
Bis(2-ethylhexyl)phthalate
Concen: 4.098 ng
RT: 21.750 min Scan# 31
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41


Tgt Ion:149 Resp: 4611


Ion	Ratio	Lower	Upper
149	100		
167	26.0	20.0	30.0
279	4.3	3.1	4.7

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.199 min Scan# 3763
Delta R.T. 0.002 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:264 Resp: 401350

Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.0	28.4
265	22.1	17.6	26.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 4.700 ng
RT: 29.030 min Scan# 44
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41


Tgt Ion:276 Resp: 13875


Ion	Ratio	Lower	Upper
276	100		
138	0.0	10.8	16.2
277	24.6	20.1	30.1

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#88
Benzo(b)fluoranthene
Concen: 4.640 ng
RT: 24.129 min Scan# 3581
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:252 Resp: 114203

Ion	Ratio	Lower	Upper
252	100		
253	21.0	17.8	26.8
125	6.7	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 4.787 ng
RT: 24.200 min Scan# 3
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:252 Resp: 11789


Ion	Ratio	Lower	Upper
252	100		
253	21.5	16.8	25.2
125	6.0	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 4.704 ng
RT: 25.040 min Scan# 3736
Delta R.T. -0.004 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:252 Resp: 106414

Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.6	26.4
125	7.7	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 4.686 ng
RT: 29.100 min Scan# 44
Delta R.T. -0.010 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:278 Resp: 11239


Ion	Ratio	Lower	Upper
278	100		
139	10.5	8.0	12.0
279	22.2	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 4.731 ng
RT: 30.216 min Scan# 4617
Delta R.T. -0.016 min
Lab File: BG064570.D
Acq: 24 Oct 2025 9:41

Tgt Ion:276 Resp: 108407

Ion	Ratio	Lower	Upper
276	100		
277	22.3	19.1	28.7
138	13.2	11.3	16.9



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064571.D
Acq On : 24 Oct 2025 11:02
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 17:59:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.218	152	36324	20.000	ng		0.00
21) Naphthalene-d8	11.044	136	151627	20.000	ng		0.00
39) Acenaphthene-d10	14.839	164	117618	20.000	ng		0.00
64) Phenanthrene-d10	17.577	188	300549	20.000	ng		0.00
76) Chrysene-d12	21.855	240	343958	20.000	ng		0.00
86) Perylene-d12	25.198	264	354939	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.750	112	43837	20.257	ng		0.00
7) Phenol-d6	7.348	99	60084	20.233	ng		0.00
23) Nitrobenzene-d5	9.381	82	43376	17.182	ng		0.00
42) 2,4,6-Tribromophenol	16.314	330	31236	18.400	ng		0.00
45) 2-Fluorobiphenyl	13.470	172	163340	21.050	ng		0.00
79) Terphenyl-d14	20.168	244	385094	21.129	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.588	88	10545	11.006	ng		97
3) Pyridine	3.999	79	29199	10.148	ng		95
4) n-Nitrosodimethylamine	3.905	42	16561	10.733	ng	#	92
6) Aniline	7.530	93	42134	10.427	ng		97
8) 2-Chlorophenol	7.777	128	21922	9.678	ng		96
9) Benzaldehyde	7.342	77	17117	10.558	ng		94
10) Phenol	7.372	94	33058	10.066	ng		87
11) bis(2-Chloroethyl)ether	7.618	93	25594	10.600	ng		91
12) 1,3-Dichlorobenzene	8.112	146	27537	10.562	ng		98
13) 1,4-Dichlorobenzene	8.253	146	28081	10.593	ng		91
14) 1,2-Dichlorobenzene	8.576	146	27577	10.785	ng		96
15) Benzyl Alcohol	8.447	79	23263	9.500	ng		97
16) 2,2'-oxybis(1-Chloropr...	8.741	45	65452	10.334	ng		97
17) 2-Methylphenol	8.647	107	22254	10.206	ng		97
18) Hexachloroethane	9.322	117	9723	10.452	ng		93
19) n-Nitroso-di-n-propyla...	9.017	70	20724	10.098	ng	#	98
20) 3+4-Methylphenols	8.970	107	30009	9.786	ng		95
22) Acetophenone	9.040	105	42851	10.338	ng	#	93
24) Nitrobenzene	9.422	77	23618	8.602	ng		98
25) Isophorone	9.951	82	60347	9.962	ng		98
26) 2-Nitrophenol	10.139	139	7208	9.491	ng	#	91
27) 2,4-Dimethylphenol	10.192	122	22259	9.879	ng		96
28) bis(2-Chloroethoxy)met...	10.433	93	35170	10.197	ng		98
29) 2,4-Dichlorophenol	10.680	162	22643	9.338	ng		94
30) 1,2,4-Trichlorobenzene	10.903	180	27947	10.143	ng		93
31) Naphthalene	11.097	128	86750	10.319	ng		99
32) Benzoic acid	10.245	122	10748m	10.637	ng		
33) 4-Chloroaniline	11.185	127	34581	10.583	ng		97
34) Hexachlorobutadiene	11.390	225	22099	10.256	ng		95
35) Caprolactam	11.931	113	8521	9.162	ng		92
36) 4-Chloro-3-methylphenol	12.289	107	28904	9.916	ng		98
37) 2-Methylnaphthalene	12.689	142	65070	10.543	ng		95
38) 1-Methylnaphthalene	12.906	142	64384	10.540	ng		95
40) 1,2,4,5-Tetrachloroben...	13.053	216	40253	10.197	ng		98
41) Hexachlorocyclopentadiene	13.036	237	15733	8.855	ng		97
43) 2,4,6-Trichlorophenol	13.276	196	19584	8.673	ng		91

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064571.D
Acq On : 24 Oct 2025 11:02
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Nov 03 17:59:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.341	196	24973	9.584	ng		94
46) 1,1'-Biphenyl	13.682	154	85287	10.200	ng		96
47) 2-Chloronaphthalene	13.723	162	63631	10.081	ng		98
48) 2-Nitroaniline	13.905	65	13764	9.284	ng		91
49) Acenaphthylene	14.563	152	101526	10.198	ng		100
50) Dimethylphthalate	14.281	163	92308	10.256	ng		96
51) 2,6-Dinitrotoluene	14.393	165	10601	9.857	ng		87
52) Acenaphthene	14.904	154	69273	10.184	ng		97
53) 3-Nitroaniline	14.722	138	12414	8.026	ng	#	92
54) 2,4-Dinitrophenol	14.922	184	5082	10.760	ng	#	1
55) Dibenzofuran	15.239	168	115990	10.548	ng		97
56) 4-Nitrophenol	15.010	139	11207	7.815	ng		86
57) 2,4-Dinitrotoluene	15.174	165	15797	9.494	ng	#	94
58) Fluorene	15.879	166	91712	10.679	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.450	232	22645	9.088	ng	#	99
60) Diethylphthalate	15.644	149	96719	10.525	ng		98
61) 4-Chlorophenyl-phenyle...	15.873	204	50818	10.544	ng		97
62) 4-Nitroaniline	15.873	138	14819	8.612	ng		95
63) Azobenzene	16.161	77	88441	10.489	ng		97
65) 4,6-Dinitro-2-methylph...	15.938	198	8190	10.919	ng		86
66) n-Nitrosodiphenylamine	16.079	169	82787	10.197	ng		99
67) 4-Bromophenyl-phenylether	16.767	248	35433	9.949	ng		96
68) Hexachlorobenzene	16.884	284	39975	9.838	ng		94
69) Atrazine	17.019	200	34185	9.785	ng		99
70) Pentachlorophenol	17.225	266	19859	7.838	ng		96
71) Phenanthrene	17.618	178	170294	10.361	ng		99
72) Anthracene	17.712	178	171860	10.279	ng		98
73) Carbazole	17.965	167	155907	10.575	ng		99
74) Di-n-butylphthalate	18.529	149	173716	10.189	ng		99
75) Fluoranthene	19.616	202	219838	10.677	ng		98
77) Benzidine	19.781	184	67427	8.750	ng		99
78) Pyrene	19.975	202	228905	10.186	ng		99
80) Butylbenzylphthalate	20.850	149	61669	9.080	ng		94
81) Benzo(a)anthracene	21.831	228	233051	10.219	ng		99
82) 3,3'-Dichlorobenzidine	21.737	252	71951	9.448	ng		94
83) Chrysene	21.902	228	222018	10.231	ng		97
84) Bis(2-ethylhexyl)phtha...	21.749	149	95500	9.410	ng	#	99
85) Di-n-octyl phthalate	23.012	149	149877	8.759	ng		99
87) Indeno(1,2,3-cd)pyrene	29.023	276	264547	10.132	ng		93
88) Benzo(b)fluoranthene	24.128	252	221197	10.163	ng		98
89) Benzo(k)fluoranthene	24.199	252	223674	10.270	ng		98
90) Benzo(a)pyrene	25.033	252	202985	10.146	ng		98
91) Dibenzo(a,h)anthracene	29.099	278	217114	10.235	ng		99
92) Benzo(g,h,i)perylene	30.210	276	206153	10.173	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064571.D
Acq On : 24 Oct 2025 11:02
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Nov 03 17:59:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.218 min Scan# 8
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:152 Resp: 3632


Ion	Ratio	Lower	Upper
152	100		
150	151.7	122.2	183.4
115	55.1	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 11.006 ng
RT: 3.588 min Scan# 85
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 88 Resp: 10545

Ion	Ratio	Lower	Upper
88	100		
58	112.0	92.4	138.6
43	39.4	33.0	49.6






#3
Pyridine
Concen: 10.148 ng
RT: 3.999 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 79 Resp: 2919


Ion	Ratio	Lower	Upper
79	100		
52	113.0	94.2	141.2
51	61.6	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 10.733 ng
RT: 3.905 min Scan# 139
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 42 Resp: 16561

Ion	Ratio	Lower	Upper
42	100		
74	103.1	76.1	114.1
44	11.4	6.3	9.5#






#5
2-Fluorophenol
Concen: 20.257 ng
RT: 5.750 min Scan# 4
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:112 Resp: 4383


Ion	Ratio	Lower	Upper
112	100		
64	69.1	60.7	91.1
63	35.4	33.4	50.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#6
Aniline
Concen: 10.427 ng
RT: 7.530 min Scan# 756
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 93 Resp: 42134

Ion	Ratio	Lower	Upper
93	100		
66	44.4	37.0	55.6
65	25.1	18.3	27.5






#7
Phenol-d6
Concen: 20.233 ng
RT: 7.348 min Scan# 7
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 99 Resp: 6008


Ion	Ratio	Lower	Upper
99	100		
42	25.0	18.3	27.5
71	35.3	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 9.678 ng
RT: 7.777 min Scan# 798
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:128 Resp: 21922

Ion	Ratio	Lower	Upper
128	100		
130	31.9	13.5	53.5
64	68.8	45.2	85.2





#9
Benzaldehyde
Concen: 10.558 ng
RT: 7.342 min Scan# 7
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02


Tgt Ion: 77 Resp: 1711

Ion	Ratio	Lower	Upper
77	100		
105	96.5	72.9	112.9
106	98.4	70.8	110.8

Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025



#10
Phenol
Concen: 10.066 ng
RT: 7.372 min Scan# 729
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 94 Resp: 33058


Ion	Ratio	Lower	Upper
94	100		
65	44.2	15.8	55.8
66	51.9	24.0	64.0



#11
bis(2-Chloroethyl)ether
Concen: 10.600 ng
RT: 7.618 min Scan# 7
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 93 Resp: 2559


Ion	Ratio	Lower	Upper
93	100		
63	101.2	91.3	131.3
95	29.6	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 10.562 ng
RT: 8.112 min Scan# 855
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:146 Resp: 27537

Ion	Ratio	Lower	Upper
146	100		
148	69.4	55.3	82.9
75	32.3	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 10.593 ng
RT: 8.253 min Scan# 8
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:146 Resp: 2808


Ion	Ratio	Lower	Upper
146	100		
148	62.8	56.2	84.4
111	39.3	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 10.785 ng
RT: 8.576 min Scan# 934
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:146 Resp: 27577

Ion	Ratio	Lower	Upper
146	100		
148	60.2	51.5	77.3
111	43.3	35.0	52.6






#15
Benzyl Alcohol
Concen: 9.500 ng
RT: 8.447 min Scan# 91
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 79 Resp: 2326


Ion	Ratio	Lower	Upper
79	100		
108	76.3	58.1	87.1
77	66.9	52.8	79.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.334 ng
RT: 8.741 min Scan# 962
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 45 Resp: 65452

Ion	Ratio	Lower	Upper
45	100		
77	6.7	0.0	28.5
79	5.8	0.0	25.6






#17
2-Methylphenol
Concen: 10.206 ng
RT: 8.647 min Scan# 94
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:107 Resp: 2225


Ion	Ratio	Lower	Upper
107	100		
108	111.9	90.0	135.0
77	50.5	37.8	56.8
79	46.6	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 10.452 ng
RT: 9.322 min Scan# 1061
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:117 Resp: 9723

Ion	Ratio	Lower	Upper
117	100		
119	87.5	75.4	113.2
201	115.2	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 10.098 ng
RT: 9.017 min Scan# 10
Delta R.T. -0.011 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 70 Resp: 2072


Ion	Ratio	Lower	Upper
70	100		
42	80.3	65.2	97.8
101	7.6	7.9	11.9#
130	16.2	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 9.786 ng
RT: 8.970 min Scan# 1001
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:107 Resp: 30009

Ion	Ratio	Lower	Upper
107	100		
108	84.8	70.3	110.3
77	35.4	14.6	54.6
79	29.1	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.044 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:136 Resp: 15162


Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.2	13.8
54	13.8	12.6	19.0
68	7.3	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 10.338 ng
RT: 9.040 min Scan# 1013
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:105 Resp: 42851

Ion	Ratio	Lower	Upper
105	100		
71	1.1	1.5	2.3#
51	37.9	33.7	50.5
120	16.4	16.1	24.1






#23
Nitrobenzene-d5
Concen: 17.182 ng
RT: 9.381 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 82 Resp: 4337


Ion	Ratio	Lower	Upper
82	100		
128	37.2	31.3	46.9
54	70.1	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 8.602 ng
RT: 9.422 min Scan# 1078
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 77 Resp: 23618

Ion	Ratio	Lower	Upper
77	100		
123	37.3	31.0	46.6
65	16.5	12.7	19.1






#25
Isophorone
Concen: 9.962 ng
RT: 9.951 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 82 Resp: 6034


Ion	Ratio	Lower	Upper
82	100		
95	8.3	6.1	9.1
138	15.1	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 9.491 ng
RT: 10.139 min Scan# 1200
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:139 Resp: 7208

Ion	Ratio	Lower	Upper
139	100		
109	40.9	24.6	37.0#
65	56.6	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 9.879 ng
RT: 10.192 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:122 Resp: 2225


Ion	Ratio	Lower	Upper
122	100		
107	121.6	103.0	154.4
121	60.2	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 10.197 ng
RT: 10.433 min Scan# 1250
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 93 Resp: 35170

Ion	Ratio	Lower	Upper
93	100		
95	33.2	25.4	38.2
123	11.0	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 9.338 ng
RT: 10.680 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:162 Resp: 2264


Ion	Ratio	Lower	Upper
162	100		
164	64.0	48.7	88.7
98	32.7	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 10.143 ng
RT: 10.903 min Scan# 1330
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:180 Resp: 27947

Ion	Ratio	Lower	Upper
180	100		
182	90.1	76.9	115.3
145	31.7	22.1	33.1






#31
Naphthalene
Concen: 10.319 ng
RT: 11.097 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:128 Resp: 8675


Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.8	13.2
127	12.4	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 10.637 ng m
RT: 10.245 min Scan# 1218
Delta R.T. -0.064 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:122 Resp: 10748

Ion	Ratio	Lower	Upper
122	100		
105	100.8	104.2	144.2#
77	80.6	70.2	110.2






#33
4-Chloroaniline
Concen: 10.583 ng
RT: 11.185 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:127 Resp: 3458


Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.5	38.3
65	39.1	32.2	48.4
92	19.3	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 10.256 ng
RT: 11.390 min Scan# 1413
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:225 Resp: 22099

Ion	Ratio	Lower	Upper
225	100		
223	57.6	48.8	73.2
227	57.0	49.5	74.3






#35
Caprolactam
Concen: 9.162 ng
RT: 11.931 min Scan# 1
Delta R.T. -0.017 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:113 Resp: 852


Ion	Ratio	Lower	Upper
113	100		
55	297.7	290.1	330.1
56	185.5	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 9.916 ng
RT: 12.289 min Scan# 1566
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:107 Resp: 28904

Ion	Ratio	Lower	Upper
107	100		
144	23.5	19.1	28.7
142	74.9	58.6	87.8






#37
2-Methylnaphthalene
Concen: 10.543 ng
RT: 12.689 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:142 Resp: 6507


Ion	Ratio	Lower	Upper
142	100		
141	90.6	68.9	103.3
115	36.0	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 10.540 ng
RT: 12.906 min Scan# 1671
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:142 Resp: 64384

Ion	Ratio	Lower	Upper
142	100		
141	87.2	73.3	109.9
116	4.4	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.839 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:164 Resp: 11761


Ion	Ratio	Lower	Upper
164	100		
162	98.4	77.0	115.6
160	46.2	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.197 ng
RT: 13.053 min Scan# 1696
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:216 Resp: 40253

Ion	Ratio	Lower	Upper
216	100		
214	75.7	61.4	92.2
179	20.2	15.4	23.0
108	15.7	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 8.855 ng
RT: 13.036 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:237 Resp: 1573


Ion	Ratio	Lower	Upper
237	100		
235	61.6	44.1	84.1
272	10.9	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 18.400 ng
RT: 16.314 min Scan# 2251
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:330 Resp: 31236

Ion	Ratio	Lower	Upper
330	100		
332	96.5	78.0	117.0
141	25.2	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 8.673 ng
RT: 13.276 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:196 Resp: 1958


Ion	Ratio	Lower	Upper
196	100		
198	104.8	76.6	115.0
200	36.2	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 9.584 ng
RT: 13.341 min Scan# 1745
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:196 Resp: 24973

Ion	Ratio	Lower	Upper
196	100		
198	85.8	74.2	111.4
97	42.4	37.0	55.6
132	22.8	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 21.050 ng
RT: 13.470 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:172 Resp: 16334


Ion	Ratio	Lower	Upper
172	100		
171	35.6	29.4	44.0
170	22.2	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 10.200 ng
RT: 13.682 min Scan# 1803
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:154 Resp: 85287

Ion	Ratio	Lower	Upper
154	100		
153	38.8	22.4	62.4
76	13.3	0.0	33.5






#47
2-Chloronaphthalene
Concen: 10.081 ng
RT: 13.723 min Scan# 18
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:162 Resp: 6363


Ion	Ratio	Lower	Upper
162	100		
127	41.2	32.2	48.2
164	33.4	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 9.284 ng
RT: 13.905 min Scan# 1841
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 65 Resp: 13764

Ion	Ratio	Lower	Upper
65	100		
92	58.5	48.3	72.5
138	81.1	74.9	112.3






#49
Acenaphthylene
Concen: 10.198 ng
RT: 14.563 min Scan# 19
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:152 Resp: 10152


Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.4	23.2
153	14.0	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 10.256 ng
RT: 14.281 min Scan# 1905
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:163 Resp: 92308

Ion	Ratio	Lower	Upper
163	100		
194	3.2	2.8	4.2
164	11.3	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 9.857 ng
RT: 14.393 min Scan# 19
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:165 Resp: 10601


Ion	Ratio	Lower	Upper
165	100		
63	60.2	57.4	86.0
89	50.8	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 10.184 ng
RT: 14.904 min Scan# 2011
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:154 Resp: 69273

Ion	Ratio	Lower	Upper
154	100		
153	110.8	92.2	138.4
152	53.8	43.2	64.8






#53
3-Nitroaniline
Concen: 8.026 ng
RT: 14.722 min Scan# 19
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:138 Resp: 1241


Ion	Ratio	Lower	Upper
138	100		
108	24.4	11.4	17.0#
92	126.2	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 10.760 ng
RT: 14.922 min Scan# 2014
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:184 Resp: 5082

Ion	Ratio	Lower	Upper
184	100		
63	95.9	56.6	85.0#
154	514.8	158.6	238.0#






#55
Dibenzofuran
Concen: 10.548 ng
RT: 15.239 min Scan# 20
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:168 Resp: 11599


Ion	Ratio	Lower	Upper
168	100		
139	32.6	27.8	41.8
169	12.4	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 7.815 ng
RT: 15.010 min Scan# 2029
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:139 Resp: 11207

Ion	Ratio	Lower	Upper
139	100		
109	78.9	67.8	107.8
65	105.0	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 9.494 ng
RT: 15.174 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:165 Resp: 1579


Ion	Ratio	Lower	Upper
165	100		
63	45.9	41.8	62.8
89	68.9	58.1	87.1
182	0.0	1.4	2.0#



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 10.679 ng
RT: 15.879 min Scan# 2177
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:166 Resp: 91712

Ion	Ratio	Lower	Upper
166	100		
165	101.4	81.5	122.3
167	13.4	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 9.088 ng
RT: 15.450 min Scan# 21
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:232 Resp: 2264


Ion	Ratio	Lower	Upper
232	100		
131	33.5	26.7	40.1
130	1.5	1.6	2.4#
166	24.3	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 10.525 ng
RT: 15.644 min Scan# 2137
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:149 Resp: 96719

Ion	Ratio	Lower	Upper
149	100		
177	21.3	18.0	27.0
150	11.8	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 10.544 ng
RT: 15.873 min Scan# 21
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:204 Resp: 5081


Ion	Ratio	Lower	Upper
204	100		
206	32.5	25.6	38.4
141	53.7	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 8.612 ng
RT: 15.873 min Scan# 2176
Delta R.T. -0.011 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:138 Resp: 14819

Ion	Ratio	Lower	Upper
138	100		
92	61.4	36.1	76.1
108	98.5	75.6	115.6






#63
Azobenzene
Concen: 10.489 ng
RT: 16.161 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion: 77 Resp: 8844


Ion	Ratio	Lower	Upper
77	100		
182	22.8	4.9	44.9
105	21.7	1.3	41.3
51	41.3	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.577 min Scan# 2466
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:188 Resp: 300549

Ion	Ratio	Lower	Upper
188	100		
94	7.7	5.7	8.5
80	8.8	6.8	10.2






#65
4,6-Dinitro-2-methylphenol
Concen: 10.919 ng
RT: 15.938 min Scan# 21
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:198 Resp: 819


Ion	Ratio	Lower	Upper
198	100		
51	36.4	24.7	64.7
105	28.3	18.0	58.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#66
n-Nitrosodiphenylamine
Concen: 10.197 ng
RT: 16.079 min Scan# 2211
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:169 Resp: 82787

Ion	Ratio	Lower	Upper
169	100		
168	69.7	56.3	84.5
167	37.4	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 9.949 ng
RT: 16.767 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:248 Resp: 3543


Ion	Ratio	Lower	Upper
248	100		
250	93.0	77.7	116.5
141	58.0	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 9.838 ng
RT: 16.884 min Scan# 2348
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:284 Resp: 39975

Ion	Ratio	Lower	Upper
284	100		
142	27.8	21.0	31.6
249	35.9	24.9	37.3






#69
Atrazine
Concen: 9.785 ng
RT: 17.019 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:200 Resp: 3418


Ion	Ratio	Lower	Upper
200	100		
173	20.7	0.9	40.9
215	47.7	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 7.838 ng
RT: 17.225 min Scan# 2406
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:266 Resp: 19859

Ion	Ratio	Lower	Upper
266	100		
268	58.9	50.3	75.5
264	61.2	47.3	70.9






#71
Phenanthrene
Concen: 10.361 ng
RT: 17.618 min Scan# 24
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:178 Resp: 170294


Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.7	22.1
179	15.4	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 10.279 ng
RT: 17.712 min Scan# 2489
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:178 Resp: 171860

Ion	Ratio	Lower	Upper
178	100		
176	17.2	14.4	21.6
179	14.5	12.4	18.6






#73
Carbazole
Concen: 10.575 ng
RT: 17.965 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:167 Resp: 15590


Ion	Ratio	Lower	Upper
167	100		
166	23.0	17.9	26.9
139	11.8	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 10.189 ng
RT: 18.529 min Scan# 2628
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:149 Resp: 173716

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.0	10.6
104	5.5	4.1	6.1






#75
Fluoranthene
Concen: 10.677 ng
RT: 19.616 min Scan# 28
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:202 Resp: 219838


Ion	Ratio	Lower	Upper
202	100		
101	8.1	0.0	28.7
203	15.8	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.855 min Scan# 3194
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:240 Resp: 343958

Ion	Ratio	Lower	Upper
240	100		
120	8.1	6.5	9.7
236	27.6	20.2	30.2





#77
Benzidine
Concen: 8.750 ng
RT: 19.781 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02


Tgt Ion:184 Resp: 6742


Ion	Ratio	Lower	Upper
184	100		
185	14.6	11.4	17.2
183	12.3	9.4	14.2

Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#78
Pyrene
Concen: 10.186 ng
RT: 19.975 min Scan# 2874
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:202 Resp: 228905

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.0	25.4
203	17.7	13.8	20.6






#79
Terphenyl-d14
Concen: 21.129 ng
RT: 20.168 min Scan# 29
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:244 Resp: 385094


Ion	Ratio	Lower	Upper
244	100		
212	7.0	6.1	9.1
122	7.2	5.6	8.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#80
Butylbenzylphthalate
Concen: 9.080 ng
RT: 20.850 min Scan# 3023
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:149 Resp: 61669

Ion	Ratio	Lower	Upper
149	100		
91	67.3	59.2	88.8
206	22.3	18.1	27.1






#81
Benzo(a)anthracene
Concen: 10.219 ng
RT: 21.831 min Scan# 31
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:228 Resp: 233051


Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.9	32.9
229	18.6	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 9.448 ng
RT: 21.737 min Scan# 3174
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:252 Resp: 71951

Ion	Ratio	Lower	Upper
252	100		
254	68.4	50.8	76.2
126	7.6	5.8	8.6






#83
Chrysene
Concen: 10.231 ng
RT: 21.902 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:228 Resp: 22201


Ion	Ratio	Lower	Upper
228	100		
226	31.5	23.3	34.9
229	19.6	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 9.410 ng
RT: 21.749 min Scan# 3176
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:149 Resp: 95500

Ion	Ratio	Lower	Upper
149	100		
167	24.9	20.0	30.0
279	5.0	3.1	4.7#






#85
Di-n-octyl phthalate
Concen: 8.759 ng
RT: 23.012 min Scan# 3
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:149 Resp: 14987


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	12.9	10.8	16.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 25.198 min Scan# 3763
Delta R.T. 0.001 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:264 Resp: 354939

Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.0	28.4
265	20.7	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 10.132 ng
RT: 29.023 min Scan# 44
Delta R.T. -0.017 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:276 Resp: 26454


Ion	Ratio	Lower	Upper
276	100		
138	11.8	10.8	16.2
277	20.9	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 10.163 ng
RT: 24.128 min Scan# 3581
Delta R.T. -0.011 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:252 Resp: 221197

Ion	Ratio	Lower	Upper
252	100		
253	21.2	17.8	26.8
125	7.0	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 10.270 ng
RT: 24.199 min Scan# 3
Delta R.T. -0.011 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:252 Resp: 22367


Ion	Ratio	Lower	Upper
252	100		
253	21.7	16.8	25.2
125	7.7	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 10.146 ng
RT: 25.033 min Scan# 3735
Delta R.T. -0.011 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:252 Resp: 202985

Ion	Ratio	Lower	Upper
252	100		
253	21.1	17.6	26.4
125	7.2	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 10.235 ng
RT: 29.099 min Scan# 44
Delta R.T. -0.011 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:278 Resp: 21711


Ion	Ratio	Lower	Upper
278	100		
139	10.4	8.0	12.0
279	24.0	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 10.173 ng
RT: 30.210 min Scan# 4616
Delta R.T. -0.023 min
Lab File: BG064571.D
Acq: 24 Oct 2025 11:02

Tgt Ion:276 Resp: 206153

Ion	Ratio	Lower	Upper
276	100		
277	25.0	19.1	28.7
138	13.9	11.3	16.9



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064572.D
Acq On : 24 Oct 2025 12:23
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 18:00:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.217	152	46818	20.000	ng		0.00
21) Naphthalene-d8	11.049	136	192551	20.000	ng		0.00
39) Acenaphthene-d10	14.845	164	147520	20.000	ng		0.00
64) Phenanthrene-d10	17.577	188	357228	20.000	ng		0.00
76) Chrysene-d12	21.854	240	408574	20.000	ng		0.00
86) Perylene-d12	25.198	264	430370	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.744	112	108123	38.764	ng		0.00
7) Phenol-d6	7.348	99	146608	38.304	ng		0.00
23) Nitrobenzene-d5	9.381	82	121107	37.777	ng		0.00
42) 2,4,6-Tribromophenol	16.320	330	80231	37.682	ng		0.00
45) 2-Fluorobiphenyl	13.470	172	385251	39.584	ng		0.00
79) Terphenyl-d14	20.168	244	870741	40.219	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.588	88	24533	19.867	ng		99
3) Pyridine	3.999	79	72660	19.593	ng		96
4) n-Nitrosodimethylamine	3.905	42	37799	19.007	ng	#	90
6) Aniline	7.530	93	99762	19.155	ng		99
8) 2-Chlorophenol	7.777	128	56018	19.187	ng		93
9) Benzaldehyde	7.342	77	46540	22.273	ng		97
10) Phenol	7.377	94	81849	19.337	ng		97
11) bis(2-Chloroethyl)ether	7.624	93	60495	19.439	ng		93
12) 1,3-Dichlorobenzene	8.112	146	65241	19.415	ng		99
13) 1,4-Dichlorobenzene	8.259	146	67506	19.757	ng		94
14) 1,2-Dichlorobenzene	8.582	146	63867	19.380	ng		99
15) Benzyl Alcohol	8.447	79	59123	18.732	ng		96
16) 2,2'-oxybis(1-Chloropr...	8.740	45	158985	19.476	ng		97
17) 2-Methylphenol	8.646	107	55174	19.632	ng		97
18) Hexachloroethane	9.322	117	23663	19.735	ng		97
19) n-Nitroso-di-n-propyla...	9.022	70	52857	19.982	ng	#	94
20) 3+4-Methylphenols	8.969	107	74026	18.730	ng		94
22) Acetophenone	9.040	105	101737	19.328	ng		96
24) Nitrobenzene	9.422	77	66292	19.012	ng		95
25) Isophorone	9.951	82	148388	19.290	ng		98
26) 2-Nitrophenol	10.145	139	21166	17.837	ng		92
27) 2,4-Dimethylphenol	10.192	122	54972	19.211	ng		90
28) bis(2-Chloroethoxy)met...	10.432	93	85225	19.458	ng		97
29) 2,4-Dichlorophenol	10.679	162	60695	19.710	ng		96
30) 1,2,4-Trichlorobenzene	10.908	180	67078	19.171	ng		97
31) Naphthalene	11.096	128	208734	19.552	ng		99
32) Benzoic acid	10.274	122	31567m	18.155	ng		
33) 4-Chloroaniline	11.190	127	79067	19.055	ng		96
34) Hexachlorobutadiene	11.390	225	54233	19.820	ng		98
35) Caprolactam	11.931	113	22010	18.635	ng		93
36) 4-Chloro-3-methylphenol	12.289	107	69582	18.798	ng		92
37) 2-Methylnaphthalene	12.689	142	153445	19.577	ng		97
38) 1-Methylnaphthalene	12.906	142	149456	19.266	ng		97
40) 1,2,4,5-Tetrachloroben...	13.053	216	95653	19.320	ng		99
41) Hexachlorocyclopentadiene	13.035	237	41178	18.479	ng		97
43) 2,4,6-Trichlorophenol	13.276	196	55295	19.525	ng		95

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064572.D
Acq On : 24 Oct 2025 12:23
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Nov 03 18:00:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.347	196	64753	19.814	ng		93
46) 1,1'-Biphenyl	13.682	154	206656	19.705	ng		94
47) 2-Chloronaphthalene	13.729	162	155356	19.625	ng		99
48) 2-Nitroaniline	13.911	65	40854	17.565	ng		98
49) Acenaphthylene	14.563	152	242914	19.454	ng		98
50) Dimethylphthalate	14.281	163	219575	19.452	ng		99
51) 2,6-Dinitrotoluene	14.393	165	28427	17.440	ng		96
52) Acenaphthene	14.904	154	164542	19.287	ng		96
53) 3-Nitroaniline	14.722	138	37234	19.193	ng	#	99
54) 2,4-Dinitrophenol	14.921	184	16182	18.744	ng	#	29
55) Dibenzofuran	15.239	168	269419	19.534	ng		98
56) 4-Nitrophenol	15.009	139	31366	17.439	ng		90
57) 2,4-Dinitrotoluene	15.180	165	44209	17.427	ng	#	96
58) Fluorene	15.885	166	210426	19.536	ng		97
59) 2,3,4,6-Tetrachlorophenol	15.456	232	59314	18.980	ng		99
60) Diethylphthalate	15.644	149	223398	19.382	ng		99
61) 4-Chlorophenyl-phenyle...	15.873	204	116865	19.333	ng		98
62) 4-Nitroaniline	15.879	138	42261	19.582	ng		99
63) Azobenzene	16.161	77	203101	19.205	ng		99
65) 4,6-Dinitro-2-methylph...	15.944	198	22843	18.316	ng		92
66) n-Nitrosodiphenylamine	16.079	169	187560	19.437	ng		98
67) 4-Bromophenyl-phenylether	16.766	248	82311	19.444	ng		98
68) Hexachlorobenzene	16.890	284	93647	19.391	ng		96
69) Atrazine	17.019	200	88832	21.392	ng		95
70) Pentachlorophenol	17.219	266	54275	18.022	ng		88
71) Phenanthrene	17.618	178	383359	19.624	ng		99
72) Anthracene	17.712	178	393312	19.792	ng		100
73) Carbazole	17.971	167	350722	20.014	ng		99
74) Di-n-butylphthalate	18.529	149	421239	20.788	ng		100
75) Fluoranthene	19.616	202	492924	20.142	ng		99
77) Benzidine	19.780	184	194070	21.201	ng		99
78) Pyrene	19.974	202	524431	19.646	ng		99
80) Butylbenzylphthalate	20.856	149	158394	19.633	ng		97
81) Benzo(a)anthracene	21.837	228	528009	19.490	ng		100
82) 3,3'-Dichlorobenzidine	21.737	252	174803	19.323	ng		99
83) Chrysene	21.901	228	502125	19.479	ng		97
84) Bis(2-ethylhexyl)phtha...	21.749	149	237524	19.703	ng		100
85) Di-n-octyl phthalate	23.012	149	380895	18.740	ng		99
87) Indeno(1,2,3-cd)pyrene	29.028	276	606702	19.163	ng		99
88) Benzo(b)fluoranthene	24.134	252	503145	19.066	ng		98
89) Benzo(k)fluoranthene	24.205	252	518915	19.650	ng		99
90) Benzo(a)pyrene	25.039	252	470556	19.398	ng		98
91) Dibenzo(a,h)anthracene	29.099	278	498395	19.377	ng		97
92) Benzo(g,h,i)perylene	30.221	276	468496	19.067	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064572.D
Acq On : 24 Oct 2025 12:23
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Nov 03 18:00:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.217 min Scan# 8
Delta R.T. -0.006 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:152 Resp: 4681


Ion	Ratio	Lower	Upper
152	100		
150	159.6	122.2	183.4
115	61.5	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 19.867 ng
RT: 3.588 min Scan# 85
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 88 Resp: 24533

Ion	Ratio	Lower	Upper
88	100		
58	114.9	92.4	138.6
43	38.9	33.0	49.6






#3
Pyridine
Concen: 19.593 ng
RT: 3.999 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 79 Resp: 7266


Ion	Ratio	Lower	Upper
79	100		
52	113.1	94.2	141.2
51	54.4	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 19.007 ng
RT: 3.905 min Scan# 139
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 42 Resp: 37799

Ion	Ratio	Lower	Upper
42	100		
74	105.0	76.1	114.1
44	10.3	6.3	9.5#






#5
2-Fluorophenol
Concen: 38.764 ng
RT: 5.744 min Scan# 4
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:112 Resp: 10812


Ion	Ratio	Lower	Upper
112	100		
64	78.2	60.7	91.1
63	40.7	33.4	50.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#6
Aniline
Concen: 19.155 ng
RT: 7.530 min Scan# 756
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 93 Resp: 99762

Ion	Ratio	Lower	Upper
93	100		
66	47.2	37.0	55.6
65	23.2	18.3	27.5






#7
Phenol-d6
Concen: 38.304 ng
RT: 7.348 min Scan# 7
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 99 Resp: 14660


Ion	Ratio	Lower	Upper
99	100		
42	23.4	18.3	27.5
71	33.1	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 19.187 ng
RT: 7.777 min Scan# 798
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:128 Resp: 56018

Ion	Ratio	Lower	Upper
128	100		
130	35.0	13.5	53.5
64	72.8	45.2	85.2






#9
Benzaldehyde
Concen: 22.273 ng
RT: 7.342 min Scan# 7
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 77 Resp: 4654


Ion	Ratio	Lower	Upper
77	100		
105	96.8	72.9	112.9
106	93.4	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 19.337 ng
RT: 7.377 min Scan# 730
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 94 Resp: 81849

Ion	Ratio	Lower	Upper
94	100		
65	34.1	15.8	55.8
66	42.5	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 19.439 ng
RT: 7.624 min Scan# 7
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 93 Resp: 6049


Ion	Ratio	Lower	Upper
93	100		
63	103.3	91.3	131.3
95	30.4	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 19.415 ng
RT: 8.112 min Scan# 855
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:146 Resp: 65241

Ion	Ratio	Lower	Upper
146	100		
148	70.0	55.3	82.9
75	35.2	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 19.757 ng
RT: 8.259 min Scan# 88
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:146 Resp: 6750


Ion	Ratio	Lower	Upper
146	100		
148	62.4	56.2	84.4
111	43.5	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 19.380 ng
RT: 8.582 min Scan# 935
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:146 Resp: 63867

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.5	77.3
111	42.5	35.0	52.6






#15
Benzyl Alcohol
Concen: 18.732 ng
RT: 8.447 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 79 Resp: 5912


Ion	Ratio	Lower	Upper
79	100		
108	70.4	58.1	87.1
77	62.4	52.8	79.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 19.476 ng
RT: 8.740 min Scan# 962
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 45 Resp: 158985

Ion	Ratio	Lower	Upper
45	100		
77	7.6	0.0	28.5
79	6.6	0.0	25.6






#17
2-Methylphenol
Concen: 19.632 ng
RT: 8.646 min Scan# 94
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:107 Resp: 5517


Ion	Ratio	Lower	Upper
107	100		
108	108.3	90.0	135.0
77	47.9	37.8	56.8
79	47.7	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 19.735 ng
RT: 9.322 min Scan# 1061
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:117 Resp: 23663

Ion	Ratio	Lower	Upper
117	100		
119	90.8	75.4	113.2
201	120.3	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 19.982 ng
RT: 9.022 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 70 Resp: 5285


Ion	Ratio	Lower	Upper
70	100		
42	75.6	65.2	97.8
101	7.6	7.9	11.9#
130	16.5	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 18.730 ng
RT: 8.969 min Scan# 1001
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:107 Resp: 74026

Ion	Ratio	Lower	Upper
107	100		
108	81.9	70.3	110.3
77	34.6	14.6	54.6
79	28.5	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.049 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:136 Resp: 19255


Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.2	13.8
54	14.8	12.6	19.0
68	6.0	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 19.328 ng
RT: 9.040 min Scan# 1013
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:105 Resp: 101737

Ion	Ratio	Lower	Upper
105	100		
71	1.6	1.5	2.3
51	38.8	33.7	50.5
120	20.3	16.1	24.1






#23
Nitrobenzene-d5
Concen: 37.777 ng
RT: 9.381 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 82 Resp: 12110


Ion	Ratio	Lower	Upper
82	100		
128	39.2	31.3	46.9
54	70.7	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 19.012 ng
RT: 9.422 min Scan# 1078
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 77 Resp: 66292

Ion	Ratio	Lower	Upper
77	100		
123	42.6	31.0	46.6
65	15.1	12.7	19.1






#25
Isophorone
Concen: 19.290 ng
RT: 9.951 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 82 Resp: 14838


Ion	Ratio	Lower	Upper
82	100		
95	7.0	6.1	9.1
138	15.1	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 17.837 ng
RT: 10.145 min Scan# 1201
Delta R.T. 0.001 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:139 Resp: 21166

Ion	Ratio	Lower	Upper
139	100		
109	28.2	24.6	37.0
65	52.4	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 19.211 ng
RT: 10.192 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:122 Resp: 5497


Ion	Ratio	Lower	Upper
122	100		
107	115.2	103.0	154.4
121	55.1	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 19.458 ng
RT: 10.432 min Scan# 1250
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 93 Resp: 85225

Ion	Ratio	Lower	Upper
93	100		
95	33.8	25.4	38.2
123	9.9	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 19.710 ng
RT: 10.679 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:162 Resp: 6069


Ion	Ratio	Lower	Upper
162	100		
164	66.4	48.7	88.7
98	33.1	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 19.171 ng
RT: 10.908 min Scan# 1331
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:180 Resp: 67078

Ion	Ratio	Lower	Upper
180	100		
182	92.6	76.9	115.3
145	27.5	22.1	33.1






#31
Naphthalene
Concen: 19.552 ng
RT: 11.096 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:128 Resp: 20873


Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.8	13.2
127	13.9	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 18.155 ng m
RT: 10.274 min Scan# 1223
Delta R.T. -0.035 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:122 Resp: 31567

Ion	Ratio	Lower	Upper
122	100		
105	135.4	104.2	144.2
77	99.0	70.2	110.2






#33
4-Chloroaniline
Concen: 19.055 ng
RT: 11.190 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:127 Resp: 7906


Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.5	38.3
65	36.8	32.2	48.4
92	20.2	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 19.820 ng
RT: 11.390 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:225 Resp: 54233

Ion	Ratio	Lower	Upper
225	100		
223	58.3	48.8	73.2
227	62.1	49.5	74.3






#35
Caprolactam
Concen: 18.635 ng
RT: 11.931 min Scan# 1
Delta R.T. -0.017 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:113 Resp: 2201


Ion	Ratio	Lower	Upper
113	100		
55	295.5	290.1	330.1
56	212.3	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 18.798 ng
RT: 12.289 min Scan# 1566
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:107 Resp: 69582

Ion	Ratio	Lower	Upper
107	100		
144	26.6	19.1	28.7
142	80.0	58.6	87.8






#37
2-Methylnaphthalene
Concen: 19.577 ng
RT: 12.689 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:142 Resp: 15344


Ion	Ratio	Lower	Upper
142	100		
141	84.1	68.9	103.3
115	35.5	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 19.266 ng
RT: 12.906 min Scan# 1671
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:142 Resp: 149456

Ion	Ratio	Lower	Upper
142	100		
141	89.0	73.3	109.9
116	3.9	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.845 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:164 Resp: 14752


Ion	Ratio	Lower	Upper
164	100		
162	99.6	77.0	115.6
160	45.6	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 19.320 ng
RT: 13.053 min Scan# 1696
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:216 Resp: 95653

Ion	Ratio	Lower	Upper
216	100		
214	78.1	61.4	92.2
179	20.4	15.4	23.0
108	14.7	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 18.479 ng
RT: 13.035 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:237 Resp: 4117


Ion	Ratio	Lower	Upper
237	100		
235	61.9	44.1	84.1
272	11.7	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 37.682 ng
RT: 16.320 min Scan# 2252
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:330 Resp: 80231

Ion	Ratio	Lower	Upper
330	100		
332	99.2	78.0	117.0
141	27.2	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 19.525 ng
RT: 13.276 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:196 Resp: 5529


Ion	Ratio	Lower	Upper
196	100		
198	90.9	76.6	115.0
200	30.7	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 19.814 ng
RT: 13.347 min Scan# 1746
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:196 Resp: 64753

Ion	Ratio	Lower	Upper
196	100		
198	101.1	74.2	111.4
97	43.6	37.0	55.6
132	21.4	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 39.584 ng
RT: 13.470 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:172 Resp: 38525


Ion	Ratio	Lower	Upper
172	100		
171	36.5	29.4	44.0
170	24.7	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 19.705 ng
RT: 13.682 min Scan# 1803
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:154 Resp: 206656

Ion	Ratio	Lower	Upper
154	100		
153	38.5	22.4	62.4
76	12.0	0.0	33.5






#47
2-Chloronaphthalene
Concen: 19.625 ng
RT: 13.729 min Scan# 18
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:162 Resp: 15535


Ion	Ratio	Lower	Upper
162	100		
127	40.6	32.2	48.2
164	33.2	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 17.565 ng
RT: 13.911 min Scan# 1842
Delta R.T. 0.001 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 65 Resp: 40854

Ion	Ratio	Lower	Upper
65	100		
92	60.5	48.3	72.5
138	90.1	74.9	112.3






#49
Acenaphthylene
Concen: 19.454 ng
RT: 14.563 min Scan# 19
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:152 Resp: 242914


Ion	Ratio	Lower	Upper
152	100		
151	18.6	15.4	23.2
153	12.9	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 19.452 ng
RT: 14.281 min Scan# 1905
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:163 Resp: 219575

Ion	Ratio	Lower	Upper
163	100		
194	3.6	2.8	4.2
164	10.2	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 17.440 ng
RT: 14.393 min Scan# 19
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:165 Resp: 2842


Ion	Ratio	Lower	Upper
165	100		
63	70.8	57.4	86.0
89	65.5	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 19.287 ng
RT: 14.904 min Scan# 2011
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:154 Resp: 164542

Ion	Ratio	Lower	Upper
154	100		
153	120.8	92.2	138.4
152	55.3	43.2	64.8






#53
3-Nitroaniline
Concen: 19.193 ng
RT: 14.722 min Scan# 19
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:138 Resp: 3723


Ion	Ratio	Lower	Upper
138	100		
108	17.6	11.4	17.0#
92	118.5	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 18.744 ng
RT: 14.921 min Scan# 2014
Delta R.T. -0.006 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:184 Resp: 16182

Ion	Ratio	Lower	Upper
184	100		
63	78.3	56.6	85.0
154	339.0	158.6	238.0#






#55
Dibenzofuran
Concen: 19.534 ng
RT: 15.239 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:168 Resp: 26941


Ion	Ratio	Lower	Upper
168	100		
139	36.1	27.8	41.8
169	13.0	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 17.439 ng
RT: 15.009 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:139 Resp: 31366

Ion	Ratio	Lower	Upper
139	100		
109	76.4	67.8	107.8
65	115.0	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 17.427 ng
RT: 15.180 min Scan# 20
Delta R.T. 0.001 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:165 Resp: 4420


Ion	Ratio	Lower	Upper
165	100		
63	50.7	41.8	62.8
89	68.7	58.1	87.1
182	2.2	1.4	2.0#



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 19.536 ng
RT: 15.885 min Scan# 2178
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:166 Resp: 210426

Ion	Ratio	Lower	Upper
166	100		
165	99.0	81.5	122.3
167	12.8	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 18.980 ng
RT: 15.456 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:232 Resp: 59314


Ion	Ratio	Lower	Upper
232	100		
131	32.5	26.7	40.1
130	1.8	1.6	2.4
166	25.9	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 19.382 ng
RT: 15.644 min Scan# 2137
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:149 Resp: 223398

Ion	Ratio	Lower	Upper
149	100		
177	23.0	18.0	27.0
150	11.9	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 19.333 ng
RT: 15.873 min Scan# 21
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:204 Resp: 11686


Ion	Ratio	Lower	Upper
204	100		
206	33.3	25.6	38.4
141	49.7	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 19.582 ng
RT: 15.879 min Scan# 2177
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:138 Resp: 42261

Ion	Ratio	Lower	Upper
138	100		
92	53.4	36.1	76.1
108	95.8	75.6	115.6






#63
Azobenzene
Concen: 19.205 ng
RT: 16.161 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion: 77 Resp: 20310


Ion	Ratio	Lower	Upper
77	100		
182	24.0	4.9	44.9
105	21.3	1.3	41.3
51	39.4	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.577 min Scan# 2466
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:188 Resp: 357228

Ion	Ratio	Lower	Upper
188	100		
94	7.6	5.7	8.5
80	8.4	6.8	10.2





#65
4,6-Dinitro-2-methylphenol
Concen: 18.316 ng
RT: 15.944 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23


Tgt Ion:198 Resp: 2284


Ion	Ratio	Lower	Upper
198	100		
51	41.3	24.7	64.7
105	30.8	18.0	58.0

Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#66
n-Nitrosodiphenylamine
Concen: 19.437 ng
RT: 16.079 min Scan# 2211
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:169 Resp: 187560

Ion	Ratio	Lower	Upper
169	100		
168	69.0	56.3	84.5
167	38.0	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 19.444 ng
RT: 16.766 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:248 Resp: 8231


Ion	Ratio	Lower	Upper
248	100		
250	94.2	77.7	116.5
141	60.9	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 19.391 ng
RT: 16.890 min Scan# 2349
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:284 Resp: 93647

Ion	Ratio	Lower	Upper
284	100		
142	25.4	21.0	31.6
249	34.6	24.9	37.3






#69
Atrazine
Concen: 21.392 ng
RT: 17.019 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:200 Resp: 8883


Ion	Ratio	Lower	Upper
200	100		
173	24.1	0.9	40.9
215	49.6	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 18.022 ng
RT: 17.219 min Scan# 2405
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:266 Resp: 54275

Ion	Ratio	Lower	Upper
266	100		
268	71.7	50.3	75.5
264	68.9	47.3	70.9






#71
Phenanthrene
Concen: 19.624 ng
RT: 17.618 min Scan# 24
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:178 Resp: 383359


Ion	Ratio	Lower	Upper
178	100		
176	17.9	14.7	22.1
179	15.3	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 19.792 ng
RT: 17.712 min Scan# 2489
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:178 Resp: 393312

Ion	Ratio	Lower	Upper
178	100		
176	17.9	14.4	21.6
179	15.7	12.4	18.6






#73
Carbazole
Concen: 20.014 ng
RT: 17.971 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:167 Resp: 35072


Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.9	26.9
139	11.9	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 20.788 ng
RT: 18.529 min Scan# 2628
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:149 Resp: 421239

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.0	10.6
104	5.4	4.1	6.1






#75
Fluoranthene
Concen: 20.142 ng
RT: 19.616 min Scan# 28
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:202 Resp: 49292


Ion	Ratio	Lower	Upper
202	100		
101	8.0	0.0	28.7
203	17.0	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.854 min Scan# 3194
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:240 Resp: 408574

Ion	Ratio	Lower	Upper
240	100		
120	7.6	6.5	9.7
236	25.9	20.2	30.2






#77
Benzidine
Concen: 21.201 ng
RT: 19.780 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:184 Resp: 19407


Ion	Ratio	Lower	Upper
184	100		
185	13.9	11.4	17.2
183	11.7	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 19.646 ng
RT: 19.974 min Scan# 2874
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:202 Resp: 524431

Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.0	25.4
203	18.0	13.8	20.6






#79
Terphenyl-d14
Concen: 40.219 ng
RT: 20.168 min Scan# 29
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:244 Resp: 870741


Ion	Ratio	Lower	Upper
244	100		
212	7.1	6.1	9.1
122	7.7	5.6	8.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#80
Butylbenzylphthalate
Concen: 19.633 ng
RT: 20.856 min Scan# 3024
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:149 Resp: 158394

Ion	Ratio	Lower	Upper
149	100		
91	70.4	59.2	88.8
206	22.3	18.1	27.1






#81
Benzo(a)anthracene
Concen: 19.490 ng
RT: 21.837 min Scan# 3
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:228 Resp: 52800


Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.9	32.9
229	19.7	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 19.323 ng
RT: 21.737 min Scan# 3174
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:252 Resp: 174803

Ion	Ratio	Lower	Upper
252	100		
254	63.9	50.8	76.2
126	8.0	5.8	8.6






#83
Chrysene
Concen: 19.479 ng
RT: 21.901 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:228 Resp: 50212


Ion	Ratio	Lower	Upper
228	100		
226	31.1	23.3	34.9
229	20.2	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 19.703 ng
RT: 21.749 min Scan# 3176
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:149 Resp: 237524

Ion	Ratio	Lower	Upper
149	100		
167	25.0	20.0	30.0
279	4.1	3.1	4.7





#85
Di-n-octyl phthalate
Concen: 18.740 ng
RT: 23.012 min Scan# 3
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23


Tgt Ion:149 Resp: 38089


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	14.0	10.8	16.2

Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.198 min Scan# 3763
Delta R.T. 0.000 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:264 Resp: 430370

Ion	Ratio	Lower	Upper
264	100		
260	23.2	19.0	28.4
265	22.2	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 19.163 ng
RT: 29.028 min Scan# 44
Delta R.T. -0.011 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:276 Resp: 606702


Ion	Ratio	Lower	Upper
276	100		
138	12.0	10.8	16.2
277	25.1	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 19.066 ng
RT: 24.134 min Scan# 3582
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:252 Resp: 503145

Ion	Ratio	Lower	Upper
252	100		
253	21.2	17.8	26.8
125	7.4	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 19.650 ng
RT: 24.205 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:252 Resp: 51891


Ion	Ratio	Lower	Upper
252	100		
253	20.5	16.8	25.2
125	7.2	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 19.398 ng
RT: 25.039 min Scan# 3736
Delta R.T. -0.005 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:252 Resp: 470556

Ion	Ratio	Lower	Upper
252	100		
253	20.6	17.6	26.4
125	7.4	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 19.377 ng
RT: 29.099 min Scan# 44
Delta R.T. -0.011 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:278 Resp: 49839


Ion	Ratio	Lower	Upper
278	100		
139	10.2	8.0	12.0
279	23.0	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 19.067 ng
RT: 30.221 min Scan# 4618
Delta R.T. -0.011 min
Lab File: BG064572.D
Acq: 24 Oct 2025 12:23

Tgt Ion:276 Resp: 468496

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.1	28.7
138	15.1	11.3	16.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064573.D
Acq On : 24 Oct 2025 13:03
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 18:00:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.223	152	40195	20.000	ng	0.00
21) Naphthalene-d8	11.049	136	164795	20.000	ng	0.00
39) Acenaphthene-d10	14.845	164	130542	20.000	ng	0.00
64) Phenanthrene-d10	17.582	188	313311	20.000	ng	0.00
76) Chrysene-d12	21.860	240	332834	20.000	ng	0.00
86) Perylene-d12	25.197	264	350897	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.749	112	207807	86.778	ng	0.00
7) Phenol-d6	7.353	99	284694	86.638	ng	0.00
23) Nitrobenzene-d5	9.386	82	248341	90.513	ng	0.00
42) 2,4,6-Tribromophenol	16.319	330	165976	88.092	ng	0.00
45) 2-Fluorobiphenyl	13.476	172	722477	83.888	ng	0.00
79) Terphenyl-d14	20.174	244	1521028	86.242	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.587	88	42215	39.818	ng	100
3) Pyridine	3.998	79	134153	42.136	ng	100
4) n-Nitrosodimethylamine	3.904	42	72367	42.385	ng	100
6) Aniline	7.530	93	191799	42.894	ng	100
8) 2-Chlorophenol	7.776	128	110414	44.051	ng	100
9) Benzaldehyde	7.342	77	75557	42.117	ng	100
10) Phenol	7.377	94	156259	42.998	ng	100
11) bis(2-Chloroethyl)ether	7.624	93	111183	41.613	ng	100
12) 1,3-Dichlorobenzene	8.111	146	122224	42.367	ng	100
13) 1,4-Dichlorobenzene	8.258	146	123314	42.037	ng	100
14) 1,2-Dichlorobenzene	8.581	146	119228	42.140	ng	100
15) Benzyl Alcohol	8.446	79	114526	42.265	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.740	45	300095	42.819	ng	100
17) 2-Methylphenol	8.652	107	102315	42.405	ng	100
18) Hexachloroethane	9.322	117	43488	42.246	ng	100
19) n-Nitroso-di-n-propyla...	9.028	70	98647	43.438	ng	100
20) 3+4-Methylphenols	8.975	107	146177	43.080	ng	100
22) Acetophenone	9.045	105	192583	42.750	ng	100
24) Nitrobenzene	9.427	77	135107	45.274	ng	100
25) Isophorone	9.956	82	286439	43.507	ng	100
26) 2-Nitrophenol	10.144	139	45134	39.772	ng	100
27) 2,4-Dimethylphenol	10.191	122	110518	45.129	ng	100
28) bis(2-Chloroethoxy)met...	10.438	93	161449	43.070	ng	100
29) 2,4-Dichlorophenol	10.679	162	118669	45.026	ng	100
30) 1,2,4-Trichlorobenzene	10.908	180	128902	43.046	ng	100
31) Naphthalene	11.096	128	388032	42.469	ng	100
32) Benzoic acid	10.309	122	73604m	40.425	ng	
33) 4-Chloroaniline	11.190	127	151231	42.585	ng	100
34) Hexachlorobutadiene	11.390	225	98683	42.139	ng	100
35) Caprolactam	11.948	113	44205	43.731	ng	100
36) 4-Chloro-3-methylphenol	12.289	107	140272	44.278	ng	100
37) 2-Methylnaphthalene	12.688	142	285089	42.499	ng	100
38) 1-Methylnaphthalene	12.906	142	281256	42.362	ng	100
40) 1,2,4,5-Tetrachloroben...	13.052	216	183376	41.856	ng	100
41) Hexachlorocyclopentadiene	13.041	237	84505	42.853	ng	100
43) 2,4,6-Trichlorophenol	13.276	196	110509	44.097	ng	100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064573.D
Acq On : 24 Oct 2025 13:03
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Nov 03 18:00:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.346	196	126360	43.694	ng	100
46) 1,1'-Biphenyl	13.681	154	388718	41.885	ng	100
47) 2-Chloronaphthalene	13.728	162	295501	42.182	ng	100
48) 2-Nitroaniline	13.910	65	90958	39.303	ng	100
49) Acenaphthylene	14.568	152	466393	42.209	ng	100
50) Dimethylphthalate	14.286	163	428475	42.894	ng	100
51) 2,6-Dinitrotoluene	14.398	165	64452	39.696	ng	100
52) Acenaphthene	14.909	154	319250	42.289	ng	100
53) 3-Nitroaniline	14.727	138	76841	44.762	ng	100
54) 2,4-Dinitrophenol	14.927	184	35677	38.390	ng	100
55) Dibenzofuran	15.238	168	512547	41.995	ng	100
56) 4-Nitrophenol	15.009	139	68064	42.765	ng	100
57) 2,4-Dinitrotoluene	15.179	165	100448	39.962	ng	100
58) Fluorene	15.884	166	399745	41.939	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.456	232	121762	44.030	ng	100
60) Diethylphthalate	15.649	149	432571	42.412	ng	100
61) 4-Chlorophenyl-phenyle...	15.879	204	226104	42.270	ng	100
62) 4-Nitroaniline	15.884	138	85297	44.662	ng	100
63) Azobenzene	16.166	77	401676	42.922	ng	100
65) 4,6-Dinitro-2-methylph...	15.943	198	51829	38.770	ng	100
66) n-Nitrosodiphenylamine	16.084	169	365624	43.201	ng	100
67) 4-Bromophenyl-phenylether	16.766	248	161744	43.563	ng	100
68) Hexachlorobenzene	16.889	284	183262	43.265	ng	100
69) Atrazine	17.024	200	153706	42.203	ng	100
70) Pentachlorophenol	17.224	266	112769	42.694	ng	100
71) Phenanthrene	17.624	178	728184	42.500	ng	100
72) Anthracene	17.712	178	743567	42.663	ng	100
73) Carbazole	17.970	167	646319	42.053	ng	100
74) Di-n-butylphthalate	18.534	149	776847	43.711	ng	100
75) Fluoranthene	19.615	202	902557	42.050	ng	100
77) Benzidine	19.780	184	325845	43.697	ng	100
78) Pyrene	19.974	202	939110	43.186	ng	100
80) Butylbenzylphthalate	20.855	149	299706	45.601	ng	100
81) Benzo(a)anthracene	21.836	228	942839	42.723	ng	100
82) 3,3'-Dichlorobenzidine	21.742	252	317851	43.131	ng	100
83) Chrysene	21.907	228	894813	42.613	ng	100
84) Bis(2-ethylhexyl)phtha...	21.748	149	446785	45.494	ng	100
85) Di-n-octyl phthalate	23.011	149	723360	43.688	ng	100
87) Indeno(1,2,3-cd)pyrene	29.040	276	1100259	42.624	ng	100
88) Benzo(b)fluoranthene	24.139	252	927550	43.108	ng	100
89) Benzo(k)fluoranthene	24.210	252	901706	41.878	ng	100
90) Benzo(a)pyrene	25.044	252	843258	42.635	ng	100
91) Dibenzo(a,h)anthracene	29.110	278	889808	42.430	ng	100
92) Benzo(g,h,i)perylene	30.232	276	851248	42.492	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064573.D
Acq On : 24 Oct 2025 13:03
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Nov 03 18:00:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.223 min Scan# 8
Delta R.T. -0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:152 Resp: 4019


Ion	Ratio	Lower	Upper
152	100		
150	152.8	122.2	183.4
115	63.5	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 39.818 ng
RT: 3.587 min Scan# 85
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 88 Resp: 42215

Ion	Ratio	Lower	Upper
88	100		
58	115.5	92.4	138.6
43	41.3	33.0	49.6






#3
Pyridine
Concen: 42.136 ng
RT: 3.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 79 Resp: 13415


Ion	Ratio	Lower	Upper
79	100		
52	117.7	94.2	141.2
51	57.6	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 42.385 ng
RT: 3.904 min Scan# 139
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 42 Resp: 72367

Ion	Ratio	Lower	Upper
42	100		
74	95.1	76.1	114.1
44	7.9	6.3	9.5






#5
2-Fluorophenol
Concen: 86.778 ng
RT: 5.749 min Scan# 4
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:112 Resp: 20780


Ion	Ratio	Lower	Upper
112	100		
64	75.9	60.7	91.1
63	41.7	33.4	50.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#6
Aniline
Concen: 42.894 ng
RT: 7.530 min Scan# 756
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 93 Resp: 191799

Ion	Ratio	Lower	Upper
93	100		
66	46.3	37.0	55.6
65	22.9	18.3	27.5






#7
Phenol-d6
Concen: 86.638 ng
RT: 7.353 min Scan# 7
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 99 Resp: 28469


Ion	Ratio	Lower	Upper
99	100		
42	22.9	18.3	27.5
71	35.8	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 44.051 ng
RT: 7.776 min Scan# 798
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:128 Resp: 110414

Ion	Ratio	Lower	Upper
128	100		
130	33.5	13.5	53.5
64	65.2	45.2	85.2






#9
Benzaldehyde
Concen: 42.117 ng
RT: 7.342 min Scan# 7
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 77 Resp: 7555


Ion	Ratio	Lower	Upper
77	100		
105	92.9	72.9	112.9
106	90.8	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 42.998 ng
RT: 7.377 min Scan# 730
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 94 Resp: 156259

Ion	Ratio	Lower	Upper
94	100		
65	35.8	15.8	55.8
66	44.0	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 41.613 ng
RT: 7.624 min Scan# 7
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 93 Resp: 11118


Ion	Ratio	Lower	Upper
93	100		
63	111.3	91.3	131.3
95	33.3	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 42.367 ng
RT: 8.111 min Scan# 855
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:146 Resp: 122224

Ion	Ratio	Lower	Upper
146	100		
148	69.1	55.3	82.9
75	35.2	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 42.037 ng
RT: 8.258 min Scan# 88
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:146 Resp: 12331


Ion	Ratio	Lower	Upper
146	100		
148	70.3	56.2	84.4
111	44.2	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 42.140 ng
RT: 8.581 min Scan# 935
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:146 Resp: 119228

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.5	77.3
111	43.8	35.0	52.6






#15
Benzyl Alcohol
Concen: 42.265 ng
RT: 8.446 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 79 Resp: 11452


Ion	Ratio	Lower	Upper
79	100		
108	72.6	58.1	87.1
77	66.0	52.8	79.2



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.819 ng
RT: 8.740 min Scan# 962
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 45 Resp: 300095

Ion	Ratio	Lower	Upper
45	100		
77	8.5	0.0	28.5
79	5.6	0.0	25.6






#17
2-Methylphenol
Concen: 42.405 ng
RT: 8.652 min Scan# 94
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:107 Resp: 10231


Ion	Ratio	Lower	Upper
107	100		
108	112.5	90.0	135.0
77	47.3	37.8	56.8
79	50.6	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 42.246 ng
RT: 9.322 min Scan# 1061
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:117 Resp: 43488

Ion	Ratio	Lower	Upper
117	100		
119	94.3	75.4	113.2
201	124.0	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 43.438 ng
RT: 9.028 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03


Tgt Ion: 70 Resp: 9864

Ion	Ratio	Lower	Upper
70	100		
42	81.5	65.2	97.8
101	9.9	7.9	11.9
130	17.9	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040


Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 43.080 ng
RT: 8.975 min Scan# 1002
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:107 Resp: 146177

Ion	Ratio	Lower	Upper
107	100		
108	90.3	70.3	110.3
77	34.6	14.6	54.6
79	31.7	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.049 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:136 Resp: 16479


Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.2	13.8
54	15.8	12.6	19.0
68	7.1	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 42.750 ng
RT: 9.045 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:105 Resp: 192583

Ion	Ratio	Lower	Upper
105	100		
71	1.9	1.5	2.3
51	42.1	33.7	50.5
120	20.1	16.1	24.1






#23
Nitrobenzene-d5
Concen: 90.513 ng
RT: 9.386 min Scan# 10
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 82 Resp: 248341


Ion	Ratio	Lower	Upper
82	100		
128	39.1	31.3	46.9
54	70.7	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 45.274 ng
RT: 9.427 min Scan# 1079
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 77 Resp: 135107

Ion	Ratio	Lower	Upper
77	100		
123	38.8	31.0	46.6
65	15.9	12.7	19.1






#25
Isophorone
Concen: 43.507 ng
RT: 9.956 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 82 Resp: 28643


Ion	Ratio	Lower	Upper
82	100		
95	7.6	6.1	9.1
138	16.2	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 39.772 ng
RT: 10.144 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:139 Resp: 45134

Ion	Ratio	Lower	Upper
139	100		
109	30.8	24.6	37.0
65	59.4	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 45.129 ng
RT: 10.191 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:122 Resp: 11051


Ion	Ratio	Lower	Upper
122	100		
107	128.7	103.0	154.4
121	59.8	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 43.070 ng
RT: 10.438 min Scan# 1251
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 93 Resp: 161449

Ion	Ratio	Lower	Upper
93	100		
95	31.8	25.4	38.2
123	10.5	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 45.026 ng
RT: 10.679 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:162 Resp: 11866


Ion	Ratio	Lower	Upper
162	100		
164	68.7	48.7	88.7
98	36.9	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 43.046 ng
RT: 10.908 min Scan# 1331
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:180 Resp: 128902

Ion	Ratio	Lower	Upper
180	100		
182	96.1	76.9	115.3
145	27.6	22.1	33.1






#31
Naphthalene
Concen: 42.469 ng
RT: 11.096 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:128 Resp: 38803


Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.2	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 40.425 ng m
RT: 10.309 min Scan# 1229
Delta R.T. -0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:122 Resp: 73604

Ion	Ratio	Lower	Upper
122	100		
105	124.2	104.2	144.2
77	90.2	70.2	110.2






#33
4-Chloroaniline
Concen: 42.585 ng
RT: 11.190 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:127 Resp: 15123


Ion	Ratio	Lower	Upper
127	100		
129	31.9	25.5	38.3
65	40.3	32.2	48.4
92	22.7	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 42.139 ng
RT: 11.390 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:225 Resp: 98683

Ion	Ratio	Lower	Upper
225	100		
223	61.0	48.8	73.2
227	61.9	49.5	74.3






#35
Caprolactam
Concen: 43.731 ng
RT: 11.948 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:113 Resp: 4420


Ion	Ratio	Lower	Upper
113	100		
55	310.1	290.1	330.1
56	203.8	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 44.278 ng
RT: 12.289 min Scan# 1566
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:107 Resp: 140272

Ion	Ratio	Lower	Upper
107	100		
144	23.9	19.1	28.7
142	73.2	58.6	87.8






#37
2-Methylnaphthalene
Concen: 42.499 ng
RT: 12.688 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:142 Resp: 28508


Ion	Ratio	Lower	Upper
142	100		
141	86.1	68.9	103.3
115	38.3	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 42.362 ng
RT: 12.906 min Scan# 1671
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:142 Resp: 281256

Ion	Ratio	Lower	Upper
142	100		
141	91.6	73.3	109.9
116	3.9	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.845 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:164 Resp: 13054


Ion	Ratio	Lower	Upper
164	100		
162	96.3	77.0	115.6
160	43.0	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.856 ng
RT: 13.052 min Scan# 1696
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:216 Resp: 183376

Ion	Ratio	Lower	Upper
216	100		
214	76.8	61.4	92.2
179	19.2	15.4	23.0
108	14.7	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 42.853 ng
RT: 13.041 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:237 Resp: 8450


Ion	Ratio	Lower	Upper
237	100		
235	64.1	44.1	84.1
272	12.4	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 88.092 ng
RT: 16.319 min Scan# 2252
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:330 Resp: 165976

Ion	Ratio	Lower	Upper
330	100		
332	97.5	78.0	117.0
141	27.2	21.8	32.6





#43
2,4,6-Trichlorophenol
Concen: 44.097 ng
RT: 13.276 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03


Tgt Ion:196 Resp: 11050


Ion	Ratio	Lower	Upper
196	100		
198	95.8	76.6	115.0
200	32.5	26.0	39.0

Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#44
2,4,5-Trichlorophenol
Concen: 43.694 ng
RT: 13.346 min Scan# 1746
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:196 Resp: 126360

Ion	Ratio	Lower	Upper
196	100		
198	92.8	74.2	111.4
97	46.3	37.0	55.6
132	23.1	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 83.888 ng
RT: 13.476 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:172 Resp: 72247


Ion	Ratio	Lower	Upper
172	100		
171	36.7	29.4	44.0
170	23.9	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 41.885 ng
RT: 13.681 min Scan# 1803
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:154 Resp: 388718

Ion	Ratio	Lower	Upper
154	100		
153	42.4	22.4	62.4
76	13.5	0.0	33.5






#47
2-Chloronaphthalene
Concen: 42.182 ng
RT: 13.728 min Scan# 18
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:162 Resp: 29550


Ion	Ratio	Lower	Upper
162	100		
127	40.2	32.2	48.2
164	34.4	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 39.303 ng
RT: 13.910 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 65 Resp: 90958

Ion	Ratio	Lower	Upper
65	100		
92	60.4	48.3	72.5
138	93.6	74.9	112.3






#49
Acenaphthylene
Concen: 42.209 ng
RT: 14.568 min Scan# 19
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:152 Resp: 46639


Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.4	23.2
153	13.7	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 42.894 ng
RT: 14.286 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:163 Resp: 428475

Ion	Ratio	Lower	Upper
163	100		
194	3.5	2.8	4.2
164	9.6	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 39.696 ng
RT: 14.398 min Scan# 19
Delta R.T. -0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:165 Resp: 6445


Ion	Ratio	Lower	Upper
165	100		
63	71.7	57.4	86.0
89	59.8	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 42.289 ng
RT: 14.909 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:154 Resp: 319250

Ion	Ratio	Lower	Upper
154	100		
153	115.3	92.2	138.4
152	54.0	43.2	64.8






#53
3-Nitroaniline
Concen: 44.762 ng
RT: 14.727 min Scan# 19
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:138 Resp: 7684


Ion	Ratio	Lower	Upper
138	100		
108	14.2	11.4	17.0
92	119.1	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 38.390 ng
RT: 14.927 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:184 Resp: 35677

Ion	Ratio	Lower	Upper
184	100		
63	70.8	56.6	85.0
154	198.3	158.6	238.0






#55
Dibenzofuran
Concen: 41.995 ng
RT: 15.238 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:168 Resp: 51254


Ion	Ratio	Lower	Upper
168	100		
139	34.8	27.8	41.8
169	12.9	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 42.765 ng
RT: 15.009 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:139 Resp: 68064

Ion	Ratio	Lower	Upper
139	100		
109	87.8	67.8	107.8
65	123.7	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 39.962 ng
RT: 15.179 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:165 Resp: 10044


Ion	Ratio	Lower	Upper
165	100		
63	52.3	41.8	62.8
89	72.6	58.1	87.1
182	1.7	1.4	2.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 41.939 ng
RT: 15.884 min Scan# 2178
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:166 Resp: 399745

Ion	Ratio	Lower	Upper
166	100		
165	101.9	81.5	122.3
167	14.2	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 44.030 ng
RT: 15.456 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:232 Resp: 12176


Ion	Ratio	Lower	Upper
232	100		
131	33.4	26.7	40.1
130	2.0	1.6	2.4
166	25.6	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 42.412 ng
RT: 15.649 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:149 Resp: 432571

Ion	Ratio	Lower	Upper
149	100		
177	22.5	18.0	27.0
150	12.8	10.2	15.4





#61
4-Chlorophenyl-phenylether
Concen: 42.270 ng
RT: 15.879 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03


Tgt Ion:204 Resp: 22610


Ion	Ratio	Lower	Upper
204	100		
206	32.0	25.6	38.4
141	50.7	40.6	60.8

Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#62
4-Nitroaniline
Concen: 44.662 ng
RT: 15.884 min Scan# 2178
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:138 Resp: 85297

Ion	Ratio	Lower	Upper
138	100		
92	56.1	36.1	76.1
108	95.6	75.6	115.6






#63
Azobenzene
Concen: 42.922 ng
RT: 16.166 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion: 77 Resp: 40167


Ion	Ratio	Lower	Upper
77	100		
182	24.9	4.9	44.9
105	21.3	1.3	41.3
51	39.0	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.582 min Scan# 2467
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:188 Resp: 313311

Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.7	8.5
80	8.5	6.8	10.2






#65
4,6-Dinitro-2-methylphenol
Concen: 38.770 ng
RT: 15.943 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:198 Resp: 5182


Ion	Ratio	Lower	Upper
198	100		
51	44.7	24.7	64.7
105	38.0	18.0	58.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#66
n-Nitrosodiphenylamine
Concen: 43.201 ng
RT: 16.084 min Scan# 2212
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:169 Resp: 365624

Ion	Ratio	Lower	Upper
169	100		
168	70.4	56.3	84.5
167	36.3	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 43.563 ng
RT: 16.766 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:248 Resp: 16174


Ion	Ratio	Lower	Upper
248	100		
250	97.1	77.7	116.5
141	60.7	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 43.265 ng
RT: 16.889 min Scan# 2349
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:284 Resp: 183262

Ion	Ratio	Lower	Upper
284	100		
142	26.3	21.0	31.6
249	31.1	24.9	37.3






#69
Atrazine
Concen: 42.203 ng
RT: 17.024 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:200 Resp: 153706


Ion	Ratio	Lower	Upper
200	100		
173	20.9	0.9	40.9
215	47.0	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 42.694 ng
RT: 17.224 min Scan# 2406
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:266 Resp: 112769

Ion	Ratio	Lower	Upper
266	100		
268	62.9	50.3	75.5
264	59.1	47.3	70.9






#71
Phenanthrene
Concen: 42.500 ng
RT: 17.624 min Scan# 24
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:178 Resp: 72818


Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.7	22.1
179	15.3	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 42.663 ng
RT: 17.712 min Scan# 2489
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:178 Resp: 743567

Ion	Ratio	Lower	Upper
178	100		
176	18.0	14.4	21.6
179	15.5	12.4	18.6






#73
Carbazole
Concen: 42.053 ng
RT: 17.970 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:167 Resp: 64631


Ion	Ratio	Lower	Upper
167	100		
166	22.4	17.9	26.9
139	12.1	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 43.711 ng
RT: 18.534 min Scan# 2629
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:149 Resp: 776847

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.0	10.6
104	5.1	4.1	6.1






#75
Fluoranthene
Concen: 42.050 ng
RT: 19.615 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:202 Resp: 90255


Ion	Ratio	Lower	Upper
202	100		
101	8.7	0.0	28.7
203	17.0	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.860 min Scan# 3195
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:240 Resp: 332834

Ion	Ratio	Lower	Upper
240	100		
120	8.1	6.5	9.7
236	25.2	20.2	30.2






#77
Benzidine
Concen: 43.697 ng
RT: 19.780 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:184 Resp: 32584


Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.4	17.2
183	11.8	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 43.186 ng
RT: 19.974 min Scan# 2874
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:202 Resp: 939110

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.0	25.4
203	17.2	13.8	20.6






#79
Terphenyl-d14
Concen: 86.242 ng
RT: 20.174 min Scan# 29
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:244 Resp: 152102


Ion	Ratio	Lower	Upper
244	100		
212	7.6	6.1	9.1
122	7.0	5.6	8.4



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#80
Butylbenzylphthalate
Concen: 45.601 ng
RT: 20.855 min Scan# 3024
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:149 Resp: 299706

Ion	Ratio	Lower	Upper
149	100		
91	74.0	59.2	88.8
206	22.6	18.1	27.1






#81
Benzo(a)anthracene
Concen: 42.723 ng
RT: 21.836 min Scan# 3
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:228 Resp: 94283


Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.9	32.9
229	19.5	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 43.131 ng
RT: 21.742 min Scan# 3175
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:252 Resp: 317851

Ion	Ratio	Lower	Upper
252	100		
254	63.5	50.8	76.2
126	7.2	5.8	8.6






#83
Chrysene
Concen: 42.613 ng
RT: 21.907 min Scan# 3
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:228 Resp: 89481


Ion	Ratio	Lower	Upper
228	100		
226	29.1	23.3	34.9
229	19.0	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 45.494 ng
RT: 21.748 min Scan# 3176
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:149 Resp: 446785

Ion	Ratio	Lower	Upper
149	100		
167	25.0	20.0	30.0
279	3.9	3.1	4.7






#85
Di-n-octyl phthalate
Concen: 43.688 ng
RT: 23.011 min Scan# 3
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:149 Resp: 72336


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	13.5	10.8	16.2



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 25.197 min Scan# 3763
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:264 Resp: 350897

Ion	Ratio	Lower	Upper
264	100		
260	23.7	19.0	28.4
265	22.0	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 42.624 ng
RT: 29.040 min Scan# 44
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:276 Resp: 110025


Ion	Ratio	Lower	Upper
276	100		
138	13.5	10.8	16.2
277	25.1	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 43.108 ng
RT: 24.139 min Scan# 3583
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:252 Resp: 927550

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.8	26.8
125	7.0	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 41.878 ng
RT: 24.210 min Scan# 3
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:252 Resp: 90170


Ion	Ratio	Lower	Upper
252	100		
253	21.0	16.8	25.2
125	6.7	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 42.635 ng
RT: 25.044 min Scan# 3737
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:252 Resp: 843258

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	7.6	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 42.430 ng
RT: 29.110 min Scan# 44
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:278 Resp: 889808


Ion	Ratio	Lower	Upper
278	100		
139	10.0	8.0	12.0
279	24.7	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 42.492 ng
RT: 30.232 min Scan# 4620
Delta R.T. 0.000 min
Lab File: BG064573.D
Acq: 24 Oct 2025 13:03

Tgt Ion:276 Resp: 851248

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.1	28.7
138	14.1	11.3	16.9



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064574.D
Acq On : 24 Oct 2025 13:44
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 18:01:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.219	152	40271	20.000	ng		0.00
21) Naphthalene-d8	11.045	136	168005	20.000	ng		0.00
39) Acenaphthene-d10	14.841	164	134994	20.000	ng		0.00
64) Phenanthrene-d10	17.579	188	324385	20.000	ng		0.00
76) Chrysene-d12	21.856	240	340896	20.000	ng		0.00
86) Perylene-d12	25.205	264	350916	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.746	112	318446	132.728	ng		0.00
7) Phenol-d6	7.355	99	450449	136.821	ng		0.00
23) Nitrobenzene-d5	9.388	82	406268	145.243	ng		0.00
42) 2,4,6-Tribromophenol	16.321	330	280190	143.807	ng		0.00
45) 2-Fluorobiphenyl	13.478	172	1120989	125.867	ng		0.00
79) Terphenyl-d14	20.176	244	2253704	124.763	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.589	88	67487	63.535	ng		98
3) Pyridine	4.001	79	209372	65.637	ng		98
4) n-Nitrosodimethylamine	3.907	42	111310	65.071	ng		94
6) Aniline	7.532	93	294791	65.803	ng		98
8) 2-Chlorophenol	7.779	128	176533	70.297	ng		98
9) Benzaldehyde	7.344	77	118559	65.963	ng		98
10) Phenol	7.379	94	244914	67.267	ng		98
11) bis(2-Chloroethyl)ether	7.626	93	174337	65.126	ng		100
12) 1,3-Dichlorobenzene	8.113	146	186290	64.452	ng		99
13) 1,4-Dichlorobenzene	8.254	146	188553	64.156	ng		95
14) 1,2-Dichlorobenzene	8.583	146	183593	64.766	ng		98
15) Benzyl Alcohol	8.448	79	185157	68.201	ng		97
16) 2,2'-oxybis(1-Chloropr...	8.742	45	459295	65.410	ng		100
17) 2-Methylphenol	8.648	107	162536	67.237	ng		98
18) Hexachloroethane	9.324	117	67237	65.193	ng		99
19) n-Nitroso-di-n-propyla...	9.030	70	155683	68.424	ng		99
20) 3+4-Methylphenols	8.983	107	228995	67.359	ng		98
22) Acetophenone	9.048	105	297413	64.759	ng	#	96
24) Nitrobenzene	9.430	77	220774	72.567	ng		98
25) Isophorone	9.958	82	449298	66.940	ng		99
26) 2-Nitrophenol	10.146	139	76368	63.945	ng		98
27) 2,4-Dimethylphenol	10.193	122	173061	69.317	ng		97
28) bis(2-Chloroethoxy)met...	10.440	93	248699	65.078	ng		99
29) 2,4-Dichlorophenol	10.681	162	188233	70.056	ng		96
30) 1,2,4-Trichlorobenzene	10.904	180	200988	65.837	ng		95
31) Naphthalene	11.098	128	605239	64.976	ng		99
32) Benzoic acid	10.340	122	130898m	65.612	ng		
33) 4-Chloroaniline	11.192	127	245101	67.699	ng		98
34) Hexachlorobutadiene	11.392	225	154303	64.630	ng		97
35) Caprolactam	11.968	113	71319	69.206	ng		96
36) 4-Chloro-3-methylphenol	12.297	107	228833	70.853	ng		100
37) 2-Methylnaphthalene	12.690	142	447497	65.435	ng		98
38) 1-Methylnaphthalene	12.908	142	442677	65.402	ng		99
40) 1,2,4,5-Tetrachloroben...	13.055	216	290173	64.049	ng		100
41) Hexachlorocyclopentadiene	13.037	237	137727	67.540	ng		99
43) 2,4,6-Trichlorophenol	13.278	196	184771	71.299	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064574.D
Acq On : 24 Oct 2025 13:44
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 18:01:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

	Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44)	2,4,5-Trichlorophenol	13.354	196	209320	69.994	ng		96
46)	1,1'-Biphenyl	13.683	154	606860	63.233	ng		98
47)	2-Chloronaphthalene	13.730	162	463792	64.022	ng		98
48)	2-Nitroaniline	13.913	65	163796	66.052	ng		95
49)	Acenaphthylene	14.571	152	739728	64.738	ng		100
50)	Dimethylphthalate	14.289	163	688128	66.616	ng		99
51)	2,6-Dinitrotoluene	14.400	165	114588	65.951	ng		95
52)	Acenaphthene	14.911	154	510499	65.392	ng		98
53)	3-Nitroaniline	14.729	138	131776	74.231	ng		91
54)	2,4-Dinitrophenol	14.929	184	66900	65.082	ng	#	75
55)	Dibenzofuran	15.240	168	804964	63.779	ng		97
56)	4-Nitrophenol	15.017	139	114408	69.512	ng		97
57)	2,4-Dinitrotoluene	15.187	165	175802	65.522	ng	#	100
58)	Fluorene	15.887	166	627437	63.656	ng		100
59)	2,3,4,6-Tetrachlorophenol	15.458	232	205529	71.870	ng		99
60)	Diethylphthalate	15.652	149	695924	65.982	ng		99
61)	4-Chlorophenyl-phenyle...	15.875	204	359759	65.038	ng		99
62)	4-Nitroaniline	15.893	138	143839	72.832	ng		98
63)	Azobenzene	16.169	77	633520	65.464	ng		99
65)	4,6-Dinitro-2-methylph...	15.945	198	94338	64.046	ng		93
66)	n-Nitrosodiphenylamine	16.086	169	577835	65.944	ng		99
67)	4-Bromophenyl-phenylether	16.768	248	256556	66.740	ng		100
68)	Hexachlorobenzene	16.891	284	293831	67.001	ng		96
69)	Atrazine	17.026	200	254070	67.378	ng		97
70)	Pentachlorophenol	17.226	266	193954	70.923	ng		91
71)	Phenanthrene	17.626	178	1147317	64.676	ng		99
72)	Anthracene	17.714	178	1171223	64.905	ng		99
73)	Carbazole	17.972	167	1000680	62.887	ng		99
74)	Di-n-butylphthalate	18.536	149	1165396	63.334	ng		99
75)	Fluoranthene	19.618	202	1364891	61.419	ng		98
77)	Benzidine	19.782	184	544918	71.348	ng		99
78)	Pyrene	19.976	202	1437669	64.549	ng		99
80)	Butylbenzylphthalate	20.857	149	477358	70.914	ng		98
81)	Benzo(a)anthracene	21.839	228	1462488	64.702	ng		99
82)	3,3'-Dichlorobenzidine	21.739	252	497174	65.868	ng		99
83)	Chrysene	21.909	228	1389821	64.621	ng		98
84)	Bis(2-ethylhexyl)phtha...	21.750	149	681649	67.768	ng		98
85)	Di-n-octyl phthalate	23.014	149	1142651	67.379	ng		100
87)	Indeno(1,2,3-cd)pyrene	29.048	276	1729035	66.979	ng		98
88)	Benzo(b)fluoranthene	24.142	252	1424640	66.207	ng		99
89)	Benzo(k)fluoranthene	24.218	252	1436312	66.703	ng		99
90)	Benzo(a)pyrene	25.052	252	1311701	66.315	ng		100
91)	Dibenzo(a,h)anthracene	29.124	278	1394234	66.479	ng		99
92)	Benzo(g,h,i)perylene	30.246	276	1347382	67.253	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064574.D
Acq On : 24 Oct 2025 13:44
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Nov 03 18:01:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.219 min Scan# 8
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:152 Resp: 4027


Ion	Ratio	Lower	Upper
152	100		
150	147.4	122.2	183.4
115	57.8	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 63.535 ng
RT: 3.589 min Scan# 85
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 88 Resp: 67487

Ion	Ratio	Lower	Upper
88	100		
58	113.3	92.4	138.6
43	40.0	33.0	49.6






#3
Pyridine
Concen: 65.637 ng
RT: 4.001 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 79 Resp: 20937


Ion	Ratio	Lower	Upper
79	100		
52	116.3	94.2	141.2
51	54.8	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 65.071 ng
RT: 3.907 min Scan# 139
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 42 Resp: 111310

Ion	Ratio	Lower	Upper
42	100		
74	100.7	76.1	114.1
44	8.7	6.3	9.5





#5
2-Fluorophenol
Concen: 132.728 ng
RT: 5.746 min Scan# 4
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44


Tgt Ion:112 Resp: 31844

Ion	Ratio	Lower	Upper
112	100		
64	77.2	60.7	91.1
63	43.1	33.4	50.0

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#6
Aniline
Concen: 65.803 ng
RT: 7.532 min Scan# 756
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 93 Resp: 294791


Ion	Ratio	Lower	Upper
93	100		
66	46.7	37.0	55.6
65	24.7	18.3	27.5



#7
Phenol-d6
Concen: 136.821 ng
RT: 7.355 min Scan# 7
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 99 Resp: 45044


Ion	Ratio	Lower	Upper
99	100		
42	24.4	18.3	27.5
71	36.7	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 70.297 ng
RT: 7.779 min Scan# 798
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:128 Resp: 176533

Ion	Ratio	Lower	Upper
128	100		
130	33.9	13.5	53.5
64	63.3	45.2	85.2






#9
Benzaldehyde
Concen: 65.963 ng
RT: 7.344 min Scan# 7
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 77 Resp: 11855


Ion	Ratio	Lower	Upper
77	100		
105	95.2	72.9	112.9
106	89.0	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 67.267 ng
RT: 7.379 min Scan# 730
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 94 Resp: 244914

Ion	Ratio	Lower	Upper
94	100		
65	35.3	15.8	55.8
66	45.9	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 65.126 ng
RT: 7.626 min Scan# 7
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 93 Resp: 17433


Ion	Ratio	Lower	Upper
93	100		
63	111.2	91.3	131.3
95	33.3	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 64.452 ng
RT: 8.113 min Scan# 855
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:146 Resp: 186290

Ion	Ratio	Lower	Upper
146	100		
148	68.4	55.3	82.9
75	36.1	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 64.156 ng
RT: 8.254 min Scan# 8
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:146 Resp: 18855


Ion	Ratio	Lower	Upper
146	100		
148	65.5	56.2	84.4
111	41.2	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 64.766 ng
RT: 8.583 min Scan# 935
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:146 Resp: 183593

Ion	Ratio	Lower	Upper
146	100		
148	65.6	51.5	77.3
111	45.4	35.0	52.6





#15
Benzyl Alcohol
Concen: 68.201 ng
RT: 8.448 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44


Tgt Ion: 79 Resp: 18515


Ion	Ratio	Lower	Upper
79	100		
108	71.1	58.1	87.1
77	61.8	52.8	79.2

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 65.410 ng
RT: 8.742 min Scan# 962
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 45 Resp: 459295

Ion	Ratio	Lower	Upper
45	100		
77	8.3	0.0	28.5
79	5.6	0.0	25.6






#17
2-Methylphenol
Concen: 67.237 ng
RT: 8.648 min Scan# 94
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:107 Resp: 16253


Ion	Ratio	Lower	Upper
107	100		
108	111.1	90.0	135.0
77	48.2	37.8	56.8
79	48.9	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 65.193 ng
RT: 9.324 min Scan# 1061
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:117 Resp: 67237

Ion	Ratio	Lower	Upper
117	100		
119	93.9	75.4	113.2
201	126.1	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 68.424 ng
RT: 9.030 min Scan# 10
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 70 Resp: 15568


Ion	Ratio	Lower	Upper
70	100		
42	80.9	65.2	97.8
101	9.4	7.9	11.9
130	15.8	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 67.359 ng
RT: 8.983 min Scan# 1003
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:107 Resp: 228995

Ion	Ratio	Lower	Upper
107	100		
108	88.4	70.3	110.3
77	35.9	14.6	54.6
79	31.5	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.045 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:136 Resp: 16800


Ion	Ratio	Lower	Upper
136	100		
137	10.8	9.2	13.8
54	15.3	12.6	19.0
68	6.2	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 64.759 ng
RT: 9.048 min Scan# 1014
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:105 Resp: 297413

Ion	Ratio	Lower	Upper
105	100		
71	2.6	1.5	2.3#
51	39.0	33.7	50.5
120	19.4	16.1	24.1






#23
Nitrobenzene-d5
Concen: 145.243 ng
RT: 9.388 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 82 Resp: 40626


Ion	Ratio	Lower	Upper
82	100		
128	38.5	31.3	46.9
54	73.8	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 72.567 ng
RT: 9.430 min Scan# 1079
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 77 Resp: 220774

Ion	Ratio	Lower	Upper
77	100		
123	40.7	31.0	46.6
65	16.3	12.7	19.1






#25
Isophorone
Concen: 66.940 ng
RT: 9.958 min Scan# 11
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 82 Resp: 449298


Ion	Ratio	Lower	Upper
82	100		
95	7.3	6.1	9.1
138	15.6	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 63.945 ng
RT: 10.146 min Scan# 1201
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:139 Resp: 76368

Ion	Ratio	Lower	Upper
139	100		
109	32.9	24.6	37.0
65	60.1	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 69.317 ng
RT: 10.193 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:122 Resp: 17306


Ion	Ratio	Lower	Upper
122	100		
107	123.8	103.0	154.4
121	61.1	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 65.078 ng
RT: 10.440 min Scan# 1251
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 93 Resp: 248699

Ion	Ratio	Lower	Upper
93	100		
95	32.5	25.4	38.2
123	9.9	8.4	12.6





#29
2,4-Dichlorophenol
Concen: 70.056 ng
RT: 10.681 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44


Tgt Ion:162 Resp: 18823


Ion	Ratio	Lower	Upper
162	100		
164	64.9	48.7	88.7
98	35.2	16.9	56.9

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#30
1,2,4-Trichlorobenzene
Concen: 65.837 ng
RT: 10.904 min Scan# 1330
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:180 Resp: 200988

Ion	Ratio	Lower	Upper
180	100		
182	89.9	76.9	115.3
145	28.5	22.1	33.1






#31
Naphthalene
Concen: 64.976 ng
RT: 11.098 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:128 Resp: 60523


Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.8	13.2
127	13.6	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 65.612 ng m
RT: 10.340 min Scan# 1234
Delta R.T. 0.031 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:122 Resp: 130898

Ion	Ratio	Lower	Upper
122	100		
105	128.3	104.2	144.2
77	99.8	70.2	110.2






#33
4-Chloroaniline
Concen: 67.699 ng
RT: 11.192 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:127 Resp: 245101


Ion	Ratio	Lower	Upper
127	100		
129	32.9	25.5	38.3
65	39.7	32.2	48.4
92	20.6	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 64.630 ng
RT: 11.392 min Scan# 1413
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:225 Resp: 154303

Ion	Ratio	Lower	Upper
225	100		
223	61.4	48.8	73.2
227	65.7	49.5	74.3






#35
Caprolactam
Concen: 69.206 ng
RT: 11.968 min Scan# 1
Delta R.T. 0.020 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:113 Resp: 7131


Ion	Ratio	Lower	Upper
113	100		
55	314.3	290.1	330.1
56	213.7	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 70.853 ng
RT: 12.297 min Scan# 1567
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:107 Resp: 228833

Ion	Ratio	Lower	Upper
107	100		
144	24.1	19.1	28.7
142	72.8	58.6	87.8






#37
2-Methylnaphthalene
Concen: 65.435 ng
RT: 12.690 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:142 Resp: 44749


Ion	Ratio	Lower	Upper
142	100		
141	87.4	68.9	103.3
115	36.7	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 65.402 ng
RT: 12.908 min Scan# 1671
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:142 Resp: 442677

Ion	Ratio	Lower	Upper
142	100		
141	90.4	73.3	109.9
116	4.0	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.841 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:164 Resp: 13499


Ion	Ratio	Lower	Upper
164	100		
162	102.5	77.0	115.6
160	43.9	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 64.049 ng
RT: 13.055 min Scan# 1696
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:216 Resp: 290173

Ion	Ratio	Lower	Upper
216	100		
214	77.0	61.4	92.2
179	18.8	15.4	23.0
108	14.5	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 67.540 ng
RT: 13.037 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:237 Resp: 13772


Ion	Ratio	Lower	Upper
237	100		
235	64.6	44.1	84.1
272	12.9	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 143.807 ng
RT: 16.321 min Scan# 2252
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:330 Resp: 280190

Ion	Ratio	Lower	Upper
330	100		
332	95.9	78.0	117.0
141	27.0	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 71.299 ng
RT: 13.278 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:196 Resp: 18477


Ion	Ratio	Lower	Upper
196	100		
198	95.9	76.6	115.0
200	30.4	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 69.994 ng
RT: 13.354 min Scan# 1747
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:196 Resp: 209320

Ion	Ratio	Lower	Upper
196	100		
198	89.8	74.2	111.4
97	42.2	37.0	55.6
132	22.0	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 125.867 ng
RT: 13.478 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:172 Resp: 1120989


Ion	Ratio	Lower	Upper
172	100		
171	37.4	29.4	44.0
170	24.6	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 63.233 ng
RT: 13.683 min Scan# 1803
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:154 Resp: 606860

Ion	Ratio	Lower	Upper
154	100		
153	41.4	22.4	62.4
76	12.9	0.0	33.5






#47
2-Chloronaphthalene
Concen: 64.022 ng
RT: 13.730 min Scan# 18
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:162 Resp: 463791


Ion	Ratio	Lower	Upper
162	100		
127	38.8	32.2	48.2
164	34.0	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 66.052 ng
RT: 13.913 min Scan# 1842
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 65 Resp: 163796

Ion	Ratio	Lower	Upper
65	100		
92	62.2	48.3	72.5
138	87.6	74.9	112.3






#49
Acenaphthylene
Concen: 64.738 ng
RT: 14.571 min Scan# 19
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:152 Resp: 73972


Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.4	23.2
153	13.5	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 66.616 ng
RT: 14.289 min Scan# 1906
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:163 Resp: 688128

Ion	Ratio	Lower	Upper
163	100		
194	3.6	2.8	4.2
164	10.1	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 65.951 ng
RT: 14.400 min Scan# 19
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:165 Resp: 114588


Ion	Ratio	Lower	Upper
165	100		
63	73.9	57.4	86.0
89	65.4	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 65.392 ng
RT: 14.911 min Scan# 2012
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:154 Resp: 510499

Ion	Ratio	Lower	Upper
154	100		
153	112.5	92.2	138.4
152	53.8	43.2	64.8






#53
3-Nitroaniline
Concen: 74.231 ng
RT: 14.729 min Scan# 19
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:138 Resp: 13177


Ion	Ratio	Lower	Upper
138	100		
108	13.1	11.4	17.0
92	130.0	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 65.082 ng
RT: 14.929 min Scan# 2015
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:184 Resp: 66900

Ion	Ratio	Lower	Upper
184	100		
63	68.5	56.6	85.0
154	149.4	158.6	238.0#






#55
Dibenzofuran
Concen: 63.779 ng
RT: 15.240 min Scan# 2
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:168 Resp: 80496


Ion	Ratio	Lower	Upper
168	100		
139	36.7	27.8	41.8
169	13.4	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 69.512 ng
RT: 15.017 min Scan# 2030
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:139 Resp: 114408

Ion	Ratio	Lower	Upper
139	100		
109	93.0	67.8	107.8
65	125.8	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 65.522 ng
RT: 15.187 min Scan# 2
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:165 Resp: 17580


Ion	Ratio	Lower	Upper
165	100		
63	52.0	41.8	62.8
89	72.9	58.1	87.1
182	2.1	1.4	2.0#



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 63.656 ng
RT: 15.887 min Scan# 2178
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:166 Resp: 627437

Ion	Ratio	Lower	Upper
166	100		
165	101.8	81.5	122.3
167	13.5	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 71.870 ng
RT: 15.458 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:232 Resp: 20552


Ion	Ratio	Lower	Upper
232	100		
131	34.7	26.7	40.1
130	2.0	1.6	2.4
166	25.6	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 65.982 ng
RT: 15.652 min Scan# 2138
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:149 Resp: 695924

Ion	Ratio	Lower	Upper
149	100		
177	22.2	18.0	27.0
150	12.4	10.2	15.4





#61
4-Chlorophenyl-phenylether
Concen: 65.038 ng
RT: 15.875 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44


Tgt Ion:204 Resp: 359759


Ion	Ratio	Lower	Upper
204	100		
206	33.3	25.6	38.4
141	50.4	40.6	60.8

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#62
4-Nitroaniline
Concen: 72.832 ng
RT: 15.893 min Scan# 2179
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:138 Resp: 143839

Ion	Ratio	Lower	Upper
138	100		
92	57.1	36.1	76.1
108	98.4	75.6	115.6






#63
Azobenzene
Concen: 65.464 ng
RT: 16.169 min Scan# 2
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion: 77 Resp: 63352


Ion	Ratio	Lower	Upper
77	100		
182	25.5	4.9	44.9
105	20.1	1.3	41.3
51	39.5	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.579 min Scan# 2466
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:188 Resp: 324385

Ion	Ratio	Lower	Upper
188	100		
94	7.2	5.7	8.5
80	9.0	6.8	10.2






#65
4,6-Dinitro-2-methylphenol
Concen: 64.046 ng
RT: 15.945 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:198 Resp: 9433


Ion	Ratio	Lower	Upper
198	100		
51	50.7	24.7	64.7
105	36.0	18.0	58.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#66
n-Nitrosodiphenylamine
Concen: 65.944 ng
RT: 16.086 min Scan# 2212
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:169 Resp: 577835

Ion	Ratio	Lower	Upper
169	100		
168	70.9	56.3	84.5
167	38.0	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 66.740 ng
RT: 16.768 min Scan# 2
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:248 Resp: 25655


Ion	Ratio	Lower	Upper
248	100		
250	96.6	77.7	116.5
141	60.8	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 67.001 ng
RT: 16.891 min Scan# 2349
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:284 Resp: 293831

Ion	Ratio	Lower	Upper
284	100		
142	25.6	21.0	31.6
249	34.5	24.9	37.3






#69
Atrazine
Concen: 67.378 ng
RT: 17.026 min Scan# 2
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:200 Resp: 25407


Ion	Ratio	Lower	Upper
200	100		
173	23.6	0.9	40.9
215	47.9	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 70.923 ng
RT: 17.226 min Scan# 2406
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:266 Resp: 193954

Ion	Ratio	Lower	Upper
266	100		
268	66.0	50.3	75.5
264	69.3	47.3	70.9






#71
Phenanthrene
Concen: 64.676 ng
RT: 17.626 min Scan# 24
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:178 Resp: 114731


Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.7	22.1
179	15.4	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 64.905 ng
RT: 17.714 min Scan# 2489
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:178 Resp: 1171223

Ion	Ratio	Lower	Upper
178	100		
176	18.2	14.4	21.6
179	15.8	12.4	18.6






#73
Carbazole
Concen: 62.887 ng
RT: 17.972 min Scan# 2
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:167 Resp: 100068


Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.9	26.9
139	12.7	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 63.334 ng
RT: 18.536 min Scan# 2629
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:149 Resp: 1165396

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.0	10.6
104	5.6	4.1	6.1






#75
Fluoranthene
Concen: 61.419 ng
RT: 19.618 min Scan# 28
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:202 Resp: 136489


Ion	Ratio	Lower	Upper
202	100		
101	9.0	0.0	28.7
203	17.8	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.856 min Scan# 3194
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:240 Resp: 340896

Ion	Ratio	Lower	Upper
240	100		
120	7.1	6.5	9.7
236	24.9	20.2	30.2






#77
Benzidine
Concen: 71.348 ng
RT: 19.782 min Scan# 28
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:184 Resp: 544918


Ion	Ratio	Lower	Upper
184	100		
185	13.9	11.4	17.2
183	11.8	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 64.549 ng
RT: 19.976 min Scan# 2874
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:202 Resp: 1437669

Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.0	25.4
203	17.7	13.8	20.6






#79
Terphenyl-d14
Concen: 124.763 ng
RT: 20.176 min Scan# 29
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:244 Resp: 225370


Ion	Ratio	Lower	Upper
244	100		
212	8.0	6.1	9.1
122	7.1	5.6	8.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#80
Butylbenzylphthalate
Concen: 70.914 ng
RT: 20.857 min Scan# 3024
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:149 Resp: 477358

Ion	Ratio	Lower	Upper
149	100		
91	72.0	59.2	88.8
206	22.8	18.1	27.1






#81
Benzo(a)anthracene
Concen: 64.702 ng
RT: 21.839 min Scan# 3
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:228 Resp: 146248


Ion	Ratio	Lower	Upper
228	100		
226	27.8	21.9	32.9
229	20.2	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 65.868 ng
RT: 21.739 min Scan# 3174
Delta R.T. -0.004 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:252 Resp: 497174

Ion	Ratio	Lower	Upper
252	100		
254	62.6	50.8	76.2
126	7.4	5.8	8.6






#83
Chrysene
Concen: 64.621 ng
RT: 21.909 min Scan# 3
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:228 Resp: 138982


Ion	Ratio	Lower	Upper
228	100		
226	30.2	23.3	34.9
229	19.9	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 67.768 ng
RT: 21.750 min Scan# 3176
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:149 Resp: 681649

Ion	Ratio	Lower	Upper
149	100		
167	25.9	20.0	30.0
279	4.4	3.1	4.7






#85
Di-n-octyl phthalate
Concen: 67.379 ng
RT: 23.014 min Scan# 3
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:149 Resp: 114265


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	13.5	10.8	16.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 25.205 min Scan# 3764
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:264 Resp: 350916

Ion	Ratio	Lower	Upper
264	100		
260	23.2	19.0	28.4
265	23.5	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 66.979 ng
RT: 29.048 min Scan# 44
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:276 Resp: 172903


Ion	Ratio	Lower	Upper
276	100		
138	11.2	10.8	16.2
277	24.9	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 66.207 ng
RT: 24.142 min Scan# 3583
Delta R.T. 0.002 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:252 Resp: 1424640

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.8	26.8
125	6.6	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 66.703 ng
RT: 24.218 min Scan# 3
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:252 Resp: 143631


Ion	Ratio	Lower	Upper
252	100		
253	21.3	16.8	25.2
125	6.9	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 66.315 ng
RT: 25.052 min Scan# 3738
Delta R.T. 0.008 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:252 Resp: 1311701

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	7.7	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 66.479 ng
RT: 29.124 min Scan# 44
Delta R.T. 0.014 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:278 Resp: 139423


Ion	Ratio	Lower	Upper
278	100		
139	10.7	8.0	12.0
279	24.1	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 67.253 ng
RT: 30.246 min Scan# 4622
Delta R.T. 0.014 min
Lab File: BG064574.D
Acq: 24 Oct 2025 13:44

Tgt Ion:276 Resp: 1347382

Ion	Ratio	Lower	Upper
276	100		
277	23.0	19.1	28.7
138	14.5	11.3	16.9



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064575.D
Acq On : 24 Oct 2025 14:24
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Nov 03 18:02:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.221	152	45466	20.000	ng		0.00
21) Naphthalene-d8	11.053	136	179569	20.000	ng		0.00
39) Acenaphthene-d10	14.842	164	133797	20.000	ng		0.00
64) Phenanthrene-d10	17.580	188	333421	20.000	ng		0.00
76) Chrysene-d12	21.863	240	363291	20.000	ng		0.00
86) Perylene-d12	25.201	264	382372	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.747	112	411587	151.948	ng		0.00
7) Phenol-d6	7.357	99	554287	149.124	ng		0.00
23) Nitrobenzene-d5	9.390	82	511341	171.034	ng		0.00
42) 2,4,6-Tribromophenol	16.323	330	328066	169.886	ng		0.00
45) 2-Fluorobiphenyl	13.479	172	1314892	148.960	ng		0.00
79) Terphenyl-d14	20.177	244	2697491	140.125	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.591	88	83655	69.757	ng		99
3) Pyridine	3.996	79	259519	72.062	ng		96
4) n-Nitrosodimethylamine	3.908	42	141185	73.105	ng		99
6) Aniline	7.533	93	364528	72.073	ng		99
8) 2-Chlorophenol	7.780	128	221112	77.988	ng		99
9) Benzaldehyde	7.339	77	120456	59.361	ng		94
10) Phenol	7.386	94	307886	74.900	ng		98
11) bis(2-Chloroethyl)ether	7.627	93	217960	72.119	ng		96
12) 1,3-Dichlorobenzene	8.109	146	236208	72.385	ng		98
13) 1,4-Dichlorobenzene	8.256	146	240570	72.502	ng		97
14) 1,2-Dichlorobenzene	8.579	146	233326	72.905	ng		98
15) Benzyl Alcohol	8.456	79	233374	76.140	ng		95
16) 2,2'-oxybis(1-Chloropr...	8.743	45	573671	72.364	ng		99
17) 2-Methylphenol	8.649	107	203667	74.625	ng		97
18) Hexachloroethane	9.325	117	87876	75.469	ng		98
19) n-Nitroso-di-n-propyla...	9.031	70	192155	74.804	ng		99
20) 3+4-Methylphenols	8.984	107	285690	74.434	ng		98
22) Acetophenone	9.049	105	365265	74.412	ng	#	99
24) Nitrobenzene	9.431	77	273231	84.025	ng		97
25) Isophorone	9.966	82	539122	75.150	ng		100
26) 2-Nitrophenol	10.148	139	103090	79.938	ng		94
27) 2,4-Dimethylphenol	10.195	122	208054	77.967	ng		97
28) bis(2-Chloroethoxy)met...	10.436	93	299758	73.388	ng		99
29) 2,4-Dichlorophenol	10.682	162	231256	80.526	ng		97
30) 1,2,4-Trichlorobenzene	10.906	180	250883	76.888	ng		97
31) Naphthalene	11.100	128	744985	74.828	ng		99
32) Benzoic acid	10.342	122	169962m	77.917	ng		
33) 4-Chloroaniline	11.188	127	290611	75.100	ng		98
34) Hexachlorobutadiene	11.393	225	196140	76.863	ng		99
35) Caprolactam	11.975	113	83266	75.596	ng		93
36) 4-Chloro-3-methylphenol	12.298	107	265909	77.031	ng		99
37) 2-Methylnaphthalene	12.692	142	542132	74.168	ng		99
38) 1-Methylnaphthalene	12.909	142	527717	72.945	ng		98
40) 1,2,4,5-Tetrachloroben...	13.056	216	351719	78.328	ng		99
41) Hexachlorocyclopentadiene	13.038	237	163461	80.876	ng		99
43) 2,4,6-Trichlorophenol	13.279	196	218318	84.998	ng		98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064575.D
Acq On : 24 Oct 2025 14:24
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 18:02:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.350	196	243241	82.064	ng		99
46) 1,1'-Biphenyl	13.685	154	713186	74.977	ng		100
47) 2-Chloronaphthalene	13.732	162	547242	76.218	ng		99
48) 2-Nitroaniline	13.914	65	196580	79.301	ng		95
49) Acenaphthylene	14.566	152	858511	75.806	ng		99
50) Dimethylphthalate	14.290	163	782708	76.450	ng		98
51) 2,6-Dinitrotoluene	14.402	165	135769	78.216	ng		98
52) Acenaphthene	14.913	154	589213	76.150	ng		99
53) 3-Nitroaniline	14.731	138	159072	90.409	ng		96
54) 2,4-Dinitrophenol	14.930	184	82655	79.755	ng	#	65
55) Dibenzofuran	15.242	168	923386	73.816	ng		99
56) 4-Nitrophenol	15.018	139	142966	87.640	ng		98
57) 2,4-Dinitrotoluene	15.189	165	212164	79.117	ng		96
58) Fluorene	15.888	166	711212	72.801	ng		98
59) 2,3,4,6-Tetrachlorophenol	15.459	232	240094	84.708	ng		99
60) Diethylphthalate	15.653	149	803621	76.874	ng		100
61) 4-Chlorophenyl-phenyle...	15.876	204	408239	74.463	ng		99
62) 4-Nitroaniline	15.894	138	173159	88.462	ng		96
63) Azobenzene	16.170	77	723988	75.482	ng		99
65) 4,6-Dinitro-2-methylph...	15.947	198	124101	80.450	ng		95
66) n-Nitrosodiphenylamine	16.088	169	666665	74.020	ng		98
67) 4-Bromophenyl-phenylether	16.769	248	299348	75.762	ng		99
68) Hexachlorobenzene	16.893	284	341338	75.724	ng		96
69) Atrazine	17.028	200	281950	72.745	ng		97
70) Pentachlorophenol	17.228	266	235472	83.771	ng		93
71) Phenanthrene	17.627	178	1340610	73.525	ng		99
72) Anthracene	17.715	178	1367044	73.704	ng		99
73) Carbazole	17.974	167	1195960	73.122	ng		99
74) Di-n-butylphthalate	18.538	149	1433446	75.790	ng		100
75) Fluoranthene	19.619	202	1663673	72.835	ng		98
77) Benzidine	19.783	184	556525	68.376	ng		98
78) Pyrene	19.977	202	1737788	73.215	ng		99
80) Butylbenzylphthalate	20.859	149	595333	82.988	ng		98
81) Benzo(a)anthracene	21.840	228	1805628	74.959	ng		99
82) 3,3'-Dichlorobenzidine	21.740	252	615546	76.524	ng		99
83) Chrysene	21.910	228	1704533	74.368	ng		97
84) Bis(2-ethylhexyl)phtha...	21.752	149	854027	79.672	ng		98
85) Di-n-octyl phthalate	23.015	149	1432956	79.289	ng		100
87) Indeno(1,2,3-cd)pyrene	29.061	276	2156428	76.663	ng		98
88) Benzo(b)fluoranthene	24.143	252	1810094	77.199	ng		99
89) Benzo(k)fluoranthene	24.219	252	1757758	74.916	ng		98
90) Benzo(a)pyrene	25.060	252	1649779	76.546	ng		99
91) Dibenzo(a,h)anthracene	29.137	278	1747890	76.486	ng		98
92) Benzo(g,h,i)perylene	30.265	276	1669976	76.498	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064575.D
Acq On : 24 Oct 2025 14:24
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 8 Sample Multiplier: 1


Quant Time: Nov 03 18:02:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.221 min Scan# 8
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:152 Resp: 4546


Ion	Ratio	Lower	Upper
152	100		
150	150.5	122.2	183.4
115	61.3	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 69.757 ng
RT: 3.591 min Scan# 85
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 88 Resp: 83655

Ion	Ratio	Lower	Upper
88	100		
58	114.3	92.4	138.6
43	42.2	33.0	49.6






#3
Pyridine
Concen: 72.062 ng
RT: 3.996 min Scan# 1
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 79 Resp: 25951


Ion	Ratio	Lower	Upper
79	100		
52	122.7	94.2	141.2
51	56.2	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 73.105 ng
RT: 3.908 min Scan# 139
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 42 Resp: 141185

Ion	Ratio	Lower	Upper
42	100		
74	96.0	76.1	114.1
44	8.7	6.3	9.5





#5
2-Fluorophenol
Concen: 151.948 ng
RT: 5.747 min Scan# 4
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24


Tgt Ion:112 Resp: 41158


Ion	Ratio	Lower	Upper
112	100		
64	80.3	60.7	91.1
63	43.5	33.4	50.0

Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#6
Aniline
Concen: 72.073 ng
RT: 7.533 min Scan# 756
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 93 Resp: 364528

Ion	Ratio	Lower	Upper
93	100		
66	46.6	37.0	55.6
65	24.3	18.3	27.5






#7
Phenol-d6
Concen: 149.124 ng
RT: 7.357 min Scan# 7
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 99 Resp: 55428


Ion	Ratio	Lower	Upper
99	100		
42	24.9	18.3	27.5
71	37.8	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 77.988 ng
RT: 7.780 min Scan# 798
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:128 Resp: 221112

Ion	Ratio	Lower	Upper
128	100		
130	32.7	13.5	53.5
64	65.9	45.2	85.2






#9
Benzaldehyde
Concen: 59.361 ng
RT: 7.339 min Scan# 7
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 77 Resp: 12045


Ion	Ratio	Lower	Upper
77	100		
105	88.4	72.9	112.9
106	84.2	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 74.900 ng
RT: 7.386 min Scan# 731
Delta R.T. 0.009 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 94 Resp: 307886

Ion	Ratio	Lower	Upper
94	100		
65	36.2	15.8	55.8
66	46.0	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 72.119 ng
RT: 7.627 min Scan# 7
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 93 Resp: 21796


Ion	Ratio	Lower	Upper
93	100		
63	106.7	91.3	131.3
95	31.7	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 72.385 ng
RT: 8.109 min Scan# 854
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:146 Resp: 236208

Ion	Ratio	Lower	Upper
146	100		
148	67.1	55.3	82.9
75	36.1	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 72.502 ng
RT: 8.256 min Scan# 8
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:146 Resp: 24057


Ion	Ratio	Lower	Upper
146	100		
148	66.4	56.2	84.4
111	43.2	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 72.905 ng
RT: 8.579 min Scan# 934
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:146 Resp: 233326

Ion	Ratio	Lower	Upper
146	100		
148	67.2	51.5	77.3
111	43.8	35.0	52.6






#15
Benzyl Alcohol
Concen: 76.140 ng
RT: 8.456 min Scan# 91
Delta R.T. 0.009 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 79 Resp: 23337


Ion	Ratio	Lower	Upper
79	100		
108	67.5	58.1	87.1
77	62.4	52.8	79.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 72.364 ng
RT: 8.743 min Scan# 962
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 45 Resp: 573671

Ion	Ratio	Lower	Upper
45	100		
77	8.4	0.0	28.5
79	5.9	0.0	25.6






#17
2-Methylphenol
Concen: 74.625 ng
RT: 8.649 min Scan# 94
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:107 Resp: 20366


Ion	Ratio	Lower	Upper
107	100		
108	115.8	90.0	135.0
77	50.4	37.8	56.8
79	50.6	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 75.469 ng
RT: 9.325 min Scan# 1061
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:117 Resp: 87876

Ion	Ratio	Lower	Upper
117	100		
119	92.8	75.4	113.2
201	122.1	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 74.804 ng
RT: 9.031 min Scan# 10
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 70 Resp: 19215


Ion	Ratio	Lower	Upper
70	100		
42	80.5	65.2	97.8
101	9.3	7.9	11.9
130	17.1	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 74.434 ng
RT: 8.984 min Scan# 1003
Delta R.T. 0.009 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:107 Resp: 285690

Ion	Ratio	Lower	Upper
107	100		
108	88.9	70.3	110.3
77	36.2	14.6	54.6
79	32.3	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.053 min Scan# 1
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:136 Resp: 179569


Ion	Ratio	Lower	Upper
136	100		
137	10.7	9.2	13.8
54	14.0	12.6	19.0
68	6.7	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 74.412 ng
RT: 9.049 min Scan# 1014
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:105 Resp: 365265

Ion	Ratio	Lower	Upper
105	100		
71	3.4	1.5	2.3#
51	41.4	33.7	50.5
120	20.8	16.1	24.1






#23
Nitrobenzene-d5
Concen: 171.034 ng
RT: 9.390 min Scan# 10
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 82 Resp: 511341


Ion	Ratio	Lower	Upper
82	100		
128	40.8	31.3	46.9
54	73.3	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 84.025 ng
RT: 9.431 min Scan# 1079
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 77 Resp: 273231

Ion	Ratio	Lower	Upper
77	100		
123	40.7	31.0	46.6
65	16.8	12.7	19.1






#25
Isophorone
Concen: 75.150 ng
RT: 9.966 min Scan# 11
Delta R.T. 0.009 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 82 Resp: 539121


Ion	Ratio	Lower	Upper
82	100		
95	7.5	6.1	9.1
138	16.0	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 79.938 ng
RT: 10.148 min Scan# 1201
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:139 Resp: 103090

Ion	Ratio	Lower	Upper
139	100		
109	34.1	24.6	37.0
65	63.6	47.5	71.3





#27
2,4-Dimethylphenol
Concen: 77.967 ng
RT: 10.195 min Scan# 1
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24


Tgt Ion:122 Resp: 20805

Ion	Ratio	Lower	Upper
122	100		
107	125.0	103.0	154.4
121	58.5	47.8	71.8

Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#28
bis(2-Chloroethoxy)methane
Concen: 73.388 ng
RT: 10.436 min Scan# 1250
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 93 Resp: 299758


Ion	Ratio	Lower	Upper
93	100		
95	32.0	25.4	38.2
123	9.7	8.4	12.6



#29
2,4-Dichlorophenol
Concen: 80.526 ng
RT: 10.682 min Scan# 1
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:162 Resp: 23125


Ion	Ratio	Lower	Upper
162	100		
164	66.5	48.7	88.7
98	34.8	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 76.888 ng
RT: 10.906 min Scan# 1330
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:180 Resp: 250883

Ion	Ratio	Lower	Upper
180	100		
182	92.0	76.9	115.3
145	27.2	22.1	33.1






#31
Naphthalene
Concen: 74.828 ng
RT: 11.100 min Scan# 1
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:128 Resp: 74498


Ion	Ratio	Lower	Upper
128	100		
129	10.4	8.8	13.2
127	13.1	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 77.917 ng m
RT: 10.342 min Scan# 1234
Delta R.T. 0.033 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:122 Resp: 169962

Ion	Ratio	Lower	Upper
122	100		
105	124.4	104.2	144.2
77	100.9	70.2	110.2






#33
4-Chloroaniline
Concen: 75.100 ng
RT: 11.188 min Scan# 1
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:127 Resp: 29061


Ion	Ratio	Lower	Upper
127	100		
129	29.5	25.5	38.3
65	40.5	32.2	48.4
92	21.9	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 76.863 ng
RT: 11.393 min Scan# 1413
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:225 Resp: 196140

Ion	Ratio	Lower	Upper
225	100		
223	60.7	48.8	73.2
227	62.8	49.5	74.3






#35
Caprolactam
Concen: 75.596 ng
RT: 11.975 min Scan# 1
Delta R.T. 0.027 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:113 Resp: 8326


Ion	Ratio	Lower	Upper
113	100		
55	318.6	290.1	330.1
56	221.7	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 77.031 ng
RT: 12.298 min Scan# 1567
Delta R.T. 0.009 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:107 Resp: 265909

Ion	Ratio	Lower	Upper
107	100		
144	23.4	19.1	28.7
142	74.0	58.6	87.8






#37
2-Methylnaphthalene
Concen: 74.168 ng
RT: 12.692 min Scan# 1
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:142 Resp: 54213


Ion	Ratio	Lower	Upper
142	100		
141	87.2	68.9	103.3
115	39.6	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 72.945 ng
RT: 12.909 min Scan# 1671
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:142 Resp: 527717

Ion	Ratio	Lower	Upper
142	100		
141	90.1	73.3	109.9
116	4.1	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.842 min Scan# 2
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:164 Resp: 13379


Ion	Ratio	Lower	Upper
164	100		
162	98.9	77.0	115.6
160	43.9	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 78.328 ng
RT: 13.056 min Scan# 1696
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:216 Resp: 351719

Ion	Ratio	Lower	Upper
216	100		
214	76.1	61.4	92.2
179	19.3	15.4	23.0
108	14.3	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 80.876 ng
RT: 13.038 min Scan# 1
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:237 Resp: 16346


Ion	Ratio	Lower	Upper
237	100		
235	63.6	44.1	84.1
272	12.3	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 169.886 ng
RT: 16.323 min Scan# 2252
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:330 Resp: 328066

Ion	Ratio	Lower	Upper
330	100		
332	96.4	78.0	117.0
141	27.1	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 84.998 ng
RT: 13.279 min Scan# 1
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:196 Resp: 21831


Ion	Ratio	Lower	Upper
196	100		
198	94.0	76.6	115.0
200	31.2	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 82.064 ng
RT: 13.350 min Scan# 1746
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:196 Resp: 243241

Ion	Ratio	Lower	Upper
196	100		
198	94.3	74.2	111.4
97	46.3	37.0	55.6
132	23.4	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 148.960 ng
RT: 13.479 min Scan# 1
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:172 Resp: 131489


Ion	Ratio	Lower	Upper
172	100		
171	36.9	29.4	44.0
170	24.9	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 74.977 ng
RT: 13.685 min Scan# 1803
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:154 Resp: 713186

Ion	Ratio	Lower	Upper
154	100		
153	42.2	22.4	62.4
76	13.2	0.0	33.5






#47
2-Chloronaphthalene
Concen: 76.218 ng
RT: 13.732 min Scan# 18
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:162 Resp: 54724


Ion	Ratio	Lower	Upper
162	100		
127	41.1	32.2	48.2
164	33.7	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 79.301 ng
RT: 13.914 min Scan# 1842
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 65 Resp: 196580

Ion	Ratio	Lower	Upper
65	100		
92	59.1	48.3	72.5
138	86.4	74.9	112.3






#49
Acenaphthylene
Concen: 75.806 ng
RT: 14.566 min Scan# 19
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:152 Resp: 858511


Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.4	23.2
153	13.5	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 76.450 ng
RT: 14.290 min Scan# 1906
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:163 Resp: 782708

Ion	Ratio	Lower	Upper
163	100		
194	3.4	2.8	4.2
164	10.5	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 78.216 ng
RT: 14.402 min Scan# 19
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:165 Resp: 135769


Ion	Ratio	Lower	Upper
165	100		
63	74.1	57.4	86.0
89	59.7	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 76.150 ng
RT: 14.913 min Scan# 2012
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:154 Resp: 589213

Ion	Ratio	Lower	Upper
154	100		
153	117.2	92.2	138.4
152	54.2	43.2	64.8






#53
3-Nitroaniline
Concen: 90.409 ng
RT: 14.731 min Scan# 19
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:138 Resp: 15907


Ion	Ratio	Lower	Upper
138	100		
108	12.0	11.4	17.0
92	122.9	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 79.755 ng
RT: 14.930 min Scan# 2015
Delta R.T. 0.003 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:184 Resp: 82655

Ion	Ratio	Lower	Upper
184	100		
63	69.0	56.6	85.0
154	128.4	158.6	238.0#






#55
Dibenzofuran
Concen: 73.816 ng
RT: 15.242 min Scan# 20
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:168 Resp: 92338


Ion	Ratio	Lower	Upper
168	100		
139	35.8	27.8	41.8
169	13.0	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 87.640 ng
RT: 15.018 min Scan# 2030
Delta R.T. 0.009 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:139 Resp: 142966

Ion	Ratio	Lower	Upper
139	100		
109	91.9	67.8	107.8
65	123.7	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 79.117 ng
RT: 15.189 min Scan# 2
Delta R.T. 0.010 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:165 Resp: 212164


Ion	Ratio	Lower	Upper
165	100		
63	50.3	41.8	62.8
89	69.0	58.1	87.1
182	1.8	1.4	2.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 72.801 ng
RT: 15.888 min Scan# 2178
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:166 Resp: 711212

Ion	Ratio	Lower	Upper
166	100		
165	104.4	81.5	122.3
167	13.8	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 84.708 ng
RT: 15.459 min Scan# 21
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:232 Resp: 240094


Ion	Ratio	Lower	Upper
232	100		
131	34.1	26.7	40.1
130	2.1	1.6	2.4
166	25.3	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 76.874 ng
RT: 15.653 min Scan# 2138
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:149 Resp: 803621

Ion	Ratio	Lower	Upper
149	100		
177	22.4	18.0	27.0
150	12.4	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 74.463 ng
RT: 15.876 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:204 Resp: 40823


Ion	Ratio	Lower	Upper
204	100		
206	32.0	25.6	38.4
141	51.7	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 88.462 ng
RT: 15.894 min Scan# 2179
Delta R.T. 0.009 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:138 Resp: 173159

Ion	Ratio	Lower	Upper
138	100		
92	55.1	36.1	76.1
108	101.4	75.6	115.6






#63
Azobenzene
Concen: 75.482 ng
RT: 16.170 min Scan# 2
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion: 77 Resp: 72398


Ion	Ratio	Lower	Upper
77	100		
182	26.3	4.9	44.9
105	21.0	1.3	41.3
51	38.9	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.580 min Scan# 2466
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:188 Resp: 333421

Ion	Ratio	Lower	Upper
188	100		
94	7.9	5.7	8.5
80	8.6	6.8	10.2






#65
4,6-Dinitro-2-methylphenol
Concen: 80.450 ng
RT: 15.947 min Scan# 21
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:198 Resp: 12410


Ion	Ratio	Lower	Upper
198	100		
51	49.5	24.7	64.7
105	36.3	18.0	58.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#66
n-Nitrosodiphenylamine
Concen: 74.020 ng
RT: 16.088 min Scan# 2212
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:169 Resp: 666665

Ion	Ratio	Lower	Upper
169	100		
168	71.0	56.3	84.5
167	38.3	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 75.762 ng
RT: 16.769 min Scan# 2
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:248 Resp: 29934


Ion	Ratio	Lower	Upper
248	100		
250	96.9	77.7	116.5
141	58.0	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 75.724 ng
RT: 16.893 min Scan# 2349
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:284 Resp: 341338

Ion	Ratio	Lower	Upper
284	100		
142	26.1	21.0	31.6
249	35.4	24.9	37.3






#69
Atrazine
Concen: 72.745 ng
RT: 17.028 min Scan# 2
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:200 Resp: 28195


Ion	Ratio	Lower	Upper
200	100		
173	22.2	0.9	40.9
215	45.2	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 83.771 ng
RT: 17.228 min Scan# 2406
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:266 Resp: 235472

Ion	Ratio	Lower	Upper
266	100		
268	67.7	50.3	75.5
264	65.8	47.3	70.9






#71
Phenanthrene
Concen: 73.525 ng
RT: 17.627 min Scan# 24
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:178 Resp: 134061


Ion	Ratio	Lower	Upper
178	100		
176	19.0	14.7	22.1
179	15.8	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 73.704 ng
RT: 17.715 min Scan# 2489
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:178 Resp: 1367044

Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.4	21.6
179	15.1	12.4	18.6






#73
Carbazole
Concen: 73.122 ng
RT: 17.974 min Scan# 2
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:167 Resp: 119596


Ion	Ratio	Lower	Upper
167	100		
166	22.5	17.9	26.9
139	13.0	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 75.790 ng
RT: 18.538 min Scan# 2629
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:149 Resp: 1433446

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.0	10.6
104	5.4	4.1	6.1






#75
Fluoranthene
Concen: 72.835 ng
RT: 19.619 min Scan# 2
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:202 Resp: 166367


Ion	Ratio	Lower	Upper
202	100		
101	8.8	0.0	28.7
203	18.2	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.863 min Scan# 3195
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:240 Resp: 363291

Ion	Ratio	Lower	Upper
240	100		
120	6.7	6.5	9.7
236	26.2	20.2	30.2






#77
Benzidine
Concen: 68.376 ng
RT: 19.783 min Scan# 2
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:184 Resp: 55652


Ion	Ratio	Lower	Upper
184	100		
185	13.5	11.4	17.2
183	11.3	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 73.215 ng
RT: 19.977 min Scan# 2874
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:202 Resp: 1737788

Ion	Ratio	Lower	Upper
202	100		
200	22.2	17.0	25.4
203	17.5	13.8	20.6






#79
Terphenyl-d14
Concen: 140.125 ng
RT: 20.177 min Scan# 29
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:244 Resp: 269749


Ion	Ratio	Lower	Upper
244	100		
212	8.1	6.1	9.1
122	7.0	5.6	8.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#80
Butylbenzylphthalate
Concen: 82.988 ng
RT: 20.859 min Scan# 3024
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:149 Resp: 595333

Ion	Ratio	Lower	Upper
149	100		
91	72.8	59.2	88.8
206	23.6	18.1	27.1






#81
Benzo(a)anthracene
Concen: 74.959 ng
RT: 21.840 min Scan# 31
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:228 Resp: 180562


Ion	Ratio	Lower	Upper
228	100		
226	28.0	21.9	32.9
229	20.1	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 76.524 ng
RT: 21.740 min Scan# 3174
Delta R.T. -0.002 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:252 Resp: 615546

Ion	Ratio	Lower	Upper
252	100		
254	63.9	50.8	76.2
126	7.7	5.8	8.6






#83
Chrysene
Concen: 74.368 ng
RT: 21.910 min Scan# 3
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:228 Resp: 170453


Ion	Ratio	Lower	Upper
228	100		
226	30.8	23.3	34.9
229	20.0	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 79.672 ng
RT: 21.752 min Scan# 3176
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:149 Resp: 854027

Ion	Ratio	Lower	Upper
149	100		
167	26.3	20.0	30.0
279	4.3	3.1	4.7





#85
Di-n-octyl phthalate
Concen: 79.289 ng
RT: 23.015 min Scan# 3
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24


Tgt Ion:149 Resp: 143295


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	13.5	10.8	16.2

Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.201 min Scan# 3763
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:264 Resp: 382372

Ion	Ratio	Lower	Upper
264	100		
260	23.6	19.0	28.4
265	21.0	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 76.663 ng
RT: 29.061 min Scan# 44
Delta R.T. 0.021 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:276 Resp: 215642


Ion	Ratio	Lower	Upper
276	100		
138	10.9	10.8	16.2
277	25.2	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 77.199 ng
RT: 24.143 min Scan# 3583
Delta R.T. 0.004 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:252 Resp: 1810094

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.8	26.8
125	6.8	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 74.916 ng
RT: 24.219 min Scan# 3
Delta R.T. 0.009 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:252 Resp: 175775


Ion	Ratio	Lower	Upper
252	100		
253	21.8	16.8	25.2
125	7.1	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 76.546 ng
RT: 25.060 min Scan# 3739
Delta R.T. 0.015 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:252 Resp: 1649779

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.6	26.4
125	7.2	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 76.486 ng
RT: 29.137 min Scan# 44
Delta R.T. 0.027 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:278 Resp: 174789


Ion	Ratio	Lower	Upper
278	100		
139	10.0	8.0	12.0
279	23.0	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 76.498 ng
RT: 30.265 min Scan# 4625
Delta R.T. 0.033 min
Lab File: BG064575.D
Acq: 24 Oct 2025 14:24

Tgt Ion:276 Resp: 1669976

Ion	Ratio	Lower	Upper
276	100		
277	23.1	19.1	28.7
138	13.9	11.3	16.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064576.D
Acq On : 24 Oct 2025 15:32
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 18:03:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.224	152	43774	20.000	ng		0.00
21) Naphthalene-d8	11.050	136	182273	20.000	ng		0.00
39) Acenaphthene-d10	14.846	164	142032	20.000	ng		0.00
64) Phenanthrene-d10	17.584	188	343365	20.000	ng		0.00
76) Chrysene-d12	21.861	240	367592	20.000	ng		0.00
86) Perylene-d12	25.210	264	385865	20.000	ng		0.01
System Monitoring Compounds							
5) 2-Fluorophenol	5.750	112	249333	95.606	ng		0.00
7) Phenol-d6	7.354	99	349586	97.687	ng		0.00
23) Nitrobenzene-d5	9.387	82	324249	106.847	ng		0.00
42) 2,4,6-Tribromophenol	16.326	330	207294	101.122	ng		0.00
45) 2-Fluorobiphenyl	13.477	172	877315	93.626	ng		0.00
79) Terphenyl-d14	20.175	244	1814661	93.162	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.588	88	51872	44.926	ng		99
3) Pyridine	3.999	79	161000	46.434	ng		96
4) n-Nitrosodimethylamine	3.905	42	85377	45.917	ng		92
6) Aniline	7.536	93	228687	46.963	ng		99
8) 2-Chlorophenol	7.777	128	134062	49.112	ng		96
9) Benzaldehyde	7.343	77	91457	46.812	ng		96
10) Phenol	7.384	94	192663	48.681	ng		98
11) bis(2-Chloroethyl)ether	7.630	93	139360	47.894	ng		95
12) 1,3-Dichlorobenzene	8.112	146	146905	46.758	ng		98
13) 1,4-Dichlorobenzene	8.259	146	151083	47.293	ng		93
14) 1,2-Dichlorobenzene	8.582	146	143986	46.729	ng		96
15) Benzyl Alcohol	8.447	79	142880	48.417	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.741	45	361264	47.332	ng		99
17) 2-Methylphenol	8.653	107	126868	48.282	ng		94
18) Hexachloroethane	9.328	117	54330	48.463	ng		94
19) n-Nitroso-di-n-propyla...	9.029	70	122051	49.350	ng		97
20) 3+4-Methylphenols	8.982	107	176415	47.740	ng		99
22) Acetophenone	9.046	105	234956	47.155	ng	#	100
24) Nitrobenzene	9.428	77	171837	52.060	ng		96
25) Isophorone	9.957	82	348174	47.813	ng		97
26) 2-Nitrophenol	10.145	139	61611	48.353	ng		90
27) 2,4-Dimethylphenol	10.198	122	135402	49.988	ng		96
28) bis(2-Chloroethoxy)met...	10.439	93	194038	46.800	ng		99
29) 2,4-Dichlorophenol	10.680	162	148850	51.062	ng		94
30) 1,2,4-Trichlorobenzene	10.909	180	159207	48.068	ng		94
31) Naphthalene	11.103	128	471891	46.695	ng		98
32) Benzoic acid	10.321	122	97970m	47.410	ng		
33) 4-Chloroaniline	11.191	127	186814	47.560	ng		98
34) Hexachlorobutadiene	11.397	225	123596	47.716	ng		98
35) Caprolactam	11.961	113	53689	48.020	ng	#	93
36) 4-Chloro-3-methylphenol	12.296	107	171120	48.836	ng		98
37) 2-Methylnaphthalene	12.695	142	351946	47.435	ng		97
38) 1-Methylnaphthalene	12.913	142	342687	46.666	ng		95
40) 1,2,4,5-Tetrachloroben...	13.054	216	224658	47.131	ng		99
41) Hexachlorocyclopentadiene	13.042	237	105411	49.131	ng		98
43) 2,4,6-Trichlorophenol	13.283	196	141505	51.898	ng		98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064576.D
Acq On : 24 Oct 2025 15:32
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025

Quant Time: Nov 03 18:03:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.353	196	157315	49.997	ng		96
46) 1,1'-Biphenyl	13.688	154	470676	46.613	ng		99
47) 2-Chloronaphthalene	13.729	162	359811	47.208	ng		99
48) 2-Nitroaniline	13.911	65	121908	47.668	ng		92
49) Acenaphthylene	14.569	152	565771	47.060	ng		99
50) Dimethylphthalate	14.293	163	519303	47.781	ng		99
51) 2,6-Dinitrotoluene	14.405	165	86665	48.306	ng		99
52) Acenaphthene	14.910	154	388012	47.239	ng		99
53) 3-Nitroaniline	14.734	138	99748	53.405	ng		96
54) 2,4-Dinitrophenol	14.928	184	49318	47.268	ng		89
55) Dibenzofuran	15.239	168	615553	46.355	ng		99
56) 4-Nitrophenol	15.016	139	88609	51.169	ng		95
57) 2,4-Dinitrotoluene	15.186	165	132374	47.759	ng	#	99
58) Fluorene	15.891	166	481625	46.441	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.457	232	157783	52.440	ng		99
60) Diethylphthalate	15.650	149	527594	47.543	ng		100
61) 4-Chlorophenyl-phenyle...	15.880	204	271068	46.576	ng		98
62) 4-Nitroaniline	15.891	138	110067	52.970	ng		98
63) Azobenzene	16.168	77	475203	46.671	ng		99
65) 4,6-Dinitro-2-methylph...	15.944	198	71672	47.500	ng		95
66) n-Nitrosodiphenylamine	16.085	169	444824	47.959	ng		99
67) 4-Bromophenyl-phenylether	16.773	248	196226	48.225	ng		96
68) Hexachlorobenzene	16.890	284	223018	48.043	ng		94
69) Atrazine	17.025	200	187862	47.066	ng		96
70) Pentachlorophenol	17.225	266	147427	50.930	ng		95
71) Phenanthrene	17.625	178	875415	46.621	ng		99
72) Anthracene	17.719	178	893629	46.785	ng		99
73) Carbazole	17.977	167	781031	46.370	ng		100
74) Di-n-butylphthalate	18.535	149	930281	47.762	ng		100
75) Fluoranthene	19.622	202	1092855	46.459	ng		99
77) Benzidine	19.787	184	382418	46.435	ng		100
78) Pyrene	19.981	202	1133345	47.190	ng		99
80) Butylbenzylphthalate	20.856	149	374611	51.609	ng		100
81) Benzo(a)anthracene	21.843	228	1151242	47.234	ng		99
82) 3,3'-Dichlorobenzidine	21.743	252	389240	47.824	ng		97
83) Chrysene	21.908	228	1091456	47.062	ng		98
84) Bis(2-ethylhexyl)phtha...	21.749	149	537716	49.576	ng		98
85) Di-n-octyl phthalate	23.018	149	896763	49.039	ng		99
87) Indeno(1,2,3-cd)pyrene	29.052	276	1346188	47.425	ng		97
88) Benzo(b)fluoranthene	24.146	252	1131250	47.810	ng		100
89) Benzo(k)fluoranthene	24.217	252	1110490	46.901	ng		98
90) Benzo(a)pyrene	25.051	252	1029605	47.339	ng		99
91) Dibenzo(a,h)anthracene	29.129	278	1090407	47.283	ng		99
92) Benzo(g,h,i)perylene	30.245	276	1039447	47.184	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064576.D
Acq On : 24 Oct 2025 15:32
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 9 Sample Multiplier: 1


Quant Time: Nov 03 18:03:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 17:56:56 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.224 min Scan# 8
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:152 Resp: 4377


Ion	Ratio	Lower	Upper
152	100		
150	158.7	122.2	183.4
115	61.5	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 44.926 ng
RT: 3.588 min Scan# 85
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 88 Resp: 51872

Ion	Ratio	Lower	Upper
88	100		
58	116.7	92.4	138.6
43	41.6	33.0	49.6





#3
Pyridine
Concen: 46.434 ng
RT: 3.999 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32


Tgt Ion: 79 Resp: 16100


Ion	Ratio	Lower	Upper
79	100		
52	121.3	94.2	141.2
51	54.1	46.1	69.1

Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#4
n-Nitrosodimethylamine
Concen: 45.917 ng
RT: 3.905 min Scan# 139
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 42 Resp: 85377

Ion	Ratio	Lower	Upper
42	100		
74	103.9	76.1	114.1
44	8.4	6.3	9.5






#5
2-Fluorophenol
Concen: 95.606 ng
RT: 5.750 min Scan# 4
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:112 Resp: 24933


Ion	Ratio	Lower	Upper
112	100		
64	79.6	60.7	91.1
63	42.3	33.4	50.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#6
Aniline
Concen: 46.963 ng
RT: 7.536 min Scan# 757
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 93 Resp: 228687

Ion	Ratio	Lower	Upper
93	100		
66	46.4	37.0	55.6
65	24.1	18.3	27.5






#7
Phenol-d6
Concen: 97.687 ng
RT: 7.354 min Scan# 7
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 99 Resp: 34958


Ion	Ratio	Lower	Upper
99	100		
42	24.0	18.3	27.5
71	36.1	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 49.112 ng
RT: 7.777 min Scan# 798
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:128 Resp: 134062

Ion	Ratio	Lower	Upper
128	100		
130	33.5	13.5	53.5
64	70.0	45.2	85.2






#9
Benzaldehyde
Concen: 46.812 ng
RT: 7.343 min Scan# 7
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 77 Resp: 9145


Ion	Ratio	Lower	Upper
77	100		
105	91.7	72.9	112.9
106	84.2	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 48.681 ng
RT: 7.384 min Scan# 731
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 94 Resp: 192663

Ion	Ratio	Lower	Upper
94	100		
65	34.1	15.8	55.8
66	44.8	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 47.894 ng
RT: 7.630 min Scan# 7
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 93 Resp: 13936


Ion	Ratio	Lower	Upper
93	100		
63	105.5	91.3	131.3
95	31.8	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 46.758 ng
RT: 8.112 min Scan# 855
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:146 Resp: 146905

Ion	Ratio	Lower	Upper
146	100		
148	67.4	55.3	82.9
75	34.9	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 47.293 ng
RT: 8.259 min Scan# 8
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:146 Resp: 15108


Ion	Ratio	Lower	Upper
146	100		
148	62.7	56.2	84.4
111	42.0	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 46.729 ng
RT: 8.582 min Scan# 935
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:146 Resp: 143986

Ion	Ratio	Lower	Upper
146	100		
148	67.0	51.5	77.3
111	46.3	35.0	52.6






#15
Benzyl Alcohol
Concen: 48.417 ng
RT: 8.447 min Scan# 91
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 79 Resp: 14288


Ion	Ratio	Lower	Upper
79	100		
108	71.2	58.1	87.1
77	66.0	52.8	79.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.332 ng
RT: 8.741 min Scan# 962
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 45 Resp: 361264

Ion	Ratio	Lower	Upper
45	100		
77	8.0	0.0	28.5
79	5.7	0.0	25.6






#17
2-Methylphenol
Concen: 48.282 ng
RT: 8.653 min Scan# 94
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:107 Resp: 126868


Ion	Ratio	Lower	Upper
107	100		
108	120.8	90.0	135.0
77	49.7	37.8	56.8
79	48.8	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 48.463 ng
RT: 9.328 min Scan# 1062
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:117 Resp: 54330

Ion	Ratio	Lower	Upper
117	100		
119	96.5	75.4	113.2
201	134.0	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 49.350 ng
RT: 9.029 min Scan# 10
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 70 Resp: 12205


Ion	Ratio	Lower	Upper
70	100		
42	78.6	65.2	97.8
101	8.8	7.9	11.9
130	18.5	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 47.740 ng
RT: 8.982 min Scan# 1003
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:107 Resp: 176415

Ion	Ratio	Lower	Upper
107	100		
108	90.4	70.3	110.3
77	34.2	14.6	54.6
79	30.0	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.050 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:136 Resp: 18227


Ion	Ratio	Lower	Upper
136	100		
137	11.7	9.2	13.8
54	14.6	12.6	19.0
68	6.8	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 47.155 ng
RT: 9.046 min Scan# 1014
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:105 Resp: 234956

Ion	Ratio	Lower	Upper
105	100		
71	2.8	1.5	2.3#
51	42.1	33.7	50.5
120	19.5	16.1	24.1






#23
Nitrobenzene-d5
Concen: 106.847 ng
RT: 9.387 min Scan# 10
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 82 Resp: 324249


Ion	Ratio	Lower	Upper
82	100		
128	37.9	31.3	46.9
54	72.9	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 52.060 ng
RT: 9.428 min Scan# 1079
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 77 Resp: 171837

Ion	Ratio	Lower	Upper
77	100		
123	41.6	31.0	46.6
65	15.5	12.7	19.1






#25
Isophorone
Concen: 47.813 ng
RT: 9.957 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 82 Resp: 34817


Ion	Ratio	Lower	Upper
82	100		
95	7.9	6.1	9.1
138	14.8	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 48.353 ng
RT: 10.145 min Scan# 1201
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:139 Resp: 61611

Ion	Ratio	Lower	Upper
139	100		
109	35.4	24.6	37.0
65	67.0	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 49.988 ng
RT: 10.198 min Scan# 1
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:122 Resp: 13540


Ion	Ratio	Lower	Upper
122	100		
107	121.2	103.0	154.4
121	59.9	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 46.800 ng
RT: 10.439 min Scan# 1251
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 93 Resp: 194038

Ion	Ratio	Lower	Upper
93	100		
95	32.5	25.4	38.2
123	9.7	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 51.062 ng
RT: 10.680 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:162 Resp: 14885


Ion	Ratio	Lower	Upper
162	100		
164	63.1	48.7	88.7
98	34.6	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 48.068 ng
RT: 10.909 min Scan# 1331
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:180 Resp: 159207

Ion	Ratio	Lower	Upper
180	100		
182	89.2	76.9	115.3
145	26.4	22.1	33.1






#31
Naphthalene
Concen: 46.695 ng
RT: 11.103 min Scan# 1
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:128 Resp: 471891


Ion	Ratio	Lower	Upper
128	100		
129	10.6	8.8	13.2
127	14.3	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 47.410 ng m
RT: 10.321 min Scan# 1231
Delta R.T. 0.012 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:122 Resp: 97970

Ion	Ratio	Lower	Upper
122	100		
105	130.6	104.2	144.2
77	105.4	70.2	110.2






#33
4-Chloroaniline
Concen: 47.560 ng
RT: 11.191 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:127 Resp: 18681


Ion	Ratio	Lower	Upper
127	100		
129	31.7	25.5	38.3
65	42.0	32.2	48.4
92	21.4	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 47.716 ng
RT: 11.397 min Scan# 1414
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:225 Resp: 123596

Ion	Ratio	Lower	Upper
225	100		
223	59.0	48.8	73.2
227	63.1	49.5	74.3






#35
Caprolactam
Concen: 48.020 ng
RT: 11.961 min Scan# 1
Delta R.T. 0.013 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:113 Resp: 5368


Ion	Ratio	Lower	Upper
113	100		
55	305.8	290.1	330.1
56	227.6	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 48.836 ng
RT: 12.296 min Scan# 1567
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:107 Resp: 171120

Ion	Ratio	Lower	Upper
107	100		
144	23.9	19.1	28.7
142	75.2	58.6	87.8






#37
2-Methylnaphthalene
Concen: 47.435 ng
RT: 12.695 min Scan# 1
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:142 Resp: 351946


Ion	Ratio	Lower	Upper
142	100		
141	83.0	68.9	103.3
115	37.5	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 46.666 ng
RT: 12.913 min Scan# 1672
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:142 Resp: 342687

Ion	Ratio	Lower	Upper
142	100		
141	86.8	73.3	109.9
116	4.4	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.846 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:164 Resp: 14203


Ion	Ratio	Lower	Upper
164	100		
162	101.1	77.0	115.6
160	47.2	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 47.131 ng
RT: 13.054 min Scan# 1696
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:216 Resp: 224658

Ion	Ratio	Lower	Upper
216	100		
214	77.5	61.4	92.2
179	19.4	15.4	23.0
108	14.6	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 49.131 ng
RT: 13.042 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:237 Resp: 10541


Ion	Ratio	Lower	Upper
237	100		
235	66.1	44.1	84.1
272	13.3	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 101.122 ng
RT: 16.326 min Scan# 2253
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:330 Resp: 207294

Ion	Ratio	Lower	Upper
330	100		
332	97.5	78.0	117.0
141	27.6	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 51.898 ng
RT: 13.283 min Scan# 1
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:196 Resp: 14150


Ion	Ratio	Lower	Upper
196	100		
198	98.3	76.6	115.0
200	31.4	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 49.997 ng
RT: 13.353 min Scan# 1747
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:196 Resp: 157315

Ion	Ratio	Lower	Upper
196	100		
198	97.6	74.2	111.4
97	43.5	37.0	55.6
132	24.0	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 93.626 ng
RT: 13.477 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:172 Resp: 87731


Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.4	44.0
170	24.5	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 46.613 ng
RT: 13.688 min Scan# 1804
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:154 Resp: 470676

Ion	Ratio	Lower	Upper
154	100		
153	42.0	22.4	62.4
76	12.8	0.0	33.5






#47
2-Chloronaphthalene
Concen: 47.208 ng
RT: 13.729 min Scan# 18
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:162 Resp: 359811


Ion	Ratio	Lower	Upper
162	100		
127	39.8	32.2	48.2
164	34.2	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 47.668 ng
RT: 13.911 min Scan# 1842
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 65 Resp: 121908

Ion	Ratio	Lower	Upper
65	100		
92	55.4	48.3	72.5
138	85.4	74.9	112.3






#49
Acenaphthylene
Concen: 47.060 ng
RT: 14.569 min Scan# 19
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:152 Resp: 56577


Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.4	23.2
153	12.6	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 47.781 ng
RT: 14.293 min Scan# 1907
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:163 Resp: 519303

Ion	Ratio	Lower	Upper
163	100		
194	3.4	2.8	4.2
164	9.3	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 48.306 ng
RT: 14.405 min Scan# 19
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:165 Resp: 8666


Ion	Ratio	Lower	Upper
165	100		
63	72.5	57.4	86.0
89	58.9	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 47.239 ng
RT: 14.910 min Scan# 2012
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:154 Resp: 388012

Ion	Ratio	Lower	Upper
154	100		
153	114.6	92.2	138.4
152	54.3	43.2	64.8






#53
3-Nitroaniline
Concen: 53.405 ng
RT: 14.734 min Scan# 19
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:138 Resp: 9974


Ion	Ratio	Lower	Upper
138	100		
108	12.6	11.4	17.0
92	114.8	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 47.268 ng
RT: 14.928 min Scan# 2015
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:184 Resp: 49318

Ion	Ratio	Lower	Upper
184	100		
63	69.5	56.6	85.0
154	219.8	158.6	238.0






#55
Dibenzofuran
Concen: 46.355 ng
RT: 15.239 min Scan# 20
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:168 Resp: 61555


Ion	Ratio	Lower	Upper
168	100		
139	35.5	27.8	41.8
169	12.5	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 51.169 ng
RT: 15.016 min Scan# 2030
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:139 Resp: 88609

Ion	Ratio	Lower	Upper
139	100		
109	95.2	67.8	107.8
65	127.2	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 47.759 ng
RT: 15.186 min Scan# 2
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:165 Resp: 13237


Ion	Ratio	Lower	Upper
165	100		
63	52.8	41.8	62.8
89	73.6	58.1	87.1
182	2.2	1.4	2.0#



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 46.441 ng
RT: 15.891 min Scan# 2179
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:166 Resp: 481625

Ion	Ratio	Lower	Upper
166	100		
165	102.6	81.5	122.3
167	13.5	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 52.440 ng
RT: 15.457 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:232 Resp: 15778


Ion	Ratio	Lower	Upper
232	100		
131	33.9	26.7	40.1
130	2.1	1.6	2.4
166	25.7	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 47.543 ng
RT: 15.650 min Scan# 2138
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:149 Resp: 527594

Ion	Ratio	Lower	Upper
149	100		
177	22.6	18.0	27.0
150	13.1	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 46.576 ng
RT: 15.880 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:204 Resp: 271068


Ion	Ratio	Lower	Upper
204	100		
206	31.7	25.6	38.4
141	53.3	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 52.970 ng
RT: 15.891 min Scan# 2179
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:138 Resp: 110067

Ion	Ratio	Lower	Upper
138	100		
92	54.3	36.1	76.1
108	97.2	75.6	115.6






#63
Azobenzene
Concen: 46.671 ng
RT: 16.168 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion: 77 Resp: 47520


Ion	Ratio	Lower	Upper
77	100		
182	25.0	4.9	44.9
105	20.2	1.3	41.3
51	38.1	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.584 min Scan# 2467
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:188 Resp: 343365

Ion	Ratio	Lower	Upper
188	100		
94	7.6	5.7	8.5
80	8.4	6.8	10.2






#65
4,6-Dinitro-2-methylphenol
Concen: 47.500 ng
RT: 15.944 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:198 Resp: 7167


Ion	Ratio	Lower	Upper
198	100		
51	50.7	24.7	64.7
105	38.3	18.0	58.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#66
n-Nitrosodiphenylamine
Concen: 47.959 ng
RT: 16.085 min Scan# 2212
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:169 Resp: 444824

Ion	Ratio	Lower	Upper
169	100		
168	69.8	56.3	84.5
167	37.5	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 48.225 ng
RT: 16.773 min Scan# 2
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:248 Resp: 19622


Ion	Ratio	Lower	Upper
248	100		
250	101.5	77.7	116.5
141	59.0	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 48.043 ng
RT: 16.890 min Scan# 2349
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:284 Resp: 223018

Ion	Ratio	Lower	Upper
284	100		
142	28.9	21.0	31.6
249	35.3	24.9	37.3






#69
Atrazine
Concen: 47.066 ng
RT: 17.025 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:200 Resp: 18786


Ion	Ratio	Lower	Upper
200	100		
173	22.9	0.9	40.9
215	44.7	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 50.930 ng
RT: 17.225 min Scan# 2406
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:266 Resp: 147427

Ion	Ratio	Lower	Upper
266	100		
268	67.9	50.3	75.5
264	62.0	47.3	70.9






#71
Phenanthrene
Concen: 46.621 ng
RT: 17.625 min Scan# 24
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:178 Resp: 87541


Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.7	22.1
179	16.1	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 46.785 ng
RT: 17.719 min Scan# 2490
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:178 Resp: 893629

Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.4	21.6
179	15.2	12.4	18.6






#73
Carbazole
Concen: 46.370 ng
RT: 17.977 min Scan# 2
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:167 Resp: 78103


Ion	Ratio	Lower	Upper
167	100		
166	22.4	17.9	26.9
139	12.3	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 47.762 ng
RT: 18.535 min Scan# 2629
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:149 Resp: 930281

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.0	10.6
104	5.4	4.1	6.1






#75
Fluoranthene
Concen: 46.459 ng
RT: 19.622 min Scan# 2
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:202 Resp: 109285


Ion	Ratio	Lower	Upper
202	100		
101	8.2	0.0	28.7
203	17.3	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.861 min Scan# 3195
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:240 Resp: 367592

Ion	Ratio	Lower	Upper
240	100		
120	7.9	6.5	9.7
236	26.3	20.2	30.2






#77
Benzidine
Concen: 46.435 ng
RT: 19.787 min Scan# 28
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:184 Resp: 382418


Ion	Ratio	Lower	Upper
184	100		
185	14.4	11.4	17.2
183	11.8	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 47.190 ng
RT: 19.981 min Scan# 2875
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:202 Resp: 1133345

Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.0	25.4
203	17.0	13.8	20.6





#79
Terphenyl-d14
Concen: 93.162 ng
RT: 20.175 min Scan# 29
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32


Tgt Ion:244 Resp: 181466


Ion	Ratio	Lower	Upper
244	100		
212	7.6	6.1	9.1
122	6.9	5.6	8.4

Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#80
Butylbenzylphthalate
Concen: 51.609 ng
RT: 20.856 min Scan# 3024
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:149 Resp: 374611

Ion	Ratio	Lower	Upper
149	100		
91	74.1	59.2	88.8
206	22.5	18.1	27.1






#81
Benzo(a)anthracene
Concen: 47.234 ng
RT: 21.843 min Scan# 3
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:228 Resp: 115124


Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.9	32.9
229	20.1	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 47.824 ng
RT: 21.743 min Scan# 3175
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:252 Resp: 389240

Ion	Ratio	Lower	Upper
252	100		
254	66.1	50.8	76.2
126	6.9	5.8	8.6






#83
Chrysene
Concen: 47.062 ng
RT: 21.908 min Scan# 3
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:228 Resp: 109145


Ion	Ratio	Lower	Upper
228	100		
226	30.1	23.3	34.9
229	19.5	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 49.576 ng
RT: 21.749 min Scan# 3176
Delta R.T. 0.001 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:149 Resp: 537716

Ion	Ratio	Lower	Upper
149	100		
167	26.1	20.0	30.0
279	4.4	3.1	4.7





#85
Di-n-octyl phthalate
Concen: 49.039 ng
RT: 23.018 min Scan# 3
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32


Tgt Ion:149 Resp: 89676


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	13.1	10.8	16.2

Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.210 min Scan# 3765
Delta R.T. 0.013 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:264 Resp: 385865

Ion	Ratio	Lower	Upper
264	100		
260	22.4	19.0	28.4
265	21.7	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 47.425 ng
RT: 29.052 min Scan# 44
Delta R.T. 0.013 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:276 Resp: 134618


Ion	Ratio	Lower	Upper
276	100		
138	10.9	10.8	16.2
277	24.8	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 47.810 ng
RT: 24.146 min Scan# 3584
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:252 Resp: 1131250

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.8	26.8
125	6.9	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 46.901 ng
RT: 24.217 min Scan# 3
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:252 Resp: 111049


Ion	Ratio	Lower	Upper
252	100		
253	22.4	16.8	25.2
125	6.8	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 47.339 ng
RT: 25.051 min Scan# 3738
Delta R.T. 0.007 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:252 Resp: 1029605

Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.6	26.4
125	7.2	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 47.283 ng
RT: 29.129 min Scan# 44
Delta R.T. 0.019 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:278 Resp: 109040


Ion	Ratio	Lower	Upper
278	100		
139	9.7	8.0	12.0
279	24.0	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 10/27/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 47.184 ng
RT: 30.245 min Scan# 4622
Delta R.T. 0.013 min
Lab File: BG064576.D
Acq: 24 Oct 2025 15:32

Tgt Ion:276 Resp: 1039447

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.1	28.7
138	15.0	11.3	16.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.222	152	37186	20.000	ng	0.00
21) Naphthalene-d8	11.048	136	157125	20.000	ng	0.00
39) Acenaphthene-d10	14.844	164	124202	20.000	ng	0.00
64) Phenanthrene-d10	17.582	188	290051	20.000	ng	0.00
76) Chrysene-d12	21.859	240	293608	20.000	ng	0.00
86) Perylene-d12	25.202	264	306886	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.743	112	156266	70.535	ng	0.00
7) Phenol-d6	7.353	99	225294	74.109	ng	0.00
23) Nitrobenzene-d5	9.386	82	200749	76.738	ng	0.00
42) 2,4,6-Tribromophenol	16.319	330	134578	75.074	ng	0.00
45) 2-Fluorobiphenyl	13.475	172	594887	72.599	ng	0.00
79) Terphenyl-d14	20.173	244	1151597	74.019	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.586	88	33172	33.820	ng	97
3) Pyridine	3.998	79	98123	33.313	ng	95
4) n-Nitrosodimethylamine	3.898	42	56903	36.025	ng	98
6) Aniline	7.529	93	157683	38.118	ng	99
8) 2-Chlorophenol	7.776	128	85755	36.981	ng	96
9) Benzaldehyde	7.341	77	84516	50.923	ng	95
10) Phenol	7.376	94	122412	36.410	ng	98
11) bis(2-Chloroethyl)ether	7.623	93	86447	34.973	ng	99
12) 1,3-Dichlorobenzene	8.111	146	96308	36.085	ng	94
13) 1,4-Dichlorobenzene	8.257	146	102132	37.634	ng	96
14) 1,2-Dichlorobenzene	8.581	146	96651	36.924	ng	97
15) Benzyl Alcohol	8.445	79	95250	37.995	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.745	45	234476	36.163	ng	98
17) 2-Methylphenol	8.651	107	85863	38.466	ng	93
18) Hexachloroethane	9.327	117	34040	35.743	ng	96
19) n-Nitroso-di-n-propyla...	9.027	70	83140	39.572	ng	96
20) 3+4-Methylphenols	8.974	107	118917	37.882	ng	96
22) Acetophenone	9.045	105	162385	37.806	ng	97
24) Nitrobenzene	9.427	77	111836	39.305	ng	97
25) Isophorone	9.955	82	236860	37.733	ng	99
26) 2-Nitrophenol	10.143	139	38013	35.498	ng	97
27) 2,4-Dimethylphenol	10.196	122	94545	40.491	ng	99
28) bis(2-Chloroethoxy)met...	10.437	93	132072	36.953	ng	99
29) 2,4-Dichlorophenol	10.684	162	95966	38.190	ng	97
30) 1,2,4-Trichlorobenzene	10.907	180	108339	37.945	ng	97
31) Naphthalene	11.101	128	311084	35.709	ng	100
32) Benzoic acid	10.296	122	48162	29.620	ng	96
33) 4-Chloroaniline	11.189	127	139799	41.287	ng	99
34) Hexachlorobutadiene	11.389	225	78691	35.242	ng	97
35) Caprolactam	11.941	113	35011	36.326	ng	93
36) 4-Chloro-3-methylphenol	12.294	107	114927	38.049	ng	98
37) 2-Methylnaphthalene	12.693	142	213545	33.388	ng	98
38) 1-Methylnaphthalene	12.911	142	224857	35.513	ng	100
40) 1,2,4,5-Tetrachloroben...	13.052	216	154838	37.146	ng	99
41) Hexachlorocyclopentadiene	13.040	237	95065	50.669	ng	98
43) 2,4,6-Trichlorophenol	13.281	196	92335	38.726	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.346	196	102433	37.228	ng		99
46) 1,1'-Biphenyl	13.686	154	329629	37.331	ng		99
47) 2-Chloronaphthalene	13.727	162	241311	36.205	ng		97
48) 2-Nitroaniline	13.910	65	78927	36.129	ng		96
49) Acenaphthylene	14.568	152	435836	41.457	ng		98
50) Dimethylphthalate	14.292	163	350600	36.890	ng		100
51) 2,6-Dinitrotoluene	14.403	165	61784	39.971	ng		100
52) Acenaphthene	14.908	154	253833	35.340	ng		98
53) 3-Nitroaniline	14.726	138	68729	42.080	ng		99
54) 2,4-Dinitrophenol	14.926	184	29394	33.991	ng	#	56
55) Dibenzofuran	15.237	168	419154	36.096	ng		97
56) 4-Nitrophenol	15.014	139	51210	33.818	ng		92
57) 2,4-Dinitrotoluene	15.185	165	85201	35.957	ng		96
58) Fluorene	15.890	166	328636	36.238	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.455	232	106626	40.525	ng		97
60) Diethylphthalate	15.649	149	348964	35.961	ng		99
61) 4-Chlorophenyl-phenyle...	15.878	204	182714	35.902	ng		99
62) 4-Nitroaniline	15.884	138	68807	37.867	ng		97
63) Azobenzene	16.172	77	323432	36.325	ng		97
65) 4,6-Dinitro-2-methylph...	15.943	198	42391	34.885	ng		94
66) n-Nitrosodiphenylamine	16.084	169	300274	38.325	ng		98
67) 4-Bromophenyl-phenylether	16.771	248	130765	38.044	ng		97
68) Hexachlorobenzene	16.888	284	144974	36.971	ng		96
69) Atrazine	17.024	200	130007	38.558	ng		96
70) Pentachlorophenol	17.223	266	85166	34.829	ng		90
71) Phenanthrene	17.623	178	569825	35.925	ng		99
72) Anthracene	17.717	178	584246	36.210	ng		99
73) Carbazole	17.975	167	498849	35.061	ng		99
74) Di-n-butylphthalate	18.534	149	570718	34.688	ng		99
75) Fluoranthene	19.621	202	676996	34.070	ng		99
77) Benzidine	19.785	184	356547	54.203	ng		99
78) Pyrene	19.979	202	693068	36.130	ng		99
80) Butylbenzylphthalate	20.854	149	224386	38.702	ng		98
81) Benzo(a)anthracene	21.836	228	706428	36.287	ng		99
82) 3,3'-Dichlorobenzidine	21.742	252	265324	40.813	ng		96
83) Chrysene	21.906	228	658217	35.533	ng		98
84) Bis(2-ethylhexyl)phtha...	21.747	149	324614	37.470	ng		96
85) Di-n-octyl phthalate	23.017	149	527097	36.087	ng		99
87) Indeno(1,2,3-cd)pyrene	29.045	276	820974	36.366	ng		99
88) Benzo(b)fluoranthene	24.139	252	681881	36.235	ng		99
89) Benzo(k)fluoranthene	24.209	252	701500	37.252	ng		99
90) Benzo(a)pyrene	25.049	252	652126	37.700	ng		98
91) Dibenzo(a,h)anthracene	29.121	278	674510	36.776	ng		96
92) Benzo(g,h,i)perylene	30.232	276	653564	37.302	ng		98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.222 min Scan# 8
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:152 Resp: 37186

Ion	Ratio	Lower	Upper
152	100		
150	144.5	122.2	183.4
115	60.5	50.8	76.2







#2
1,4-Dioxane
Concen: 33.820 ng
RT: 3.586 min Scan# 85
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion: 88 Resp: 33172

Ion	Ratio	Lower	Upper
88	100		
58	111.8	92.4	138.6
43	40.1	33.0	49.6







#3
Pyridine
Concen: 33.313 ng
RT: 3.998 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 79 Resp: 98123

Ion	Ratio	Lower	Upper
79	100		
52	124.1	94.2	141.2
51	55.8	46.1	69.1







#4
n-Nitrosodimethylamine
Concen: 36.025 ng
RT: 3.898 min Scan# 138
Delta R.T. -0.007 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion: 42 Resp: 56903

Ion	Ratio	Lower	Upper
42	100		
74	97.0	76.1	114.1
44	9.0	6.3	9.5







#5
2-Fluorophenol
Concen: 70.535 ng
RT: 5.743 min Scan# 4
Delta R.T. -0.007 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:112 Resp: 156266

Ion	Ratio	Lower	Upper
112	100		
64	78.8	60.7	91.1
63	39.5	33.4	50.0







#6
Aniline
Concen: 38.118 ng
RT: 7.529 min Scan# 756
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion: 93 Resp: 157683

Ion	Ratio	Lower	Upper
93	100		
66	46.0	37.0	55.6
65	24.3	18.3	27.5







#7
Phenol-d6
Concen: 74.109 ng
RT: 7.353 min Scan# 7
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 99 Resp: 225294

Ion	Ratio	Lower	Upper
99	100		
42	23.9	18.3	27.5
71	36.5	28.6	43.0







#8
2-Chlorophenol
Concen: 36.981 ng
RT: 7.776 min Scan# 798
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:128 Resp: 85755

Ion	Ratio	Lower	Upper
128	100		
130	31.0	13.5	53.5
64	67.9	45.2	85.2







#9
Benzaldehyde
Concen: 50.923 ng
RT: 7.341 min Scan# 7
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 77 Resp: 84516

Ion	Ratio	Lower	Upper
77	100		
105	90.1	72.9	112.9
106	83.4	70.8	110.8







#10
Phenol
Concen: 36.410 ng
RT: 7.376 min Scan# 730
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion: 94 Resp: 122412

Ion	Ratio	Lower	Upper
94	100		
65	34.6	15.8	55.8
66	45.6	24.0	64.0







#11
bis(2-Chloroethyl)ether
Concen: 34.973 ng
RT: 7.623 min Scan# 7
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 93 Resp: 86447

Ion	Ratio	Lower	Upper
93	100		
63	111.9	91.3	131.3
95	32.5	13.3	53.3







#12
1,3-Dichlorobenzene
Concen: 36.085 ng
RT: 8.111 min Scan# 855
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:146 Resp: 96308

Ion	Ratio	Lower	Upper
146	100		
148	63.9	55.3	82.9
75	32.5	28.2	42.2







#13
1,4-Dichlorobenzene
Concen: 37.634 ng
RT: 8.257 min Scan# 88
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:146 Resp: 102132

Ion	Ratio	Lower	Upper
146	100		
148	67.2	56.2	84.4
111	40.6	35.4	53.0







#14
1,2-Dichlorobenzene
Concen: 36.924 ng
RT: 8.581 min Scan# 935
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:146 Resp: 96651

Ion	Ratio	Lower	Upper
146	100		
148	67.4	51.5	77.3
111	44.5	35.0	52.6







#15
Benzyl Alcohol
Concen: 37.995 ng
RT: 8.445 min Scan# 91
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 79 Resp: 95250

Ion	Ratio	Lower	Upper
79	100		
108	70.7	58.1	87.1
77	64.5	52.8	79.2







#16
2,2'-oxybis(1-Chloropropane)
Concen: 36.163 ng
RT: 8.745 min Scan# 963
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion: 45 Resp: 234476

Ion	Ratio	Lower	Upper
45	100		
77	7.8	0.0	28.5
79	6.3	0.0	25.6







#17
2-Methylphenol
Concen: 38.466 ng
RT: 8.651 min Scan# 94
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:107 Resp: 85863

Ion	Ratio	Lower	Upper
107	100		
108	102.3	90.0	135.0
77	49.0	37.8	56.8
79	47.0	40.5	60.7







#18
Hexachloroethane
Concen: 35.743 ng
RT: 9.327 min Scan# 1062
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:117 Resp: 34040

Ion	Ratio	Lower	Upper
117	100		
119	92.3	75.4	113.2
201	129.6	99.2	148.8







#19
n-Nitroso-di-n-propylamine
Concen: 39.572 ng
RT: 9.027 min Scan# 10
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 70 Resp: 83140

Ion	Ratio	Lower	Upper
70	100		
42	76.9	65.2	97.8
101	9.1	7.9	11.9
130	17.6	14.3	21.5







#20
3+4-Methylphenols
Concen: 37.882 ng
RT: 8.974 min Scan# 1002
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:107 Resp: 118917

Ion	Ratio	Lower	Upper
107	100		
108	85.3	70.3	110.3
77	33.8	14.6	54.6
79	29.8	11.7	51.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.048 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:136 Resp: 157125

Ion	Ratio	Lower	Upper
136	100		
137	12.4	9.2	13.8
54	15.6	12.6	19.0
68	6.3	5.7	8.5







#22
Acetophenone
Concen: 37.806 ng
RT: 9.045 min Scan# 1014
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:105 Resp: 162385

Ion	Ratio	Lower	Upper
105	100		
71	1.7	1.5	2.3
51	39.4	33.7	50.5
120	19.7	16.1	24.1







#23
Nitrobenzene-d5
Concen: 76.738 ng
RT: 9.386 min Scan# 10
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 82 Resp: 200749

Ion	Ratio	Lower	Upper
82	100		
128	39.4	31.3	46.9
54	72.3	56.6	84.8







#24
Nitrobenzene
Concen: 39.305 ng
RT: 9.427 min Scan# 1079
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion: 77 Resp: 111836

Ion	Ratio	Lower	Upper
77	100		
123	40.6	31.0	46.6
65	14.3	12.7	19.1







#25
Isophorone
Concen: 37.733 ng
RT: 9.955 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 82 Resp: 236860

Ion	Ratio	Lower	Upper
82	100		
95	7.5	6.1	9.1
138	15.4	13.0	19.4







#26
2-Nitrophenol
Concen: 35.498 ng
RT: 10.143 min Scan# 1201
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:139 Resp: 38013

Ion	Ratio	Lower	Upper
139	100		
109	32.9	24.6	37.0
65	61.3	47.5	71.3







#27
2,4-Dimethylphenol
Concen: 40.491 ng
RT: 10.196 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:122 Resp: 94545

Ion	Ratio	Lower	Upper
122	100		
107	130.3	103.0	154.4
121	60.4	47.8	71.8







#28
bis(2-Chloroethoxy)methane
Concen: 36.953 ng
RT: 10.437 min Scan# 1251
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion: 93 Resp: 132072

Ion	Ratio	Lower	Upper
93	100		
95	31.4	25.4	38.2
123	9.7	8.4	12.6







#29
2,4-Dichlorophenol
Concen: 38.190 ng
RT: 10.684 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:162 Resp: 95966

Ion	Ratio	Lower	Upper
162	100		
164	71.5	48.7	88.7
98	35.9	16.9	56.9







#30
1,2,4-Trichlorobenzene
Concen: 37.945 ng
RT: 10.907 min Scan# 1331
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:180 Resp: 108339

Ion	Ratio	Lower	Upper
180	100		
182	99.3	76.9	115.3
145	26.3	22.1	33.1







#31
Naphthalene
Concen: 35.709 ng
RT: 11.101 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:128 Resp: 311084

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.8	13.2
127	13.4	10.6	15.8







#32
Benzoic acid
Concen: 29.620 ng
RT: 10.296 min Scan# 1227
Delta R.T. -0.013 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:122 Resp: 48162

Ion	Ratio	Lower	Upper
122	100		
105	124.6	104.2	144.2
77	98.8	70.2	110.2







#33
4-Chloroaniline
Concen: 41.287 ng
RT: 11.189 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:127 Resp: 139799

Ion	Ratio	Lower	Upper
127	100		
129	31.4	25.5	38.3
65	40.0	32.2	48.4
92	21.0	18.2	27.2







#34
Hexachlorobutadiene
Concen: 35.242 ng
RT: 11.389 min Scan# 1413
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:225 Resp: 78691

Ion	Ratio	Lower	Upper
225	100		
223	59.7	48.8	73.2
227	59.3	49.5	74.3







#35
Caprolactam
Concen: 36.326 ng
RT: 11.941 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:113 Resp: 35011

Ion	Ratio	Lower	Upper
113	100		
55	299.0	290.1	330.1
56	217.0	183.8	223.8







#36
4-Chloro-3-methylphenol
Concen: 38.049 ng
RT: 12.294 min Scan# 1567
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:107 Resp: 114927

Ion	Ratio	Lower	Upper
107	100		
144	23.8	19.1	28.7
142	75.7	58.6	87.8







#37
2-Methylnaphthalene
Concen: 33.388 ng
RT: 12.693 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:142 Resp: 213545

Ion	Ratio	Lower	Upper
142	100		
141	84.0	68.9	103.3
115	36.8	30.6	46.0







#38
1-Methylnaphthalene
Concen: 35.513 ng
RT: 12.911 min Scan# 1672
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:142 Resp: 224857

Ion	Ratio	Lower	Upper
142	100		
141	91.9	73.3	109.9
116	3.8	3.1	4.7







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.844 min Scan# 20
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:164 Resp: 124202

Ion	Ratio	Lower	Upper
164	100		
162	97.8	77.0	115.6
160	44.9	34.4	51.6







#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.146 ng
RT: 13.052 min Scan# 1696
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:216 Resp: 154838

Ion	Ratio	Lower	Upper
216	100		
214	77.4	61.4	92.2
179	18.9	15.4	23.0
108	14.9	11.8	17.6







#41
Hexachlorocyclopentadiene
Concen: 50.669 ng
RT: 13.040 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:237 Resp: 95065

Ion	Ratio	Lower	Upper
237	100		
235	62.9	44.1	84.1
272	13.5	0.0	32.4







#42
2,4,6-Tribromophenol
Concen: 75.074 ng
RT: 16.319 min Scan# 2252
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:330 Resp: 134578

Ion	Ratio	Lower	Upper
330	100		
332	97.4	78.0	117.0
141	27.0	21.8	32.6







#43
2,4,6-Trichlorophenol
Concen: 38.726 ng
RT: 13.281 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:196 Resp: 92335

Ion	Ratio	Lower	Upper
196	100		
198	94.8	76.6	115.0
200	31.7	26.0	39.0







#44
2,4,5-Trichlorophenol
Concen: 37.228 ng
RT: 13.346 min Scan# 1746
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:196 Resp: 102433

Ion	Ratio	Lower	Upper
196	100		
198	93.9	74.2	111.4
97	47.0	37.0	55.6
132	23.1	18.5	27.7







#45
2-Fluorobiphenyl
Concen: 72.599 ng
RT: 13.475 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:172 Resp: 594887

Ion	Ratio	Lower	Upper
172	100		
171	35.6	29.4	44.0
170	24.2	19.1	28.7







#46
1,1'-Biphenyl
Concen: 37.331 ng
RT: 13.686 min Scan# 1804
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:154 Resp: 329629

Ion	Ratio	Lower	Upper
154	100		
153	42.8	22.4	62.4
76	12.9	0.0	33.5







#47
2-Chloronaphthalene
Concen: 36.205 ng
RT: 13.727 min Scan# 18
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:162 Resp: 241311

Ion	Ratio	Lower	Upper
162	100		
127	41.5	32.2	48.2
164	36.6	27.5	41.3







#48
2-Nitroaniline
Concen: 36.129 ng
RT: 13.910 min Scan# 1842
Delta R.T. -0.000 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion: 65 Resp: 78927

Ion	Ratio	Lower	Upper
65	100		
92	60.2	48.3	72.5
138	87.2	74.9	112.3







#49
Acenaphthylene
Concen: 41.457 ng
RT: 14.568 min Scan# 19
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:152 Resp: 435836

Ion	Ratio	Lower	Upper
152	100		
151	20.5	15.4	23.2
153	13.2	11.0	16.4







#50
Dimethylphthalate
Concen: 36.890 ng
RT: 14.292 min Scan# 1907
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:163 Resp: 350600

Ion	Ratio	Lower	Upper
163	100		
194	3.2	2.8	4.2
164	9.5	7.7	11.5







#51
2,6-Dinitrotoluene
Concen: 39.971 ng
RT: 14.403 min Scan# 19
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:165 Resp: 61784

Ion	Ratio	Lower	Upper
165	100		
63	71.5	57.4	86.0
89	59.7	47.8	71.8







#52
Acenaphthene
Concen: 35.340 ng
RT: 14.908 min Scan# 2012
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:154 Resp: 253833

Ion	Ratio	Lower	Upper
154	100		
153	118.0	92.2	138.4
152	54.6	43.2	64.8







#53
3-Nitroaniline
Concen: 42.080 ng
RT: 14.726 min Scan# 19
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:138 Resp: 68729

Ion	Ratio	Lower	Upper
138	100		
108	12.5	11.4	17.0
92	120.1	95.3	142.9







#54
2,4-Dinitrophenol
Concen: 33.991 ng
RT: 14.926 min Scan# 2015
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:184 Resp: 29394

Ion	Ratio	Lower	Upper
184	100		
63	75.8	56.6	85.0
154	285.9	158.6	238.0#







#55
Dibenzofuran
Concen: 36.096 ng
RT: 15.237 min Scan# 20
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:168 Resp: 419154

Ion	Ratio	Lower	Upper
168	100		
139	36.7	27.8	41.8
169	13.6	10.3	15.5







#56
4-Nitrophenol
Concen: 33.818 ng
RT: 15.014 min Scan# 2030
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:139 Resp: 51210

Ion	Ratio	Lower	Upper
139	100		
109	96.8	67.8	107.8
65	132.2	103.7	143.7







#57
2,4-Dinitrotoluene
Concen: 35.957 ng
RT: 15.185 min Scan# 2
Delta R.T. 0.006 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:165 Resp: 85201

Ion	Ratio	Lower	Upper
165	100		
63	50.9	41.8	62.8
89	67.3	58.1	87.1
182	1.7	1.4	2.0







#58
Fluorene
Concen: 36.238 ng
RT: 15.890 min Scan# 2179
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:166 Resp: 328636

Ion	Ratio	Lower	Upper
166	100		
165	102.6	81.5	122.3
167	13.1	11.4	17.0







#59
2,3,4,6-Tetrachlorophenol
Concen: 40.525 ng
RT: 15.455 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:232 Resp: 106626

Ion	Ratio	Lower	Upper
232	100		
131	35.6	26.7	40.1
130	1.7	1.6	2.4
166	24.7	20.5	30.7







#60
Diethylphthalate
Concen: 35.961 ng
RT: 15.649 min Scan# 2138
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:149 Resp: 348964

Ion	Ratio	Lower	Upper
149	100		
177	22.5	18.0	27.0
150	12.2	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 35.902 ng
RT: 15.878 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:204 Resp: 182714

Ion	Ratio	Lower	Upper
204	100		
206	32.0	25.6	38.4
141	51.5	40.6	60.8







#62
4-Nitroaniline
Concen: 37.867 ng
RT: 15.884 min Scan# 2178
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:138 Resp: 68807

Ion	Ratio	Lower	Upper
138	100		
92	52.5	36.1	76.1
108	97.9	75.6	115.6







#63
Azobenzene
Concen: 36.325 ng
RT: 16.172 min Scan# 2
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion: 77 Resp: 323432

Ion	Ratio	Lower	Upper
77	100		
182	25.7	4.9	44.9
105	20.1	1.3	41.3
51	36.5	19.0	59.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.582 min Scan# 2467
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:188 Resp: 290051

Ion	Ratio	Lower	Upper
188	100		
94	7.4	5.7	8.5
80	8.2	6.8	10.2







#65
4,6-Dinitro-2-methylphenol
Concen: 34.885 ng
RT: 15.943 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:198 Resp: 42391

Ion	Ratio	Lower	Upper
198	100		
51	49.5	24.7	64.7
105	35.9	18.0	58.0







#66
n-Nitrosodiphenylamine
Concen: 38.325 ng
RT: 16.084 min Scan# 2212
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:169 Resp: 300274

Ion	Ratio	Lower	Upper
169	100		
168	68.4	56.3	84.5
167	37.3	29.0	43.6







#67
4-Bromophenyl-phenylether
Concen: 38.044 ng
RT: 16.771 min Scan# 2
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:248 Resp: 130765

Ion	Ratio	Lower	Upper
248	100		
250	97.6	77.7	116.5
141	56.3	48.6	72.8







#68
Hexachlorobenzene
Concen: 36.971 ng
RT: 16.888 min Scan# 2349
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:284 Resp: 144974

Ion	Ratio	Lower	Upper
284	100		
142	28.3	21.0	31.6
249	33.5	24.9	37.3







#69
Atrazine
Concen: 38.558 ng
RT: 17.024 min Scan# 2
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:200 Resp: 130007

Ion	Ratio	Lower	Upper
200	100		
173	22.2	0.9	40.9
215	44.3	27.0	67.0







#70
Pentachlorophenol
Concen: 34.829 ng
RT: 17.223 min Scan# 2406
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:266 Resp: 85166

Ion	Ratio	Lower	Upper
266	100		
268	66.7	50.3	75.5
264	70.5	47.3	70.9







#71
Phenanthrene
Concen: 35.925 ng
RT: 17.623 min Scan# 24
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:178 Resp: 569825

Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.7	22.1
179	15.6	12.2	18.4







#72
Anthracene
Concen: 36.210 ng
RT: 17.717 min Scan# 2490
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:178 Resp: 584246

Ion	Ratio	Lower	Upper
178	100		
176	17.6	14.4	21.6
179	15.1	12.4	18.6







#73
Carbazole
Concen: 35.061 ng
RT: 17.975 min Scan# 2
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:167 Resp: 498849

Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.9	26.9
139	12.1	9.7	14.5







#74
Di-n-butylphthalate
Concen: 34.688 ng
RT: 18.534 min Scan# 2629
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:149 Resp: 570718

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.0	10.6
104	5.3	4.1	6.1







#75
Fluoranthene
Concen: 34.070 ng
RT: 19.621 min Scan# 2
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:202 Resp: 676996

Ion	Ratio	Lower	Upper
202	100		
101	8.0	0.0	28.7
203	16.7	0.0	37.0







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.859 min Scan# 3195
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:240 Resp: 293608

Ion	Ratio	Lower	Upper
240	100		
120	7.0	6.5	9.7
236	26.0	20.2	30.2







#77
Benzidine
Concen: 54.203 ng
RT: 19.785 min Scan# 2
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:184 Resp: 356547

Ion	Ratio	Lower	Upper
184	100		
185	14.8	11.4	17.2
183	12.2	9.4	14.2







#78
Pyrene
Concen: 36.130 ng
RT: 19.979 min Scan# 2875
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:202 Resp: 693068

Ion	Ratio	Lower	Upper
202	100		
200	21.9	17.0	25.4
203	17.7	13.8	20.6







#79
Terphenyl-d14
Concen: 74.019 ng
RT: 20.173 min Scan# 29
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:244 Resp: 1151597

Ion	Ratio	Lower	Upper
244	100		
212	7.2	6.1	9.1
122	6.7	5.6	8.4







#80
Butylbenzylphthalate
Concen: 38.702 ng
RT: 20.854 min Scan# 3024
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:149 Resp: 224386

Ion	Ratio	Lower	Upper
149	100		
91	72.1	59.2	88.8
206	22.6	18.1	27.1







#81
Benzo(a)anthracene
Concen: 36.287 ng
RT: 21.836 min Scan# 31
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:228 Resp: 706428

Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.9	32.9
229	19.2	15.6	23.4







#82
3,3'-Dichlorobenzidine
Concen: 40.813 ng
RT: 21.742 min Scan# 3175
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:252 Resp: 265324

Ion	Ratio	Lower	Upper
252	100		
254	66.8	50.8	76.2
126	6.7	5.8	8.6







#83
Chrysene
Concen: 35.533 ng
RT: 21.906 min Scan# 3
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:228 Resp: 658217

Ion	Ratio	Lower	Upper
228	100		
226	30.4	23.3	34.9
229	19.5	15.2	22.8







#84
Bis(2-ethylhexyl)phthalate
Concen: 37.470 ng
RT: 21.747 min Scan# 3176
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:149 Resp: 324614

Ion	Ratio	Lower	Upper
149	100		
167	27.1	20.0	30.0
279	4.3	3.1	4.7







#85
Di-n-octyl phthalate
Concen: 36.087 ng
RT: 23.017 min Scan# 3
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:149 Resp: 527097

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	13.3	10.8	16.2







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.202 min Scan# 3764
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:264 Resp: 306886

Ion	Ratio	Lower	Upper
264	100		
260	24.0	19.0	28.4
265	20.8	17.6	26.4







#87
Indeno(1,2,3-cd)pyrene
Concen: 36.366 ng
RT: 29.045 min Scan# 44
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:276 Resp: 820974

Ion	Ratio	Lower	Upper
276	100		
138	12.6	10.8	16.2
277	25.0	20.1	30.1







#88
Benzo(b)fluoranthene
Concen: 36.235 ng
RT: 24.139 min Scan# 3583
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:252 Resp: 681881

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.8	26.8
125	6.8	5.6	8.4







#89
Benzo(k)fluoranthene
Concen: 37.252 ng
RT: 24.209 min Scan# 3
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:252 Resp: 701500

Ion	Ratio	Lower	Upper
252	100		
253	21.2	16.8	25.2
125	7.3	5.4	8.0







#90
Benzo(a)pyrene
Concen: 37.700 ng
RT: 25.049 min Scan# 3738
Delta R.T. 0.005 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:252 Resp: 652126

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.6	26.4
125	6.9	6.1	9.1







#91
Dibenzo(a,h)anthracene
Concen: 36.776 ng
RT: 29.121 min Scan# 44
Delta R.T. 0.011 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Tgt Ion:278 Resp: 674510

Ion	Ratio	Lower	Upper
278	100		
139	9.4	8.0	12.0
279	21.9	19.8	29.6






#92
Benzo(g,h,i)perylene
Concen: 37.302 ng
RT: 30.232 min Scan# 4620
Delta R.T. -0.001 min
Lab File: BG064577.D
Acq: 24 Oct 2025 16:47

Tgt Ion:276 Resp: 653564

Ion	Ratio	Lower	Upper
276	100		
277	23.3	19.1	28.7
138	15.1	11.3	16.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2		1,4-Dioxane	40.000	33.820	15.4	79	0.00
3		Pyridine	40.000	33.313	16.7	73	0.00
4		n-Nitrosodimethylamine	40.000	36.025	9.9	79	0.00
5	S	2-Fluorophenol	80.000	70.535	11.8	75	0.00
6		Aniline	40.000	38.118	4.7	82	0.00
7	S	Phenol-d6	80.000	74.109	7.4	79	0.00
8		2-Chlorophenol	40.000	36.981	7.5	78	0.00
9		Benzaldehyde	40.000	50.923	-27.3#	112	0.00
10	C	Phenol	40.000	36.410	9.0	78	0.00
11		bis(2-Chloroethyl)ether	40.000	34.973	12.6	78	0.00
12		1,3-Dichlorobenzene	40.000	36.085	9.8	79	0.00
13	C	1,4-Dichlorobenzene	40.000	37.634	5.9	83	0.00
14		1,2-Dichlorobenzene	40.000	36.924	7.7	81	0.00
15		Benzyl Alcohol	40.000	37.995	5.0	83	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	36.163	9.6	78	0.00
17		2-Methylphenol	40.000	38.466	3.8	84	0.00
18		Hexachloroethane	40.000	35.743	10.6	78	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	39.572	1.1	84	0.00
20		3+4-Methylphenols	40.000	37.882	5.3	81	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22		Acetophenone	40.000	37.806	5.5	84	0.00
23	S	Nitrobenzene-d5	80.000	76.738	4.1	81	0.00
24		Nitrobenzene	40.000	39.305	1.7	83	0.00
25		Isophorone	40.000	37.733	5.7	83	0.00
26	C	2-Nitrophenol	40.000	35.498	11.3	84	0.00
27		2,4-Dimethylphenol	40.000	40.491	-1.2	86	0.00
28		bis(2-Chloroethoxy)methane	40.000	36.953	7.6	82	0.00
29	C	2,4-Dichlorophenol	40.000	38.190	4.5	81	0.00
30		1,2,4-Trichlorobenzene	40.000	37.945	5.1	84	0.00
31		Naphthalene	40.000	35.709	10.7	80	0.00
32		Benzoic acid	40.000	29.620	25.9#	65	-0.01
33		4-Chloroaniline	40.000	41.287	-3.2	92	0.00
34	C	Hexachlorobutadiene	40.000	35.242	11.9	80	0.00
35		Caprolactam	40.000	36.326	9.2	79	0.00
36	C	4-Chloro-3-methylphenol	40.000	38.049	4.9	82	0.00
37		2-Methylnaphthalene	40.000	33.388	16.5	75	0.00
38		1-Methylnaphthalene	40.000	35.513	11.2	80	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	37.146	7.1	84	0.00
41	P	Hexachlorocyclopentadiene	40.000	50.669	-26.7#	112	0.00
42	S	2,4,6-Tribromophenol	80.000	75.074	6.2	81	0.00
43	C	2,4,6-Trichlorophenol	40.000	38.726	3.2	84	0.00
44		2,4,5-Trichlorophenol	40.000	37.228	6.9	81	0.00
45	S	2-Fluorobiphenyl	80.000	72.599	9.3	82	0.00
46		1,1'-Biphenyl	40.000	37.331	6.7	85	0.00
47		2-Chloronaphthalene	40.000	36.205	9.5	82	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	36.129	9.7	87	0.00
49		Acenaphthylene	40.000	41.457	-3.6	93	0.00
50		Dimethylphthalate	40.000	36.890	7.8	82	0.00
51		2,6-Dinitrotoluene	40.000	39.971	0.1	96	0.00
52	C	Acenaphthene	40.000	35.340	11.6	80	0.00
53		3-Nitroaniline	40.000	42.080	-5.2	89	0.00
54	P	2,4-Dinitrophenol	40.000	33.991	15.0	82	0.00
55		Dibenzofuran	40.000	36.096	9.8	82	0.00
56	P	4-Nitrophenol	40.000	33.818	15.5	75	0.00
57		2,4-Dinitrotoluene	40.000	35.957	10.1	85	0.00
58		Fluorene	40.000	36.238	9.4	82	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	40.525	-1.3	88	0.00
60		Diethylphthalate	40.000	35.961	10.1	81	0.00
61		4-Chlorophenyl-phenylether	40.000	35.902	10.2	81	0.00
62		4-Nitroaniline	40.000	37.867	5.3	81	0.00
63		Azobenzene	40.000	36.325	9.2	81	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65		4,6-Dinitro-2-methylphenol	40.000	34.885	12.8	82	0.00
66	c	n-Nitrosodiphenylamine	40.000	38.325	4.2	82	0.00
67		4-Bromophenyl-phenylether	40.000	38.044	4.9	81	0.00
68		Hexachlorobenzene	40.000	36.971	7.6	79	0.00
69		Atrazine	40.000	38.558	3.6	85	0.00
70	C	Pentachlorophenol	40.000	34.829	12.9	76	0.00
71		Phenanthrene	40.000	35.925	10.2	78	0.00
72		Anthracene	40.000	36.210	9.5	79	0.00
73		Carbazole	40.000	35.061	12.3	77	0.00
74		Di-n-butylphthalate	40.000	34.688	13.3	73	0.00
75	C	Fluoranthene	40.000	34.070	14.8	75	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	88	0.00
77		Benzidine	40.000	54.203	-35.5#	109	0.00
78		Pyrene	40.000	36.130	9.7	74	0.00
79	S	Terphenyl-d14	80.000	74.019	7.5	76	0.00
80		Butylbenzylphthalate	40.000	38.702	3.2	75	0.00
81		Benzo(a)anthracene	40.000	36.287	9.3	75	0.00
82		3,3'-Dichlorobenzidine	40.000	40.813	-2.0	83	0.00
83		Chrysene	40.000	35.533	11.2	74	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	37.470	6.3	73	0.00
85	c	Di-n-octyl phthalate	40.000	36.087	9.8	73	0.00
86	I	Perylene-d12	20.000	20.000	0.0	87	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	36.366	9.1	75	0.00
88		Benzo(b)fluoranthene	40.000	36.235	9.4	74	0.00
89		Benzo(k)fluoranthene	40.000	37.252	6.9	78	0.00
90	C	Benzo(a)pyrene	40.000	37.700	5.7	77	0.00
91		Dibenzo(a,h)anthracene	40.000	36.776	8.1	76	0.01
92		Benzo(g,h,i)perylene	40.000	37.302	6.7	77	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2		1,4-Dioxane	0.528	0.446	15.5	79	0.00
3		Pyridine	1.584	1.319	16.7	73	0.00
4		n-Nitrosodimethylamine	0.850	0.765	10.0	79	0.00
5	S	2-Fluorophenol	1.192	1.051	11.8	75	0.00
6		Aniline	2.225	2.120	4.7	82	0.00
7	S	Phenol-d6	1.635	1.515	7.3	79	0.00
8		2-Chlorophenol	1.247	1.153	7.5	78	0.00
9		Benzaldehyde	0.893	1.136	-27.2#	112	0.00
10	C	Phenol	1.808	1.646	9.0	78	0.00
11		bis(2-Chloroethyl)ether	1.329	1.162	12.6	78	0.00
12		1,3-Dichlorobenzene	1.435	1.295	9.8	79	0.00
13	C	1,4-Dichlorobenzene	1.460	1.373	6.0	83	0.00
14		1,2-Dichlorobenzene	1.408	1.300	7.7	81	0.00
15		Benzyl Alcohol	1.348	1.281	5.0	83	0.00
16		2,2'-oxybis(1-Chloropropane	3.487	3.153	9.6	78	0.00
17		2-Methylphenol	1.201	1.155	3.8	84	0.00
18		Hexachloroethane	0.512	0.458	10.5	78	0.00
19	P	n-Nitroso-di-n-propylamine	1.130	1.118	1.1	84	0.00
20		3+4-Methylphenols	1.688	1.599	5.3	81	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
22		Acetophenone	0.547	0.517	5.5	84	0.00
23	S	Nitrobenzene-d5	0.333	0.319	4.2	81	0.00
24		Nitrobenzene	0.362	0.356	1.7	83	0.00
25		Isophorone	0.799	0.754	5.6	83	0.00
26	C	2-Nitrophenol	0.121	0.121	0.0	84	0.00
27		2,4-Dimethylphenol	0.297	0.301	-1.3	86	0.00
28		bis(2-Chloroethoxy)methane	0.455	0.420	7.7	82	0.00
29	C	2,4-Dichlorophenol	0.320	0.305	4.7	81	0.00
30		1,2,4-Trichlorobenzene	0.363	0.345	5.0	84	0.00
31		Naphthalene	1.109	0.990	10.7	80	0.00
32		Benzoic acid	0.207	0.153	26.1#	65	-0.01
33		4-Chloroaniline	0.431	0.445	-3.2	92	0.00
34	C	Hexachlorobutadiene	0.284	0.250	12.0	80	0.00
35		Caprolactam	0.123	0.111	9.8	79	0.00
36	C	4-Chloro-3-methylphenol	0.384	0.366	4.7	82	0.00
37		2-Methylnaphthalene	0.814	0.680	16.5	75	0.00
38		1-Methylnaphthalene	0.806	0.716	11.2	80	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
40		1,2,4,5-Tetrachlorobenzene	0.671	0.623	7.2	84	0.00
41	P	Hexachlorocyclopentadiene	0.302	0.383	-26.8#	112	0.00
42	S	2,4,6-Tribromophenol	0.289	0.271	6.2	81	0.00
43	C	2,4,6-Trichlorophenol	0.384	0.372	3.1	84	0.00
44		2,4,5-Trichlorophenol	0.443	0.412	7.0	81	0.00
45	S	2-Fluorobiphenyl	1.319	1.197	9.2	82	0.00
46		1,1'-Biphenyl	1.422	1.327	6.7	85	0.00
47		2-Chloronaphthalene	1.073	0.971	9.5	82	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.311	0.318	-2.3	87	0.00
49		Acenaphthylene	1.693	1.755	-3.7	93	0.00
50		Dimethylphthalate	1.530	1.411	7.8	82	0.00
51		2,6-Dinitrotoluene	0.218	0.249	-14.2	96	0.00
52	C	Acenaphthene	1.157	1.022	11.7	80	0.00
53		3-Nitroaniline	0.263	0.277	-5.3	89	0.00
54	P	2,4-Dinitrophenol	0.132	0.118	10.6	82	0.00
55		Dibenzofuran	1.870	1.687	9.8	82	0.00
56	P	4-Nitrophenol	0.244	0.206	15.6	75	0.00
57		2,4-Dinitrotoluene	0.340	0.343	-0.9	85	0.00
58		Fluorene	1.460	1.323	9.4	82	0.00
59		2,3,4,6-Tetrachlorophenol	0.424	0.429	-1.2	88	0.00
60		Diethylphthalate	1.563	1.405	10.1	81	0.00
61		4-Chlorophenyl-phenylether	0.820	0.736	10.2	81	0.00
62		4-Nitroaniline	0.293	0.277	5.5	81	0.00
63		Azobenzene	1.434	1.302	9.2	81	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
65		4,6-Dinitro-2-methylphenol	0.079	0.073	7.6	82	0.00
66	c	n-Nitrosodiphenylamine	0.540	0.518	4.1	82	0.00
67		4-Bromophenyl-phenylether	0.237	0.225	5.1	81	0.00
68		Hexachlorobenzene	0.270	0.250	7.4	79	0.00
69		Atrazine	0.232	0.224	3.4	85	0.00
70	C	Pentachlorophenol	0.169	0.147	13.0	76	0.00
71		Phenanthrene	1.094	0.982	10.2	78	0.00
72		Anthracene	1.113	1.007	9.5	79	0.00
73		Carbazole	0.981	0.860	12.3	77	0.00
74		Di-n-butylphthalate	1.134	0.984	13.2	73	0.00
75	C	Fluoranthene	1.370	1.167	14.8	75	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	88	0.00
77		Benzidine	0.448	0.607	-35.5#	109	0.00
78		Pyrene	1.307	1.180	9.7	74	0.00
79	S	Terphenyl-d14	1.060	0.981	7.5	76	0.00
80		Butylbenzylphthalate	0.395	0.382	3.3	75	0.00
81		Benzo(a)anthracene	1.326	1.203	9.3	75	0.00
82		3,3'-Dichlorobenzidine	0.443	0.452	-2.0	83	0.00
83		Chrysene	1.262	1.121	11.2	74	0.00
84		Bis(2-ethylhexyl)phthalate	0.590	0.553	6.3	73	0.00
85	c	Di-n-octyl phthalate	0.995	0.898	9.7	73	0.00
86	I	Perylene-d12	1.000	1.000	0.0	87	0.00
87		Indeno(1,2,3-cd)pyrene	1.471	1.338	9.0	75	0.00
88		Benzo(b)fluoranthene	1.226	1.111	9.4	74	0.00
89		Benzo(k)fluoranthene	1.227	1.143	6.8	78	0.00
90	C	Benzo(a)pyrene	1.127	1.062	5.8	77	0.00
91		Dibenzo(a,h)anthracene	1.195	1.099	8.0	76	0.01
92		Benzo(g,h,i)perylene	1.142	1.065	6.7	77	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064577.D
Acq On : 24 Oct 2025 16:47
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG102425

Quant Time: Nov 04 00:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG No.: Q3434

Instrument ID: BNA_F Calibration Date/Time: 10/28/2025 13:17

Lab File ID: BF144097.D Init. Calib. Date(s): 10/24/2025 10/24/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:13 16:24

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.264	1.303		3.1	
Benzaldehyde	1.084	1.105		1.9	
Phenol-d6	1.396	1.466		5.0	
Phenol	1.722	1.775		3.1	20.0
bis(2-Chloroethyl)ether	1.245	1.329		6.7	
2-Chlorophenol	1.245	1.300		4.4	
2-Methylphenol	0.957	1.028		7.4	
2,2-oxybis(1-Chloropropane)	2.339	2.513		7.4	
Acetophenone	0.432	0.450		4.2	
3+4-Methylphenols	1.132	1.216		7.4	
n-Nitroso-di-n-propylamine	0.903	1.004	0.050	11.2	
Nitrobenzene-d5	0.365	0.371		1.6	
Hexachloroethane	0.541	0.558		3.1	
Nitrobenzene	0.386	0.408		5.7	
Isophorone	0.638	0.714		11.9	
2-Nitrophenol	0.184	0.203		10.3	20.0
2,4-Dimethylphenol	0.273	0.303		11.0	
bis(2-Chloroethoxy)methane	0.398	0.444		11.6	
2,4-Dichlorophenol	0.317	0.338		6.6	20.0
Naphthalene	0.944	0.987		4.6	
4-Chloroaniline	0.329	0.364		10.6	
Hexachlorobutadiene	0.267	0.269		0.7	20.0
Caprolactam	0.080	0.097		21.3	
4-Chloro-3-methylphenol	0.280	0.322		15.0	20.0
2-Methylnaphthalene	0.622	0.666		7.1	
Hexachlorocyclopentadiene	0.212	0.238	0.050	12.3	
2,4,6-Trichlorophenol	0.447	0.474		6.0	20.0
2-Fluorobiphenyl	1.421	1.379		-3.0	
2,4,5-Trichlorophenol	0.469	0.522		11.3	
1,1-Biphenyl	1.432	1.391		-2.9	
2-Chloronaphthalene	1.199	1.218		1.6	
2-Nitroaniline	0.356	0.388		9.0	
Dimethylphthalate	1.311	1.432		9.2	
Acenaphthylene	1.617	1.684		4.1	
2,6-Dinitrotoluene	0.259	0.297		14.7	
3-Nitroaniline	0.288	0.326		13.2	
Acenaphthene	1.080	1.111		2.9	20.0
2,4-Dinitrophenol	0.140	0.166	0.050	18.6	
4-Nitrophenol	0.195	0.221	0.050	13.3	

Form VII SV-1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG No.: Q3434

Instrument ID: BNA_F Calibration Date/Time: 10/28/2025 13:17

Lab File ID: BF144097.D Init. Calib. Date(s): 10/24/2025 10/24/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:13 16:24

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.499	1.565		4.4	
2,4-Dinitrotoluene	0.350	0.407		16.3	
Diethylphthalate	1.233	1.355		9.9	
4-Chlorophenyl-phenylether	0.643	0.693		7.8	
Fluorene	1.143	1.207		5.6	
4-Nitroaniline	0.273	0.304		11.4	
4,6-Dinitro-2-methylphenol	0.128	0.136		6.3	
n-Nitrosodiphenylamine	0.639	0.653		2.2	20.0
2,4,6-Tribromophenol	0.255	0.287		12.5	
4-Bromophenyl-phenylether	0.250	0.266		6.4	
Hexachlorobenzene	0.307	0.322		4.9	
Atrazine	0.210	0.208		-1.0	
Pentachlorophenol	0.146	0.170		16.4	20.0
Phenanthrene	0.994	1.013		1.9	
Anthracene	1.026	1.035		0.9	
Carbazole	0.874	0.869		-0.6	
Di-n-butylphthalate	1.071	1.082		1.0	
Fluoranthene	1.098	1.000		-8.9	20.0
Pyrene	1.438	2.068		43.8	
Terphenyl-d14	1.144	1.573		37.5	
Butylbenzylphthalate	0.544	0.651		19.7	
3,3-Dichlorobenzidine	0.470	0.517		10.0	
Benzo(a)anthracene	1.352	1.480		9.5	
Chrysene	1.268	1.334		5.2	
Bis(2-ethylhexyl)phthalate	0.743	0.775		4.3	
Di-n-octyl phthalate	1.277	1.323		3.6	20.0
Benzo(b)fluoranthene	1.131	1.126		-0.4	
Benzo(k)fluoranthene	1.043	1.047		0.4	
Benzo(a)pyrene	1.014	1.094		7.9	20.0
Indeno(1,2,3-cd)pyrene	1.373	1.546		12.6	
Dibenzo(a,h)anthracene	1.087	1.218		12.1	
Benzo(g,h,i)perylene	1.093	1.265		15.7	
1,2,4,5-Tetrachlorobenzene	0.687	0.668		-2.8	
1,4-Dioxane	0.546	0.520		-4.8	20.0
2,3,4,6-Tetrachlorophenol	0.330	0.372		12.7	

All other compounds must meet a minimum RRF of 0.010.

Form VII SV-1

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	66521	20.000	ng		0.00
21) Naphthalene-d8	8.163	136	250146	20.000	ng		0.00
39) Acenaphthene-d10	9.916	164	147935	20.000	ng		0.00
64) Phenanthrene-d10	11.410	188	264042	20.000	ng		0.00
76) Chrysene-d12	14.057	240	128248	20.000	ng		0.00
86) Perylene-d12	15.539	264	169914	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.498	112	346602	82.434	ng		0.00
7) Phenol-d6	6.510	99	390159	84.022	ng		0.00
23) Nitrobenzene-d5	7.445	82	371128	81.285	ng		0.00
42) 2,4,6-Tribromophenol	10.710	330	169601	90.002	ng		0.00
45) 2-Fluorobiphenyl	9.239	172	816145	77.635	ng		0.00
79) Terphenyl-d14	12.992	244	806709	109.967	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.628	88	69149	38.059	ng		96
3) Pyridine	3.404	79	202882	41.710	ng		96
4) n-Nitrosodimethylamine	3.351	42	112807	37.743	ng		91
6) Aniline	6.545	93	275642	41.492	ng		99
8) 2-Chlorophenol	6.663	128	172921	41.762	ng		100
9) Benzaldehyde	6.434	77	147003	40.774	ng		98
10) Phenol	6.528	94	236104	41.214	ng		92
11) bis(2-Chloroethyl)ether	6.616	93	176755	42.699	ng		96
12) 1,3-Dichlorobenzene	6.822	146	212619	41.627	ng		98
13) 1,4-Dichlorobenzene	6.898	146	210959	41.909	ng		98
14) 1,2-Dichlorobenzene	7.051	146	203316	41.744	ng		98
15) Benzyl Alcohol	7.022	79	164325	43.521	ng		97
16) 2,2'-oxybis(1-Chloropr...	7.151	45	334275	42.966	ng		96
17) 2-Methylphenol	7.134	107	136788	42.952	ng		97
18) Hexachloroethane	7.392	117	74260	41.263	ng		98
19) n-Nitroso-di-n-propyla...	7.292	70	133622	44.499	ng		92
20) 3+4-Methylphenols	7.287	107	161750	42.964	ng	#	86
22) Acetophenone	7.292	105	224937	41.609	ng		95
24) Nitrobenzene	7.463	77	204243	42.287	ng		99
25) Isophorone	7.704	82	357344	44.766	ng		99
26) 2-Nitrophenol	7.781	139	101537	44.191	ng		95
27) 2,4-Dimethylphenol	7.816	122	151833	44.467	ng		98
28) bis(2-Chloroethoxy)met...	7.910	93	222369	44.670	ng		97
29) 2,4-Dichlorophenol	8.022	162	169254	42.625	ng		97
30) 1,2,4-Trichlorobenzene	8.104	180	172640	42.262	ng		99
31) Naphthalene	8.186	128	493628	41.811	ng		98
32) Benzoic acid	7.934	122	115219	53.198	ng	#	84
33) 4-Chloroaniline	8.234	127	182107	44.308	ng		98
34) Hexachlorobutadiene	8.298	225	134816	40.361	ng		99
35) Caprolactam	8.610	113	48594	48.802	ng		91
36) 4-Chloro-3-methylphenol	8.716	107	161239	46.014	ng		96
37) 2-Methylnaphthalene	8.875	142	333422	42.835	ng		96
38) 1-Methylnaphthalene	8.975	142	320892	43.513	ng		100
40) 1,2,4,5-Tetrachloroben...	9.039	216	197569	38.859	ng		98
41) Hexachlorocyclopentadiene	9.028	237	70289	44.760	ng		97
43) 2,4,6-Trichlorophenol	9.151	196	140181	42.418	ng		98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	154404	44.519	ng	96
46) 1,1'-Biphenyl	9.339	154	411658	38.873	ng	100
47) 2-Chloronaphthalene	9.363	162	360439	40.627	ng	97
48) 2-Nitroaniline	9.463	65	114922	43.622	ng	98
49) Acenaphthylene	9.780	152	498378	41.661	ng	99
50) Dimethylphthalate	9.633	163	423610	43.685	ng	100
51) 2,6-Dinitrotoluene	9.704	165	87885	45.846	ng	92
52) Acenaphthene	9.951	154	328569	41.127	ng	98
53) 3-Nitroaniline	9.875	138	96397	45.229	ng	98
54) 2,4-Dinitrophenol	9.980	184	49169	47.423	ng	98
55) Dibenzofuran	10.128	168	463130	41.757	ng	97
56) 4-Nitrophenol	10.033	139	65421	45.329	ng	92
57) 2,4-Dinitrotoluene	10.110	165	120319	46.505	ng	95
58) Fluorene	10.469	166	357025	42.247	ng	97
59) 2,3,4,6-Tetrachlorophenol	10.245	232	109986	45.002	ng	98
60) Diethylphthalate	10.339	149	401002	43.974	ng	100
61) 4-Chlorophenyl-phenyle...	10.457	204	205056	43.147	ng	97
62) 4-Nitroaniline	10.486	138	89874	44.504	ng	96
63) Azobenzene	10.622	77	376152	43.581	ng	98
65) 4,6-Dinitro-2-methylph...	10.522	198	71598	42.264	ng	98
66) n-Nitrosodiphenylamine	10.580	169	344740	40.864	ng	98
67) 4-Bromophenyl-phenylether	10.951	248	140573	42.521	ng	97
68) Hexachlorobenzene	11.022	284	169947	41.916	ng	96
69) Atrazine	11.104	200	109902	39.700	ng	96
70) Pentachlorophenol	11.216	266	89552	46.401	ng	96
71) Phenanthrene	11.433	178	534720	40.767	ng	99
72) Anthracene	11.486	178	546494	40.333	ng	98
73) Carbazole	11.639	167	458674	39.771	ng	98
74) Di-n-butylphthalate	11.969	149	571521	40.420	ng	100
75) Fluoranthene	12.621	202	528275	36.455	ng	99
77) Benzidine	12.751	184	201207	50.411	ng	99
78) Pyrene	12.857	202	530531	57.540	ng	99
80) Butylbenzylphthalate	13.468	149	167106	47.907	ng	98
81) Benzo(a)anthracene	14.045	228	379691	43.808	ng	99
82) 3,3'-Dichlorobenzidine	14.004	252	132543	43.932	ng	100
83) Chrysene	14.080	228	342254	42.081	ng	99
84) Bis(2-ethylhexyl)phtha...	14.027	149	198759	41.739	ng	99
85) Di-n-octyl phthalate	14.639	149	339385	41.443	ng	100
87) Indeno(1,2,3-cd)pyrene	17.045	276	525344	45.034	ng	98
88) Benzo(b)fluoranthene	15.104	252	382543	39.824	ng	100
89) Benzo(k)fluoranthene	15.133	252	355845	40.151	ng	100
90) Benzo(a)pyrene	15.474	252	371712	43.134	ng	99
91) Dibenzo(a,h)anthracene	17.062	278	413954	44.829	ng	99
92) Benzo(g,h,i)perylene	17.503	276	429795	46.294	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 66521

Ion	Ratio	Lower	Upper
152	100		
150	160.2	124.4	186.6
115	59.5	49.4	74.0







#2
1,4-Dioxane
Concen: 38.059 ng
RT: 2.628 min Scan# 74
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion: 88 Resp: 69149

Ion	Ratio	Lower	Upper
88	100		
58	91.4	76.5	114.7
43	41.2	35.3	52.9







#3
Pyridine
Concen: 41.710 ng
RT: 3.404 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 202882

Ion	Ratio	Lower	Upper
79	100		
52	80.3	67.5	101.3
51	40.5	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 37.743 ng
RT: 3.351 min Scan# 197
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion: 42 Resp: 112807

Ion	Ratio	Lower	Upper
42	100		
74	105.1	76.4	114.6
44	6.5	5.2	7.8







#5
2-Fluorophenol
Concen: 82.434 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 346602

Ion	Ratio	Lower	Upper
112	100		
64	58.8	49.8	74.8
63	31.4	27.4	41.2







#6
Aniline
Concen: 41.492 ng
RT: 6.545 min Scan# 740
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion: 93 Resp: 275642

Ion	Ratio	Lower	Upper
93	100		
66	43.3	35.7	53.5
65	21.2	17.0	25.6







#7
Phenol-d6
Concen: 84.022 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 390159

Ion	Ratio	Lower	Upper
99	100		
42	27.7	23.9	35.9
71	41.7	34.4	51.6







#8
2-Chlorophenol
Concen: 41.762 ng
RT: 6.663 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:128 Resp: 172921

Ion	Ratio	Lower	Upper
128	100		
130	30.7	10.7	50.7
64	54.4	34.7	74.7







#9
Benzaldehyde
Concen: 40.774 ng
RT: 6.434 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 147003

Ion	Ratio	Lower	Upper
77	100		
105	94.9	72.2	112.2
106	86.5	67.9	107.9







#10
Phenol
Concen: 41.214 ng
RT: 6.528 min Scan# 737
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion: 94 Resp: 236104

Ion	Ratio	Lower	Upper
94	100		
65	36.6	13.1	53.1
66	57.4	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 42.699 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 176755

Ion	Ratio	Lower	Upper
93	100		
63	75.1	60.2	100.2
95	32.4	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 41.627 ng
RT: 6.822 min Scan# 787
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:146 Resp: 212619

Ion	Ratio	Lower	Upper
146	100		
148	63.7	50.6	75.8
75	30.9	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 41.909 ng
RT: 6.898 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 210959

Ion	Ratio	Lower	Upper
146	100		
148	62.8	51.0	76.6
111	41.3	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 41.744 ng
RT: 7.051 min Scan# 826
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:146 Resp: 203316

Ion	Ratio	Lower	Upper
146	100		
148	62.9	50.8	76.2
111	41.6	34.7	52.1







#15
Benzyl Alcohol
Concen: 43.521 ng
RT: 7.022 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 164325

Ion	Ratio	Lower	Upper
79	100		
108	70.1	54.8	82.2
77	65.3	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.966 ng
RT: 7.151 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion: 45 Resp: 334275

Ion	Ratio	Lower	Upper
45	100		
77	12.6	0.0	30.9
79	9.1	0.0	28.2







#17
2-Methylphenol
Concen: 42.952 ng
RT: 7.134 min Scan# 84
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 136788

Ion	Ratio	Lower	Upper
107	100		
108	115.1	96.4	144.6
77	59.2	46.8	70.2
79	57.6	45.8	68.6







#18
Hexachloroethane
Concen: 41.263 ng
RT: 7.392 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:117 Resp: 74260

Ion	Ratio	Lower	Upper
117	100		
119	96.3	76.6	115.0
201	123.3	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 44.499 ng
RT: 7.292 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 133622

Ion	Ratio	Lower	Upper
70	100		
42	75.4	67.4	101.0
101	8.3	7.1	10.7
130	15.7	13.5	20.3







#20
3+4-Methylphenols
Concen: 42.964 ng
RT: 7.287 min Scan# 866
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:107 Resp: 161750

Ion	Ratio	Lower	Upper
107	100		
108	91.0	68.0	108.0
77	157.2	108.7	148.7#
79	37.0	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 250146

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	13.0
54	8.4	6.6	9.8
68	5.4	4.6	6.8







#22
Acetophenone
Concen: 41.609 ng
RT: 7.292 min Scan# 867
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:105 Resp: 224937

Ion	Ratio	Lower	Upper
105	100		
71	9.6	6.9	10.3
51	46.2	40.6	60.8
120	21.9	16.8	25.2







#23
Nitrobenzene-d5
Concen: 81.285 ng
RT: 7.445 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 371128

Ion	Ratio	Lower	Upper
82	100		
128	42.6	32.4	48.6
54	66.0	54.0	81.0







#24
Nitrobenzene
Concen: 42.287 ng
RT: 7.463 min Scan# 896
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion: 77 Resp: 204243

Ion	Ratio	Lower	Upper
77	100		
123	43.1	35.0	52.6
65	15.3	12.2	18.2







#25
Isophorone
Concen: 44.766 ng
RT: 7.704 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 357344

Ion	Ratio	Lower	Upper
82	100		
95	7.8	6.6	9.8
138	17.9	13.8	20.8







#26
2-Nitrophenol
Concen: 44.191 ng
RT: 7.781 min Scan# 950
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:139 Resp: 101537

Ion	Ratio	Lower	Upper
139	100		
109	29.1	24.6	37.0
65	53.7	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 44.467 ng
RT: 7.816 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 151833

Ion	Ratio	Lower	Upper
122	100		
107	113.4	91.8	137.8
121	54.7	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 44.670 ng
RT: 7.910 min Scan# 972
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion: 93 Resp: 222369

Ion	Ratio	Lower	Upper
93	100		
95	32.7	24.6	36.8
123	10.9	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 42.625 ng
RT: 8.022 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 169254

Ion	Ratio	Lower	Upper
162	100		
164	65.0	42.1	82.1
98	34.8	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 42.262 ng
RT: 8.104 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:180 Resp: 172640

Ion	Ratio	Lower	Upper
180	100		
182	108.2	87.1	130.7
145	31.4	25.3	37.9







#31
Naphthalene
Concen: 41.811 ng
RT: 8.186 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 493628

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.4	12.6
127	13.0	9.9	14.9







#32
Benzoic acid
Concen: 53.198 ng
RT: 7.934 min Scan# 976
Delta R.T. 0.018 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:122 Resp: 115219

Ion	Ratio	Lower	Upper
122	100		
105	115.8	111.3	151.3
77	96.1	96.7	136.7#







#33
4-Chloroaniline
Concen: 44.308 ng
RT: 8.234 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 182107

Ion	Ratio	Lower	Upper
127	100		
129	33.3	25.6	38.4
65	37.3	30.9	46.3
92	23.1	18.9	28.3







#34
Hexachlorobutadiene
Concen: 40.361 ng
RT: 8.298 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:225 Resp: 134816

Ion	Ratio	Lower	Upper
225	100		
223	63.8	50.6	76.0
227	63.8	51.8	77.8







#35
Caprolactam
Concen: 48.802 ng
RT: 8.610 min Scan# 1
Delta R.T. 0.018 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 48594

Ion	Ratio	Lower	Upper
113	100		
55	193.0	160.1	200.1
56	157.1	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 46.014 ng
RT: 8.716 min Scan# 1109
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:107 Resp: 161239

Ion	Ratio	Lower	Upper
107	100		
144	28.4	21.6	32.4
142	77.3	65.4	98.0







#37
2-Methylnaphthalene
Concen: 42.835 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 333422

Ion	Ratio	Lower	Upper
142	100		
141	88.3	67.6	101.4







#38
1-Methylnaphthalene
Concen: 43.513 ng
RT: 8.975 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:142 Resp: 320892

Ion	Ratio	Lower	Upper
142	100		
141	91.5	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 147935

Ion	Ratio	Lower	Upper
164	100		
162	101.3	79.4	119.2
160	47.0	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.859 ng
RT: 9.039 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:216 Resp: 197569

Ion	Ratio	Lower	Upper
216	100		
214	80.2	63.2	94.8
179	20.0	15.2	22.8
108	13.8	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 44.760 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 70289

Ion	Ratio	Lower	Upper
237	100		
235	66.9	44.3	84.3
272	14.8	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 90.002 ng
RT: 10.710 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:330 Resp: 169601

Ion	Ratio	Lower	Upper
330	100		
332	94.4	77.0	115.6
141	22.8	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 42.418 ng
RT: 9.151 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 140181

Ion	Ratio	Lower	Upper
196	100		
198	95.6	74.4	111.6
200	30.2	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 44.519 ng
RT: 9.198 min Scan# 1191
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:196 Resp: 154404

Ion	Ratio	Lower	Upper
196	100		
198	95.9	78.0	117.0
97	40.6	37.7	56.5
132	23.8	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 77.635 ng
RT: 9.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 816145

Ion	Ratio	Lower	Upper
172	100		
171	34.2	27.5	41.3
170	23.5	19.0	28.6







#46
1,1'-Biphenyl
Concen: 38.873 ng
RT: 9.339 min Scan# 1215
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:154 Resp: 411658

Ion	Ratio	Lower	Upper
154	100		
153	39.8	19.7	59.7
76	11.1	0.0	31.1







#47
2-Chloronaphthalene
Concen: 40.627 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 360439

Ion	Ratio	Lower	Upper
162	100		
127	37.6	31.6	47.4
164	32.0	26.5	39.7







#48
2-Nitroaniline
Concen: 43.622 ng
RT: 9.463 min Scan# 1236
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion: 65 Resp: 114922

Ion	Ratio	Lower	Upper
65	100		
92	64.5	52.3	78.5
138	101.0	78.3	117.5







#49
Acenaphthylene
Concen: 41.661 ng
RT: 9.780 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 498378

Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.9	23.9
153	12.8	10.4	15.6







#50
Dimethylphthalate
Concen: 43.685 ng
RT: 9.633 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:163 Resp: 423610

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.0
164	9.7	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 45.846 ng
RT: 9.704 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 87885

Ion	Ratio	Lower	Upper
165	100		
63	55.4	50.4	75.6
89	47.4	41.4	62.2







#52
Acenaphthene
Concen: 41.127 ng
RT: 9.951 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:154 Resp: 328569

Ion	Ratio	Lower	Upper
154	100		
153	114.5	89.9	134.9
152	53.7	42.0	63.0







#53
3-Nitroaniline
Concen: 45.229 ng
RT: 9.875 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 96397

Ion	Ratio	Lower	Upper
138	100		
108	9.6	7.9	11.9
92	112.9	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 47.423 ng
RT: 9.980 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:184 Resp: 49169

Ion	Ratio	Lower	Upper
184	100		
63	64.4	52.1	78.1
154	61.9	47.9	71.9







#55
Dibenzofuran
Concen: 41.757 ng
RT: 10.128 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 463130

Ion	Ratio	Lower	Upper
168	100		
139	39.4	33.2	49.8
169	15.6	12.6	19.0







#56
4-Nitrophenol
Concen: 45.329 ng
RT: 10.033 min Scan# 1333
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:139 Resp: 65421

Ion	Ratio	Lower	Upper
139	100		
109	68.3	53.8	93.8
65	106.5	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 46.505 ng
RT: 10.110 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 120319

Ion	Ratio	Lower	Upper
165	100		
63	48.4	35.4	53.2
89	62.0	52.7	79.1
182	2.5	2.3	3.5







#58
Fluorene
Concen: 42.247 ng
RT: 10.469 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:166 Resp: 357025

Ion	Ratio	Lower	Upper
166	100		
165	98.9	77.0	115.4
167	13.4	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 45.002 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 109986

Ion	Ratio	Lower	Upper
232	100		
131	41.7	31.9	47.9
130	1.7	1.1	1.7
166	26.5	21.0	31.4







#60
Diethylphthalate
Concen: 43.974 ng
RT: 10.339 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:149 Resp: 401002

Ion	Ratio	Lower	Upper
149	100		
177	22.1	17.8	26.6
150	12.8	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 43.147 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 205056

Ion	Ratio	Lower	Upper
204	100		
206	33.0	27.8	41.6
141	55.1	45.4	68.2







#62
4-Nitroaniline
Concen: 44.504 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:138 Resp: 89874

Ion	Ratio	Lower	Upper
138	100		
92	50.4	30.5	70.5
108	70.2	56.2	96.2







#63
Azobenzene
Concen: 43.581 ng
RT: 10.622 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 376152

Ion	Ratio	Lower	Upper
77	100		
182	29.8	9.0	49.0
105	21.4	1.4	41.4
51	34.9	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:188 Resp: 264042

Ion	Ratio	Lower	Upper
188	100		
94	6.8	5.4	8.0
80	7.4	6.0	9.0







#65
4,6-Dinitro-2-methylphenol
Concen: 42.264 ng
RT: 10.522 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 71598

Ion	Ratio	Lower	Upper
198	100		
51	45.7	25.8	65.8
105	34.3	17.2	57.2







#66
n-Nitrosodiphenylamine
Concen: 40.864 ng
RT: 10.580 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:169 Resp: 344740

Ion	Ratio	Lower	Upper
169	100		
168	57.2	47.4	71.2
167	32.9	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 42.521 ng
RT: 10.951 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 140573

Ion	Ratio	Lower	Upper
248	100		
250	100.1	77.8	116.6
141	61.1	50.5	75.7







#68
Hexachlorobenzene
Concen: 41.916 ng
RT: 11.022 min Scan# 1501
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:284 Resp: 169947

Ion	Ratio	Lower	Upper
284	100		
142	21.5	20.0	30.0
249	27.2	22.2	33.2







#69
Atrazine
Concen: 39.700 ng
RT: 11.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 109902

Ion	Ratio	Lower	Upper
200	100		
173	26.5	4.9	44.9
215	51.3	28.2	68.2







#70
Pentachlorophenol
Concen: 46.401 ng
RT: 11.216 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:266 Resp: 89552

Ion	Ratio	Lower	Upper
266	100		
268	65.0	48.2	72.4
264	63.6	49.4	74.0







#71
Phenanthrene
Concen: 40.767 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 534720

Ion	Ratio	Lower	Upper
178	100		
176	19.7	16.3	24.5
179	15.4	12.6	18.8







#72
Anthracene
Concen: 40.333 ng
RT: 11.486 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:178 Resp: 546494

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.0	22.4
179	15.4	11.8	17.6







#73
Carbazole
Concen: 39.771 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 458674

Ion	Ratio	Lower	Upper
167	100		
166	21.1	16.1	24.1
139	13.0	9.8	14.6







#74
Di-n-butylphthalate
Concen: 40.420 ng
RT: 11.969 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:149 Resp: 571521

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.3	10.9
104	4.3	3.6	5.4







#75
Fluoranthene
Concen: 36.455 ng
RT: 12.621 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 528275

Ion	Ratio	Lower	Upper
202	100		
101	8.4	0.0	27.7
203	17.7	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:240 Resp: 128248

Ion	Ratio	Lower	Upper
240	100		
120	7.8	6.5	9.7
236	26.3	20.1	30.1







#77
Benzidine
Concen: 50.411 ng
RT: 12.751 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 201207

Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.6	17.4
183	10.2	8.5	12.7







#78
Pyrene
Concen: 57.540 ng
RT: 12.857 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:202 Resp: 530531

Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.2	25.8
203	17.4	13.5	20.3







#79
Terphenyl-d14
Concen: 109.967 ng
RT: 12.992 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 806709

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.8	8.6
122	7.1	4.9	7.3







#80
Butylbenzylphthalate
Concen: 47.907 ng
RT: 13.468 min Scan# 1917
Delta R.T. -0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:149 Resp: 167106

Ion	Ratio	Lower	Upper
149	100		
91	68.3	56.6	84.8
206	22.9	18.3	27.5







#81
Benzo(a)anthracene
Concen: 43.808 ng
RT: 14.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 379691

Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.4	32.2
229	19.8	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 43.932 ng
RT: 14.004 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:252 Resp: 132543

Ion	Ratio	Lower	Upper
252	100		
254	62.3	49.9	74.9
126	8.1	6.5	9.7







#83
Chrysene
Concen: 42.081 ng
RT: 14.080 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 342254

Ion	Ratio	Lower	Upper
228	100		
226	28.7	23.8	35.6
229	19.8	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 41.739 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:149 Resp: 198759

Ion	Ratio	Lower	Upper
149	100		
167	27.5	22.5	33.7
279	4.6	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 41.443 ng
RT: 14.639 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 339385

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	11.5	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:264 Resp: 169914

Ion	Ratio	Lower	Upper
264	100		
260	21.2	17.0	25.4
265	21.2	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 45.034 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 525344

Ion	Ratio	Lower	Upper
276	100		
138	16.4	13.0	19.6
277	25.5	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 39.824 ng
RT: 15.104 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:252 Resp: 382543

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.2	25.8
125	6.3	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 40.151 ng
RT: 15.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 355845

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.2
125	6.0	4.9	7.3







#90
Benzo(a)pyrene
Concen: 43.134 ng
RT: 15.474 min Scan# 2258
Delta R.T. -0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:252 Resp: 371712

Ion	Ratio	Lower	Upper
252	100		
253	20.7	17.1	25.7
125	7.0	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 44.829 ng
RT: 17.062 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 413954

Ion	Ratio	Lower	Upper
278	100		
139	10.5	8.5	12.7
279	24.2	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 46.294 ng
RT: 17.503 min Scan# 2603
Delta R.T. 0.006 min
Lab File: BF144097.D
Acq: 28 Oct 2025 13:17

Tgt Ion:276 Resp: 429795

Ion	Ratio	Lower	Upper
276	100		
277	21.8	17.9	26.9
138	14.8	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2		1,4-Dioxane	0.546	0.520	4.8	127	0.00
3		Pyridine	1.462	1.525	-4.3	134	0.00
4		n-Nitrosodimethylamine	0.899	0.848	5.7	125	0.00
5	S	2-Fluorophenol	1.264	1.303	-3.1	134	0.00
6		Aniline	1.997	2.072	-3.8	138	0.00
7	S	Phenol-d6	1.396	1.466	-5.0	140	0.00
8		2-Chlorophenol	1.245	1.300	-4.4	140	0.00
9		Benzaldehyde	1.084	1.105	-1.9	142	0.00
10	C	Phenol	1.722	1.775	-3.1	138	0.00
11		bis(2-Chloroethyl)ether	1.245	1.329	-6.7	143	0.00
12		1,3-Dichlorobenzene	1.536	1.598	-4.0	137	0.00
13	C	1,4-Dichlorobenzene	1.513	1.586	-4.8	137	0.00
14		1,2-Dichlorobenzene	1.464	1.528	-4.4	139	0.00
15		Benzyl Alcohol	1.135	1.235	-8.8	145	0.00
16		2,2'-oxybis(1-Chloropropane	2.339	2.513	-7.4	142	0.00
17		2-Methylphenol	0.957	1.028	-7.4	149	0.00
18		Hexachloroethane	0.541	0.558	-3.1	138	0.00
19	P	n-Nitroso-di-n-propylamine	0.903	1.004	-11.2	154#	0.00
20		3+4-Methylphenols	1.132	1.216	-7.4	140	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	140	0.00
22		Acetophenone	0.432	0.450	-4.2	145	0.00
23	S	Nitrobenzene-d5	0.365	0.371	-1.6	142	0.00
24		Nitrobenzene	0.386	0.408	-5.7	146	0.00
25		Isophorone	0.638	0.714	-11.9	156#	0.00
26	C	2-Nitrophenol	0.184	0.203	-10.3	149	0.00
27		2,4-Dimethylphenol	0.273	0.303	-11.0	148	0.00
28		bis(2-Chloroethoxy)methane	0.398	0.444	-11.6	154#	0.00
29	C	2,4-Dichlorophenol	0.317	0.338	-6.6	147	0.00
30		1,2,4-Trichlorobenzene	0.327	0.345	-5.5	145	0.00
31		Naphthalene	0.944	0.987	-4.6	143	0.00
32		Benzoic acid	0.173	0.230	-32.9#	222#	0.02
33		4-Chloroaniline	0.329	0.364	-10.6	152#	0.00
34	C	Hexachlorobutadiene	0.267	0.269	-0.7	140	0.00
35		Caprolactam	0.080	0.097	-21.3	161#	0.02
36	C	4-Chloro-3-methylphenol	0.280	0.322	-15.0	161#	0.00
37		2-Methylnaphthalene	0.622	0.666	-7.1	145	0.00
38		1-Methylnaphthalene	0.590	0.641	-8.6	152#	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	162#	0.00
40		1,2,4,5-Tetrachlorobenzene	0.687	0.668	2.8	153#	0.00
41	P	Hexachlorocyclopentadiene	0.212	0.238	-12.3	176#	0.00
42	S	2,4,6-Tribromophenol	0.255	0.287	-12.5	177#	0.00
43	C	2,4,6-Trichlorophenol	0.447	0.474	-6.0	162#	0.00
44		2,4,5-Trichlorophenol	0.469	0.522	-11.3	181#	0.00
45	S	2-Fluorobiphenyl	1.421	1.379	3.0	154#	0.00
46		1,1'-Biphenyl	1.432	1.391	2.9	153#	0.00
47		2-Chloronaphthalene	1.199	1.218	-1.6	161#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.356	0.388	-9.0	165#	0.00
49		Acenaphthylene	1.617	1.684	-4.1	160#	0.00
50		Dimethylphthalate	1.311	1.432	-9.2	169#	0.00
51		2,6-Dinitrotoluene	0.259	0.297	-14.7	174#	0.00
52	C	Acenaphthene	1.080	1.111	-2.9	161#	0.00
53		3-Nitroaniline	0.288	0.326	-13.2	163#	0.00
54	P	2,4-Dinitrophenol	0.140	0.166	-18.6	217#	0.00
55		Dibenzofuran	1.499	1.565	-4.4	162#	0.00
56	P	4-Nitrophenol	0.195	0.221	-13.3	191#	0.00
57		2,4-Dinitrotoluene	0.350	0.407	-16.3	170#	0.00
58		Fluorene	1.143	1.207	-5.6	164#	0.00
59		2,3,4,6-Tetrachlorophenol	0.330	0.372	-12.7	167#	0.00
60		Diethylphthalate	1.233	1.355	-9.9	170#	0.00
61		4-Chlorophenyl-phenylether	0.643	0.693	-7.8	166#	0.00
62		4-Nitroaniline	0.273	0.304	-11.4	166#	0.00
63		Azobenzene	1.167	1.271	-8.9	166#	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	165#	0.00
65		4,6-Dinitro-2-methylphenol	0.128	0.136	-6.3	183#	0.00
66	c	n-Nitrosodiphenylamine	0.639	0.653	-2.2	168#	0.00
67		4-Bromophenyl-phenylether	0.250	0.266	-6.4	169#	0.00
68		Hexachlorobenzene	0.307	0.322	-4.9	166#	0.00
69		Atrazine	0.210	0.208	1.0	165#	0.00
70	C	Pentachlorophenol	0.146	0.170	-16.4	187#	0.00
71		Phenanthrene	0.994	1.013	-1.9	165#	0.00
72		Anthracene	1.026	1.035	-0.9	163#	0.00
73		Carbazole	0.874	0.869	0.6	159#	0.00
74		Di-n-butylphthalate	1.071	1.082	-1.0	160#	0.00
75	C	Fluoranthene	1.098	1.000	8.9	147	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77		Benzidine	0.622	0.784	-26.0#	122	0.00
78		Pyrene	1.438	2.068	-43.8#	141	0.00
79	S	Terphenyl-d14	1.144	1.573	-37.5#	135	0.00
80		Butylbenzylphthalate	0.544	0.651	-19.7	112	0.00
81		Benzo(a)anthracene	1.352	1.480	-9.5	101	0.00
82		3,3'-Dichlorobenzidine	0.470	0.517	-10.0	105	0.00
83		Chrysene	1.268	1.334	-5.2	105	0.00
84		Bis(2-ethylhexyl)phthalate	0.743	0.775	-4.3	101	0.00
85	c	Di-n-octyl phthalate	1.277	1.323	-3.6	98	0.00
86	I	Perylene-d12	1.000	1.000	0.0	111	0.00
87		Indeno(1,2,3-cd)pyrene	1.373	1.546	-12.6	116	0.00
88		Benzo(b)fluoranthene	1.131	1.126	0.4	99	0.00
89		Benzo(k)fluoranthene	1.043	1.047	-0.4	110	0.00
90	C	Benzo(a)pyrene	1.014	1.094	-7.9	112	0.00
91		Dibenzo(a,h)anthracene	1.087	1.218	-12.1	116	0.00
92		Benzo(g,h,i)perylene	1.093	1.265	-15.7	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	135	0.00
2		1,4-Dioxane	40.000	38.059	4.9	127	0.00
3		Pyridine	40.000	41.710	-4.3	134	0.00
4		n-Nitrosodimethylamine	40.000	37.743	5.6	125	0.00
5	S	2-Fluorophenol	80.000	82.434	-3.0	134	0.00
6		Aniline	40.000	41.492	-3.7	138	0.00
7	S	Phenol-d6	80.000	84.022	-5.0	140	0.00
8		2-Chlorophenol	40.000	41.762	-4.4	140	0.00
9		Benzaldehyde	40.000	40.774	-1.9	142	0.00
10	C	Phenol	40.000	41.214	-3.0	138	0.00
11		bis(2-Chloroethyl)ether	40.000	42.699	-6.7	143	0.00
12		1,3-Dichlorobenzene	40.000	41.627	-4.1	137	0.00
13	C	1,4-Dichlorobenzene	40.000	41.909	-4.8	137	0.00
14		1,2-Dichlorobenzene	40.000	41.744	-4.4	139	0.00
15		Benzyl Alcohol	40.000	43.521	-8.8	145	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	42.966	-7.4	142	0.00
17		2-Methylphenol	40.000	42.952	-7.4	149	0.00
18		Hexachloroethane	40.000	41.263	-3.2	138	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	44.499	-11.2	154	0.00
20		3+4-Methylphenols	40.000	42.964	-7.4	140	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	140	0.00
22		Acetophenone	40.000	41.609	-4.0	145	0.00
23	S	Nitrobenzene-d5	80.000	81.285	-1.6	142	0.00
24		Nitrobenzene	40.000	42.287	-5.7	146	0.00
25		Isophorone	40.000	44.766	-11.9	156	0.00
26	C	2-Nitrophenol	40.000	44.191	-10.5	149	0.00
27		2,4-Dimethylphenol	40.000	44.467	-11.2	148	0.00
28		bis(2-Chloroethoxy)methane	40.000	44.670	-11.7	154	0.00
29	C	2,4-Dichlorophenol	40.000	42.625	-6.6	147	0.00
30		1,2,4-Trichlorobenzene	40.000	42.262	-5.7	145	0.00
31		Naphthalene	40.000	41.811	-4.5	143	0.00
32		Benzoic acid	40.000	53.198	-33.0#	222	0.02
33		4-Chloroaniline	40.000	44.308	-10.8	152	0.00
34	C	Hexachlorobutadiene	40.000	40.361	-0.9	140	0.00
35		Caprolactam	40.000	48.802	-22.0	161	0.02
36	C	4-Chloro-3-methylphenol	40.000	46.014	-15.0	161	0.00
37		2-Methylnaphthalene	40.000	42.835	-7.1	145	0.00
38		1-Methylnaphthalene	40.000	43.513	-8.8	152	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	162	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	38.859	2.9	153	0.00
41	P	Hexachlorocyclopentadiene	40.000	44.760	-11.9	176	0.00
42	S	2,4,6-Tribromophenol	80.000	90.002	-12.5	177	0.00
43	C	2,4,6-Trichlorophenol	40.000	42.418	-6.0	162	0.00
44		2,4,5-Trichlorophenol	40.000	44.519	-11.3	181	0.00
45	S	2-Fluorobiphenyl	80.000	77.635	3.0	154	0.00
46		1,1'-Biphenyl	40.000	38.873	2.8	153	0.00
47		2-Chloronaphthalene	40.000	40.627	-1.6	161	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	43.622	-9.1	165	0.00
49		Acenaphthylene	40.000	41.661	-4.2	160	0.00
50		Dimethylphthalate	40.000	43.685	-9.2	169	0.00
51		2,6-Dinitrotoluene	40.000	45.846	-14.6	174	0.00
52	C	Acenaphthene	40.000	41.127	-2.8	161	0.00
53		3-Nitroaniline	40.000	45.229	-13.1	163	0.00
54	P	2,4-Dinitrophenol	40.000	47.423	-18.6	217	0.00
55		Dibenzofuran	40.000	41.757	-4.4	162	0.00
56	P	4-Nitrophenol	40.000	45.329	-13.3	191	0.00
57		2,4-Dinitrotoluene	40.000	46.505	-16.3	170	0.00
58		Fluorene	40.000	42.247	-5.6	164	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	45.002	-12.5	167	0.00
60		Diethylphthalate	40.000	43.974	-9.9	170	0.00
61		4-Chlorophenyl-phenylether	40.000	43.147	-7.9	166	0.00
62		4-Nitroaniline	40.000	44.504	-11.3	166	0.00
63		Azobenzene	40.000	43.581	-9.0	166	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	165	0.00
65		4,6-Dinitro-2-methylphenol	40.000	42.264	-5.7	183	0.00
66	c	n-Nitrosodiphenylamine	40.000	40.864	-2.2	168	0.00
67		4-Bromophenyl-phenylether	40.000	42.521	-6.3	169	0.00
68		Hexachlorobenzene	40.000	41.916	-4.8	166	0.00
69		Atrazine	40.000	39.700	0.7	165	0.00
70	C	Pentachlorophenol	40.000	46.401	-16.0	187	0.00
71		Phenanthrene	40.000	40.767	-1.9	165	0.00
72		Anthracene	40.000	40.333	-0.8	163	0.00
73		Carbazole	40.000	39.771	0.6	159	0.00
74		Di-n-butylphthalate	40.000	40.420	-1.1	160	0.00
75	C	Fluoranthene	40.000	36.455	8.9	147	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77		Benzidine	40.000	50.411	-26.0#	122	0.00
78		Pyrene	40.000	57.540	-43.9#	141	0.00
79	S	Terphenyl-d14	80.000	109.967	-37.5#	135	0.00
80		Butylbenzylphthalate	40.000	47.907	-19.8	112	0.00
81		Benzo(a)anthracene	40.000	43.808	-9.5	101	0.00
82		3,3'-Dichlorobenzidine	40.000	43.932	-9.8	105	0.00
83		Chrysene	40.000	42.081	-5.2	105	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	41.739	-4.3	101	0.00
85	c	Di-n-octyl phthalate	40.000	41.443	-3.6	98	0.00
86	I	Perylene-d12	20.000	20.000	0.0	111	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	45.034	-12.6	116	0.00
88		Benzo(b)fluoranthene	40.000	39.824	0.4	99	0.00
89		Benzo(k)fluoranthene	40.000	40.151	-0.4	110	0.00
90	C	Benzo(a)pyrene	40.000	43.134	-7.8	112	0.00
91		Dibenzo(a,h)anthracene	40.000	44.829	-12.1	116	0.00
92		Benzo(g,h,i)perylene	40.000	46.294	-15.7	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144097.D
Acq On : 28 Oct 2025 13:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 14:10:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG No.: Q3434

Instrument ID: BNA_F Calibration Date/Time: 10/29/2025 11:30

Lab File ID: BF144121.D Init. Calib. Date(s): 10/24/2025 10/24/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:13 16:24

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.264	1.346		6.5	
Benzaldehyde	1.084	1.051		-3.0	
Phenol-d6	1.396	1.460		4.6	
Phenol	1.722	1.843		7.0	20.0
bis(2-Chloroethyl)ether	1.245	1.332		7.0	
2-Chlorophenol	1.245	1.284		3.1	
2-Methylphenol	0.957	1.014		6.0	
2,2-oxybis(1-Chloropropane)	2.339	2.460		5.2	
Acetophenone	0.432	0.429		-0.7	
3+4-Methylphenols	1.132	1.224		8.1	
n-Nitroso-di-n-propylamine	0.903	0.927	0.050	2.7	
Nitrobenzene-d5	0.365	0.361		-1.1	
Hexachloroethane	0.541	0.545		0.7	
Nitrobenzene	0.386	0.394		2.1	
Isophorone	0.638	0.649		1.7	
2-Nitrophenol	0.184	0.193		4.9	20.0
2,4-Dimethylphenol	0.273	0.290		6.2	
bis(2-Chloroethoxy)methane	0.398	0.409		2.8	
2,4-Dichlorophenol	0.317	0.328		3.5	20.0
Naphthalene	0.944	0.957		1.4	
4-Chloroaniline	0.329	0.349		6.1	
Hexachlorobutadiene	0.267	0.260		-2.6	20.0
Caprolactam	0.080	0.084		5.0	
4-Chloro-3-methylphenol	0.280	0.296		5.7	20.0
2-Methylnaphthalene	0.622	0.630		1.3	
Hexachlorocyclopentadiene	0.212	0.181	0.050	-14.6	
2,4,6-Trichlorophenol	0.447	0.466		4.3	20.0
2-Fluorobiphenyl	1.421	1.368		-3.7	
2,4,5-Trichlorophenol	0.469	0.490		4.5	
1,1-Biphenyl	1.432	1.430		-0.1	
2-Chloronaphthalene	1.199	1.242		3.6	
2-Nitroaniline	0.356	0.357		0.3	
Dimethylphthalate	1.311	1.355		3.4	
Acenaphthylene	1.617	1.679		3.8	
2,6-Dinitrotoluene	0.259	0.272		5.0	
3-Nitroaniline	0.288	0.311		8.0	
Acenaphthene	1.080	1.076		-0.4	20.0
2,4-Dinitrophenol	0.140	0.135	0.050	-3.6	
4-Nitrophenol	0.195	0.193	0.050	-1.0	

Form VII SV-1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG No.: Q3434

Instrument ID: BNA_F Calibration Date/Time: 10/29/2025 11:30

Lab File ID: BF144121.D Init. Calib. Date(s): 10/24/2025 10/24/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:13 16:24

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.499	1.529		2.0	
2,4-Dinitrotoluene	0.350	0.369		5.4	
Diethylphthalate	1.233	1.231		-0.2	
4-Chlorophenyl-phenylether	0.643	0.660		2.6	
Fluorene	1.143	1.160		1.5	
4-Nitroaniline	0.273	0.277		1.5	
4,6-Dinitro-2-methylphenol	0.128	0.128		0.0	
n-Nitrosodiphenylamine	0.639	0.669		4.7	20.0
2,4,6-Tribromophenol	0.255	0.252		-1.2	
4-Bromophenyl-phenylether	0.250	0.265		6.0	
Hexachlorobenzene	0.307	0.310		1.0	
Atrazine	0.210	0.208		-1.0	
Pentachlorophenol	0.146	0.162		11.0	20.0
Phenanthrene	0.994	1.031		3.7	
Anthracene	1.026	1.051		2.4	
Carbazole	0.874	0.888		1.6	
Di-n-butylphthalate	1.071	1.096		2.3	
Fluoranthene	1.098	1.097		-0.1	20.0
Pyrene	1.438	1.509		4.9	
Terphenyl-d14	1.144	1.166		1.9	
Butylbenzylphthalate	0.544	0.562		3.3	
3,3-Dichlorobenzidine	0.470	0.520		10.6	
Benzo(a)anthracene	1.352	1.439		6.4	
Chrysene	1.268	1.339		5.6	
Bis(2-ethylhexyl)phthalate	0.743	0.791		6.5	
Di-n-octyl phthalate	1.277	1.456		14.0	20.0
Benzo(b)fluoranthene	1.131	1.146		1.3	
Benzo(k)fluoranthene	1.043	1.131		8.4	
Benzo(a)pyrene	1.014	1.085		7.0	20.0
Indeno(1,2,3-cd)pyrene	1.373	1.487		8.3	
Dibenzo(a,h)anthracene	1.087	1.186		9.1	
Benzo(g,h,i)perylene	1.093	1.183		8.2	
1,2,4,5-Tetrachlorobenzene	0.687	0.695		1.2	
1,4-Dioxane	0.546	0.560		2.6	20.0
2,3,4,6-Tetrachlorophenol	0.330	0.352		6.7	

All other compounds must meet a minimum RRF of 0.010.

Form VII SV-1

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	61447	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	232371	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	124035	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	207074	20.000	ng	0.00
76) Chrysene-d12	14.057	240	153131	20.000	ng	0.00
86) Perylene-d12	15.539	264	204043	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	330929	85.206	ng	0.00
7) Phenol-d6	6.504	99	358957	83.686	ng	0.00
23) Nitrobenzene-d5	7.445	82	335980	79.215	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	124844	79.017	ng	0.00
45) 2-Fluorobiphenyl	9.234	172	678951	77.029	ng	0.00
79) Terphenyl-d14	12.992	244	714068	81.522	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.610	88	68837	41.016	ng	98
3) Pyridine	3.381	79	193889	43.152	ng	98
4) n-Nitrosodimethylamine	3.328	42	101090	36.615	ng	93
6) Aniline	6.540	93	257784	42.009	ng	99
8) 2-Chlorophenol	6.663	128	157734	41.239	ng	99
9) Benzaldehyde	6.428	77	129140	38.777	ng	98
10) Phenol	6.522	94	226466	42.796	ng	97
11) bis(2-Chloroethyl)ether	6.616	93	163647	42.797	ng	94
12) 1,3-Dichlorobenzene	6.822	146	200122	42.415	ng	98
13) 1,4-Dichlorobenzene	6.898	146	196909	42.348	ng	98
14) 1,2-Dichlorobenzene	7.045	146	186585	41.472	ng	98
15) Benzyl Alcohol	7.022	79	144053	41.303	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.145	45	302282	42.062	ng	92
17) 2-Methylphenol	7.128	107	124615	42.361	ng	96
18) Hexachloroethane	7.392	117	66998	40.302	ng	99
19) n-Nitroso-di-n-propyla...	7.287	70	113931	41.075	ng	95
20) 3+4-Methylphenols	7.281	107	150418	43.254	ng	96
22) Acetophenone	7.287	105	199329	39.693	ng	100
24) Nitrobenzene	7.463	77	182876	40.760	ng	99
25) Isophorone	7.698	82	301404	40.647	ng	99
26) 2-Nitrophenol	7.775	139	89681	42.017	ng	98
27) 2,4-Dimethylphenol	7.810	122	134777	42.492	ng	98
28) bis(2-Chloroethoxy)met...	7.910	93	190207	41.132	ng	98
29) 2,4-Dichlorophenol	8.016	162	152408	41.318	ng	99
30) 1,2,4-Trichlorobenzene	8.104	180	155305	40.926	ng	98
31) Naphthalene	8.187	128	444899	40.566	ng	98
32) Benzoic acid	7.922	122	87292	43.387	ng	86
33) 4-Chloroaniline	8.234	127	162085	42.453	ng	96
34) Hexachlorobutadiene	8.298	225	120689	38.895	ng	99
35) Caprolactam	8.598	113	39199	42.378	ng	93
36) 4-Chloro-3-methylphenol	8.710	107	137486	42.237	ng	96
37) 2-Methylnaphthalene	8.875	142	292579	40.463	ng	95
38) 1-Methylnaphthalene	8.975	142	278578	40.665	ng	99
40) 1,2,4,5-Tetrachloroben...	9.039	216	172358	40.433	ng	99
41) Hexachlorocyclopentadiene	9.028	237	44780	34.011	ng	99
43) 2,4,6-Trichlorophenol	9.151	196	115708	41.760	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	121526	41.791	ng	96
46) 1,1'-Biphenyl	9.339	154	354773	39.957	ng	99
47) 2-Chloronaphthalene	9.363	162	307997	41.405	ng	96
48) 2-Nitroaniline	9.457	65	88438	40.037	ng	95
49) Acenaphthylene	9.781	152	416607	41.536	ng	99
50) Dimethylphthalate	9.634	163	336124	41.342	ng	99
51) 2,6-Dinitrotoluene	9.698	165	67580	42.046	ng	95
52) Acenaphthene	9.951	154	266968	39.855	ng	98
53) 3-Nitroaniline	9.869	138	77266	43.238	ng	99
54) 2,4-Dinitrophenol	9.981	184	33608	38.660	ng	90
55) Dibenzofuran	10.122	168	379280	40.786	ng	98
56) 4-Nitrophenol	10.034	139	47859	39.550	ng	89
57) 2,4-Dinitrotoluene	10.104	165	91495	42.178	ng	95
58) Fluorene	10.469	166	287737	40.609	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.239	232	87320	42.612	ng	98
60) Diethylphthalate	10.333	149	305410	39.944	ng	99
61) 4-Chlorophenyl-phenyle...	10.457	204	163745	41.094	ng	96
62) 4-Nitroaniline	10.486	138	68617	40.525	ng	94
63) Azobenzene	10.616	77	289198	39.963	ng	98
65) 4,6-Dinitro-2-methylph...	10.516	198	52916	39.830	ng	98
66) n-Nitrosodiphenylamine	10.575	169	277049	41.875	ng	99
67) 4-Bromophenyl-phenylether	10.951	248	109675	42.302	ng	96
68) Hexachlorobenzene	11.016	284	128445	40.395	ng	98
69) Atrazine	11.104	200	86008	39.616	ng	98
70) Pentachlorophenol	11.210	266	67248	44.431	ng	97
71) Phenanthrene	11.433	178	427157	41.526	ng	98
72) Anthracene	11.486	178	435076	40.944	ng	99
73) Carbazole	11.639	167	367780	40.663	ng	99
74) Di-n-butylphthalate	11.963	149	453989	40.941	ng	99
75) Fluoranthene	12.622	202	454294	39.974	ng	99
77) Benzidine	12.745	184	201262	42.231	ng	99
78) Pyrene	12.851	202	462034	41.969	ng	100
80) Butylbenzylphthalate	13.469	149	172222	41.351	ng	99
81) Benzo(a)anthracene	14.045	228	440710	42.586	ng	99
82) 3,3'-Dichlorobenzidine	14.004	252	159124	44.172	ng	100
83) Chrysene	14.080	228	410191	42.239	ng	99
84) Bis(2-ethylhexyl)phtha...	14.027	149	242344	42.623	ng	98
85) Di-n-octyl phthalate	14.639	149	446067	45.619	ng	99
87) Indeno(1,2,3-cd)pyrene	17.045	276	606674	43.307	ng	100
88) Benzo(b)fluoranthene	15.104	252	467688	40.544	ng	100
89) Benzo(k)fluoranthene	15.133	252	461643	43.376	ng	99
90) Benzo(a)pyrene	15.474	252	442928	42.801	ng	99
91) Dibenzo(a,h)anthracene	17.062	278	484044	43.651	ng	100
92) Benzo(g,h,i)perylene	17.504	276	482661	43.293	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 61447

Ion	Ratio	Lower	Upper
152	100		
150	165.8	124.4	186.6
115	61.9	49.4	74.0







#2
1,4-Dioxane
Concen: 41.016 ng
RT: 2.610 min Scan# 71
Delta R.T. -0.017 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion: 88 Resp: 68837

Ion	Ratio	Lower	Upper
88	100		
58	94.6	76.5	114.7
43	41.5	35.3	52.9







#3
Pyridine
Concen: 43.152 ng
RT: 3.381 min Scan# 2
Delta R.T. -0.018 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 193889

Ion	Ratio	Lower	Upper
79	100		
52	82.6	67.5	101.3
51	41.4	33.6	50.4







#4
n-Nitrosodimethylamine
Concen: 36.615 ng
RT: 3.328 min Scan# 193
Delta R.T. -0.017 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion: 42 Resp: 101090

Ion	Ratio	Lower	Upper
42	100		
74	102.5	76.4	114.6
44	6.5	5.2	7.8







#5
2-Fluorophenol
Concen: 85.206 ng
RT: 5.493 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 330929

Ion	Ratio	Lower	Upper
112	100		
64	59.1	49.8	74.8
63	31.6	27.4	41.2







#6
Aniline
Concen: 42.009 ng
RT: 6.540 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion: 93 Resp: 257784

Ion	Ratio	Lower	Upper
93	100		
66	43.5	35.7	53.5
65	21.3	17.0	25.6







#7
Phenol-d6
Concen: 83.686 ng
RT: 6.504 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 358957

Ion	Ratio	Lower	Upper
99	100		
42	27.4	23.9	35.9
71	41.4	34.4	51.6







#8
2-Chlorophenol
Concen: 41.239 ng
RT: 6.663 min Scan# 760
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:128 Resp: 157734

Ion	Ratio	Lower	Upper
128	100		
130	31.5	10.7	50.7
64	54.3	34.7	74.7







#9
Benzaldehyde
Concen: 38.777 ng
RT: 6.428 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 129140

Ion	Ratio	Lower	Upper
77	100		
105	89.9	72.2	112.2
106	89.8	67.9	107.9







#10
Phenol
Concen: 42.796 ng
RT: 6.522 min Scan# 736
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion: 94 Resp: 226466

Ion	Ratio	Lower	Upper
94	100		
65	31.9	13.1	53.1
66	48.1	30.6	70.6







#11
bis(2-Chloroethyl)ether
Concen: 42.797 ng
RT: 6.616 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 163647

Ion	Ratio	Lower	Upper
93	100		
63	73.6	60.2	100.2
95	31.3	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 42.415 ng
RT: 6.822 min Scan# 787
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:146 Resp: 200122

Ion	Ratio	Lower	Upper
146	100		
148	63.0	50.6	75.8
75	29.8	26.6	40.0







#13
1,4-Dichlorobenzene
Concen: 42.348 ng
RT: 6.898 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 196909

Ion	Ratio	Lower	Upper
146	100		
148	63.0	51.0	76.6
111	41.6	35.0	52.6







#14
1,2-Dichlorobenzene
Concen: 41.472 ng
RT: 7.045 min Scan# 825
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:146 Resp: 186585

Ion	Ratio	Lower	Upper
146	100		
148	61.9	50.8	76.2
111	44.3	34.7	52.1







#15
Benzyl Alcohol
Concen: 41.303 ng
RT: 7.022 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 144053

Ion	Ratio	Lower	Upper
79	100		
108	71.1	54.8	82.2
77	61.9	49.4	74.0







#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.062 ng
RT: 7.145 min Scan# 842
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion: 45 Resp: 302282

Ion	Ratio	Lower	Upper
45	100		
77	14.1	0.0	30.9
79	11.1	0.0	28.2







#17
2-Methylphenol
Concen: 42.361 ng
RT: 7.128 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 124615

Ion	Ratio	Lower	Upper
107	100		
108	115.6	96.4	144.6
77	55.0	46.8	70.2
79	54.7	45.8	68.6







#18
Hexachloroethane
Concen: 40.302 ng
RT: 7.392 min Scan# 884
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:117 Resp: 66998

Ion	Ratio	Lower	Upper
117	100		
119	97.6	76.6	115.0
201	125.8	100.9	151.3







#19
n-Nitroso-di-n-propylamine
Concen: 41.075 ng
RT: 7.287 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 113931

Ion	Ratio	Lower	Upper
70	100		
42	79.3	67.4	101.0
101	8.1	7.1	10.7
130	16.0	13.5	20.3







#20
3+4-Methylphenols
Concen: 43.254 ng
RT: 7.281 min Scan# 865
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:107 Resp: 150418

Ion	Ratio	Lower	Upper
107	100		
108	90.5	68.0	108.0
77	123.6	108.7	148.7
79	32.6	16.7	56.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 232371

Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.6	13.0
54	8.3	6.6	9.8
68	5.3	4.6	6.8







#22
Acetophenone
Concen: 39.693 ng
RT: 7.287 min Scan# 866
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:105 Resp: 199329

Ion	Ratio	Lower	Upper
105	100		
71	8.9	6.9	10.3
51	50.5	40.6	60.8
120	21.3	16.8	25.2







#23
Nitrobenzene-d5
Concen: 79.215 ng
RT: 7.445 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 335980

Ion	Ratio	Lower	Upper
82	100		
128	42.0	32.4	48.6
54	64.0	54.0	81.0







#24
Nitrobenzene
Concen: 40.760 ng
RT: 7.463 min Scan# 896
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion: 77 Resp: 182876

Ion	Ratio	Lower	Upper
77	100		
123	44.1	35.0	52.6
65	14.1	12.2	18.2







#25
Isophorone
Concen: 40.647 ng
RT: 7.698 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 301404

Ion	Ratio	Lower	Upper
82	100		
95	7.8	6.6	9.8
138	17.1	13.8	20.8







#26
2-Nitrophenol
Concen: 42.017 ng
RT: 7.775 min Scan# 949
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:139 Resp: 89681

Ion	Ratio	Lower	Upper
139	100		
109	28.8	24.6	37.0
65	57.6	46.6	69.8







#27
2,4-Dimethylphenol
Concen: 42.492 ng
RT: 7.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 134777

Ion	Ratio	Lower	Upper
122	100		
107	112.4	91.8	137.8
121	57.3	45.2	67.8







#28
bis(2-Chloroethoxy)methane
Concen: 41.132 ng
RT: 7.910 min Scan# 972
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion: 93 Resp: 190207

Ion	Ratio	Lower	Upper
93	100		
95	31.8	24.6	36.8
123	11.8	9.1	13.7







#29
2,4-Dichlorophenol
Concen: 41.318 ng
RT: 8.016 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 152408

Ion	Ratio	Lower	Upper
162	100		
164	62.1	42.1	82.1
98	34.7	16.0	56.0







#30
1,2,4-Trichlorobenzene
Concen: 40.926 ng
RT: 8.104 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:180 Resp: 155305

Ion	Ratio	Lower	Upper
180	100		
182	110.8	87.1	130.7
145	30.9	25.3	37.9







#31
Naphthalene
Concen: 40.566 ng
RT: 8.187 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 444899

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.4	12.6
127	13.2	9.9	14.9







#32
Benzoic acid
Concen: 43.387 ng
RT: 7.922 min Scan# 974
Delta R.T. 0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:122 Resp: 87292

Ion	Ratio	Lower	Upper
122	100		
105	117.6	111.3	151.3
77	98.7	96.7	136.7







#33
4-Chloroaniline
Concen: 42.453 ng
RT: 8.234 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 162085

Ion	Ratio	Lower	Upper
127	100		
129	31.0	25.6	38.4
65	35.2	30.9	46.3
92	22.2	18.9	28.3







#34
Hexachlorobutadiene
Concen: 38.895 ng
RT: 8.298 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:225 Resp: 120689

Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.6	76.0
227	63.2	51.8	77.8







#35
Caprolactam
Concen: 42.378 ng
RT: 8.598 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 39199

Ion	Ratio	Lower	Upper
113	100		
55	188.9	160.1	200.1
56	154.9	124.9	164.9







#36
4-Chloro-3-methylphenol
Concen: 42.237 ng
RT: 8.710 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:107 Resp: 137486

Ion	Ratio	Lower	Upper
107	100		
144	27.0	21.6	32.4
142	77.5	65.4	98.0







#37
2-Methylnaphthalene
Concen: 40.463 ng
RT: 8.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 292579

Ion	Ratio	Lower	Upper
142	100		
141	89.1	67.6	101.4







#38
1-Methylnaphthalene
Concen: 40.665 ng
RT: 8.975 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:142 Resp: 278578

Ion	Ratio	Lower	Upper
142	100		
141	92.5	73.4	110.2







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.916 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 124035

Ion	Ratio	Lower	Upper
164	100		
162	99.3	79.4	119.2
160	46.2	36.7	55.1







#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.433 ng
RT: 9.039 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:216 Resp: 172358

Ion	Ratio	Lower	Upper
216	100		
214	79.3	63.2	94.8
179	18.8	15.2	22.8
108	13.8	11.9	17.9







#41
Hexachlorocyclopentadiene
Concen: 34.011 ng
RT: 9.028 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 44780

Ion	Ratio	Lower	Upper
237	100		
235	65.1	44.3	84.3
272	14.3	0.0	35.2







#42
2,4,6-Tribromophenol
Concen: 79.017 ng
RT: 10.710 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:330 Resp: 124844

Ion	Ratio	Lower	Upper
330	100		
332	101.3	77.0	115.6
141	23.0	19.4	29.0







#43
2,4,6-Trichlorophenol
Concen: 41.760 ng
RT: 9.151 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 115708

Ion	Ratio	Lower	Upper
196	100		
198	97.5	74.4	111.6
200	29.6	23.8	35.8







#44
2,4,5-Trichlorophenol
Concen: 41.791 ng
RT: 9.192 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:196 Resp: 121526

Ion	Ratio	Lower	Upper
196	100		
198	93.4	78.0	117.0
97	43.6	37.7	56.5
132	24.9	20.2	30.4







#45
2-Fluorobiphenyl
Concen: 77.029 ng
RT: 9.234 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 678951

Ion	Ratio	Lower	Upper
172	100		
171	34.9	27.5	41.3
170	23.6	19.0	28.6







#46
1,1'-Biphenyl
Concen: 39.957 ng
RT: 9.339 min Scan# 1215
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:154 Resp: 354773

Ion	Ratio	Lower	Upper
154	100		
153	39.3	19.7	59.7
76	11.0	0.0	31.1







#47
2-Chloronaphthalene
Concen: 41.405 ng
RT: 9.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 307997

Ion	Ratio	Lower	Upper
162	100		
127	36.8	31.6	47.4
164	30.8	26.5	39.7







#48
2-Nitroaniline
Concen: 40.037 ng
RT: 9.457 min Scan# 1235
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion: 65 Resp: 88438

Ion	Ratio	Lower	Upper
65	100		
92	69.8	52.3	78.5
138	102.0	78.3	117.5







#49
Acenaphthylene
Concen: 41.536 ng
RT: 9.781 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 416607

Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.9	23.9
153	12.8	10.4	15.6







#50
Dimethylphthalate
Concen: 41.342 ng
RT: 9.634 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:163 Resp: 336124

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.0
164	10.1	7.8	11.8







#51
2,6-Dinitrotoluene
Concen: 42.046 ng
RT: 9.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 67580

Ion	Ratio	Lower	Upper
165	100		
63	58.3	50.4	75.6
89	49.9	41.4	62.2







#52
Acenaphthene
Concen: 39.855 ng
RT: 9.951 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:154 Resp: 266968

Ion	Ratio	Lower	Upper
154	100		
153	114.9	89.9	134.9
152	53.5	42.0	63.0







#53
3-Nitroaniline
Concen: 43.238 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 77266

Ion	Ratio	Lower	Upper
138	100		
108	10.6	7.9	11.9
92	113.9	92.0	138.0







#54
2,4-Dinitrophenol
Concen: 38.660 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:184 Resp: 33608

Ion	Ratio	Lower	Upper
184	100		
63	54.4	52.1	78.1
154	55.5	47.9	71.9







#55
Dibenzofuran
Concen: 40.786 ng
RT: 10.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 379280

Ion	Ratio	Lower	Upper
168	100		
139	39.7	33.2	49.8
169	15.4	12.6	19.0







#56
4-Nitrophenol
Concen: 39.550 ng
RT: 10.034 min Scan# 1333
Delta R.T. 0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:139 Resp: 47859

Ion	Ratio	Lower	Upper
139	100		
109	66.8	53.8	93.8
65	101.7	95.9	135.9







#57
2,4-Dinitrotoluene
Concen: 42.178 ng
RT: 10.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 91495

Ion	Ratio	Lower	Upper
165	100		
63	45.2	35.4	53.2
89	60.1	52.7	79.1
182	2.5	2.3	3.5







#58
Fluorene
Concen: 40.609 ng
RT: 10.469 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:166 Resp: 287737

Ion	Ratio	Lower	Upper
166	100		
165	97.5	77.0	115.4
167	13.4	10.2	15.4







#59
2,3,4,6-Tetrachlorophenol
Concen: 42.612 ng
RT: 10.239 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 87320

Ion	Ratio	Lower	Upper
232	100		
131	38.5	31.9	47.9
130	1.5	1.1	1.7
166	26.9	21.0	31.4







#60
Diethylphthalate
Concen: 39.944 ng
RT: 10.333 min Scan# 1384
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:149 Resp: 305410

Ion	Ratio	Lower	Upper
149	100		
177	23.0	17.8	26.6
150	12.9	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 41.094 ng
RT: 10.457 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 163745

Ion	Ratio	Lower	Upper
204	100		
206	32.2	27.8	41.6
141	53.7	45.4	68.2







#62
4-Nitroaniline
Concen: 40.525 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:138 Resp: 68617

Ion	Ratio	Lower	Upper
138	100		
92	52.2	30.5	70.5
108	68.6	56.2	96.2







#63
Azobenzene
Concen: 39.963 ng
RT: 10.616 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 289198

Ion	Ratio	Lower	Upper
77	100		
182	29.0	9.0	49.0
105	21.3	1.4	41.4
51	33.6	16.9	56.9







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1567
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:188 Resp: 207074

Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.4	8.0
80	6.9	6.0	9.0







#65
4,6-Dinitro-2-methylphenol
Concen: 39.830 ng
RT: 10.516 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 52916

Ion	Ratio	Lower	Upper
198	100		
51	45.5	25.8	65.8
105	34.6	17.2	57.2







#66
n-Nitrosodiphenylamine
Concen: 41.875 ng
RT: 10.575 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:169 Resp: 277049

Ion	Ratio	Lower	Upper
169	100		
168	58.2	47.4	71.2
167	31.2	25.5	38.3







#67
4-Bromophenyl-phenylether
Concen: 42.302 ng
RT: 10.951 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 109675

Ion	Ratio	Lower	Upper
248	100		
250	96.9	77.8	116.6
141	56.3	50.5	75.7







#68
Hexachlorobenzene
Concen: 40.395 ng
RT: 11.016 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:284 Resp: 128445

Ion	Ratio	Lower	Upper
284	100		
142	24.2	20.0	30.0
249	28.5	22.2	33.2







#69
Atrazine
Concen: 39.616 ng
RT: 11.104 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 86008

Ion	Ratio	Lower	Upper
200	100		
173	25.4	4.9	44.9
215	49.8	28.2	68.2







#70
Pentachlorophenol
Concen: 44.431 ng
RT: 11.210 min Scan# 1533
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:266 Resp: 67248

Ion	Ratio	Lower	Upper
266	100		
268	62.7	48.2	72.4
264	63.5	49.4	74.0







#71
Phenanthrene
Concen: 41.526 ng
RT: 11.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 427157

Ion	Ratio	Lower	Upper
178	100		
176	19.7	16.3	24.5
179	14.9	12.6	18.8







#72
Anthracene
Concen: 40.944 ng
RT: 11.486 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:178 Resp: 435076

Ion	Ratio	Lower	Upper
178	100		
176	18.5	15.0	22.4
179	14.9	11.8	17.6







#73
Carbazole
Concen: 40.663 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 367780

Ion	Ratio	Lower	Upper
167	100		
166	20.7	16.1	24.1
139	12.8	9.8	14.6







#74
Di-n-butylphthalate
Concen: 40.941 ng
RT: 11.963 min Scan# 1661
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:149 Resp: 453989

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.3	10.9
104	4.2	3.6	5.4







#75
Fluoranthene
Concen: 39.974 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 454294

Ion	Ratio	Lower	Upper
202	100		
101	8.6	0.0	27.7
203	17.3	0.0	37.3







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:240 Resp: 153131

Ion	Ratio	Lower	Upper
240	100		
120	7.9	6.5	9.7
236	25.9	20.1	30.1







#77
Benzidine
Concen: 42.231 ng
RT: 12.745 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 201262

Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.6	17.4
183	11.6	8.5	12.7







#78
Pyrene
Concen: 41.969 ng
RT: 12.851 min Scan# 1812
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:202 Resp: 462034

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	16.7	13.5	20.3







#79
Terphenyl-d14
Concen: 81.522 ng
RT: 12.992 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 714068

Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.8	8.6
122	6.9	4.9	7.3







#80
Butylbenzylphthalate
Concen: 41.351 ng
RT: 13.469 min Scan# 1917
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:149 Resp: 172222

Ion	Ratio	Lower	Upper
149	100		
91	69.5	56.6	84.8
206	23.7	18.3	27.5







#81
Benzo(a)anthracene
Concen: 42.586 ng
RT: 14.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 440710

Ion	Ratio	Lower	Upper
228	100		
226	26.2	21.4	32.2
229	19.6	15.3	22.9







#82
3,3'-Dichlorobenzidine
Concen: 44.172 ng
RT: 14.004 min Scan# 2008
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:252 Resp: 159124

Ion	Ratio	Lower	Upper
252	100		
254	62.6	49.9	74.9
126	7.6	6.5	9.7







#83
Chrysene
Concen: 42.239 ng
RT: 14.080 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 410191

Ion	Ratio	Lower	Upper
228	100		
226	28.4	23.8	35.6
229	19.6	15.6	23.4







#84
Bis(2-ethylhexyl)phthalate
Concen: 42.623 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:149 Resp: 242344

Ion	Ratio	Lower	Upper
149	100		
167	27.1	22.5	33.7
279	4.8	3.6	5.4







#85
Di-n-octyl phthalate
Concen: 45.619 ng
RT: 14.639 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 446067

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	11.3	9.3	13.9







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.539 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:264 Resp: 204043

Ion	Ratio	Lower	Upper
264	100		
260	22.5	17.0	25.4
265	21.9	16.6	24.8







#87
Indeno(1,2,3-cd)pyrene
Concen: 43.307 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 606674

Ion	Ratio	Lower	Upper
276	100		
138	16.4	13.0	19.6
277	24.4	19.4	29.0







#88
Benzo(b)fluoranthene
Concen: 40.544 ng
RT: 15.104 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:252 Resp: 467688

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.2	25.8
125	6.0	5.0	7.6







#89
Benzo(k)fluoranthene
Concen: 43.376 ng
RT: 15.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 461643

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.4	26.2
125	6.3	4.9	7.3







#90
Benzo(a)pyrene
Concen: 42.801 ng
RT: 15.474 min Scan# 2258
Delta R.T. -0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:252 Resp: 442928

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.1	25.7
125	7.2	5.4	8.2







#91
Dibenzo(a,h)anthracene
Concen: 43.651 ng
RT: 17.062 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 484044

Ion	Ratio	Lower	Upper
278	100		
139	10.6	8.5	12.7
279	23.6	18.6	28.0






#92
Benzo(g,h,i)perylene
Concen: 43.293 ng
RT: 17.504 min Scan# 2603
Delta R.T. 0.006 min
Lab File: BF144121.D
Acq: 29 Oct 2025 11:30

Tgt Ion:276 Resp: 482661

Ion	Ratio	Lower	Upper
276	100		
277	23.4	17.9	26.9
138	15.2	12.0	18.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2		1,4-Dioxane	0.546	0.560	-2.6	127	-0.02
3		Pyridine	1.462	1.578	-7.9	128	-0.02
4		n-Nitrosodimethylamine	0.899	0.823	8.5	112	-0.02
5	S	2-Fluorophenol	1.264	1.346	-6.5	128	0.00
6		Aniline	1.997	2.098	-5.1	129	0.00
7	S	Phenol-d6	1.396	1.460	-4.6	129	0.00
8		2-Chlorophenol	1.245	1.283	-3.1	127	0.00
9		Benzaldehyde	1.084	1.051	3.0	125	0.00
10	C	Phenol	1.722	1.843	-7.0	132	0.00
11		bis(2-Chloroethyl)ether	1.245	1.332	-7.0	133	0.00
12		1,3-Dichlorobenzene	1.536	1.628	-6.0	129	0.00
13	C	1,4-Dichlorobenzene	1.513	1.602	-5.9	128	0.00
14		1,2-Dichlorobenzene	1.464	1.518	-3.7	127	0.00
15		Benzyl Alcohol	1.135	1.172	-3.3	127	0.00
16		2,2'-oxybis(1-Chloropropane	2.339	2.460	-5.2	129	0.00
17		2-Methylphenol	0.957	1.014	-6.0	135	0.00
18		Hexachloroethane	0.541	0.545	-0.7	124	0.00
19	P	n-Nitroso-di-n-propylamine	0.903	0.927	-2.7	131	0.00
20		3+4-Methylphenols	1.132	1.224	-8.1	130	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
22		Acetophenone	0.432	0.429	0.7	128	0.00
23	S	Nitrobenzene-d5	0.365	0.361	1.1	128	0.00
24		Nitrobenzene	0.386	0.394	-2.1	131	0.00
25		Isophorone	0.638	0.649	-1.7	132	0.00
26	C	2-Nitrophenol	0.184	0.193	-4.9	132	0.00
27		2,4-Dimethylphenol	0.273	0.290	-6.2	132	0.00
28		bis(2-Chloroethoxy)methane	0.398	0.409	-2.8	132	0.00
29	C	2,4-Dichlorophenol	0.317	0.328	-3.5	132	0.00
30		1,2,4-Trichlorobenzene	0.327	0.334	-2.1	130	0.00
31		Naphthalene	0.944	0.957	-1.4	129	0.00
32		Benzoic acid	0.173	0.188	-8.7	168#	0.00
33		4-Chloroaniline	0.329	0.349	-6.1	135	0.00
34	C	Hexachlorobutadiene	0.267	0.260	2.6	126	0.00
35		Caprolactam	0.080	0.084	-5.0	130	0.00
36	C	4-Chloro-3-methylphenol	0.280	0.296	-5.7	137	0.00
37		2-Methylnaphthalene	0.622	0.630	-1.3	128	0.00
38		1-Methylnaphthalene	0.590	0.599	-1.5	132	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	135	0.00
40		1,2,4,5-Tetrachlorobenzene	0.687	0.695	-1.2	133	0.00
41	P	Hexachlorocyclopentadiene	0.212	0.181	14.6	112	0.00
42	S	2,4,6-Tribromophenol	0.255	0.252	1.2	130	0.00
43	C	2,4,6-Trichlorophenol	0.447	0.466	-4.3	134	0.00
44		2,4,5-Trichlorophenol	0.469	0.490	-4.5	142	0.00
45	S	2-Fluorobiphenyl	1.421	1.368	3.7	128	0.00
46		1,1'-Biphenyl	1.432	1.430	0.1	132	0.00
47		2-Chloronaphthalene	1.199	1.242	-3.6	138	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.356	0.357	-0.3	127	0.00
49		Acenaphthylene	1.617	1.679	-3.8	134	0.00
50		Dimethylphthalate	1.311	1.355	-3.4	134	0.00
51		2,6-Dinitrotoluene	0.259	0.272	-5.0	134	0.00
52	C	Acenaphthene	1.080	1.076	0.4	131	0.00
53		3-Nitroaniline	0.288	0.311	-8.0	131	0.00
54	P	2,4-Dinitrophenol	0.140	0.135	3.6	148	0.00
55		Dibenzofuran	1.499	1.529	-2.0	133	0.00
56	P	4-Nitrophenol	0.195	0.193	1.0	140	0.00
57		2,4-Dinitrotoluene	0.350	0.369	-5.4	130	0.00
58		Fluorene	1.143	1.160	-1.5	132	0.00
59		2,3,4,6-Tetrachlorophenol	0.330	0.352	-6.7	132	0.00
60		Diethylphthalate	1.233	1.231	0.2	129	0.00
61		4-Chlorophenyl-phenylether	0.643	0.660	-2.6	132	0.00
62		4-Nitroaniline	0.273	0.277	-1.5	127	0.00
63		Azobenzene	1.167	1.166	0.1	128	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
65		4,6-Dinitro-2-methylphenol	0.128	0.128	0.0	135	0.00
66	c	n-Nitrosodiphenylamine	0.639	0.669	-4.7	135	0.00
67		4-Bromophenyl-phenylether	0.250	0.265	-6.0	132	0.00
68		Hexachlorobenzene	0.307	0.310	-1.0	125	0.00
69		Atrazine	0.210	0.208	1.0	129	0.00
70	C	Pentachlorophenol	0.146	0.162	-11.0	141	0.00
71		Phenanthrene	0.994	1.031	-3.7	132	0.00
72		Anthracene	1.026	1.051	-2.4	129	0.00
73		Carbazole	0.874	0.888	-1.6	127	0.00
74		Di-n-butylphthalate	1.071	1.096	-2.3	127	0.00
75	C	Fluoranthene	1.098	1.097	0.1	127	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	124	0.00
77		Benzidine	0.622	0.657	-5.6	122	0.00
78		Pyrene	1.438	1.509	-4.9	123	0.00
79	S	Terphenyl-d14	1.144	1.166	-1.9	119	0.00
80		Butylbenzylphthalate	0.544	0.562	-3.3	116	0.00
81		Benzo(a)anthracene	1.352	1.439	-6.4	118	0.00
82		3,3'-Dichlorobenzidine	0.470	0.520	-10.6	126	0.00
83		Chrysene	1.268	1.339	-5.6	126	0.00
84		Bis(2-ethylhexyl)phthalate	0.743	0.791	-6.5	123	0.00
85	c	Di-n-octyl phthalate	1.277	1.456	-14.0	129	0.00
86	I	Perylene-d12	1.000	1.000	0.0	133	0.00
87		Indeno(1,2,3-cd)pyrene	1.373	1.487	-8.3	134	0.00
88		Benzo(b)fluoranthene	1.131	1.146	-1.3	121	0.00
89		Benzo(k)fluoranthene	1.043	1.131	-8.4	143	0.00
90	C	Benzo(a)pyrene	1.014	1.085	-7.0	133	0.00
91		Dibenzo(a,h)anthracene	1.087	1.186	-9.1	136	0.00
92		Benzo(g,h,i)perylene	1.093	1.183	-8.2	132	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	124	0.00
2		1,4-Dioxane	40.000	41.016	-2.5	127	-0.02
3		Pyridine	40.000	43.152	-7.9	128	-0.02
4		n-Nitrosodimethylamine	40.000	36.615	8.5	112	-0.02
5	S	2-Fluorophenol	80.000	85.206	-6.5	128	0.00
6		Aniline	40.000	42.009	-5.0	129	0.00
7	S	Phenol-d6	80.000	83.686	-4.6	129	0.00
8		2-Chlorophenol	40.000	41.239	-3.1	127	0.00
9		Benzaldehyde	40.000	38.777	3.1	125	0.00
10	C	Phenol	40.000	42.796	-7.0	132	0.00
11		bis(2-Chloroethyl)ether	40.000	42.797	-7.0	133	0.00
12		1,3-Dichlorobenzene	40.000	42.415	-6.0	129	0.00
13	C	1,4-Dichlorobenzene	40.000	42.348	-5.9	128	0.00
14		1,2-Dichlorobenzene	40.000	41.472	-3.7	127	0.00
15		Benzyl Alcohol	40.000	41.303	-3.3	127	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	42.062	-5.2	129	0.00
17		2-Methylphenol	40.000	42.361	-5.9	135	0.00
18		Hexachloroethane	40.000	40.302	-0.8	124	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	41.075	-2.7	131	0.00
20		3+4-Methylphenols	40.000	43.254	-8.1	130	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	130	0.00
22		Acetophenone	40.000	39.693	0.8	128	0.00
23	S	Nitrobenzene-d5	80.000	79.215	1.0	128	0.00
24		Nitrobenzene	40.000	40.760	-1.9	131	0.00
25		Isophorone	40.000	40.647	-1.6	132	0.00
26	C	2-Nitrophenol	40.000	42.017	-5.0	132	0.00
27		2,4-Dimethylphenol	40.000	42.492	-6.2	132	0.00
28		bis(2-Chloroethoxy)methane	40.000	41.132	-2.8	132	0.00
29	C	2,4-Dichlorophenol	40.000	41.318	-3.3	132	0.00
30		1,2,4-Trichlorobenzene	40.000	40.926	-2.3	130	0.00
31		Naphthalene	40.000	40.566	-1.4	129	0.00
32		Benzoic acid	40.000	43.387	-8.5	168	0.00
33		4-Chloroaniline	40.000	42.453	-6.1	135	0.00
34	C	Hexachlorobutadiene	40.000	38.895	2.8	126	0.00
35		Caprolactam	40.000	42.378	-5.9	130	0.00
36	C	4-Chloro-3-methylphenol	40.000	42.237	-5.6	137	0.00
37		2-Methylnaphthalene	40.000	40.463	-1.2	128	0.00
38		1-Methylnaphthalene	40.000	40.665	-1.7	132	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	135	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	40.433	-1.1	133	0.00
41	P	Hexachlorocyclopentadiene	40.000	34.011	15.0	112	0.00
42	S	2,4,6-Tribromophenol	80.000	79.017	1.2	130	0.00
43	C	2,4,6-Trichlorophenol	40.000	41.760	-4.4	134	0.00
44		2,4,5-Trichlorophenol	40.000	41.791	-4.5	142	0.00
45	S	2-Fluorobiphenyl	80.000	77.029	3.7	128	0.00
46		1,1'-Biphenyl	40.000	39.957	0.1	132	0.00
47		2-Chloronaphthalene	40.000	41.405	-3.5	138	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	40.037	-0.1	127	0.00
49		Acenaphthylene	40.000	41.536	-3.8	134	0.00
50		Dimethylphthalate	40.000	41.342	-3.4	134	0.00
51		2,6-Dinitrotoluene	40.000	42.046	-5.1	134	0.00
52	C	Acenaphthene	40.000	39.855	0.4	131	0.00
53		3-Nitroaniline	40.000	43.238	-8.1	131	0.00
54	P	2,4-Dinitrophenol	40.000	38.660	3.4	148	0.00
55		Dibenzofuran	40.000	40.786	-2.0	133	0.00
56	P	4-Nitrophenol	40.000	39.550	1.1	140	0.00
57		2,4-Dinitrotoluene	40.000	42.178	-5.4	130	0.00
58		Fluorene	40.000	40.609	-1.5	132	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	42.612	-6.5	132	0.00
60		Diethylphthalate	40.000	39.944	0.1	129	0.00
61		4-Chlorophenyl-phenylether	40.000	41.094	-2.7	132	0.00
62		4-Nitroaniline	40.000	40.525	-1.3	127	0.00
63		Azobenzene	40.000	39.963	0.1	128	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	130	0.00
65		4,6-Dinitro-2-methylphenol	40.000	39.830	0.4	135	0.00
66	c	n-Nitrosodiphenylamine	40.000	41.875	-4.7	135	0.00
67		4-Bromophenyl-phenylether	40.000	42.302	-5.8	132	0.00
68		Hexachlorobenzene	40.000	40.395	-1.0	125	0.00
69		Atrazine	40.000	39.616	1.0	129	0.00
70	C	Pentachlorophenol	40.000	44.431	-11.1	141	0.00
71		Phenanthrene	40.000	41.526	-3.8	132	0.00
72		Anthracene	40.000	40.944	-2.4	129	0.00
73		Carbazole	40.000	40.663	-1.7	127	0.00
74		Di-n-butylphthalate	40.000	40.941	-2.4	127	0.00
75	C	Fluoranthene	40.000	39.974	0.1	127	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	124	0.00
77		Benzidine	40.000	42.231	-5.6	122	0.00
78		Pyrene	40.000	41.969	-4.9	123	0.00
79	S	Terphenyl-d14	80.000	81.522	-1.9	119	0.00
80		Butylbenzylphthalate	40.000	41.351	-3.4	116	0.00
81		Benzo(a)anthracene	40.000	42.586	-6.5	118	0.00
82		3,3'-Dichlorobenzidine	40.000	44.172	-10.4	126	0.00
83		Chrysene	40.000	42.239	-5.6	126	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	42.623	-6.6	123	0.00
85	c	Di-n-octyl phthalate	40.000	45.619	-14.0	129	0.00
86	I	Perylene-d12	20.000	20.000	0.0	133	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	43.307	-8.3	134	0.00
88		Benzo(b)fluoranthene	40.000	40.544	-1.4	121	0.00
89		Benzo(k)fluoranthene	40.000	43.376	-8.4	143	0.00
90	C	Benzo(a)pyrene	40.000	42.801	-7.0	133	0.00
91		Dibenzo(a,h)anthracene	40.000	43.651	-9.1	136	0.00
92		Benzo(g,h,i)perylene	40.000	43.293	-8.2	132	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144121.D
Acq On : 29 Oct 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 12:22:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG No.: Q3434

Instrument ID: BNA_G Calibration Date/Time: 10/29/2025 11:22

Lab File ID: BG064618.D Init. Calib. Date(s): 10/24/2025 10/24/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:00 15:32

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.192	1.266		6.2	
Benzaldehyde	0.893	0.838		-6.2	
Phenol-d6	1.635	1.656		1.3	
Phenol	1.808	1.838		1.7	20.0
bis(2-Chloroethyl)ether	1.329	1.382		4.0	
2-Chlorophenol	1.247	1.377		10.4	
2-Methylphenol	1.201	1.261		5.0	
2,2-oxybis(1-Chloropropane)	3.487	3.674		5.4	
Acetophenone	0.547	0.581		6.2	
3+4-Methylphenols	1.688	1.742		3.2	
n-Nitroso-di-n-propylamine	1.130	1.223	0.050	8.2	
Nitrobenzene-d5	0.333	0.401		20.4	
Hexachloroethane	0.512	0.567		10.7	
Nitrobenzene	0.362	0.429		18.5	
Isophorone	0.799	0.833		4.3	
2-Nitrophenol	0.121	0.168		38.8	20.0
2,4-Dimethylphenol	0.297	0.311		4.7	
bis(2-Chloroethoxy)methane	0.455	0.480		5.5	
2,4-Dichlorophenol	0.320	0.352		10.0	20.0
Naphthalene	1.109	1.159		4.5	
4-Chloroaniline	0.431	0.435		0.9	
Hexachlorobutadiene	0.284	0.322		13.4	20.0
Caprolactam	0.123	0.108		-12.2	
4-Chloro-3-methylphenol	0.384	0.389		1.3	20.0
2-Methylnaphthalene	0.814	0.836		2.7	
Hexachlorocyclopentadiene	0.302	0.400	0.050	32.5	
2,4,6-Trichlorophenol	0.384	0.455		18.5	20.0
2-Fluorobiphenyl	1.319	1.473		11.7	
2,4,5-Trichlorophenol	0.443	0.492		11.1	
1,1-Biphenyl	1.422	1.535		7.9	
2-Chloronaphthalene	1.073	1.175		9.5	
2-Nitroaniline	0.311	0.378		21.5	
Dimethylphthalate	1.530	1.552		1.4	
Acenaphthylene	1.693	1.770		4.5	
2,6-Dinitrotoluene	0.218	0.272		24.8	
3-Nitroaniline	0.263	0.288		9.5	
Acenaphthene	1.157	1.225		5.9	20.0
2,4-Dinitrophenol	0.132	0.164	0.050	24.2	
4-Nitrophenol	0.244	0.241	0.050	-1.2	

Form VII SV-1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG No.: Q3434

Instrument ID: BNA_G Calibration Date/Time: 10/29/2025 11:22

Lab File ID: BG064618.D Init. Calib. Date(s): 10/24/2025 10/24/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:00 15:32

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.870	1.930		3.2	
2,4-Dinitrotoluene	0.340	0.396		16.5	
Diethylphthalate	1.563	1.507		-3.6	
4-Chlorophenyl-phenylether	0.820	0.840		2.4	
Fluorene	1.460	1.476		1.1	
4-Nitroaniline	0.293	0.306		4.4	
4,6-Dinitro-2-methylphenol	0.079	0.104		31.6	
n-Nitrosodiphenylamine	0.540	0.582		7.8	20.0
2,4,6-Tribromophenol	0.289	0.312		8.0	
4-Bromophenyl-phenylether	0.237	0.262		10.5	
Hexachlorobenzene	0.270	0.300		11.1	
Atrazine	0.232	0.233		0.4	
Pentachlorophenol	0.169	0.189		11.8	20.0
Phenanthrene	1.094	1.132		3.5	
Anthracene	1.113	1.125		1.1	
Carbazole	0.981	0.970		-1.1	
Di-n-butylphthalate	1.135	1.140		0.4	
Fluoranthene	1.370	1.343		-2.0	20.0
Pyrene	1.307	1.322		1.1	
Terphenyl-d14	1.060	1.074		1.3	
Butylbenzylphthalate	0.395	0.462		17.0	
3,3-Dichlorobenzidine	0.443	0.484		9.3	
Benzo(a)anthracene	1.326	1.395		5.2	
Chrysene	1.262	1.298		2.9	
Bis(2-ethylhexyl)phthalate	0.590	0.671		13.7	
Di-n-octyl phthalate	0.995	1.203		20.9	20.0
Benzo(b)fluoranthene	1.226	1.236		0.8	
Benzo(k)fluoranthene	1.227	1.235		0.7	
Benzo(a)pyrene	1.127	1.157		2.7	20.0
Indeno(1,2,3-cd)pyrene	1.471	1.519		3.3	
Dibenzo(a,h)anthracene	1.195	1.241		3.8	
Benzo(g,h,i)perylene	1.142	1.182		3.5	
1,2,4,5-Tetrachlorobenzene	0.671	0.754		12.4	
1,4-Dioxane	0.528	0.560		6.1	20.0
2,3,4,6-Tetrachlorophenol	0.424	0.469		10.6	

All other compounds must meet a minimum RRF of 0.010.

Form VII SV-1

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.223	152	36613	20.000	ng		0.00
21) Naphthalene-d8	11.043	136	148961	20.000	ng		0.00
39) Acenaphthene-d10	14.839	164	108882	20.000	ng		0.00
64) Phenanthrene-d10	17.577	188	241683	20.000	ng		0.00
76) Chrysene-d12	21.854	240	261016	20.000	ng		0.00
86) Perylene-d12	25.197	264	298207	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.743	112	185390	84.991	ng		0.00
7) Phenol-d6	7.353	99	242475	81.009	ng		0.00
23) Nitrobenzene-d5	9.380	82	238914	96.333	ng		0.00
42) 2,4,6-Tribromophenol	16.319	330	135782	86.403	ng		0.00
45) 2-Fluorobiphenyl	13.470	172	641606	89.318	ng		0.00
79) Terphenyl-d14	20.168	244	1121232	81.066	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.587	88	41038	42.495	ng		97
3) Pyridine	3.998	79	118806	40.966	ng		97
4) n-Nitrosodimethylamine	3.904	42	63036	40.532	ng		100
6) Aniline	7.529	93	166421	40.860	ng		98
8) 2-Chlorophenol	7.776	128	100847	44.170	ng		99
9) Benzaldehyde	7.341	77	61374	37.558	ng		96
10) Phenol	7.377	94	134608	40.664	ng		96
11) bis(2-Chloroethyl)ether	7.623	93	101216	41.589	ng		96
12) 1,3-Dichlorobenzene	8.111	146	113188	43.073	ng		98
13) 1,4-Dichlorobenzene	8.258	146	113697	42.551	ng		98
14) 1,2-Dichlorobenzene	8.575	146	111091	43.105	ng		98
15) Benzyl Alcohol	8.446	79	104162	42.201	ng		96
16) 2,2'-oxybis(1-Chloropr...	8.740	45	269007	42.138	ng		98
17) 2-Methylphenol	8.646	107	92368	42.028	ng		98
18) Hexachloroethane	9.321	117	41522	44.282	ng		95
19) n-Nitroso-di-n-propyla...	9.022	70	89519	43.275	ng		97
20) 3+4-Methylphenols	8.975	107	127586	41.279	ng		99
22) Acetophenone	9.039	105	173153	42.523	ng	#	97
24) Nitrobenzene	9.427	77	127739	47.355	ng		97
25) Isophorone	9.956	82	248133	41.695	ng		99
26) 2-Nitrophenol	10.144	139	49935	47.979	ng		96
27) 2,4-Dimethylphenol	10.191	122	92582	41.823	ng		93
28) bis(2-Chloroethoxy)met...	10.432	93	142887	42.170	ng		99
29) 2,4-Dichlorophenol	10.679	162	104960	44.058	ng		95
30) 1,2,4-Trichlorobenzene	10.902	180	118900	43.927	ng		96
31) Naphthalene	11.096	128	345379	41.819	ng		99
32) Benzoic acid	10.309	122	60590	37.459	ng		96
33) 4-Chloroaniline	11.184	127	129600	40.373	ng		98
34) Hexachlorobutadiene	11.384	225	95906	45.306	ng		98
35) Caprolactam	11.948	113	32311	35.362	ng		91
36) 4-Chloro-3-methylphenol	12.289	107	115865	40.462	ng		98
37) 2-Methylnaphthalene	12.688	142	249124	41.085	ng		98
38) 1-Methylnaphthalene	12.900	142	245754	40.940	ng		95
40) 1,2,4,5-Tetrachloroben...	13.047	216	164299	44.962	ng		100
41) Hexachlorocyclopentadiene	13.035	237	87148	52.985	ng		99
43) 2,4,6-Trichlorophenol	13.276	196	99057	47.391	ng		96

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.346	196	107132	44.415	ng		96
46) 1,1'-Biphenyl	13.681	154	334243	43.180	ng		98
47) 2-Chloronaphthalene	13.728	162	255946	43.804	ng		98
48) 2-Nitroaniline	13.910	65	82275	42.351	ng		96
49) Acenaphthylene	14.562	152	385513	41.830	ng		99
50) Dimethylphthalate	14.280	163	338032	40.572	ng		99
51) 2,6-Dinitrotoluene	14.398	165	59140	43.351	ng		97
52) Acenaphthene	14.903	154	266835	42.377	ng		96
53) 3-Nitroaniline	14.721	138	62676	43.773	ng		97
54) 2,4-Dinitrophenol	14.927	184	35680	44.922	ng	#	58
55) Dibenzofuran	15.238	168	420237	41.281	ng		99
56) 4-Nitrophenol	15.015	139	52490	39.540	ng		94
57) 2,4-Dinitrotoluene	15.179	165	86302	41.073	ng	#	99
58) Fluorene	15.884	166	321438	40.432	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.455	232	102176	44.298	ng		99
60) Diethylphthalate	15.644	149	328215	38.582	ng		99
61) 4-Chlorophenyl-phenyle...	15.873	204	182832	40.979	ng		99
62) 4-Nitroaniline	15.884	138	66563	41.786	ng		99
63) Azobenzene	16.161	77	305844	39.183	ng		99
65) 4,6-Dinitro-2-methylph...	15.943	198	50385	47.448	ng		91
66) n-Nitrosodiphenylamine	16.078	169	281278	43.085	ng		98
67) 4-Bromophenyl-phenylether	16.766	248	126485	44.163	ng		98
68) Hexachlorobenzene	16.889	284	145122	44.415	ng		96
69) Atrazine	17.018	200	112465	40.031	ng		96
70) Pentachlorophenol	17.224	266	91563	44.939	ng		94
71) Phenanthrene	17.618	178	547128	41.397	ng		99
72) Anthracene	17.712	178	543644	40.436	ng		99
73) Carbazole	17.970	167	468954	39.556	ng		99
74) Di-n-butylphthalate	18.528	149	550901	40.184	ng		100
75) Fluoranthene	19.615	202	648927	39.193	ng		98
77) Benzidine	19.780	184	265517	45.404	ng		99
78) Pyrene	19.974	202	690102	40.467	ng		99
80) Butylbenzylphthalate	20.849	149	241218	46.801	ng		96
81) Benzo(a)anthracene	21.836	228	728489	42.093	ng		100
82) 3,3'-Dichlorobenzidine	21.736	252	252767	43.737	ng		98
83) Chrysene	21.901	228	677717	41.154	ng		99
84) Bis(2-ethylhexyl)phtha...	21.742	149	350537	45.515	ng		96
85) Di-n-octyl phthalate	23.000	149	627808	48.350	ng		99
87) Indeno(1,2,3-cd)pyrene	29.040	276	906170	41.308	ng		98
88) Benzo(b)fluoranthene	24.134	252	736965	40.302	ng		99
89) Benzo(k)fluoranthene	24.210	252	736695	40.260	ng		99
90) Benzo(a)pyrene	25.038	252	690279	41.067	ng		100
91) Dibenzo(a,h)anthracene	29.104	278	740147	41.529	ng		99
92) Benzo(g,h,i)perylene	30.232	276	705048	41.412	ng		98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.223 min Scan# 8
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 36613

Ion	Ratio	Lower	Upper
152	100		
150	155.3	122.2	183.4
115	59.3	50.8	76.2







#2
1,4-Dioxane
Concen: 42.495 ng
RT: 3.587 min Scan# 85
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion: 88 Resp: 41038

Ion	Ratio	Lower	Upper
88	100		
58	110.9	92.4	138.6
43	40.4	33.0	49.6







#3
Pyridine
Concen: 40.966 ng
RT: 3.998 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 118806

Ion	Ratio	Lower	Upper
79	100		
52	121.2	94.2	141.2
51	58.5	46.1	69.1







#4
n-Nitrosodimethylamine
Concen: 40.532 ng
RT: 3.904 min Scan# 139
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion: 42 Resp: 63036

Ion	Ratio	Lower	Upper
42	100		
74	95.4	76.1	114.1
44	8.4	6.3	9.5







#5
2-Fluorophenol
Concen: 84.991 ng
RT: 5.743 min Scan# 4
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 185390

Ion	Ratio	Lower	Upper
112	100		
64	76.3	60.7	91.1
63	41.8	33.4	50.0







#6
Aniline
Concen: 40.860 ng
RT: 7.529 min Scan# 756
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion: 93 Resp: 166421

Ion	Ratio	Lower	Upper
93	100		
66	45.3	37.0	55.6
65	24.3	18.3	27.5







#7
Phenol-d6
Concen: 81.009 ng
RT: 7.353 min Scan# 7
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 242475

Ion	Ratio	Lower	Upper
99	100		
42	24.1	18.3	27.5
71	38.5	28.6	43.0







#8
2-Chlorophenol
Concen: 44.170 ng
RT: 7.776 min Scan# 798
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:128 Resp: 100847

Ion	Ratio	Lower	Upper
128	100		
130	34.8	13.5	53.5
64	65.7	45.2	85.2







#9
Benzaldehyde
Concen: 37.558 ng
RT: 7.341 min Scan# 7
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 61374

Ion	Ratio	Lower	Upper
77	100		
105	98.7	72.9	112.9
106	92.6	70.8	110.8







#10
Phenol
Concen: 40.664 ng
RT: 7.377 min Scan# 730
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion: 94 Resp: 134608

Ion	Ratio	Lower	Upper
94	100		
65	36.7	15.8	55.8
66	47.6	24.0	64.0







#11
bis(2-Chloroethyl)ether
Concen: 41.589 ng
RT: 7.623 min Scan# 7
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 101216

Ion	Ratio	Lower	Upper
93	100		
63	106.9	91.3	131.3
95	32.3	13.3	53.3







#12
1,3-Dichlorobenzene
Concen: 43.073 ng
RT: 8.111 min Scan# 855
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:146 Resp: 113188

Ion	Ratio	Lower	Upper
146	100		
148	67.1	55.3	82.9
75	33.8	28.2	42.2







#13
1,4-Dichlorobenzene
Concen: 42.551 ng
RT: 8.258 min Scan# 8
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 113697

Ion	Ratio	Lower	Upper
146	100		
148	68.5	56.2	84.4
111	43.9	35.4	53.0







#14
1,2-Dichlorobenzene
Concen: 43.105 ng
RT: 8.575 min Scan# 934
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:146 Resp: 111091

Ion	Ratio	Lower	Upper
146	100		
148	65.8	51.5	77.3
111	44.8	35.0	52.6







#15
Benzyl Alcohol
Concen: 42.201 ng
RT: 8.446 min Scan# 91
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 104162

Ion	Ratio	Lower	Upper
79	100		
108	74.5	58.1	87.1
77	62.0	52.8	79.2







#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.138 ng
RT: 8.740 min Scan# 962
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion: 45 Resp: 269007

Ion	Ratio	Lower	Upper
45	100		
77	9.1	0.0	28.5
79	6.3	0.0	25.6







#17
2-Methylphenol
Concen: 42.028 ng
RT: 8.646 min Scan# 94
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 92368

Ion	Ratio	Lower	Upper
107	100		
108	110.4	90.0	135.0
77	49.2	37.8	56.8
79	50.0	40.5	60.7







#18
Hexachloroethane
Concen: 44.282 ng
RT: 9.321 min Scan# 1061
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:117 Resp: 41522

Ion	Ratio	Lower	Upper
117	100		
119	99.2	75.4	113.2
201	130.3	99.2	148.8







#19
n-Nitroso-di-n-propylamine
Concen: 43.275 ng
RT: 9.022 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 89519

Ion	Ratio	Lower	Upper
70	100		
42	78.6	65.2	97.8
101	9.1	7.9	11.9
130	18.8	14.3	21.5







#20
3+4-Methylphenols
Concen: 41.279 ng
RT: 8.975 min Scan# 1002
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:107 Resp: 127586

Ion	Ratio	Lower	Upper
107	100		
108	90.0	70.3	110.3
77	36.9	14.6	54.6
79	30.9	11.7	51.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.043 min Scan# 11
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 148961

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	14.0	12.6	19.0
68	6.5	5.7	8.5







#22
Acetophenone
Concen: 42.523 ng
RT: 9.039 min Scan# 1013
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:105 Resp: 173153

Ion	Ratio	Lower	Upper
105	100		
71	3.5	1.5	2.3#
51	39.3	33.7	50.5
120	19.6	16.1	24.1







#23
Nitrobenzene-d5
Concen: 96.333 ng
RT: 9.380 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 238914

Ion	Ratio	Lower	Upper
82	100		
128	39.9	31.3	46.9
54	71.7	56.6	84.8







#24
Nitrobenzene
Concen: 47.355 ng
RT: 9.427 min Scan# 1079
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion: 77 Resp: 127739

Ion	Ratio	Lower	Upper
77	100		
123	40.7	31.0	46.6
65	15.0	12.7	19.1







#25
Isophorone
Concen: 41.695 ng
RT: 9.956 min Scan# 11
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 248133

Ion	Ratio	Lower	Upper
82	100		
95	7.4	6.1	9.1
138	16.7	13.0	19.4







#26
2-Nitrophenol
Concen: 47.979 ng
RT: 10.144 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:139 Resp: 49935

Ion	Ratio	Lower	Upper
139	100		
109	35.8	24.6	37.0
65	60.9	47.5	71.3







#27
2,4-Dimethylphenol
Concen: 41.823 ng
RT: 10.191 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 92582

Ion	Ratio	Lower	Upper
122	100		
107	119.1	103.0	154.4
121	57.0	47.8	71.8







#28
bis(2-Chloroethoxy)methane
Concen: 42.170 ng
RT: 10.432 min Scan# 1250
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion: 93 Resp: 142887

Ion	Ratio	Lower	Upper
93	100		
95	31.6	25.4	38.2
123	10.2	8.4	12.6







#29
2,4-Dichlorophenol
Concen: 44.058 ng
RT: 10.679 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 104960

Ion	Ratio	Lower	Upper
162	100		
164	63.8	48.7	88.7
98	34.2	16.9	56.9







#30
1,2,4-Trichlorobenzene
Concen: 43.927 ng
RT: 10.902 min Scan# 1330
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:180 Resp: 118900

Ion	Ratio	Lower	Upper
180	100		
182	92.8	76.9	115.3
145	30.5	22.1	33.1







#31
Naphthalene
Concen: 41.819 ng
RT: 11.096 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 345379

Ion	Ratio	Lower	Upper
128	100		
129	10.3	8.8	13.2
127	13.4	10.6	15.8







#32
Benzoic acid
Concen: 37.459 ng
RT: 10.309 min Scan# 1229
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:122 Resp: 60590

Ion	Ratio	Lower	Upper
122	100		
105	126.4	104.2	144.2
77	96.1	70.2	110.2







#33
4-Chloroaniline
Concen: 40.373 ng
RT: 11.184 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 129600

Ion	Ratio	Lower	Upper
127	100		
129	31.2	25.5	38.3
65	39.0	32.2	48.4
92	21.2	18.2	27.2







#34
Hexachlorobutadiene
Concen: 45.306 ng
RT: 11.384 min Scan# 1412
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:225 Resp: 95906

Ion	Ratio	Lower	Upper
225	100		
223	61.5	48.8	73.2
227	64.3	49.5	74.3







#35
Caprolactam
Concen: 35.362 ng
RT: 11.948 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 32311

Ion	Ratio	Lower	Upper
113	100		
55	292.9	290.1	330.1
56	217.3	183.8	223.8







#36
4-Chloro-3-methylphenol
Concen: 40.462 ng
RT: 12.289 min Scan# 1566
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:107 Resp: 115865

Ion	Ratio	Lower	Upper
107	100		
144	23.0	19.1	28.7
142	71.7	58.6	87.8







#37
2-Methylnaphthalene
Concen: 41.085 ng
RT: 12.688 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 249124

Ion	Ratio	Lower	Upper
142	100		
141	84.2	68.9	103.3
115	40.0	30.6	46.0







#38
1-Methylnaphthalene
Concen: 40.940 ng
RT: 12.900 min Scan# 1670
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:142 Resp: 245754

Ion	Ratio	Lower	Upper
142	100		
141	86.5	73.3	109.9
116	4.2	3.1	4.7







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.839 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 108882

Ion	Ratio	Lower	Upper
164	100		
162	104.0	77.0	115.6
160	44.7	34.4	51.6







#40
1,2,4,5-Tetrachlorobenzene
Concen: 44.962 ng
RT: 13.047 min Scan# 1695
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:216 Resp: 164299

Ion	Ratio	Lower	Upper
216	100		
214	77.0	61.4	92.2
179	18.6	15.4	23.0
108	14.6	11.8	17.6







#41
Hexachlorocyclopentadiene
Concen: 52.985 ng
RT: 13.035 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 87148

Ion	Ratio	Lower	Upper
237	100		
235	63.5	44.1	84.1
272	12.5	0.0	32.4







#42
2,4,6-Tribromophenol
Concen: 86.403 ng
RT: 16.319 min Scan# 2252
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:330 Resp: 135782

Ion	Ratio	Lower	Upper
330	100		
332	96.5	78.0	117.0
141	27.8	21.8	32.6







#43
2,4,6-Trichlorophenol
Concen: 47.391 ng
RT: 13.276 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 99057

Ion	Ratio	Lower	Upper
196	100		
198	91.5	76.6	115.0
200	30.3	26.0	39.0







#44
2,4,5-Trichlorophenol
Concen: 44.415 ng
RT: 13.346 min Scan# 1746
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:196 Resp: 107132

Ion	Ratio	Lower	Upper
196	100		
198	96.8	74.2	111.4
97	48.4	37.0	55.6
132	23.8	18.5	27.7







#45
2-Fluorobiphenyl
Concen: 89.318 ng
RT: 13.470 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 641606

Ion	Ratio	Lower	Upper
172	100		
171	36.4	29.4	44.0
170	23.5	19.1	28.7







#46
1,1'-Biphenyl
Concen: 43.180 ng
RT: 13.681 min Scan# 1803
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:154 Resp: 334243

Ion	Ratio	Lower	Upper
154	100		
153	41.3	22.4	62.4
76	11.8	0.0	33.5







#47
2-Chloronaphthalene
Concen: 43.804 ng
RT: 13.728 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 255946

Ion	Ratio	Lower	Upper
162	100		
127	38.4	32.2	48.2
164	33.7	27.5	41.3







#48
2-Nitroaniline
Concen: 42.351 ng
RT: 13.910 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion: 65 Resp: 82275

Ion	Ratio	Lower	Upper
65	100		
92	59.5	48.3	72.5
138	88.3	74.9	112.3







#49
Acenaphthylene
Concen: 41.830 ng
RT: 14.562 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 385513

Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.4	23.2
153	13.2	11.0	16.4







#50
Dimethylphthalate
Concen: 40.572 ng
RT: 14.280 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:163 Resp: 338032

Ion	Ratio	Lower	Upper
163	100		
194	3.8	2.8	4.2
164	9.3	7.7	11.5







#51
2,6-Dinitrotoluene
Concen: 43.351 ng
RT: 14.398 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 59140

Ion	Ratio	Lower	Upper
165	100		
63	69.6	57.4	86.0
89	56.6	47.8	71.8







#52
Acenaphthene
Concen: 42.377 ng
RT: 14.903 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:154 Resp: 266835

Ion	Ratio	Lower	Upper
154	100		
153	120.6	92.2	138.4
152	56.4	43.2	64.8







#53
3-Nitroaniline
Concen: 43.773 ng
RT: 14.721 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 62676

Ion	Ratio	Lower	Upper
138	100		
108	12.2	11.4	17.0
92	115.5	95.3	142.9







#54
2,4-Dinitrophenol
Concen: 44.922 ng
RT: 14.927 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:184 Resp: 35680

Ion	Ratio	Lower	Upper
184	100		
63	63.4	56.6	85.0
154	116.6	158.6	238.0#







#55
Dibenzofuran
Concen: 41.281 ng
RT: 15.238 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 420237

Ion	Ratio	Lower	Upper
168	100		
139	34.4	27.8	41.8
169	13.2	10.3	15.5







#56
4-Nitrophenol
Concen: 39.540 ng
RT: 15.015 min Scan# 2030
Delta R.T. 0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:139 Resp: 52490

Ion	Ratio	Lower	Upper
139	100		
109	96.1	67.8	107.8
65	119.8	103.7	143.7







#57
2,4-Dinitrotoluene
Concen: 41.073 ng
RT: 15.179 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 86302

Ion	Ratio	Lower	Upper
165	100		
63	51.4	41.8	62.8
89	71.9	58.1	87.1
182	2.2	1.4	2.0#







#58
Fluorene
Concen: 40.432 ng
RT: 15.884 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:166 Resp: 321438

Ion	Ratio	Lower	Upper
166	100		
165	101.3	81.5	122.3
167	13.0	11.4	17.0







#59
2,3,4,6-Tetrachlorophenol
Concen: 44.298 ng
RT: 15.455 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 102176

Ion	Ratio	Lower	Upper
232	100		
131	34.0	26.7	40.1
130	2.2	1.6	2.4
166	24.7	20.5	30.7







#60
Diethylphthalate
Concen: 38.582 ng
RT: 15.644 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:149 Resp: 328215

Ion	Ratio	Lower	Upper
149	100		
177	21.8	18.0	27.0
150	12.2	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 40.979 ng
RT: 15.873 min Scan# 21
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 182832

Ion	Ratio	Lower	Upper
204	100		
206	32.4	25.6	38.4
141	51.8	40.6	60.8







#62
4-Nitroaniline
Concen: 41.786 ng
RT: 15.884 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:138 Resp: 66563

Ion	Ratio	Lower	Upper
138	100		
92	53.3	36.1	76.1
108	95.7	75.6	115.6







#63
Azobenzene
Concen: 39.183 ng
RT: 16.161 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 305844

Ion	Ratio	Lower	Upper
77	100		
182	25.0	4.9	44.9
105	20.8	1.3	41.3
51	38.6	19.0	59.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.577 min Scan# 2466
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:188 Resp: 241683

Ion	Ratio	Lower	Upper
188	100		
94	8.1	5.7	8.5
80	8.6	6.8	10.2







#65
4,6-Dinitro-2-methylphenol
Concen: 47.448 ng
RT: 15.943 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 50385

Ion	Ratio	Lower	Upper
198	100		
51	51.9	24.7	64.7
105	41.5	18.0	58.0







#66
n-Nitrosodiphenylamine
Concen: 43.085 ng
RT: 16.078 min Scan# 2211
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:169 Resp: 281278

Ion	Ratio	Lower	Upper
169	100		
168	71.4	56.3	84.5
167	38.2	29.0	43.6







#67
4-Bromophenyl-phenylether
Concen: 44.163 ng
RT: 16.766 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 126485

Ion	Ratio	Lower	Upper
248	100		
250	96.3	77.7	116.5
141	58.3	48.6	72.8







#68
Hexachlorobenzene
Concen: 44.415 ng
RT: 16.889 min Scan# 2349
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:284 Resp: 145122

Ion	Ratio	Lower	Upper
284	100		
142	24.3	21.0	31.6
249	32.9	24.9	37.3







#69
Atrazine
Concen: 40.031 ng
RT: 17.018 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 112465

Ion	Ratio	Lower	Upper
200	100		
173	24.1	0.9	40.9
215	45.3	27.0	67.0







#70
Pentachlorophenol
Concen: 44.939 ng
RT: 17.224 min Scan# 2406
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:266 Resp: 91563

Ion	Ratio	Lower	Upper
266	100		
268	67.2	50.3	75.5
264	63.9	47.3	70.9







#71
Phenanthrene
Concen: 41.397 ng
RT: 17.618 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 547128

Ion	Ratio	Lower	Upper
178	100		
176	19.1	14.7	22.1
179	15.4	12.2	18.4







#72
Anthracene
Concen: 40.436 ng
RT: 17.712 min Scan# 2489
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:178 Resp: 543644

Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.4	21.6
179	15.1	12.4	18.6







#73
Carbazole
Concen: 39.556 ng
RT: 17.970 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 468954

Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.9	26.9
139	12.2	9.7	14.5







#74
Di-n-butylphthalate
Concen: 40.184 ng
RT: 18.528 min Scan# 2628
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:149 Resp: 550901

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.0	10.6
104	5.3	4.1	6.1







#75
Fluoranthene
Concen: 39.193 ng
RT: 19.615 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 648927

Ion	Ratio	Lower	Upper
202	100		
101	8.1	0.0	28.7
203	17.7	0.0	37.0







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.854 min Scan# 3194
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:240 Resp: 261016

Ion	Ratio	Lower	Upper
240	100		
120	7.1	6.5	9.7
236	25.3	20.2	30.2







#77
Benzidine
Concen: 45.404 ng
RT: 19.780 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 265517

Ion	Ratio	Lower	Upper
184	100		
185	13.8	11.4	17.2
183	11.4	9.4	14.2







#78
Pyrene
Concen: 40.467 ng
RT: 19.974 min Scan# 2874
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:202 Resp: 690102

Ion	Ratio	Lower	Upper
202	100		
200	20.6	17.0	25.4
203	17.1	13.8	20.6







#79
Terphenyl-d14
Concen: 81.066 ng
RT: 20.168 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 1121232

Ion	Ratio	Lower	Upper
244	100		
212	7.2	6.1	9.1
122	6.8	5.6	8.4







#80
Butylbenzylphthalate
Concen: 46.801 ng
RT: 20.849 min Scan# 3023
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:149 Resp: 241218

Ion	Ratio	Lower	Upper
149	100		
91	69.9	59.2	88.8
206	22.1	18.1	27.1







#81
Benzo(a)anthracene
Concen: 42.093 ng
RT: 21.836 min Scan# 3
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 728489

Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.9	32.9
229	19.4	15.6	23.4







#82
3,3'-Dichlorobenzidine
Concen: 43.737 ng
RT: 21.736 min Scan# 3174
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:252 Resp: 252767

Ion	Ratio	Lower	Upper
252	100		
254	65.6	50.8	76.2
126	7.2	5.8	8.6







#83
Chrysene
Concen: 41.154 ng
RT: 21.901 min Scan# 3
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 677717

Ion	Ratio	Lower	Upper
228	100		
226	29.9	23.3	34.9
229	19.3	15.2	22.8







#84
Bis(2-ethylhexyl)phthalate
Concen: 45.515 ng
RT: 21.742 min Scan# 3175
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:149 Resp: 350537

Ion	Ratio	Lower	Upper
149	100		
167	27.1	20.0	30.0
279	4.3	3.1	4.7







#85
Di-n-octyl phthalate
Concen: 48.350 ng
RT: 23.000 min Scan# 3
Delta R.T. -0.012 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 627808

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	13.0	10.8	16.2







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.197 min Scan# 3763
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:264 Resp: 298207

Ion	Ratio	Lower	Upper
264	100		
260	22.9	19.0	28.4
265	21.9	17.6	26.4







#87
Indeno(1,2,3-cd)pyrene
Concen: 41.308 ng
RT: 29.040 min Scan# 4
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 906170

Ion	Ratio	Lower	Upper
276	100		
138	11.7	10.8	16.2
277	25.3	20.1	30.1







#88
Benzo(b)fluoranthene
Concen: 40.302 ng
RT: 24.134 min Scan# 3582
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:252 Resp: 736965

Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.8	26.8
125	6.4	5.6	8.4







#89
Benzo(k)fluoranthene
Concen: 40.260 ng
RT: 24.210 min Scan# 3
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 736695

Ion	Ratio	Lower	Upper
252	100		
253	21.6	16.8	25.2
125	6.1	5.4	8.0







#90
Benzo(a)pyrene
Concen: 41.067 ng
RT: 25.038 min Scan# 3736
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:252 Resp: 690279

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.6	26.4
125	7.8	6.1	9.1







#91
Dibenzo(a,h)anthracene
Concen: 41.529 ng
RT: 29.104 min Scan# 4
Delta R.T. -0.006 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 740147

Ion	Ratio	Lower	Upper
278	100		
139	10.4	8.0	12.0
279	23.9	19.8	29.6






#92
Benzo(g,h,i)perylene
Concen: 41.412 ng
RT: 30.232 min Scan# 4620
Delta R.T. -0.000 min
Lab File: BG064618.D
Acq: 29 Oct 2025 11:22

Tgt Ion:276 Resp: 705048

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.1	28.7
138	12.4	11.3	16.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2		1,4-Dioxane	0.528	0.560	-6.1	97	0.00
3		Pyridine	1.584	1.622	-2.4	89	0.00
4		n-Nitrosodimethylamine	0.850	0.861	-1.3	87	0.00
5	S	2-Fluorophenol	1.192	1.266	-6.2	89	0.00
6		Aniline	2.225	2.273	-2.2	87	0.00
7	S	Phenol-d6	1.635	1.656	-1.3	85	0.00
8		2-Chlorophenol	1.247	1.377	-10.4	91	0.00
9		Benzaldehyde	0.893	0.838	6.2	81	0.00
10	C	Phenol	1.808	1.838	-1.7	86	0.00
11		bis(2-Chloroethyl)ether	1.329	1.382	-4.0	91	0.00
12		1,3-Dichlorobenzene	1.435	1.546	-7.7	93	0.00
13	C	1,4-Dichlorobenzene	1.460	1.553	-6.4	92	0.00
14		1,2-Dichlorobenzene	1.408	1.517	-7.7	93	0.00
15		Benzyl Alcohol	1.348	1.422	-5.5	91	0.00
16		2,2'-oxybis(1-Chloropropane	3.487	3.674	-5.4	90	0.00
17		2-Methylphenol	1.201	1.261	-5.0	90	0.00
18		Hexachloroethane	0.512	0.567	-10.7	95	0.00
19	P	n-Nitroso-di-n-propylamine	1.130	1.223	-8.2	91	0.00
20		3+4-Methylphenols	1.688	1.742	-3.2	87	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
22		Acetophenone	0.547	0.581	-6.2	90	0.00
23	S	Nitrobenzene-d5	0.333	0.401	-20.4	96	0.00
24		Nitrobenzene	0.362	0.429	-18.5	95	0.00
25		Isophorone	0.799	0.833	-4.3	87	0.00
26	C	2-Nitrophenol	0.121	0.168	-38.8#	111	0.00
27		2,4-Dimethylphenol	0.297	0.311	-4.7	84	0.00
28		bis(2-Chloroethoxy)methane	0.455	0.480	-5.5	89	0.00
29	C	2,4-Dichlorophenol	0.320	0.352	-10.0	88	0.00
30		1,2,4-Trichlorobenzene	0.363	0.399	-9.9	92	0.00
31		Naphthalene	1.109	1.159	-4.5	89	0.00
32		Benzoic acid	0.207	0.203	1.9	82	0.00
33		4-Chloroaniline	0.431	0.435	-0.9	86	0.00
34	C	Hexachlorobutadiene	0.284	0.322	-13.4	97	0.00
35		Caprolactam	0.123	0.108	12.2	73	0.00
36	C	4-Chloro-3-methylphenol	0.384	0.389	-1.3	83	0.00
37		2-Methylnaphthalene	0.814	0.836	-2.7	87	0.00
38		1-Methylnaphthalene	0.806	0.825	-2.4	87	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
40		1,2,4,5-Tetrachlorobenzene	0.671	0.754	-12.4	90	0.00
41	P	Hexachlorocyclopentadiene	0.302	0.400	-32.5#	103	0.00
42	S	2,4,6-Tribromophenol	0.289	0.312	-8.0	82	0.00
43	C	2,4,6-Trichlorophenol	0.384	0.455	-18.5	90	0.00
44		2,4,5-Trichlorophenol	0.443	0.492	-11.1	85	0.00
45	S	2-Fluorobiphenyl	1.319	1.473	-11.7	89	0.00
46		1,1'-Biphenyl	1.422	1.535	-7.9	86	0.00
47		2-Chloronaphthalene	1.073	1.175	-9.5	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.311	0.378	-21.5	90	0.00
49		Acenaphthylene	1.693	1.770	-4.5	83	0.00
50		Dimethylphthalate	1.530	1.552	-1.4	79	0.00
51		2,6-Dinitrotoluene	0.218	0.272	-24.8	92	0.00
52	C	Acenaphthene	1.157	1.225	-5.9	84	0.00
53		3-Nitroaniline	0.263	0.288	-9.5	82	0.00
54	P	2,4-Dinitrophenol	0.132	0.164	-24.2	100	0.00
55		Dibenzofuran	1.870	1.930	-3.2	82	0.00
56	P	4-Nitrophenol	0.244	0.241	1.2	77	0.00
57		2,4-Dinitrotoluene	0.340	0.396	-16.5	86	0.00
58		Fluorene	1.460	1.476	-1.1	80	0.00
59		2,3,4,6-Tetrachlorophenol	0.424	0.469	-10.6	84	0.00
60		Diethylphthalate	1.563	1.507	3.6	76	0.00
61		4-Chlorophenyl-phenylether	0.820	0.840	-2.4	81	0.00
62		4-Nitroaniline	0.293	0.306	-4.4	78	0.00
63		Azobenzene	1.434	1.404	2.1	76	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	77	0.00
65		4,6-Dinitro-2-methylphenol	0.079	0.104	-31.6#	97	0.00
66	c	n-Nitrosodiphenylamine	0.540	0.582	-7.8	77	0.00
67		4-Bromophenyl-phenylether	0.237	0.262	-10.5	78	0.00
68		Hexachlorobenzene	0.270	0.300	-11.1	79	0.00
69		Atrazine	0.232	0.233	-0.4	73	0.00
70	C	Pentachlorophenol	0.169	0.189	-11.8	81	0.00
71		Phenanthrene	1.094	1.132	-3.5	75	0.00
72		Anthracene	1.113	1.125	-1.1	73	0.00
73		Carbazole	0.981	0.970	1.1	73	0.00
74		Di-n-butylphthalate	1.134	1.140	-0.5	71	0.00
75	C	Fluoranthene	1.370	1.343	2.0	72	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	78	0.00
77		Benzidine	0.448	0.509	-13.6	81	0.00
78		Pyrene	1.307	1.322	-1.1	73	0.00
79	S	Terphenyl-d14	1.060	1.074	-1.3	74	0.00
80		Butylbenzylphthalate	0.395	0.462	-17.0	80	0.00
81		Benzo(a)anthracene	1.326	1.395	-5.2	77	0.00
82		3,3'-Dichlorobenzidine	0.443	0.484	-9.3	80	0.00
83		Chrysene	1.262	1.298	-2.9	76	0.00
84		Bis(2-ethylhexyl)phthalate	0.590	0.671	-13.7	78	0.00
85	c	Di-n-octyl phthalate	0.995	1.203	-20.9#	87	-0.01
86	I	Perylene-d12	1.000	1.000	0.0	85	0.00
87		Indeno(1,2,3-cd)pyrene	1.471	1.519	-3.3	82	0.00
88		Benzo(b)fluoranthene	1.226	1.236	-0.8	79	0.00
89		Benzo(k)fluoranthene	1.227	1.235	-0.7	82	0.00
90	C	Benzo(a)pyrene	1.127	1.157	-2.7	82	0.00
91		Dibenzo(a,h)anthracene	1.195	1.241	-3.8	83	0.00
92		Benzo(g,h,i)perylene	1.142	1.182	-3.5	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 2

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2		1,4-Dioxane	40.000	42.495	-6.2	97	0.00
3		Pyridine	40.000	40.966	-2.4	89	0.00
4		n-Nitrosodimethylamine	40.000	40.532	-1.3	87	0.00
5	S	2-Fluorophenol	80.000	84.991	-6.2	89	0.00
6		Aniline	40.000	40.860	-2.1	87	0.00
7	S	Phenol-d6	80.000	81.009	-1.3	85	0.00
8		2-Chlorophenol	40.000	44.170	-10.4	91	0.00
9		Benzaldehyde	40.000	37.558	6.1	81	0.00
10	C	Phenol	40.000	40.664	-1.7	86	0.00
11		bis(2-Chloroethyl)ether	40.000	41.589	-4.0	91	0.00
12		1,3-Dichlorobenzene	40.000	43.073	-7.7	93	0.00
13	C	1,4-Dichlorobenzene	40.000	42.551	-6.4	92	0.00
14		1,2-Dichlorobenzene	40.000	43.105	-7.8	93	0.00
15		Benzyl Alcohol	40.000	42.201	-5.5	91	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	42.138	-5.3	90	0.00
17		2-Methylphenol	40.000	42.028	-5.1	90	0.00
18		Hexachloroethane	40.000	44.282	-10.7	95	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	43.275	-8.2	91	0.00
20		3+4-Methylphenols	40.000	41.279	-3.2	87	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	90	0.00
22		Acetophenone	40.000	42.523	-6.3	90	0.00
23	S	Nitrobenzene-d5	80.000	96.333	-20.4	96	0.00
24		Nitrobenzene	40.000	47.355	-18.4	95	0.00
25		Isophorone	40.000	41.695	-4.2	87	0.00
26	C	2-Nitrophenol	40.000	47.979	-19.9	111	0.00
27		2,4-Dimethylphenol	40.000	41.823	-4.6	84	0.00
28		bis(2-Chloroethoxy)methane	40.000	42.170	-5.4	89	0.00
29	C	2,4-Dichlorophenol	40.000	44.058	-10.1	88	0.00
30		1,2,4-Trichlorobenzene	40.000	43.927	-9.8	92	0.00
31		Naphthalene	40.000	41.819	-4.5	89	0.00
32		Benzoic acid	40.000	37.459	6.4	82	0.00
33		4-Chloroaniline	40.000	40.373	-0.9	86	0.00
34	C	Hexachlorobutadiene	40.000	45.306	-13.3	97	0.00
35		Caprolactam	40.000	35.362	11.6	73	0.00
36	C	4-Chloro-3-methylphenol	40.000	40.462	-1.2	83	0.00
37		2-Methylnaphthalene	40.000	41.085	-2.7	87	0.00
38		1-Methylnaphthalene	40.000	40.940	-2.3	87	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	83	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	44.962	-12.4	90	0.00
41	P	Hexachlorocyclopentadiene	40.000	52.985	-32.5#	103	0.00
42	S	2,4,6-Tribromophenol	80.000	86.403	-8.0	82	0.00
43	C	2,4,6-Trichlorophenol	40.000	47.391	-18.5	90	0.00
44		2,4,5-Trichlorophenol	40.000	44.415	-11.0	85	0.00
45	S	2-Fluorobiphenyl	80.000	89.318	-11.6	89	0.00
46		1,1'-Biphenyl	40.000	43.180	-7.9	86	0.00
47		2-Chloronaphthalene	40.000	43.804	-9.5	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	42.351	-5.9	90	0.00
49		Acenaphthylene	40.000	41.830	-4.6	83	0.00
50		Dimethylphthalate	40.000	40.572	-1.4	79	0.00
51		2,6-Dinitrotoluene	40.000	43.351	-8.4	92	0.00
52	C	Acenaphthene	40.000	42.377	-5.9	84	0.00
53		3-Nitroaniline	40.000	43.773	-9.4	82	0.00
54	P	2,4-Dinitrophenol	40.000	44.922	-12.3	100	0.00
55		Dibenzofuran	40.000	41.281	-3.2	82	0.00
56	P	4-Nitrophenol	40.000	39.540	1.2	77	0.00
57		2,4-Dinitrotoluene	40.000	41.073	-2.7	86	0.00
58		Fluorene	40.000	40.432	-1.1	80	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	44.298	-10.7	84	0.00
60		Diethylphthalate	40.000	38.582	3.5	76	0.00
61		4-Chlorophenyl-phenylether	40.000	40.979	-2.4	81	0.00
62		4-Nitroaniline	40.000	41.786	-4.5	78	0.00
63		Azobenzene	40.000	39.183	2.0	76	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	77	0.00
65		4,6-Dinitro-2-methylphenol	40.000	47.448	-18.6	97	0.00
66	c	n-Nitrosodiphenylamine	40.000	43.085	-7.7	77	0.00
67		4-Bromophenyl-phenylether	40.000	44.163	-10.4	78	0.00
68		Hexachlorobenzene	40.000	44.415	-11.0	79	0.00
69		Atrazine	40.000	40.031	-0.1	73	0.00
70	C	Pentachlorophenol	40.000	44.939	-12.3	81	0.00
71		Phenanthrene	40.000	41.397	-3.5	75	0.00
72		Anthracene	40.000	40.436	-1.1	73	0.00
73		Carbazole	40.000	39.556	1.1	73	0.00
74		Di-n-butylphthalate	40.000	40.184	-0.5	71	0.00
75	C	Fluoranthene	40.000	39.193	2.0	72	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	78	0.00
77		Benzidine	40.000	45.404	-13.5	81	0.00
78		Pyrene	40.000	40.467	-1.2	73	0.00
79	S	Terphenyl-d14	80.000	81.066	-1.3	74	0.00
80		Butylbenzylphthalate	40.000	46.801	-17.0	80	0.00
81		Benzo(a)anthracene	40.000	42.093	-5.2	77	0.00
82		3,3'-Dichlorobenzidine	40.000	43.737	-9.3	80	0.00
83		Chrysene	40.000	41.154	-2.9	76	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	45.515	-13.8	78	0.00
85	c	Di-n-octyl phthalate	40.000	48.350	-20.9#	87	-0.01
86	I	Perylene-d12	20.000	20.000	0.0	85	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	41.308	-3.3	82	0.00
88		Benzo(b)fluoranthene	40.000	40.302	-0.8	79	0.00
89		Benzo(k)fluoranthene	40.000	40.260	-0.6	82	0.00
90	C	Benzo(a)pyrene	40.000	41.067	-2.7	82	0.00
91		Dibenzo(a,h)anthracene	40.000	41.529	-3.8	83	0.00
92		Benzo(g,h,i)perylene	40.000	41.412	-3.5	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 05:47:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG No.: Q3434

Instrument ID: BNA_G Calibration Date/Time: 10/30/2025 11:36

Lab File ID: BG064636.D Init. Calib. Date(s): 10/24/2025 10/24/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:00 15:32

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.192	1.233		3.4	
Benzaldehyde	0.893	0.861		-3.6	
Phenol-d6	1.635	1.603		-2.0	
Phenol	1.808	1.761		-2.6	20.0
bis(2-Chloroethyl)ether	1.329	1.296		-2.5	
2-Chlorophenol	1.247	1.369		9.8	
2-Methylphenol	1.201	1.221		1.7	
2,2-oxybis(1-Chloropropane)	3.487	3.562		2.2	
Acetophenone	0.547	0.558		2.0	
3+4-Methylphenols	1.688	1.620		-4.0	
n-Nitroso-di-n-propylamine	1.130	1.116	0.050	-1.2	
Nitrobenzene-d5	0.333	0.420		26.1	
Hexachloroethane	0.512	0.586		14.5	
Nitrobenzene	0.362	0.438		21.0	
Isophorone	0.799	0.808		1.1	
2-Nitrophenol	0.121	0.180		48.8	20.0
2,4-Dimethylphenol	0.297	0.320		7.7	
bis(2-Chloroethoxy)methane	0.455	0.460		1.1	
2,4-Dichlorophenol	0.320	0.361		12.8	20.0
Naphthalene	1.109	1.158		4.4	
4-Chloroaniline	0.431	0.422		-2.1	
Hexachlorobutadiene	0.284	0.320		12.7	20.0
Caprolactam	0.123	0.106		-13.8	
4-Chloro-3-methylphenol	0.384	0.390		1.6	20.0
2-Methylnaphthalene	0.814	0.842		3.4	
Hexachlorocyclopentadiene	0.302	0.413	0.050	36.8	
2,4,6-Trichlorophenol	0.384	0.457		19.0	20.0
2-Fluorobiphenyl	1.319	1.427		8.2	
2,4,5-Trichlorophenol	0.443	0.495		11.7	
1,1-Biphenyl	1.422	1.507		6.0	
2-Chloronaphthalene	1.073	1.145		6.7	
2-Nitroaniline	0.311	0.392		26.0	
Dimethylphthalate	1.530	1.509		-1.4	
Acenaphthylene	1.693	1.756		3.7	
2,6-Dinitrotoluene	0.218	0.281		28.9	
3-Nitroaniline	0.263	0.304		15.6	
Acenaphthene	1.157	1.205		4.1	20.0
2,4-Dinitrophenol	0.132	0.181	0.050	37.1	
4-Nitrophenol	0.244	0.267	0.050	9.4	

Form VII SV-1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02

Lab Code: ACE SDG No.: Q3434

Instrument ID: BNA_G Calibration Date/Time: 10/30/2025 11:36

Lab File ID: BG064636.D Init. Calib. Date(s): 10/24/2025 10/24/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:00 15:32

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.870	1.869		-0.1	
2,4-Dinitrotoluene	0.340	0.413		21.5	
Diethylphthalate	1.563	1.474		-5.7	
4-Chlorophenyl-phenylether	0.820	0.831		1.3	
Fluorene	1.460	1.442		-1.2	
4-Nitroaniline	0.293	0.318		8.5	
4,6-Dinitro-2-methylphenol	0.079	0.117		48.1	
n-Nitrosodiphenylamine	0.540	0.569		5.4	20.0
2,4,6-Tribromophenol	0.289	0.317		9.7	
4-Bromophenyl-phenylether	0.237	0.261		10.1	
Hexachlorobenzene	0.270	0.291		7.8	
Atrazine	0.232	0.232		0.0	
Pentachlorophenol	0.169	0.195		15.4	20.0
Phenanthrene	1.094	1.114		1.8	
Anthracene	1.113	1.144		2.8	
Carbazole	0.981	1.006		2.5	
Di-n-butylphthalate	1.135	1.131		-0.4	
Fluoranthene	1.370	1.451		5.9	20.0
Pyrene	1.307	1.247		-4.6	
Terphenyl-d14	1.060	1.043		-1.6	
Butylbenzylphthalate	0.395	0.445		12.7	
3,3-Dichlorobenzidine	0.443	0.493		11.3	
Benzo(a)anthracene	1.326	1.410		6.3	
Chrysene	1.262	1.324		4.9	
Bis(2-ethylhexyl)phthalate	0.590	0.652		10.5	
Di-n-octyl phthalate	0.995	1.171		17.7	20.0
Benzo(b)fluoranthene	1.226	1.272		3.8	
Benzo(k)fluoranthene	1.227	1.285		4.7	
Benzo(a)pyrene	1.127	1.190		5.6	20.0
Indeno(1,2,3-cd)pyrene	1.471	1.593		8.3	
Dibenzo(a,h)anthracene	1.195	1.283		7.4	
Benzo(g,h,i)perylene	1.142	1.243		8.8	
1,2,4,5-Tetrachlorobenzene	0.671	0.762		13.6	
1,4-Dioxane	0.528	0.546		3.4	20.0
2,3,4,6-Tetrachlorophenol	0.424	0.473		11.6	

All other compounds must meet a minimum RRF of 0.010.

Form VII SV-1

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.219	152	32501	20.000	ng		0.00
21) Naphthalene-d8	11.045	136	127691	20.000	ng		0.00
39) Acenaphthene-d10	14.841	164	94503	20.000	ng		0.00
64) Phenanthrene-d10	17.579	188	210658	20.000	ng		0.00
76) Chrysene-d12	21.856	240	257813	20.000	ng		0.00
86) Perylene-d12	25.199	264	295961	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.745	112	160336	82.805	ng		0.00
7) Phenol-d6	7.355	99	208374	78.424	ng		0.00
23) Nitrobenzene-d5	9.382	82	214287	100.795	ng		0.00
42) 2,4,6-Tribromophenol	16.315	330	119915	87.917	ng		0.00
45) 2-Fluorobiphenyl	13.472	172	539428	86.520	ng		0.00
79) Terphenyl-d14	20.170	244	1075391	78.718	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.589	88	35491	41.400	ng		98
3) Pyridine	3.995	79	105689	41.054	ng		98
4) n-Nitrosodimethylamine	3.906	42	56991	41.282	ng		99
6) Aniline	7.532	93	136655	37.797	ng		97
8) 2-Chlorophenol	7.778	128	88994	43.910	ng		99
9) Benzaldehyde	7.338	77	55986	38.596	ng		93
10) Phenol	7.379	94	114461	38.953	ng		96
11) bis(2-Chloroethyl)ether	7.626	93	84261	39.002	ng		97
12) 1,3-Dichlorobenzene	8.107	146	101335	43.441	ng		93
13) 1,4-Dichlorobenzene	8.254	146	100871	42.528	ng		95
14) 1,2-Dichlorobenzene	8.577	146	98266	42.953	ng		99
15) Benzyl Alcohol	8.448	79	90352	41.237	ng		94
16) 2,2'-oxybis(1-Chloropr...	8.736	45	231528	40.856	ng		99
17) 2-Methylphenol	8.648	107	79347	40.671	ng		97
18) Hexachloroethane	9.318	117	38071	45.739	ng		95
19) n-Nitroso-di-n-propyla...	9.024	70	72573	39.522	ng		96
20) 3+4-Methylphenols	8.977	107	105306	38.381	ng		97
22) Acetophenone	9.042	105	142610	40.856	ng		97
24) Nitrobenzene	9.424	77	111858	48.375	ng		98
25) Isophorone	9.952	82	206469	40.473	ng		98
26) 2-Nitrophenol	10.140	139	46050	51.379	ng	#	90
27) 2,4-Dimethylphenol	10.193	122	81767	43.090	ng		91
28) bis(2-Chloroethoxy)met...	10.434	93	117360	40.406	ng		99
29) 2,4-Dichlorophenol	10.681	162	92269	45.182	ng		97
30) 1,2,4-Trichlorobenzene	10.904	180	101552	43.767	ng		97
31) Naphthalene	11.098	128	295608	41.755	ng		99
32) Benzoic acid	10.299	122	58855	41.641	ng		98
33) 4-Chloroaniline	11.186	127	107772	39.165	ng		97
34) Hexachlorobutadiene	11.386	225	81632	44.987	ng		93
35) Caprolactam	11.950	113	27064	34.554	ng	#	88
36) 4-Chloro-3-methylphenol	12.291	107	99554	40.557	ng		97
37) 2-Methylnaphthalene	12.684	142	215078	41.379	ng		97
38) 1-Methylnaphthalene	12.902	142	208518	40.524	ng		98
40) 1,2,4,5-Tetrachloroben...	13.049	216	144011	45.407	ng		99
41) Hexachlorocyclopentadiene	13.031	237	78048	54.673	ng		96
43) 2,4,6-Trichlorophenol	13.278	196	86362	47.604	ng		98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.348	196	93561	44.690	ng		98
46) 1,1'-Biphenyl	13.677	154	284894	42.404	ng		94
47) 2-Chloronaphthalene	13.724	162	216419	42.675	ng		99
48) 2-Nitroaniline	13.906	65	74045	43.795	ng		92
49) Acenaphthylene	14.565	152	331834	41.484	ng		99
50) Dimethylphthalate	14.283	163	285287	39.451	ng		97
51) 2,6-Dinitrotoluene	14.394	165	53162	44.784	ng		98
52) Acenaphthene	14.905	154	227801	41.683	ng		97
53) 3-Nitroaniline	14.723	138	57551	46.309	ng		97
54) 2,4-Dinitrophenol	14.923	184	34138	48.950	ng		85
55) Dibenzofuran	15.234	168	353239	39.980	ng		98
56) 4-Nitrophenol	15.011	139	50480	43.812	ng		97
57) 2,4-Dinitrotoluene	15.176	165	78142	42.716	ng	#	98
58) Fluorene	15.881	166	272466	39.487	ng		98
59) 2,3,4,6-Tetrachlorophenol	15.452	232	89397	44.655	ng		98
60) Diethylphthalate	15.640	149	278640	37.738	ng		99
61) 4-Chlorophenyl-phenyle...	15.869	204	157055	40.558	ng		97
62) 4-Nitroaniline	15.881	138	60132	43.493	ng		93
63) Azobenzene	16.163	77	257601	38.024	ng		97
65) 4,6-Dinitro-2-methylph...	15.939	198	49140	52.445	ng		96
66) n-Nitrosodiphenylamine	16.080	169	239610	42.108	ng		98
67) 4-Bromophenyl-phenylether	16.762	248	109900	44.024	ng		99
68) Hexachlorobenzene	16.885	284	122806	43.121	ng		96
69) Atrazine	17.015	200	97691	39.893	ng		97
70) Pentachlorophenol	17.220	266	82346	46.367	ng		92
71) Phenanthrene	17.620	178	469425	40.749	ng		99
72) Anthracene	17.708	178	482051	41.136	ng		99
73) Carbazole	17.972	167	423668	40.999	ng		99
74) Di-n-butylphthalate	18.525	149	476308	39.860	ng		99
75) Fluoranthene	19.612	202	611318	42.360	ng		99
77) Benzidine	19.782	184	240984	41.721	ng		98
78) Pyrene	19.976	202	643136	38.182	ng		100
80) Butylbenzylphthalate	20.851	149	229670	45.114	ng		96
81) Benzo(a)anthracene	21.832	228	726865	42.521	ng		98
82) 3,3'-Dichlorobenzidine	21.738	252	254141	44.521	ng		100
83) Chrysene	21.903	228	682939	41.987	ng		98
84) Bis(2-ethylhexyl)phtha...	21.738	149	336059	44.177	ng		99
85) Di-n-octyl phthalate	22.996	149	603742	47.074	ng		98
87) Indeno(1,2,3-cd)pyrene	29.036	276	943017	43.313	ng		98
88) Benzo(b)fluoranthene	24.136	252	753080	41.496	ng		98
89) Benzo(k)fluoranthene	24.206	252	760727	41.889	ng		97
90) Benzo(a)pyrene	25.041	252	704164	42.211	ng		100
91) Dibenzo(a,h)anthracene	29.106	278	759427	42.934	ng		98
92) Benzo(g,h,i)perylene	30.229	276	735814	43.547	ng		97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.219 min Scan# 8
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 32501

Ion	Ratio	Lower	Upper
152	100		
150	161.4	122.2	183.4
115	59.3	50.8	76.2







#2
1,4-Dioxane
Concen: 41.400 ng
RT: 3.589 min Scan# 85
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion: 88 Resp: 35491

Ion	Ratio	Lower	Upper
88	100		
58	117.3	92.4	138.6
43	43.0	33.0	49.6







#3
Pyridine
Concen: 41.054 ng
RT: 3.995 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 105689

Ion	Ratio	Lower	Upper
79	100		
52	121.4	94.2	141.2
51	57.5	46.1	69.1







#4
n-Nitrosodimethylamine
Concen: 41.282 ng
RT: 3.906 min Scan# 139
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion: 42 Resp: 56991

Ion	Ratio	Lower	Upper
42	100		
74	96.5	76.1	114.1
44	8.6	6.3	9.5







#5
2-Fluorophenol
Concen: 82.805 ng
RT: 5.745 min Scan# 4
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 160336

Ion	Ratio	Lower	Upper
112	100		
64	77.0	60.7	91.1
63	40.6	33.4	50.0







#6
Aniline
Concen: 37.797 ng
RT: 7.532 min Scan# 756
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion: 93 Resp: 136655

Ion	Ratio	Lower	Upper
93	100		
66	47.5	37.0	55.6
65	25.4	18.3	27.5







#7
Phenol-d6
Concen: 78.424 ng
RT: 7.355 min Scan# 7
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 208374

Ion	Ratio	Lower	Upper
99	100		
42	23.6	18.3	27.5
71	37.4	28.6	43.0







#8
2-Chlorophenol
Concen: 43.910 ng
RT: 7.778 min Scan# 798
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:128 Resp: 88994

Ion	Ratio	Lower	Upper
128	100		
130	33.4	13.5	53.5
64	64.6	45.2	85.2







#9
Benzaldehyde
Concen: 38.596 ng
RT: 7.338 min Scan# 7
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 55986

Ion	Ratio	Lower	Upper
77	100		
105	87.9	72.9	112.9
106	82.5	70.8	110.8







#10
Phenol
Concen: 38.953 ng
RT: 7.379 min Scan# 730
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion: 94 Resp: 114461

Ion	Ratio	Lower	Upper
94	100		
65	34.4	15.8	55.8
66	47.3	24.0	64.0







#11
bis(2-Chloroethyl)ether
Concen: 39.002 ng
RT: 7.626 min Scan# 7
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 84261

Ion	Ratio	Lower	Upper
93	100		
63	106.9	91.3	131.3
95	32.6	13.3	53.3







#12
1,3-Dichlorobenzene
Concen: 43.441 ng
RT: 8.107 min Scan# 854
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:146 Resp: 101335

Ion	Ratio	Lower	Upper
146	100		
148	62.6	55.3	82.9
75	32.5	28.2	42.2







#13
1,4-Dichlorobenzene
Concen: 42.528 ng
RT: 8.254 min Scan# 8
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 100871

Ion	Ratio	Lower	Upper
146	100		
148	64.4	56.2	84.4
111	43.8	35.4	53.0







#14
1,2-Dichlorobenzene
Concen: 42.953 ng
RT: 8.577 min Scan# 934
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:146 Resp: 98266

Ion	Ratio	Lower	Upper
146	100		
148	65.0	51.5	77.3
111	44.0	35.0	52.6







#15
Benzyl Alcohol
Concen: 41.237 ng
RT: 8.448 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 90352

Ion	Ratio	Lower	Upper
79	100		
108	67.8	58.1	87.1
77	61.5	52.8	79.2







#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.856 ng
RT: 8.736 min Scan# 961
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion: 45 Resp: 231528

Ion	Ratio	Lower	Upper
45	100		
77	8.1	0.0	28.5
79	5.8	0.0	25.6







#17
2-Methylphenol
Concen: 40.671 ng
RT: 8.648 min Scan# 94
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 79347

Ion	Ratio	Lower	Upper
107	100		
108	109.3	90.0	135.0
77	48.4	37.8	56.8
79	49.1	40.5	60.7







#18
Hexachloroethane
Concen: 45.739 ng
RT: 9.318 min Scan# 1060
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:117 Resp: 38071

Ion	Ratio	Lower	Upper
117	100		
119	87.7	75.4	113.2
201	120.0	99.2	148.8







#19
n-Nitroso-di-n-propylamine
Concen: 39.522 ng
RT: 9.024 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 72573

Ion	Ratio	Lower	Upper
70	100		
42	85.5	65.2	97.8
101	9.6	7.9	11.9
130	17.5	14.3	21.5







#20
3+4-Methylphenols
Concen: 38.381 ng
RT: 8.977 min Scan# 1002
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:107 Resp: 105306

Ion	Ratio	Lower	Upper
107	100		
108	92.2	70.3	110.3
77	37.6	14.6	54.6
79	33.4	11.7	51.7







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.045 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 127691

Ion	Ratio	Lower	Upper
136	100		
137	10.7	9.2	13.8
54	14.4	12.6	19.0
68	6.0	5.7	8.5







#22
Acetophenone
Concen: 40.856 ng
RT: 9.042 min Scan# 1013
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:105 Resp: 142610

Ion	Ratio	Lower	Upper
105	100		
71	2.1	1.5	2.3
51	39.6	33.7	50.5
120	20.4	16.1	24.1







#23
Nitrobenzene-d5
Concen: 100.795 ng
RT: 9.382 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 214287

Ion	Ratio	Lower	Upper
82	100		
128	38.3	31.3	46.9
54	70.5	56.6	84.8







#24
Nitrobenzene
Concen: 48.375 ng
RT: 9.424 min Scan# 1078
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion: 77 Resp: 111858

Ion	Ratio	Lower	Upper
77	100		
123	40.6	31.0	46.6
65	15.7	12.7	19.1







#25
Isophorone
Concen: 40.473 ng
RT: 9.952 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 206469

Ion	Ratio	Lower	Upper
82	100		
95	7.8	6.1	9.1
138	15.2	13.0	19.4







#26
2-Nitrophenol
Concen: 51.379 ng
RT: 10.140 min Scan# 1200
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:139 Resp: 46050

Ion	Ratio	Lower	Upper
139	100		
109	37.8	24.6	37.0#
65	65.4	47.5	71.3







#27
2,4-Dimethylphenol
Concen: 43.090 ng
RT: 10.193 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 81767

Ion	Ratio	Lower	Upper
122	100		
107	115.0	103.0	154.4
121	57.5	47.8	71.8







#28
bis(2-Chloroethoxy)methane
Concen: 40.406 ng
RT: 10.434 min Scan# 1250
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion: 93 Resp: 117360

Ion	Ratio	Lower	Upper
93	100		
95	31.7	25.4	38.2
123	9.4	8.4	12.6







#29
2,4-Dichlorophenol
Concen: 45.182 ng
RT: 10.681 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 92269

Ion	Ratio	Lower	Upper
162	100		
164	66.8	48.7	88.7
98	34.7	16.9	56.9







#30
1,2,4-Trichlorobenzene
Concen: 43.767 ng
RT: 10.904 min Scan# 1330
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:180 Resp: 101552

Ion	Ratio	Lower	Upper
180	100		
182	93.2	76.9	115.3
145	26.0	22.1	33.1







#31
Naphthalene
Concen: 41.755 ng
RT: 11.098 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 295608

Ion	Ratio	Lower	Upper
128	100		
129	10.3	8.8	13.2
127	12.8	10.6	15.8







#32
Benzoic acid
Concen: 41.641 ng
RT: 10.299 min Scan# 1227
Delta R.T. -0.010 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:122 Resp: 58855

Ion	Ratio	Lower	Upper
122	100		
105	122.7	104.2	144.2
77	93.9	70.2	110.2







#33
4-Chloroaniline
Concen: 39.165 ng
RT: 11.186 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 107772

Ion	Ratio	Lower	Upper
127	100		
129	34.3	25.5	38.3
65	41.5	32.2	48.4
92	21.1	18.2	27.2







#34
Hexachlorobutadiene
Concen: 44.987 ng
RT: 11.386 min Scan# 1412
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:225 Resp: 81632

Ion	Ratio	Lower	Upper
225	100		
223	64.5	48.8	73.2
227	68.7	49.5	74.3







#35
Caprolactam
Concen: 34.554 ng
RT: 11.950 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 27064

Ion	Ratio	Lower	Upper
113	100		
55	273.8	290.1	330.1#
56	206.8	183.8	223.8







#36
4-Chloro-3-methylphenol
Concen: 40.557 ng
RT: 12.291 min Scan# 1566
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:107 Resp: 99554

Ion	Ratio	Lower	Upper
107	100		
144	24.3	19.1	28.7
142	76.5	58.6	87.8







#37
2-Methylnaphthalene
Concen: 41.379 ng
RT: 12.684 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 215078

Ion	Ratio	Lower	Upper
142	100		
141	84.4	68.9	103.3
115	35.3	30.6	46.0







#38
1-Methylnaphthalene
Concen: 40.524 ng
RT: 12.902 min Scan# 1670
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:142 Resp: 208518

Ion	Ratio	Lower	Upper
142	100		
141	89.2	73.3	109.9
116	4.0	3.1	4.7







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.841 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 94503

Ion	Ratio	Lower	Upper
164	100		
162	98.1	77.0	115.6
160	42.0	34.4	51.6







#40
1,2,4,5-Tetrachlorobenzene
Concen: 45.407 ng
RT: 13.049 min Scan# 1695
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:216 Resp: 144011

Ion	Ratio	Lower	Upper
216	100		
214	76.5	61.4	92.2
179	18.3	15.4	23.0
108	14.8	11.8	17.6







#41
Hexachlorocyclopentadiene
Concen: 54.673 ng
RT: 13.031 min Scan# 1
Delta R.T. -0.010 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 78048

Ion	Ratio	Lower	Upper
237	100		
235	60.2	44.1	84.1
272	11.6	0.0	32.4







#42
2,4,6-Tribromophenol
Concen: 87.917 ng
RT: 16.315 min Scan# 2251
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:330 Resp: 119915

Ion	Ratio	Lower	Upper
330	100		
332	96.2	78.0	117.0
141	29.0	21.8	32.6







#43
2,4,6-Trichlorophenol
Concen: 47.604 ng
RT: 13.278 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 86362

Ion	Ratio	Lower	Upper
196	100		
198	98.4	76.6	115.0
200	32.0	26.0	39.0







#44
2,4,5-Trichlorophenol
Concen: 44.690 ng
RT: 13.348 min Scan# 1746
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:196 Resp: 93561

Ion	Ratio	Lower	Upper
196	100		
198	95.5	74.2	111.4
97	45.8	37.0	55.6
132	21.7	18.5	27.7







#45
2-Fluorobiphenyl
Concen: 86.520 ng
RT: 13.472 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 539428

Ion	Ratio	Lower	Upper
172	100		
171	36.3	29.4	44.0
170	24.9	19.1	28.7







#46
1,1'-Biphenyl
Concen: 42.404 ng
RT: 13.677 min Scan# 1802
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:154 Resp: 284894

Ion	Ratio	Lower	Upper
154	100		
153	38.5	22.4	62.4
76	11.1	0.0	33.5







#47
2-Chloronaphthalene
Concen: 42.675 ng
RT: 13.724 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 216419

Ion	Ratio	Lower	Upper
162	100		
127	39.2	32.2	48.2
164	34.8	27.5	41.3







#48
2-Nitroaniline
Concen: 43.795 ng
RT: 13.906 min Scan# 1841
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion: 65 Resp: 74045

Ion	Ratio	Lower	Upper
65	100		
92	59.0	48.3	72.5
138	81.9	74.9	112.3







#49
Acenaphthylene
Concen: 41.484 ng
RT: 14.565 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 331834

Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.4	23.2
153	12.7	11.0	16.4







#50
Dimethylphthalate
Concen: 39.451 ng
RT: 14.283 min Scan# 1905
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:163 Resp: 285287

Ion	Ratio	Lower	Upper
163	100		
194	3.8	2.8	4.2
164	11.2	7.7	11.5







#51
2,6-Dinitrotoluene
Concen: 44.784 ng
RT: 14.394 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 53162

Ion	Ratio	Lower	Upper
165	100		
63	72.8	57.4	86.0
89	62.5	47.8	71.8







#52
Acenaphthene
Concen: 41.683 ng
RT: 14.905 min Scan# 2011
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:154 Resp: 227801

Ion	Ratio	Lower	Upper
154	100		
153	118.7	92.2	138.4
152	55.5	43.2	64.8







#53
3-Nitroaniline
Concen: 46.309 ng
RT: 14.723 min Scan# 1
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 57551

Ion	Ratio	Lower	Upper
138	100		
108	12.8	11.4	17.0
92	122.7	95.3	142.9







#54
2,4-Dinitrophenol
Concen: 48.950 ng
RT: 14.923 min Scan# 2014
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:184 Resp: 34138

Ion	Ratio	Lower	Upper
184	100		
63	77.8	56.6	85.0
154	171.2	158.6	238.0







#55
Dibenzofuran
Concen: 39.980 ng
RT: 15.234 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 353239

Ion	Ratio	Lower	Upper
168	100		
139	35.9	27.8	41.8
169	13.5	10.3	15.5







#56
4-Nitrophenol
Concen: 43.812 ng
RT: 15.011 min Scan# 2029
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:139 Resp: 50480

Ion	Ratio	Lower	Upper
139	100		
109	91.6	67.8	107.8
65	121.5	103.7	143.7







#57
2,4-Dinitrotoluene
Concen: 42.716 ng
RT: 15.176 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 78142

Ion	Ratio	Lower	Upper
165	100		
63	54.6	41.8	62.8
89	73.9	58.1	87.1
182	2.4	1.4	2.0#







#58
Fluorene
Concen: 39.487 ng
RT: 15.881 min Scan# 2177
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:166 Resp: 272466

Ion	Ratio	Lower	Upper
166	100		
165	100.3	81.5	122.3
167	13.6	11.4	17.0







#59
2,3,4,6-Tetrachlorophenol
Concen: 44.655 ng
RT: 15.452 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 89397

Ion	Ratio	Lower	Upper
232	100		
131	35.5	26.7	40.1
130	2.0	1.6	2.4
166	25.8	20.5	30.7







#60
Diethylphthalate
Concen: 37.738 ng
RT: 15.640 min Scan# 2136
Delta R.T. -0.010 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:149 Resp: 278640

Ion	Ratio	Lower	Upper
149	100		
177	22.3	18.0	27.0
150	12.0	10.2	15.4







#61
4-Chlorophenyl-phenylether
Concen: 40.558 ng
RT: 15.869 min Scan# 21
Delta R.T. -0.010 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 157055

Ion	Ratio	Lower	Upper
204	100		
206	32.1	25.6	38.4
141	54.2	40.6	60.8







#62
4-Nitroaniline
Concen: 43.493 ng
RT: 15.881 min Scan# 2177
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:138 Resp: 60132

Ion	Ratio	Lower	Upper
138	100		
92	55.2	36.1	76.1
108	105.1	75.6	115.6







#63
Azobenzene
Concen: 38.024 ng
RT: 16.163 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 257601

Ion	Ratio	Lower	Upper
77	100		
182	27.4	4.9	44.9
105	20.3	1.3	41.3
51	37.3	19.0	59.0







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.579 min Scan# 2466
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:188 Resp: 210658

Ion	Ratio	Lower	Upper
188	100		
94	6.6	5.7	8.5
80	7.7	6.8	10.2







#65
4,6-Dinitro-2-methylphenol
Concen: 52.445 ng
RT: 15.939 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 49140

Ion	Ratio	Lower	Upper
198	100		
51	48.8	24.7	64.7
105	37.2	18.0	58.0







#66
n-Nitrosodiphenylamine
Concen: 42.108 ng
RT: 16.080 min Scan# 2211
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:169 Resp: 239610

Ion	Ratio	Lower	Upper
169	100		
168	68.2	56.3	84.5
167	37.2	29.0	43.6







#67
4-Bromophenyl-phenylether
Concen: 44.024 ng
RT: 16.762 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 109900

Ion	Ratio	Lower	Upper
248	100		
250	98.3	77.7	116.5
141	61.2	48.6	72.8







#68
Hexachlorobenzene
Concen: 43.121 ng
RT: 16.885 min Scan# 2348
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:284 Resp: 122806

Ion	Ratio	Lower	Upper
284	100		
142	27.7	21.0	31.6
249	34.3	24.9	37.3







#69
Atrazine
Concen: 39.893 ng
RT: 17.015 min Scan# 2
Delta R.T. -0.010 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 97691

Ion	Ratio	Lower	Upper
200	100		
173	23.8	0.9	40.9
215	46.3	27.0	67.0







#70
Pentachlorophenol
Concen: 46.367 ng
RT: 17.220 min Scan# 2405
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:266 Resp: 82346

Ion	Ratio	Lower	Upper
266	100		
268	67.0	50.3	75.5
264	67.1	47.3	70.9







#71
Phenanthrene
Concen: 40.749 ng
RT: 17.620 min Scan# 24
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 469425

Ion	Ratio	Lower	Upper
178	100		
176	19.1	14.7	22.1
179	15.7	12.2	18.4







#72
Anthracene
Concen: 41.136 ng
RT: 17.708 min Scan# 2488
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:178 Resp: 482051

Ion	Ratio	Lower	Upper
178	100		
176	17.7	14.4	21.6
179	15.1	12.4	18.6







#73
Carbazole
Concen: 40.999 ng
RT: 17.972 min Scan# 2
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 423668

Ion	Ratio	Lower	Upper
167	100		
166	22.7	17.9	26.9
139	11.8	9.7	14.5







#74
Di-n-butylphthalate
Concen: 39.860 ng
RT: 18.525 min Scan# 2627
Delta R.T. -0.010 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:149 Resp: 476308

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.0	10.6
104	5.3	4.1	6.1







#75
Fluoranthene
Concen: 42.360 ng
RT: 19.612 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 611318

Ion	Ratio	Lower	Upper
202	100		
101	7.9	0.0	28.7
203	17.0	0.0	37.0







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.856 min Scan# 3194
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:240 Resp: 257813

Ion	Ratio	Lower	Upper
240	100		
120	6.6	6.5	9.7
236	26.0	20.2	30.2







#77
Benzidine
Concen: 41.721 ng
RT: 19.782 min Scan# 2
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 240984

Ion	Ratio	Lower	Upper
184	100		
185	13.5	11.4	17.2
183	12.2	9.4	14.2







#78
Pyrene
Concen: 38.182 ng
RT: 19.976 min Scan# 2874
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:202 Resp: 643136

Ion	Ratio	Lower	Upper
202	100		
200	21.0	17.0	25.4
203	17.2	13.8	20.6







#79
Terphenyl-d14
Concen: 78.718 ng
RT: 20.170 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 1075391

Ion	Ratio	Lower	Upper
244	100		
212	6.9	6.1	9.1
122	6.1	5.6	8.4







#80
Butylbenzylphthalate
Concen: 45.114 ng
RT: 20.851 min Scan# 3023
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:149 Resp: 229670

Ion	Ratio	Lower	Upper
149	100		
91	70.1	59.2	88.8
206	23.2	18.1	27.1







#81
Benzo(a)anthracene
Concen: 42.521 ng
RT: 21.832 min Scan# 3
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 726865

Ion	Ratio	Lower	Upper
228	100		
226	26.1	21.9	32.9
229	18.6	15.6	23.4







#82
3,3'-Dichlorobenzidine
Concen: 44.521 ng
RT: 21.738 min Scan# 3174
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:252 Resp: 254141

Ion	Ratio	Lower	Upper
252	100		
254	63.1	50.8	76.2
126	7.2	5.8	8.6







#83
Chrysene
Concen: 41.987 ng
RT: 21.903 min Scan# 3
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 682939

Ion	Ratio	Lower	Upper
228	100		
226	29.9	23.3	34.9
229	19.8	15.2	22.8







#84
Bis(2-ethylhexyl)phthalate
Concen: 44.177 ng
RT: 21.738 min Scan# 3174
Delta R.T. -0.010 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:149 Resp: 336059

Ion	Ratio	Lower	Upper
149	100		
167	25.4	20.0	30.0
279	4.0	3.1	4.7







#85
Di-n-octyl phthalate
Concen: 47.074 ng
RT: 22.996 min Scan# 3
Delta R.T. -0.016 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 603742

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	12.6	10.8	16.2







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.199 min Scan# 3763
Delta R.T. 0.002 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:264 Resp: 295961

Ion	Ratio	Lower	Upper
264	100		
260	22.7	19.0	28.4
265	20.5	17.6	26.4







#87
Indeno(1,2,3-cd)pyrene
Concen: 43.313 ng
RT: 29.036 min Scan# 4
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 943017

Ion	Ratio	Lower	Upper
276	100		
138	11.3	10.8	16.2
277	25.0	20.1	30.1







#88
Benzo(b)fluoranthene
Concen: 41.496 ng
RT: 24.136 min Scan# 3582
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:252 Resp: 753080

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.8	26.8
125	6.3	5.6	8.4







#89
Benzo(k)fluoranthene
Concen: 41.889 ng
RT: 24.206 min Scan# 3
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 760727

Ion	Ratio	Lower	Upper
252	100		
253	22.5	16.8	25.2
125	6.6	5.4	8.0







#90
Benzo(a)pyrene
Concen: 42.211 ng
RT: 25.041 min Scan# 3736
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:252 Resp: 704164

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.6	26.4
125	7.3	6.1	9.1







#91
Dibenzo(a,h)anthracene
Concen: 42.934 ng
RT: 29.106 min Scan# 44
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 759427

Ion	Ratio	Lower	Upper
278	100		
139	10.3	8.0	12.0
279	23.3	19.8	29.6






#92
Benzo(g,h,i)perylene
Concen: 43.547 ng
RT: 30.229 min Scan# 4619
Delta R.T. -0.004 min
Lab File: BG064636.D
Acq: 30 Oct 2025 11:36

Tgt Ion:276 Resp: 735814

Ion	Ratio	Lower	Upper
276	100		
277	23.1	19.1	28.7
138	12.2	11.3	16.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2		1,4-Dioxane	0.528	0.546	-3.4	84	0.00
3		Pyridine	1.584	1.626	-2.7	79	0.00
4		n-Nitrosodimethylamine	0.850	0.877	-3.2	79	0.00
5	S	2-Fluorophenol	1.192	1.233	-3.4	77	0.00
6		Aniline	2.225	2.102	5.5	71	0.00
7	S	Phenol-d6	1.635	1.603	2.0	73	0.00
8		2-Chlorophenol	1.247	1.369	-9.8	81	0.00
9		Benzaldehyde	0.893	0.861	3.6	74	0.00
10	C	Phenol	1.808	1.761	2.6	73	0.00
11		bis(2-Chloroethyl)ether	1.329	1.296	2.5	76	0.00
12		1,3-Dichlorobenzene	1.435	1.559	-8.6	83	0.00
13	C	1,4-Dichlorobenzene	1.460	1.552	-6.3	82	0.00
14		1,2-Dichlorobenzene	1.408	1.512	-7.4	82	0.00
15		Benzyl Alcohol	1.348	1.390	-3.1	79	0.00
16		2,2'-oxybis(1-Chloropropane	3.487	3.562	-2.2	77	0.00
17		2-Methylphenol	1.201	1.221	-1.7	78	0.00
18		Hexachloroethane	0.512	0.586	-14.5	88	0.00
19	P	n-Nitroso-di-n-propylamine	1.130	1.116	1.2	74	0.00
20		3+4-Methylphenols	1.688	1.620	4.0	72	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
22		Acetophenone	0.547	0.558	-2.0	74	0.00
23	S	Nitrobenzene-d5	0.333	0.420	-26.1#	86	0.00
24		Nitrobenzene	0.362	0.438	-21.0	83	0.00
25		Isophorone	0.799	0.808	-1.1	72	0.00
26	C	2-Nitrophenol	0.121	0.180	-48.8#	102	0.00
27		2,4-Dimethylphenol	0.297	0.320	-7.7	74	0.00
28		bis(2-Chloroethoxy)methane	0.455	0.460	-1.1	73	0.00
29	C	2,4-Dichlorophenol	0.320	0.361	-12.8	78	0.00
30		1,2,4-Trichlorobenzene	0.363	0.398	-9.6	79	0.00
31		Naphthalene	1.109	1.158	-4.4	76	0.00
32		Benzoic acid	0.207	0.230	-11.1	80	-0.01
33		4-Chloroaniline	0.431	0.422	2.1	71	0.00
34	C	Hexachlorobutadiene	0.284	0.320	-12.7	83	0.00
35		Caprolactam	0.123	0.106	13.8	61	0.00
36	C	4-Chloro-3-methylphenol	0.384	0.390	-1.6	71	0.00
37		2-Methylnaphthalene	0.814	0.842	-3.4	75	0.00
38		1-Methylnaphthalene	0.806	0.816	-1.2	74	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
40		1,2,4,5-Tetrachlorobenzene	0.671	0.762	-13.6	79	0.00
41	P	Hexachlorocyclopentadiene	0.302	0.413	-36.8#	92	0.00
42	S	2,4,6-Tribromophenol	0.289	0.317	-9.7	72	0.00
43	C	2,4,6-Trichlorophenol	0.384	0.457	-19.0	78	0.00
44		2,4,5-Trichlorophenol	0.443	0.495	-11.7	74	0.00
45	S	2-Fluorobiphenyl	1.319	1.427	-8.2	75	0.00
46		1,1'-Biphenyl	1.422	1.507	-6.0	73	0.00
47		2-Chloronaphthalene	1.073	1.145	-6.7	73	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.311	0.392	-26.0#	81	0.00
49		Acenaphthylene	1.693	1.756	-3.7	71	0.00
50		Dimethylphthalate	1.530	1.509	1.4	67	0.00
51		2,6-Dinitrotoluene	0.218	0.281	-28.9#	82	0.00
52	C	Acenaphthene	1.157	1.205	-4.1	71	0.00
53		3-Nitroaniline	0.263	0.304	-15.6	75	0.00
54	P	2,4-Dinitrophenol	0.132	0.181	-37.1#	96	0.00
55		Dibenzofuran	1.870	1.869	0.1	69	0.00
56	P	4-Nitrophenol	0.244	0.267	-9.4	74	0.00
57		2,4-Dinitrotoluene	0.340	0.413	-21.5	78	0.00
58		Fluorene	1.460	1.442	1.2	68	0.00
59		2,3,4,6-Tetrachlorophenol	0.424	0.473	-11.6	73	0.00
60		Diethylphthalate	1.563	1.474	5.7	64	0.00
61		4-Chlorophenyl-phenylether	0.820	0.831	-1.3	69	0.00
62		4-Nitroaniline	0.293	0.318	-8.5	70	0.00
63		Azobenzene	1.434	1.363	5.0	64	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
65		4,6-Dinitro-2-methylphenol	0.079	0.117	-48.1#	95	0.00
66	c	n-Nitrosodiphenylamine	0.540	0.569	-5.4	66	0.00
67		4-Bromophenyl-phenylether	0.237	0.261	-10.1	68	0.00
68		Hexachlorobenzene	0.270	0.291	-7.8	67	0.00
69		Atrazine	0.232	0.232	0.0	64	0.00
70	C	Pentachlorophenol	0.169	0.195	-15.4	73	0.00
71		Phenanthrene	1.094	1.114	-1.8	64	0.00
72		Anthracene	1.113	1.144	-2.8	65	0.00
73		Carbazole	0.981	1.006	-2.5	66	0.00
74		Di-n-butylphthalate	1.134	1.131	0.3	61	0.00
75	C	Fluoranthene	1.370	1.451	-5.9	68	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	77	0.00
77		Benzidine	0.448	0.467	-4.2	74	0.00
78		Pyrene	1.307	1.247	4.6	68	0.00
79	S	Terphenyl-d14	1.060	1.043	1.6	71	0.00
80		Butylbenzylphthalate	0.395	0.445	-12.7	77	0.00
81		Benzo(a)anthracene	1.326	1.410	-6.3	77	0.00
82		3,3'-Dichlorobenzidine	0.443	0.493	-11.3	80	0.00
83		Chrysene	1.262	1.324	-4.9	76	0.00
84		Bis(2-ethylhexyl)phthalate	0.590	0.652	-10.5	75	0.00
85	c	Di-n-octyl phthalate	0.995	1.171	-17.7	83	-0.02
86	I	Perylene-d12	1.000	1.000	0.0	84	0.00
87		Indeno(1,2,3-cd)pyrene	1.471	1.593	-8.3	86	0.00
88		Benzo(b)fluoranthene	1.226	1.272	-3.8	81	0.00
89		Benzo(k)fluoranthene	1.227	1.285	-4.7	84	0.00
90	C	Benzo(a)pyrene	1.127	1.190	-5.6	84	0.00
91		Dibenzo(a,h)anthracene	1.195	1.283	-7.4	85	0.00
92		Benzo(g,h,i)perylene	1.142	1.243	-8.8	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	81	0.00
2		1,4-Dioxane	40.000	41.400	-3.5	84	0.00
3		Pyridine	40.000	41.054	-2.6	79	0.00
4		n-Nitrosodimethylamine	40.000	41.282	-3.2	79	0.00
5	S	2-Fluorophenol	80.000	82.805	-3.5	77	0.00
6		Aniline	40.000	37.797	5.5	71	0.00
7	S	Phenol-d6	80.000	78.424	2.0	73	0.00
8		2-Chlorophenol	40.000	43.910	-9.8	81	0.00
9		Benzaldehyde	40.000	38.596	3.5	74	0.00
10	C	Phenol	40.000	38.953	2.6	73	0.00
11		bis(2-Chloroethyl)ether	40.000	39.002	2.5	76	0.00
12		1,3-Dichlorobenzene	40.000	43.441	-8.6	83	0.00
13	C	1,4-Dichlorobenzene	40.000	42.528	-6.3	82	0.00
14		1,2-Dichlorobenzene	40.000	42.953	-7.4	82	0.00
15		Benzyl Alcohol	40.000	41.237	-3.1	79	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	40.856	-2.1	77	0.00
17		2-Methylphenol	40.000	40.671	-1.7	78	0.00
18		Hexachloroethane	40.000	45.739	-14.3	88	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	39.522	1.2	74	0.00
20		3+4-Methylphenols	40.000	38.381	4.0	72	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	77	0.00
22		Acetophenone	40.000	40.856	-2.1	74	0.00
23	S	Nitrobenzene-d5	80.000	100.795	-26.0#	86	0.00
24		Nitrobenzene	40.000	48.375	-20.9	83	0.00
25		Isophorone	40.000	40.473	-1.2	72	0.00
26	C	2-Nitrophenol	40.000	51.379	-28.4#	102	0.00
27		2,4-Dimethylphenol	40.000	43.090	-7.7	74	0.00
28		bis(2-Chloroethoxy)methane	40.000	40.406	-1.0	73	0.00
29	C	2,4-Dichlorophenol	40.000	45.182	-13.0	78	0.00
30		1,2,4-Trichlorobenzene	40.000	43.767	-9.4	79	0.00
31		Naphthalene	40.000	41.755	-4.4	76	0.00
32		Benzoic acid	40.000	41.641	-4.1	80	-0.01
33		4-Chloroaniline	40.000	39.165	2.1	71	0.00
34	C	Hexachlorobutadiene	40.000	44.987	-12.5	83	0.00
35		Caprolactam	40.000	34.554	13.6	61	0.00
36	C	4-Chloro-3-methylphenol	40.000	40.557	-1.4	71	0.00
37		2-Methylnaphthalene	40.000	41.379	-3.4	75	0.00
38		1-Methylnaphthalene	40.000	40.524	-1.3	74	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	72	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	45.407	-13.5	79	0.00
41	P	Hexachlorocyclopentadiene	40.000	54.673	-36.7#	92	0.00
42	S	2,4,6-Tribromophenol	80.000	87.917	-9.9	72	0.00
43	C	2,4,6-Trichlorophenol	40.000	47.604	-19.0	78	0.00
44		2,4,5-Trichlorophenol	40.000	44.690	-11.7	74	0.00
45	S	2-Fluorobiphenyl	80.000	86.520	-8.1	75	0.00
46		1,1'-Biphenyl	40.000	42.404	-6.0	73	0.00
47		2-Chloronaphthalene	40.000	42.675	-6.7	73	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	43.795	-9.5	81	0.00
49		Acenaphthylene	40.000	41.484	-3.7	71	0.00
50		Dimethylphthalate	40.000	39.451	1.4	67	0.00
51		2,6-Dinitrotoluene	40.000	44.784	-12.0	82	0.00
52	C	Acenaphthene	40.000	41.683	-4.2	71	0.00
53		3-Nitroaniline	40.000	46.309	-15.8	75	0.00
54	P	2,4-Dinitrophenol	40.000	48.950	-22.4	96	0.00
55		Dibenzofuran	40.000	39.980	0.1	69	0.00
56	P	4-Nitrophenol	40.000	43.812	-9.5	74	0.00
57		2,4-Dinitrotoluene	40.000	42.716	-6.8	78	0.00
58		Fluorene	40.000	39.487	1.3	68	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	44.655	-11.6	73	0.00
60		Diethylphthalate	40.000	37.738	5.7	64	0.00
61		4-Chlorophenyl-phenylether	40.000	40.558	-1.4	69	0.00
62		4-Nitroaniline	40.000	43.493	-8.7	70	0.00
63		Azobenzene	40.000	38.024	4.9	64	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	67	0.00
65		4,6-Dinitro-2-methylphenol	40.000	52.445	-31.1#	95	0.00
66	c	n-Nitrosodiphenylamine	40.000	42.108	-5.3	66	0.00
67		4-Bromophenyl-phenylether	40.000	44.024	-10.1	68	0.00
68		Hexachlorobenzene	40.000	43.121	-7.8	67	0.00
69		Atrazine	40.000	39.893	0.3	64	0.00
70	C	Pentachlorophenol	40.000	46.367	-15.9	73	0.00
71		Phenanthrene	40.000	40.749	-1.9	64	0.00
72		Anthracene	40.000	41.136	-2.8	65	0.00
73		Carbazole	40.000	40.999	-2.5	66	0.00
74		Di-n-butylphthalate	40.000	39.860	0.4	61	0.00
75	C	Fluoranthene	40.000	42.360	-5.9	68	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	77	0.00
77		Benzidine	40.000	41.721	-4.3	74	0.00
78		Pyrene	40.000	38.182	4.5	68	0.00
79	S	Terphenyl-d14	80.000	78.718	1.6	71	0.00
80		Butylbenzylphthalate	40.000	45.114	-12.8	77	0.00
81		Benzo(a)anthracene	40.000	42.521	-6.3	77	0.00
82		3,3'-Dichlorobenzidine	40.000	44.521	-11.3	80	0.00
83		Chrysene	40.000	41.987	-5.0	76	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	44.177	-10.4	75	0.00
85	c	Di-n-octyl phthalate	40.000	47.074	-17.7	83	-0.02
86	I	Perylene-d12	20.000	20.000	0.0	84	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	43.313	-8.3	86	0.00
88		Benzo(b)fluoranthene	40.000	41.496	-3.7	81	0.00
89		Benzo(k)fluoranthene	40.000	41.889	-4.7	84	0.00
90	C	Benzo(a)pyrene	40.000	42.211	-5.5	84	0.00
91		Dibenzo(a,h)anthracene	40.000	42.934	-7.3	85	0.00
92		Benzo(g,h,i)perylene	40.000	43.547	-8.9	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064636.D
Acq On : 30 Oct 2025 11:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Nov 04 04:21:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 1


QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144055.D
Acq On : 24 Oct 2025 09:45
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 24 17:30:25 2025




AutoFind: Scans 1609, 1610, 1611; Background Corrected with Scan 1603


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	1.6	333	PASS
69	69	100	100	100.0	20363	PASS
70	69	0.00	2	0.0	0	PASS
197	198	0.00	2	0.2	133	PASS
198	198	100	100	100.0	67712	PASS
199	198	5	9	6.1	4142	PASS
365	198	1	100	4.5	3052	PASS
441	443	0.01	150	89.0	14677	PASS
442	442	100	100	100.0	87547	PASS
443	442	15	24	18.8	16486	PASS


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144055.D
Acq On : 24 Oct 2025 09:45
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 24 10:43:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 06 15:29:47 2025
Response via : Initial Calibration







TIC: BF144055.D\data.ms

(70) Pentachlorophenol (C)

11.216min (-0.006) 43857.19 ng


response 76701


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.70	64.29
264.00	64.70	62.39
0.00	0.00	0.00


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102425\
Data File : BF144055.D
Acq On : 24 Oct 2025 09:45
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 24 10:43:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 06 15:29:47 2025
Response via : Initial Calibration







TIC: BF144055.D\data.ms

(77) Benzidine

12.751min (-0.006) 0.00 ng

response 416414

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.70	14.23
183.00	11.00	11.07
0.00	0.00	0.00

Instrument :
BNA_F
ClientSampleId :
DFTPP

DDT Breakdown


Date	Instrument Name	DFTPP Data File
10/24/2025	BNA_F	BF144055.D
Compound Name	**Response**	**Retention Time**
DDT	192449	13.574
DDD	3266	13.274
DDE	1346	12.939
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
4612	197061	2.34

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144096.D
Acq On : 28 Oct 2025 12:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 24 17:30:25 2025



AutoFind: Scans 1608, 1609, 1610; Background Corrected with Scan 1602


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	1.7	551	PASS
69	69	100	100	100.0	33349	PASS
70	69	0.00	2	0.0	0	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	102923	PASS
199	198	5	9	6.6	6806	PASS
365	198	1	100	4.1	4261	PASS
441	443	0.01	150	88.4	22452	PASS
442	442	100	100	100.0	137480	PASS
443	442	15	24	18.5	25399	PASS


Sample Results: **BF144096.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144096.D
Acq On : 28 Oct 2025 12:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 28 14:08:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







TIC: BF144096.D\data.ms

(70) Pentachlorophenol (C)

11.216min (+ 0.000) 40706.49 ng m


response 152337


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	60.30	61.51
264.00	61.70	60.62
0.00	0.00	0.00


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102825\
Data File : BF144096.D
Acq On : 28 Oct 2025 12:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 28 14:08:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







TIC: BF144096.D\data.ms

(77) Benzidine

12.751min (+ 0.000) 31665.53 ng

response 613959

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.04
183.00	10.60	10.96
0.00	0.00	0.00

Instrument :
BNA_F
ClientSampleId :
DFTPP

DDT Breakdown


Date	Instrument Name	DFTPP Data File
10/28/2025	BNA_F	BF144096.D
Compound Name	**Response**	**Retention Time**
DDT	289356	13.569
DDD	3978	13.274
DDE	557	12.933
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
4535	293891	1.54

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144120.D
Acq On : 29 Oct 2025 11:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 24 17:30:25 2025




AutoFind: Scans 1608, 1609, 1610; Background Corrected with Scan 1602


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	1.3	329	PASS
69	69	100	100	100.0	24539	PASS
70	69	0.00	2	0.0	0	PASS
197	198	0.00	2	0.4	309	PASS
198	198	100	100	100.0	78523	PASS
199	198	5	9	6.9	5389	PASS
365	198	1	100	4.2	3266	PASS
441	443	0.01	150	84.9	15712	PASS
442	442	100	100	100.0	102557	PASS
443	442	15	24	18.0	18498	PASS


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144120.D
Acq On : 29 Oct 2025 11:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 29 12:26:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







TIC: BF144120.D\data.ms

(70) Pentachlorophenol (C)

11.210min (-0.006) 40226.81 ng


response 108496


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	60.30	59.00
264.00	61.70	59.22
0.00	0.00	0.00


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102925\
Data File : BF144120.D
Acq On : 29 Oct 2025 11:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 29 12:26:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 17:30:25 2025
Response via : Initial Calibration







TIC: BF144120.D\data.ms

(77) Benzidine

12.745min (-0.006) 267936.77 ng

response 458627

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	13.88
183.00	10.60	10.97
0.00	0.00	0.00

Instrument :
BNA_F
ClientSampleId :
DFTPP

DDT Breakdown


Date	Instrument Name	DFTPP Data File
10/29/2025	BNA_F	BF144120.D
Compound Name	**Response**	**Retention Time**
DDT	197272	13.569
DDD	5071	13.269
DDE	134	12.933
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
5205	202477	2.57

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064568.D
Acq On : 24 Oct 2025 8:20
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 24 16:40:26 2025




AutoFind: Scans 2539, 2540, 2541; Background Corrected with Scan 2533


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	0.0	0	PASS
69	69	100	100	100.0	35203	PASS
70	69	0.00	2	0.2	70	PASS
197	198	0.00	2	0.4	428	PASS
198	198	100	100	100.0	99864	PASS
199	198	5	9	7.2	7222	PASS
365	198	1	100	5.1	5094	PASS
441	443	0.01	150	86.3	15073	PASS
442	442	100	100	100.0	92525	PASS
443	442	15	24	18.9	17468	PASS


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064568.D
Acq On : 24 Oct 2025 8:20
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 24 12:18:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 12:16:12 2025
Response via : Initial Calibration







TIC: BG064568.D\data.ms

(70) Pentachlorophenol (C)

17.226min (+ 0.001) 15878.32 ng


response 79720


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	69.10	64.11
264.00	61.90	64.80
0.00	0.00	0.00


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102425\
Data File : BG064568.D
Acq On : 24 Oct 2025 8:20
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 24 12:18:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 12:16:12 2025
Response via : Initial Calibration







TIC: BG064568.D\data.ms

(77) Benzidine

19.787min (+ 0.001) 0.00 ng

response 632825

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	13.68
183.00	12.10	11.84
0.00	0.00	0.00

Instrument :
BNA_G
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/24/2025	BNA_G	BG064568.D
Compound Name	**Response**	**Retention Time**
DDT	249894	21.033
DDD	4771	20.586
DDE	827	20.087
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
5598	255492	2.19


Sample Results: **BG064617.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064617.D
Acq On : 29 Oct 2025 10:41
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 24 16:40:26 2025




AutoFind: Scans 2538, 2539, 2540; Background Corrected with Scan 2532


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	1.0	229	PASS
69	69	100	100	100.0	21821	PASS
70	69	0.00	2	0.6	135	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	68565	PASS
199	198	5	9	7.3	4984	PASS
365	198	1	100	5.3	3612	PASS
441	443	0.01	150	91.7	11008	PASS
442	442	100	100	100.0	66685	PASS
443	442	15	24	18.0	12010	PASS


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064617.D
Acq On : 29 Oct 2025 10:41
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 29 11:21:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration







TIC: BG064617.D\data.ms

(70) Pentachlorophenol (C)

17.219min (-0.009) 35052.04 ng


response 97221


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.90	65.04
264.00	59.10	69.72
0.00	0.00	0.00


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064617.D
Acq On : 29 Oct 2025 10:41
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 29 11:21:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration







TIC: BG064617.D\data.ms

(77) Benzidine

19.780min (-0.003) 345166.57 ng

response 533588

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.30	13.80
183.00	11.80	12.33
0.00	0.00	0.00

Instrument :
BNA_G
ClientSampleId :
DFTPP

DDT Breakdown


Date	Instrument Name	DFTPP Data File
10/29/2025	BNA_G	BG064617.D
Compound Name	**Response**	**Retention Time**
DDT	244663	21.026
DDD	3569	20.573
DDE	0	19.706
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
3569	248232	1.44

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064635.D
Acq On : 30 Oct 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 24 16:40:26 2025



AutoFind: Scans 2537, 2538, 2539; Background Corrected with Scan 2531


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	0.7	175	PASS
69	69	100	100	100.0	25608	PASS
70	69	0.00	2	0.0	0	PASS
197	198	0.00	2	0.5	368	PASS
198	198	100	100	100.0	74355	PASS
199	198	5	9	7.3	5401	PASS
365	198	1	100	5.1	3813	PASS
441	443	0.01	150	80.3	11250	PASS
442	442	100	100	100.0	70075	PASS
443	442	15	24	20.0	14011	PASS


Sample Results: **BG064635.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064635.D
Acq On : 30 Oct 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 30 11:34:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration







TIC: BG064635.D\data.ms

(70) Pentachlorophenol (C)

17.221min (-0.007) 25561.00 ng


response 128648


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.90	68.13
264.00	59.10	64.51
0.00	0.00	0.00


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064635.D
Acq On : 30 Oct 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 30 11:34:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration







TIC: BG064635.D\data.ms

(77) Benzidine

19.783min (-0.001) 0.00 ng

response 690984

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.30	14.22
183.00	11.80	11.74
0.00	0.00	0.00

Instrument :
BNA_G
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/30/2025	BNA_G	BG064635.D
Compound Name	**Response**	**Retention Time**
DDT	367195	21.023
DDD	3406	20.576
DDE	398	20.077
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
3804	370999	1.03

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170215BL	SDG No.:	Q3434
Lab Sample ID:	PB170215BL	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	100
Sample Wt/Vol:	30.01 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	330	U	1	160	330	ug/Kg	10/29/25 13:25	PB170215
108-95-2	Phenol	170	U	1	22.1	170	ug/Kg	10/29/25 13:25	PB170215
111-44-4	bis(2-Chloroethyl)ether	170	U	1	24.3	170	ug/Kg	10/29/25 13:25	PB170215
95-57-8	2-Chlorophenol	170	U	1	24.4	170	ug/Kg	10/29/25 13:25	PB170215
95-48-7	2-Methylphenol	170	U	1	29.9	170	ug/Kg	10/29/25 13:25	PB170215
108-60-1	2,2-oxybis(1-Chloropropane)	170	U	1	37.5	170	ug/Kg	10/29/25 13:25	PB170215
98-86-2	Acetophenone	170	U	1	29.5	170	ug/Kg	10/29/25 13:25	PB170215
65794-96-9	3+4-Methylphenols	330	U	1	41.1	330	ug/Kg	10/29/25 13:25	PB170215
621-64-7	n-Nitroso-di-n-propylamine	80.0	U	1	47.4	80.0	ug/Kg	10/29/25 13:25	PB170215
67-72-1	Hexachloroethane	170	U	1	17.6	170	ug/Kg	10/29/25 13:25	PB170215
98-95-3	Nitrobenzene	170	U	1	18.3	170	ug/Kg	10/29/25 13:25	PB170215
78-59-1	Isophorone	170	U	1	32.8	170	ug/Kg	10/29/25 13:25	PB170215
88-75-5	2-Nitrophenol	170	U	1	58.2	170	ug/Kg	10/29/25 13:25	PB170215
105-67-9	2,4-Dimethylphenol	170	U	1	64.8	170	ug/Kg	10/29/25 13:25	PB170215
111-91-1	bis(2-Chloroethoxy)methane	170	U	1	30.8	170	ug/Kg	10/29/25 13:25	PB170215
120-83-2	2,4-Dichlorophenol	170	U	1	28.3	170	ug/Kg	10/29/25 13:25	PB170215
91-20-3	Naphthalene	170	U	1	22.7	170	ug/Kg	10/29/25 13:25	PB170215
106-47-8	4-Chloroaniline	170	U	1	35.4	170	ug/Kg	10/29/25 13:25	PB170215
87-68-3	Hexachlorobutadiene	170	U	1	25.3	170	ug/Kg	10/29/25 13:25	PB170215
105-60-2	Caprolactam	330	U	1	52.1	330	ug/Kg	10/29/25 13:25	PB170215
59-50-7	4-Chloro-3-methylphenol	170	U	1	28.7	170	ug/Kg	10/29/25 13:25	PB170215
91-57-6	2-Methylnaphthalene	170	U	1	25.6	170	ug/Kg	10/29/25 13:25	PB170215
77-47-4	Hexachlorocyclopentadiene	330	U	1	120	330	ug/Kg	10/29/25 13:25	PB170215
88-06-2	2,4,6-Trichlorophenol	170	U	1	19.8	170	ug/Kg	10/29/25 13:25	PB170215
95-95-4	2,4,5-Trichlorophenol	170	U	1	29.1	170	ug/Kg	10/29/25 13:25	PB170215
92-52-4	1,1-Biphenyl	170	U	1	21.8	170	ug/Kg	10/29/25 13:25	PB170215
91-58-7	2-Chloronaphthalene	170	U	1	22.5	170	ug/Kg	10/29/25 13:25	PB170215
88-74-4	2-Nitroaniline	170	U	1	48.1	170	ug/Kg	10/29/25 13:25	PB170215
131-11-3	Dimethylphthalate	170	U	1	27.1	170	ug/Kg	10/29/25 13:25	PB170215
208-96-8	Acenaphthylene	170	U	1	28.9	170	ug/Kg	10/29/25 13:25	PB170215
606-20-2	2,6-Dinitrotoluene	170	U	1	33.6	170	ug/Kg	10/29/25 13:25	PB170215
99-09-2	3-Nitroaniline	170	U	1	46.0	170	ug/Kg	10/29/25 13:25	PB170215
83-32-9	Acenaphthene	170	U	1	21.3	170	ug/Kg	10/29/25 13:25	PB170215
51-28-5	2,4-Dinitrophenol	330	U	1	230	330	ug/Kg	10/29/25 13:25	PB170215
100-02-7	4-Nitrophenol	330	U	1	110	330	ug/Kg	10/29/25 13:25	PB170215
132-64-9	Dibenzofuran	170	U	1	22.7	170	ug/Kg	10/29/25 13:25	PB170215
121-14-2	2,4-Dinitrotoluene	170	U	1	50.1	170	ug/Kg	10/29/25 13:25	PB170215
84-66-2	Diethylphthalate	170	U	1	28.3	170	ug/Kg	10/29/25 13:25	PB170215
7005-72-3	4-Chlorophenyl-phenylether	170	U	1	26.7	170	ug/Kg	10/29/25 13:25	PB170215
86-73-7	Fluorene	170	U	1	25.3	170	ug/Kg	10/29/25 13:25	PB170215
100-01-6	4-Nitroaniline	170	U	1	64.2	170	ug/Kg	10/29/25 13:25	PB170215
534-52-1	4,6-Dinitro-2-methylphenol	330	U	1	100	330	ug/Kg	10/29/25 13:25	PB170215

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM			Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206			Date Received:	
Client Sample ID:	PB170215BL			SDG No.:	Q3434
Lab Sample ID:	PB170215BL			Matrix:	SOIL
Analytical Method:	8270E	Level:	LOW	% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol:	1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541	Prep Date:	10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	170	U	1	32.9	170	ug/Kg	10/29/25 13:25	PB170215
101-55-3	4-Bromophenyl-phenylether	170	U	1	27.8	170	ug/Kg	10/29/25 13:25	PB170215
118-74-1	Hexachlorobenzene	170	U	1	25.3	170	ug/Kg	10/29/25 13:25	PB170215
1912-24-9	Atrazine	170	U	1	34.0	170	ug/Kg	10/29/25 13:25	PB170215
87-86-5	Pentachlorophenol	330	U	1	51.3	330	ug/Kg	10/29/25 13:25	PB170215
85-01-8	Phenanthrene	170	U	1	20.9	170	ug/Kg	10/29/25 13:25	PB170215
120-12-7	Anthracene	170	U	1	33.3	170	ug/Kg	10/29/25 13:25	PB170215
86-74-8	Carbazole	170	U	1	31.2	170	ug/Kg	10/29/25 13:25	PB170215
84-74-2	Di-n-butylphthalate	170	U	1	47.9	170	ug/Kg	10/29/25 13:25	PB170215
206-44-0	Fluoranthene	170	U	1	30.0	170	ug/Kg	10/29/25 13:25	PB170215
129-00-0	Pyrene	170	U	1	36.0	170	ug/Kg	10/29/25 13:25	PB170215
85-68-7	Butylbenzylphthalate	170	U	1	71.4	170	ug/Kg	10/29/25 13:25	PB170215
91-94-1	3,3-Dichlorobenzidine	330	U	1	36.7	330	ug/Kg	10/29/25 13:25	PB170215
56-55-3	Benzo(a)anthracene	170	U	1	23.0	170	ug/Kg	10/29/25 13:25	PB170215
218-01-9	Chrysene	170	U	1	19.9	170	ug/Kg	10/29/25 13:25	PB170215
117-81-7	Bis(2-ethylhexyl)phthalate	170	U	1	59.2	170	ug/Kg	10/29/25 13:25	PB170215
117-84-0	Di-n-octyl phthalate	330	U	1	86.8	330	ug/Kg	10/29/25 13:25	PB170215
205-99-2	Benzo(b)fluoranthene	170	U	1	19.0	170	ug/Kg	10/29/25 13:25	PB170215
207-08-9	Benzo(k)fluoranthene	170	U	1	22.4	170	ug/Kg	10/29/25 13:25	PB170215
50-32-8	Benzo(a)pyrene	170	U	1	29.5	170	ug/Kg	10/29/25 13:25	PB170215
193-39-5	Indeno(1,2,3-cd)pyrene	170	U	1	29.1	170	ug/Kg	10/29/25 13:25	PB170215
53-70-3	Dibenzo(a,h)anthracene	170	U	1	27.4	170	ug/Kg	10/29/25 13:25	PB170215
191-24-2	Benzo(g,h,i)perylene	170	U	1	25.7	170	ug/Kg	10/29/25 13:25	PB170215
95-94-3	1,2,4,5-Tetrachlorobenzene	170	U	1	25.6	170	ug/Kg	10/29/25 13:25	PB170215
123-91-1	1,4-Dioxane	170	U	1	45.2	170	ug/Kg	10/29/25 13:25	PB170215
58-90-2	2,3,4,6-Tetrachlorophenol	170	U	1	27.4	170	ug/Kg	10/29/25 13:25	PB170215
SURROGATES									
367-12-4	2-Fluorophenol	139			18 - 112	92%	SPK: 150		
13127-88-3	Phenol-d6	139			15 - 107	92%	SPK: 150		
4165-60-0	Nitrobenzene-d5	110	*		18 - 107	110%	SPK: 100		
321-60-8	2-Fluorobiphenyl	96.8			20 - 109	97%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	161			10 - 116	108%	SPK: 150		
1718-51-0	Terphenyl-d14	90.8			10 - 105	91%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	34000							
1146-65-2	Naphthalene-d8	131000							
15067-26-2	Acenaphthene-d10	98300							
1517-22-2	Phenanthrene-d10	248000							
1719-03-5	Chrysene-d12	318000							
1520-96-3	Perylene-d12	369000							

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM			Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206			Date Received:	
Client Sample ID:	PB170215BL			SDG No.:	Q3434
Lab Sample ID:	PB170215BL			Matrix:	SOIL
Analytical Method:	8270E	Level:	LOW	% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol:	1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541	Prep Date:	10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064621.D
Acq On : 29 Oct 2025 13:25
Operator : RC/JU
Sample : PB170215BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Quant Time: Oct 29 14:03:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.216	152	33974	20.000	ng	0.00
21) Naphthalene-d8	11.043	136	131027	20.000	ng	0.00
39) Acenaphthene-d10	14.838	164	98338	20.000	ng	0.00
64) Phenanthrene-d10	17.576	188	248117	20.000	ng	0.00
76) Chrysene-d12	21.848	240	318243	20.000	ng	-0.02
86) Perylene-d12	25.191	264	369003	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.749	112	280581	138.622	ng	0.00
7) Phenol-d6	7.353	99	385276	138.716	ng	0.00
23) Nitrobenzene-d5	9.380	82	240197	110.106	ng	-0.01
42) 2,4,6-Tribromophenol	16.319	330	229169	161.465	ng	0.00
45) 2-Fluorobiphenyl	13.469	172	628128	96.817	ng	0.00
79) Terphenyl-d14	20.167	244	1531012	90.788	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064621.D
Acq On : 29 Oct 2025 13:25
Operator : RC/JU
Sample : PB170215BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Quant Time: Oct 29 14:03:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.216 min Scan# 8
Delta R.T. -0.005 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Tgt Ion:152 Resp: 33974

Ion	Ratio	Lower	Upper
152	100		
150	161.9	122.2	183.4
115	62.1	50.8	76.2







#5
2-Fluorophenol
Concen: 138.622 ng
RT: 5.749 min Scan# 453
Delta R.T. 0.002 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Tgt Ion:112 Resp: 280581

Ion	Ratio	Lower	Upper
112	100		
64	81.0	60.7	91.1
63	43.3	33.4	50.0







#7
Phenol-d6
Concen: 138.716 ng
RT: 7.353 min Scan# 7
Delta R.T. -0.004 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Tgt Ion: 99 Resp: 385276

Ion	Ratio	Lower	Upper
99	100		
42	24.1	18.3	27.5
71	37.5	28.6	43.0







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.043 min Scan# 1354
Delta R.T. -0.010 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Tgt Ion:136 Resp: 131027

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	13.2	12.6	19.0
68	6.7	5.7	8.5







#23
Nitrobenzene-d5
Concen: 110.106 ng
RT: 9.380 min Scan# 1
Delta R.T. -0.010 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Tgt Ion: 82 Resp: 240197

Ion	Ratio	Lower	Upper
82	100		
128	38.9	31.3	46.9
54	71.4	56.6	84.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.838 min Scan# 2000
Delta R.T. -0.004 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Tgt Ion:164 Resp: 98338

Ion	Ratio	Lower	Upper
164	100		
162	99.3	77.0	115.6
160	47.9	34.4	51.6







#42
2,4,6-Tribromophenol
Concen: 161.465 ng
RT: 16.319 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Tgt Ion:330 Resp: 229169

Ion	Ratio	Lower	Upper
330	100		
332	97.3	78.0	117.0
141	27.1	21.8	32.6







#45
2-Fluorobiphenyl
Concen: 96.817 ng
RT: 13.469 min Scan# 1767
Delta R.T. -0.010 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Tgt Ion:172 Resp: 628128

Ion	Ratio	Lower	Upper
172	100		
171	35.8	29.4	44.0
170	24.8	19.1	28.7







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.576 min Scan# 24
Delta R.T. -0.004 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Tgt Ion:188 Resp: 248117

Ion	Ratio	Lower	Upper
188	100		
94	7.2	5.7	8.5
80	7.6	6.8	10.2







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.848 min Scan# 3193
Delta R.T. -0.016 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Tgt Ion:240 Resp: 318243

Ion	Ratio	Lower	Upper
240	100		
120	7.6	6.5	9.7
236	25.4	20.2	30.2







#79
Terphenyl-d14
Concen: 90.788 ng
RT: 20.167 min Scan# 2
Delta R.T. -0.010 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Tgt Ion:244 Resp: 1531012

Ion	Ratio	Lower	Upper
244	100		
212	7.1	6.1	9.1
122	7.0	5.6	8.4






#86
Perylene-d12
Concen: 20.000 ng
RT: 25.191 min Scan# 3762
Delta R.T. -0.010 min
Lab File: BG064621.D
Acq: 29 Oct 2025 13:25

Tgt Ion:264 Resp: 369003

Ion	Ratio	Lower	Upper
264	100		
260	23.3	19.0	28.4
265	21.4	17.6	26.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064621.D
Acq On : 29 Oct 2025 13:25
Operator : RC/JU
Sample : PB170215BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Integration Parameters: rteint.p
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC: BG064621.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.749	445	453	461	rBV	736299	1176882	29.66%	8.360%
2	7.353	717	726	733	rBV	746346	1274418	32.12%	9.053%
3	8.216	866	873	880	rVB	126297	212505	5.36%	1.510%
4	9.380	1063	1071	1078	rBV	442409	777277	19.59%	5.522%
5	11.043	1344	1354	1362	rBV	162352	294281	7.42%	2.090%
6	13.469	1760	1767	1774	rBV	1202173	1855560	46.77%	13.181%
7	14.838	1993	2000	2008	rBV	270056	418534	10.55%	2.973%
8	16.313	2244	2251	2263	rBV	1042193	1646496	41.50%	11.696%
9	17.576	2459	2466	2473	rBV	394044	587050	14.80%	4.170%
10	20.167	2901	2907	2917	rBV	3054784	3967474	100.00%	28.184%
11	21.848	3186	3193	3201	rBV2	492613	821231	20.70%	5.834%
12	24.286	3601	3608	3615	rVB3	25469	62345	1.57%	0.443%
13	25.191	3751	3762	3772	rBV	308414	911115	22.96%	6.472%
14	26.495	3977	3984	3996	rVB5	25920	72023	1.82%	0.512%

Sum of corrected areas: 14077191


Sample Results: **BG064621.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064621.D
Acq On : 29 Oct 2025 13:25
Operator : RC/JU
Sample : PB170215BL
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
PB170215BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P







Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064621.D
Acq On : 29 Oct 2025 13:25
Operator : RC/JU
Sample : PB170215BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected
****

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064621.D
Acq On : 29 Oct 2025 13:25
Operator : RC/JU
Sample : PB170215BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170215BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

					--Internal Standard---			
TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170286BL	SDG No.:	Q3434
Lab Sample ID:	PB170286BL	Matrix:	Water
Analytical Method:	8270E Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C Prep Date: 10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	10.0	U	1	3.90	10.0	ug/L	10/30/25 12:17	PB170286
108-95-2	Phenol	5.00	U	1	0.91	5.00	ug/L	10/30/25 12:17	PB170286
111-44-4	bis(2-Chloroethyl)ether	5.00	U	1	0.81	5.00	ug/L	10/30/25 12:17	PB170286
95-57-8	2-Chlorophenol	5.00	U	1	0.58	5.00	ug/L	10/30/25 12:17	PB170286
95-48-7	2-Methylphenol	5.00	U	1	1.10	5.00	ug/L	10/30/25 12:17	PB170286
108-60-1	2,2-oxybis(1-Chloropropane)	5.00	U	1	1.30	5.00	ug/L	10/30/25 12:17	PB170286
98-86-2	Acetophenone	5.00	U	1	0.74	5.00	ug/L	10/30/25 12:17	PB170286
65794-96-9	3+4-Methylphenols	10.0	U	1	1.10	10.0	ug/L	10/30/25 12:17	PB170286
621-64-7	n-Nitroso-di-n-propylamine	2.50	U	1	1.40	2.50	ug/L	10/30/25 12:17	PB170286
67-72-1	Hexachloroethane	5.00	U	1	0.65	5.00	ug/L	10/30/25 12:17	PB170286
98-95-3	Nitrobenzene	5.00	U	1	0.76	5.00	ug/L	10/30/25 12:17	PB170286
78-59-1	Isophorone	5.00	U	1	0.75	5.00	ug/L	10/30/25 12:17	PB170286
88-75-5	2-Nitrophenol	5.00	U	1	1.80	5.00	ug/L	10/30/25 12:17	PB170286
105-67-9	2,4-Dimethylphenol	5.00	U	1	1.90	5.00	ug/L	10/30/25 12:17	PB170286
111-91-1	bis(2-Chloroethoxy)methane	5.00	U	1	0.68	5.00	ug/L	10/30/25 12:17	PB170286
120-83-2	2,4-Dichlorophenol	5.00	U	1	0.52	5.00	ug/L	10/30/25 12:17	PB170286
91-20-3	Naphthalene	5.00	U	1	0.50	5.00	ug/L	10/30/25 12:17	PB170286
106-47-8	4-Chloroaniline	5.00	U	1	0.84	5.00	ug/L	10/30/25 12:17	PB170286
87-68-3	Hexachlorobutadiene	5.00	U	1	0.54	5.00	ug/L	10/30/25 12:17	PB170286
105-60-2	Caprolactam	10.0	U	1	1.10	10.0	ug/L	10/30/25 12:17	PB170286
59-50-7	4-Chloro-3-methylphenol	5.00	U	1	0.59	5.00	ug/L	10/30/25 12:17	PB170286
91-57-6	2-Methylnaphthalene	5.00	U	1	0.56	5.00	ug/L	10/30/25 12:17	PB170286
77-47-4	Hexachlorocyclopentadiene	10.0	U	1	3.60	10.0	ug/L	10/30/25 12:17	PB170286
88-06-2	2,4,6-Trichlorophenol	5.00	U	1	0.51	5.00	ug/L	10/30/25 12:17	PB170286
95-95-4	2,4,5-Trichlorophenol	5.00	U	1	0.62	5.00	ug/L	10/30/25 12:17	PB170286
92-52-4	1,1-Biphenyl	5.00	U	1	0.53	5.00	ug/L	10/30/25 12:17	PB170286
91-58-7	2-Chloronaphthalene	5.00	U	1	0.61	5.00	ug/L	10/30/25 12:17	PB170286
88-74-4	2-Nitroaniline	5.00	U	1	1.30	5.00	ug/L	10/30/25 12:17	PB170286
131-11-3	Dimethylphthalate	5.00	U	1	0.61	5.00	ug/L	10/30/25 12:17	PB170286
208-96-8	Acenaphthylene	5.00	U	1	0.75	5.00	ug/L	10/30/25 12:17	PB170286
606-20-2	2,6-Dinitrotoluene	5.00	U	1	0.92	5.00	ug/L	10/30/25 12:17	PB170286
99-09-2	3-Nitroaniline	5.00	U	1	1.10	5.00	ug/L	10/30/25 12:17	PB170286
83-32-9	Acenaphthene	5.00	U	1	0.55	5.00	ug/L	10/30/25 12:17	PB170286
51-28-5	2,4-Dinitrophenol	10.0	U	1	6.00	10.0	ug/L	10/30/25 12:17	PB170286
100-02-7	4-Nitrophenol	10.0	U	1	2.40	10.0	ug/L	10/30/25 12:17	PB170286
132-64-9	Dibenzofuran	5.00	U	1	0.61	5.00	ug/L	10/30/25 12:17	PB170286
121-14-2	2,4-Dinitrotoluene	5.00	U	1	1.20	5.00	ug/L	10/30/25 12:17	PB170286
84-66-2	Diethylphthalate	5.00	U	1	0.69	5.00	ug/L	10/30/25 12:17	PB170286
7005-72-3	4-Chlorophenyl-phenylether	5.00	U	1	0.68	5.00	ug/L	10/30/25 12:17	PB170286
86-73-7	Fluorene	5.00	U	1	0.63	5.00	ug/L	10/30/25 12:17	PB170286
100-01-6	4-Nitroaniline	5.00	U	1	1.50	5.00	ug/L	10/30/25 12:17	PB170286
534-52-1	4,6-Dinitro-2-methylphenol	10.0	U	1	2.90	10.0	ug/L	10/30/25 12:17	PB170286

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170286BL	SDG No.:	Q3434
Lab Sample ID:	PB170286BL	Matrix:	Water
Analytical Method:	8270E Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C Prep Date: 10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	5.00	U	1	0.58	5.00	ug/L	10/30/25 12:17	PB170286
101-55-3	4-Bromophenyl-phenylether	5.00	U	1	0.40	5.00	ug/L	10/30/25 12:17	PB170286
118-74-1	Hexachlorobenzene	5.00	U	1	0.52	5.00	ug/L	10/30/25 12:17	PB170286
1912-24-9	Atrazine	5.00	U	1	1.00	5.00	ug/L	10/30/25 12:17	PB170286
87-86-5	Pentachlorophenol	10.0	U	1	1.60	10.0	ug/L	10/30/25 12:17	PB170286
85-01-8	Phenanthrene	5.00	U	1	0.50	5.00	ug/L	10/30/25 12:17	PB170286
120-12-7	Anthracene	5.00	U	1	0.61	5.00	ug/L	10/30/25 12:17	PB170286
86-74-8	Carbazole	5.00	U	1	0.72	5.00	ug/L	10/30/25 12:17	PB170286
84-74-2	Di-n-butylphthalate	5.00	U	1	1.20	5.00	ug/L	10/30/25 12:17	PB170286
206-44-0	Fluoranthene	5.00	U	1	0.82	5.00	ug/L	10/30/25 12:17	PB170286
129-00-0	Pyrene	5.00	U	1	0.50	5.00	ug/L	10/30/25 12:17	PB170286
85-68-7	Butylbenzylphthalate	5.00	U	1	1.90	5.00	ug/L	10/30/25 12:17	PB170286
91-94-1	3,3-Dichlorobenzidine	10.0	U	1	0.93	10.0	ug/L	10/30/25 12:17	PB170286
56-55-3	Benzo(a)anthracene	5.00	U	1	0.45	5.00	ug/L	10/30/25 12:17	PB170286
218-01-9	Chrysene	5.00	U	1	0.44	5.00	ug/L	10/30/25 12:17	PB170286
117-81-7	Bis(2-ethylhexyl)phthalate	5.00	U	1	1.60	5.00	ug/L	10/30/25 12:17	PB170286
117-84-0	Di-n-octyl phthalate	10.0	U	1	2.30	10.0	ug/L	10/30/25 12:17	PB170286
205-99-2	Benzo(b)fluoranthene	5.00	U	1	0.49	5.00	ug/L	10/30/25 12:17	PB170286
207-08-9	Benzo(k)fluoranthene	5.00	U	1	0.48	5.00	ug/L	10/30/25 12:17	PB170286
50-32-8	Benzo(a)pyrene	5.00	U	1	0.55	5.00	ug/L	10/30/25 12:17	PB170286
193-39-5	Indeno(1,2,3-cd)pyrene	5.00	U	1	0.59	5.00	ug/L	10/30/25 12:17	PB170286
53-70-3	Dibenzo(a,h)anthracene	5.00	U	1	0.67	5.00	ug/L	10/30/25 12:17	PB170286
191-24-2	Benzo(g,h,i)perylene	5.00	U	1	0.69	5.00	ug/L	10/30/25 12:17	PB170286
95-94-3	1,2,4,5-Tetrachlorobenzene	5.00	U	1	0.52	5.00	ug/L	10/30/25 12:17	PB170286
123-91-1	1,4-Dioxane	5.00	U	1	1.00	5.00	ug/L	10/30/25 12:17	PB170286
58-90-2	2,3,4,6-Tetrachlorophenol	5.00	U	1	0.72	5.00	ug/L	10/30/25 12:17	PB170286
SURROGATES									
367-12-4	2-Fluorophenol	133			23 - 138	88%	SPK: 150		
13127-88-3	Phenol-d6	126			10 - 134	84%	SPK: 150		
4165-60-0	Nitrobenzene-d5	102			67 - 132	102%	SPK: 100		
321-60-8	2-Fluorobiphenyl	87.4			52 - 132	87%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	148			44 - 137	99%	SPK: 150		
1718-51-0	Terphenyl-d14	79.4			42 - 152	79%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	25100							
1146-65-2	Naphthalene-d8	99200							
15067-26-2	Acenaphthene-d10	73700							
1517-22-2	Phenanthrene-d10	192000							
1719-03-5	Chrysene-d12	261000							
1520-96-3	Perylene-d12	303000							

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM			Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206			Date Received:	
Client Sample ID:	PB170286BL			SDG No.:	Q3434
Lab Sample ID:	PB170286BL			Matrix:	Water
Analytical Method:	8270E	Level:	LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C	Prep Date:	10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064637.D
Acq On : 30 Oct 2025 12:17
Operator : RC/JU
Sample : PB170286BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Quant Time: Oct 30 12:52:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.217	152	25136	20.000	ng	0.00
21) Naphthalene-d8	11.043	136	99198	20.000	ng	0.00
39) Acenaphthene-d10	14.839	164	73729	20.000	ng	0.00
64) Phenanthrene-d10	17.577	188	191571	20.000	ng	0.00
76) Chrysene-d12	21.848	240	261407	20.000	ng	-0.02
86) Perylene-d12	25.191	264	302809	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.749	112	198539	132.577	ng	0.00
7) Phenol-d6	7.353	99	259688	126.373	ng	0.00
23) Nitrobenzene-d5	9.380	82	168217	101.852	ng	0.00
42) 2,4,6-Tribromophenol	16.319	330	157646	148.145	ng	0.00
45) 2-Fluorobiphenyl	13.470	172	425146	87.403	ng	0.00
79) Terphenyl-d14	20.168	244	1099593	79.383	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064637.D
Acq On : 30 Oct 2025 12:17
Operator : RC/JU
Sample : PB170286BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Quant Time: Oct 30 12:52:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.217 min Scan# 8
Delta R.T. -0.004 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Tgt Ion:152 Resp: 25136

Ion	Ratio	Lower	Upper
152	100		
150	157.9	122.2	183.4
115	64.6	50.8	76.2







#5
2-Fluorophenol
Concen: 132.577 ng
RT: 5.749 min Scan# 453
Delta R.T. 0.002 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Tgt Ion:112 Resp: 198539

Ion	Ratio	Lower	Upper
112	100		
64	79.2	60.7	91.1
63	44.3	33.4	50.0







#7
Phenol-d6
Concen: 126.373 ng
RT: 7.353 min Scan# 7
Delta R.T. -0.003 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Tgt Ion: 99 Resp: 259688

Ion	Ratio	Lower	Upper
99	100		
42	23.5	18.3	27.5
71	37.9	28.6	43.0







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.043 min Scan# 1354
Delta R.T. -0.009 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Tgt Ion:136 Resp: 99198

Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.2	13.8
54	14.6	12.6	19.0
68	6.6	5.7	8.5







#23
Nitrobenzene-d5
Concen: 101.852 ng
RT: 9.380 min Scan# 1
Delta R.T. -0.010 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Tgt Ion: 82 Resp: 168217

Ion	Ratio	Lower	Upper
82	100		
128	38.3	31.3	46.9
54	71.0	56.6	84.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.839 min Scan# 2000
Delta R.T. -0.003 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Tgt Ion:164 Resp: 73729

Ion	Ratio	Lower	Upper
164	100		
162	99.5	77.0	115.6
160	43.5	34.4	51.6







#42
2,4,6-Tribromophenol
Concen: 148.145 ng
RT: 16.319 min Scan# 2
Delta R.T. -0.004 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Tgt Ion:330 Resp: 157646

Ion	Ratio	Lower	Upper
330	100		
332	96.6	78.0	117.0
141	27.0	21.8	32.6







#45
2-Fluorobiphenyl
Concen: 87.403 ng
RT: 13.470 min Scan# 1767
Delta R.T. -0.009 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Tgt Ion:172 Resp: 425146

Ion	Ratio	Lower	Upper
172	100		
171	35.9	29.4	44.0
170	24.0	19.1	28.7







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.577 min Scan# 24
Delta R.T. -0.003 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Tgt Ion:188 Resp: 191571

Ion	Ratio	Lower	Upper
188	100		
94	6.5	5.7	8.5
80	8.2	6.8	10.2







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.848 min Scan# 3193
Delta R.T. -0.015 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Tgt Ion:240 Resp: 261407

Ion	Ratio	Lower	Upper
240	100		
120	7.4	6.5	9.7
236	25.3	20.2	30.2







#79
Terphenyl-d14
Concen: 79.383 ng
RT: 20.168 min Scan# 2
Delta R.T. -0.009 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Tgt Ion:244 Resp: 1099593

Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.1	9.1
122	6.5	5.6	8.4






#86
Perylene-d12
Concen: 20.000 ng
RT: 25.191 min Scan# 3762
Delta R.T. -0.009 min
Lab File: BG064637.D
Acq: 30 Oct 2025 12:17

Tgt Ion:264 Resp: 302809

Ion	Ratio	Lower	Upper
264	100		
260	23.5	19.0	28.4
265	21.2	17.6	26.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064637.D
Acq On : 30 Oct 2025 12:17
Operator : RC/JU
Sample : PB170286BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Integration Parameters: rteint.p
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC: BG064637.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.749	444	453	464	rBV	528450	833334	29.47%	8.359%
2	7.353	718	726	735	rBV	500002	849437	30.04%	8.521%
3	8.217	866	873	880	rBV	93110	156411	5.53%	1.569%
4	9.380	1063	1071	1080	rBV	307213	540815	19.13%	5.425%
5	11.043	1347	1354	1363	rVB	124300	220900	7.81%	2.216%
6	13.470	1760	1767	1774	rBV	790004	1250233	44.22%	12.541%
7	14.839	1993	2000	2007	rVB	205833	315418	11.16%	3.164%
8	16.313	2244	2251	2266	rBV	740874	1127251	39.87%	11.307%
9	17.577	2459	2466	2473	rBV	301398	452463	16.00%	4.539%
10	20.168	2900	2907	2914	rBV	2149344	2827435	100.00%	28.361%
11	21.848	3187	3193	3200	rBV2	402931	660745	23.37%	6.628%
12	25.191	3751	3762	3773	rBV	242022	734877	25.99%	7.371%

Sum of corrected areas: 9969319


Sample Results: **BG064637.D**


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064637.D
Acq On : 30 Oct 2025 12:17
Operator : RC/JU
Sample : PB170286BL
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
PB170286BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P







Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064637.D
Acq On : 30 Oct 2025 12:17
Operator : RC/JU
Sample : PB170286BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected
****

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064637.D
Acq On : 30 Oct 2025 12:17
Operator : RC/JU
Sample : PB170286BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170286BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

					--Internal Standard---			
TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170215BS	SDG No.:	Q3434
Lab Sample ID:	PB170215BS	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	100
Sample Wt/Vol:	30.03 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	1100		1	160	330	ug/Kg	10/29/25 14:06	PB170215
108-95-2	Phenol	1500		1	22.1	170	ug/Kg	10/29/25 14:06	PB170215
111-44-4	bis(2-Chloroethyl)ether	1300		1	24.3	170	ug/Kg	10/29/25 14:06	PB170215
95-57-8	2-Chlorophenol	1600		1	24.4	170	ug/Kg	10/29/25 14:06	PB170215
95-48-7	2-Methylphenol	1500		1	29.9	170	ug/Kg	10/29/25 14:06	PB170215
108-60-1	2,2-oxybis(1-Chloropropane)	1300		1	37.5	170	ug/Kg	10/29/25 14:06	PB170215
98-86-2	Acetophenone	1600		1	29.5	170	ug/Kg	10/29/25 14:06	PB170215
65794-96-9	3+4-Methylphenols	1500		1	41.1	330	ug/Kg	10/29/25 14:06	PB170215
621-64-7	n-Nitroso-di-n-propylamine	1400		1	47.4	79.9	ug/Kg	10/29/25 14:06	PB170215
67-72-1	Hexachloroethane	1500		1	17.6	170	ug/Kg	10/29/25 14:06	PB170215
98-95-3	Nitrobenzene	1600		1	18.3	170	ug/Kg	10/29/25 14:06	PB170215
78-59-1	Isophorone	1400		1	32.8	170	ug/Kg	10/29/25 14:06	PB170215
88-75-5	2-Nitrophenol	1700		1	58.1	170	ug/Kg	10/29/25 14:06	PB170215
105-67-9	2,4-Dimethylphenol	1700		1	64.7	170	ug/Kg	10/29/25 14:06	PB170215
111-91-1	bis(2-Chloroethoxy)methane	1400		1	30.8	170	ug/Kg	10/29/25 14:06	PB170215
120-83-2	2,4-Dichlorophenol	1600		1	28.3	170	ug/Kg	10/29/25 14:06	PB170215
91-20-3	Naphthalene	1400		1	22.7	170	ug/Kg	10/29/25 14:06	PB170215
106-47-8	4-Chloroaniline	790		1	35.4	170	ug/Kg	10/29/25 14:06	PB170215
87-68-3	Hexachlorobutadiene	1500		1	25.3	170	ug/Kg	10/29/25 14:06	PB170215
105-60-2	Caprolactam	1500		1	52.0	330	ug/Kg	10/29/25 14:06	PB170215
59-50-7	4-Chloro-3-methylphenol	1600		1	28.7	170	ug/Kg	10/29/25 14:06	PB170215
91-57-6	2-Methylnaphthalene	1300		1	25.6	170	ug/Kg	10/29/25 14:06	PB170215
77-47-4	Hexachlorocyclopentadiene	4100	E	1	120	330	ug/Kg	10/29/25 14:06	PB170215
88-06-2	2,4,6-Trichlorophenol	1700		1	19.8	170	ug/Kg	10/29/25 14:06	PB170215
95-95-4	2,4,5-Trichlorophenol	1600		1	29.1	170	ug/Kg	10/29/25 14:06	PB170215
92-52-4	1,1-Biphenyl	1600		1	21.8	170	ug/Kg	10/29/25 14:06	PB170215
91-58-7	2-Chloronaphthalene	1400		1	22.5	170	ug/Kg	10/29/25 14:06	PB170215
88-74-4	2-Nitroaniline	1600		1	48.1	170	ug/Kg	10/29/25 14:06	PB170215
131-11-3	Dimethylphthalate	1500		1	27.1	170	ug/Kg	10/29/25 14:06	PB170215
208-96-8	Acenaphthylene	1700		1	28.9	170	ug/Kg	10/29/25 14:06	PB170215
606-20-2	2,6-Dinitrotoluene	1700		1	33.6	170	ug/Kg	10/29/25 14:06	PB170215
99-09-2	3-Nitroaniline	1200		1	46.0	170	ug/Kg	10/29/25 14:06	PB170215
83-32-9	Acenaphthene	1600		1	21.3	170	ug/Kg	10/29/25 14:06	PB170215
51-28-5	2,4-Dinitrophenol	3600	E	1	230	330	ug/Kg	10/29/25 14:06	PB170215
100-02-7	4-Nitrophenol	3400	E	1	110	330	ug/Kg	10/29/25 14:06	PB170215
132-64-9	Dibenzofuran	1500		1	22.7	170	ug/Kg	10/29/25 14:06	PB170215
121-14-2	2,4-Dinitrotoluene	1700		1	50.0	170	ug/Kg	10/29/25 14:06	PB170215
84-66-2	Diethylphthalate	1500		1	28.3	170	ug/Kg	10/29/25 14:06	PB170215
7005-72-3	4-Chlorophenyl-phenylether	1500		1	26.7	170	ug/Kg	10/29/25 14:06	PB170215
86-73-7	Fluorene	1500		1	25.3	170	ug/Kg	10/29/25 14:06	PB170215
100-01-6	4-Nitroaniline	1700		1	64.1	170	ug/Kg	10/29/25 14:06	PB170215
534-52-1	4,6-Dinitro-2-methylphenol	1800		1	100	330	ug/Kg	10/29/25 14:06	PB170215

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM			Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206			Date Received:	
Client Sample ID:	PB170215BS			SDG No.:	Q3434
Lab Sample ID:	PB170215BS			Matrix:	SOIL
Analytical Method:	8270E	Level:	LOW	% Solid:	100
Sample Wt/Vol:	30.03 g	Final Vol:	1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541	Prep Date:	10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	1500		1	32.9	170	ug/Kg	10/29/25 14:06	PB170215
101-55-3	4-Bromophenyl-phenylether	1500		1	27.8	170	ug/Kg	10/29/25 14:06	PB170215
118-74-1	Hexachlorobenzene	1500		1	25.3	170	ug/Kg	10/29/25 14:06	PB170215
1912-24-9	Atrazine	1600		1	34.0	170	ug/Kg	10/29/25 14:06	PB170215
87-86-5	Pentachlorophenol	3300	E	1	51.2	330	ug/Kg	10/29/25 14:06	PB170215
85-01-8	Phenanthrene	1500		1	20.9	170	ug/Kg	10/29/25 14:06	PB170215
120-12-7	Anthracene	1500		1	33.3	170	ug/Kg	10/29/25 14:06	PB170215
86-74-8	Carbazole	1500		1	31.2	170	ug/Kg	10/29/25 14:06	PB170215
84-74-2	Di-n-butylphthalate	1500		1	47.9	170	ug/Kg	10/29/25 14:06	PB170215
206-44-0	Fluoranthene	1500		1	30.0	170	ug/Kg	10/29/25 14:06	PB170215
129-00-0	Pyrene	1400		1	36.0	170	ug/Kg	10/29/25 14:06	PB170215
85-68-7	Butylbenzylphthalate	1600		1	71.3	170	ug/Kg	10/29/25 14:06	PB170215
91-94-1	3,3-Dichlorobenzidine	1100		1	36.7	330	ug/Kg	10/29/25 14:06	PB170215
56-55-3	Benzo(a)anthracene	1500		1	23.0	170	ug/Kg	10/29/25 14:06	PB170215
218-01-9	Chrysene	1500		1	19.9	170	ug/Kg	10/29/25 14:06	PB170215
117-81-7	Bis(2-ethylhexyl)phthalate	1600		1	59.1	170	ug/Kg	10/29/25 14:06	PB170215
117-84-0	Di-n-octyl phthalate	1700		1	86.7	330	ug/Kg	10/29/25 14:06	PB170215
205-99-2	Benzo(b)fluoranthene	1500		1	19.0	170	ug/Kg	10/29/25 14:06	PB170215
207-08-9	Benzo(k)fluoranthene	1500		1	22.4	170	ug/Kg	10/29/25 14:06	PB170215
50-32-8	Benzo(a)pyrene	1600		1	29.5	170	ug/Kg	10/29/25 14:06	PB170215
193-39-5	Indeno(1,2,3-cd)pyrene	1500		1	29.1	170	ug/Kg	10/29/25 14:06	PB170215
53-70-3	Dibenzo(a,h)anthracene	1500		1	27.4	170	ug/Kg	10/29/25 14:06	PB170215
191-24-2	Benzo(g,h,i)perylene	1500		1	25.7	170	ug/Kg	10/29/25 14:06	PB170215
95-94-3	1,2,4,5-Tetrachlorobenzene	1600		1	25.6	170	ug/Kg	10/29/25 14:06	PB170215
123-91-1	1,4-Dioxane	1100		1	45.2	170	ug/Kg	10/29/25 14:06	PB170215
58-90-2	2,3,4,6-Tetrachlorophenol	1800		1	27.4	170	ug/Kg	10/29/25 14:06	PB170215
SURROGATES									
367-12-4	2-Fluorophenol	125			18 - 112	83%	SPK: 150		
13127-88-3	Phenol-d6	125			15 - 107	83%	SPK: 150		
4165-60-0	Nitrobenzene-d5	95.4			18 - 107	95%	SPK: 100		
321-60-8	2-Fluorobiphenyl	79.6			20 - 109	80%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	144			10 - 116	96%	SPK: 150		
1718-51-0	Terphenyl-d14	83.3			10 - 105	83%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	30900							
1146-65-2	Naphthalene-d8	126000							
15067-26-2	Acenaphthene-d10	103000							
1517-22-2	Phenanthrene-d10	253000							
1719-03-5	Chrysene-d12	286000							
1520-96-3	Perylene-d12	311000							

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM			Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206			Date Received:	
Client Sample ID:	PB170215BS			SDG No.:	Q3434
Lab Sample ID:	PB170215BS			Matrix:	SOIL
Analytical Method:	8270E	Level:	LOW	% Solid:	100
Sample Wt/Vol:	30.03 g	Final Vol:	1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541	Prep Date:	10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064622.D
Acq On : 29 Oct 2025 14:06
Operator : RC/JU
Sample : PB170215BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Nov 04 06:35:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.217	152	30857	20.000	ng		0.00
21) Naphthalene-d8	11.044	136	126314	20.000	ng		0.00
39) Acenaphthene-d10	14.839	164	102616	20.000	ng		0.00
64) Phenanthrene-d10	17.577	188	252758	20.000	ng		0.00
76) Chrysene-d12	21.854	240	286271	20.000	ng		0.00
86) Perylene-d12	25.192	264	311364	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.750	112	229807	125.006	ng		0.00
7) Phenol-d6	7.354	99	315661	125.132	ng		0.00
23) Nitrobenzene-d5	9.381	82	200721	95.443	ng		0.00
42) 2,4,6-Tribromophenol	16.320	330	213604	144.224	ng		0.00
45) 2-Fluorobiphenyl	13.470	172	539175	79.642	ng		0.00
79) Terphenyl-d14	20.168	244	1263324	83.281	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.582	88	27367	33.625	ng		97
3) Pyridine	3.987	79	80236	32.828	ng		99
4) n-Nitrosodimethylamine	3.899	42	55779	42.556	ng		96
6) Aniline	7.524	93	111247	32.409	ng		98
8) 2-Chlorophenol	7.777	128	90171	46.861	ng		96
9) Benzaldehyde	7.336	77	44662	32.430	ng		99
10) Phenol	7.377	94	125167	44.866	ng		96
11) bis(2-Chloroethyl)ether	7.618	93	82115	40.034	ng		94
12) 1,3-Dichlorobenzene	8.106	146	89896	40.591	ng		96
13) 1,4-Dichlorobenzene	8.253	146	95840	42.559	ng		92
14) 1,2-Dichlorobenzene	8.576	146	95566	43.998	ng		95
15) Benzyl Alcohol	8.447	79	97685	46.959	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.746	45	208165	38.690	ng		97
17) 2-Methylphenol	8.646	107	82384	44.477	ng		95
18) Hexachloroethane	9.322	117	34789	44.022	ng		96
19) n-Nitroso-di-n-propyla...	9.022	70	75603	43.365	ng		98
20) 3+4-Methylphenols	8.975	107	115498	44.339	ng		95
22) Acetophenone	9.040	105	161222	46.691	ng		96
24) Nitrobenzene	9.422	77	109545	47.891	ng		97
25) Isophorone	9.951	82	215090	42.623	ng		100
26) 2-Nitrophenol	10.139	139	44164	49.908	ng	#	85
27) 2,4-Dimethylphenol	10.192	122	95696	50.981	ng		97
28) bis(2-Chloroethoxy)met...	10.433	93	121990	42.458	ng		99
29) 2,4-Dichlorophenol	10.679	162	99058	49.036	ng		97
30) 1,2,4-Trichlorobenzene	10.903	180	106824	46.541	ng		97
31) Naphthalene	11.091	128	295744	42.229	ng		99
32) Benzoic acid	10.309	122	73385	50.777	ng		94
33) 4-Chloroaniline	11.185	127	64445	23.675	ng		97
34) Hexachlorobutadiene	11.384	225	78836	43.919	ng		97
35) Caprolactam	11.943	113	35241m	45.484	ng		
36) 4-Chloro-3-methylphenol	12.289	107	115355	47.506	ng		98
37) 2-Methylnaphthalene	12.683	142	200867	39.066	ng		99
38) 1-Methylnaphthalene	12.900	142	210799	41.413	ng		95
40) 1,2,4,5-Tetrachloroben...	13.047	216	160584	46.629	ng		99
41) Hexachlorocyclopentadiene	13.035	237	193126	124.589	ng		99
43) 2,4,6-Trichlorophenol	13.276	196	97819	49.656	ng		100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064622.D
Acq On : 29 Oct 2025 14:06
Operator : RC/JU
Sample : PB170215BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Nov 04 06:35:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.347	196	111336	48.976	ng		99
46) 1,1'-Biphenyl	13.682	154	341429	46.801	ng		97
47) 2-Chloronaphthalene	13.723	162	234788	42.637	ng		96
48) 2-Nitroaniline	13.911	65	87853	47.555	ng		94
49) Acenaphthylene	14.563	152	430487	49.562	ng		98
50) Dimethylphthalate	14.287	163	347656	44.275	ng		99
51) 2,6-Dinitrotoluene	14.398	165	65227	50.188	ng		96
52) Acenaphthene	14.904	154	290908	49.021	ng		97
53) 3-Nitroaniline	14.722	138	49122	36.402	ng	#	99
54) 2,4-Dinitrophenol	14.927	184	88044	108.602	ng	#	42
55) Dibenzofuran	15.239	168	418469	43.618	ng		99
56) 4-Nitrophenol	15.015	139	127758	102.115	ng		94
57) 2,4-Dinitrotoluene	15.180	165	100857	50.198	ng		98
58) Fluorene	15.879	166	335610	44.792	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.456	232	114603	52.720	ng		98
60) Diethylphthalate	15.644	149	357386	44.576	ng		99
61) 4-Chlorophenyl-phenyle...	15.873	204	190016	45.190	ng		97
62) 4-Nitroaniline	15.885	138	74872	49.873	ng		95
63) Azobenzene	16.161	77	323701	44.003	ng		99
65) 4,6-Dinitro-2-methylph...	15.944	198	61637	54.580	ng		96
66) n-Nitrosodiphenylamine	16.079	169	307367	45.018	ng		96
67) 4-Bromophenyl-phenylether	16.766	248	133834	44.682	ng		97
68) Hexachlorobenzene	16.884	284	155883	45.618	ng		96
69) Atrazine	17.019	200	140535	47.830	ng		98
70) Pentachlorophenol	17.219	266	214000	100.429	ng		95
71) Phenanthrene	17.618	178	612283	44.297	ng		99
72) Anthracene	17.712	178	631323	44.900	ng		100
73) Carbazole	17.971	167	558592	45.052	ng		99
74) Di-n-butylphthalate	18.529	149	636589	44.400	ng		100
75) Fluoranthene	19.616	202	768700	44.393	ng		99
77) Benzidine	19.780	184	240455	37.491	ng		99
78) Pyrene	19.974	202	804895	43.035	ng		99
80) Butylbenzylphthalate	20.850	149	279278	49.405	ng		97
81) Benzo(a)anthracene	21.831	228	850571	44.811	ng		99
82) 3,3'-Dichlorobenzidine	21.737	252	205029	32.347	ng		98
83) Chrysene	21.901	228	799015	44.240	ng		99
84) Bis(2-ethylhexyl)phtha...	21.737	149	407169	48.204	ng		99
85) Di-n-octyl phthalate	23.000	149	709236	49.802	ng		99
87) Indeno(1,2,3-cd)pyrene	29.034	276	1033332	45.114	ng	#	97
88) Benzo(b)fluoranthene	24.134	252	872834	45.715	ng		99
89) Benzo(k)fluoranthene	24.205	252	867424	45.401	ng		98
90) Benzo(a)pyrene	25.039	252	830855	47.341	ng		100
91) Dibenzo(a,h)anthracene	29.099	278	858540	46.137	ng		99
92) Benzo(g,h,i)perylene	30.233	276	825558	46.441	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064622.D
Acq On : 29 Oct 2025 14:06
Operator : RC/JU
Sample : PB170215BS
Misc :
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Nov 04 06:35:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.217 min Scan# 8
Delta R.T. -0.006 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:152 Resp: 3085


Ion	Ratio	Lower	Upper
152	100		
150	148.2	122.2	183.4
115	56.8	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 33.625 ng
RT: 3.582 min Scan# 84
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 88 Resp: 27367

Ion	Ratio	Lower	Upper
88	100		
58	117.6	92.4	138.6
43	44.4	33.0	49.6






#3
Pyridine
Concen: 32.828 ng
RT: 3.987 min Scan# 1
Delta R.T. -0.011 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 79 Resp: 8023


Ion	Ratio	Lower	Upper
79	100		
52	118.1	94.2	141.2
51	58.8	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 42.556 ng
RT: 3.899 min Scan# 138
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 42 Resp: 55779

Ion	Ratio	Lower	Upper
42	100		
74	98.7	76.1	114.1
44	8.8	6.3	9.5





#5
2-Fluorophenol
Concen: 125.006 ng
RT: 5.750 min Scan# 4
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06


Tgt Ion:112 Resp: 22980


Ion	Ratio	Lower	Upper
112	100		
64	79.8	60.7	91.1
63	43.5	33.4	50.0

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#6
Aniline
Concen: 32.409 ng
RT: 7.524 min Scan# 755
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 93 Resp: 111247

Ion	Ratio	Lower	Upper
93	100		
66	44.6	37.0	55.6
65	22.9	18.3	27.5






#7
Phenol-d6
Concen: 125.132 ng
RT: 7.354 min Scan# 7
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 99 Resp: 31566


Ion	Ratio	Lower	Upper
99	100		
42	25.1	18.3	27.5
71	38.4	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 46.861 ng
RT: 7.777 min Scan# 798
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:128 Resp: 90171

Ion	Ratio	Lower	Upper
128	100		
130	29.9	13.5	53.5
64	62.8	45.2	85.2






#9
Benzaldehyde
Concen: 32.430 ng
RT: 7.336 min Scan# 7
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 77 Resp: 4466


Ion	Ratio	Lower	Upper
77	100		
105	91.9	72.9	112.9
106	91.6	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 44.866 ng
RT: 7.377 min Scan# 730
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 94 Resp: 125167

Ion	Ratio	Lower	Upper
94	100		
65	35.5	15.8	55.8
66	48.7	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 40.034 ng
RT: 7.618 min Scan# 7
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 93 Resp: 8211


Ion	Ratio	Lower	Upper
93	100		
63	104.9	91.3	131.3
95	29.8	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 40.591 ng
RT: 8.106 min Scan# 854
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:146 Resp: 89896

Ion	Ratio	Lower	Upper
146	100		
148	65.5	55.3	82.9
75	34.2	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 42.559 ng
RT: 8.253 min Scan# 8
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:146 Resp: 9584


Ion	Ratio	Lower	Upper
146	100		
148	63.7	56.2	84.4
111	39.7	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 43.998 ng
RT: 8.576 min Scan# 934
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:146 Resp: 95566

Ion	Ratio	Lower	Upper
146	100		
148	69.9	51.5	77.3
111	45.3	35.0	52.6






#15
Benzyl Alcohol
Concen: 46.959 ng
RT: 8.447 min Scan# 9
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 79 Resp: 9768


Ion	Ratio	Lower	Upper
79	100		
108	70.9	58.1	87.1
77	65.6	52.8	79.2



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.690 ng
RT: 8.746 min Scan# 963
Delta R.T. 0.006 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 45 Resp: 208165

Ion	Ratio	Lower	Upper
45	100		
77	9.4	0.0	28.5
79	7.2	0.0	25.6






#17
2-Methylphenol
Concen: 44.477 ng
RT: 8.646 min Scan# 94
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:107 Resp: 8238


Ion	Ratio	Lower	Upper
107	100		
108	105.5	90.0	135.0
77	50.2	37.8	56.8
79	48.3	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 44.022 ng
RT: 9.322 min Scan# 1061
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:117 Resp: 34789

Ion	Ratio	Lower	Upper
117	100		
119	93.6	75.4	113.2
201	130.9	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 43.365 ng
RT: 9.022 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 70 Resp: 7560


Ion	Ratio	Lower	Upper
70	100		
42	79.8	65.2	97.8
101	9.0	7.9	11.9
130	18.0	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 44.339 ng
RT: 8.975 min Scan# 1002
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:107 Resp: 115498

Ion	Ratio	Lower	Upper
107	100		
108	83.0	70.3	110.3
77	35.8	14.6	54.6
79	31.6	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.044 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:136 Resp: 12631


Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.2	13.8
54	14.5	12.6	19.0
68	5.8	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 46.691 ng
RT: 9.040 min Scan# 1013
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:105 Resp: 161222

Ion	Ratio	Lower	Upper
105	100		
71	1.8	1.5	2.3
51	38.9	33.7	50.5
120	19.3	16.1	24.1





#23
Nitrobenzene-d5
Concen: 95.443 ng
RT: 9.381 min Scan# 1(
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06


Tgt Ion: 82 Resp: 20072


Ion	Ratio	Lower	Upper
82	100		
128	38.7	31.3	46.9
54	70.8	56.6	84.8

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#24
Nitrobenzene
Concen: 47.891 ng
RT: 9.422 min Scan# 1078
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 77 Resp: 109545

Ion	Ratio	Lower	Upper
77	100		
123	40.7	31.0	46.6
65	16.5	12.7	19.1






#25
Isophorone
Concen: 42.623 ng
RT: 9.951 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 82 Resp: 21509


Ion	Ratio	Lower	Upper
82	100		
95	7.5	6.1	9.1
138	16.3	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 49.908 ng
RT: 10.139 min Scan# 1200
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:139 Resp: 44164

Ion	Ratio	Lower	Upper
139	100		
109	41.8	24.6	37.0#
65	68.2	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 50.981 ng
RT: 10.192 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:122 Resp: 9569


Ion	Ratio	Lower	Upper
122	100		
107	124.2	103.0	154.4
121	58.7	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 42.458 ng
RT: 10.433 min Scan# 1250
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 93 Resp: 121990

Ion	Ratio	Lower	Upper
93	100		
95	31.0	25.4	38.2
123	10.3	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 49.036 ng
RT: 10.679 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:162 Resp: 9905


Ion	Ratio	Lower	Upper
162	100		
164	67.5	48.7	88.7
98	34.2	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 46.541 ng
RT: 10.903 min Scan# 1330
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:180 Resp: 106824

Ion	Ratio	Lower	Upper
180	100		
182	91.9	76.9	115.3
145	27.5	22.1	33.1






#31
Naphthalene
Concen: 42.229 ng
RT: 11.091 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:128 Resp: 29574


Ion	Ratio	Lower	Upper
128	100		
129	10.6	8.8	13.2
127	12.9	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 50.777 ng
RT: 10.309 min Scan# 1229
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:122 Resp: 73385

Ion	Ratio	Lower	Upper
122	100		
105	126.3	104.2	144.2
77	100.4	70.2	110.2






#33
4-Chloroaniline
Concen: 23.675 ng
RT: 11.185 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:127 Resp: 6444


Ion	Ratio	Lower	Upper
127	100		
129	29.0	25.5	38.3
65	40.8	32.2	48.4
92	20.7	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 43.919 ng
RT: 11.384 min Scan# 1412
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:225 Resp: 78836

Ion	Ratio	Lower	Upper
225	100		
223	57.4	48.8	73.2
227	62.7	49.5	74.3






#35
Caprolactam
Concen: 45.484 ng m
RT: 11.943 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:113 Resp: 3524


Ion	Ratio	Lower	Upper
113	100		
55	290.8	290.1	330.1
56	200.3	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 47.506 ng
RT: 12.289 min Scan# 1566
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:107 Resp: 115355

Ion	Ratio	Lower	Upper
107	100		
144	24.9	19.1	28.7
142	74.6	58.6	87.8






#37
2-Methylnaphthalene
Concen: 39.066 ng
RT: 12.683 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:142 Resp: 20086


Ion	Ratio	Lower	Upper
142	100		
141	87.5	68.9	103.3
115	38.0	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 41.413 ng
RT: 12.900 min Scan# 1670
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:142 Resp: 210799

Ion	Ratio	Lower	Upper
142	100		
141	86.7	73.3	109.9
116	4.3	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.839 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:164 Resp: 10261


Ion	Ratio	Lower	Upper
164	100		
162	100.2	77.0	115.6
160	41.8	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 46.629 ng
RT: 13.047 min Scan# 1695
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:216 Resp: 160584

Ion	Ratio	Lower	Upper
216	100		
214	77.1	61.4	92.2
179	19.2	15.4	23.0
108	16.3	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 124.589 ng
RT: 13.035 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:237 Resp: 19312


Ion	Ratio	Lower	Upper
237	100		
235	63.6	44.1	84.1
272	12.1	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 144.224 ng
RT: 16.320 min Scan# 2252
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:330 Resp: 213604

Ion	Ratio	Lower	Upper
330	100		
332	97.7	78.0	117.0
141	27.7	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 49.656 ng
RT: 13.276 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:196 Resp: 9781


Ion	Ratio	Lower	Upper
196	100		
198	96.2	76.6	115.0
200	32.5	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 48.976 ng
RT: 13.347 min Scan# 1746
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:196 Resp: 111336

Ion	Ratio	Lower	Upper
196	100		
198	93.3	74.2	111.4
97	44.8	37.0	55.6
132	24.0	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 79.642 ng
RT: 13.470 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:172 Resp: 53917


Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.4	44.0
170	24.4	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 46.801 ng
RT: 13.682 min Scan# 1803
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:154 Resp: 341429

Ion	Ratio	Lower	Upper
154	100		
153	41.1	22.4	62.4
76	11.4	0.0	33.5






#47
2-Chloronaphthalene
Concen: 42.637 ng
RT: 13.723 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:162 Resp: 23478


Ion	Ratio	Lower	Upper
162	100		
127	37.9	32.2	48.2
164	32.4	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 47.555 ng
RT: 13.911 min Scan# 1842
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion: 65 Resp: 87853

Ion	Ratio	Lower	Upper
65	100		
92	58.4	48.3	72.5
138	86.3	74.9	112.3






#49
Acenaphthylene
Concen: 49.562 ng
RT: 14.563 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:152 Resp: 43048


Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.4	23.2
153	12.5	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 44.275 ng
RT: 14.287 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:163 Resp: 347656

Ion	Ratio	Lower	Upper
163	100		
194	3.4	2.8	4.2
164	10.2	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 50.188 ng
RT: 14.398 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:165 Resp: 6522


Ion	Ratio	Lower	Upper
165	100		
63	66.6	57.4	86.0
89	60.1	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 49.021 ng
RT: 14.904 min Scan# 2011
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:154 Resp: 290908

Ion	Ratio	Lower	Upper
154	100		
153	112.6	92.2	138.4
152	52.1	43.2	64.8






#53
3-Nitroaniline
Concen: 36.402 ng
RT: 14.722 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:138 Resp: 4912


Ion	Ratio	Lower	Upper
138	100		
108	11.2	11.4	17.0#
92	118.7	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 108.602 ng
RT: 14.927 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:184 Resp: 88044

Ion	Ratio	Lower	Upper
184	100		
63	68.2	56.6	85.0
154	81.9	158.6	238.0#






#55
Dibenzofuran
Concen: 43.618 ng
RT: 15.239 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:168 Resp: 41846


Ion	Ratio	Lower	Upper
168	100		
139	35.5	27.8	41.8
169	12.8	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 102.115 ng
RT: 15.015 min Scan# 2030
Delta R.T. 0.006 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:139 Resp: 127758

Ion	Ratio	Lower	Upper
139	100		
109	94.1	67.8	107.8
65	129.1	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 50.198 ng
RT: 15.180 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:165 Resp: 10085


Ion	Ratio	Lower	Upper
165	100		
63	53.3	41.8	62.8
89	74.2	58.1	87.1
182	1.8	1.4	2.0



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 44.792 ng
RT: 15.879 min Scan# 2177
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:166 Resp: 335610

Ion	Ratio	Lower	Upper
166	100		
165	101.3	81.5	122.3
167	13.3	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 52.720 ng
RT: 15.456 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:232 Resp: 11460


Ion	Ratio	Lower	Upper
232	100		
131	35.2	26.7	40.1
130	1.8	1.6	2.4
166	25.4	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 44.576 ng
RT: 15.644 min Scan# 2137
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:149 Resp: 357386

Ion	Ratio	Lower	Upper
149	100		
177	22.2	18.0	27.0
150	11.7	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 45.190 ng
RT: 15.873 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:204 Resp: 19001


Ion	Ratio	Lower	Upper
204	100		
206	33.6	25.6	38.4
141	48.8	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 49.873 ng
RT: 15.885 min Scan# 2178
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:138 Resp: 74872

Ion	Ratio	Lower	Upper
138	100		
92	51.2	36.1	76.1
108	99.4	75.6	115.6






#63
Azobenzene
Concen: 44.003 ng
RT: 16.161 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06


Tgt Ion: 77 Resp: 32370

Ion	Ratio	Lower	Upper
77	100		
182	25.8	4.9	44.9
105	20.7	1.3	41.3
51	39.4	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
PB170215BS


Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.577 min Scan# 2466
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:188 Resp: 252758

Ion	Ratio	Lower	Upper
188	100		
94	7.9	5.7	8.5
80	8.6	6.8	10.2






#65
4,6-Dinitro-2-methylphenol
Concen: 54.580 ng
RT: 15.944 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06


Tgt Ion:198 Resp: 6163

Ion	Ratio	Lower	Upper
198	100		
51	48.1	24.7	64.7
105	35.9	18.0	58.0



Instrument :
BNA_G
ClientSampleId :
PB170215BS


Manual Integrations
APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#66
n-Nitrosodiphenylamine
Concen: 45.018 ng
RT: 16.079 min Scan# 2211
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:169 Resp: 307367

Ion	Ratio	Lower	Upper
169	100		
168	73.0	56.3	84.5
167	39.0	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 44.682 ng
RT: 16.766 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:248 Resp: 13383


Ion	Ratio	Lower	Upper
248	100		
250	96.3	77.7	116.5
141	56.6	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 45.618 ng
RT: 16.884 min Scan# 2348
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:284 Resp: 155883

Ion	Ratio	Lower	Upper
284	100		
142	26.8	21.0	31.6
249	34.6	24.9	37.3





#69
Atrazine
Concen: 47.830 ng
RT: 17.019 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06


Tgt Ion:200 Resp: 14053


Ion	Ratio	Lower	Upper
200	100		
173	23.0	0.9	40.9
215	47.9	27.0	67.0

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#70
Pentachlorophenol
Concen: 100.429 ng
RT: 17.219 min Scan# 2405
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:266 Resp: 214000

Ion	Ratio	Lower	Upper
266	100		
268	66.3	50.3	75.5
264	63.9	47.3	70.9





#71
Phenanthrene
Concen: 44.297 ng
RT: 17.618 min Scan# 24
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06


Tgt Ion:178 Resp: 61228


Ion	Ratio	Lower	Upper
178	100		
176	19.2	14.7	22.1
179	15.2	12.2	18.4

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#72
Anthracene
Concen: 44.900 ng
RT: 17.712 min Scan# 2489
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:178 Resp: 631323

Ion	Ratio	Lower	Upper
178	100		
176	18.0	14.4	21.6
179	15.2	12.4	18.6





#73
Carbazole
Concen: 45.052 ng
RT: 17.971 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06


Tgt Ion:167 Resp: 55859


Ion	Ratio	Lower	Upper
167	100		
166	22.2	17.9	26.9
139	12.4	9.7	14.5

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#74
Di-n-butylphthalate
Concen: 44.400 ng
RT: 18.529 min Scan# 2628
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:149 Resp: 636589

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.0	10.6
104	5.4	4.1	6.1





#75
Fluoranthene
Concen: 44.393 ng
RT: 19.616 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06


Tgt Ion:202 Resp: 76870


Ion	Ratio	Lower	Upper
202	100		
101	8.2	0.0	28.7
203	17.5	0.0	37.0

Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.854 min Scan# 3194
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:240 Resp: 286271

Ion	Ratio	Lower	Upper
240	100		
120	7.1	6.5	9.7
236	26.8	20.2	30.2






#77
Benzidine
Concen: 37.491 ng
RT: 19.780 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:184 Resp: 24045


Ion	Ratio	Lower	Upper
184	100		
185	13.7	11.4	17.2
183	12.2	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 43.035 ng
RT: 19.974 min Scan# 2874
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:202 Resp: 804895

Ion	Ratio	Lower	Upper
202	100		
200	21.0	17.0	25.4
203	16.9	13.8	20.6






#79
Terphenyl-d14
Concen: 83.281 ng
RT: 20.168 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:244 Resp: 126332


Ion	Ratio	Lower	Upper
244	100		
212	7.1	6.1	9.1
122	6.5	5.6	8.4



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#80
Butylbenzylphthalate
Concen: 49.405 ng
RT: 20.850 min Scan# 3023
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:149 Resp: 279278

Ion	Ratio	Lower	Upper
149	100		
91	70.8	59.2	88.8
206	22.7	18.1	27.1






#81
Benzo(a)anthracene
Concen: 44.811 ng
RT: 21.831 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:228 Resp: 85057


Ion	Ratio	Lower	Upper
228	100		
226	27.6	21.9	32.9
229	20.3	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 32.347 ng
RT: 21.737 min Scan# 3174
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:252 Resp: 205029

Ion	Ratio	Lower	Upper
252	100		
254	64.8	50.8	76.2
126	6.8	5.8	8.6






#83
Chrysene
Concen: 44.240 ng
RT: 21.901 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:228 Resp: 79901


Ion	Ratio	Lower	Upper
228	100		
226	29.8	23.3	34.9
229	19.5	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 48.204 ng
RT: 21.737 min Scan# 3174
Delta R.T. -0.011 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:149 Resp: 407169

Ion	Ratio	Lower	Upper
149	100		
167	25.7	20.0	30.0
279	4.0	3.1	4.7






#85
Di-n-octyl phthalate
Concen: 49.802 ng
RT: 23.000 min Scan# 3
Delta R.T. -0.011 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:149 Resp: 70923


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	13.1	10.8	16.2



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 25.192 min Scan# 3762
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:264 Resp: 311364

Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.0	28.4
265	21.6	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 45.114 ng
RT: 29.034 min Scan# 44
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:276 Resp: 103333


Ion	Ratio	Lower	Upper
276	100		
138	10.3	10.8	16.2#
277	25.3	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 45.715 ng
RT: 24.134 min Scan# 3582
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:252 Resp: 872834

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.8	26.8
125	6.2	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 45.401 ng
RT: 24.205 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:252 Resp: 86742


Ion	Ratio	Lower	Upper
252	100		
253	22.1	16.8	25.2
125	6.7	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 47.341 ng
RT: 25.039 min Scan# 3736
Delta R.T. -0.005 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:252 Resp: 830855

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	7.4	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 46.137 ng
RT: 29.099 min Scan# 4
Delta R.T. -0.011 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:278 Resp: 85854


Ion	Ratio	Lower	Upper
278	100		
139	9.8	8.0	12.0
279	24.0	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
PB170215BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 46.441 ng
RT: 30.233 min Scan# 4620
Delta R.T. 0.000 min
Lab File: BG064622.D
Acq: 29 Oct 2025 14:06

Tgt Ion:276 Resp: 825558

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.1	28.7
138	12.7	11.3	16.9



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170286BS	SDG No.:	Q3434
Lab Sample ID:	PB170286BS	Matrix:	Water
Analytical Method:	8270E Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C Prep Date: 10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	11.4		1	3.90	10.0	ug/L	10/30/25 12:58	PB170286
108-95-2	Phenol	38.9		1	0.91	5.00	ug/L	10/30/25 12:58	PB170286
111-44-4	bis(2-Chloroethyl)ether	35.9		1	0.81	5.00	ug/L	10/30/25 12:58	PB170286
95-57-8	2-Chlorophenol	43.7		1	0.58	5.00	ug/L	10/30/25 12:58	PB170286
95-48-7	2-Methylphenol	40.8		1	1.10	5.00	ug/L	10/30/25 12:58	PB170286
108-60-1	2,2-oxybis(1-Chloropropane)	35.8		1	1.30	5.00	ug/L	10/30/25 12:58	PB170286
98-86-2	Acetophenone	40.5		1	0.74	5.00	ug/L	10/30/25 12:58	PB170286
65794-96-9	3+4-Methylphenols	38.9		1	1.10	10.0	ug/L	10/30/25 12:58	PB170286
621-64-7	n-Nitroso-di-n-propylamine	38.1		1	1.40	2.50	ug/L	10/30/25 12:58	PB170286
67-72-1	Hexachloroethane	40.6		1	0.65	5.00	ug/L	10/30/25 12:58	PB170286
98-95-3	Nitrobenzene	43.9		1	0.76	5.00	ug/L	10/30/25 12:58	PB170286
78-59-1	Isophorone	37.1		1	0.75	5.00	ug/L	10/30/25 12:58	PB170286
88-75-5	2-Nitrophenol	47.8		1	1.80	5.00	ug/L	10/30/25 12:58	PB170286
105-67-9	2,4-Dimethylphenol	45.1		1	1.90	5.00	ug/L	10/30/25 12:58	PB170286
111-91-1	bis(2-Chloroethoxy)methane	38.1		1	0.68	5.00	ug/L	10/30/25 12:58	PB170286
120-83-2	2,4-Dichlorophenol	44.2		1	0.52	5.00	ug/L	10/30/25 12:58	PB170286
91-20-3	Naphthalene	38.0		1	0.50	5.00	ug/L	10/30/25 12:58	PB170286
106-47-8	4-Chloroaniline	19.7		1	0.84	5.00	ug/L	10/30/25 12:58	PB170286
87-68-3	Hexachlorobutadiene	40.7		1	0.54	5.00	ug/L	10/30/25 12:58	PB170286
105-60-2	Caprolactam	39.1		1	1.10	10.0	ug/L	10/30/25 12:58	PB170286
59-50-7	4-Chloro-3-methylphenol	41.8		1	0.59	5.00	ug/L	10/30/25 12:58	PB170286
91-57-6	2-Methylnaphthalene	34.6		1	0.56	5.00	ug/L	10/30/25 12:58	PB170286
77-47-4	Hexachlorocyclopentadiene	120	E	1	3.60	10.0	ug/L	10/30/25 12:58	PB170286
88-06-2	2,4,6-Trichlorophenol	47.1		1	0.51	5.00	ug/L	10/30/25 12:58	PB170286
95-95-4	2,4,5-Trichlorophenol	44.9		1	0.62	5.00	ug/L	10/30/25 12:58	PB170286
92-52-4	1,1-Biphenyl	42.2		1	0.53	5.00	ug/L	10/30/25 12:58	PB170286
91-58-7	2-Chloronaphthalene	39.4		1	0.61	5.00	ug/L	10/30/25 12:58	PB170286
88-74-4	2-Nitroaniline	45.5		1	1.30	5.00	ug/L	10/30/25 12:58	PB170286
131-11-3	Dimethylphthalate	40.5		1	0.61	5.00	ug/L	10/30/25 12:58	PB170286
208-96-8	Acenaphthylene	44.6		1	0.75	5.00	ug/L	10/30/25 12:58	PB170286
606-20-2	2,6-Dinitrotoluene	50.0		1	0.92	5.00	ug/L	10/30/25 12:58	PB170286
99-09-2	3-Nitroaniline	33.5		1	1.10	5.00	ug/L	10/30/25 12:58	PB170286
83-32-9	Acenaphthene	45.1		1	0.55	5.00	ug/L	10/30/25 12:58	PB170286
51-28-5	2,4-Dinitrophenol	110	E	1	6.00	10.0	ug/L	10/30/25 12:58	PB170286
100-02-7	4-Nitrophenol	95.8	E	1	2.40	10.0	ug/L	10/30/25 12:58	PB170286
132-64-9	Dibenzofuran	40.5		1	0.61	5.00	ug/L	10/30/25 12:58	PB170286
121-14-2	2,4-Dinitrotoluene	48.4		1	1.20	5.00	ug/L	10/30/25 12:58	PB170286
84-66-2	Diethylphthalate	39.9		1	0.69	5.00	ug/L	10/30/25 12:58	PB170286
7005-72-3	4-Chlorophenyl-phenylether	40.7		1	0.68	5.00	ug/L	10/30/25 12:58	PB170286
86-73-7	Fluorene	41.0		1	0.63	5.00	ug/L	10/30/25 12:58	PB170286
100-01-6	4-Nitroaniline	47.7		1	1.50	5.00	ug/L	10/30/25 12:58	PB170286
534-52-1	4,6-Dinitro-2-methylphenol	53.3		1	2.90	10.0	ug/L	10/30/25 12:58	PB170286

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM			Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206			Date Received:	
Client Sample ID:	PB170286BS			SDG No.:	Q3434
Lab Sample ID:	PB170286BS			Matrix:	Water
Analytical Method:	8270E	Level:	LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C	Prep Date:	10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	41.6		1	0.58	5.00	ug/L	10/30/25 12:58	PB170286
101-55-3	4-Bromophenyl-phenylether	41.5		1	0.40	5.00	ug/L	10/30/25 12:58	PB170286
118-74-1	Hexachlorobenzene	41.8		1	0.52	5.00	ug/L	10/30/25 12:58	PB170286
1912-24-9	Atrazine	42.1		1	1.00	5.00	ug/L	10/30/25 12:58	PB170286
87-86-5	Pentachlorophenol	93.8	E	1	1.60	10.0	ug/L	10/30/25 12:58	PB170286
85-01-8	Phenanthrene	40.4		1	0.50	5.00	ug/L	10/30/25 12:58	PB170286
120-12-7	Anthracene	40.2		1	0.61	5.00	ug/L	10/30/25 12:58	PB170286
86-74-8	Carbazole	41.3		1	0.72	5.00	ug/L	10/30/25 12:58	PB170286
84-74-2	Di-n-butylphthalate	39.4		1	1.20	5.00	ug/L	10/30/25 12:58	PB170286
206-44-0	Fluoranthene	41.6		1	0.82	5.00	ug/L	10/30/25 12:58	PB170286
129-00-0	Pyrene	38.0		1	0.50	5.00	ug/L	10/30/25 12:58	PB170286
85-68-7	Butylbenzylphthalate	43.4		1	1.90	5.00	ug/L	10/30/25 12:58	PB170286
91-94-1	3,3-Dichlorobenzidine	30.5		1	0.93	10.0	ug/L	10/30/25 12:58	PB170286
56-55-3	Benzo(a)anthracene	40.7		1	0.45	5.00	ug/L	10/30/25 12:58	PB170286
218-01-9	Chrysene	40.2		1	0.44	5.00	ug/L	10/30/25 12:58	PB170286
117-81-7	Bis(2-ethylhexyl)phthalate	42.9		1	1.60	5.00	ug/L	10/30/25 12:58	PB170286
117-84-0	Di-n-octyl phthalate	44.5		1	2.30	10.0	ug/L	10/30/25 12:58	PB170286
205-99-2	Benzo(b)fluoranthene	43.3		1	0.49	5.00	ug/L	10/30/25 12:58	PB170286
207-08-9	Benzo(k)fluoranthene	43.1		1	0.48	5.00	ug/L	10/30/25 12:58	PB170286
50-32-8	Benzo(a)pyrene	45.1		1	0.55	5.00	ug/L	10/30/25 12:58	PB170286
193-39-5	Indeno(1,2,3-cd)pyrene	43.3		1	0.59	5.00	ug/L	10/30/25 12:58	PB170286
53-70-3	Dibenzo(a,h)anthracene	43.7		1	0.67	5.00	ug/L	10/30/25 12:58	PB170286
191-24-2	Benzo(g,h,i)perylene	45.2		1	0.69	5.00	ug/L	10/30/25 12:58	PB170286
95-94-3	1,2,4,5-Tetrachlorobenzene	43.1		1	0.52	5.00	ug/L	10/30/25 12:58	PB170286
123-91-1	1,4-Dioxane	30.4		1	1.00	5.00	ug/L	10/30/25 12:58	PB170286
58-90-2	2,3,4,6-Tetrachlorophenol	50.0		1	0.72	5.00	ug/L	10/30/25 12:58	PB170286
SURROGATES									
367-12-4	2-Fluorophenol	121			23 - 138	81%	SPK: 150		
13127-88-3	Phenol-d6	116			10 - 134	78%	SPK: 150		
4165-60-0	Nitrobenzene-d5	93.9			67 - 132	94%	SPK: 100		
321-60-8	2-Fluorobiphenyl	79.9			52 - 132	80%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	144			44 - 137	96%	SPK: 150		
1718-51-0	Terphenyl-d14	79.5			42 - 152	79%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	30800							
1146-65-2	Naphthalene-d8	125000							
15067-26-2	Acenaphthene-d10	97000							
1517-22-2	Phenanthrene-d10	243000							
1719-03-5	Chrysene-d12	288000							
1520-96-3	Perylene-d12	302000							

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM			Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206			Date Received:	
Client Sample ID:	PB170286BS			SDG No.:	Q3434
Lab Sample ID:	PB170286BS			Matrix:	Water
Analytical Method:	8270E	Level:	LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C	Prep Date:	10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064638.D
Acq On : 30 Oct 2025 12:58
Operator : RC/JU
Sample : PB170286BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Nov 04 04:23:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.217	152	30784	20.000	ng	0.00
21) Naphthalene-d8	11.043	136	124677	20.000	ng	0.00
39) Acenaphthene-d10	14.839	164	97021	20.000	ng	0.00
64) Phenanthrene-d10	17.577	188	242858	20.000	ng	0.00
76) Chrysene-d12	21.854	240	288177	20.000	ng	0.00
86) Perylene-d12	25.197	264	302049	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.749	112	222225	121.168	ng	0.00
7) Phenol-d6	7.359	99	293062	116.449	ng	0.00
23) Nitrobenzene-d5	9.380	82	194877	93.881	ng	0.00
42) 2,4,6-Tribromophenol	16.319	330	201434	143.850	ng	0.00
45) 2-Fluorobiphenyl	13.470	172	511406	79.896	ng	0.00
79) Terphenyl-d14	20.168	244	1213413	79.462	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.581	88	24696	30.415	ng	92
3) Pyridine	3.987	79	71577	29.354	ng	95
4) n-Nitrosodimethylamine	3.898	42	50644	38.730	ng	92
6) Aniline	7.529	93	88719	25.907	ng	96
8) 2-Chlorophenol	7.776	128	83962	43.738	ng	96
9) Benzaldehyde	7.336	77	15716	11.439	ng	98
10) Phenol	7.383	94	108355	38.932	ng	94
11) bis(2-Chloroethyl)ether	7.623	93	73434	35.887	ng	94
12) 1,3-Dichlorobenzene	8.111	146	86606	39.198	ng	94
13) 1,4-Dichlorobenzene	8.252	146	85572	38.090	ng	97
14) 1,2-Dichlorobenzene	8.575	146	83239	38.414	ng	95
15) Benzyl Alcohol	8.446	79	82132	39.576	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.746	45	192190	35.806	ng	99
17) 2-Methylphenol	8.646	107	75352	40.777	ng	95
18) Hexachloroethane	9.321	117	32047	40.649	ng	95
19) n-Nitroso-di-n-propyla...	9.022	70	66214	38.070	ng	97
20) 3+4-Methylphenols	8.975	107	101160	38.927	ng	98
22) Acetophenone	9.039	105	138177	40.543	ng	97
24) Nitrobenzene	9.421	77	99054	43.873	ng	99
25) Isophorone	9.950	82	184933	37.128	ng	98
26) 2-Nitrophenol	10.144	139	41607	47.778	ng	90
27) 2,4-Dimethylphenol	10.191	122	83626	45.136	ng	96
28) bis(2-Chloroethoxy)met...	10.432	93	108016	38.088	ng	97
29) 2,4-Dichlorophenol	10.679	162	88062	44.165	ng	95
30) 1,2,4-Trichlorobenzene	10.902	180	95114	41.983	ng	99
31) Naphthalene	11.096	128	262425	37.964	ng	100
32) Benzoic acid	10.309	122	63329	45.241	ng	96
33) 4-Chloroaniline	11.184	127	52856	19.673	ng	97
34) Hexachlorobutadiene	11.390	225	72053	40.668	ng	94
35) Caprolactam	11.942	113	29879m	39.070	ng	
36) 4-Chloro-3-methylphenol	12.289	107	100287	41.843	ng	99
37) 2-Methylnaphthalene	12.688	142	175641	34.608	ng	99
38) 1-Methylnaphthalene	12.906	142	188034	37.426	ng	92
40) 1,2,4,5-Tetrachloroben...	13.052	216	140180	43.052	ng	99
41) Hexachlorocyclopentadiene	13.035	237	172388	117.624	ng	100
43) 2,4,6-Trichlorophenol	13.276	196	87768	47.123	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064638.D
Acq On : 30 Oct 2025 12:58
Operator : RC/JU
Sample : PB170286BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Nov 04 04:23:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.346	196	96559	44.925	ng		96
46) 1,1'-Biphenyl	13.681	154	291306	42.233	ng		97
47) 2-Chloronaphthalene	13.722	162	204969	39.368	ng		100
48) 2-Nitroaniline	13.910	65	79131	45.457	ng		94
49) Acenaphthylene	14.568	152	366587	44.639	ng		99
50) Dimethylphthalate	14.286	163	300984	40.542	ng		99
51) 2,6-Dinitrotoluene	14.398	165	61387	49.972	ng		99
52) Acenaphthene	14.903	154	253184	45.125	ng		96
53) 3-Nitroaniline	14.721	138	42706	33.472	ng		94
54) 2,4-Dinitrophenol	14.927	184	83799	109.290	ng	#	43
55) Dibenzofuran	15.238	168	366959	40.455	ng		99
56) 4-Nitrophenol	15.021	139	113344	95.819	ng		90
57) 2,4-Dinitrotoluene	15.179	165	91677	48.378	ng	#	100
58) Fluorene	15.884	166	290135	40.956	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.455	232	102856	50.044	ng		99
60) Diethylphthalate	15.644	149	302705	39.933	ng		98
61) 4-Chlorophenyl-phenyle...	15.873	204	161828	40.706	ng		97
62) 4-Nitroaniline	15.884	138	67726	47.714	ng		94
63) Azobenzene	16.161	77	278632	40.061	ng		100
65) 4,6-Dinitro-2-methylph...	15.943	198	57652	53.276	ng		94
66) n-Nitrosodiphenylamine	16.078	169	272742	41.575	ng		98
67) 4-Bromophenyl-phenylether	16.766	248	119558	41.543	ng		99
68) Hexachlorobenzene	16.889	284	137259	41.805	ng		95
69) Atrazine	17.018	200	118963	42.139	ng		96
70) Pentachlorophenol	17.224	266	191995	93.775	ng		97
71) Phenanthrene	17.618	178	536917	40.428	ng		99
72) Anthracene	17.712	178	543314	40.216	ng		100
73) Carbazole	17.970	167	491851	41.286	ng		99
74) Di-n-butylphthalate	18.528	149	543306	39.438	ng		99
75) Fluoranthene	19.615	202	691810	41.581	ng		98
77) Benzidine	19.780	184	165366	25.613	ng		99
78) Pyrene	19.974	202	715895	38.023	ng		99
80) Butylbenzylphthalate	20.849	149	246800	43.371	ng		95
81) Benzo(a)anthracene	21.830	228	777040	40.666	ng		99
82) 3,3'-Dichlorobenzidine	21.736	252	194405	30.468	ng		99
83) Chrysene	21.901	228	730189	40.161	ng		98
84) Bis(2-ethylhexyl)phtha...	21.742	149	364913	42.916	ng		99
85) Di-n-octyl phthalate	23.000	149	638034	44.506	ng		98
87) Indeno(1,2,3-cd)pyrene	29.028	276	962160	43.302	ng	#	96
88) Benzo(b)fluoranthene	24.134	252	801805	43.290	ng		99
89) Benzo(k)fluoranthene	24.204	252	798213	43.067	ng		99
90) Benzo(a)pyrene	25.044	252	768387	45.132	ng		98
91) Dibenzo(a,h)anthracene	29.104	278	789221	43.720	ng		97
92) Benzo(g,h,i)perylene	30.232	276	779694	45.214	ng		97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064638.D
Acq On : 30 Oct 2025 12:58
Operator : RC/JU
Sample : PB170286BS
Misc :
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Nov 04 04:23:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.217 min Scan# 8
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:152 Resp: 3078


Ion	Ratio	Lower	Upper
152	100		
150	153.8	122.2	183.4
115	64.9	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 30.415 ng
RT: 3.581 min Scan# 84
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 88 Resp: 24696

Ion	Ratio	Lower	Upper
88	100		
58	123.6	92.4	138.6
43	46.2	33.0	49.6






#3
Pyridine
Concen: 29.354 ng
RT: 3.987 min Scan# 1
Delta R.T. -0.012 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 79 Resp: 7157


Ion	Ratio	Lower	Upper
79	100		
52	111.6	94.2	141.2
51	53.7	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 38.730 ng
RT: 3.898 min Scan# 138
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 42 Resp: 50644

Ion	Ratio	Lower	Upper
42	100		
74	103.3	76.1	114.1
44	7.8	6.3	9.5





#5
2-Fluorophenol
Concen: 121.168 ng
RT: 5.749 min Scan# 4
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58


Tgt Ion:112 Resp: 22222


Ion	Ratio	Lower	Upper
112	100		
64	80.3	60.7	91.1
63	44.4	33.4	50.0

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#6
Aniline
Concen: 25.907 ng
RT: 7.529 min Scan# 756
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 93 Resp: 88719

Ion	Ratio	Lower	Upper
93	100		
66	48.3	37.0	55.6
65	25.1	18.3	27.5






#7
Phenol-d6
Concen: 116.449 ng
RT: 7.359 min Scan# 7
Delta R.T. 0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 99 Resp: 29306


Ion	Ratio	Lower	Upper
99	100		
42	23.6	18.3	27.5
71	37.6	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 43.738 ng
RT: 7.776 min Scan# 798
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:128 Resp: 83962

Ion	Ratio	Lower	Upper
128	100		
130	31.9	13.5	53.5
64	62.1	45.2	85.2






#9
Benzaldehyde
Concen: 11.439 ng
RT: 7.336 min Scan# 7
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 77 Resp: 1571


Ion	Ratio	Lower	Upper
77	100		
105	91.3	72.9	112.9
106	88.1	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 38.932 ng
RT: 7.383 min Scan# 731
Delta R.T. 0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 94 Resp: 108355

Ion	Ratio	Lower	Upper
94	100		
65	35.8	15.8	55.8
66	50.5	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 35.887 ng
RT: 7.623 min Scan# 7
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 93 Resp: 7343


Ion	Ratio	Lower	Upper
93	100		
63	103.5	91.3	131.3
95	33.3	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 39.198 ng
RT: 8.111 min Scan# 855
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:146 Resp: 86606

Ion	Ratio	Lower	Upper
146	100		
148	64.5	55.3	82.9
75	32.1	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 38.090 ng
RT: 8.252 min Scan# 8
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:146 Resp: 8557


Ion	Ratio	Lower	Upper
146	100		
148	67.6	56.2	84.4
111	45.0	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 38.414 ng
RT: 8.575 min Scan# 934
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:146 Resp: 83239

Ion	Ratio	Lower	Upper
146	100		
148	70.1	51.5	77.3
111	44.5	35.0	52.6





#15
Benzyl Alcohol
Concen: 39.576 ng
RT: 8.446 min Scan# 9
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58


Tgt Ion: 79 Resp: 8213


Ion	Ratio	Lower	Upper
79	100		
108	71.3	58.1	87.1
77	67.0	52.8	79.2

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 35.806 ng
RT: 8.746 min Scan# 963
Delta R.T. 0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 45 Resp: 192190

Ion	Ratio	Lower	Upper
45	100		
77	8.6	0.0	28.5
79	6.0	0.0	25.6






#17
2-Methylphenol
Concen: 40.777 ng
RT: 8.646 min Scan# 94
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:107 Resp: 7535


Ion	Ratio	Lower	Upper
107	100		
108	105.0	90.0	135.0
77	48.5	37.8	56.8
79	47.3	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 40.649 ng
RT: 9.321 min Scan# 1061
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:117 Resp: 32047

Ion	Ratio	Lower	Upper
117	100		
119	99.9	75.4	113.2
201	129.7	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 38.070 ng
RT: 9.022 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 70 Resp: 6621


Ion	Ratio	Lower	Upper
70	100		
42	78.5	65.2	97.8
101	9.8	7.9	11.9
130	16.9	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 38.927 ng
RT: 8.975 min Scan# 1002
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:107 Resp: 101160

Ion	Ratio	Lower	Upper
107	100		
108	89.7	70.3	110.3
77	37.0	14.6	54.6
79	30.7	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.043 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:136 Resp: 12467


Ion	Ratio	Lower	Upper
136	100		
137	11.9	9.2	13.8
54	14.9	12.6	19.0
68	6.1	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 40.543 ng
RT: 9.039 min Scan# 1013
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:105 Resp: 138177

Ion	Ratio	Lower	Upper
105	100		
71	1.9	1.5	2.3
51	40.1	33.7	50.5
120	19.2	16.1	24.1






#23
Nitrobenzene-d5
Concen: 93.881 ng
RT: 9.380 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 82 Resp: 19487


Ion	Ratio	Lower	Upper
82	100		
128	38.3	31.3	46.9
54	71.9	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 43.873 ng
RT: 9.421 min Scan# 1078
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 77 Resp: 99054

Ion	Ratio	Lower	Upper
77	100		
123	39.9	31.0	46.6
65	15.6	12.7	19.1






#25
Isophorone
Concen: 37.128 ng
RT: 9.950 min Scan# 11
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 82 Resp: 18493


Ion	Ratio	Lower	Upper
82	100		
95	7.4	6.1	9.1
138	15.1	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 47.778 ng
RT: 10.144 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:139 Resp: 41607

Ion	Ratio	Lower	Upper
139	100		
109	36.9	24.6	37.0
65	66.0	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 45.136 ng
RT: 10.191 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:122 Resp: 8362


Ion	Ratio	Lower	Upper
122	100		
107	123.1	103.0	154.4
121	61.1	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 38.088 ng
RT: 10.432 min Scan# 1250
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 93 Resp: 108016

Ion	Ratio	Lower	Upper
93	100		
95	30.1	25.4	38.2
123	9.9	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 44.165 ng
RT: 10.679 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:162 Resp: 8806


Ion	Ratio	Lower	Upper
162	100		
164	65.6	48.7	88.7
98	33.1	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 41.983 ng
RT: 10.902 min Scan# 1330
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:180 Resp: 95114

Ion	Ratio	Lower	Upper
180	100		
182	94.4	76.9	115.3
145	27.3	22.1	33.1






#31
Naphthalene
Concen: 37.964 ng
RT: 11.096 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:128 Resp: 26242


Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.8	13.2
127	13.3	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 45.241 ng
RT: 10.309 min Scan# 1229
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:122 Resp: 63329

Ion	Ratio	Lower	Upper
122	100		
105	124.4	104.2	144.2
77	100.0	70.2	110.2






#33
4-Chloroaniline
Concen: 19.673 ng
RT: 11.184 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:127 Resp: 5285


Ion	Ratio	Lower	Upper
127	100		
129	31.1	25.5	38.3
65	38.2	32.2	48.4
92	21.4	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 40.668 ng
RT: 11.390 min Scan# 1413
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:225 Resp: 72053

Ion	Ratio	Lower	Upper
225	100		
223	57.0	48.8	73.2
227	66.5	49.5	74.3






#35
Caprolactam
Concen: 39.070 ng m
RT: 11.942 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:113 Resp: 2987


Ion	Ratio	Lower	Upper
113	100		
55	314.2	290.1	330.1
56	227.3	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 41.843 ng
RT: 12.289 min Scan# 1566
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:107 Resp: 100287

Ion	Ratio	Lower	Upper
107	100		
144	23.1	19.1	28.7
142	73.7	58.6	87.8






#37
2-Methylnaphthalene
Concen: 34.608 ng
RT: 12.688 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:142 Resp: 17564


Ion	Ratio	Lower	Upper
142	100		
141	86.2	68.9	103.3
115	40.8	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 37.426 ng
RT: 12.906 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:142 Resp: 188034

Ion	Ratio	Lower	Upper
142	100		
141	83.9	73.3	109.9
116	4.4	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.839 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:164 Resp: 9702


Ion	Ratio	Lower	Upper
164	100		
162	98.5	77.0	115.6
160	43.5	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 43.052 ng
RT: 13.052 min Scan# 1696
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:216 Resp: 140180

Ion	Ratio	Lower	Upper
216	100		
214	76.4	61.4	92.2
179	18.6	15.4	23.0
108	15.7	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 117.624 ng
RT: 13.035 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:237 Resp: 17238


Ion	Ratio	Lower	Upper
237	100		
235	64.2	44.1	84.1
272	12.2	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 143.850 ng
RT: 16.319 min Scan# 2252
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:330 Resp: 201434

Ion	Ratio	Lower	Upper
330	100		
332	97.2	78.0	117.0
141	27.8	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 47.123 ng
RT: 13.276 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:196 Resp: 8776


Ion	Ratio	Lower	Upper
196	100		
198	91.9	76.6	115.0
200	29.5	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 44.925 ng
RT: 13.346 min Scan# 1746
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:196 Resp: 96559

Ion	Ratio	Lower	Upper
196	100		
198	97.1	74.2	111.4
97	42.5	37.0	55.6
132	24.0	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 79.896 ng
RT: 13.470 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:172 Resp: 51140


Ion	Ratio	Lower	Upper
172	100		
171	36.0	29.4	44.0
170	24.8	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 42.233 ng
RT: 13.681 min Scan# 1803
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:154 Resp: 291306

Ion	Ratio	Lower	Upper
154	100		
153	40.7	22.4	62.4
76	11.8	0.0	33.5






#47
2-Chloronaphthalene
Concen: 39.368 ng
RT: 13.722 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:162 Resp: 20496


Ion	Ratio	Lower	Upper
162	100		
127	40.1	32.2	48.2
164	33.9	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 45.457 ng
RT: 13.910 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 65 Resp: 79131

Ion	Ratio	Lower	Upper
65	100		
92	56.2	48.3	72.5
138	87.2	74.9	112.3






#49
Acenaphthylene
Concen: 44.639 ng
RT: 14.568 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:152 Resp: 36658


Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.4	23.2
153	12.9	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 40.542 ng
RT: 14.286 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:163 Resp: 300984

Ion	Ratio	Lower	Upper
163	100		
194	4.0	2.8	4.2
164	9.7	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 49.972 ng
RT: 14.398 min Scan# 1
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:165 Resp: 6138


Ion	Ratio	Lower	Upper
165	100		
63	70.7	57.4	86.0
89	59.6	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 45.125 ng
RT: 14.903 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:154 Resp: 253184

Ion	Ratio	Lower	Upper
154	100		
153	120.8	92.2	138.4
152	55.4	43.2	64.8





#53
3-Nitroaniline
Concen: 33.472 ng
RT: 14.721 min Scan# 1
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58


Tgt Ion:138 Resp: 4270


Ion	Ratio	Lower	Upper
138	100		
108	12.3	11.4	17.0
92	126.5	95.3	142.9

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#54
2,4-Dinitrophenol
Concen: 109.290 ng
RT: 14.927 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:184 Resp: 83799

Ion	Ratio	Lower	Upper
184	100		
63	70.0	56.6	85.0
154	81.6	158.6	238.0#






#55
Dibenzofuran
Concen: 40.455 ng
RT: 15.238 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:168 Resp: 36695


Ion	Ratio	Lower	Upper
168	100		
139	35.4	27.8	41.8
169	12.7	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 95.819 ng
RT: 15.021 min Scan# 2031
Delta R.T. 0.012 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:139 Resp: 113344

Ion	Ratio	Lower	Upper
139	100		
109	99.6	67.8	107.8
65	132.4	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 48.378 ng
RT: 15.179 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:165 Resp: 9167


Ion	Ratio	Lower	Upper
165	100		
63	52.4	41.8	62.8
89	72.1	58.1	87.1
182	2.6	1.4	2.0#



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 40.956 ng
RT: 15.884 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:166 Resp: 290135

Ion	Ratio	Lower	Upper
166	100		
165	101.0	81.5	122.3
167	13.0	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 50.044 ng
RT: 15.455 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:232 Resp: 10285


Ion	Ratio	Lower	Upper
232	100		
131	33.9	26.7	40.1
130	2.1	1.6	2.4
166	25.3	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 39.933 ng
RT: 15.644 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:149 Resp: 302705

Ion	Ratio	Lower	Upper
149	100		
177	23.4	18.0	27.0
150	13.1	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 40.706 ng
RT: 15.873 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:204 Resp: 16182


Ion	Ratio	Lower	Upper
204	100		
206	30.4	25.6	38.4
141	48.3	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 47.714 ng
RT: 15.884 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:138 Resp: 67726

Ion	Ratio	Lower	Upper
138	100		
92	54.3	36.1	76.1
108	103.9	75.6	115.6






#63
Azobenzene
Concen: 40.061 ng
RT: 16.161 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion: 77 Resp: 27863


Ion	Ratio	Lower	Upper
77	100		
182	25.1	4.9	44.9
105	21.2	1.3	41.3
51	38.7	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.577 min Scan# 2466
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:188 Resp: 242858

Ion	Ratio	Lower	Upper
188	100		
94	7.3	5.7	8.5
80	8.6	6.8	10.2





#65
4,6-Dinitro-2-methylphenol
Concen: 53.276 ng
RT: 15.943 min Scan# 2
Delta R.T. 0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58


Tgt Ion:198 Resp: 5765


Ion	Ratio	Lower	Upper
198	100		
51	49.4	24.7	64.7
105	35.5	18.0	58.0

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#66
n-Nitrosodiphenylamine
Concen: 41.575 ng
RT: 16.078 min Scan# 2211
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:169 Resp: 272742

Ion	Ratio	Lower	Upper
169	100		
168	69.3	56.3	84.5
167	35.1	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 41.543 ng
RT: 16.766 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:248 Resp: 11955


Ion	Ratio	Lower	Upper
248	100		
250	96.9	77.7	116.5
141	58.8	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 41.805 ng
RT: 16.889 min Scan# 2349
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:284 Resp: 137259

Ion	Ratio	Lower	Upper
284	100		
142	23.5	21.0	31.6
249	33.3	24.9	37.3






#69
Atrazine
Concen: 42.139 ng
RT: 17.018 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:200 Resp: 11896


Ion	Ratio	Lower	Upper
200	100		
173	24.9	0.9	40.9
215	45.8	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 93.775 ng
RT: 17.224 min Scan# 2406
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:266 Resp: 191995

Ion	Ratio	Lower	Upper
266	100		
268	64.1	50.3	75.5
264	61.8	47.3	70.9






#71
Phenanthrene
Concen: 40.428 ng
RT: 17.618 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:178 Resp: 53691


Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.7	22.1
179	15.1	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 40.216 ng
RT: 17.712 min Scan# 2489
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:178 Resp: 543314

Ion	Ratio	Lower	Upper
178	100		
176	18.0	14.4	21.6
179	15.3	12.4	18.6






#73
Carbazole
Concen: 41.286 ng
RT: 17.970 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:167 Resp: 49185


Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.9	26.9
139	11.7	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 39.438 ng
RT: 18.528 min Scan# 2628
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:149 Resp: 543306

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.0	10.6
104	5.6	4.1	6.1






#75
Fluoranthene
Concen: 41.581 ng
RT: 19.615 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:202 Resp: 69181


Ion	Ratio	Lower	Upper
202	100		
101	7.3	0.0	28.7
203	17.4	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.854 min Scan# 3194
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:240 Resp: 288177

Ion	Ratio	Lower	Upper
240	100		
120	6.5	6.5	9.7
236	26.1	20.2	30.2






#77
Benzidine
Concen: 25.613 ng
RT: 19.780 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:184 Resp: 16536


Ion	Ratio	Lower	Upper
184	100		
185	13.8	11.4	17.2
183	11.7	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 38.023 ng
RT: 19.974 min Scan# 2874
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:202 Resp: 715895

Ion	Ratio	Lower	Upper
202	100		
200	21.3	17.0	25.4
203	17.6	13.8	20.6





#79
Terphenyl-d14
Concen: 79.462 ng
RT: 20.168 min Scan# 2
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58


Tgt Ion:244 Resp: 121341


Ion	Ratio	Lower	Upper
244	100		
212	7.4	6.1	9.1
122	6.8	5.6	8.4

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#80
Butylbenzylphthalate
Concen: 43.371 ng
RT: 20.849 min Scan# 3023
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:149 Resp: 246800

Ion	Ratio	Lower	Upper
149	100		
91	69.4	59.2	88.8
206	23.9	18.1	27.1






#81
Benzo(a)anthracene
Concen: 40.666 ng
RT: 21.830 min Scan# 3
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:228 Resp: 77704


Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.9	32.9
229	19.5	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 30.468 ng
RT: 21.736 min Scan# 3174
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:252 Resp: 194405

Ion	Ratio	Lower	Upper
252	100		
254	62.9	50.8	76.2
126	6.3	5.8	8.6






#83
Chrysene
Concen: 40.161 ng
RT: 21.901 min Scan# 3
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:228 Resp: 73018


Ion	Ratio	Lower	Upper
228	100		
226	30.4	23.3	34.9
229	20.0	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 42.916 ng
RT: 21.742 min Scan# 3175
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:149 Resp: 364913

Ion	Ratio	Lower	Upper
149	100		
167	25.8	20.0	30.0
279	4.1	3.1	4.7






#85
Di-n-octyl phthalate
Concen: 44.506 ng
RT: 23.000 min Scan# 3
Delta R.T. -0.012 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:149 Resp: 63803


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.3	1.9
43	12.7	10.8	16.2



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 25.197 min Scan# 3763
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:264 Resp: 302049

Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.0	28.4
265	21.2	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 43.302 ng
RT: 29.028 min Scan# 4
Delta R.T. -0.012 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58


Tgt Ion:276 Resp: 96216

Ion	Ratio	Lower	Upper
276	100		
138	9.2	10.8	16.2
277	25.2	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
PB170286BS


Manual Integrations APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 43.290 ng
RT: 24.134 min Scan# 3582
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:252 Resp: 801805

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.8	26.8
125	6.2	5.6	8.4





#89
Benzo(k)fluoranthene
Concen: 43.067 ng
RT: 24.204 min Scan# 3
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58


Tgt Ion:252 Resp: 79821


Ion	Ratio	Lower	Upper
252	100		
253	21.4	16.8	25.2
125	6.3	5.4	8.0

Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#90
Benzo(a)pyrene
Concen: 45.132 ng
RT: 25.044 min Scan# 3737
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:252 Resp: 768387

Ion	Ratio	Lower	Upper
252	100		
253	21.2	17.6	26.4
125	6.9	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 43.720 ng
RT: 29.104 min Scan# 44
Delta R.T. -0.006 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:278 Resp: 78922


Ion	Ratio	Lower	Upper
278	100		
139	8.8	8.0	12.0
279	23.4	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
PB170286BS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 45.214 ng
RT: 30.232 min Scan# 4620
Delta R.T. -0.000 min
Lab File: BG064638.D
Acq: 30 Oct 2025 12:58

Tgt Ion:276 Resp: 779694

Ion	Ratio	Lower	Upper
276	100		
277	22.5	19.1	28.7
138	13.1	11.3	16.9



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170286BSD	SDG No.:	Q3434
Lab Sample ID:	PB170286BSD	Matrix:	Water
Analytical Method:	8270E Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C Prep Date: 10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	12.8		1	3.90	10.0	ug/L	10/30/25 13:39	PB170286
108-95-2	Phenol	40.5		1	0.91	5.00	ug/L	10/30/25 13:39	PB170286
111-44-4	bis(2-Chloroethyl)ether	36.7		1	0.81	5.00	ug/L	10/30/25 13:39	PB170286
95-57-8	2-Chlorophenol	42.6		1	0.58	5.00	ug/L	10/30/25 13:39	PB170286
95-48-7	2-Methylphenol	41.6		1	1.10	5.00	ug/L	10/30/25 13:39	PB170286
108-60-1	2,2-oxybis(1-Chloropropane)	36.9		1	1.30	5.00	ug/L	10/30/25 13:39	PB170286
98-86-2	Acetophenone	42.7		1	0.74	5.00	ug/L	10/30/25 13:39	PB170286
65794-96-9	3+4-Methylphenols	40.6		1	1.10	10.0	ug/L	10/30/25 13:39	PB170286
621-64-7	n-Nitroso-di-n-propylamine	40.3		1	1.40	2.50	ug/L	10/30/25 13:39	PB170286
67-72-1	Hexachloroethane	42.4		1	0.65	5.00	ug/L	10/30/25 13:39	PB170286
98-95-3	Nitrobenzene	45.1		1	0.76	5.00	ug/L	10/30/25 13:39	PB170286
78-59-1	Isophorone	39.1		1	0.75	5.00	ug/L	10/30/25 13:39	PB170286
88-75-5	2-Nitrophenol	51.5		1	1.80	5.00	ug/L	10/30/25 13:39	PB170286
105-67-9	2,4-Dimethylphenol	47.6		1	1.90	5.00	ug/L	10/30/25 13:39	PB170286
111-91-1	bis(2-Chloroethoxy)methane	39.6		1	0.68	5.00	ug/L	10/30/25 13:39	PB170286
120-83-2	2,4-Dichlorophenol	45.9		1	0.52	5.00	ug/L	10/30/25 13:39	PB170286
91-20-3	Naphthalene	39.4		1	0.50	5.00	ug/L	10/30/25 13:39	PB170286
106-47-8	4-Chloroaniline	21.2		1	0.84	5.00	ug/L	10/30/25 13:39	PB170286
87-68-3	Hexachlorobutadiene	43.1		1	0.54	5.00	ug/L	10/30/25 13:39	PB170286
105-60-2	Caprolactam	40.3		1	1.10	10.0	ug/L	10/30/25 13:39	PB170286
59-50-7	4-Chloro-3-methylphenol	43.6		1	0.59	5.00	ug/L	10/30/25 13:39	PB170286
91-57-6	2-Methylnaphthalene	36.4		1	0.56	5.00	ug/L	10/30/25 13:39	PB170286
77-47-4	Hexachlorocyclopentadiene	120	E	1	3.60	10.0	ug/L	10/30/25 13:39	PB170286
88-06-2	2,4,6-Trichlorophenol	46.3		1	0.51	5.00	ug/L	10/30/25 13:39	PB170286
95-95-4	2,4,5-Trichlorophenol	44.9		1	0.62	5.00	ug/L	10/30/25 13:39	PB170286
92-52-4	1,1-Biphenyl	41.0		1	0.53	5.00	ug/L	10/30/25 13:39	PB170286
91-58-7	2-Chloronaphthalene	38.9		1	0.61	5.00	ug/L	10/30/25 13:39	PB170286
88-74-4	2-Nitroaniline	45.9		1	1.30	5.00	ug/L	10/30/25 13:39	PB170286
131-11-3	Dimethylphthalate	40.2		1	0.61	5.00	ug/L	10/30/25 13:39	PB170286
208-96-8	Acenaphthylene	44.5		1	0.75	5.00	ug/L	10/30/25 13:39	PB170286
606-20-2	2,6-Dinitrotoluene	49.0		1	0.92	5.00	ug/L	10/30/25 13:39	PB170286
99-09-2	3-Nitroaniline	33.1		1	1.10	5.00	ug/L	10/30/25 13:39	PB170286
83-32-9	Acenaphthene	45.3		1	0.55	5.00	ug/L	10/30/25 13:39	PB170286
51-28-5	2,4-Dinitrophenol	110	E	1	6.00	10.0	ug/L	10/30/25 13:39	PB170286
100-02-7	4-Nitrophenol	95.8	E	1	2.40	10.0	ug/L	10/30/25 13:39	PB170286
132-64-9	Dibenzofuran	39.8		1	0.61	5.00	ug/L	10/30/25 13:39	PB170286
121-14-2	2,4-Dinitrotoluene	47.9		1	1.20	5.00	ug/L	10/30/25 13:39	PB170286
84-66-2	Diethylphthalate	38.7		1	0.69	5.00	ug/L	10/30/25 13:39	PB170286
7005-72-3	4-Chlorophenyl-phenylether	40.1		1	0.68	5.00	ug/L	10/30/25 13:39	PB170286
86-73-7	Fluorene	40.0		1	0.63	5.00	ug/L	10/30/25 13:39	PB170286
100-01-6	4-Nitroaniline	47.3		1	1.50	5.00	ug/L	10/30/25 13:39	PB170286
534-52-1	4,6-Dinitro-2-methylphenol	55.1		1	2.90	10.0	ug/L	10/30/25 13:39	PB170286

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID:	PB170286BSD	SDG No.:	Q3434
Lab Sample ID:	PB170286BSD	Matrix:	Water
Analytical Method:	8270E Level: LOW	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C Prep Date: 10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	41.8		1	0.58	5.00	ug/L	10/30/25 13:39	PB170286
101-55-3	4-Bromophenyl-phenylether	41.3		1	0.40	5.00	ug/L	10/30/25 13:39	PB170286
118-74-1	Hexachlorobenzene	41.3		1	0.52	5.00	ug/L	10/30/25 13:39	PB170286
1912-24-9	Atrazine	42.2		1	1.00	5.00	ug/L	10/30/25 13:39	PB170286
87-86-5	Pentachlorophenol	93.3	E	1	1.60	10.0	ug/L	10/30/25 13:39	PB170286
85-01-8	Phenanthrene	40.3		1	0.50	5.00	ug/L	10/30/25 13:39	PB170286
120-12-7	Anthracene	40.1		1	0.61	5.00	ug/L	10/30/25 13:39	PB170286
86-74-8	Carbazole	41.3		1	0.72	5.00	ug/L	10/30/25 13:39	PB170286
84-74-2	Di-n-butylphthalate	39.2		1	1.20	5.00	ug/L	10/30/25 13:39	PB170286
206-44-0	Fluoranthene	41.3		1	0.82	5.00	ug/L	10/30/25 13:39	PB170286
129-00-0	Pyrene	37.9		1	0.50	5.00	ug/L	10/30/25 13:39	PB170286
85-68-7	Butylbenzylphthalate	43.3		1	1.90	5.00	ug/L	10/30/25 13:39	PB170286
91-94-1	3,3-Dichlorobenzidine	31.1		1	0.93	10.0	ug/L	10/30/25 13:39	PB170286
56-55-3	Benzo(a)anthracene	40.7		1	0.45	5.00	ug/L	10/30/25 13:39	PB170286
218-01-9	Chrysene	40.5		1	0.44	5.00	ug/L	10/30/25 13:39	PB170286
117-81-7	Bis(2-ethylhexyl)phthalate	42.3		1	1.60	5.00	ug/L	10/30/25 13:39	PB170286
117-84-0	Di-n-octyl phthalate	44.3		1	2.30	10.0	ug/L	10/30/25 13:39	PB170286
205-99-2	Benzo(b)fluoranthene	43.9		1	0.49	5.00	ug/L	10/30/25 13:39	PB170286
207-08-9	Benzo(k)fluoranthene	43.0		1	0.48	5.00	ug/L	10/30/25 13:39	PB170286
50-32-8	Benzo(a)pyrene	44.8		1	0.55	5.00	ug/L	10/30/25 13:39	PB170286
193-39-5	Indeno(1,2,3-cd)pyrene	43.3		1	0.59	5.00	ug/L	10/30/25 13:39	PB170286
53-70-3	Dibenzo(a,h)anthracene	44.4		1	0.67	5.00	ug/L	10/30/25 13:39	PB170286
191-24-2	Benzo(g,h,i)perylene	44.6		1	0.69	5.00	ug/L	10/30/25 13:39	PB170286
95-94-3	1,2,4,5-Tetrachlorobenzene	42.2		1	0.52	5.00	ug/L	10/30/25 13:39	PB170286
123-91-1	1,4-Dioxane	31.6		1	1.00	5.00	ug/L	10/30/25 13:39	PB170286
58-90-2	2,3,4,6-Tetrachlorophenol	49.8		1	0.72	5.00	ug/L	10/30/25 13:39	PB170286
SURROGATES									
367-12-4	2-Fluorophenol	122			23 - 138	82%	SPK: 150		
13127-88-3	Phenol-d6	122			10 - 134	81%	SPK: 150		
4165-60-0	Nitrobenzene-d5	97.0			67 - 132	97%	SPK: 100		
321-60-8	2-Fluorobiphenyl	79.8			52 - 132	80%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	143			44 - 137	95%	SPK: 150		
1718-51-0	Terphenyl-d14	80.6			42 - 152	81%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	30200							
1146-65-2	Naphthalene-d8	122000							
15067-26-2	Acenaphthene-d10	101000							
1517-22-2	Phenanthrene-d10	246000							
1719-03-5	Chrysene-d12	289000							
1520-96-3	Perylene-d12	302000							

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM			Date Collected:	
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206			Date Received:	
Client Sample ID:	PB170286BSD			SDG No.:	Q3434
Lab Sample ID:	PB170286BSD			Matrix:	Water
Analytical Method:	8270E	Level:	LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3510C	Prep Date:	10/28/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064639.D
Acq On : 30 Oct 2025 13:39
Operator : RC/JU
Sample : PB170286BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Nov 04 04:25:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.218	152	30191	20.000	ng		0.00
21) Naphthalene-d8	11.044	136	121649	20.000	ng		0.00
39) Acenaphthene-d10	14.839	164	100532	20.000	ng		0.00
64) Phenanthrene-d10	17.577	188	246375	20.000	ng		0.00
76) Chrysene-d12	21.855	240	288758	20.000	ng		0.00
86) Perylene-d12	25.192	264	302411	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.750	112	220132	122.384	ng		0.00
7) Phenol-d6	7.354	99	301103	121.994	ng		0.00
23) Nitrobenzene-d5	9.381	82	196502	97.020	ng		0.00
42) 2,4,6-Tribromophenol	16.320	330	207601	143.076	ng		0.00
45) 2-Fluorobiphenyl	13.470	172	529120	79.777	ng		0.00
79) Terphenyl-d14	20.168	244	1233810	80.635	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.582	88	25147	31.579	ng		92
3) Pyridine	3.987	79	72694	30.398	ng		98
4) n-Nitrosodimethylamine	3.899	42	52092	40.620	ng		98
6) Aniline	7.530	93	87639	26.094	ng		97
8) 2-Chlorophenol	7.777	128	80216	42.607	ng		98
9) Benzaldehyde	7.336	77	17308	12.845	ng		88
10) Phenol	7.383	94	110417	40.452	ng		94
11) bis(2-Chloroethyl)ether	7.618	93	73580	36.664	ng		98
12) 1,3-Dichlorobenzene	8.106	146	83501	38.535	ng		100
13) 1,4-Dichlorobenzene	8.253	146	86307	39.171	ng		98
14) 1,2-Dichlorobenzene	8.576	146	87499	41.173	ng		98
15) Benzyl Alcohol	8.447	79	85819	42.165	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.740	45	194319	36.913	ng		98
17) 2-Methylphenol	8.652	107	75318	41.559	ng		95
18) Hexachloroethane	9.316	117	32767	42.378	ng		94
19) n-Nitroso-di-n-propyla...	9.022	70	68699	40.275	ng		99
20) 3+4-Methylphenols	8.975	107	103473	40.599	ng		96
22) Acetophenone	9.040	105	141859	42.659	ng		94
24) Nitrobenzene	9.422	77	99379	45.113	ng		93
25) Isophorone	9.951	82	189958	39.086	ng		98
26) 2-Nitrophenol	10.145	139	43956	51.473	ng	#	94
27) 2,4-Dimethylphenol	10.192	122	86068	47.610	ng		98
28) bis(2-Chloroethoxy)met...	10.433	93	109620	39.615	ng		99
29) 2,4-Dichlorophenol	10.679	162	89329	45.915	ng		98
30) 1,2,4-Trichlorobenzene	10.903	180	98058	44.360	ng		99
31) Naphthalene	11.097	128	265592	39.378	ng		98
32) Benzoic acid	10.309	122	68669	49.533	ng		94
33) 4-Chloroaniline	11.191	127	55506	21.173	ng		97
34) Hexachlorobutadiene	11.384	225	74494	43.092	ng		96
35) Caprolactam	11.943	113	30084m	40.317	ng		
36) 4-Chloro-3-methylphenol	12.289	107	102058	43.642	ng		99
37) 2-Methylnaphthalene	12.689	142	180411	36.433	ng		98
38) 1-Methylnaphthalene	12.900	142	192239	39.215	ng		92
40) 1,2,4,5-Tetrachloroben...	13.047	216	142310	42.179	ng		99
41) Hexachlorocyclopentadiene	13.035	237	175492	115.560	ng		98
43) 2,4,6-Trichlorophenol	13.276	196	89394	46.320	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064639.D
Acq On : 30 Oct 2025 13:39
Operator : RC/JU
Sample : PB170286BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Nov 04 04:25:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.347	196	100082	44.938	ng		99
46) 1,1'-Biphenyl	13.682	154	293305	41.038	ng		98
47) 2-Chloronaphthalene	13.723	162	209827	38.894	ng		97
48) 2-Nitroaniline	13.911	65	82776	45.859	ng		91
49) Acenaphthylene	14.563	152	378803	44.515	ng		98
50) Dimethylphthalate	14.287	163	308978	40.165	ng		98
51) 2,6-Dinitrotoluene	14.399	165	62233	48.961	ng		98
52) Acenaphthene	14.904	154	263186	45.269	ng		98
53) 3-Nitroaniline	14.722	138	43752	33.095	ng		99
54) 2,4-Dinitrophenol	14.927	184	87034	109.532	ng	#	48
55) Dibenzofuran	15.239	168	373821	39.772	ng		99
56) 4-Nitrophenol	15.021	139	117470	95.838	ng		95
57) 2,4-Dinitrotoluene	15.180	165	93928	47.869	ng	#	100
58) Fluorene	15.885	166	293455	39.978	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.456	232	106145	49.841	ng		99
60) Diethylphthalate	15.644	149	304241	38.734	ng		98
61) 4-Chlorophenyl-phenyle...	15.873	204	165276	40.121	ng		98
62) 4-Nitroaniline	15.885	138	69573	47.304	ng		94
63) Azobenzene	16.161	77	277544	38.511	ng		97
65) 4,6-Dinitro-2-methylph...	15.944	198	60745	55.123	ng		95
66) n-Nitrosodiphenylamine	16.079	169	278158	41.795	ng		98
67) 4-Bromophenyl-phenylether	16.766	248	120639	41.320	ng		98
68) Hexachlorobenzene	16.890	284	137499	41.281	ng		96
69) Atrazine	17.019	200	120979	42.241	ng		94
70) Pentachlorophenol	17.225	266	193782	93.297	ng		89
71) Phenanthrene	17.618	178	542808	40.288	ng		99
72) Anthracene	17.712	178	549560	40.098	ng		99
73) Carbazole	17.971	167	499007	41.289	ng		100
74) Di-n-butylphthalate	18.529	149	547593	39.182	ng		99
75) Fluoranthene	19.616	202	696333	41.256	ng		98
77) Benzidine	19.780	184	166962	25.808	ng		99
78) Pyrene	19.974	202	715626	37.932	ng		99
80) Butylbenzylphthalate	20.850	149	246906	43.302	ng		100
81) Benzo(a)anthracene	21.831	228	779257	40.700	ng		100
82) 3,3'-Dichlorobenzidine	21.737	252	198719	31.081	ng		97
83) Chrysene	21.901	228	737890	40.503	ng		99
84) Bis(2-ethylhexyl)phtha...	21.737	149	360268	42.284	ng		100
85) Di-n-octyl phthalate	23.000	149	636901	44.338	ng		98
87) Indeno(1,2,3-cd)pyrene	29.028	276	964238	43.344	ng	#	97
88) Benzo(b)fluoranthene	24.134	252	814847	43.942	ng		100
89) Benzo(k)fluoranthene	24.205	252	798463	43.029	ng		99
90) Benzo(a)pyrene	25.039	252	763198	44.774	ng		99
91) Dibenzo(a,h)anthracene	29.099	278	802843	44.421	ng		99
92) Benzo(g,h,i)perylene	30.221	276	770738	44.641	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG103025\
Data File : BG064639.D
Acq On : 30 Oct 2025 13:39
Operator : RC/JU
Sample : PB170286BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Nov 04 04:25:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.218 min Scan# 8
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:152 Resp: 3019


Ion	Ratio	Lower	Upper
152	100		
150	159.2	122.2	183.4
115	60.7	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 31.579 ng
RT: 3.582 min Scan# 84
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 88 Resp: 25147

Ion	Ratio	Lower	Upper
88	100		
58	122.9	92.4	138.6
43	47.6	33.0	49.6






#3
Pyridine
Concen: 30.398 ng
RT: 3.987 min Scan# 1
Delta R.T. -0.011 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 79 Resp: 7269


Ion	Ratio	Lower	Upper
79	100		
52	119.5	94.2	141.2
51	55.8	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 40.620 ng
RT: 3.899 min Scan# 138
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 42 Resp: 52092

Ion	Ratio	Lower	Upper
42	100		
74	92.6	76.1	114.1
44	7.3	6.3	9.5





#5
2-Fluorophenol
Concen: 122.384 ng
RT: 5.750 min Scan# 4
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39


Tgt Ion:112 Resp: 22013


Ion	Ratio	Lower	Upper
112	100		
64	80.2	60.7	91.1
63	41.7	33.4	50.0

Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#6
Aniline
Concen: 26.094 ng
RT: 7.530 min Scan# 756
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 93 Resp: 87639

Ion	Ratio	Lower	Upper
93	100		
66	47.6	37.0	55.6
65	24.7	18.3	27.5






#7
Phenol-d6
Concen: 121.994 ng
RT: 7.354 min Scan# 7
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 99 Resp: 30110


Ion	Ratio	Lower	Upper
99	100		
42	24.7	18.3	27.5
71	38.8	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 42.607 ng
RT: 7.777 min Scan# 798
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:128 Resp: 80216

Ion	Ratio	Lower	Upper
128	100		
130	33.8	13.5	53.5
64	63.1	45.2	85.2






#9
Benzaldehyde
Concen: 12.845 ng
RT: 7.336 min Scan# 7
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 77 Resp: 1730


Ion	Ratio	Lower	Upper
77	100		
105	86.6	72.9	112.9
106	74.8	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 40.452 ng
RT: 7.383 min Scan# 731
Delta R.T. 0.006 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 94 Resp: 110417

Ion	Ratio	Lower	Upper
94	100		
65	35.4	15.8	55.8
66	50.9	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 36.664 ng
RT: 7.618 min Scan# 7
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 93 Resp: 7358


Ion	Ratio	Lower	Upper
93	100		
63	108.3	91.3	131.3
95	33.3	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 38.535 ng
RT: 8.106 min Scan# 854
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:146 Resp: 83501

Ion	Ratio	Lower	Upper
146	100		
148	68.9	55.3	82.9
75	35.3	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 39.171 ng
RT: 8.253 min Scan# 8
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:146 Resp: 8630


Ion	Ratio	Lower	Upper
146	100		
148	67.7	56.2	84.4
111	44.0	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 41.173 ng
RT: 8.576 min Scan# 934
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:146 Resp: 87499

Ion	Ratio	Lower	Upper
146	100		
148	65.8	51.5	77.3
111	42.1	35.0	52.6






#15
Benzyl Alcohol
Concen: 42.165 ng
RT: 8.447 min Scan# 9
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 79 Resp: 8581


Ion	Ratio	Lower	Upper
79	100		
108	71.7	58.1	87.1
77	67.0	52.8	79.2



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 36.913 ng
RT: 8.740 min Scan# 962
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 45 Resp: 194319

Ion	Ratio	Lower	Upper
45	100		
77	7.6	0.0	28.5
79	6.2	0.0	25.6






#17
2-Methylphenol
Concen: 41.559 ng
RT: 8.652 min Scan# 94
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:107 Resp: 7531


Ion	Ratio	Lower	Upper
107	100		
108	105.8	90.0	135.0
77	47.5	37.8	56.8
79	46.0	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 42.378 ng
RT: 9.316 min Scan# 1060
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:117 Resp: 32767

Ion	Ratio	Lower	Upper
117	100		
119	99.2	75.4	113.2
201	132.3	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 40.275 ng
RT: 9.022 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 70 Resp: 6869


Ion	Ratio	Lower	Upper
70	100		
42	80.6	65.2	97.8
101	8.5	7.9	11.9
130	17.4	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 40.599 ng
RT: 8.975 min Scan# 1002
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:107 Resp: 103473

Ion	Ratio	Lower	Upper
107	100		
108	93.5	70.3	110.3
77	37.9	14.6	54.6
79	33.9	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.044 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:136 Resp: 12164


Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.2	13.8
54	16.2	12.6	19.0
68	6.6	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 42.659 ng
RT: 9.040 min Scan# 1013
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:105 Resp: 141859

Ion	Ratio	Lower	Upper
105	100		
71	1.5	1.5	2.3
51	37.3	33.7	50.5
120	19.0	16.1	24.1






#23
Nitrobenzene-d5
Concen: 97.020 ng
RT: 9.381 min Scan# 1(
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 82 Resp: 19650


Ion	Ratio	Lower	Upper
82	100		
128	41.4	31.3	46.9
54	72.1	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 45.113 ng
RT: 9.422 min Scan# 1078
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 77 Resp: 99379

Ion	Ratio	Lower	Upper
77	100		
123	43.2	31.0	46.6
65	18.0	12.7	19.1






#25
Isophorone
Concen: 39.086 ng
RT: 9.951 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 82 Resp: 18995


Ion	Ratio	Lower	Upper
82	100		
95	7.6	6.1	9.1
138	14.9	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 51.473 ng
RT: 10.145 min Scan# 1201
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:139 Resp: 43956

Ion	Ratio	Lower	Upper
139	100		
109	38.1	24.6	37.0#
65	60.6	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 47.610 ng
RT: 10.192 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:122 Resp: 8606


Ion	Ratio	Lower	Upper
122	100		
107	125.8	103.0	154.4
121	59.7	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 39.615 ng
RT: 10.433 min Scan# 1250
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 93 Resp: 109620

Ion	Ratio	Lower	Upper
93	100		
95	31.7	25.4	38.2
123	9.6	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 45.915 ng
RT: 10.679 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:162 Resp: 8932


Ion	Ratio	Lower	Upper
162	100		
164	68.2	48.7	88.7
98	33.7	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 44.360 ng
RT: 10.903 min Scan# 1330
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:180 Resp: 98058

Ion	Ratio	Lower	Upper
180	100		
182	94.5	76.9	115.3
145	27.6	22.1	33.1






#31
Naphthalene
Concen: 39.378 ng
RT: 11.097 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:128 Resp: 26559


Ion	Ratio	Lower	Upper
128	100		
129	11.5	8.8	13.2
127	14.2	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 49.533 ng
RT: 10.309 min Scan# 1229
Delta R.T. 0.000 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:122 Resp: 68669

Ion	Ratio	Lower	Upper
122	100		
105	124.1	104.2	144.2
77	103.4	70.2	110.2






#33
4-Chloroaniline
Concen: 21.173 ng
RT: 11.191 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:127 Resp: 5550


Ion	Ratio	Lower	Upper
127	100		
129	29.7	25.5	38.3
65	38.8	32.2	48.4
92	20.9	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 43.092 ng
RT: 11.384 min Scan# 1412
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:225 Resp: 74494

Ion	Ratio	Lower	Upper
225	100		
223	62.0	48.8	73.2
227	67.3	49.5	74.3






#35
Caprolactam
Concen: 40.317 ng m
RT: 11.943 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:113 Resp: 3008


Ion	Ratio	Lower	Upper
113	100		
55	294.5	290.1	330.1
56	211.4	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 43.642 ng
RT: 12.289 min Scan# 1566
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:107 Resp: 102058

Ion	Ratio	Lower	Upper
107	100		
144	24.8	19.1	28.7
142	74.1	58.6	87.8






#37
2-Methylnaphthalene
Concen: 36.433 ng
RT: 12.689 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:142 Resp: 18041


Ion	Ratio	Lower	Upper
142	100		
141	88.3	68.9	103.3
115	39.5	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 39.215 ng
RT: 12.900 min Scan# 1670
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:142 Resp: 192239

Ion	Ratio	Lower	Upper
142	100		
141	84.0	73.3	109.9
116	4.0	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.839 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:164 Resp: 10053


Ion	Ratio	Lower	Upper
164	100		
162	93.0	77.0	115.6
160	42.0	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.179 ng
RT: 13.047 min Scan# 1695
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:216 Resp: 142310

Ion	Ratio	Lower	Upper
216	100		
214	77.2	61.4	92.2
179	18.9	15.4	23.0
108	15.9	11.8	17.6






#41
Hexachlorocyclopentadiene
Concen: 115.560 ng
RT: 13.035 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:237 Resp: 17549


Ion	Ratio	Lower	Upper
237	100		
235	62.4	44.1	84.1
272	12.7	0.0	32.4



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#42
2,4,6-Tribromophenol
Concen: 143.076 ng
RT: 16.320 min Scan# 2252
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:330 Resp: 207601

Ion	Ratio	Lower	Upper
330	100		
332	97.7	78.0	117.0
141	27.3	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 46.320 ng
RT: 13.276 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:196 Resp: 8939


Ion	Ratio	Lower	Upper
196	100		
198	96.9	76.6	115.0
200	31.9	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 44.938 ng
RT: 13.347 min Scan# 1746
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:196 Resp: 100082

Ion	Ratio	Lower	Upper
196	100		
198	93.1	74.2	111.4
97	45.0	37.0	55.6
132	22.4	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 79.777 ng
RT: 13.470 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:172 Resp: 52912


Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.4	44.0
170	23.5	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 41.038 ng
RT: 13.682 min Scan# 1803
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:154 Resp: 293305

Ion	Ratio	Lower	Upper
154	100		
153	41.4	22.4	62.4
76	12.0	0.0	33.5






#47
2-Chloronaphthalene
Concen: 38.894 ng
RT: 13.723 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:162 Resp: 20982


Ion	Ratio	Lower	Upper
162	100		
127	38.5	32.2	48.2
164	32.0	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 45.859 ng
RT: 13.911 min Scan# 1842
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 65 Resp: 82776

Ion	Ratio	Lower	Upper
65	100		
92	58.4	48.3	72.5
138	81.3	74.9	112.3





#49
Acenaphthylene
Concen: 44.515 ng
RT: 14.563 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39


Tgt Ion:152 Resp: 37880


Ion	Ratio	Lower	Upper
152	100		
151	18.6	15.4	23.2
153	12.6	11.0	16.4

Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#50
Dimethylphthalate
Concen: 40.165 ng
RT: 14.287 min Scan# 1906
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:163 Resp: 308978

Ion	Ratio	Lower	Upper
163	100		
194	3.4	2.8	4.2
164	10.3	7.7	11.5






#51
2,6-Dinitrotoluene
Concen: 48.961 ng
RT: 14.399 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:165 Resp: 6223


Ion	Ratio	Lower	Upper
165	100		
63	69.0	57.4	86.0
89	60.1	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#52
Acenaphthene
Concen: 45.269 ng
RT: 14.904 min Scan# 2011
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:154 Resp: 263186

Ion	Ratio	Lower	Upper
154	100		
153	113.2	92.2	138.4
152	54.7	43.2	64.8






#53
3-Nitroaniline
Concen: 33.095 ng
RT: 14.722 min Scan# 1
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:138 Resp: 4375


Ion	Ratio	Lower	Upper
138	100		
108	12.0	11.4	17.0
92	119.8	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 109.532 ng
RT: 14.927 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:184 Resp: 87034

Ion	Ratio	Lower	Upper
184	100		
63	70.8	56.6	85.0
154	91.4	158.6	238.0#






#55
Dibenzofuran
Concen: 39.772 ng
RT: 15.239 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:168 Resp: 37382


Ion	Ratio	Lower	Upper
168	100		
139	35.7	27.8	41.8
169	12.9	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 95.838 ng
RT: 15.021 min Scan# 2031
Delta R.T. 0.012 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:139 Resp: 117470

Ion	Ratio	Lower	Upper
139	100		
109	97.6	67.8	107.8
65	125.4	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 47.869 ng
RT: 15.180 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:165 Resp: 9392


Ion	Ratio	Lower	Upper
165	100		
63	51.9	41.8	62.8
89	72.4	58.1	87.1
182	2.3	1.4	2.0#



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 39.978 ng
RT: 15.885 min Scan# 2178
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:166 Resp: 293455

Ion	Ratio	Lower	Upper
166	100		
165	101.4	81.5	122.3
167	12.8	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 49.841 ng
RT: 15.456 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:232 Resp: 10614


Ion	Ratio	Lower	Upper
232	100		
131	33.7	26.7	40.1
130	1.6	1.6	2.4
166	25.1	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 38.734 ng
RT: 15.644 min Scan# 2137
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:149 Resp: 304241

Ion	Ratio	Lower	Upper
149	100		
177	24.0	18.0	27.0
150	12.6	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 40.121 ng
RT: 15.873 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:204 Resp: 16527


Ion	Ratio	Lower	Upper
204	100		
206	32.3	25.6	38.4
141	52.7	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 47.304 ng
RT: 15.885 min Scan# 2178
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:138 Resp: 69573

Ion	Ratio	Lower	Upper
138	100		
92	52.5	36.1	76.1
108	102.5	75.6	115.6






#63
Azobenzene
Concen: 38.511 ng
RT: 16.161 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion: 77 Resp: 27754


Ion	Ratio	Lower	Upper
77	100		
182	27.7	4.9	44.9
105	20.4	1.3	41.3
51	40.3	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.577 min Scan# 2466
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:188 Resp: 246375

Ion	Ratio	Lower	Upper
188	100		
94	6.6	5.7	8.5
80	8.1	6.8	10.2





#65
4,6-Dinitro-2-methylphenol
Concen: 55.123 ng
RT: 15.944 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39


Tgt Ion:198 Resp: 6074


Ion	Ratio	Lower	Upper
198	100		
51	47.9	24.7	64.7
105	35.0	18.0	58.0

Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#66
n-Nitrosodiphenylamine
Concen: 41.795 ng
RT: 16.079 min Scan# 2211
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:169 Resp: 278158

Ion	Ratio	Lower	Upper
169	100		
168	69.0	56.3	84.5
167	37.9	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 41.320 ng
RT: 16.766 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:248 Resp: 12063


Ion	Ratio	Lower	Upper
248	100		
250	95.7	77.7	116.5
141	58.3	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 41.281 ng
RT: 16.890 min Scan# 2349
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:284 Resp: 137499

Ion	Ratio	Lower	Upper
284	100		
142	25.2	21.0	31.6
249	33.9	24.9	37.3






#69
Atrazine
Concen: 42.241 ng
RT: 17.019 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:200 Resp: 12097


Ion	Ratio	Lower	Upper
200	100		
173	24.3	0.9	40.9
215	50.5	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 93.297 ng
RT: 17.225 min Scan# 2406
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:266 Resp: 193782

Ion	Ratio	Lower	Upper
266	100		
268	69.3	50.3	75.5
264	69.6	47.3	70.9






#71
Phenanthrene
Concen: 40.288 ng
RT: 17.618 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:178 Resp: 54280


Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.7	22.1
179	15.7	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 40.098 ng
RT: 17.712 min Scan# 2489
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:178 Resp: 549560

Ion	Ratio	Lower	Upper
178	100		
176	18.2	14.4	21.6
179	14.5	12.4	18.6






#73
Carbazole
Concen: 41.289 ng
RT: 17.971 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:167 Resp: 49900


Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.9	26.9
139	12.1	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 39.182 ng
RT: 18.529 min Scan# 2628
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:149 Resp: 547593

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.0	10.6
104	5.5	4.1	6.1






#75
Fluoranthene
Concen: 41.256 ng
RT: 19.616 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:202 Resp: 69633


Ion	Ratio	Lower	Upper
202	100		
101	7.4	0.0	28.7
203	17.2	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.855 min Scan# 3194
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:240 Resp: 288758

Ion	Ratio	Lower	Upper
240	100		
120	6.9	6.5	9.7
236	25.8	20.2	30.2






#77
Benzidine
Concen: 25.808 ng
RT: 19.780 min Scan# 2
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39


Tgt Ion:184 Resp: 16696

Ion	Ratio	Lower	Upper
184	100		
185	13.9	11.4	17.2
183	11.7	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
PB170286BSD


Manual Integrations
APPROVED
Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 37.932 ng
RT: 19.974 min Scan# 2874
Delta R.T. 0.001 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:202 Resp: 715626

Ion	Ratio	Lower	Upper
202	100		
200	21.4	17.0	25.4
203	17.5	13.8	20.6






#79
Terphenyl-d14
Concen: 80.635 ng
RT: 20.168 min Scan# 2
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:244 Resp: 123381


Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.1	9.1
122	6.6	5.6	8.4



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#80
Butylbenzylphthalate
Concen: 43.302 ng
RT: 20.850 min Scan# 3023
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:149 Resp: 246906

Ion	Ratio	Lower	Upper
149	100		
91	74.0	59.2	88.8
206	22.5	18.1	27.1






#81
Benzo(a)anthracene
Concen: 40.700 ng
RT: 21.831 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:228 Resp: 77925


Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.9	32.9
229	19.9	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 31.081 ng
RT: 21.737 min Scan# 3174
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:252 Resp: 198719

Ion	Ratio	Lower	Upper
252	100		
254	65.9	50.8	76.2
126	6.8	5.8	8.6






#83
Chrysene
Concen: 40.503 ng
RT: 21.901 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:228 Resp: 73789


Ion	Ratio	Lower	Upper
228	100		
226	30.0	23.3	34.9
229	19.3	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 42.284 ng
RT: 21.737 min Scan# 3174
Delta R.T. -0.011 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:149 Resp: 360268

Ion	Ratio	Lower	Upper
149	100		
167	25.2	20.0	30.0
279	4.1	3.1	4.7





#85
Di-n-octyl phthalate
Concen: 44.338 ng
RT: 23.000 min Scan# 3
Delta R.T. -0.011 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39


Tgt Ion:149 Resp: 63690


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	12.6	10.8	16.2

Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.192 min Scan# 3762
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:264 Resp: 302411

Ion	Ratio	Lower	Upper
264	100		
260	23.4	19.0	28.4
265	21.7	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 43.344 ng
RT: 29.028 min Scan# 44
Delta R.T. -0.011 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:276 Resp: 96423


Ion	Ratio	Lower	Upper
276	100		
138	10.0	10.8	16.2
277	24.9	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 43.942 ng
RT: 24.134 min Scan# 3582
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:252 Resp: 814847

Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.8	26.8
125	6.8	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 43.029 ng
RT: 24.205 min Scan# 3
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:252 Resp: 79846


Ion	Ratio	Lower	Upper
252	100		
253	21.3	16.8	25.2
125	6.3	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 44.774 ng
RT: 25.039 min Scan# 3736
Delta R.T. -0.005 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:252 Resp: 763198

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.6	26.4
125	6.8	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 44.421 ng
RT: 29.099 min Scan# 44
Delta R.T. -0.011 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:278 Resp: 80284


Ion	Ratio	Lower	Upper
278	100		
139	9.7	8.0	12.0
279	24.3	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
PB170286BSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 44.641 ng
RT: 30.221 min Scan# 4618
Delta R.T. -0.011 min
Lab File: BG064639.D
Acq: 30 Oct 2025 13:39

Tgt Ion:276 Resp: 770738

Ion	Ratio	Lower	Upper
276	100		
277	23.4	19.1	28.7
138	13.0	11.3	16.9



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/21/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/21/25
Client Sample ID:	PR-132-S16-015094-20251021MS	SDG No.:	Q3434
Lab Sample ID:	Q3419-02MS	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	54.4
Sample Wt/Vol:	30.03 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	2100		1	290	610	ug/Kg	10/29/25 17:34	PB170215
108-95-2	Phenol	3000		1	40.6	310	ug/Kg	10/29/25 17:34	PB170215
111-44-4	bis(2-Chloroethyl)ether	2800		1	44.6	310	ug/Kg	10/29/25 17:34	PB170215
95-57-8	2-Chlorophenol	3200		1	44.8	310	ug/Kg	10/29/25 17:34	PB170215
95-48-7	2-Methylphenol	3100		1	54.9	310	ug/Kg	10/29/25 17:34	PB170215
108-60-1	2,2-oxybis(1-Chloropropane)	2800		1	68.9	310	ug/Kg	10/29/25 17:34	PB170215
98-86-2	Acetophenone	3300		1	54.2	310	ug/Kg	10/29/25 17:34	PB170215
65794-96-9	3+4-Methylphenols	3000		1	75.5	610	ug/Kg	10/29/25 17:34	PB170215
621-64-7	n-Nitroso-di-n-propylamine	3000		1	87.0	150	ug/Kg	10/29/25 17:34	PB170215
67-72-1	Hexachloroethane	3200		1	32.3	310	ug/Kg	10/29/25 17:34	PB170215
98-95-3	Nitrobenzene	3400		1	33.6	310	ug/Kg	10/29/25 17:34	PB170215
78-59-1	Isophorone	3000		1	60.2	310	ug/Kg	10/29/25 17:34	PB170215
88-75-5	2-Nitrophenol	3800		1	110	310	ug/Kg	10/29/25 17:34	PB170215
105-67-9	2,4-Dimethylphenol	3500		1	120	310	ug/Kg	10/29/25 17:34	PB170215
111-91-1	bis(2-Chloroethoxy)methane	2900		1	56.6	310	ug/Kg	10/29/25 17:34	PB170215
120-83-2	2,4-Dichlorophenol	3500		1	52.0	310	ug/Kg	10/29/25 17:34	PB170215
91-20-3	Naphthalene	3000		1	41.7	310	ug/Kg	10/29/25 17:34	PB170215
106-47-8	4-Chloroaniline	1300		1	65.0	310	ug/Kg	10/29/25 17:34	PB170215
87-68-3	Hexachlorobutadiene	3200		1	46.5	310	ug/Kg	10/29/25 17:34	PB170215
105-60-2	Caprolactam	3100		1	95.7	610	ug/Kg	10/29/25 17:34	PB170215
59-50-7	4-Chloro-3-methylphenol	3200		1	52.7	310	ug/Kg	10/29/25 17:34	PB170215
91-57-6	2-Methylnaphthalene	2800		1	47.0	310	ug/Kg	10/29/25 17:34	PB170215
77-47-4	Hexachlorocyclopentadiene	9500	E	1	210	610	ug/Kg	10/29/25 17:34	PB170215
88-06-2	2,4,6-Trichlorophenol	3600		1	36.4	310	ug/Kg	10/29/25 17:34	PB170215
95-95-4	2,4,5-Trichlorophenol	3400		1	53.4	310	ug/Kg	10/29/25 17:34	PB170215
92-52-4	1,1-Biphenyl	3400		1	40.0	310	ug/Kg	10/29/25 17:34	PB170215
91-58-7	2-Chloronaphthalene	3100		1	41.3	310	ug/Kg	10/29/25 17:34	PB170215
88-74-4	2-Nitroaniline	3400		1	88.3	310	ug/Kg	10/29/25 17:34	PB170215
131-11-3	Dimethylphthalate	3100		1	49.8	310	ug/Kg	10/29/25 17:34	PB170215
208-96-8	Acenaphthylene	3500		1	53.1	310	ug/Kg	10/29/25 17:34	PB170215
606-20-2	2,6-Dinitrotoluene	3700		1	61.7	310	ug/Kg	10/29/25 17:34	PB170215
99-09-2	3-Nitroaniline	2500		1	84.5	310	ug/Kg	10/29/25 17:34	PB170215
83-32-9	Acenaphthene	3500		1	39.1	310	ug/Kg	10/29/25 17:34	PB170215
51-28-5	2,4-Dinitrophenol	8100	E	1	420	610	ug/Kg	10/29/25 17:34	PB170215
100-02-7	4-Nitrophenol	7500	E	1	200	610	ug/Kg	10/29/25 17:34	PB170215
132-64-9	Dibenzofuran	3100		1	41.7	310	ug/Kg	10/29/25 17:34	PB170215
121-14-2	2,4-Dinitrotoluene	3600		1	92.0	310	ug/Kg	10/29/25 17:34	PB170215
84-66-2	Diethylphthalate	3100		1	52.0	310	ug/Kg	10/29/25 17:34	PB170215
7005-72-3	4-Chlorophenyl-phenylether	3100		1	49.0	310	ug/Kg	10/29/25 17:34	PB170215
86-73-7	Fluorene	3100		1	46.5	310	ug/Kg	10/29/25 17:34	PB170215
100-01-6	4-Nitroaniline	3600		1	120	310	ug/Kg	10/29/25 17:34	PB170215
534-52-1	4,6-Dinitro-2-methylphenol	4000		1	190	610	ug/Kg	10/29/25 17:34	PB170215

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/21/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/21/25
Client Sample ID:	PR-132-S16-015094-20251021MS	SDG No.:	Q3434
Lab Sample ID:	Q3419-02MS	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	54.4
Sample Wt/Vol:	30.03 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	3100		1	60.4	310	ug/Kg	10/29/25 17:34	PB170215
101-55-3	4-Bromophenyl-phenylether	3200		1	51.1	310	ug/Kg	10/29/25 17:34	PB170215
118-74-1	Hexachlorobenzene	3200		1	46.5	310	ug/Kg	10/29/25 17:34	PB170215
1912-24-9	Atrazine	3400		1	62.4	310	ug/Kg	10/29/25 17:34	PB170215
87-86-5	Pentachlorophenol	7000	E	1	94.2	610	ug/Kg	10/29/25 17:34	PB170215
85-01-8	Phenanthrene	3100		1	38.4	310	ug/Kg	10/29/25 17:34	PB170215
120-12-7	Anthracene	3100		1	61.2	310	ug/Kg	10/29/25 17:34	PB170215
86-74-8	Carbazole	3100		1	57.3	310	ug/Kg	10/29/25 17:34	PB170215
84-74-2	Di-n-butylphthalate	3200		1	88.0	310	ug/Kg	10/29/25 17:34	PB170215
206-44-0	Fluoranthene	3100		1	55.1	310	ug/Kg	10/29/25 17:34	PB170215
129-00-0	Pyrene	2900		1	66.1	310	ug/Kg	10/29/25 17:34	PB170215
85-68-7	Butylbenzylphthalate	3500		1	130	310	ug/Kg	10/29/25 17:34	PB170215
91-94-1	3,3-Dichlorobenzidine	1700		1	67.4	610	ug/Kg	10/29/25 17:34	PB170215
56-55-3	Benzo(a)anthracene	3100		1	42.2	310	ug/Kg	10/29/25 17:34	PB170215
218-01-9	Chrysene	3100		1	36.5	310	ug/Kg	10/29/25 17:34	PB170215
117-81-7	Bis(2-ethylhexyl)phthalate	3400		1	110	310	ug/Kg	10/29/25 17:34	PB170215
117-84-0	Di-n-octyl phthalate	3500		1	160	610	ug/Kg	10/29/25 17:34	PB170215
205-99-2	Benzo(b)fluoranthene	3100		1	34.9	310	ug/Kg	10/29/25 17:34	PB170215
207-08-9	Benzo(k)fluoranthene	3200		1	41.1	310	ug/Kg	10/29/25 17:34	PB170215
50-32-8	Benzo(a)pyrene	3200		1	54.2	310	ug/Kg	10/29/25 17:34	PB170215
193-39-5	Indeno(1,2,3-cd)pyrene	3100		1	53.4	310	ug/Kg	10/29/25 17:34	PB170215
53-70-3	Dibenzo(a,h)anthracene	3200		1	50.3	310	ug/Kg	10/29/25 17:34	PB170215
191-24-2	Benzo(g,h,i)perylene	3200		1	47.2	310	ug/Kg	10/29/25 17:34	PB170215
95-94-3	1,2,4,5-Tetrachlorobenzene	3500		1	47.0	310	ug/Kg	10/29/25 17:34	PB170215
123-91-1	1,4-Dioxane	2500		1	83.0	310	ug/Kg	10/29/25 17:34	PB170215
58-90-2	2,3,4,6-Tetrachlorophenol	3800		1	50.3	310	ug/Kg	10/29/25 17:34	PB170215
SURROGATES									
367-12-4	2-Fluorophenol	95.5			18 - 112	64%	SPK: 150		
13127-88-3	Phenol-d6	99.7			15 - 107	66%	SPK: 150		
4165-60-0	Nitrobenzene-d5	76.9			18 - 107	77%	SPK: 100		
321-60-8	2-Fluorobiphenyl	60.3			20 - 109	60%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	123			10 - 116	82%	SPK: 150		
1718-51-0	Terphenyl-d14	60.4			10 - 105	60%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	30500							
1146-65-2	Naphthalene-d8	121000							
15067-26-2	Acenaphthene-d10	93600							
1517-22-2	Phenanthrene-d10	230000							
1719-03-5	Chrysene-d12	272000							
1520-96-3	Perylene-d12	297000							

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/21/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/21/25
Client Sample ID:	PR-132-S16-015094-20251021MS	SDG No.:	Q3434
Lab Sample ID:	Q3419-02MS	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	54.4
Sample Wt/Vol:	30.03 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064627.D
Acq On : 29 Oct 2025 17:34
Operator : RC/JU
Sample : Q3419-02MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Nov 04 06:46:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.216	152	30488	20.000	ng	0.00
21) Naphthalene-d8	11.042	136	121320	20.000	ng	0.00
39) Acenaphthene-d10	14.838	164	93614	20.000	ng	0.00
64) Phenanthrene-d10	17.576	188	229619	20.000	ng	0.00
76) Chrysene-d12	21.853	240	271892	20.000	ng	# 0.00
86) Perylene-d12	25.191	264	296767	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.749	112	173396	95.462	ng	0.00
7) Phenol-d6	7.353	99	248444	99.678	ng	0.00
23) Nitrobenzene-d5	9.380	82	155421	76.945	ng	0.00
42) 2,4,6-Tribromophenol	16.319	330	166810	123.459	ng	0.00
45) 2-Fluorobiphenyl	13.469	172	372604	60.330	ng	0.00
79) Terphenyl-d14	20.167	244	870859	60.445	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.581	88	32986	41.019	ng	99
3) Pyridine	3.992	79	91699	37.972	ng	98
4) n-Nitrosodimethylamine	3.898	42	64506	49.810	ng	98
6) Aniline	7.529	93	110647	32.624	ng	97
8) 2-Chlorophenol	7.776	128	100885	53.064	ng	94
9) Benzaldehyde	7.335	77	47095	34.610	ng	96
10) Phenol	7.382	94	133081	48.280	ng	96
11) bis(2-Chloroethyl)ether	7.623	93	91218	45.010	ng	94
12) 1,3-Dichlorobenzene	8.111	146	106607	48.719	ng	100
13) 1,4-Dichlorobenzene	8.257	146	108820	48.908	ng	96
14) 1,2-Dichlorobenzene	8.575	146	107105	49.907	ng	97
15) Benzyl Alcohol	8.445	79	100831	49.058	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.745	45	245238	46.132	ng	99
17) 2-Methylphenol	8.645	107	92198	50.378	ng	96
18) Hexachloroethane	9.321	117	40540	51.921	ng	97
19) n-Nitroso-di-n-propyla...	9.021	70	83332	48.377	ng	98
20) 3+4-Methylphenols	8.974	107	125652	48.821	ng	98
22) Acetophenone	9.039	105	180237	54.347	ng	95
24) Nitrobenzene	9.421	77	121561	55.332	ng	95
25) Isophorone	9.955	82	240580	49.637	ng	98
26) 2-Nitrophenol	10.138	139	54118	62.810	ng	93
27) 2,4-Dimethylphenol	10.196	122	103937	57.650	ng	88
28) bis(2-Chloroethoxy)met...	10.431	93	132986	48.190	ng	97
29) 2,4-Dichlorophenol	10.678	162	109647	56.512	ng	99
30) 1,2,4-Trichlorobenzene	10.901	180	120137	54.496	ng	96
31) Naphthalene	11.095	128	328094	48.777	ng	100
32) Benzoic acid	10.320	122	89978	62.736	ng	90
33) 4-Chloroaniline	11.183	127	56519	21.618	ng	98
34) Hexachlorobutadiene	11.383	225	89758	52.062	ng	93
35) Caprolactam	11.947	113	37506m	50.400	ng	
36) 4-Chloro-3-methylphenol	12.288	107	122640	52.585	ng	96
37) 2-Methylnaphthalene	12.688	142	224108	45.380	ng	100
38) 1-Methylnaphthalene	12.905	142	231684	47.390	ng	95
40) 1,2,4,5-Tetrachloroben...	13.046	216	181893	57.895	ng	99
41) Hexachlorocyclopentadiene	13.034	237	218323	154.388	ng	99
43) 2,4,6-Trichlorophenol	13.275	196	105398	58.649	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064627.D
Acq On : 29 Oct 2025 17:34
Operator : RC/JU
Sample : Q3419-02MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Nov 04 06:46:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.346	196	114236	55.084	ng		99
46) 1,1'-Biphenyl	13.680	154	374605	56.287	ng		98
47) 2-Chloronaphthalene	13.722	162	251555	50.074	ng		98
48) 2-Nitroaniline	13.910	65	95513	56.055	ng		95
49) Acenaphthylene	14.568	152	452550	57.112	ng		99
50) Dimethylphthalate	14.286	163	362240	50.569	ng		98
51) 2,6-Dinitrotoluene	14.397	165	72717	60.623	ng		97
52) Acenaphthene	14.903	154	312776	57.775	ng		99
53) 3-Nitroaniline	14.720	138	50027	40.637	ng		98
54) 2,4-Dinitrophenol	14.926	184	99300	132.948	ng	#	46
55) Dibenzofuran	15.237	168	444561	50.793	ng		100
56) 4-Nitrophenol	15.020	139	139148	121.914	ng		90
57) 2,4-Dinitrotoluene	15.179	165	108193	58.491	ng		99
58) Fluorene	15.884	166	345737	50.581	ng		98
59) 2,3,4,6-Tetrachlorophenol	15.455	232	122956	62.001	ng		99
60) Diethylphthalate	15.643	149	365971	50.036	ng		98
61) 4-Chlorophenyl-phenyle...	15.872	204	195464	50.956	ng		99
62) 4-Nitroaniline	15.884	138	81463	59.481	ng		93
63) Azobenzene	16.160	77	358981	53.492	ng		100
65) 4,6-Dinitro-2-methylph...	15.943	198	67553	64.726	ng		94
66) n-Nitrosodiphenylamine	16.078	169	317977	51.265	ng		98
67) 4-Bromophenyl-phenylether	16.765	248	141564	52.025	ng		96
68) Hexachlorobenzene	16.888	284	160884	51.826	ng		96
69) Atrazine	17.018	200	146978	55.064	ng		99
70) Pentachlorophenol	17.223	266	220117	113.709	ng		96
71) Phenanthrene	17.623	178	634368	50.519	ng		99
72) Anthracene	17.711	178	639545	50.069	ng		100
73) Carbazole	17.970	167	577061	51.232	ng		99
74) Di-n-butylphthalate	18.528	149	670737	51.496	ng		98
75) Fluoranthene	19.615	202	802773	51.033	ng		98
77) Benzidine	19.779	184	255069	41.873	ng		98
78) Pyrene	19.973	202	842015	47.400	ng		99
80) Butylbenzylphthalate	20.849	149	309969	57.734	ng		99
81) Benzo(a)anthracene	21.836	228	915343	50.774	ng		99
82) 3,3'-Dichlorobenzidine	21.736	252	166879	27.720	ng		96
83) Chrysene	21.900	228	868130	50.608	ng		98
84) Bis(2-ethylhexyl)phtha...	21.742	149	450725	56.183	ng		98
85) Di-n-octyl phthalate	22.999	149	774721	57.277	ng		99
87) Indeno(1,2,3-cd)pyrene	29.033	276	1117747	51.200	ng	#	97
88) Benzo(b)fluoranthene	24.133	252	914206	50.237	ng		99
89) Benzo(k)fluoranthene	24.209	252	937496	51.482	ng		97
90) Benzo(a)pyrene	25.044	252	883679	52.828	ng		100
91) Dibenzo(a,h)anthracene	29.109	278	916893	51.696	ng		98
92) Benzo(g,h,i)perylene	30.232	276	889128	52.478	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064627.D
Acq On : 29 Oct 2025 17:34
Operator : RC/JU
Sample : Q3419-02MS
Misc :
ALS Vial : 11 Sample Multiplier: 1


Quant Time: Nov 04 06:46:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.216 min Scan# 8
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:152 Resp: 3048


Ion	Ratio	Lower	Upper
152	100		
150	153.9	122.2	183.4
115	61.5	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 41.019 ng
RT: 3.581 min Scan# 84
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 88 Resp: 32986

Ion	Ratio	Lower	Upper
88	100		
58	115.1	92.4	138.6
43	44.0	33.0	49.6






#3
Pyridine
Concen: 37.972 ng
RT: 3.992 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 79 Resp: 9169


Ion	Ratio	Lower	Upper
79	100		
52	115.8	94.2	141.2
51	55.3	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 49.810 ng
RT: 3.898 min Scan# 138
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 42 Resp: 64506

Ion	Ratio	Lower	Upper
42	100		
74	96.6	76.1	114.1
44	7.4	6.3	9.5





#5
2-Fluorophenol
Concen: 95.462 ng
RT: 5.749 min Scan# 4
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:112 Resp: 17339


Ion	Ratio	Lower	Upper
112	100		
64	79.6	60.7	91.1
63	44.4	33.4	50.0

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#6
Aniline
Concen: 32.624 ng
RT: 7.529 min Scan# 756
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 93 Resp: 110647

Ion	Ratio	Lower	Upper
93	100		
66	44.5	37.0	55.6
65	24.7	18.3	27.5






#7
Phenol-d6
Concen: 99.678 ng
RT: 7.353 min Scan# 7
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 99 Resp: 24844


Ion	Ratio	Lower	Upper
99	100		
42	24.4	18.3	27.5
71	39.2	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 53.064 ng
RT: 7.776 min Scan# 798
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:128 Resp: 100885

Ion	Ratio	Lower	Upper
128	100		
130	31.3	13.5	53.5
64	59.5	45.2	85.2






#9
Benzaldehyde
Concen: 34.610 ng
RT: 7.335 min Scan# 7
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 77 Resp: 4709


Ion	Ratio	Lower	Upper
77	100		
105	90.0	72.9	112.9
106	85.1	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 48.280 ng
RT: 7.382 min Scan# 731
Delta R.T. 0.005 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 94 Resp: 133081

Ion	Ratio	Lower	Upper
94	100		
65	36.5	15.8	55.8
66	47.8	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 45.010 ng
RT: 7.623 min Scan# 7
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 93 Resp: 9121


Ion	Ratio	Lower	Upper
93	100		
63	103.4	91.3	131.3
95	33.3	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 48.719 ng
RT: 8.111 min Scan# 855
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:146 Resp: 106607

Ion	Ratio	Lower	Upper
146	100		
148	69.2	55.3	82.9
75	34.8	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 48.908 ng
RT: 8.257 min Scan# 8
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:146 Resp: 10882


Ion	Ratio	Lower	Upper
146	100		
148	66.6	56.2	84.4
111	42.7	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 49.907 ng
RT: 8.575 min Scan# 934
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:146 Resp: 107105

Ion	Ratio	Lower	Upper
146	100		
148	67.1	51.5	77.3
111	45.3	35.0	52.6





#15
Benzyl Alcohol
Concen: 49.058 ng
RT: 8.445 min Scan# 9
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion: 79 Resp: 10083


Ion	Ratio	Lower	Upper
79	100		
108	75.0	58.1	87.1
77	66.5	52.8	79.2

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 46.132 ng
RT: 8.745 min Scan# 963
Delta R.T. 0.005 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 45 Resp: 245238

Ion	Ratio	Lower	Upper
45	100		
77	8.4	0.0	28.5
79	5.9	0.0	25.6






#17
2-Methylphenol
Concen: 50.378 ng
RT: 8.645 min Scan# 94
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:107 Resp: 9219


Ion	Ratio	Lower	Upper
107	100		
108	107.8	90.0	135.0
77	49.0	37.8	56.8
79	47.1	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 51.921 ng
RT: 9.321 min Scan# 1061
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:117 Resp: 40540

Ion	Ratio	Lower	Upper
117	100		
119	92.7	75.4	113.2
201	128.0	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 48.377 ng
RT: 9.021 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 70 Resp: 8333


Ion	Ratio	Lower	Upper
70	100		
42	83.9	65.2	97.8
101	9.9	7.9	11.9
130	18.9	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 48.821 ng
RT: 8.974 min Scan# 1002
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:107 Resp: 125652

Ion	Ratio	Lower	Upper
107	100		
108	91.2	70.3	110.3
77	37.3	14.6	54.6
79	31.8	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.042 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:136 Resp: 12132


Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.2	13.8
54	15.6	12.6	19.0
68	6.1	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 54.347 ng
RT: 9.039 min Scan# 1013
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:105 Resp: 180237

Ion	Ratio	Lower	Upper
105	100		
71	2.3	1.5	2.3
51	38.6	33.7	50.5
120	17.5	16.1	24.1






#23
Nitrobenzene-d5
Concen: 76.945 ng
RT: 9.380 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 82 Resp: 15542


Ion	Ratio	Lower	Upper
82	100		
128	40.2	31.3	46.9
54	73.2	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 55.332 ng
RT: 9.421 min Scan# 1078
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 77 Resp: 121561

Ion	Ratio	Lower	Upper
77	100		
123	43.1	31.0	46.6
65	16.4	12.7	19.1






#25
Isophorone
Concen: 49.637 ng
RT: 9.955 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 82 Resp: 24058


Ion	Ratio	Lower	Upper
82	100		
95	7.2	6.1	9.1
138	15.5	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 62.810 ng
RT: 10.138 min Scan# 1200
Delta R.T. -0.006 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:139 Resp: 54118

Ion	Ratio	Lower	Upper
139	100		
109	35.9	24.6	37.0
65	64.3	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 57.650 ng
RT: 10.196 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:122 Resp: 10393


Ion	Ratio	Lower	Upper
122	100		
107	114.0	103.0	154.4
121	52.2	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 48.190 ng
RT: 10.431 min Scan# 1250
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 93 Resp: 132986

Ion	Ratio	Lower	Upper
93	100		
95	30.0	25.4	38.2
123	10.3	8.4	12.6






#29
2,4-Dichlorophenol
Concen: 56.512 ng
RT: 10.678 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:162 Resp: 10964


Ion	Ratio	Lower	Upper
162	100		
164	68.8	48.7	88.7
98	34.6	16.9	56.9



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#30
1,2,4-Trichlorobenzene
Concen: 54.496 ng
RT: 10.901 min Scan# 1330
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:180 Resp: 120137

Ion	Ratio	Lower	Upper
180	100		
182	92.2	76.9	115.3
145	29.1	22.1	33.1






#31
Naphthalene
Concen: 48.777 ng
RT: 11.095 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:128 Resp: 32809


Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.3	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 62.736 ng
RT: 10.320 min Scan# 1231
Delta R.T. 0.011 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:122 Resp: 89978

Ion	Ratio	Lower	Upper
122	100		
105	129.9	104.2	144.2
77	107.0	70.2	110.2






#33
4-Chloroaniline
Concen: 21.618 ng
RT: 11.183 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:127 Resp: 5651


Ion	Ratio	Lower	Upper
127	100		
129	32.6	25.5	38.3
65	39.3	32.2	48.4
92	20.3	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 52.062 ng
RT: 11.383 min Scan# 1412
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:225 Resp: 89758

Ion	Ratio	Lower	Upper
225	100		
223	64.6	48.8	73.2
227	68.7	49.5	74.3






#35
Caprolactam
Concen: 50.400 ng m
RT: 11.947 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:113 Resp: 3750


Ion	Ratio	Lower	Upper
113	100		
55	284.8	290.1	330.1
56	191.4	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 52.585 ng
RT: 12.288 min Scan# 1566
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:107 Resp: 122640

Ion	Ratio	Lower	Upper
107	100		
144	23.0	19.1	28.7
142	77.0	58.6	87.8






#37
2-Methylnaphthalene
Concen: 45.380 ng
RT: 12.688 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:142 Resp: 22410


Ion	Ratio	Lower	Upper
142	100		
141	85.8	68.9	103.3
115	38.1	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 47.390 ng
RT: 12.905 min Scan# 1671
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:142 Resp: 231684

Ion	Ratio	Lower	Upper
142	100		
141	86.8	73.3	109.9
116	4.0	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.838 min Scan# 2
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:164 Resp: 9361


Ion	Ratio	Lower	Upper
164	100		
162	96.9	77.0	115.6
160	46.8	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 57.895 ng
RT: 13.046 min Scan# 1695
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:216 Resp: 181893

Ion	Ratio	Lower	Upper
216	100		
214	76.0	61.4	92.2
179	19.5	15.4	23.0
108	16.1	11.8	17.6





#41
Hexachlorocyclopentadiene
Concen: 154.388 ng
RT: 13.034 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:237 Resp: 21832


Ion	Ratio	Lower	Upper
237	100		
235	62.8	44.1	84.1
272	12.4	0.0	32.4

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#42
2,4,6-Tribromophenol
Concen: 123.459 ng
RT: 16.319 min Scan# 2252
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:330 Resp: 166810

Ion	Ratio	Lower	Upper
330	100		
332	97.3	78.0	117.0
141	27.3	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 58.649 ng
RT: 13.275 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:196 Resp: 10539


Ion	Ratio	Lower	Upper
196	100		
198	98.0	76.6	115.0
200	35.8	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 55.084 ng
RT: 13.346 min Scan# 1746
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:196 Resp: 114236

Ion	Ratio	Lower	Upper
196	100		
198	92.2	74.2	111.4
97	46.1	37.0	55.6
132	23.0	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 60.330 ng
RT: 13.469 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:172 Resp: 37260


Ion	Ratio	Lower	Upper
172	100		
171	36.6	29.4	44.0
170	24.9	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 56.287 ng
RT: 13.680 min Scan# 1803
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:154 Resp: 374605

Ion	Ratio	Lower	Upper
154	100		
153	41.4	22.4	62.4
76	11.4	0.0	33.5






#47
2-Chloronaphthalene
Concen: 50.074 ng
RT: 13.722 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:162 Resp: 25155


Ion	Ratio	Lower	Upper
162	100		
127	39.6	32.2	48.2
164	32.7	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 56.055 ng
RT: 13.910 min Scan# 1842
Delta R.T. -0.000 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 65 Resp: 95513

Ion	Ratio	Lower	Upper
65	100		
92	58.7	48.3	72.5
138	87.5	74.9	112.3





#49
Acenaphthylene
Concen: 57.112 ng
RT: 14.568 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:152 Resp: 45255


Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.4	23.2
153	13.5	11.0	16.4

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#50
Dimethylphthalate
Concen: 50.569 ng
RT: 14.286 min Scan# 1906
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:163 Resp: 362240

Ion	Ratio	Lower	Upper
163	100		
194	3.7	2.8	4.2
164	10.5	7.7	11.5





#51
2,6-Dinitrotoluene
Concen: 60.623 ng
RT: 14.397 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:165 Resp: 7271


Ion	Ratio	Lower	Upper
165	100		
63	68.6	57.4	86.0
89	58.5	47.8	71.8

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#52
Acenaphthene
Concen: 57.775 ng
RT: 14.903 min Scan# 2011
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:154 Resp: 312776

Ion	Ratio	Lower	Upper
154	100		
153	117.0	92.2	138.4
152	54.2	43.2	64.8






#53
3-Nitroaniline
Concen: 40.637 ng
RT: 14.720 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:138 Resp: 5002


Ion	Ratio	Lower	Upper
138	100		
108	12.7	11.4	17.0
92	120.7	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 132.948 ng
RT: 14.926 min Scan# 2015
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:184 Resp: 99300

Ion	Ratio	Lower	Upper
184	100		
63	69.9	56.6	85.0
154	88.2	158.6	238.0#






#55
Dibenzofuran
Concen: 50.793 ng
RT: 15.237 min Scan# 2
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:168 Resp: 44456


Ion	Ratio	Lower	Upper
168	100		
139	34.8	27.8	41.8
169	13.1	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 121.914 ng
RT: 15.020 min Scan# 2031
Delta R.T. 0.011 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:139 Resp: 139148

Ion	Ratio	Lower	Upper
139	100		
109	100.0	67.8	107.8
65	132.3	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 58.491 ng
RT: 15.179 min Scan# 2
Delta R.T. -0.000 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:165 Resp: 10819


Ion	Ratio	Lower	Upper
165	100		
63	53.1	41.8	62.8
89	71.8	58.1	87.1
182	1.8	1.4	2.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 50.581 ng
RT: 15.884 min Scan# 2178
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:166 Resp: 345737

Ion	Ratio	Lower	Upper
166	100		
165	100.5	81.5	122.3
167	13.0	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 62.001 ng
RT: 15.455 min Scan# 2
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:232 Resp: 12295


Ion	Ratio	Lower	Upper
232	100		
131	34.4	26.7	40.1
130	2.1	1.6	2.4
166	25.2	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 50.036 ng
RT: 15.643 min Scan# 2137
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:149 Resp: 365971

Ion	Ratio	Lower	Upper
149	100		
177	21.9	18.0	27.0
150	11.9	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 50.956 ng
RT: 15.872 min Scan# 21
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:204 Resp: 19546


Ion	Ratio	Lower	Upper
204	100		
206	32.8	25.6	38.4
141	50.2	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 59.481 ng
RT: 15.884 min Scan# 2178
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:138 Resp: 81463

Ion	Ratio	Lower	Upper
138	100		
92	51.6	36.1	76.1
108	102.3	75.6	115.6






#63
Azobenzene
Concen: 53.492 ng
RT: 16.160 min Scan# 2
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion: 77 Resp: 35898


Ion	Ratio	Lower	Upper
77	100		
182	24.7	4.9	44.9
105	20.7	1.3	41.3
51	38.9	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.576 min Scan# 2466
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:188 Resp: 229619

Ion	Ratio	Lower	Upper
188	100		
94	7.4	5.7	8.5
80	8.6	6.8	10.2





#65
4,6-Dinitro-2-methylphenol
Concen: 64.726 ng
RT: 15.943 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:198 Resp: 6755


Ion	Ratio	Lower	Upper
198	100		
51	51.6	24.7	64.7
105	37.2	18.0	58.0

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#66
n-Nitrosodiphenylamine
Concen: 51.265 ng
RT: 16.078 min Scan# 2211
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:169 Resp: 317977

Ion	Ratio	Lower	Upper
169	100		
168	71.6	56.3	84.5
167	38.1	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 52.025 ng
RT: 16.765 min Scan# 2
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:248 Resp: 14156


Ion	Ratio	Lower	Upper
248	100		
250	100.5	77.7	116.5
141	56.9	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 51.826 ng
RT: 16.888 min Scan# 2349
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:284 Resp: 160884

Ion	Ratio	Lower	Upper
284	100		
142	25.1	21.0	31.6
249	34.2	24.9	37.3






#69
Atrazine
Concen: 55.064 ng
RT: 17.018 min Scan# 2
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:200 Resp: 14697


Ion	Ratio	Lower	Upper
200	100		
173	22.8	0.9	40.9
215	46.9	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 113.709 ng
RT: 17.223 min Scan# 2406
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:266 Resp: 220117

Ion	Ratio	Lower	Upper
266	100		
268	65.7	50.3	75.5
264	62.3	47.3	70.9






#71
Phenanthrene
Concen: 50.519 ng
RT: 17.623 min Scan# 24
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:178 Resp: 63436


Ion	Ratio	Lower	Upper
178	100		
176	19.1	14.7	22.1
179	14.8	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 50.069 ng
RT: 17.711 min Scan# 2489
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:178 Resp: 639545

Ion	Ratio	Lower	Upper
178	100		
176	17.6	14.4	21.6
179	15.4	12.4	18.6





#73
Carbazole
Concen: 51.232 ng
RT: 17.970 min Scan# 2
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:167 Resp: 57706


Ion	Ratio	Lower	Upper
167	100		
166	22.9	17.9	26.9
139	12.4	9.7	14.5

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#74
Di-n-butylphthalate
Concen: 51.496 ng
RT: 18.528 min Scan# 2628
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:149 Resp: 670737

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.0	10.6
104	5.9	4.1	6.1





#75
Fluoranthene
Concen: 51.033 ng
RT: 19.615 min Scan# 2
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:202 Resp: 80277


Ion	Ratio	Lower	Upper
202	100		
101	7.7	0.0	28.7
203	17.5	0.0	37.0

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.853 min Scan# 3194
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:240 Resp: 271892

Ion	Ratio	Lower	Upper
240	100		
120	6.3	6.5	9.7#
236	24.8	20.2	30.2






#77
Benzidine
Concen: 41.873 ng
RT: 19.779 min Scan# 2
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:184 Resp: 25506


Ion	Ratio	Lower	Upper
184	100		
185	13.3	11.4	17.2
183	12.1	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 47.400 ng
RT: 19.973 min Scan# 2874
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:202 Resp: 842015

Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.0	25.4
203	17.1	13.8	20.6





#79
Terphenyl-d14
Concen: 60.445 ng
RT: 20.167 min Scan# 2
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:244 Resp: 87085


Ion	Ratio	Lower	Upper
244	100		
212	7.1	6.1	9.1
122	6.3	5.6	8.4

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#80
Butylbenzylphthalate
Concen: 57.734 ng
RT: 20.849 min Scan# 3023
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:149 Resp: 309969

Ion	Ratio	Lower	Upper
149	100		
91	73.3	59.2	88.8
206	23.6	18.1	27.1






#81
Benzo(a)anthracene
Concen: 50.774 ng
RT: 21.836 min Scan# 3
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:228 Resp: 91534


Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.9	32.9
229	20.1	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 27.720 ng
RT: 21.736 min Scan# 3174
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:252 Resp: 166879

Ion	Ratio	Lower	Upper
252	100		
254	66.5	50.8	76.2
126	6.7	5.8	8.6






#83
Chrysene
Concen: 50.608 ng
RT: 21.900 min Scan# 3
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:228 Resp: 86813


Ion	Ratio	Lower	Upper
228	100		
226	30.5	23.3	34.9
229	19.8	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 56.183 ng
RT: 21.742 min Scan# 3175
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:149 Resp: 450725

Ion	Ratio	Lower	Upper
149	100		
167	25.9	20.0	30.0
279	4.4	3.1	4.7





#85
Di-n-octyl phthalate
Concen: 57.277 ng
RT: 22.999 min Scan# 3
Delta R.T. -0.012 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34


Tgt Ion:149 Resp: 77472


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	12.8	10.8	16.2

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.191 min Scan# 3762
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:264 Resp: 296767

Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.0	28.4
265	21.2	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 51.200 ng
RT: 29.033 min Scan# 44
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:276 Resp: 111774


Ion	Ratio	Lower	Upper
276	100		
138	10.1	10.8	16.2
277	25.3	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 50.237 ng
RT: 24.133 min Scan# 3582
Delta R.T. -0.007 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:252 Resp: 914206

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.8	26.8
125	6.6	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 51.482 ng
RT: 24.209 min Scan# 3
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:252 Resp: 93749


Ion	Ratio	Lower	Upper
252	100		
253	22.4	16.8	25.2
125	6.1	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 52.828 ng
RT: 25.044 min Scan# 3737
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:252 Resp: 883679

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	7.7	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 51.696 ng
RT: 29.109 min Scan# 4
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:278 Resp: 91689


Ion	Ratio	Lower	Upper
278	100		
139	9.4	8.0	12.0
279	23.9	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 52.478 ng
RT: 30.232 min Scan# 4620
Delta R.T. -0.001 min
Lab File: BG064627.D
Acq: 29 Oct 2025 17:34

Tgt Ion:276 Resp: 889128

Ion	Ratio	Lower	Upper
276	100		
277	24.1	19.1	28.7
138	13.7	11.3	16.9



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/21/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/21/25
Client Sample ID:	PR-132-S16-015094-20251021MSD	SDG No.:	Q3434
Lab Sample ID:	Q3419-03MSD	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	54.4
Sample Wt/Vol:	30.05 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
100-52-7	Benzaldehyde	2100		1	290	610	ug/Kg	10/29/25 18:15	PB170215
108-95-2	Phenol	2900		1	40.6	310	ug/Kg	10/29/25 18:15	PB170215
111-44-4	bis(2-Chloroethyl)ether	2600		1	44.6	310	ug/Kg	10/29/25 18:15	PB170215
95-57-8	2-Chlorophenol	3000		1	44.8	310	ug/Kg	10/29/25 18:15	PB170215
95-48-7	2-Methylphenol	2900		1	54.9	310	ug/Kg	10/29/25 18:15	PB170215
108-60-1	2,2-oxybis(1-Chloropropane)	2700		1	68.8	310	ug/Kg	10/29/25 18:15	PB170215
98-86-2	Acetophenone	3100		1	54.1	310	ug/Kg	10/29/25 18:15	PB170215
65794-96-9	3+4-Methylphenols	2900		1	75.4	610	ug/Kg	10/29/25 18:15	PB170215
621-64-7	n-Nitroso-di-n-propylamine	2700		1	87.0	150	ug/Kg	10/29/25 18:15	PB170215
67-72-1	Hexachloroethane	3000		1	32.3	310	ug/Kg	10/29/25 18:15	PB170215
98-95-3	Nitrobenzene	3300		1	33.6	310	ug/Kg	10/29/25 18:15	PB170215
78-59-1	Isophorone	2800		1	60.2	310	ug/Kg	10/29/25 18:15	PB170215
88-75-5	2-Nitrophenol	3800		1	110	310	ug/Kg	10/29/25 18:15	PB170215
105-67-9	2,4-Dimethylphenol	3400		1	120	310	ug/Kg	10/29/25 18:15	PB170215
111-91-1	bis(2-Chloroethoxy)methane	2800		1	56.5	310	ug/Kg	10/29/25 18:15	PB170215
120-83-2	2,4-Dichlorophenol	3300		1	51.9	310	ug/Kg	10/29/25 18:15	PB170215
91-20-3	Naphthalene	2900		1	41.7	310	ug/Kg	10/29/25 18:15	PB170215
106-47-8	4-Chloroaniline	1200		1	65.0	310	ug/Kg	10/29/25 18:15	PB170215
87-68-3	Hexachlorobutadiene	3100		1	46.4	310	ug/Kg	10/29/25 18:15	PB170215
105-60-2	Caprolactam	2800		1	95.6	610	ug/Kg	10/29/25 18:15	PB170215
59-50-7	4-Chloro-3-methylphenol	3100		1	52.7	310	ug/Kg	10/29/25 18:15	PB170215
91-57-6	2-Methylnaphthalene	2700		1	47.0	310	ug/Kg	10/29/25 18:15	PB170215
77-47-4	Hexachlorocyclopentadiene	8900	E	1	210	610	ug/Kg	10/29/25 18:15	PB170215
88-06-2	2,4,6-Trichlorophenol	3500		1	36.3	310	ug/Kg	10/29/25 18:15	PB170215
95-95-4	2,4,5-Trichlorophenol	3400		1	53.4	310	ug/Kg	10/29/25 18:15	PB170215
92-52-4	1,1-Biphenyl	3300		1	40.0	310	ug/Kg	10/29/25 18:15	PB170215
91-58-7	2-Chloronaphthalene	3000		1	41.3	310	ug/Kg	10/29/25 18:15	PB170215
88-74-4	2-Nitroaniline	3300		1	88.3	310	ug/Kg	10/29/25 18:15	PB170215
131-11-3	Dimethylphthalate	2900		1	49.7	310	ug/Kg	10/29/25 18:15	PB170215
208-96-8	Acenaphthylene	3300		1	53.0	310	ug/Kg	10/29/25 18:15	PB170215
606-20-2	2,6-Dinitrotoluene	3600		1	61.7	310	ug/Kg	10/29/25 18:15	PB170215
99-09-2	3-Nitroaniline	2400		1	84.4	310	ug/Kg	10/29/25 18:15	PB170215
83-32-9	Acenaphthene	3400		1	39.1	310	ug/Kg	10/29/25 18:15	PB170215
51-28-5	2,4-Dinitrophenol	7800	E	1	420	610	ug/Kg	10/29/25 18:15	PB170215
100-02-7	4-Nitrophenol	6900	E	1	200	610	ug/Kg	10/29/25 18:15	PB170215
132-64-9	Dibenzofuran	2900		1	41.7	310	ug/Kg	10/29/25 18:15	PB170215
121-14-2	2,4-Dinitrotoluene	3300		1	91.9	310	ug/Kg	10/29/25 18:15	PB170215
84-66-2	Diethylphthalate	2900		1	51.9	310	ug/Kg	10/29/25 18:15	PB170215
7005-72-3	4-Chlorophenyl-phenylether	3000		1	49.0	310	ug/Kg	10/29/25 18:15	PB170215
86-73-7	Fluorene	3000		1	46.4	310	ug/Kg	10/29/25 18:15	PB170215
100-01-6	4-Nitroaniline	3500		1	120	310	ug/Kg	10/29/25 18:15	PB170215
534-52-1	4,6-Dinitro-2-methylphenol	3900		1	190	610	ug/Kg	10/29/25 18:15	PB170215

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/21/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/21/25
Client Sample ID:	PR-132-S16-015094-20251021MSD	SDG No.:	Q3434
Lab Sample ID:	Q3419-03MSD	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	54.4
Sample Wt/Vol:	30.05 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
86-30-6	n-Nitrosodiphenylamine	3100		1	60.4	310	ug/Kg	10/29/25 18:15	PB170215
101-55-3	4-Bromophenyl-phenylether	3000		1	51.0	310	ug/Kg	10/29/25 18:15	PB170215
118-74-1	Hexachlorobenzene	3100		1	46.4	310	ug/Kg	10/29/25 18:15	PB170215
1912-24-9	Atrazine	3300		1	62.4	310	ug/Kg	10/29/25 18:15	PB170215
87-86-5	Pentachlorophenol	6800	E	1	94.1	610	ug/Kg	10/29/25 18:15	PB170215
85-01-8	Phenanthrene	3000		1	38.4	310	ug/Kg	10/29/25 18:15	PB170215
120-12-7	Anthracene	2900		1	61.1	310	ug/Kg	10/29/25 18:15	PB170215
86-74-8	Carbazole	3000		1	57.3	310	ug/Kg	10/29/25 18:15	PB170215
84-74-2	Di-n-butylphthalate	3100		1	87.9	310	ug/Kg	10/29/25 18:15	PB170215
206-44-0	Fluoranthene	3100		1	55.1	310	ug/Kg	10/29/25 18:15	PB170215
129-00-0	Pyrene	2700		1	66.1	310	ug/Kg	10/29/25 18:15	PB170215
85-68-7	Butylbenzylphthalate	3400		1	130	310	ug/Kg	10/29/25 18:15	PB170215
91-94-1	3,3-Dichlorobenzidine	1700		1	67.4	610	ug/Kg	10/29/25 18:15	PB170215
56-55-3	Benzo(a)anthracene	3000		1	42.2	310	ug/Kg	10/29/25 18:15	PB170215
218-01-9	Chrysene	2900		1	36.5	310	ug/Kg	10/29/25 18:15	PB170215
117-81-7	Bis(2-ethylhexyl)phthalate	3400		1	110	310	ug/Kg	10/29/25 18:15	PB170215
117-84-0	Di-n-octyl phthalate	3300		1	160	610	ug/Kg	10/29/25 18:15	PB170215
205-99-2	Benzo(b)fluoranthene	3000		1	34.9	310	ug/Kg	10/29/25 18:15	PB170215
207-08-9	Benzo(k)fluoranthene	3000		1	41.1	310	ug/Kg	10/29/25 18:15	PB170215
50-32-8	Benzo(a)pyrene	3100		1	54.1	310	ug/Kg	10/29/25 18:15	PB170215
193-39-5	Indeno(1,2,3-cd)pyrene	3000		1	53.4	310	ug/Kg	10/29/25 18:15	PB170215
53-70-3	Dibenzo(a,h)anthracene	3000		1	50.3	310	ug/Kg	10/29/25 18:15	PB170215
191-24-2	Benzo(g,h,i)perylene	3100		1	47.2	310	ug/Kg	10/29/25 18:15	PB170215
95-94-3	1,2,4,5-Tetrachlorobenzene	3400		1	47.0	310	ug/Kg	10/29/25 18:15	PB170215
123-91-1	1,4-Dioxane	2300		1	83.0	310	ug/Kg	10/29/25 18:15	PB170215
58-90-2	2,3,4,6-Tetrachlorophenol	3600		1	50.3	310	ug/Kg	10/29/25 18:15	PB170215
SURROGATES									
367-12-4	2-Fluorophenol	92.9			18 - 112	62%	SPK: 150		
13127-88-3	Phenol-d6	92.6			15 - 107	62%	SPK: 150		
4165-60-0	Nitrobenzene-d5	74.9			18 - 107	75%	SPK: 100		
321-60-8	2-Fluorobiphenyl	58.4			20 - 109	58%	SPK: 100		
118-79-6	2,4,6-Tribromophenol	118			10 - 116	79%	SPK: 150		
1718-51-0	Terphenyl-d14	57.1			10 - 105	57%	SPK: 100		
INTERNAL STANDARDS		**Area Count**							
3855-82-1	1,4-Dichlorobenzene-d4	31800							
1146-65-2	Naphthalene-d8	125000							
15067-26-2	Acenaphthene-d10	95000							
1517-22-2	Phenanthrene-d10	231000							
1719-03-5	Chrysene-d12	282000							
1520-96-3	Perylene-d12	309000							

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/21/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/21/25
Client Sample ID:	PR-132-S16-015094-20251021MSD	SDG No.:	Q3434
Lab Sample ID:	Q3419-03MSD	Matrix:	SOIL
Analytical Method:	8270E Level: LOW	% Solid:	54.4
Sample Wt/Vol:	30.05 g Final Vol: 1000 uL	Test:	SVOC-TCL BNA -20
Prep Method :	SW3541 Prep Date: 10/23/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064628.D
Acq On : 29 Oct 2025 18:15
Operator : RC/JU
Sample : Q3419-03MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Quant Time: Nov 04 06:47:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.214	152	31803	20.000	ng		0.00
21) Naphthalene-d8	11.040	136	124534	20.000	ng		0.00
39) Acenaphthene-d10	14.842	164	94989	20.000	ng		0.00
64) Phenanthrene-d10	17.580	188	230638	20.000	ng		0.00
76) Chrysene-d12	21.851	240	282250	20.000	ng		0.00
86) Perylene-d12	25.189	264	308600	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.747	112	176100	92.942	ng		0.00
7) Phenol-d6	7.351	99	240854	92.637	ng		0.00
23) Nitrobenzene-d5	9.378	82	155218	74.861	ng		0.00
42) 2,4,6-Tribromophenol	16.317	330	162190	118.302	ng		0.00
45) 2-Fluorobiphenyl	13.467	172	366213	58.437	ng		0.00
79) Terphenyl-d14	20.165	244	853453	57.063	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.579	88	31292	37.303	ng		94
3) Pyridine	3.990	79	91183	36.197	ng		99
4) n-Nitrosodimethylamine	3.896	42	64553	47.785	ng		95
6) Aniline	7.527	93	107880	30.493	ng		98
8) 2-Chlorophenol	7.774	128	98657	49.746	ng		100
9) Benzaldehyde	7.339	77	47639	33.562	ng		94
10) Phenol	7.380	94	134929	46.926	ng		97
11) bis(2-Chloroethyl)ether	7.621	93	90025	42.585	ng		95
12) 1,3-Dichlorobenzene	8.109	146	105133	46.058	ng		95
13) 1,4-Dichlorobenzene	8.250	146	106063	45.698	ng		95
14) 1,2-Dichlorobenzene	8.573	146	107759	48.136	ng		98
15) Benzyl Alcohol	8.444	79	100962	47.091	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.743	45	240700	43.406	ng		99
17) 2-Methylphenol	8.649	107	91314	47.832	ng		99
18) Hexachloroethane	9.319	117	40307	49.488	ng		96
19) n-Nitroso-di-n-propyla...	9.019	70	80697	44.910	ng		98
20) 3+4-Methylphenols	8.972	107	125897	46.893	ng		98
22) Acetophenone	9.037	105	172795	50.758	ng	#	96
24) Nitrobenzene	9.425	77	122938	54.514	ng		99
25) Isophorone	9.954	82	230244	46.278	ng		100
26) 2-Nitrophenol	10.142	139	54798	62.001	ng	#	90
27) 2,4-Dimethylphenol	10.194	122	102164	55.204	ng		95
28) bis(2-Chloroethoxy)met...	10.429	93	131967	46.587	ng		99
29) 2,4-Dichlorophenol	10.676	162	105842	53.143	ng		99
30) 1,2,4-Trichlorobenzene	10.899	180	119635	52.868	ng		97
31) Naphthalene	11.093	128	325077	47.081	ng		99
32) Benzoic acid	10.318	122	89736	61.183	ng		92
33) 4-Chloroaniline	11.187	127	54442	20.286	ng		96
34) Hexachlorobutadiene	11.387	225	89857	50.775	ng		98
35) Caprolactam	11.945	113	35512m	46.489	ng		
36) 4-Chloro-3-methylphenol	12.292	107	122597	51.210	ng		98
37) 2-Methylnaphthalene	12.686	142	219604	43.321	ng		98
38) 1-Methylnaphthalene	12.903	142	228542	45.541	ng		96
40) 1,2,4,5-Tetrachloroben...	13.050	216	178313	55.934	ng		99
41) Hexachlorocyclopentadiene	13.032	237	208614	145.386	ng		98
43) 2,4,6-Trichlorophenol	13.279	196	103636	56.833	ng		97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064628.D
Acq On : 29 Oct 2025 18:15
Operator : RC/JU
Sample : Q3419-03MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Quant Time: Nov 04 06:47:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	13.344	196	115660	54.963	ng		96
46) 1,1'-Biphenyl	13.679	154	365386	54.107	ng		97
47) 2-Chloronaphthalene	13.726	162	249530	48.952	ng		98
48) 2-Nitroaniline	13.908	65	93564	54.228	ng		94
49) Acenaphthylene	14.566	152	439745	54.693	ng		99
50) Dimethylphthalate	14.284	163	347153	47.761	ng		99
51) 2,6-Dinitrotoluene	14.395	165	70596	58.141	ng		99
52) Acenaphthene	14.907	154	301997	54.976	ng		98
53) 3-Nitroaniline	14.724	138	48285	38.655	ng		96
54) 2,4-Dinitrophenol	14.930	184	96369	127.399	ng	#	39
55) Dibenzofuran	15.236	168	423783	47.718	ng		97
56) 4-Nitrophenol	15.018	139	131383	113.444	ng		91
57) 2,4-Dinitrotoluene	15.183	165	102203	54.664	ng	#	97
58) Fluorene	15.882	166	334548	48.236	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.453	232	117236	58.261	ng		99
60) Diethylphthalate	15.641	149	350180	47.184	ng		99
61) 4-Chlorophenyl-phenyle...	15.870	204	189500	48.686	ng		98
62) 4-Nitroaniline	15.888	138	78479	56.473	ng		95
63) Azobenzene	16.164	77	349461	51.319	ng		99
65) 4,6-Dinitro-2-methylph...	15.941	198	67333	64.272	ng		91
66) n-Nitrosodiphenylamine	16.082	169	311250	49.959	ng		98
67) 4-Bromophenyl-phenylether	16.763	248	136215	49.838	ng		98
68) Hexachlorobenzene	16.887	284	156412	50.163	ng		99
69) Atrazine	17.022	200	143329	53.460	ng		97
70) Pentachlorophenol	17.221	266	215900	111.038	ng		96
71) Phenanthrene	17.621	178	614558	48.725	ng		99
72) Anthracene	17.709	178	618291	48.191	ng		100
73) Carbazole	17.968	167	555268	49.079	ng		99
74) Di-n-butylphthalate	18.526	149	660445	50.481	ng		99
75) Fluoranthene	19.613	202	798695	50.549	ng		98
77) Benzidine	19.777	184	266267	42.107	ng		99
78) Pyrene	19.971	202	826087	44.797	ng		99
80) Butylbenzylphthalate	20.852	149	307248	55.127	ng		95
81) Benzo(a)anthracene	21.834	228	915839	48.937	ng		99
82) 3,3'-Dichlorobenzidine	21.734	252	172028	27.527	ng		99
83) Chrysene	21.904	228	856872	48.119	ng		98
84) Bis(2-ethylhexyl)phtha...	21.740	149	456370	54.799	ng		99
85) Di-n-octyl phthalate	22.997	149	762527	54.307	ng		99
87) Indeno(1,2,3-cd)pyrene	29.031	276	1124787	49.546	ng	#	97
88) Benzo(b)fluoranthene	24.131	252	922791	48.765	ng		99
89) Benzo(k)fluoranthene	24.202	252	941575	49.723	ng		98
90) Benzo(a)pyrene	25.036	252	881844	50.697	ng		99
91) Dibenzo(a,h)anthracene	29.102	278	919177	49.838	ng		97
92) Benzo(g,h,i)perylene	30.230	276	894509	50.771	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064628.D
Acq On : 29 Oct 2025 18:15
Operator : RC/JU
Sample : Q3419-03MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1


Quant Time: Nov 04 06:47:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 03 18:16:26 2025
Response via : Initial Calibration

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.214 min Scan# 8
Delta R.T. -0.009 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:152 Resp: 3180


Ion	Ratio	Lower	Upper
152	100		
150	152.5	122.2	183.4
115	61.4	50.8	76.2



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#2
1,4-Dioxane
Concen: 37.303 ng
RT: 3.579 min Scan# 83
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 88 Resp: 31292

Ion	Ratio	Lower	Upper
88	100		
58	123.9	92.4	138.6
43	41.9	33.0	49.6






#3
Pyridine
Concen: 36.197 ng
RT: 3.990 min Scan# 1
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 79 Resp: 9118


Ion	Ratio	Lower	Upper
79	100		
52	117.1	94.2	141.2
51	55.6	46.1	69.1



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#4
n-Nitrosodimethylamine
Concen: 47.785 ng
RT: 3.896 min Scan# 137
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 42 Resp: 64553

Ion	Ratio	Lower	Upper
42	100		
74	99.8	76.1	114.1
44	8.0	6.3	9.5






#5
2-Fluorophenol
Concen: 92.942 ng
RT: 5.747 min Scan# 4
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:112 Resp: 17610


Ion	Ratio	Lower	Upper
112	100		
64	78.8	60.7	91.1
63	42.2	33.4	50.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#6
Aniline
Concen: 30.493 ng
RT: 7.527 min Scan# 755
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 93 Resp: 107880

Ion	Ratio	Lower	Upper
93	100		
66	48.0	37.0	55.6
65	23.4	18.3	27.5






#7
Phenol-d6
Concen: 92.637 ng
RT: 7.351 min Scan# 7
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 99 Resp: 24085


Ion	Ratio	Lower	Upper
99	100		
42	25.2	18.3	27.5
71	39.1	28.6	43.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#8
2-Chlorophenol
Concen: 49.746 ng
RT: 7.774 min Scan# 797
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:128 Resp: 98657

Ion	Ratio	Lower	Upper
128	100		
130	32.9	13.5	53.5
64	65.2	45.2	85.2






#9
Benzaldehyde
Concen: 33.562 ng
RT: 7.339 min Scan# 7
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 77 Resp: 4763


Ion	Ratio	Lower	Upper
77	100		
105	90.0	72.9	112.9
106	82.1	70.8	110.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#10
Phenol
Concen: 46.926 ng
RT: 7.380 min Scan# 730
Delta R.T. 0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 94 Resp: 134929

Ion	Ratio	Lower	Upper
94	100		
65	33.6	15.8	55.8
66	45.7	24.0	64.0






#11
bis(2-Chloroethyl)ether
Concen: 42.585 ng
RT: 7.621 min Scan# 7
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 93 Resp: 9002


Ion	Ratio	Lower	Upper
93	100		
63	106.7	91.3	131.3
95	30.3	13.3	53.3



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#12
1,3-Dichlorobenzene
Concen: 46.058 ng
RT: 8.109 min Scan# 854
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:146 Resp: 105133

Ion	Ratio	Lower	Upper
146	100		
148	64.8	55.3	82.9
75	33.0	28.2	42.2






#13
1,4-Dichlorobenzene
Concen: 45.698 ng
RT: 8.250 min Scan# 8
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:146 Resp: 10606


Ion	Ratio	Lower	Upper
146	100		
148	66.1	56.2	84.4
111	41.8	35.4	53.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#14
1,2-Dichlorobenzene
Concen: 48.136 ng
RT: 8.573 min Scan# 933
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:146 Resp: 107759

Ion	Ratio	Lower	Upper
146	100		
148	63.6	51.5	77.3
111	41.4	35.0	52.6





#15
Benzyl Alcohol
Concen: 47.091 ng
RT: 8.444 min Scan# 9
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15


Tgt Ion: 79 Resp: 10096


Ion	Ratio	Lower	Upper
79	100		
108	73.5	58.1	87.1
77	65.1	52.8	79.2

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 43.406 ng
RT: 8.743 min Scan# 962
Delta R.T. 0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 45 Resp: 240700

Ion	Ratio	Lower	Upper
45	100		
77	8.1	0.0	28.5
79	6.1	0.0	25.6






#17
2-Methylphenol
Concen: 47.832 ng
RT: 8.649 min Scan# 94
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:107 Resp: 91314


Ion	Ratio	Lower	Upper
107	100		
108	112.2	90.0	135.0
77	47.9	37.8	56.8
79	47.9	40.5	60.7



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#18
Hexachloroethane
Concen: 49.488 ng
RT: 9.319 min Scan# 1060
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:117 Resp: 40307

Ion	Ratio	Lower	Upper
117	100		
119	97.0	75.4	113.2
201	130.5	99.2	148.8






#19
n-Nitroso-di-n-propylamine
Concen: 44.910 ng
RT: 9.019 min Scan# 1
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 70 Resp: 8069


Ion	Ratio	Lower	Upper
70	100		
42	79.2	65.2	97.8
101	8.7	7.9	11.9
130	18.1	14.3	21.5



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#20
3+4-Methylphenols
Concen: 46.893 ng
RT: 8.972 min Scan# 1001
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:107 Resp: 125897

Ion	Ratio	Lower	Upper
107	100		
108	91.4	70.3	110.3
77	37.6	14.6	54.6
79	32.1	11.7	51.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.040 min Scan# 1
Delta R.T. -0.009 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:136 Resp: 12453


Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.2	13.8
54	14.8	12.6	19.0
68	6.5	5.7	8.5



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#22
Acetophenone
Concen: 50.758 ng
RT: 9.037 min Scan# 1012
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:105 Resp: 172795

Ion	Ratio	Lower	Upper
105	100		
71	2.6	1.5	2.3#
51	38.8	33.7	50.5
120	19.6	16.1	24.1






#23
Nitrobenzene-d5
Concen: 74.861 ng
RT: 9.378 min Scan# 1
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 82 Resp: 15521


Ion	Ratio	Lower	Upper
82	100		
128	39.7	31.3	46.9
54	72.8	56.6	84.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#24
Nitrobenzene
Concen: 54.514 ng
RT: 9.425 min Scan# 1078
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 77 Resp: 122938

Ion	Ratio	Lower	Upper
77	100		
123	39.5	31.0	46.6
65	15.0	12.7	19.1






#25
Isophorone
Concen: 46.278 ng
RT: 9.954 min Scan# 11
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 82 Resp: 23024


Ion	Ratio	Lower	Upper
82	100		
95	7.6	6.1	9.1
138	16.2	13.0	19.4



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#26
2-Nitrophenol
Concen: 62.001 ng
RT: 10.142 min Scan# 1200
Delta R.T. -0.002 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:139 Resp: 54798

Ion	Ratio	Lower	Upper
139	100		
109	39.5	24.6	37.0#
65	65.2	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 55.204 ng
RT: 10.194 min Scan# 1
Delta R.T. 0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:122 Resp: 10216


Ion	Ratio	Lower	Upper
122	100		
107	121.8	103.0	154.4
121	61.0	47.8	71.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#28
bis(2-Chloroethoxy)methane
Concen: 46.587 ng
RT: 10.429 min Scan# 1249
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 93 Resp: 131967

Ion	Ratio	Lower	Upper
93	100		
95	31.1	25.4	38.2
123	10.0	8.4	12.6





#29
2,4-Dichlorophenol
Concen: 53.143 ng
RT: 10.676 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15


Tgt Ion:162 Resp: 10584


Ion	Ratio	Lower	Upper
162	100		
164	67.2	48.7	88.7
98	37.2	16.9	56.9

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#30
1,2,4-Trichlorobenzene
Concen: 52.868 ng
RT: 10.899 min Scan# 1329
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:180 Resp: 119635

Ion	Ratio	Lower	Upper
180	100		
182	92.6	76.9	115.3
145	26.6	22.1	33.1






#31
Naphthalene
Concen: 47.081 ng
RT: 11.093 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:128 Resp: 32507


Ion	Ratio	Lower	Upper
128	100		
129	10.6	8.8	13.2
127	12.7	10.6	15.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#32
Benzoic acid
Concen: 61.183 ng
RT: 10.318 min Scan# 1230
Delta R.T. 0.009 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:122 Resp: 89736

Ion	Ratio	Lower	Upper
122	100		
105	116.5	104.2	144.2
77	99.7	70.2	110.2






#33
4-Chloroaniline
Concen: 20.286 ng
RT: 11.187 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:127 Resp: 5444


Ion	Ratio	Lower	Upper
127	100		
129	29.8	25.5	38.3
65	38.0	32.2	48.4
92	19.8	18.2	27.2



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#34
Hexachlorobutadiene
Concen: 50.775 ng
RT: 11.387 min Scan# 1412
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:225 Resp: 89857

Ion	Ratio	Lower	Upper
225	100		
223	60.3	48.8	73.2
227	60.1	49.5	74.3






#35
Caprolactam
Concen: 46.489 ng m
RT: 11.945 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:113 Resp: 3551


Ion	Ratio	Lower	Upper
113	100		
55	303.2	290.1	330.1
56	209.1	183.8	223.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#36
4-Chloro-3-methylphenol
Concen: 51.210 ng
RT: 12.292 min Scan# 1566
Delta R.T. 0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:107 Resp: 122597

Ion	Ratio	Lower	Upper
107	100		
144	22.3	19.1	28.7
142	75.1	58.6	87.8






#37
2-Methylnaphthalene
Concen: 43.321 ng
RT: 12.686 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:142 Resp: 21960


Ion	Ratio	Lower	Upper
142	100		
141	83.2	68.9	103.3
115	38.8	30.6	46.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#38
1-Methylnaphthalene
Concen: 45.541 ng
RT: 12.903 min Scan# 1670
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:142 Resp: 228542

Ion	Ratio	Lower	Upper
142	100		
141	87.2	73.3	109.9
116	4.1	3.1	4.7






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.842 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:164 Resp: 9498


Ion	Ratio	Lower	Upper
164	100		
162	99.9	77.0	115.6
160	44.5	34.4	51.6



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 55.934 ng
RT: 13.050 min Scan# 1695
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:216 Resp: 178313

Ion	Ratio	Lower	Upper
216	100		
214	76.5	61.4	92.2
179	18.7	15.4	23.0
108	15.4	11.8	17.6





#41
Hexachlorocyclopentadiene
Concen: 145.386 ng
RT: 13.032 min Scan# 1
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15


Tgt Ion:237 Resp: 20861


Ion	Ratio	Lower	Upper
237	100		
235	65.8	44.1	84.1
272	12.6	0.0	32.4

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#42
2,4,6-Tribromophenol
Concen: 118.302 ng
RT: 16.317 min Scan# 2251
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:330 Resp: 162190

Ion	Ratio	Lower	Upper
330	100		
332	96.7	78.0	117.0
141	27.0	21.8	32.6






#43
2,4,6-Trichlorophenol
Concen: 56.833 ng
RT: 13.279 min Scan# 1
Delta R.T. 0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:196 Resp: 10363


Ion	Ratio	Lower	Upper
196	100		
198	92.9	76.6	115.0
200	31.2	26.0	39.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#44
2,4,5-Trichlorophenol
Concen: 54.963 ng
RT: 13.344 min Scan# 1745
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:196 Resp: 115660

Ion	Ratio	Lower	Upper
196	100		
198	97.4	74.2	111.4
97	44.5	37.0	55.6
132	22.5	18.5	27.7






#45
2-Fluorobiphenyl
Concen: 58.437 ng
RT: 13.467 min Scan# 1
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:172 Resp: 36621


Ion	Ratio	Lower	Upper
172	100		
171	36.9	29.4	44.0
170	24.0	19.1	28.7



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#46
1,1'-Biphenyl
Concen: 54.107 ng
RT: 13.679 min Scan# 1802
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:154 Resp: 365386

Ion	Ratio	Lower	Upper
154	100		
153	40.6	22.4	62.4
76	11.5	0.0	33.5






#47
2-Chloronaphthalene
Concen: 48.952 ng
RT: 13.726 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:162 Resp: 24953


Ion	Ratio	Lower	Upper
162	100		
127	38.6	32.2	48.2
164	34.2	27.5	41.3



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#48
2-Nitroaniline
Concen: 54.228 ng
RT: 13.908 min Scan# 1841
Delta R.T. -0.002 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 65 Resp: 93564

Ion	Ratio	Lower	Upper
65	100		
92	58.2	48.3	72.5
138	86.5	74.9	112.3






#49
Acenaphthylene
Concen: 54.693 ng
RT: 14.566 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:152 Resp: 43974


Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.4	23.2
153	13.5	11.0	16.4



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#50
Dimethylphthalate
Concen: 47.761 ng
RT: 14.284 min Scan# 1905
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:163 Resp: 347153

Ion	Ratio	Lower	Upper
163	100		
194	3.5	2.8	4.2
164	10.1	7.7	11.5





#51
2,6-Dinitrotoluene
Concen: 58.141 ng
RT: 14.395 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15


Tgt Ion:165 Resp: 7059


Ion	Ratio	Lower	Upper
165	100		
63	69.6	57.4	86.0
89	59.8	47.8	71.8

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#52
Acenaphthene
Concen: 54.976 ng
RT: 14.907 min Scan# 2011
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:154 Resp: 301997

Ion	Ratio	Lower	Upper
154	100		
153	116.8	92.2	138.4
152	52.4	43.2	64.8






#53
3-Nitroaniline
Concen: 38.655 ng
RT: 14.724 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:138 Resp: 4828


Ion	Ratio	Lower	Upper
138	100		
108	13.4	11.4	17.0
92	123.2	95.3	142.9



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#54
2,4-Dinitrophenol
Concen: 127.399 ng
RT: 14.930 min Scan# 2015
Delta R.T. 0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:184 Resp: 96369

Ion	Ratio	Lower	Upper
184	100		
63	69.8	56.6	85.0
154	74.7	158.6	238.0#






#55
Dibenzofuran
Concen: 47.718 ng
RT: 15.236 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:168 Resp: 42378


Ion	Ratio	Lower	Upper
168	100		
139	37.1	27.8	41.8
169	13.5	10.3	15.5



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#56
4-Nitrophenol
Concen: 113.444 ng
RT: 15.018 min Scan# 2030
Delta R.T. 0.009 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:139 Resp: 131383

Ion	Ratio	Lower	Upper
139	100		
109	98.3	67.8	107.8
65	132.7	103.7	143.7






#57
2,4-Dinitrotoluene
Concen: 54.664 ng
RT: 15.183 min Scan# 2
Delta R.T. 0.004 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:165 Resp: 10220


Ion	Ratio	Lower	Upper
165	100		
63	50.7	41.8	62.8
89	69.4	58.1	87.1
182	2.4	1.4	2.0#



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#58
Fluorene
Concen: 48.236 ng
RT: 15.882 min Scan# 2177
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:166 Resp: 334548

Ion	Ratio	Lower	Upper
166	100		
165	102.6	81.5	122.3
167	13.8	11.4	17.0






#59
2,3,4,6-Tetrachlorophenol
Concen: 58.261 ng
RT: 15.453 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:232 Resp: 11723(


Ion	Ratio	Lower	Upper
232	100		
131	34.6	26.7	40.1
130	2.1	1.6	2.4
166	25.6	20.5	30.7



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#60
Diethylphthalate
Concen: 47.184 ng
RT: 15.641 min Scan# 2136
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:149 Resp: 350180

Ion	Ratio	Lower	Upper
149	100		
177	21.7	18.0	27.0
150	13.1	10.2	15.4






#61
4-Chlorophenyl-phenylether
Concen: 48.686 ng
RT: 15.870 min Scan# 2
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:204 Resp: 18950


Ion	Ratio	Lower	Upper
204	100		
206	32.9	25.6	38.4
141	51.7	40.6	60.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#62
4-Nitroaniline
Concen: 56.473 ng
RT: 15.888 min Scan# 2178
Delta R.T. 0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:138 Resp: 78479

Ion	Ratio	Lower	Upper
138	100		
92	50.6	36.1	76.1
108	98.4	75.6	115.6






#63
Azobenzene
Concen: 51.319 ng
RT: 16.164 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion: 77 Resp: 34946


Ion	Ratio	Lower	Upper
77	100		
182	25.9	4.9	44.9
105	21.2	1.3	41.3
51	38.4	19.0	59.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.580 min Scan# 2466
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:188 Resp: 230638

Ion	Ratio	Lower	Upper
188	100		
94	6.2	5.7	8.5
80	7.9	6.8	10.2





#65
4,6-Dinitro-2-methylphenol
Concen: 64.272 ng
RT: 15.941 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15


Tgt Ion:198 Resp: 6733


Ion	Ratio	Lower	Upper
198	100		
51	54.1	24.7	64.7
105	36.8	18.0	58.0

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#66
n-Nitrosodiphenylamine
Concen: 49.959 ng
RT: 16.082 min Scan# 2211
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:169 Resp: 311250

Ion	Ratio	Lower	Upper
169	100		
168	68.4	56.3	84.5
167	37.5	29.0	43.6






#67
4-Bromophenyl-phenylether
Concen: 49.838 ng
RT: 16.763 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:248 Resp: 13621


Ion	Ratio	Lower	Upper
248	100		
250	95.2	77.7	116.5
141	58.8	48.6	72.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#68
Hexachlorobenzene
Concen: 50.163 ng
RT: 16.887 min Scan# 2348
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:284 Resp: 156412

Ion	Ratio	Lower	Upper
284	100		
142	25.5	21.0	31.6
249	30.6	24.9	37.3






#69
Atrazine
Concen: 53.460 ng
RT: 17.022 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:200 Resp: 14332


Ion	Ratio	Lower	Upper
200	100		
173	22.7	0.9	40.9
215	49.0	27.0	67.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#70
Pentachlorophenol
Concen: 111.038 ng
RT: 17.221 min Scan# 2405
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:266 Resp: 215900

Ion	Ratio	Lower	Upper
266	100		
268	67.4	50.3	75.5
264	61.2	47.3	70.9






#71
Phenanthrene
Concen: 48.725 ng
RT: 17.621 min Scan# 24
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:178 Resp: 61455


Ion	Ratio	Lower	Upper
178	100		
176	18.8	14.7	22.1
179	15.8	12.2	18.4



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#72
Anthracene
Concen: 48.191 ng
RT: 17.709 min Scan# 2488
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:178 Resp: 618291

Ion	Ratio	Lower	Upper
178	100		
176	18.1	14.4	21.6
179	15.8	12.4	18.6






#73
Carbazole
Concen: 49.079 ng
RT: 17.968 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:167 Resp: 55526


Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.9	26.9
139	12.8	9.7	14.5



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#74
Di-n-butylphthalate
Concen: 50.481 ng
RT: 18.526 min Scan# 2627
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:149 Resp: 660445

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.0	10.6
104	6.0	4.1	6.1






#75
Fluoranthene
Concen: 50.549 ng
RT: 19.613 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:202 Resp: 79869


Ion	Ratio	Lower	Upper
202	100		
101	7.5	0.0	28.7
203	17.8	0.0	37.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.851 min Scan# 3193
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:240 Resp: 282250

Ion	Ratio	Lower	Upper
240	100		
120	7.6	6.5	9.7
236	26.3	20.2	30.2






#77
Benzidine
Concen: 42.107 ng
RT: 19.777 min Scan# 2
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:184 Resp: 26626


Ion	Ratio	Lower	Upper
184	100		
185	13.8	11.4	17.2
183	12.0	9.4	14.2



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#78
Pyrene
Concen: 44.797 ng
RT: 19.971 min Scan# 2873
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:202 Resp: 826087

Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.0	25.4
203	17.0	13.8	20.6





#79
Terphenyl-d14
Concen: 57.063 ng
RT: 20.165 min Scan# 2
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15


Tgt Ion:244 Resp: 85345


Ion	Ratio	Lower	Upper
244	100		
212	7.2	6.1	9.1
122	6.7	5.6	8.4

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#80
Butylbenzylphthalate
Concen: 55.127 ng
RT: 20.852 min Scan# 3023
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:149 Resp: 307248

Ion	Ratio	Lower	Upper
149	100		
91	68.7	59.2	88.8
206	23.1	18.1	27.1






#81
Benzo(a)anthracene
Concen: 48.937 ng
RT: 21.834 min Scan# 3
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:228 Resp: 91583


Ion	Ratio	Lower	Upper
228	100		
226	27.6	21.9	32.9
229	20.2	15.6	23.4



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 27.527 ng
RT: 21.734 min Scan# 3173
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:252 Resp: 172028

Ion	Ratio	Lower	Upper
252	100		
254	62.5	50.8	76.2
126	7.8	5.8	8.6






#83
Chrysene
Concen: 48.119 ng
RT: 21.904 min Scan# 3
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:228 Resp: 85687


Ion	Ratio	Lower	Upper
228	100		
226	30.0	23.3	34.9
229	19.8	15.2	22.8



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 54.799 ng
RT: 21.740 min Scan# 3174
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:149 Resp: 456370

Ion	Ratio	Lower	Upper
149	100		
167	25.2	20.0	30.0
279	4.7	3.1	4.7





#85
Di-n-octyl phthalate
Concen: 54.307 ng
RT: 22.997 min Scan# 3
Delta R.T. -0.014 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15


Tgt Ion:149 Resp: 76252


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	13.3	10.8	16.2

Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 25.189 min Scan# 3761
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:264 Resp: 308600

Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.0	28.4
265	21.5	17.6	26.4






#87
Indeno(1,2,3-cd)pyrene
Concen: 49.546 ng
RT: 29.031 min Scan# 44
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:276 Resp: 112478


Ion	Ratio	Lower	Upper
276	100		
138	10.3	10.8	16.2#
277	25.1	20.1	30.1



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#88
Benzo(b)fluoranthene
Concen: 48.765 ng
RT: 24.131 min Scan# 3581
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:252 Resp: 922791

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.8	26.8
125	6.8	5.6	8.4






#89
Benzo(k)fluoranthene
Concen: 49.723 ng
RT: 24.202 min Scan# 3
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:252 Resp: 94157


Ion	Ratio	Lower	Upper
252	100		
253	22.4	16.8	25.2
125	6.6	5.4	8.0



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025




#90
Benzo(a)pyrene
Concen: 50.697 ng
RT: 25.036 min Scan# 3735
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:252 Resp: 881844

Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.6	26.4
125	7.4	6.1	9.1






#91
Dibenzo(a,h)anthracene
Concen: 49.838 ng
RT: 29.102 min Scan# 44
Delta R.T. -0.008 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:278 Resp: 91917


Ion	Ratio	Lower	Upper
278	100		
139	9.8	8.0	12.0
279	22.8	19.8	29.6



Instrument :
BNA_G
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/04/2025
Supervised By :mohammad ahmed 11/05/2025



#92
Benzo(g,h,i)perylene
Concen: 50.771 ng
RT: 30.230 min Scan# 4619
Delta R.T. -0.003 min
Lab File: BG064628.D
Acq: 29 Oct 2025 18:15

Tgt Ion:276 Resp: 894509

Ion	Ratio	Lower	Upper
276	100		
277	23.3	19.1	28.7
138	13.7	11.3	16.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF102425	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF102825	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF102925	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BG102425	Instrument	BNA_g

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BG064570.D	1,4-Dichlorobenzene	rahul	10/27/2025 8:32:20 AM	mohammad	11/5/2025 12:11:16 AM	Peak Integrated by Software
SSTDICC005	BG064570.D	1-Methylnaphthalene	rahul	10/27/2025 8:32:20 AM	mohammad	11/5/2025 12:11:16 AM	Peak Integrated by Software
SSTDICC010	BG064571.D	Benzoic acid	rahul	10/27/2025 8:32:22 AM	mohammad	11/5/2025 12:11:16 AM	Peak Integrated by Software
SSTDICC020	BG064572.D	Benzoic acid	rahul	10/27/2025 8:32:28 AM	mohammad	11/5/2025 12:11:16 AM	Peak Integrated by Software
SSTDICCC040	BG064573.D	Benzoic acid	rahul	10/27/2025 8:33:01 AM	mohammad	11/5/2025 12:11:16 AM	Peak Integrated by Software
SSTDICC060	BG064574.D	Benzoic acid	Jagrut	11/4/2025 3:47:57 PM	mohammad	11/5/2025 12:11:16 AM	Peak Integrated by Software
SSTDICC080	BG064575.D	Benzoic acid	Jagrut	11/4/2025 3:48:00 PM	mohammad	11/5/2025 12:11:16 AM	Peak Integrated by Software
SSTDICC050	BG064576.D	Benzoic acid	rahul	10/27/2025 8:33:12 AM	mohammad	11/5/2025 12:11:16 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BG102925	Instrument	BNA_g

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB170215BS	BG064622.D	Caprolactam	Jagrut	11/4/2025 3:51:58 PM	mohammad	11/5/2025 12:11:38 AM	Peak Integrated by Software
Q3419-02MS	BG064627.D	Caprolactam	Jagrut	11/4/2025 3:52:06 PM	mohammad	11/5/2025 12:11:38 AM	Peak Integrated by Software
Q3419-03MSD	BG064628.D	Caprolactam	Jagrut	11/4/2025 3:52:09 PM	mohammad	11/5/2025 12:11:38 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BG103025	Instrument	BNA_g

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB170286BS	BG064638.D	Caprolactam	Jagrut	11/4/2025 3:53:35 PM	mohammad	11/5/2025 12:11:43 AM	Peak Integrated by Software
PB170286BSD	BG064639.D	Caprolactam	Jagrut	11/4/2025 3:53:38 PM	mohammad	11/5/2025 12:11:43 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102425

Review By	rahul	Review On	10/27/2025 10:16:31 AM		
Supervise By	Jagrut	Supervise On	10/27/2025 4:26:19 PM		
SubDirectory	BF102425	HP Acquire Method	BNA_F	HP Processing Method	BF102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13179,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF144055.D	24 Oct 2025 09:45	RC/JU	Ok
2	SSTDICC2.5	BF144056.D	24 Oct 2025 10:13	RC/JU	Ok
3	SSTDICC005	BF144057.D	24 Oct 2025 10:42	RC/JU	Ok
4	SSTDICC010	BF144058.D	24 Oct 2025 11:39	RC/JU	Ok
5	SSTDICC020	BF144059.D	24 Oct 2025 12:36	RC/JU	Ok
6	SSTDICCC040	BF144060.D	24 Oct 2025 13:05	RC/JU	Ok
7	SSTDICC050	BF144061.D	24 Oct 2025 13:34	RC/JU	Ok
8	SSTDICC060	BF144062.D	24 Oct 2025 14:03	RC/JU	Not Ok
9	SSTDICC080	BF144063.D	24 Oct 2025 14:32	RC/JU	Ok
10	SSTDICC060	BF144064.D	24 Oct 2025 16:24	RC/JU	Ok
11	SSTDICV040	BF144065.D	24 Oct 2025 17:22	RC/JU	Ok
12	PB170222BL	BF144066.D	24 Oct 2025 17:51	RC/JU	Ok
13	Q3431-01	BF144067.D	24 Oct 2025 18:20	RC/JU	Ok,M
14	Q3431-01DL	BF144068.D	24 Oct 2025 18:48	RC/JU	Not Ok
15	Q3405-01	BF144069.D	24 Oct 2025 19:17	RC/JU	Ok
16	Q3406-01	BF144070.D	24 Oct 2025 19:46	RC/JU	Ok
17	Q3412-01	BF144071.D	24 Oct 2025 20:15	RC/JU	Dilution
18	Q3445-01	BF144072.D	24 Oct 2025 20:44	RC/JU	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102825

Review By	rahul	Review On	10/28/2025 3:29:21 PM				
Supervise By	Jagrut	Supervise On	10/29/2025 3:24:40 PM				
SubDirectory	BF102825	HP Acquire Method	BNA_F		HP Processing Method	BF102425	

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF144096.D	28 Oct 2025 12:48	RC/JU	Ok
2	SSTDCCC040	BF144097.D	28 Oct 2025 13:17	RC/JU	Ok
3	PB170267BL	BF144098.D	28 Oct 2025 13:47	RC/JU	Ok
4	PB170267BS	BF144099.D	28 Oct 2025 14:16	RC/JU	Ok,M
5	PB170202TB	BF144100.D	28 Oct 2025 14:46	RC/JU	Not Ok
6	Q3424-01	BF144101.D	28 Oct 2025 15:16	RC/JU	Ok
7	Q3424-02	BF144102.D	28 Oct 2025 15:45	RC/JU	Ok
8	Q3424-03	BF144103.D	28 Oct 2025 16:15	RC/JU	Ok
9	Q3395-01	BF144104.D	28 Oct 2025 16:44	RC/JU	Ok
10	Q3419-04	BF144105.D	28 Oct 2025 17:14	RC/JU	ReRun
11	Q3434-01	BF144106.D	28 Oct 2025 17:44	RC/JU	Ok
12	Q3460-09	BF144107.D	28 Oct 2025 18:14	RC/JU	Ok
13	Q3439-08	BF144108.D	28 Oct 2025 18:44	RC/JU	Ok
14	Q3440-04	BF144109.D	28 Oct 2025 19:13	RC/JU	Ok
15	Q3460-04	BF144110.D	28 Oct 2025 19:43	RC/JU	Ok
16	Q3460-08	BF144111.D	28 Oct 2025 20:13	RC/JU	Ok
17	Q3460-12	BF144112.D	28 Oct 2025 20:43	RC/JU	Ok
18	Q3427-04	BF144113.D	28 Oct 2025 21:12	RC/JU	Ok
19	Q3439-04	BF144114.D	28 Oct 2025 21:42	RC/JU	Ok
20	Q3395-06	BF144115.D	28 Oct 2025 22:11	RC/JU	Ok
21	Q3447-01DL	BF144116.D	28 Oct 2025 22:40	RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102825

Review By	rahul	Review On	10/28/2025 3:29:21 PM				
Supervise By	Jagrut	Supervise On	10/29/2025 3:24:40 PM				
SubDirectory	BF102825	HP Acquire Method	BNA_F		HP Processing Method	BF102425	

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q3395-05	BF144117.D	28 Oct 2025 23:10	RC/JU	Ok
23	Q3451-08	BF144118.D	28 Oct 2025 23:39	RC/JU	Ok
24	Q3447-06	BF144119.D	29 Oct 2025 00:09	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102925

Review By	rahul	Review On	10/30/2025 8:49:10 AM				
Supervise By	Jagrut	Supervise On	10/30/2025 4:42:33 PM				
SubDirectory	BF102925	HP Acquire Method	BNA_F		HP Processing Method	BF102425	

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF144120.D	29 Oct 2025 11:00	RC/JU	Ok
2	SSTDCCC040	BF144121.D	29 Oct 2025 11:30	RC/JU	Ok
3	PB170202TB	BF144122.D	29 Oct 2025 12:00	RC/JU	Ok
4	PB170259BL	BF144123.D	29 Oct 2025 12:30	RC/JU	Ok
5	PB170259BS	BF144124.D	29 Oct 2025 13:00	RC/JU	Ok,M
6	Q3460-17	BF144125.D	29 Oct 2025 13:33	RC/JU	Ok,M
7	Q3419-04	BF144126.D	29 Oct 2025 14:03	RC/JU	Ok,M
8	Q3427-04MS	BF144127.D	29 Oct 2025 14:33	RC/JU	Ok,M
9	Q3427-04MSD	BF144128.D	29 Oct 2025 15:02	RC/JU	Ok,M
10	Q3469-01	BF144129.D	29 Oct 2025 15:32	RC/JU	Ok,M
11	Q3469-01MS	BF144130.D	29 Oct 2025 16:02	RC/JU	Ok,M
12	Q3469-01MSD	BF144131.D	29 Oct 2025 16:32	RC/JU	Ok,M
13	Q3434-02	BF144132.D	29 Oct 2025 17:02	RC/JU	Ok
14	Q3480-01	BF144133.D	29 Oct 2025 17:32	RC/JU	Ok
15	Q3476-01	BF144134.D	29 Oct 2025 18:01	RC/JU	Ok
16	Q3466-02	BF144135.D	29 Oct 2025 18:31	RC/JU	Ok,M
17	Q3466-07	BF144136.D	29 Oct 2025 19:01	RC/JU	Not Ok
18	Q3481-01	BF144137.D	29 Oct 2025 19:30	RC/JU	Ok,M
19	Q3472-04	BF144138.D	29 Oct 2025 20:00	RC/JU	Ok
20	Q3472-06	BF144139.D	29 Oct 2025 20:29	RC/JU	Ok
21	Q3472-08	BF144140.D	29 Oct 2025 20:59	RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102925

Review By	rahul	Review On	10/30/2025 8:49:10 AM				
Supervise By	Jagrut	Supervise On	10/30/2025 4:42:33 PM				
SubDirectory	BF102925	HP Acquire Method	BNA_F		HP Processing Method	BF102425	

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q3472-10	BF144141.D	29 Oct 2025 21:28	RC/JU	Ok,M
23	Q3472-12	BF144142.D	29 Oct 2025 21:57	RC/JU	Ok
24	Q3472-14	BF144143.D	29 Oct 2025 22:27	RC/JU	Dilution

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QCBatch ID # BG102425

Review By	rahul	Review On	10/27/2025 10:16:00 AM		
Supervise By	mohammad	Supervise On	11/5/2025 12:11:16 AM		
SubDirectory	BG102425	HP Acquire Method	BNA_G	HP Processing Method	BG102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13179,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BG064568.D	24 Oct 2025 8:20	RC/JU	Ok
2	SSTDICC2.5	BG064569.D	24 Oct 2025 9:00	RC/JU	Ok
3	SSTDICC005	BG064570.D	24 Oct 2025 9:41	RC/JU	Ok,M
4	SSTDICC010	BG064571.D	24 Oct 2025 11:02	RC/JU	Ok,M
5	SSTDICC020	BG064572.D	24 Oct 2025 12:23	RC/JU	Ok,M
6	SSTDICCC040	BG064573.D	24 Oct 2025 13:03	RC/JU	Ok,M
7	SSTDICC060	BG064574.D	24 Oct 2025 13:44	RC/JU	Ok,M
8	SSTDICC080	BG064575.D	24 Oct 2025 14:24	RC/JU	Ok,M
9	SSTDICC050	BG064576.D	24 Oct 2025 15:32	RC/JU	Ok,M
10	SSTDICV040	BG064577.D	24 Oct 2025 16:47	RC/JU	Ok
11	PB170222BL	BG064578.D	24 Oct 2025 17:28	RC/JU	Not Ok
12	Q3405-02	BG064579.D	24 Oct 2025 18:08	RC/JU	Ok
13	Q3405-03	BG064580.D	24 Oct 2025 18:49	RC/JU	Ok
14	Q3405-04	BG064581.D	24 Oct 2025 19:30	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QCBatch ID # BG102925

Review By	rahul	Review On	10/30/2025 8:51:46 AM				
Supervise By	mohammad	Supervise On	11/5/2025 12:11:38 AM				
SubDirectory	BG102925	HP Acquire Method	BNA_G		HP Processing Method	BG102425	

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BG064617.D	29 Oct 2025 10:41	RC/JU	Ok
2	SSTDCCC040	BG064618.D	29 Oct 2025 11:22	RC/JU	Ok
3	PB170285BL	BG064619.D	29 Oct 2025 12:03	RC/JU	Ok
4	PB170285BS	BG064620.D	29 Oct 2025 12:44	RC/JU	Ok,M
5	PB170215BL	BG064621.D	29 Oct 2025 13:25	RC/JU	Ok
6	PB170215BS	BG064622.D	29 Oct 2025 14:06	RC/JU	Ok,M
7	Q3477-01	BG064623.D	29 Oct 2025 14:50	RC/JU	Ok
8	Q3477-01MS	BG064624.D	29 Oct 2025 15:31	RC/JU	Ok,M
9	Q3477-01MSD	BG064625.D	29 Oct 2025 16:12	RC/JU	Ok,M
10	Q3419-01	BG064626.D	29 Oct 2025 16:53	RC/JU	Ok
11	Q3419-02MS	BG064627.D	29 Oct 2025 17:34	RC/JU	Ok,M
12	Q3419-03MSD	BG064628.D	29 Oct 2025 18:15	RC/JU	Ok,M
13	Q3468-01	BG064629.D	29 Oct 2025 18:56	RC/JU	Ok,M
14	Q3468-02	BG064630.D	29 Oct 2025 19:37	RC/JU	Ok,M
15	Q3468-03	BG064631.D	29 Oct 2025 20:18	RC/JU	Ok
16	Q3468-04	BG064632.D	29 Oct 2025 20:59	RC/JU	Ok
17	Q3468-05	BG064633.D	29 Oct 2025 21:40	RC/JU	Ok
18	Q3468-06	BG064634.D	29 Oct 2025 22:21	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QCBatch ID # BG103025

Review By	rahul	Review On	10/30/2025 4:42:36 PM				
Supervise By	mohammad	Supervise On	11/5/2025 12:11:43 AM				
SubDirectory	BG103025	HP Acquire Method	BNA_G		HP Processing Method	BG102425	

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BG064635.D	30 Oct 2025 10:55	RC/JU	Ok
2	SSTDCCC040	BG064636.D	30 Oct 2025 11:36	RC/JU	Ok
3	PB170286BL	BG064637.D	30 Oct 2025 12:17	RC/JU	Ok
4	PB170286BS	BG064638.D	30 Oct 2025 12:58	RC/JU	Ok,M
5	PB170286BSD	BG064639.D	30 Oct 2025 13:39	RC/JU	Ok,M
6	PB170306BL	BG064640.D	30 Oct 2025 14:20	RC/JU	Ok
7	PB170306BS	BG064641.D	30 Oct 2025 15:00	RC/JU	Ok,M
8	PB170323BL	BG064642.D	30 Oct 2025 15:41	RC/JU	Ok
9	PB170323BS	BG064643.D	30 Oct 2025 16:23	RC/JU	Not Ok
10	PB170295TB	BG064644.D	30 Oct 2025 17:04	RC/JU	Ok
11	Q3477-02	BG064645.D	30 Oct 2025 17:48	RC/JU	Ok
12	Q3477-02MS	BG064646.D	30 Oct 2025 18:29	RC/JU	Ok,M
13	Q3477-02MSD	BG064647.D	30 Oct 2025 19:10	RC/JU	Ok,M
14	Q3480-02	BG064648.D	30 Oct 2025 19:51	RC/JU	Ok
15	Q3482-01	BG064649.D	30 Oct 2025 20:32	RC/JU	Ok
16	Q3482-03	BG064650.D	30 Oct 2025 21:13	RC/JU	Ok
17	Q3488-06	BG064651.D	30 Oct 2025 21:54	RC/JU	Ok
18	Q3482-05	BG064652.D	30 Oct 2025 22:35	RC/JU	Ok

M : Manual Integration

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102425

Review By	rahul	Review On	10/27/2025 10:16:31 AM		
Supervise By	Jagrut	Supervise On	10/27/2025 4:26:19 PM		
SubDirectory	BF102425	HP Acquire Method	BNA_F	HP Processing Method	BF102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13179,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144055.D	24 Oct 2025 09:45		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF144056.D	24 Oct 2025 10:13		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF144057.D	24 Oct 2025 10:42		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF144058.D	24 Oct 2025 11:39		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF144059.D	24 Oct 2025 12:36		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF144060.D	24 Oct 2025 13:05		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF144061.D	24 Oct 2025 13:34		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF144062.D	24 Oct 2025 14:03	Not Used	RC/JU	Not Ok
9	SSTDICC080	SSTDICC080	BF144063.D	24 Oct 2025 14:32		RC/JU	Ok
10	SSTDICC060	SSTDICC060	BF144064.D	24 Oct 2025 16:24		RC/JU	Ok
11	SSTDICV040	ICVBF102425	BF144065.D	24 Oct 2025 17:22		RC/JU	Ok
12	PB170222BL	PB170222BL	BF144066.D	24 Oct 2025 17:51		RC/JU	Ok
13	Q3431-01	EO-02-10222025	BF144067.D	24 Oct 2025 18:20		RC/JU	Ok,M
14	Q3431-01DL	EO-02-10222025DL	BF144068.D	24 Oct 2025 18:48	Not Required	RC/JU	Not Ok
15	Q3405-01	PAD-10-20-25-A	BF144069.D	24 Oct 2025 19:17		RC/JU	Ok
16	Q3406-01	ARS20-0037	BF144070.D	24 Oct 2025 19:46		RC/JU	Ok
17	Q3412-01	TP-6	BF144071.D	24 Oct 2025 20:15	Need 5X Dilution	RC/JU	Dilution
18	Q3445-01	EO-01-102325	BF144072.D	24 Oct 2025 20:44		RC/JU	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102425

Review By	rahul	Review On	10/27/2025 10:16:31 AM		
Supervise By	Jagrut	Supervise On	10/27/2025 4:26:19 PM		
SubDirectory	BF102425	HP Acquire Method	BNA_F	HP Processing Method	BF102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13179,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102825

Review By	rahul	Review On	10/28/2025 3:29:21 PM		
Supervise By	Jagrut	Supervise On	10/29/2025 3:24:40 PM		
SubDirectory	BF102825	HP Acquire Method	BNA_F	HP Processing Method	BF102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144096.D	28 Oct 2025 12:48		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144097.D	28 Oct 2025 13:17		RC/JU	Ok
3	PB170267BL	PB170267BL	BF144098.D	28 Oct 2025 13:47		RC/JU	Ok
4	PB170267BS	PB170267BS	BF144099.D	28 Oct 2025 14:16		RC/JU	Ok,M
5	PB170202TB	PB170202TB	BF144100.D	28 Oct 2025 14:46	Internal standard fail	RC/JU	Not Ok
6	Q3424-01	PR132-DP3-20251021	BF144101.D	28 Oct 2025 15:16		RC/JU	Ok
7	Q3424-02	PR132-S11-010080-20	BF144102.D	28 Oct 2025 15:45		RC/JU	Ok
8	Q3424-03	PR132-S14-000094-20	BF144103.D	28 Oct 2025 16:15		RC/JU	Ok
9	Q3395-01	PR132-S09-000051-20	BF144104.D	28 Oct 2025 16:44		RC/JU	Ok
10	Q3419-04	PR132-S16-119133-20	BF144105.D	28 Oct 2025 17:14	Hit for pyrene & it is biased high in the CCC.	RC/JU	ReRun
11	Q3434-01	PR132-WC1-20251022	BF144106.D	28 Oct 2025 17:44	Hit for pyrene below CRQL	RC/JU	Ok
12	Q3460-09	WC-3	BF144107.D	28 Oct 2025 18:14		RC/JU	Ok
13	Q3439-08	TP-10	BF144108.D	28 Oct 2025 18:44		RC/JU	Ok
14	Q3440-04	JB-2	BF144109.D	28 Oct 2025 19:13		RC/JU	Ok
15	Q3460-04	WC-1	BF144110.D	28 Oct 2025 19:43		RC/JU	Ok
16	Q3460-08	WC-2	BF144111.D	28 Oct 2025 20:13		RC/JU	Ok
17	Q3460-12	WC-3	BF144112.D	28 Oct 2025 20:43		RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102825

Review By	rahul	Review On	10/28/2025 3:29:21 PM		
Supervise By	Jagrut	Supervise On	10/29/2025 3:24:40 PM		
SubDirectory	BF102825	HP Acquire Method	BNA_F	HP Processing Method	BF102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q3427-04	TP-8	BF144113.D	28 Oct 2025 21:12		RC/JU	Ok
19	Q3439-04	TP-9	BF144114.D	28 Oct 2025 21:42		RC/JU	Ok
20	Q3395-06	PR132-COMP03-2025-	BF144115.D	28 Oct 2025 22:11		RC/JU	Ok
21	Q3447-01DL	P001-TW1-251023-01D	BF144116.D	28 Oct 2025 22:40		RC/JU	Ok
22	Q3395-05	PR132-S07-102116-20	BF144117.D	28 Oct 2025 23:10		RC/JU	Ok
23	Q3451-08	TP-12	BF144118.D	28 Oct 2025 23:39		RC/JU	Ok
24	Q3447-06	P001-TW3-251023-01	BF144119.D	29 Oct 2025 00:09		RC/JU	Ok

M : Manual Integration

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102925

Review By	rahul	Review On	10/30/2025 8:49:10 AM				
Supervise By	Jagrut	Supervise On	10/30/2025 4:42:33 PM				
SubDirectory	BF102925	HP Acquire Method	BNA_F	HP Processing Method	BF102425		

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144120.D	29 Oct 2025 11:00		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144121.D	29 Oct 2025 11:30		RC/JU	Ok
3	PB170202TB	PB170202TB	BF144122.D	29 Oct 2025 12:00		RC/JU	Ok
4	PB170259BL	PB170259BL	BF144123.D	29 Oct 2025 12:30		RC/JU	Ok
5	PB170259BS	PB170259BS	BF144124.D	29 Oct 2025 13:00		RC/JU	Ok,M
6	Q3460-17	WC-5	BF144125.D	29 Oct 2025 13:33		RC/JU	Ok,M
7	Q3419-04	PR132-S16-119133-201	BF144126.D	29 Oct 2025 14:03		RC/JU	Ok,M
8	Q3427-04MS	TP-8MS	BF144127.D	29 Oct 2025 14:33		RC/JU	Ok,M
9	Q3427-04MSD	TP-8MSD	BF144128.D	29 Oct 2025 15:02		RC/JU	Ok,M
10	Q3469-01	TP-13	BF144129.D	29 Oct 2025 15:32		RC/JU	Ok,M
11	Q3469-01MS	TP-13MS	BF144130.D	29 Oct 2025 16:02		RC/JU	Ok,M
12	Q3469-01MSD	TP-13MSD	BF144131.D	29 Oct 2025 16:32		RC/JU	Ok,M
13	Q3434-02	PR132-WC2-20251022	BF144132.D	29 Oct 2025 17:02		RC/JU	Ok
14	Q3480-01	RT3449	BF144133.D	29 Oct 2025 17:32		RC/JU	Ok
15	Q3476-01	NB-07-102825	BF144134.D	29 Oct 2025 18:01		RC/JU	Ok
16	Q3466-02	B8-0.0-1.0-20251023	BF144135.D	29 Oct 2025 18:31		RC/JU	Ok,M
17	Q3466-07	B7-1.0-2.0-20251023	BF144136.D	29 Oct 2025 19:01	Need straight Run	RC/JU	Not Ok
18	Q3481-01	OR-02-102825	BF144137.D	29 Oct 2025 19:30		RC/JU	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102925

Review By	rahul	Review On	10/30/2025 8:49:10 AM		
Supervise By	Jagrut	Supervise On	10/30/2025 4:42:33 PM		
SubDirectory	BF102925	HP Acquire Method	BNA_F	HP Processing Method	BF102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q3472-04	COMP-1	BF144138.D	29 Oct 2025 20:00		RC/JU	Ok
20	Q3472-06	COMP-2	BF144139.D	29 Oct 2025 20:29	Internal standard fail	RC/JU	Ok
21	Q3472-08	COMP-3	BF144140.D	29 Oct 2025 20:59	Internal standard fail	RC/JU	Ok
22	Q3472-10	COMP-4	BF144141.D	29 Oct 2025 21:28	Surrogate Fail,	RC/JU	Ok,M
23	Q3472-12	COMP-5	BF144142.D	29 Oct 2025 21:57		RC/JU	Ok
24	Q3472-14	155	BF144143.D	29 Oct 2025 22:27	need further 10X	RC/JU	Dilution

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QCBatch ID # BG102425

Review By	rahul	Review On	10/27/2025 10:16:00 AM		
Supervise By	mohammad	Supervise On	11/5/2025 12:11:16 AM		
SubDirectory	BG102425	HP Acquire Method	BNA_G	HP Processing Method	BG102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13179,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064568.D	24 Oct 2025 8:20		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BG064569.D	24 Oct 2025 9:00		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BG064570.D	24 Oct 2025 9:41		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BG064571.D	24 Oct 2025 11:02		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BG064572.D	24 Oct 2025 12:23	Compound #32 kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BG064573.D	24 Oct 2025 13:03	Compound #26,48,51,54,57,65 are kept on LR	RC/JU	Ok,M
7	SSTDICC060	SSTDICC060	BG064574.D	24 Oct 2025 13:44		RC/JU	Ok,M
8	SSTDICC080	SSTDICC080	BG064575.D	24 Oct 2025 14:24		RC/JU	Ok,M
9	SSTDICC050	SSTDICC050	BG064576.D	24 Oct 2025 15:32		RC/JU	Ok,M
10	SSTDICV040	ICVBG102425	BG064577.D	24 Oct 2025 16:47	compound #77 failed in the ICV.	RC/JU	Ok
11	PB170222BL	PB170222BL	BG064578.D	24 Oct 2025 17:28	Analyzed for contamination check	RC/JU	Not Ok
12	Q3405-02	PAD-10-20-25-B	BG064579.D	24 Oct 2025 18:08		RC/JU	Ok
13	Q3405-03	PAD-10-20-25-C	BG064580.D	24 Oct 2025 18:49		RC/JU	Ok
14	Q3405-04	SW-10-20-25	BG064581.D	24 Oct 2025 19:30		RC/JU	Ok

M : Manual Integration

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QCBatch ID # BG102925

Review By	rahul	Review On	10/30/2025 8:51:46 AM		
Supervise By	mohammad	Supervise On	11/5/2025 12:11:38 AM		
SubDirectory	BG102925	HP Acquire Method	BNA_G	HP Processing Method	BG102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064617.D	29 Oct 2025 10:41		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064618.D	29 Oct 2025 11:22		RC/JU	Ok
3	PB170285BL	PB170285BL	BG064619.D	29 Oct 2025 12:03		RC/JU	Ok
4	PB170285BS	PB170285BS	BG064620.D	29 Oct 2025 12:44		RC/JU	Ok,M
5	PB170215BL	PB170215BL	BG064621.D	29 Oct 2025 13:25		RC/JU	Ok
6	PB170215BS	PB170215BS	BG064622.D	29 Oct 2025 14:06		RC/JU	Ok,M
7	Q3477-01	SOUTH-MAHWAH-SOI	BG064623.D	29 Oct 2025 14:50		RC/JU	Ok
8	Q3477-01MS	SOUTH-MAHWAH-SOI	BG064624.D	29 Oct 2025 15:31		RC/JU	Ok,M
9	Q3477-01MSD	SOUTH-MAHWAH-SOI	BG064625.D	29 Oct 2025 16:12		RC/JU	Ok,M
10	Q3419-01	PR-132-S16-015094-2(	BG064626.D	29 Oct 2025 16:53		RC/JU	Ok
11	Q3419-02MS	PR-132-S16-015094-2(	BG064627.D	29 Oct 2025 17:34		RC/JU	Ok,M
12	Q3419-03MSD	PR-132-S16-015094-2(	BG064628.D	29 Oct 2025 18:15		RC/JU	Ok,M
13	Q3468-01	MW-06	BG064629.D	29 Oct 2025 18:56		RC/JU	Ok,M
14	Q3468-02	MW-08	BG064630.D	29 Oct 2025 19:37		RC/JU	Ok,M
15	Q3468-03	MW-10	BG064631.D	29 Oct 2025 20:18		RC/JU	Ok
16	Q3468-04	MW-11	BG064632.D	29 Oct 2025 20:59		RC/JU	Ok
17	Q3468-05	MW-13	BG064633.D	29 Oct 2025 21:40		RC/JU	Ok
18	Q3468-06	MW-12	BG064634.D	29 Oct 2025 22:21		RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QCBatch ID # BG102925

Review By	rahul	Review On	10/30/2025 8:51:46 AM		
Supervise By	mohammad	Supervise On	11/5/2025 12:11:38 AM		
SubDirectory	BG102925	HP Acquire Method	BNA_G	HP Processing Method	BG102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QCBatch ID # BG103025

Review By	rahul	Review On	10/30/2025 4:42:36 PM		
Supervise By	mohammad	Supervise On	11/5/2025 12:11:43 AM		
SubDirectory	BG103025	HP Acquire Method	BNA_G	HP Processing Method	BG102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064635.D	30 Oct 2025 10:55		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064636.D	30 Oct 2025 11:36		RC/JU	Ok
3	PB170286BL	PB170286BL	BG064637.D	30 Oct 2025 12:17		RC/JU	Ok
4	PB170286BS	PB170286BS	BG064638.D	30 Oct 2025 12:58		RC/JU	Ok,M
5	PB170286BSD	PB170286BSD	BG064639.D	30 Oct 2025 13:39		RC/JU	Ok,M
6	PB170306BL	PB170306BL	BG064640.D	30 Oct 2025 14:20		RC/JU	Ok
7	PB170306BS	PB170306BS	BG064641.D	30 Oct 2025 15:00		RC/JU	Ok,M
8	PB170323BL	PB170323BL	BG064642.D	30 Oct 2025 15:41		RC/JU	Ok
9	PB170323BS	PB170323BS	BG064643.D	30 Oct 2025 16:23	I S Low & therefore Recoveries failed	RC/JU	Not Ok
10	PB170295TB	PB170295TB	BG064644.D	30 Oct 2025 17:04		RC/JU	Ok
11	Q3477-02	SOUTH-MAHWAH-SOI	BG064645.D	30 Oct 2025 17:48		RC/JU	Ok
12	Q3477-02MS	SOUTH-MAHWAH-SOI	BG064646.D	30 Oct 2025 18:29		RC/JU	Ok,M
13	Q3477-02MSD	SOUTH-MAHWAH-SOI	BG064647.D	30 Oct 2025 19:10		RC/JU	Ok,M
14	Q3480-02	RT3449	BG064648.D	30 Oct 2025 19:51		RC/JU	Ok
15	Q3482-01	WC-1	BG064649.D	30 Oct 2025 20:32		RC/JU	Ok
16	Q3482-03	WC-3	BG064650.D	30 Oct 2025 21:13		RC/JU	Ok
17	Q3488-06	MOO-25-0311-0313	BG064651.D	30 Oct 2025 21:54		RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QCBatch ID # BG103025

Review By	rahul	Review On	10/30/2025 4:42:36 PM		
Supervise By	mohammad	Supervise On	11/5/2025 12:11:43 AM		
SubDirectory	BG103025	HP Acquire Method	BNA_G	HP Processing Method	BG102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q3482-05	WC-5	BG064652.D	30 Oct 2025 22:35		RC/JU	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/23/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/22/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/23/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC: LB137614

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3419-01	PR-132-S16-015094-20251021	1	1.15	11.01	12.16	7.14	54.4	
Q3419-02	Q3419-01MS	2	1.15	11.01	12.16	7.14	54.4	
Q3419-03	Q3419-01MSD	3	1.15	11.01	12.16	7.14	54.4	
Q3419-04	PR-132-S16-119133-20251021	4	1.16	10.07	11.23	6.37	51.7	
Q3419-05	PR-132-S16-094119-20251021	5	1.19	11.07	12.26	6.38	46.9	
Q3422-01	PR132-S15-000008-20251017	6	1.18	11.02	12.2	7.58	58.1	
Q3424-01	PR132-DP3-20251021	7	1.15	10.36	11.51	5.91	45.9	
Q3424-02	PR132-S11-010080-20251021	8	1.12	10.72	11.84	6.49	50.1	
Q3424-03	PR132-S14-00094-20251021	9	1.14	11.34	12.48	6.82	50.1	
Q3427-01	TP-8	10	1.11	10.77	11.88	10.84	90.3	
Q3427-02	TP-8-EPH	11	1.18	10.97	12.15	10.98	89.3	
Q3427-03	TP-8-VOC	12	1.19	10.26	11.45	10.52	90.9	
Q3429-01	40309	13	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3429-02	40310	14	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3430-01	BUR-25-0095	15	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3431-01	EO-02-10222025	16	1.11	10.55	11.66	10.64	90.3	
Q3431-02	EO-02-10222025-E2	17	1.17	10.59	11.76	10.59	89.0	
Q3434-01	PR132-WC1-20251022	18	1.13	10.43	11.56	7.17	57.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

VB 137614

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-102225 **WorkList ID :** 192597 **Department :** Wet-Chemistry **Date :** 10-22-2025 09:03:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3419-01	PR-132-S16-015094-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3419-02	Q3419-01MS	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3419-03	Q3419-01MSD	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3419-04	PR-132-S16-119133-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3419-05	PR-132-S16-094119-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3422-01	PR132-S15-000008-20251017	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3424-01	PR132-DP3-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/21/2025	Chemtech -SO
Q3424-02	PR132-S11-010080-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/21/2025	Chemtech -SO
Q3424-03	PR132-S14-00094-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/21/2025	Chemtech -SO
Q3427-01	TP-8	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3427-02	TP-8-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3427-03	TP-8-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3429-01	40309	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	10/22/2025	Chemtech -SO
Q3429-02	40310	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	10/22/2025	Chemtech -SO
Q3430-01	BUR-25-0095	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	10/22/2025	Chemtech -SO
Q3431-01	EO-02-10222025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	10/22/2025	Chemtech -SO
Q3431-02	EO-02-10222025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	10/22/2025	Chemtech -SO
Q3434-01	PR132-WC1-20251022	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/22/2025	Chemtech -SO

Date/Time 10/22/25 15:00

Raw Sample Received by: JB WC

Raw Sample Relinquished by:

Date/Time 10/22/25 17:25

Raw Sample Received by:

Raw Sample Relinquished by: JB WC

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB170215

SOP ID:	M3541-ASE Extraction-15				
Clean Up SOP #:	N/A			**Extraction Start Date :**	10/23/2025
Matrix :	Solid			**Extraction Start Time :**	09:15
Weigh By:	EH	**Extraction By:**	RJ	**Extraction End Date :**	10/23/2025
Balance check:	EH	**Filter By:**	RJ	**Extraction End Time :**	12:45
Balance ID:	EX-SC-2	**pH Meter ID:**	N/A	**Concentration By:**	EH
pH Strip Lot#:	N/A	**Hood ID:**	3,7	**Supervisor By :**	RUPESH

Extraction Method: ☐ Seperatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☑ Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6865
Surrogate	1.0ML	100/150 PPM	SP6869
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
MeCl2/Acetone/1:1	N/A	EP2641
Baked Na2SO4	N/A	EP2652
Sand	N/A	E3951
Methylene Chloride	N/A	E3980
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5ML Vial Lot # 2210443.

KD Bath ID:	N/A	**Envap ID:**	N-EVAP-02
KD Bath Temperature:	N/A	**Envap Temperature:**	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/23/25	RS (Ext-lab)	RC/SVOC
12:50	**Preparation Group**	**Analysis Group**

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB170215

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 10/23/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170215BL	SBLK215	SVOC-TCL BNA -20	30.01	N/A	ritesh	Evelyn	1			U1-1
PB170215BS	SLCS215	SVOC-TCL BNA -20	30.03	N/A	ritesh	Evelyn	1			2
Q3395-01	PR132-S09-000051-2025 1017	SVOC-TCL BNA -20	30.04	N/A	ritesh	Evelyn	1	E		3
Q3395-02	PR132-S07-000102-2025 1017	SVOC-TCL BNA -20	30.09	N/A	ritesh	Evelyn	1	E		4
Q3395-03	Q3395-02MS	SVOC-TCL BNA -20	30.08	N/A	ritesh	Evelyn	1	E		5
Q3395-04	Q3395-02MSD	SVOC-TCL BNA -20	30.04	N/A	ritesh	Evelyn	1	E		6
Q3395-05	PR132-S07-102116-2025 1017	SVOC-TCL BNA -20	30.02	N/A	ritesh	Evelyn	1	E		U2-1
Q3395-06	PR132-COMP03-2025101 7	SVOC-TCL BNA -20	30.05	N/A	ritesh	Evelyn	1	E		2
~~Q3395-07~~	~~PR132-S08-038048-2025 1017~~	~~SVOC-TCL BNA -20~~		~~N/A~~	ritesh	Evelyn	~~1~~			
Q3419-01	PR-132-S16-015094-202 51021	SVOC-TCL BNA -20	30.06	N/A	ritesh	Evelyn	1	E		3
Q3419-02	Q3419-01MS	SVOC-TCL BNA -20	30.03	N/A	ritesh	Evelyn	1	E		4
Q3419-03	Q3419-01MSD	SVOC-TCL BNA -20	30.05	N/A	ritesh	Evelyn	1	E		5
Q3419-04	PR-132-S16-119133-202 51021	SVOC-TCL BNA -20	30.01	N/A	ritesh	Evelyn	1	E		6
Q3419-05	PR-132-S16-094119-202 51021	SVOC-TCL BNA -20	30.06	N/A	ritesh	Evelyn	1	E		U3-1
Q3422-01	PR132-S15-000008-2025 1017	SVOC-TCL BNA -20	30.08	N/A	ritesh	Evelyn	1	F		2
Q3424-01	PR132-DP3-20251021	SVOC-TCL BNA -20	30.02	N/A	ritesh	Evelyn	1	E		3
Q3424-02	PR132-S11-010080-2025 1021	SVOC-TCL BNA -20	30.05	N/A	ritesh	Evelyn	1	F		4
Q3424-03	PR132-S14-00094-20251 021	SVOC-TCL BNA -20	30.06	N/A	ritesh	Evelyn	1	E		5
Q3427-01	TP-8	SVOC-TCL BNA -20	50.04	N/A	ritesh	Evelyn	1	E		6
Q3431-01	EO-02-10222025	SVOC-TCL BNA -20	50.07	N/A	ritesh	Evelyn	1	E		U4-1
Q3434-01	PR132-WC1-20251022	SVOC-TCL BNA -20	30.01	N/A	ritesh	Evelyn	1	E		2

* **Extracts relinguished on the same date as received.**

Q3434-SVOC-TCL BNA -20

10/23/25 9:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3422 **WorkList ID :** 192632 **Department :** Extraction **Date :** 10-23-2025 08:59:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3395-01	PR132-S09-000051-20251017	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/17/2025	8270E
Q3395-02	PR132-S07-000102-20251017	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/17/2025	8270E
Q3395-03	Q3395-02MS	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/17/2025	8270E
Q3395-04	Q3395-02MSD	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/17/2025	8270E
Q3395-05	PR132-S07-102116-20251017	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/17/2025	8270E
Q3395-06	PR132-COMP03-20251017	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/17/2025	8270E
Q3395-07	PR132-S08-038048-20251017	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/17/2025	8270E
Q3419-01	PR-132-S16-015094-20251021	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D41	10/21/2025	8270E
Q3419-02	Q3419-01MS	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D41	10/21/2025	8270E
Q3419-03	Q3419-01MSD	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D41	10/21/2025	8270E
Q3419-04	PR-132-S16-119133-20251021	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D41	10/21/2025	8270E
Q3419-05	PR-132-S16-094119-20251021	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D41	10/21/2025	8270E
Q3422-01	PR132-S15-000008-20251017	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D41	10/21/2025	8270E
Q3424-01	PR132-DP3-20251021	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/21/2025	8270E
Q3424-02	PR132-S11-010080-20251021	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/21/2025	8270E
Q3424-03	PR132-S14-00094-20251021	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D31	10/21/2025	8270E
Q3427-01	TP-8	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	D41	10/21/2025	8270E
Q3431-01	EO-02-10222025	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG05	D41	10/22/2025	8270E
Q3434-01	PR132-WC1-20251022	Solid	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D41	10/22/2025	8270E

Date/Time 10/23/25 9:10
Raw Sample Received by: RJ (Ext-lab)
Raw Sample Relinquished by:

Date/Time 10/23/25 9:45
Raw Sample Received by:
Raw Sample Relinquished by: RJ (Ext-lab)

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB170286

SOP ID:	M3510C,3580A-Extraction SVOC-21				
Clean Up SOP #:	N/A			**Extraction Start Date :**	10/28/2025
Matrix :	Water			**Extraction Start Time :**	08:50
Weigh By:	N/A	**Extraction By:**	RS	**Extraction End Date :**	10/28/2025
Balance check:	N/A	**Filter By:**	RS	**Extraction End Time :**	13:50
Balance ID:	N/A	**pH Meter ID:**	N/A	**Concentration By:**	EH
pH Strip Lot#:	E3953	**Hood ID:**	4,6,7	**Supervisor By :**	RUPESH

Extraction Method: ☑ Seperatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6865
Surrogate	1.0ML	100/150 PPM	SP6869
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3980
Baked Na2SO4	N/A	EP2655
H2SO4 1:1	N/A	EP2610
10N NaoH	N/A	EP2609
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

pH Adjusted to < 2 with 1:1 H2SO4 & >12 with 10N NaOH, 1.5ML Vial Lot # 2210443.

KD Bath ID:	WATER BATH-1	**Envap ID:**	NE VAP-02
KD Bath Temperature:	60 °C	**Envap Temperature:**	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/28/25	RS (Ext lab)	JH/ SVOC
13:55	**Preparation Group**	**Analysis Group**

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB170286

Analytical Method: M3510C,3580A-Extraction SVOC-21

Concentration Date: 10/28/2025

Sample ID	Client Sample ID	Test	g /mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170286BL	SBLK286	SVOCMS Group1	1000	6	RUPESH	ritesh	1			SEP-1
PB170286BS	SLCS286	SVOCMS Group1	1000	6	RUPESH	ritesh	1			2
PB170286BSD	SLCSD286	SVOCMS Group1	1000	6	RUPESH	ritesh	1			3
Q3434-02	PR132-WC2-20251022	SVOC-TCL BNA -20	880	6	RUPESH	ritesh	1	F		4
Q3468-01	MW-06	SVOCMS Group1	980	6	RUPESH	ritesh	1	E		5
Q3468-02	MW-08	SVOCMS Group1	990	6	RUPESH	ritesh	1	E		6
Q3468-03	MW-10	SVOCMS Group1	970	6	RUPESH	ritesh	1	E		7
Q3468-04	MW-11	SVOCMS Group1	990	6	RUPESH	ritesh	1	E		8
Q3468-05	MW-13	SVOCMS Group1	1000	6	RUPESH	ritesh	1	E		9
Q3468-06	MW-12	SVOCMS Group1	950	12	RUPESH	ritesh	1	E		10

RS 10/28

* **Extracts relinquished on the same date as received.**

Q3434-SVOC-TCL BNA -20

Q3770286
8:50

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3468 **WorkList ID :** 192726 **Department :** Extraction **Date :** 10-28-2025 08:42:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3434-02	PR132-WC2-20251022	Water	SVOC-TCL BNA -20	Cool 4 deg C	AECO02	D41	10/22/2025	8270E
Q3468-01	MW-06	Water	SVOCMS Group1	Cool 4 deg C	LIRO01	D41	10/23/2025	8270E
Q3468-02	MW-08	Water	SVOCMS Group1	Cool 4 deg C	LIRO01	D41	10/23/2025	8270E
Q3468-03	MW-10	Water	SVOCMS Group1	Cool 4 deg C	LIRO01	D41	10/23/2025	8270E
Q3468-04	MW-11	Water	SVOCMS Group1	Cool 4 deg C	LIRO01	D41	10/23/2025	8270E
Q3468-05	MW-13	Water	SVOCMS Group1	Cool 4 deg C	LIRO01	D41	10/23/2025	8270E
Q3468-06	MW-12	Water	SVOCMS Group1	Cool 4 deg C	LIRO01	D41	10/23/2025	8270E

Date/Time 10/28/25 8:45
Raw Sample Received by: RS (Ext-Lab)
Raw Sample Relinquished by:

Date/Time 10/28/25 9:15
Raw Sample Received by:
Raw Sample Relinquished by: RS (Ext-Lab)

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q3434

Test : SVOC-TCL BNA -20

Prepbatch ID : PB170215,PB170286,

Sequence ID/Qc Batch ID: BF102425,BF102725,BF102825,BF102925,BG102725,BG102825,BG102925,BG102925,BG103025,

Standard ID :
EP2609,EP2610,EP2641,EP2652,EP2655,SP6856,SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864,SP6865,SP6869,SP6871,SP6872,SP6875,

Chemical ID :
10ul/1000ul sample,E3657,E3875,E3932,E3951,E3954,E3965,E3972,E3973,E3980,M6157,S11073,S11484,S11652,S11807,S11808,S12197,S12199,S12200,S12201,S12220,S12221,S12245,S12306,S12307,S12507,S12508,S12552,S12553,S12554,S12555,S12556,S12557,S12558,S12559,S12560,S12577,S12739,S12776,S12903,S12904,S12905,S13089,S13090,S13091,S13092,S13093,S13094,S13095,S13096,S13097,S13118,S13119,S13120,S13149,S13160,S13170,S13175,S13179,S13180,S13181,S13207,S13214,S13233,S13239,S13240,S13241,S13242,S13243,S13244,S13269,W3112,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1874	10 N SODIUM HYDROXIDE SOLN	EP2609	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025

FROM 1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
2017	1:1 ACETONE/METHYLENE CHLORIDE	EP2641	09/16/2025	03/16/2026	Evelyn Huang	None	None	Riteshkumar Patel 09/16/2025

FROM 8000.00000ml of E3972 + 8000.00000ml of E3973 = Final Quantity: 16000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2652	10/10/2025	01/28/2026	Evelyn Huang	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/10/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2655	10/24/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/24/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4215	New 8270E/625.1 ICAL Stock Solution 100 ng	SP6856	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.04000ml of S13207 + 0.20000ml of S12197 + 0.30000ml of S12306 + 0.40000ml of S12220 + 0.50000ml of S11807 + 0.50000ml of S12245 + 0.50000ml of S12776 + 0.50000ml of S13214 + 0.50000ml of S13233 + 0.70000ml of S12307 + 1.00000ml of S12739 + 4.86000ml of E3954 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4216	80ng ICC	SP6857	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.01000ml of S13170 + 0.20000ml of E3954 + 0.80000ml of SP6856 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4217	60ng ICC	SP6858	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.01000ml of S13170 + 0.40000ml of E3954 + 0.60000ml of SP6856 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4218	50ng ICC	SP6859	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.01000ml of S13170 + 0.50000ml of E3954 + 0.50000ml of SP6856 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4219	40ng ICC	SP6860	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.01000ml of S13170 + 0.60000ml of E3954 + 0.40000ml of SP6856 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4220	20ng ICC	SP6861	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.01000ml of S13170 + 0.80000ml of E3954 + 0.20000ml of SP6856 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4221	10ng ICC	SP6862	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.01000ml of S13170 + 0.75000ml of E3954 + 0.25000ml of SP6860 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4222	5ng ICC	SP6863	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.01000ml of S13170 + 0.87500ml of E3954 + 0.12500ml of SP6860 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4223	2.5ng ICC	SP6864	08/12/2025	12/16/2025	Jagrut Upadhyay	None	None	mohammad ahmed 08/22/2025

FROM 0.01000ml of S13170 + 0.50000ml of E3954 + 0.50000ml of SP6863 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
171	8270/625 Spike Solution, 50/100 PPM	SP6865	08/25/2025	10/30/2025	Jagrut Upadhyay	None	None	Rahul Chavli 08/26/2025

FROM 0.40000ml of S11484 + 0.40000ml of S11652 + 0.40000ml of S12552 + 0.40000ml of S13149 + 0.40000ml of S13160 + 0.40000ml of S13207 + 0.90000ml of S12507 + 0.90000ml of S13089 + 1.10000ml of S13120 + 1.20000ml of S12553 + 1.20000ml of S12554 + 1.20000ml of S12555 + 1.20000ml of S12556 + 1.20000ml of S12557 + 1.20000ml of S12558 + 1.20000ml of S12559 + 1.20000ml of S13269 + 1.30000ml of S11808 + 1.30000ml of S12508 + 1.30000ml of S13090 + 1.30000ml of S13091 + 1.30000ml of S13092 + 1.30000ml of S13093 + 1.30000ml of S13094 + 1.30000ml of S13095 + 1.30000ml of S13096 + 1.30000ml of S13118 + 1.30000ml of S13119 + 1.30000ml of S13239 + 1.30000ml of S13240 + 1.30000ml of S13241 + 1.30000ml of S13242 + 1.30000ml of S13243 + 1.30000ml of S13244 + 163.00000ml of E3932 = Final Quantity: 200.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
19	8270/CLP Surrogate Solution, 100 PPM BN/150 PPM ACID	SP6869	09/10/2025	01/02/2026	Jagrut Upadhyay	None	None	Rahul Chavli 09/16/2025

FROM 3.00000ml of S12197 + 3.00000ml of S12220 + 5.60000ml of S12221 + 5.60000ml of S12903 + 5.80000ml of S12904 + 6.00000ml of S12199 + 6.00000ml of S12200 + 965.00000ml of E3965 = Final Quantity: 1000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
18	Second Source Calibration Stock Standard, 100 PPM, (8270/625/CLP)	SP6871	09/22/2025	11/16/2025	Jagrut Upadhyay	None	None	Yogesh Patel 09/27/2025

FROM 0.04000ml of S12201 + 0.08000ml of S12905 + 0.10000ml of S11073 + 0.20000ml of S12560 + 0.20000ml of S13097 + 0.20000ml of S13244 + 1.18000ml of E3973 = Final Quantity: 2.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
416	40 ng BNA ICV, 40 PPM	SP6872	09/22/2025	11/16/2025	Jagrut Upadhyay	None	None	Yogesh Patel 09/27/2025

FROM 0.01000ml of S13175 + 0.60000ml of E3973 + 0.40000ml of SP6871 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3895	50 ug/ml DFTPP 8270E	SP6875	09/30/2025	03/30/2026	Rahul Chavli	None	None	Jagrut Upadhyay 09/30/2025

FROM 1.00000ml of S12577 + 19.00000ml of E3973 = Final Quantity: 20.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	07/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3932

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	25A2756718	12/31/2028	07/09/2025 / RUPESH	04/28/2020 / RUPESH	E3951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25B1862001	03/19/2026	07/14/2025 / RUPESH	06/11/2025 / RUPESH	E3954

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	08/22/2025 / RUPESH	08/20/2025 / RUPESH	E3965

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	09/15/2025 / Riteshkumar	09/15/2025 / Riteshkumar	E3973

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	10/10/2025 / RUPESH	10/10/2025 / RUPESH	E3980

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0187043	11/16/2025	05/16/2025 / Jagrut	02/06/2023 / Christian	S11073

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555870 / Custom Standard, 2,4-dinitrophenol Std [CS 5328-3]	A0200549	10/30/2025	04/30/2025 / Rahul	08/10/2023 / yogesh	S11484

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	01/30/2026	07/30/2025 / Rahul	11/09/2023 / Yogesh	S11652

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0200655	01/01/2026	07/01/2025 / Rahul	11/21/2023 / rahul	S11807

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0200655	02/25/2026	08/25/2025 / Jagrut	11/21/2023 / rahul	S11808

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	01/02/2026	07/02/2025 / Jagrut	03/15/2024 / Rahul	S12197

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	03/10/2026	09/10/2025 / Jagrut	03/15/2024 / Rahul	S12199

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	03/10/2026	09/10/2025 / Jagrut	03/15/2024 / Rahul	S12200

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	03/22/2026	09/22/2025 / Jagrut	03/15/2024 / Rahul	S12201

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	01/02/2026	07/02/2025 / Jagrut	03/15/2024 / Rahul	S12220

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	03/10/2026	09/10/2025 / Jagrut	03/15/2024 / Rahul	S12221

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30409 / Pyridine, 2000 PPM in P & T Methanol	A0206650	12/16/2025	06/16/2025 / Jagrut	05/14/2024 / Rahul	S12245

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31902 / CLP/SVOA Additions Mix (Atrazine, Benzaldehyde, Caprolactam) 1000ug/mL	A0206859	01/31/2026	08/12/2025 / Jagrut	05/30/2024 / Rahul	S12306

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31902 / CLP/SVOA Additions Mix (Atrazine, Benzaldehyde, Caprolactam) 1000ug/mL	A0206859	01/31/2026	08/12/2025 / Jagrut	05/30/2024 / Rahul	S12307

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	01/01/2026	07/01/2025 / Rahul	07/23/2024 / RAHUL	S12507

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	02/25/2026	08/25/2025 / Jagrut	07/23/2024 / RAHUL	S12508

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	01/01/2026	07/01/2025 / Rahul	07/23/2024 / RAHUL	S12552

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	02/25/2026	08/25/2025 / Jagrut	07/23/2024 / RAHUL	S12553

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	02/25/2026	08/25/2025 / Jagrut	07/23/2024 / RAHUL	S12554

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	02/25/2026	08/25/2025 / Jagrut	07/23/2024 / RAHUL	S12555

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	02/25/2026	08/25/2025 / Jagrut	07/23/2024 / RAHUL	S12556

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	02/25/2026	08/25/2025 / Jagrut	07/23/2024 / RAHUL	S12557

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	02/25/2026	08/25/2025 / Jagrut	07/23/2024 / RAHUL	S12558

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	02/25/2026	08/25/2025 / Jagrut	07/23/2024 / RAHUL	S12559

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	03/22/2026	09/22/2025 / Jagrut	07/23/2024 / RAHUL	S12560

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL,	A0212955	06/30/2027	03/31/2025 / Rahul	08/01/2024 / Rahul	S12577

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31900 / SOM01.1 Mega Mix, 500-1000 ug/ml	A0215529	02/12/2026	08/12/2025 / Jagrut	10/08/2024 / anahy	S12739

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	90494 / 1-Methylnaphthalene, 2000 ug/mL, in methylene chloride	061323	02/12/2026	08/12/2025 / Jagrut	11/08/2024 / anahy	S12776

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0216785	03/10/2026	09/10/2025 / Jagrut	12/09/2024 / anahy	S12903

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0216785	03/10/2026	09/10/2025 / Jagrut	12/09/2024 / anahy	S12904

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0216785	03/22/2026	09/22/2025 / Jagrut	12/09/2024 / anahy	S12905

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	07/01/2025 / Rahul	05/20/2025 / Rahul	S13089

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13090

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13091

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13092

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13093

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13094

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13095

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13096

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0221014	11/30/2025	09/03/2025 / Jagrut	05/20/2025 / Rahul	S13097

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0218894	02/25/2026	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13118

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0218894	02/25/2026	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13119

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0218894	02/25/2026	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13120

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555869 / Custom Standard, hexachlorocyclopentadiene Std [CS 5328-2]	A0201702	02/25/2026	08/25/2025 / Jagrut	11/13/2023 / Rahul	S13149

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4]	A0226283	02/25/2026	08/25/2025 / Jagrut	06/04/2025 / Rahul	S13160

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0224359	02/11/2026	08/11/2025 / Rahul	06/02/2025 / anahy	S13170

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0224359	03/15/2026	09/15/2025 / rahul	06/02/2025 / anahy	S13175

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0224359	04/17/2026	10/17/2025 / rahul	06/02/2025 / anahy	S13179

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0224359	03/31/2031	03/31/2025 / anahy	06/02/2025 / anahy	S13180

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0224359	03/31/2031	03/31/2025 / anahy	06/02/2025 / anahy	S13181

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555868 / Custom Standard, Benzidine Std [CS 5328-1]	A0226493	02/12/2026	08/12/2025 / Jagrut	06/11/2025 / anahy	S13207

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31879 / Benzoic Acid, 2000 µg/mL, Methylene Chloride, 1 mL/ampul	A0221395	02/12/2026	08/12/2025 / Jagrut	07/10/2025 / anahy	S13214

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	582978 / Custom Calibration Standard - C8 2,000µg/mL, Methylene chloride, 1mL/ampul, Chromatographic, CS-347186	A0228192	02/12/2026	08/12/2025 / Jagrut	08/01/2025 / Rahul	S13233

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0228451	02/25/2026	08/25/2025 / Jagrut	08/06/2025 / Rahul	S13239

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0228451	02/25/2026	08/25/2025 / Jagrut	08/06/2025 / Rahul	S13240

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0228451	02/25/2026	08/25/2025 / Jagrut	08/06/2025 / Rahul	S13241

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0228451	02/25/2026	08/25/2025 / Jagrut	08/06/2025 / Rahul	S13242

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0228451	02/25/2026	08/25/2025 / Jagrut	08/06/2025 / Rahul	S13243

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0228451	02/25/2026	08/25/2025 / Jagrut	08/06/2025 / Rahul	S13244

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0228494	02/25/2026	08/25/2025 / Jagrut	08/06/2025 / Rahul	S13269

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on 02/06/23 by CG

S11071 to S11075

Catalog No. : 31853 **Lot No.:** A0187043

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	1,4-Dioxane	2,019.0 µg/mL	+/- 11.8486	µg/mL	Gravimetric
	CAS # 123-91-1 (Lot SHBN5929)		+/- 43.2570	µg/mL	Unstressed
	Purity 99%		+/- 44.5129	µg/mL	Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)


Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Brittany Federinko - Operations Tech I

Date Mixed: 07-Jul-2022 **Balance:** 1128360905

Marlina Cowan – Operations Tech II ARM QC

Date Passed: 12-Jul-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Rec on 8/20/25

E3965

JCroak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

VWR®

Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Pellets

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

E 3657	E 3659
E 3654	E 3660

PRODUCTOS QUIMICOS MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE **RMB3375**)	FORMULA :	Na_2SO_4
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na_2SO_4)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO_4)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreing matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD RS

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3932

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

VWR® BDH® CHEMICALS

Certificate of Analysis

Material	BDH9274-2.5KG
Material Description	BDH SAND STDD OTTAWA W+I 2.5KG
Grade	NOT APPLICABLE
Batch	25A2756718
Reassay Date	12/31/2028
CAS Number	14808-60-7
Molecular Formula	SiO2
Molecular Mass	60.09
Date of Manufacture	12/05/2024
Storage	Room Temperature

Characteristics	Specifications	Measured Values
Appearance	Beige granules.	Beige granules.
Moisture	<= 0.1 %	0.1 %
Particle Size 30-40 mesh	>= 80 %	99 %
CUSTOMER PART # BDH9274-2.5KG		

Received on 7/9/25.

E3951

Internal ID #: 793

Signature	Additional Information
We certify that this batch conforms to the specifications listed above.	Analysis may have been rounded to significant digits in specification limits
This document has been electronically produced and is valid without a signature.	Product meets analytical specifications of the grades listed.
Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC. 28600 Fountain Parkway, Solon OH 44139 USA	

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™ | J.T.Baker

Material No.: 9266-A4
Batch No.: 25B1862001
Manufactured Date: 2024-12-18
Expiration Date:2026-03-19
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	99.9 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.3 ppm
Titrable Acid (µeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

RS 7/14/25

E3954

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 25C1262005
Manufactured Date: 2025-01-15
Expiration Date:2026-04-16
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Titrable Acid (µeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Rec. on 6/10/25 PJ

E3980

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

M615 MB

avantor™ | J.T.Baker

Material No.: 9673-33
Batch No.: 24I1262013
Manufactured Date: 2024-08-07
Retest Date:2029-08-06
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H_2SO_4)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO2)	<= 2 ppm	<2 ppm
Ammonium (NH_4)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO_3)	<= 0.2 ppm	0.1 ppm
Phosphate (PO_4)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™ | J.T.Baker

Material No.: 9673-33
Batch No.: 24I1262013

Test	Specification	Result
Trace Impurities - Sodium (Na)	<= 500.0 ppb	<5.0 ppb
Trace Impurities - Strontium (Sr)	<= 5.0 ppb	<1.0 ppb
Trace Impurities - Tin (Sn)	<= 5.0 ppb	1.1 ppb
Trace Impurities - Zinc (Zn)	<= 5.0 ppb	<1.0 ppb

For Laboratory,Research,or Manufacturing Use

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

J Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555870 **Lot No.:** A0200549

Description : Custom 2,4-Dinitrophenol Standard

Custom 2,4-Dinitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

S11484 ↓ S11493 Y.P. 08/10/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	25,008.0 µg/mL	+/- 777.3323

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Tom Suckar - Mix Technician **Date Mixed:** 02-Aug-2023 **Balance:** 1128342314

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0201728

Description : Custom Pentachlorophenol Standard

Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

511649 ↓ 511658 Y.P. 11/13/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023 **Balance:** B251644995

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0200655

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11795 ↓ S11808 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBQ1693	99%	2,007.0 µg/mL	+/- 24.9775

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 06-Aug-2023 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 08-Aug-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

S12187 ↓ S12206 } RC/ 03/18/24

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

S12207 ↓ S12221 RC/ 03/18/24

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30409 **Lot No.:** A0206650

Description : Pyridine Standard

Pyridine 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

S12242 ↓ S12254 RC/ 5/15/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	2,020.0 µg/mL	+/- 33.0924

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 16-Jan-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 18-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

S12302 ↓ S12311 RC 5/30/24

Catalog No. : 31902 **Lot No.:** A0206859

Description : Additions Standard

Additions Standard 1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.5419
2	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.8 µg/mL	+/- 29.6521
3	Atrazine	1912-24-9	5FYWL	99%	1,008.8 µg/mL	+/- 29.6521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 23-Jan-2024 **Balance Serial #** B442140311

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31615 **Lot No.:** A0212955

Description : GC/MS Tuning Mixture

GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2027 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,004.5 µg/mL	+/- 44.8902
2	DFTPP (Decafluorotriphenylphosphine)	5074-71-5	Q117-147	99%	1,004.5 µg/mL	+/- 44.8902
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 44.9572
4	4,4'-DDT	50-29-3	S240530RSR	97%	1,000.1 µg/mL	+/- 44.6922

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12577 ↓ S12579 RC 8/2/24

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 19-Jun-2024 **Balance Serial #** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 26-Jun-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31900 **Lot No.:** A0215529

Description : OLM 01.1 Revised SV MegaMix

OLM 01.1 Revised SV MegaMix 500-1000 µg/mL, Methylene chloride, 1mL/ampul

S12736 ↓ S12754 AC 10/9/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Phenol	108-95-2	MKCK1120	99%	1,008.6 µg/mL	+/- 19.3211
2	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,002.3 µg/mL	+/- 19.2013
3	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.2 µg/mL	+/- 19.2564
4	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,006.3 µg/mL	+/- 19.2768
5	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,009.3 µg/mL	+/- 19.3354
6	Acetophenone	98-86-2	STBH8205	99%	1,003.8 µg/mL	+/- 18.5851
7	Hexachloroethane	67-72-1	QTORH	99%	1,000.6 µg/mL	+/- 19.1690
8	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,004.5 µg/mL	+/- 19.2599
9	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	500.3 µg/mL	+/- 9.5854
10	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 9.6213
11	Nitrobenzene	98-95-3	10224044	99%	1,003.2 µg/mL	+/- 19.2181
12	Isophorone	78-59-1	MKCC9506	99%	1,007.0 µg/mL	+/- 19.2911
13	2-Nitrophenol	88-75-5	RP230509C	99%	1,003.4 µg/mL	+/- 19.2217
14	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,008.1 µg/mL	+/- 19.3115
15	Bis(2-chloroethoxy)methane	111-91-1	15174900	99%	1,002.3 µg/mL	+/- 19.2001
16	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,002.3 µg/mL	+/- 19.2001

17	Naphthalene	91-20-3	STBL1057	99%	1,003.6	µg/mL	+/- 19.2253
18	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,005.5	µg/mL	+/- 19.2791
19	Hexachlorobutadiene	87-68-3	RP240110CTH	97%	1,003.9	µg/mL	+/- 19.2316
20	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,005.1	µg/mL	+/- 19.2718
21	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.8	µg/mL	+/- 19.2301
22	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,000.5	µg/mL	+/- 19.1832
23	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 19.1811
24	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,004.1	µg/mL	+/- 19.2349
25	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,008.6	µg/mL	+/- 19.3221
26	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,002.3	µg/mL	+/- 19.2001
27	Biphenyl	92-52-4	MKCS5928	99%	1,001.3	µg/mL	+/- 18.5388
28	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,000.5	µg/mL	+/- 19.1832
29	Acenaphthylene	208-96-8	214935V16F	97%	999.9	µg/mL	+/- 19.1561
30	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.8	µg/mL	+/- 19.2672
31	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,001.2	µg/mL	+/- 19.1798
32	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 19.1570
33	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,005.5	µg/mL	+/- 19.2791
34	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,008.8	µg/mL	+/- 19.3259
35	Dibenzofuran	132-64-9	MKCD9952	99%	1,002.5	µg/mL	+/- 18.5619
36	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,002.5	µg/mL	+/- 19.2049
37	4-Nitrophenol	100-02-7	RP230627	99%	1,004.4	µg/mL	+/- 19.2421
38	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,001.5	µg/mL	+/- 19.2024
39	Fluorene	86-73-7	10241100	99%	1,004.2	µg/mL	+/- 19.2373
40	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.0	µg/mL	+/- 19.2145
41	Diethylphthalate	84-66-2	BCCJ6241	99%	1,002.1	µg/mL	+/- 19.1978
42	4-Nitroaniline	100-01-6	RP240510RSR	99%	1,005.0	µg/mL	+/- 19.2695
43	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S240410RSR	99%	1,001.1	µg/mL	+/- 19.1786
44	Diphenylamine	122-39-4	MKCT1512	99%	1,004.5	µg/mL	+/- 19.2599
45	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,005.9	µg/mL	+/- 19.2708
46	Hexachlorobenzene	118-74-1	15458400	99%	1,009.2	µg/mL	+/- 19.3331
47	Pentachlorophenol	87-86-5	RP240411RSR	99%	1,008.9	µg/mL	+/- 19.3283
48	Phenanthrene	85-01-8	MKCS5188	99%	1,006.2	µg/mL	+/- 19.2756
49	Anthracene	120-12-7	101492T18R	99%	1,001.6	µg/mL	+/- 19.1882
50	Carbazole	86-74-8	15276700	99%	1,004.0	µg/mL	+/- 19.2503
51	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,002.5	µg/mL	+/- 19.2049
52	Fluoranthene	206-44-0	MKCQ4728	99%	1,008.7	µg/mL	+/- 19.3235

53	Pyrene	129-00-0	BCCK2592	99%	1,002.9	µg/mL	+/- 19.2121
54	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,004.6	µg/mL	+/- 19.2444
55	Benz(a)anthracene	56-55-3	I50012022BAA	99%	1,009.1	µg/mL	+/- 19.3307
56	Chrysene	218-01-9	RP240719RSR	99%	1,005.9	µg/mL	+/- 19.2708
57	3,3'-Dichlorobenzidine	91-94-1	S231019RSR	99%	1,001.5	µg/mL	+/- 19.2024
58	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,006.3	µg/mL	+/- 19.2768
59	Di-n-octyl phthalate	117-84-0	15276800	99%	1,008.9	µg/mL	+/- 19.3283
60	Benzo(b)fluoranthene	205-99-2	022013B	99%	1,006.4	µg/mL	+/- 19.2792
61	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,001.9	µg/mL	+/- 19.1942
62	Benzo(a)pyrene	50-32-8	O45GL	98%	1,003.9	µg/mL	+/- 19.2327
63	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,003.7	µg/mL	+/- 19.2281
64	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,004.2	µg/mL	+/- 19.2373
65	Benzo(g,h,i)perylene	191-24-2	RP240625RSR	97%	1,001.5	µg/mL	+/- 19.1863

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 90494
Lot Number: 061323
Description: 1-Methylnaphthalene

Expiration Date: 061328
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 6UTB


Solvent(s): **Lot#**
Methylene chloride C21F09CAS0000DCM


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 100.0 0.031 Flask Uncertainty

Formulated By:	Prashant Chauhan	061323 DATE
Reviewed By:	Pedro L. Rentas	061323 DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 1-Methylnaphthalene	313	04413BX	2000	98	0.2	0.20417	0.20430	2001.2	8.3	90-12-0	N/A	orl-rat 1840mg/kg

Method GC8MSD-3.M: Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 200°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Gina McLane.





S12771 ↓ S12780 } AC 11/8/24

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0221014

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S13088 ↓ S13117 RC 5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBR4811	99%	1,005.2 µg/mL	+/- 36.5730
2	N-Nitrosodimethylamine	62-75-9	S241226RSR	99%	1,005.5 µg/mL	+/- 36.5848
3	Phenol	108-95-2	MKCK1120	99%	1,005.3 µg/mL	+/- 36.5780
4	Aniline	62-53-3	X22F726	99%	1,005.4 µg/mL	+/- 36.5816
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,004.7 µg/mL	+/- 36.5562
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.3 µg/mL	+/- 36.5766
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.6 µg/mL	+/- 36.5530
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,006.3 µg/mL	+/- 36.6125
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,005.5 µg/mL	+/- 36.5853
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,004.6 µg/mL	+/- 36.5507
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,004.8 µg/mL	+/- 36.5575
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,005.4 µg/mL	+/- 36.5803
13	3-Methylphenol (m-cresol)	108-39-4	STBL3873	99%	502.7 µg/mL	+/- 18.2888
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	502.6 µg/mL	+/- 18.2869
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,004.6 µg/mL	+/- 36.5502
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.6 µg/mL	+/- 36.5530
17	Nitrobenzene	98-95-3	10224044	99%	1,005.3 µg/mL	+/- 36.5780

18	Isophorone	78-59-1	MKCR3249	99%	1,004.6	µg/mL	+/-	36.5511
19	2-Nitrophenol	88-75-5	RP230710	99%	1,005.7	µg/mL	+/-	36.5916
20	2,4-Dimethylphenol	105-67-9	DIRAF	99%	1,004.9	µg/mL	+/-	36.5612
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,004.3	µg/mL	+/-	36.5402
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,004.7	µg/mL	+/-	36.5571
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,004.6	µg/mL	+/-	36.5516
24	Naphthalene	91-20-3	STBL1057	99%	1,006.8	µg/mL	+/-	36.6335
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,005.9	µg/mL	+/-	36.5980
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,004.1	µg/mL	+/-	36.5328
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,005.4	µg/mL	+/-	36.5793
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,006.3	µg/mL	+/-	36.6144
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/-	36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063P13G	99%	1,005.9	µg/mL	+/-	36.5975
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,004.7	µg/mL	+/-	36.5566
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.7	µg/mL	+/-	36.5571
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,005.3	µg/mL	+/-	36.5757
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.8	µg/mL	+/-	36.5584
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,004.5	µg/mL	+/-	36.5475
36	Acenaphthylene	208-96-8	214935V18H	95%	1,000.1	µg/mL	+/-	36.3888
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,004.9	µg/mL	+/-	36.5616
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,004.5	µg/mL	+/-	36.5489
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,005.2	µg/mL	+/-	36.5734
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,005.9	µg/mL	+/-	36.6003
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/-	36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/-	36.5525
43	2,4-Dinitrophenol	51-28-5	D240927RSR	99%	1,005.0	µg/mL	+/-	36.5657
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,005.4	µg/mL	+/-	36.5803
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,005.4	µg/mL	+/-	36.5803
46	4-Nitrophenol	100-02-7	20241120-1-AN	99%	1,005.7	µg/mL	+/-	36.5930
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,004.9	µg/mL	+/-	36.5616
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP240523RSR	99%	1,005.2	µg/mL	+/-	36.5739
49	Fluorene	86-73-7	10246250	98%	1,005.8	µg/mL	+/-	36.5948
50	4-Chlorophenyl phenyl ether	7005-72-3	002531K02D	99%	1,004.8	µg/mL	+/-	36.5584
51	Diethylphthalate	84-66-2	STBL3611	99%	1,005.3	µg/mL	+/-	36.5762
52	4-Nitroaniline	100-01-6	RP240830RSR	99%	1,005.5	µg/mL	+/-	36.5844
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,004.8	µg/mL	+/-	36.5589

54	Diphenylamine	122-39-4	MKCT1512	99%	1,005.2	µg/mL	+/- 36.5725
55	Azobenzene	103-33-3	BCCL3292	99%	1,004.7	µg/mL	+/- 36.5566
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,005.6	µg/mL	+/- 36.5875
57	Hexachlorobenzene	118-74-1	15828800	99%	1,005.3	µg/mL	+/- 36.5789
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.4	µg/mL	+/- 36.5825
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.8	µg/mL	+/- 36.5584
60	Anthracene	120-12-7	MKCW9141	99%	1,005.5	µg/mL	+/- 36.5834
61	Carbazole	86-74-8	15630800	99%	1,005.3	µg/mL	+/- 36.5789
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,005.5	µg/mL	+/- 36.5848
63	Fluoranthene	206-44-0	A0458721	99%	1,005.4	µg/mL	+/- <0.0001
64	Pyrene	129-00-0	BCCL8032	99%	1,005.2	µg/mL	+/- 36.5734
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,004.8	µg/mL	+/- 36.5584
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCR8567	99%	1,004.5	µg/mL	+/- 36.5480
67	Benz(a)anthracene	56-55-3	I60012022BAA	99%	1,005.5	µg/mL	+/- 36.5844
68	Chrysene	218-01-9	RP241212RSR	99%	1,005.5	µg/mL	+/- 36.5848
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,004.7	µg/mL	+/- 36.5548
70	Di-n-octyl phthalate	117-84-0	15817300	99%	1,004.8	µg/mL	+/- 36.5607
71	Benzo(b)fluoranthene	205-99-2	022013B	99%	1,005.1	µg/mL	+/- 36.5716
72	Benzo(k)fluoranthene	207-08-9	012022K	98%	1,004.6	µg/mL	+/- 36.5511
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,004.2	µg/mL	+/- 36.5377
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,004.1	µg/mL	+/- 36.5337
75	Dibenz(a,h)anthracene	53-70-3	712061504-1-1	99%	1,005.7	µg/mL	+/- 36.5930
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,005.4	µg/mL	+/- 36.5814

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 12-Jan-2025 **Balance Serial #** 1128360905

Dillan Murphy - Operations Technician I

Date Passed: 21-Jan-2025

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0218894

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2029 **Storage:** 0°C or colder

Ship: Ambient

513118 ↓ 513147 RC/ 5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBQ1693	99%	2,002.4 µg/mL	+/- 24.9202

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 07-Nov-2024 **Balance Serial #** 1128360905

Dillan Murphy - Operations Technician I

Date Passed: 11-Nov-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555871 **Lot No.:** A0226283

Description : Custom 4-Nitrophenol Standard

Custom 4-Nitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

S13158 ↓ S13167 } RC/ 6/4/25

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	4-Nitrophenol	100-02-7	20241120-1-AN	99%	25,192.0 µg/mL	+/- 783.0517

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Morgan Craighead - Mix Technician

Date Mixed: 02-Jun-2025 **Balance:** C322230531

REVIEWED By Dillon Murphy at 11:21 am, Jun 03, 2025

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

S13198 ↓ S13207 } AC 6/11/25

Catalog No. : 555868 **Lot No.:** A0226493

Description : Custom Benzidine Standard

Custom Benzidine Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2028 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Benzidine	92-87-5	S250227ECS	99%	25,004.0 µg/mL	+/- 495.8040

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Laith Clemente - Operations Technician I

Date Mixed: 09-Jun-2025 **Balance:** 1122030677

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31879 **Lot No.:** A0221395

Description : Benzoic Acid Mix

Benzoic Acid 2000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2029 **Storage:** 10°C or colder

Ship: Ambient

S13214 ↓ S13218 AC 7/10/25

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Benzoic acid	65-85-0	MKCR2694	99%	2,002.2 µg/mL	+/- 60.5426

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 582978 **Lot No.:** A0228192

513229 ↓ 513238 } RC/ 8/1/25

Description : Custom Calibration Standard - C8

Custom Calibration Standard - C8 2,000µg/mL, Methylene chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	N-Nitrosodimethylamine	62-75-9	S241226RSR	99%	2,000.0 µg/mL	+/- 35.7537
2	Aniline	62-53-3	X22F726	99%	2,002.5 µg/mL	+/- 35.7984
3	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	2,005.0 µg/mL	+/- 35.8431
4	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	2,005.0 µg/mL	+/- 35.8431
5	Benzyl alcohol	100-51-6	094986W07G	99%	2,015.0 µg/mL	+/- 36.0218
6	1,2-Dichlorobenzene	95-50-1	SHBR4446	99%	2,015.0 µg/mL	+/- 36.0218
7	1,2,4-Trichlorobenzene	120-82-1	SHBR1701	99%	2,020.0 µg/mL	+/- 36.1112
8	Azobenzene	103-33-3	BCCL3292	99%	2,005.0 µg/mL	+/- 35.8431

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Wilner Torres - Operation Tech I **Date Mixed:** 27-Jul-2025 **Balance Serial #** B345965662

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 30-Jul-2025 REVIEWED

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0228451

S13239 ↓ S13268 } RC 08/06/25

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2027 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S250226RSR	99%	1,006.0 µg/mL	+/- 23.0947
2	Atrazine	1912-24-9	5FYWL	99%	1,007.0 µg/mL	+/- 23.1176
3	Benzidine	92-87-5	S250227ECS	99%	1,003.0 µg/mL	+/- 23.0258
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,003.0 µg/mL	+/- 23.0258

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tom Suckar - Mix Technician **Date Mixed:** 01-Aug-2025 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0228451

S13239 ↓ S13268 } RC 08/06/25

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2027 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S250226RSR	99%	1,006.0 µg/mL	+/- 23.0947
2	Atrazine	1912-24-9	5FYWL	99%	1,007.0 µg/mL	+/- 23.1176
3	Benzidine	92-87-5	S250227ECS	99%	1,003.0 µg/mL	+/- 23.0258
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,003.0 µg/mL	+/- 23.0258

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tom Suckar - Mix Technician **Date Mixed:** 01-Aug-2025 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0228494

S13269 ↓ S13298 } RC 08/06/25

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,002.0 µg/mL	+/- 29.453715
2	Acetophenone	98-86-2	STBH8205	99%	1,001.0 µg/mL	+/- 29.424320
3	Benzaldehyde	100-52-7	RD250319RSR	99%	1,003.0 µg/mL	+/- 29.483110
4	Benzoic acid	65-85-0	MKCX1578	99%	1,001.0 µg/mL	+/- 29.424320
5	Biphenyl	92-52-4	MKCS5928	99%	1,002.0 µg/mL	+/- 29.453715

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Laith Clemente - Operations Technician I **Date Mixed:** 04-Aug-2025 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555869 **Lot No.:** A0201702

Description : Custom Hexachlorocyclopentadiene Standard

Custom Hexachlorocyclopentadiene Standard 25,000µg/mL, Methanol, 1mL/ampul

513148 ↓ 513157 RC 11/13/23

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Hexachlorocyclopentadiene	77-47-4	099063P13G	99%	25,244.0 µg/mL	+/- 450.6896

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Brittany Federinko - Operations Tech I

Date Mixed: 05-Sep-2023 **Balance:** B707717271

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

SHIPPING DOCUMENTS

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 · Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO.
QUOTE NO. Q3434
COC Number 2045129

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: AECOM
ADDRESS: 605 3rd Ave
CITY New York STATE: NY ZIP: 10158
ATTENTION: Robert Forstner
PHONE: 212-377-8721 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Port Richmond 132
PROJECT NO.: 60693795 LOCATION: Staten Isl
PROJECT MANAGER: Rob Forstner
e-mail: Robert.Forstner@aecom.com
PHONE: 11 FAX:

CLIENT BILLING INFORMATION

BILL TO: AECOM PO#:
ADDRESS: 11
CITY STATE: ZIP:
ATTENTION: Rob Forstner PHONE: 11

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC + Raw Data) ☐ NYS ASP A ☑ NYS ASP B
☐ Other _____
☐ EDD FORMAT _____

ANALYSIS

1 SVOC
2 Pesticides TCL
3 Hg Metals
4 Cr Cr6 Cu Pb Ni Zn
5 TPH
6 VOC
7 Flash Point
8
9

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HN03 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	1	2	3	4	5	6	7	8	9
1.	PR132-WC1-20251022	S		X	10/22	10:18	9	X	X	X	X	X	X			
2.	PR132-WC2-20251022	A		X	10/22	10:30	7	X	x	X	X	X	X	X		
3.	TB	A			10/22		3									
4.																
5.																
6.																
7.																
8.																
9.																
10.																

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. Jim Alderson	10/22/25 1pm	1. 1300 10.22.25
2.		2.
3.	10-22-25 1416	3.

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.8 °C
Comments:

Page ____ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete ☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3434 AECO02
Client Name : AECOM
Client Contact : Rob Forstner
Invoice Name : AECOM
Invoice Contact : Rob Forstner

Order Date : 10/22/2025 1:40:00 PM
Project Name : AE1-CTY 3.2.Z-PR-DOCK
Receive DateTime : 10/22/2025 2:16:00 PM
Purchase Order :

Project Mgr :
Report Type : NYS ASP B
EDD Type : EQUIS
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3434-01	PR132-WC1-20251022	Solid	10/22/2025	10:18					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q3434-02	PR132-WC2-20251022	Water	10/22/2025	10:30					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q3434-03	TB	Water	10/22/2025	10:00					
					VOC-TCLVOA-10		8260D		10 Bus. Days

Relinguished By :
Date / Time : 10-22-25 1505

Received By : Sam
Date / Time : 10/22/25 1515

Storage Area : VOA Refridgerator Room

Rq +6
Rq +2
Rq +5