

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3434

Client ID : SSTDCCC040

Date Analyzed: 10/28/2025

Lab File ID: BF144097.D

Time Analyzed: 13:17

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	66521	6.881	250146	8.16	147935	9.92
UPPER LIMIT	133042	7.381	500292	8.663	295870	10.416
LOWER LIMIT	33260.5	6.381	125073	7.663	73967.5	9.416
EPA SAMPLE NO.						
01 PR132-WC1-20251022	65453	6.88	248016	8.16	145676	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.