

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3434

Client ID : SSTDCCC040

Date Analyzed: 10/30/2025

Lab File ID: BG064636.D

Time Analyzed: 11:36

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	32501	8.219	127691	11.05	94503	14.84
UPPER LIMIT	65002	8.719	255382	11.545	189006	15.341
LOWER LIMIT	16250.5	7.719	63845.5	10.545	47251.5	14.341
EPA SAMPLE NO.						
01 PB170286BL	25136	8.22	99198	11.04	73729	14.84
02 PB170286BS	30784	8.22	124677	11.04	97021	14.84
03 PB170286BSD	30191	8.22	121649	11.04	100532	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.