

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3434
 Lab File ID: VV039306.D Date Analyzed: 11/04/2025
 Instrument ID: MSVOA_V Time Analyzed: 11:21
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	637264	4.60	1024760	5.53	1022020	8.77
UPPER LIMIT	1274530	5.1	2049510	6.032	2044030	9.273
LOWER LIMIT	318632	4.1	512378	5.032	511008	8.273
EPA SAMPLE NO.						
PR132-WC2-20251022	659851	4.60	1187277	5.53	1121779	8.77
PR132-WC2-20251022RE	632494	4.60	1126953	5.53	1080628	8.77
TB	659982	4.60	1223717	5.53	1158276	8.77
VV1104WBL01	1064589	4.60	1972523	5.53	1793346	8.77
VV1104WBS01	733220	4.60	1149929	5.53	1129519	8.77
VV1104WBSD01	706395	4.60	1110419	5.53	1079972	8.77

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.