

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3434
 Lab File ID: VV039306.D Date Analyzed: 11/04/2025
 Instrument ID: MSVOA_V Time Analyzed: 11:21
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	572200	11.172				
UPPER LIMIT	1144400	11.672				
LOWER LIMIT	286100	10.672				
EPA SAMPLE NO.						
PR132-WC2-20251022	499329	11.17				
PR132-WC2-20251022RE	492663	11.17				
TB	511635	11.17				
VV1104WBL01	828938	11.17				
VV1104WBS01	636116	11.17				
VV1104WBSD01	618050	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.