

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

|                |                   |                |                   |
|----------------|-------------------|----------------|-------------------|
| Lab Name:      | <u>Alliance</u>   | Contract:      | <u>AECO02</u>     |
| Lab Code:      | <u>ACE</u>        | SDG NO.:       | <u>Q3434</u>      |
| Lab File ID:   | <u>VY023559.D</u> | Date Analyzed: | <u>10/24/2025</u> |
| Instrument ID: | <u>MSVOA_Y</u>    | Time Analyzed: | <u>08:56</u>      |
| GC Column:     | <u>RXI-624</u>    | ID:            | <u>0.25</u> (mm)  |
|                |                   | Heated Purge:  | (Y/N) <u>Y</u>    |

|                    | IS4<br>AREA # | RT #   |  |  |  |  |
|--------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD        | 380279        | 13.346 |  |  |  |  |
| UPPER LIMIT        | 760558        | 13.846 |  |  |  |  |
| LOWER LIMIT        | 190140        | 12.846 |  |  |  |  |
| EPA SAMPLE NO.     |               |        |  |  |  |  |
| PR132-WC1-20251022 | 505336        | 13.35  |  |  |  |  |
| VY1024SBL01        | 469443        | 13.35  |  |  |  |  |
| VY1024SBS01        | 380713        | 13.35  |  |  |  |  |
| VY1024SBSD01       | 360721        | 13.35  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.