

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/09/2025 10/09/2025

Continuing Calib Time: 09:47

Initial Calibration Time(s): 14:14 22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.49       | 5.49      | 5.39      | 5.59  | 0.00       |
| Aroclor-1248-2 (2)   | 5.76       | 5.76      | 5.66      | 5.86  | 0.00       |
| Aroclor-1248-3 (3)   | 5.96       | 5.96      | 5.86      | 6.06  | 0.00       |
| Aroclor-1248-4 (4)   | 6.36       | 6.36      | 6.26      | 6.46  | 0.00       |
| Aroclor-1248-5 (5)   | 6.40       | 6.40      | 6.30      | 6.50  | 0.00       |
| Tetrachloro-m-xylene | 4.35       | 4.36      | 4.26      | 4.46  | 0.01       |
| Decachlorobiphenyl   | 9.94       | 9.93      | 9.83      | 10.03 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/09/2025 10/09/2025

Continuing Calib Time: 09:47

Initial Calibration Time(s): 14:14 22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.73       | 4.74      | 4.64      | 4.84 | 0.01       |
| Aroclor-1248-2 (2)   | 4.97       | 4.98      | 4.88      | 5.08 | 0.01       |
| Aroclor-1248-3 (3)   | 5.01       | 5.02      | 4.92      | 5.12 | 0.01       |
| Aroclor-1248-4 (4)   | 5.18       | 5.19      | 5.09      | 5.29 | 0.01       |
| Aroclor-1248-5 (5)   | 5.57       | 5.58      | 5.48      | 5.68 | 0.01       |
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56      | 3.76 | 0.01       |
| Decachlorobiphenyl   | 8.65       | 8.67      | 8.57      | 8.77 | 0.02       |