

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 15:31

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.33	6.33	6.23	6.43	0.00
Aroclor-1254-2	(2)	6.55	6.55	6.45	6.65	0.00
Aroclor-1254-3	(3)	6.91	6.91	6.81	7.01	0.00
Aroclor-1254-4	(4)	7.19	7.19	7.09	7.29	0.00
Aroclor-1254-5	(5)	7.61	7.61	7.51	7.71	0.00
Tetrachloro-m-xylene		4.35	4.36	4.26	4.46	0.01
Decachlorobiphenyl		9.94	9.93	9.83	10.03	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 15:31

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.53	5.54	5.44	5.64	0.01
Aroclor-1254-2 (2)	5.68	5.69	5.59	5.79	0.01
Aroclor-1254-3 (3)	6.08	6.09	5.99	6.19	0.01
Aroclor-1254-4 (4)	6.31	6.32	6.22	6.42	0.01
Aroclor-1254-5 (5)	6.73	6.74	6.64	6.84	0.01
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.01
Decachlorobiphenyl	8.65	8.67	8.57	8.77	0.02