

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 20:03

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1248-2 (2)	6.07	6.07	5.97	6.17	0.00
Aroclor-1248-3 (3)	6.27	6.27	6.17	6.37	0.00
Aroclor-1248-4 (4)	6.67	6.67	6.57	6.77	0.00
Aroclor-1248-5 (5)	6.71	6.71	6.61	6.81	0.01
Tetrachloro-m-xylene	4.64	4.65	4.55	4.75	0.01
Decachlorobiphenyl	10.41	10.42	10.32	10.52	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3437

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 20:03

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1248-2 (2)	5.10	5.10	5.00	5.20	0.00
Aroclor-1248-3 (3)	5.14	5.14	5.04	5.24	0.00
Aroclor-1248-4 (4)	5.31	5.31	5.21	5.41	0.00
Aroclor-1248-5 (5)	5.70	5.70	5.60	5.80	0.00
Tetrachloro-m-xylene	3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl	8.78	8.78	8.68	8.88	0.00